



Alberta Environment and Parks (AEP)
Air.Reporting@gov.ab.ca

February 22, 2018

Subject: Monthly Report Submission for the LICA Cold Lake South station

Lakeland Industry & Community Association (LICA) is pleased to submit the ambient air monitoring monthly report for the LICA Cold Lake South AQM Station in the month of March 2017.

The air monitoring program consists of continuous air monitoring, passive sampling and intermittent sampling, including both VOC, PAH and Partisol sampling programs. All the air monitoring activities were conducted by contractors.

Sampling Program	Monitoring Activities Conducted By	Sample Analysis Conducted By	Data/Report Review and Prepared By	Electronic Submission Conducted By
Continuous ambient air	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics
Passive	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics
Intermittent	Maxxam Analytics	InnoTech Alberta Inc	InnoTech Alberta Inc	Not Applicable

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

All data collected in March 2017 was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

As the LICA Environmental Program Manager and Data & Reporting Specialist, we certify that we have reviewed and verified this report and that the information is complete, accurate and representative of the monitoring results, reporting timeframe and the specified analysis, summarization and reporting requirements. We also certify all air data that are required by the AMD to be electronically submitted to AEP and Alberta's Ambient Air Quality Data Warehouse have been submitted by the time of this report submission.

Should you have any questions, please don't hesitate to contact us.

Respectfully,



Lakeland Industry & Community Association
5107 50 St
Bonnyville, AB T9N 2J7

A handwritten signature in blue ink that reads "Michael Bisaga".

Michael Bisaga
Technical Program Manager
Lakeland Industry & Community Association
780-266-7068
mbisaga@otonabee.ca

A handwritten signature in blue ink that reads "Lily Lin".

Lily Lin
Data & Reporting Specialist
587-225-2248
rebbacaa@gmail.com



MAXXAM ANALYTICS
#1 2080 39 Ave. NE, Calgary, AB
T2E 6P7

maxxam.ca
Toll Free 800-386-7247
Fax 403-219-3673

AMBIENT AIR MONITORING MONTHLY DATA REPORT
LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
COLD LAKE CONTINUOUS MONITORING STATION

JOB #: 2833-2017-03-1-C

March 2017

Prepared for:

LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
402 - 19 ST NW
CALGARY, ALBERTA
T2N 2J1

Attention: MIKE BISAGA

DATE: **May 10, 2017**

Prepared by: *Maram Ghaleb*

Maram Ghaleb, B.Sc.
Project Manager, Customer Service, Air Services

Reviewed by: *Wunmi Adekanmbi*

Wunmi Adekanmbi, M.Sc., EPT.
Project Manager, Customer Service, Air Services

SUMMARY

In March 2017, Maxxam Analytics was contracted to manage the ambient air quality monitoring and maintenance activities at the Cold Lake Continuous Monitoring Station, near Bonnyville, Alberta. The monitoring station provides continuous meteorological measurements and air quality data for non-compliance parameters, as requested by the Lakeland Industry & Community Association.

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

Annual Maintenance: A scheduled annual station maintenance was completed between March 6 and March 7 for the following analyzers: SO₂, TRS, THC, and O₃.

Two hours of downtime were recorded each for SO₂ and TRS, and nineteen hours were recorded for O₃ due to this maintenance event.

All Parameters: On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

O₃: A repeat span check was triggered on March 6 to address a biased high span drift. One hour of downtime was recorded due to the additional quality check. Annual maintenance was scheduled immediately. Following a shut-down calibration, the sample pump and the zero air pump were rebuilt and the optical cells were cleaned. The analyzer was allowed to stabilize overnight and a successful post-repair calibration was completed on March 7. Sixteen hours of downtime were attributed to the maintenance event.

THC: On March 6, the zero air system was replaced; and the channel was placed in "maintenance mode" for sample manifold cleaning. Three hours of downtime were incurred due to these events.

The analyzer recorded lower than historical concentrations between March 18 and March 19. Twenty two hours of suspect data were invalidated as a result.

NO_x/NO/NO₂: The NO_x gas concentration 50.7 ppm labelled as "Calculated NO_x" on the calibration record is not the actual concentration on the certificate of analysis, which is 50.9 ppm. A sample of affected calculations has been rerun and the error has no significant effect on the calibration. The NO_x calibration still meets the AMD calibration criteria.



PM_{2.5}: Eighteen hours of data were recorded at concentrations lower than $-3 \mu\text{g}/\text{m}^3$ this month, rendering the data invalid.

The summary of results is presented on the following pages.

Any deviations or modifications made to the sampling or analytical methods are outlined in Section 1.0, Discussion. On this basis, Maxxam Analytics is issuing this completed report to Lakeland Industry & Community Association, Cold Lake Continuous Monitoring Station.

Should you have any questions concerning the results or if we can be of further assistance, please contact us at 403-219-3677 or toll-free at 1-800-386-7247.

Pollutants		Monthly Records		1-Hour Records					24-Hour Records			
Name	Unit	Avg. Conc.	Uptime	Maximum			AAAQO Objective	Exceed. Hours	Maximum		AAAQO Objective	Exceed. Days
				Conc.	Date	Hour			Conc.	Date		
SO ₂	ppb	0.2	99.5%	1.3	March 15	13	172	0	0.5	March 7	48	0
TRS	ppb	0.3	99.5%	0.5	March 2	8	-	-	0.4	March 2	-	-
THC	ppm	2.12	96.4%	2.64	March 23	6	-	-	2.26	March 14	-	-
NO _x	ppb	4.0	99.7%	26.4	March 3	6	-	-	9.5	March 9	-	-
NO	ppb	0.6	99.7%	9.6	March 23	7	-	-	1.7	March 9	-	-
NO ₂	ppb	3.4	99.7%	23.7	March 3	6	159	0	7.9	March 9	-	-
O ₃	ppb	31.9	97.4%	48.3	March 17	15	82	0	39.8	March 17	-	-
PM _{2.5}	µg/m ³	3.7	97.3%	21.5	March 12	9	80	0	6.2	March 2	30	0
WS	%	2.3	99.7%	17.3	March 20	1	-	-	12.0	March 19	-	-
WD	degree	89 (E)	99.7%	-	-	-	-	-	-	-	-	-
RH	mm	70	99.7%	100	March 29	6	-	-	91	March 29	-	-
AmbTPX	°C	-6.4	99.7%	11.6	March 31	13	-	-	5.3	March 31	-	-



Monthly Update

- * All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.
- * All data collected this month were within the objectives outlined in the AMD 2016 and AAAQO 2016.
- * The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above 90%.

Operational Issues

Annual Maintenance: A scheduled annual station maintenance was completed between March 6 and March 7 for the following analyzers: SO₂, TRS, THC, and O₃. Two hours of downtime were recorded each for SO₂ and TRS, and nineteen hours were recorded for O₃ due to this maintenance event.

All Parameters: On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

O₃: A repeat span check was triggered on March 6 to address a biased high span drift. One hour of downtime was recorded due to the additional quality check. Annual maintenance was scheduled immediately. Following a shut-down calibration, the sample pump and the zero air pump were rebuilt and the optical cells were cleaned. The analyzer was allowed to stabilize overnight and a successful post-repair calibration was completed on March 7. Sixteen hours of downtime were attributed to the maintenance event.

THC: On March 6, the zero air system was replaced; and the channel was placed in "maintenance mode" for sample manifold cleaning. Three hours of downtime were incurred due to these events. The analyzer recorded lower than historical concentrations between March 18 and March 19. Twenty two hours of suspect data were invalidated as a result.

PM_{2.5}: Eighteen hours of data were recorded at concentrations lower than -3 µg/m³ this month, rendering the data invalid.

Monthly Continuous Data Summary

Lakeland Industry & Community Association Cold Lake Continuous Monitoring Station						MAXIMUM VALUES							OPERATIONAL TIME (%)
PARAMETER	OBJECTIVES		EXCEEDANCES		MONTHLY AVERAGE	READING	DAY	1-HOUR			24-HOUR		
	1-hr	24-hr	1-hr	24-hr				HOUR	WIND SPEED (kph)	WIND DIRECTION (sector)	READING	DAY	
SO ₂ (ppb)	172	48	0	0	0.2	1.3	15	13	2.2	SW	0.5	7	99.5
TRS (ppb)	-	-	-	-	0.3	0.5	2	8	3.2	E	0.4	2	99.5
THC (ppm)	-	-	-	-	2.12	2.64	23	6	0.6	WNW	2.26	14	96.4
NO ₂ (ppb)	159	-	0	-	3.4	23.7	3	6	2.6	E	7.9	9	99.7
NO (ppb)	-	-	-	-	0.6	9.6	23	7	1	NE	1.7	9	99.7
NO _x (ppb)	-	-	-	-	4.0	26.4	3	6	2.6	E	9.5	9	99.7
O ₃ (ppb)	82	-	0	-	31.9	48.3	17	15	7.4	SSW	39.8	17	97.4
PM _{2.5} (µg/m ³)	80	30	0	0	3.7	21.5	12	9	6.8	ESE	6.2	2	97.3
RELATIVE HUMIDITY (%)	-	-	-	-	70	100	29	6	2.3	ESE	91	29	99.7
AMBIENT TEMPERATURE (°C)	-	-	-	-	-6.4	11.6	31	13	8.5	WSW	5.3	31	99.7
VECTOR WS (kph)	-	-	-	-	2.3	17.3	20	1	-	NNW	12.0	19	99.7
VECTOR WD (sec)	-	-	-	-	89 (E)	-	-	-	-	-	-	-	99.7

Exceedance Summary Report

SO₂ 1-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 1-hour AAAQO of 172 ppb.

SO₂ 24-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 24-hour AAAQO of 48.0 ppb.

NO₂ 1-Hour Exceedances

Measured concentrations of nitrogen dioxide were below the 1-hour AAAQO of 159 ppb.

PM_{2.5} 1-Hour Exceedances

Measured concentrations of fine particulate matter were below the 1-hour AAAQO of 80 µg/m³.

PM_{2.5} 24-Hour Exceedances

Measured concentrations of fine particulate matter were below the 24-hour AAAQO of 30 µg/m³.

O₃ 1-Hour Exceedances

Measured concentrations of ozone were below the 1-hour AAAQO of 82 ppb.

In accordance with EPEA and the Substance Release Regulation.

In accordance with A Guide to Release Reporting and the Alberta Ambient Air Quality Objectives and Guidelines Summary.

Passive Sampler Summary

	Sulphur Dioxide (ppb)
Mean	0.5
Minimum	0.2
Maximum	1.3

Note: Access papers for stations #12 and #25 were not provided and access to the station #11 was blocked by snow. As a result, data is not available for these stations.

	Hydrogen Sulphide (ppb)
Mean	0.13
Minimum	0.08
Maximum	0.26

Note: Access papers for stations #12 and #25 were not provided and access to the station #11 was blocked by snow. As a result, data is not available for these stations.

	Nitrogen Dioxide (ppb)
Mean	1.0
Minimum	0.4
Maximum	3.0

Note: Access papers for stations #12 were not provided and access to the station #11 was blocked by snow. As a result, data is not available for these stations.

	Ozone (ppb)
Mean	34.1
Minimum	25.9
Maximum	43.4

Note: Access papers for stations #12 were not provided and access to the station #11 was blocked by snow. As a result, data is not available for these stations.

Volatile Organics (VOCs) Data Summary

Sample Collection Date	Maximum Reading (ppb)	Volatile Organic Compound
March 2, 2017	1.6	Acetone
March 8, 2017	2.0	Naphthalene
March 14, 2017	2.75	n-Butane
March 20, 2017	2.7	Acetone
March 26, 2017	3.5	Acetone

Note: NA

Polycyclic Aromatic Hydrocarbons (PAHs) Data Summary

Sample Collection Date	Maximum Reading ($\mu\text{g}/\text{puf}$)	Semi-Volatile Organic
March 2, 2017	0.33	Naphthalene
March 8, 2017	1.12	Naphthalene
March 14, 2017	0.25	Phenanthrene
March 20, 2017	0.14	Naphthalene
March 26, 2017	0.84	Phenanthrene

Note: NA

Partisol Sampler Summary

Sample Collection Date	Concentration ($\mu\text{g}/\text{puf}$)
March 2, 2017	0.110
March 8 2017	0.037
March 14, 2017	0.123
March 20, 2017	0.033
March 26, 2017	0.126

Note: NA

TABLE OF CONTENTS

<u>Title</u>	<u>Page</u>
SUMMARY	1
MONTHLY DASHBOARD	2
MONTHLY CONTINUOUS DATA SUMMARY REPORT	3
EXCEEDANCE SUMMARY REPORT	4
PASSIVE SAMPLE SUMMARY REPORT	5
VOC SAMPLE SUMMARY REPORT	6
PAH SAMPLE SUMMARY REPORT	7
PARTISOL SAMPLE SUMMARY REPORT	8
TABLE OF CONTENTS	9
1.0 Discussion	11
2.0 Project Personnel	15
3.0 Plant Monthly Required AMD Summary	15
4.0 Calculations and Results	15
5.0 Methods and Procedures	16
Appendix I	19
Continuous Monitoring Data Results	19
Sulphur Dioxide	20
Total Reduced Sulphur	28
Total Hydrocarbon	36
Oxides of Nitrogen	44
Nitric Oxides	52
Nitrogen Dioxide	59
Ozone	67
Particulate Matter 2.5	75
Wind Speed	80
Wind Direction	87
Standard Deviation Wind Direction	90
Relative Humidity	93
Ambient Temperature	96

Appendix II	Non-Continuous Monitoring Data Results	99
	Passive Results	100
	VOC Results	110
	PAH Results	125
	Partisol Results	136
Appendix III	Equipment Calibration Results	143
	Sulphur Dioxide	144
	Total Reduced Sulphur	148
	Total Hydrocarbon	152
	Nitrogen Dioxide	156
	Ozone	160
	Particulate Matter	165
	Wind System	168
	VOC Sampler	170
	Calibrators	172
	Calibration Gases	175
Appendix IV	Analytical Results	180
	Passive Samples	181
	VOCs Samples	188
	PAHs Samples	214
	Partisol Samples	225
Appendix V	Report Certification Form	231
Appendix VI	Data Validation Certification Form	233

1.0 Discussion

This monthly report consists of continuous monitoring results for the following parameters: Sulphur Dioxide (SO₂), Total Reduced Sulphur (TRS), Total Hydrocarbon (THC), Oxides of Nitrogen (NO_x), Nitric Oxides (NO), Nitrogen Dioxide (NO₂), Ozone (O₃), Particulate Matter 2.5 (PM_{2.5}), Relative Humidity (RH), Ambient Temperature (AmbTPX), Wind Speed (WS), Wind Direction (WD) and Standard Deviation Wind Direction (STDWD). The non-continuous monitoring data results for Partisol, VOCs, PAHs and Passive samples are also included in this report.

Sample filters for all continuous air monitors are changed before the calibration begins. The sample manifold is cleaned during the site visit each month.

Control checks, consisting of a zero and span, are conducted daily on all continuous air monitors. In place of the air sample, zero air (from scrubbed air or gas cylinders) is used for zero checks, and a known concentration of the pollutant being analyzed is used for span checks. These checks are controlled by automatic timers and valves. The total zero span cycle is completed within an hour, the commencement of the zero span cycle is at the beginning of the hour.

Multipoint calibrations are done a minimum of once a month for each continuous air monitor. An additional calibration is required under the following conditions: 1) within three days after the initial start-up and stabilization of a newly installed instrument, 2) prior to shut-down or moving of an instrument which has been working to specification, and 3) when major repair has been done on the instrument.

Time during the first multi-point calibration is not considered downtime (Data is flagged as C). If more than one calibration is performed during the month, the time during the additional calibration is considered as downtime (Data is flagged as C1).

Only one zero/span check is run per day. Time during the zero/span check is not considered as downtime (Data is flagged as S). If an extra zero/span check is performed, the time during the additional check is considered as downtime (Data is flagged as S1).

The AMD requires each instrument and accompanying data recording system to be operational 90% of the time, at a minimum, for each monthly monitoring period.

All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.

Data contained in this monthly report has undergone the verification and validation based on the requirements of the AMD Chapter 6: Ambient Data Quality (December, 2016). The descriptions of the data verification and validation process can be found in Section 5 of this report. Instantaneous data, where applicable, is provided for reference purposes and has not undergone zero correction. The minimum and maximum statistics are highlighted in the data table and are for reference only. The highlighted cells are based on the software's interpretation of the exact position of the minimum or maximum value. The visual presentation of these statistics may not be the obvious choice in a data range due to rounding, truncating or analyzer specifications.

Hourly/minute data have been reviewed based on daily zero/span results and multi-point calibration results. Data may be considered invalid if a zero-corrected span check in excess of +/- 10% of the span concentration (established by the previous multi-point calibration) is encountered and/or significant differences in the calibration factor occurs (greater than 10%).

SULPHUR DIOXIDE (SO₂)

- Operational time, for the monitoring period was 99.5%, equivalent to four hours of downtime.
- A shut-down calibration was performed on March 6, prior to completing annual maintenance on the analyzer. The sample pump was rebuilt and a leak check was performed. A successful post-repair calibration was completed afterwards. Two hours of downtime were attributed to the maintenance event.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- One hour of maximum instantaneous data was invalidated on March 31, at hour 7:00, due to a brief power failure.

TOTAL REDUCED SULPHUR (TRS)

- Operational time, for the monitoring period was 99.5%, equivalent to four hours of downtime.
- A shut-down calibration was performed on March 6, prior to completing annual maintenance on the analyzer. The sample pump was rebuilt, a leak check was performed and the converter tubing was renewed. A successful post-repair calibration was completed afterwards. Two hours of downtime were attributed to the maintenance event.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- One hour of maximum instantaneous data was invalidated on March 31, at hour 7:00, due to a brief power failure.

TOTAL HYDROCARBONS (THC)

- Operational time, for the monitoring period was 96.4%, equivalent to twenty-seven hours of downtime.
- On March 6, the zero air system was replaced; and the channel was placed in "maintenance mode" for sample manifold cleaning. Three hours of downtime were incurred due to these events.
- A shut-down calibration was performed on March 7, prior to completing annual maintenance on the analyzer. The hydrogen fuel tubing was renewed and the fuel pressure was adjusted. LICA's Zero Air generator (s/n: 4027) was installed back after it was repaired and Maxxam's Zero Air (s/n: 133) was removed. A successful post-repair calibration was completed afterwards.
- The analyzer recorded lower than historical concentrations between March 18 and March 19. Twenty two hours of suspect data were invalidated as a result.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- One hour of maximum instantaneous data was invalidated on March 31, at hour 7:00, due to a brief power failure.

OXIDES OF NITROGEN (NO_x), NITRIC OXIDE (NO) and NITROGEN DIOXIDE (NO₂)

- Operational time, for the monitoring period was 99.7%, equivalent to two hours of downtime.
- The routine monthly calibration was performed on March 6.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- One hour of maximum instantaneous data was invalidated on March 31, at hour 7:00, due to a brief power failure.
- The NO_x gas concentration 50.7 ppm labelled as "Calculated NO_x" on the calibration record is not the actual concentration on the certificate of analysis, which is 50.9 ppm. A sample of affected calculations has been rerun and the error has no significant effect on the calibration. The NO_x calibration still meets the AMD calibration criteria.

OZONE (O₃)

- Operational time, for the monitoring period was 97.4%, equivalent to nineteen hours of downtime.
- The analyzer spanned close to the upper acceptance limit on March 5. A repeat span check was triggered on March 6 and the result confirmed the drift. Annual maintenance was scheduled immediately. Following a shut-down calibration, the sample pump and the zero air pump were rebuilt and the optical cells were cleaned. The analyzer was allowed to stabilize overnight and a successful post-repair calibration was completed on March 7. As the analyzer passed a shutdown calibration, no data were discarded. Seventeen hours of downtime were attributed to the maintenance event and the repeat span check.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- One hour of maximum instantaneous data was invalidated on March 31, at hour 7:00, due to a brief power failure.

PARTICULATE MATTER < 2.5 MICRONS (PM_{2.5})

- Operational time, for the monitoring period was 97.3%, equivalent to twenty hours of downtime.
- Two routine TEOM audits were performed this month. The first was completed on March 7, and the second on March 27.
- Data was corrected in accordance with AMD (2016), Chapter 6, Table 2, *Zero Adjustment Criteria*. Data recorded between 0 and $-3 \mu\text{g}/\text{m}^3$ was corrected to $0 \mu\text{g}/\text{m}^3$. Data recorded below $-3 \mu\text{g}/\text{m}^3$ was invalidated. Eighteen hours of data were invalidated as the data was below $-3 \mu\text{g}/\text{m}^3$ this month.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

WIND SPEED (WS), WIND DIRECTION (WD) and STANDARD DEVIATION WIND DIRECTION (STDWD)

- Operational time, for the monitoring period was 99.7%, equivalent to two hours of downtime.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- Wind data is reported as vector wind speed and vector wind direction. Wind direction is defined as the direction from which the wind is blowing from and is measured in degrees from true north.

RELATIVE HUMIDITY (RH)

- Operational time, for the monitoring period was 99.7%, equivalent to two hours of downtime.
- A function check was performed on the relative humidity sensor on March 6. The sensor was determined to operate properly.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

AMBIENT TEMPERATURE (AmbTPX)

- Operational time, for the monitoring period was 99.7%, equivalent to two hours of downtime.
- A function check was performed on the temperature sensor on March 6. The sensor was determined to operate properly.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

VOC SAMPLES

- The sampler was programmed to run for 24 hours every 6th day per the NAPS (North American Pollution Surveillance Schedule).
- Samples were collected, as scheduled, on March 2, 8, 14, 20 and 26. Analysis was provided by InnoTech Alberta, results are included in this report.
- The routine quarterly audit for the VOC sampler was completed on March 6.

PAH SAMPLES

- The sampler was programmed to run for 24 hours every 6th day per the NAPS (North American Pollution Surveillance Schedule).
- Samples were collected, as scheduled, on March 2, 8, 14, 20 and 26. Analysis was provided by InnoTech Alberta, results are included in this report.

PARTISOL SAMPLES

- The sampler was programmed to run for 24 hours every 6th day per the NAPS (North American Pollution Surveillance Schedule).
- Samples were collected, as scheduled, on March 2, 8, 14, 20 and 26. Analysis was provided by InnoTech Alberta, results are included in this report.

PASSIVE SAMPLES

- Samples were collected over the months of February and March. Samples were collected at all designated stations, except stations #12 and #25 as access documents were not provided by client. A sample was not collected at station #11 as the access to the station was blocked by snow. Analytical results are included in this report. Passive results are reported in ppb.

2.0 Project Personnel

Mike Bisaga was the contact for Lakeland Industry & Community Association and the Maxxam field technicians were Alexander Yakupov.

3.0 Plant Monthly Required AMD Summary

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

4.0 Calculations and Results

All calculations and reporting of results follow the methods described in the AMD, 2016.

5.0 Methods and Procedures

The following methods and procedures were used to complete the monitoring program:

- Maxxam AIR SOP-00007: TISCH PUF Sampler Operating, Calibration and Maintenance Procedures
- Maxxam AIR SOP-00209: Ambient Sulphur Monitoring
- Maxxam AIR SOP-00212: Ambient O₃ Monitoring
- Maxxam AIR SOP-00213: Ambient NO/NO₂/NO_x Monitoring
- Maxxam AIR SOP-00214: Ambient Hydrocarbon (THC) Monitoring
- Maxxam AIR SOP-00215: TEOM Operation
- Maxxam AIR SOP-00225: The Collection of VOCs in Ambient Air Using Canister and Xontech
- Maxxam PTC SOP-00148: Monitoring NO₂ in the Atm. by using All-Season Passive
- Maxxam PTC SOP-00149: Monitoring SO₂ in the Atm. by using All-Season Passive
- Maxxam PTC SOP-00150: Monitoring H₂S in the Atm. by using All-Season Passive
- Maxxam PTC SOP-00151: Mass Determination of Particulate Matter (PM_{2.5} and PM₁₀)
- Maxxam PTC SOP-00197: Monitoring O₃ in the Atm. by Using Maxxam All-Season Passive

There were no deviations from the prescribed methods.

The following instruments were used to perform the test program:

- Sulphur Dioxide - Thermo 43i UV Fluorescent Analyzer
- Total Reduced Sulphur - Thermo 450i UV Fluorescent Analyzer
- Total Hydrocarbons - Thermo 51C FID Analyzer
- Oxides of Nitrogen - Thermo 42i Chemiluminescent Analyzer
- Ozone - Thermo 49i Photometric Analyzer
- Particulate Matter (PM_{2.5}) - R&P 1405F TEOM Unit
- Wind System - Met One Unit
- Relative Humidity - Met One Unit
- Ambient Temperature - Met One Unit
- PAH - TISCH PUF Plus
- VOC - XONTECH 910A Gaseous Air Sampler
- Datalogger - ESC 8832
- Partisol - R&P 2000H Unit

The following steps were used to complete the data verification and validation process:

Level 0 Preliminary Verification

Level 0 data are raw data obtained directly from the data acquisition system (DAS). Under the step of Level 0, these data undergo a certain amount of manual or automated screening and flagging. It included a) identification of periods of missing data; b) verification of time stamps against reference time; c) verification that instrument diagnostics/datalogger flags indicate normal operation; d) comparison of data to upper and lower limits; e) rate of change flagging indicating that data changed too rapidly or not at all; and f) verification that zero, span and multipoint performance checks are within specifications. This level of verification is performed on a daily basis.

Level 1 Primary Validation

Validation actions under the step of Level 1 include a) review of all screening flags assigned during preliminary verification; b) review of all supporting site information and documentation; c) review of operational acceptance limits for each parameter/analyzer; d) review of daily zero/span and monthly calibration results for all gaseous parameters; and e) application of any necessary adjustments to data (e.g. baseline adjustments, below zero adjustments). This level of validation is performed on a monthly basis.

Level 2 Final Validation


The purpose of Level 2 validation is to verify that there are no inconsistencies among related data, or among regional data measured at nearby sites.

Level 3 Independent Data Review

Level 3 validation is the last step of data review, and it is completed by an individual that is independent of both field operations and primary data validation. A final independent QA review and endorsement is performed during this step before data is submitted to Alberta Environment.

Post-Final Validation

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. Any data issues or patterns which were not clear on a monthly basis are highlighted during this step. This validation is performed on an annual basis.



Source: Air Monitoring Directive (December 2016), Chapter 6, Ambient Data Quality; Figure 1 Data Collection and Management Process Flow Chart

APPENDIX I
CONTINUOUS MONITORING DATA RESULTS

SULPHUR DIOXIDE

SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.		
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.			
DAY																														
1	0.1	0.0	0.0	S	0.1	0.1	0.1	0.0	0.1	0.0	0.1	0.1	0.0	0.0	0.1	0.0	0.0	0.1	0.1	0.1	0.1	0.0	0.1	0.1	0.0	0.1	0.1	0.0	0.1	24
2	0.1	0.1	S	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.1	0.1	0.1	0.3	0.6	0.5	0.7	0.7	0.5	0.4	0.3	0.3	0.1	0.7	0.3	24
3	0.2	S	0.1	0.0	0.1	0.1	0.1	0.0	0.0	0.2	0.2	0.3	0.3	0.5	0.6	0.4	0.4	0.3	0.2	0.1	0.0	0.1	0.3	0.0	0.0	0.6	0.6	0.2	24	
4	S	0.0	0.1	0.6	0.4	0.0	0.1	0.0	0.1	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.0	0.1	0.1	0.0	0.1	0.1	S	0.0	0.6	0.1	24		
5	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.0	0.1	0.1	0.0	0.1	0.1	0.1	0.1	0.0	0.1	0.0	0.1	S	0.1	0.0	0.1	0.1	24	
6	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.1	C	C	C	Y	Y	C	C	C	C	0.2	0.1	S	0.2	0.2	0.1	0.2	0.1	22		
7	0.1	0.1	0.2	0.1	0.1	0.2	0.4	0.3	0.4	0.5	0.5	0.6	0.7	0.4	0.5	0.7	0.9	1.0	1.0	0.9	S	0.7	0.6	0.4	0.1	1.0	0.5	24		
8	0.3	0.2	0.2	0.2	0.1	0.1	0.2	0.1	0.1	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	S	0.1	0.1	0.1	0.1	0.3	0.1	24		
9	0.1	0.2	0.2	0.1	0.0	0.0	0.0	0.0	0.1	0.1	0.1	0.1	0.2	0.1	0.3	0.3	0.5	0.4	S	0.1	0.1	0.0	0.0	0.0	0.0	0.5	0.1	24		
10	0.0	0.1	0.0	0.1	0.1	0.2	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.3	0.2	0.3	S	0.2	0.2	0.2	0.2	0.1	0.2	0.2	0.0	0.3	0.2	24		
11	0.1	0.1	0.1	0.1	0.1	0.1	0.2	X	X	0.1	0.1	0.1	0.1	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.1	22		
12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.3	0.2	0.0	0.3	0.0	24	
13	0.1	0.1	0.1	0.1	0.1	0.2	0.1	0.1	0.3	0.8	1.2	1.1	0.7	0.5	S	0.5	0.6	0.7	1.0	0.9	0.5	0.4	0.4	0.4	0.1	1.2	0.5	24		
14	0.3	0.2	0.3	0.3	0.3	0.4	0.3	0.4	0.5	0.7	0.5	0.7	1.0	S	0.7	0.8	0.8	0.9	0.7	0.6	0.5	0.4	0.3	0.3	0.2	1.0	0.5	24		
15	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.4	0.3	0.5	S	1.3	0.9	0.7	0.5	0.3	0.2	0.1	0.2	0.2	0.2	0.2	0.1	1.3	0.3	24		
16	0.3	0.2	0.2	0.1	0.1	0.1	0.1	0.0	0.0	0.1	0.1	S	0.0	0.1	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.1	0.0	0.1	0.0	0.3	0.1	24		
17	0.1	0.1	0.1	0.1	0.0	0.1	0.1	0.1	0.2	0.2	S	0.2	0.2	0.2	0.4	0.4	0.3	0.3	0.4	0.3	0.3	0.3	0.1	0.1	0.0	0.4	0.2	24		
18	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.0	0.1	S	0.2	0.2	0.1	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.0	0.4	0.0	0.4	0.1	24		
19	0.2	0.1	0.1	0.1	0.1	0.0	0.1	0.1	S	0.1	0.1	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.0	0.0	0.1	0.1	0.0	0.6	0.0	0.6	0.1	24		
20	0.3	0.3	0.5	0.2	0.4	0.5	0.5	S	0.9	0.6	0.6	0.4	0.4	0.3	0.3	0.3	0.2	0.3	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.9	0.3	24		
21	0.0	0.0	0.0	0.0	0.0	0.0	S	0.1	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.1	0.2	0.2	0.1	0.1	0.1	0.1	0.2	0.1	0.0	0.3	0.1	24		
22	0.2	0.3	0.3	0.3	0.2	S	0.1	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.0	0.1	0.2	0.2	0.2	0.1	0.0	0.3	0.2	24		
23	0.0	0.1	0.0	0.0	S	0.1	0.1	0.2	0.4	0.2	0.1	0.1	0.1	0.0	0.0	0.1	0.0	0.1	0.2	0.3	0.1	0.1	0.0	0.0	0.0	0.4	0.1	24		
24	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.0	0.1	0.0	24		
25	0.1	0.0	S	0.1	0.0	0.1	0.1	0.1	0.1	0.2	1.3	0.9	0.4	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.0	1.3	0.2	24		
26	0.0	S	0.1	0.0	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.4	0.5	0.2	0.2	0.2	0.2	0.2	0.0	0.5	0.2	24		
27	S	0.1	0.1	0.1	0.1	0.1	0.1	0.0	0.0	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.1	0.0	24		
28	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	24		
29	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	24		
30	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	24		
31	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.3	0.4	0.3	0.5	0.2	0.0	0.2	0.0	0.0	S	0.0	0.0	0.2	0.1	0.0	0.5	0.1	24		
HOURLY MAX	0.3	0.3	0.5	0.6	0.4	0.5	0.5	0.4	0.9	0.8	1.3	1.1	1.0	1.3	0.9	0.8	0.9	1.0	1.0	0.9	0.5	0.7	0.6	0.6						
HOURLY AVG	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.2						

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 172 ppb | 24-HR 48 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0				
NUMBER OF 24-HR EXCEEDANCES:	0				
NUMBER OF NON-ZERO READINGS:	483				
MINIMUM 1-HR AVERAGE:	0.0 ppb	@ HOUR(S)	1	ON DAY(S)	1
MAXIMUM 1-HR AVERAGE:	1.3 ppb	@ HOUR(S)	13	ON DAY(S)	15
MAXIMUM 24-HR AVERAGE:	0.5 ppb			ON DAY(S)	7
				VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	740 hrs		
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	99.5 %		
STANDARD DEVIATION:	0.2	MONTHLY AVERAGE:	0.2 ppb		

24 HR AVERAGES March 2017



SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake Continuous Monitoring Station - March 2017

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	0.5	0.5	0.5	S	0.6	0.6	0.5	0.5	0.6	0.3	0.5	0.5	0.5	0.3	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.6	0.3	0.6	0.5	24	
2	0.5	0.6	S	0.5	0.5	0.4	0.5	0.5	0.6	0.5	0.6	0.7	0.6	0.6	0.6	0.8	1.1	1.1	1.1	1.1	0.9	1.1	0.7	0.9	0.4	1.1	0.7	24	
3	0.6	S	0.6	0.6	0.5	0.6	0.6	0.5	0.5	0.6	0.8	0.8	0.8	1.2	1.1	0.9	0.9	0.9	0.6	0.6	0.6	0.8	1.1	0.6	0.5	1.2	0.7	24	
4	S	0.6	0.6	1.4	0.9	0.5	0.6	0.6	0.6	0.6	0.6	0.6	0.6	0.6	0.5	0.6	0.5	0.5	0.5	0.5	0.5	0.5	0.5	S	0.5	1.4	0.6	24	
5	0.6	0.5	0.5	0.4	0.5	0.5	0.5	0.5	0.5	0.5	0.3	0.5	0.5	0.5	0.5	0.6	0.6	0.5	0.5	0.3	0.5	0.3	S	0.5	0.3	0.6	0.5	24	
6	0.5	0.5	0.5	0.5	0.5	0.5	0.3	0.5	0.5	0.5	C	C	C	Y	Y	C	C	C	C	0.5	0.5	S	0.6	0.5	0.3	0.6	0.5	22	
7	0.5	0.5	0.3	0.3	0.3	0.5	0.6	0.6	0.6	0.6	0.6	0.8	0.9	0.6	0.7	0.9	1.1	1.1	1.1	1.4	S	0.8	0.7	0.6	0.3	1.4	0.7	24	
8	0.5	0.5	0.3	0.3	0.3	0.3	0.5	0.5	0.3	0.4	0.4	0.3	0.3	0.3	0.3	0.4	0.3	0.5	0.3	S	0.3	0.5	0.3	0.3	0.3	0.5	0.4	24	
9	0.3	0.5	0.5	0.5	0.3	0.3	0.3	0.3	0.4	0.4	0.4	0.6	0.6	0.4	0.5	0.8	0.9	0.8	S	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.9	0.4	24
10	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.6	0.3	0.4	0.3	0.5	0.3	0.5	0.3	0.3	S	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.6	0.3	24
11	0.3	0.2	0.2	0.3	0.3	0.5	0.3	X	X	0.3	0.5	0.5	0.3	0.2	0.3	0.5	S	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.5	0.2	0.5	0.3	22
12	0.5	0.3	0.3	0.3	0.5	0.5	0.3	0.3	0.5	0.5	0.5	0.3	0.2	0.3	0.4	S	0.3	0.5	0.3	0.3	0.5	0.3	0.6	0.6	0.2	0.6	0.4	24	
13	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.7	1.2	1.5	1.4	1.1	0.9	S	0.9	0.9	1.1	1.2	1.4	0.7	0.6	0.7	0.6	0.5	1.5	0.8	24	
14	0.5	0.5	0.5	0.6	0.6	0.6	0.5	0.5	0.6	0.9	0.6	0.8	1.1	S	0.9	0.9	0.9	0.9	0.8	0.6	0.6	0.6	0.6	0.5	0.5	1.1	0.7	24	
15	0.5	0.3	0.5	0.3	0.5	0.5	0.3	0.5	0.5	0.5	0.6	1.2	S	1.7	1.5	1.1	1.1	0.5	0.5	0.3	0.5	0.5	0.5	0.5	0.3	1.7	0.6	24	
16	0.5	0.5	0.5	0.4	0.3	0.4	0.5	0.3	0.3	0.4	0.5	S	0.4	0.3	0.5	0.5	0.5	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.5	0.4	24
17	0.3	0.5	0.3	0.3	0.3	0.3	0.3	0.3	0.4	0.3	S	0.5	0.4	0.5	0.6	0.7	0.6	0.7	0.6	0.5	0.5	0.4	0.3	0.3	0.3	0.7	0.4	24	
18	0.3	0.5	0.3	0.3	0.3	0.3	0.5	0.2	0.3	S	0.5	0.5	0.3	0.6	0.3	0.3	0.3	0.5	0.1	0.3	0.3	0.5	0.3	0.9	0.1	0.9	0.4	24	
19	0.4	0.2	0.3	0.3	0.3	0.2	0.2	0.3	S	0.3	0.3	0.5	0.3	0.3	0.3	0.4	0.5	0.3	0.3	0.1	0.3	0.3	0.3	1.2	0.1	1.2	0.3	24	
20	1.1	0.6	0.9	0.6	0.6	1.1	0.9	S	1.4	0.9	1.1	0.7	0.7	0.6	0.7	0.6	0.6	0.6	0.6	0.5	0.3	0.5	0.5	0.3	0.3	1.4	0.7	24	
21	0.3	0.3	0.5	0.5	0.3	0.3	S	0.6	0.9	0.7	0.6	0.5	0.5	0.5	0.4	0.5	0.6	0.5	0.3	0.5	0.5	0.3	0.5	0.3	0.3	0.9	0.5	24	
22	0.5	0.5	0.6	0.5	0.5	S	0.3	0.3	0.5	0.5	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.5	0.5	0.5	0.3	0.3	0.6	0.4	24
23	0.3	0.3	0.2	0.3	S	0.3	0.3	0.6	0.9	0.6	0.3	0.5	0.5	0.5	0.3	0.5	0.5	0.5	0.6	0.7	0.5	0.5	0.5	0.3	0.2	0.9	0.5	24	
24	0.3	0.3	0.5	S	0.3	0.3	0.3	0.3	0.3	0.5	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.3	0.2	0.3	0.3	0.2	0.5	0.3	24
25	0.3	0.2	S	0.3	0.2	0.3	0.3	0.3	0.3	0.6	2.3	1.8	0.8	0.5	0.5	0.3	0.4	0.3	0.2	0.3	0.2	0.3	0.3	0.3	0.2	2.3	0.5	24	
26	0.3	S	0.3	0.3	0.3	0.3	0.5	0.3	0.3	0.4	0.5	0.3	0.3	0.5	0.5	0.5	0.6	0.7	0.5	0.3	0.3	0.3	0.3	0.3	0.3	0.7	0.4	24	
27	S	0.5	0.3	0.3	0.4	0.3	0.3	0.3	0.3	0.3	0.3	0.4	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.3	S	0.2	0.5	0.3	24	
28	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	S	0.3	0.2	0.3	0.3	24	
29	0.3	0.3	0.3	0.3	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.5	0.3	S	0.3	0.3	0.2	0.5	0.3	24	
30	0.1	0.3	0.3	0.3	0.3	0.3	0.5	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.3	0.1	0.1	0.3	0.3	0.2	S	0.3	0.3	0.1	0.5	0.3	24	
31	0.4	0.3	0.3	0.2	0.3	0.5	0.3	P	0.3	0.6	0.7	0.7	0.9	0.9	1.1	0.5	0.6	0.3	0.3	S	0.3	0.3	0.6	0.5	0.2	1.1	0.5	23	
HOURLY MAX	1.1	0.6	0.9	1.4	0.9	1.1	0.9	0.6	1.4	1.2	2.3	1.8	1.1	1.7	1.5	1.1	1.1	1.1	1.2	1.4	0.9	1.1	1.1	1.2					
HOURLY AVG	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.5	0.5	0.6	0.6	0.5	0.5	0.5	0.5	0.6	0.5	0.5	0.5	0.4	0.4	0.5	0.5					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	699
MAXIMUM INSTANTANEOUS VALUE:	2.3 ppb @ HOUR(S) 10 ON DAY(S) 25
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	7 hrs
OPERATIONAL TIME:	739 hrs
STANDARD DEVIATION:	0.2

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)



Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-SO2[ppb]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 12.16%

Calm Avg: 0.07 [ppb]


Direction	0.0-0.4	0.4-0.8	0.8-1.1	1.1-1.5	1.5-1.9	>1.9	Total
N	4.0	1.0	0.6	0.0	0.0	0.0	5.6
NE	15.9	0.3	0.3	0.0	0.0	0.0	16.5
E	21.2	0.6	0.1	0.0	0.0	0.0	21.9
SE	23.2	3.2	0.7	0.1	0.0	0.0	27.2
S	0.6	0.4	0.0	0.0	0.0	0.0	1.0
SW	3.7	0.6	0.0	0.1	0.0	0.0	4.4
W	5.3	0.6	0.3	0.1	0.0	0.0	6.3
NW	3.2	1.7	0.1	0.0	0.0	0.0	5.0
Summary	77.0	8.3	2.2	0.4	0.0	0.0	87.8

% Icon Classes (ppb) 77 0.0-0.4 8 0.4-0.8 2 0.8-1.1 0 1.1-1.5 0 1.5-1.9 0 >1.9

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-SO2[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 12.16% Calm Poll Avg: 0.07[ppb]



SO2[ppb] Calibration: LICA COLD LAKE SOUTH Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

TOTAL REDUCED SULPHUR


TOTAL REDUCED SULPHUR Hourly Averages (TRS ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.																						
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.																							
DAY																																																		
1	0.2	0.2	0.2	S	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.4	0.3	0.3	0.4	0.4	0.4	0.4	0.2	0.4	0.3	24																						
2	0.4	0.4	S	0.4	0.4	0.4	0.4	0.4	0.5	0.4	0.3	0.4	0.4	0.3	0.3	0.3	0.3	0.4	0.4	0.4	0.3	0.3	0.3	0.3	0.3	0.5	0.4	24																						
3	0.4	S	0.3	0.2	0.3	0.4	0.4	0.4	0.4	0.4	0.3	0.3	0.3	0.4	0.3	0.3	0.3	0.3	0.3	0.4	0.3	0.3	0.4	0.3	0.2	0.4	0.3	24																						
4	S	0.3	0.3	0.4	0.4	0.4	0.3	0.3	0.3	0.2	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	S	0.2	0.4	0.3	24																						
5	0.2	0.2	0.2	0.3	0.2	0.2	0.2	0.3	0.3	0.3	0.2	0.3	0.2	0.2	0.3	0.3	0.4	0.3	0.3	0.3	0.3	0.3	S	0.3	0.2	0.4	0.3	24																						
6	0.4	0.3	0.3	0.3	0.3	0.3	0.3	0.4	0.3	0.3	C	C	C	Y	Y	C	C	C	C	0.3	0.4	S	0.3	0.3	0.3	0.4	0.3	22																						
7	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.3	0.2	0.2	0.3	0.3	0.3	S	0.2	0.3	0.3	0.2	0.3	0.3	24																						
8	0.3	0.2	0.2	0.2	0.3	0.3	0.3	0.2	0.3	0.2	0.2	0.2	0.3	0.2	0.2	0.2	0.3	0.3	0.3	S	0.3	0.3	0.3	0.2	0.2	0.3	0.3	24																						
9	0.2	0.2	0.3	0.3	0.2	0.2	0.3	0.2	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	S	0.2	0.2	0.3	0.3	0.3	0.2	0.3	0.2	24																						
10	0.3	0.3	0.3	0.2	0.3	0.3	0.3	0.4	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.3	S	0.3	0.4	0.3	0.4	0.3	0.3	0.2	0.4	0.3	24																						
11	0.3	0.3	0.3	0.3	0.3	0.3	X	X	0.3	0.3	0.3	0.3	0.3	0.3	0.3	S	0.2	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.3	0.3	22																						
12	0.4	0.3	0.3	0.3	0.3	0.2	0.3	0.3	0.3	0.3	0.3	0.4	0.3	0.3	0.3	S	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.4	0.3	24																						
13	0.3	0.3	0.3	0.3	0.3	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	S	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.3	0.2	0.3	24																						
14	0.3	0.4	0.3	0.3	0.3	0.3	0.4	0.3	0.3	0.3	0.3	0.3	0.2	S	0.3	0.3	0.3	0.3	0.4	0.3	0.3	0.3	0.3	0.2	0.3	0.2	0.4	0.3	24																					
15	0.2	0.3	0.3	0.3	0.3	0.2	0.3	0.3	0.3	0.2	0.3	0.3	S	0.3	0.3	0.3	0.3	0.3	0.2	0.3	0.2	0.3	0.3	0.2	0.2	0.3	0.3	24																						
16	0.3	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	S	0.3	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.3	24																						
17	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.2	0.2	0.2	S	0.2	0.2	0.2	0.3	0.3	0.3	0.2	0.3	0.3	0.3	0.3	0.3	0.2	0.3	0.2	0.3	24																						
18	0.3	0.3	0.3	0.3	0.2	0.3	0.3	0.3	0.2	S	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.2	0.3	0.2	0.1	0.3	0.2	24																						
19	0.2	0.2	0.1	0.3	0.2	0.1	0.1	0.1	S	0.1	0.0	0.1	0.1	0.1	0.1	0.1	0.2	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.0	0.3	0.1	24																						
20	0.2	0.2	0.2	0.2	0.2	0.2	0.2	S	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.2	0.2	0.2	0.2	0.3	24																						
21	0.2	0.2	0.3	0.3	0.2	0.2	S	0.3	0.3	0.3	0.3	0.2	0.3	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.3	0.3	24																						
22	0.3	0.4	0.4	0.4	0.4	S	0.3	0.4	0.3	0.3	0.3	0.3	0.2	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.3	0.3	0.2	0.2	0.4	0.3	24																						
23	0.2	0.2	0.2	0.2	S	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.2	0.1	0.1	0.1	0.2	0.1	0.1	0.2	0.1	0.1	0.2	0.1	0.2	0.1	0.2	24																						
24	0.2	0.2	0.2	S	0.1	0.1	0.2	0.2	0.2	0.3	0.3	0.3	0.2	0.2	0.3	0.3	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.1	0.3	0.2	24																						
25	0.3	0.4	S	0.4	0.4	0.4	0.4	0.4	0.3	0.4	0.4	0.4	0.4	0.5	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.5	0.4	0.3	0.5	0.4	24																						
26	0.4	S	0.5	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.5	0.3	24																						
27	S	0.1	0.1	0.1	0.2	0.2	0.2	0.1	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	S	0.1	0.2	0.1	24																					
28	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.4	0.3	S	0.3	0.2	0.4	0.3	24																					
29	0.3	0.3	0.3	0.3	0.3	0.4	0.4	0.4	0.3	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	S	0.4	0.4	0.3	0.4	0.4	24																					
30	0.4	0.4	0.4	0.5	0.4	0.4	0.3	0.4	0.3	0.3	0.3	0.3	0.3	0.3	0.4	0.3	0.2	0.3	0.2	0.2	S	0.3	0.2	0.2	0.2	0.5	0.3	24																						
31	0.2	0.2	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.3	0.2	0.2	0.1	0.1	0.1	0.1	0.1	S	0.0	0.1	0.1	0.2	0.0	0.3	0.2	24																						
HOURLY MAX	0.4	0.4	0.5	0.5	0.4	0.4	0.4	0.4	0.5	0.4	0.4	0.4	0.4	0.5	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.5	0.4																									
HOURLY AVG	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3																									

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	698				
MINIMUM 1-HR AVERAGE:	0.0 ppb	@ HOUR(S)	10	ON DAY(S)	19
MAXIMUM 1-HR AVERAGE:	0.5 ppb	@ HOUR(S)	8	ON DAY(S)	2
MAXIMUM 24-HR AVERAGE:	0.4 ppb			ON DAY(S)	2
				VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	740 hrs		
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	99.5 %		
STANDARD DEVIATION:	0.1	MONTHLY AVERAGE:	0.3 ppb		

TOTAL REDUCED SULPHUR Hourly Averages (TRS ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake Continuous Monitoring Station - March 2017

TOTAL REDUCED SULPHUR Instantaneous Maximum (TRS ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	0.7	0.9	1.0	S	1.0	0.9	0.9	1.0	0.9	1.0	0.8	0.9	0.9	1.0	0.8	0.9	0.9	0.9	1.0	0.8	0.8	0.9	1.0	0.8	0.7	1.0	0.9	24
2	0.9	0.9	S	0.8	0.9	0.9	1.0	1.3	1.2	0.9	0.9	0.9	0.9	0.8	0.9	0.9	0.9	1.0	1.1	1.0	0.9	0.9	0.9	0.8	1.3	0.9	24	
3	0.9	S	1.0	0.9	0.8	0.9	1.0	1.1	1.0	1.0	0.9	1.0	0.9	1.1	0.9	0.9	0.9	0.9	1.0	0.9	0.9	1.0	0.9	0.8	1.1	0.9	24	
4	S	0.9	0.9	1.1	1.0	1.0	0.9	0.8	0.9	0.9	0.8	0.9	0.8	1.0	0.8	0.8	1.0	0.9	0.9	1.0	0.9	0.8	1.0	S	0.8	1.1	0.9	24
5	0.9	0.8	0.9	0.9	0.8	0.9	0.9	0.8	0.9	0.8	0.7	1.0	0.8	0.8	0.9	0.9	0.9	0.8	0.8	0.9	0.8	0.9	S	0.8	0.7	1.0	0.9	24
6	1.0	0.9	0.9	1.0	0.9	0.8	0.9	0.9	0.9	0.9	C	C	C	Y	Y	C	C	C	C	0.9	0.8	S	0.8	0.7	0.7	1.0	0.9	22
7	0.7	0.7	0.7	0.9	0.9	0.8	0.9	0.9	0.8	0.8	0.9	0.9	0.9	0.7	0.9	0.9	0.7	1.0	0.9	0.8	S	0.7	0.9	0.9	0.7	1.0	0.8	24
8	1.0	0.7	0.8	0.9	0.8	0.8	0.8	0.7	0.8	1.0	0.8	0.7	0.8	0.8	0.9	0.7	0.8	0.8	0.8	S	0.8	0.8	0.9	0.9	0.7	1.0	0.8	24
9	0.9	0.8	0.9	1.0	0.9	0.8	1.0	0.8	0.9	0.8	0.7	0.8	1.0	0.7	0.8	0.9	0.8	0.9	S	0.7	0.8	0.7	0.9	0.8	0.7	1.0	0.8	24
10	0.8	0.9	0.8	0.7	0.8	0.7	0.8	0.9	0.7	0.6	0.8	0.8	0.8	0.8	0.8	0.7	0.8	S	0.9	1.1	0.8	0.8	0.8	0.7	0.6	1.1	0.8	24
11	0.7	0.7	0.8	0.7	0.8	0.8	0.8	X	X	0.9	0.9	0.7	0.7	0.8	0.6	0.8	S	0.9	0.8	1.0	0.9	0.8	0.9	0.7	0.6	1.0	0.8	22
12	0.9	0.9	0.9	0.9	0.9	0.9	0.8	0.8	0.9	0.8	0.9	0.9	0.9	0.7	0.8	S	0.7	0.7	0.8	0.9	0.8	0.8	0.8	1.1	0.7	1.1	0.9	24
13	0.7	0.8	0.8	0.8	0.8	0.7	0.9	0.8	0.8	0.8	0.7	0.8	0.7	0.9	S	0.8	0.9	0.9	1.0	0.8	0.8	0.8	0.9	0.9	0.7	1.0	0.8	24
14	0.8	1.0	0.9	0.8	0.8	0.9	1.1	0.8	0.9	0.9	0.9	0.8	0.9	S	0.7	0.9	1.0	0.9	1.0	0.9	0.8	0.9	0.8	0.8	0.7	1.1	0.9	24
15	0.9	0.9	0.9	1.0	0.9	0.9	0.8	0.9	1.0	0.9	0.8	0.8	S	0.8	0.9	0.9	0.8	0.9	0.9	0.9	0.7	0.9	0.8	0.9	0.7	1.0	0.9	24
16	0.9	0.8	0.8	0.8	0.9	0.8	0.9	0.9	0.9	0.9	0.8	S	0.8	0.8	0.8	1.0	0.9	0.8	0.9	0.8	0.9	0.9	0.8	0.8	0.8	1.0	0.9	24
17	0.8	0.8	0.8	0.9	0.8	0.8	1.0	0.8	0.9	0.9	S	1.0	0.8	0.8	0.9	0.9	0.8	0.9	1.0	0.8	0.8	0.9	0.8	0.8	0.8	1.0	0.9	24
18	0.8	0.9	1.1	0.9	0.8	0.9	0.9	0.9	0.7	S	0.8	1.0	0.9	0.8	0.9	0.8	0.9	0.8	0.8	0.8	0.8	0.9	1.3	1.2	0.7	1.3	0.9	24
19	1.0	1.1	0.8	0.9	1.0	0.9	0.9	0.9	S	0.9	0.7	0.8	0.9	0.9	0.9	0.8	0.9	0.9	0.9	0.9	0.8	0.8	0.9	0.8	0.7	1.1	0.9	24
20	0.8	0.9	0.9	0.7	0.9	0.7	0.8	S	0.8	0.9	0.7	0.8	0.8	0.8	0.7	1.0	0.8	0.9	1.1	0.7	1.0	0.9	0.8	0.9	0.7	1.1	0.8	24
21	0.8	0.8	0.9	0.9	0.8	0.8	S	0.9	0.9	0.9	0.9	0.8	1.0	0.9	0.7	1.0	0.8	0.9	0.9	0.9	0.8	0.9	0.8	0.7	1.0	0.9	24	
22	0.7	0.9	0.8	1.0	0.9	S	0.9	0.9	0.8	0.9	0.7	0.9	0.9	0.9	0.9	0.8	0.9	1.0	0.9	1.1	0.9	0.9	1.1	1.0	0.7	1.1	0.9	24
23	0.9	0.9	0.9	0.9	S	1.0	0.8	0.9	0.9	0.9	1.0	0.8	0.8	0.8	0.8	1.0	0.8	0.9	0.9	0.9	0.8	0.8	0.8	0.8	0.8	1.0	0.9	24
24	0.9	1.0	0.9	S	0.9	0.7	0.9	0.9	0.9	1.0	0.8	0.9	0.9	1.0	0.8	1.1	0.9	0.9	0.8	0.8	0.7	1.0	0.8	1.0	0.7	1.1	0.9	24
25	0.8	0.9	S	1.0	1.0	0.9	0.9	0.8	1.0	0.9	0.9	0.8	0.8	1.0	1.0	0.8	1.0	1.0	0.9	0.8	0.9	0.9	1.1	0.9	0.8	1.1	0.9	24
26	1.0	S	0.9	1.0	0.9	1.0	0.8	0.9	0.9	1.0	0.9	0.8	0.9	0.9	0.9	0.9	0.8	0.9	0.8	1.1	0.9	0.9	1.0	0.9	0.8	1.1	0.9	24
27	S	0.9	0.9	0.9	0.9	0.9	0.9	0.9	1.0	0.8	0.9	0.7	0.8	0.9	0.9	0.9	0.9	1.0	0.9	0.8	0.7	0.9	0.9	S	0.7	1.0	0.9	24
28	1.0	0.9	1.0	1.0	0.9	1.0	1.2	1.2	1.0	1.0	1.0	0.9	1.1	1.0	1.0	1.0	1.0	1.0	1.0	1.0	1.1	1.0	S	1.1	0.9	1.2	1.0	24
29	1.1	1.1	1.0	1.0	1.0	0.9	0.9	1.0	1.1	1.0	1.0	1.1	1.0	1.0	0.9	0.9	1.0	1.0	1.2	1.0	1.0	S	1.0	1.0	0.9	1.2	1.0	24
30	1.0	1.2	1.0	1.1	1.1	1.1	0.8	0.9	1.1	1.0	0.9	0.9	0.9	0.9	0.9	0.9	0.9	0.9	0.9	0.9	S	1.1	0.9	0.9	0.8	1.2	1.0	24
31	0.7	0.9	0.9	0.9	0.9	1.0	0.9	P	1.2	1.0	1.0	1.0	1.0	1.0	0.9	0.8	0.9	0.9	0.8	S	0.9	1.0	0.9	1.1	0.7	1.2	0.9	23
HOURLY MAX	1.1	1.2	1.1	1.1	1.1	1.1	1.2	1.3	1.2	1.0	1.0	1.1	1.1	1.1	1.0	1.1	1.0	1.0	1.2	1.1	1.1	1.1	1.3	1.2				
HOURLY AVG	0.9	0.9	0.9	0.9	0.9	0.9	0.9	0.9	0.9	0.9	0.8	0.9	0.9	0.9	0.8	0.9	0.9	0.9	0.9	0.9	0.9	0.9	0.9	0.9				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	699
MAXIMUM INSTANTANEOUS VALUE:	1.3 ppb @ HOUR(S) 7 ON DAY(S) 2
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	7 hrs
OPERATIONAL TIME:	739 hrs
STANDARD DEVIATION:	0.1

TOTAL REDUCED SULPHUR Instantaneous Maximum (TRS ppb)




Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-TRS[ppb]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 12.16% Calm Avg: 0.27 [ppb]


Direction	0.0-0.6	0.6-1.3	1.3-1.9	>1.9	Total
N	5.6	0.0	0.0	0.0	5.6
NE	16.5	0.0	0.0	0.0	16.5
E	21.9	0.0	0.0	0.0	21.9
SE	27.2	0.0	0.0	0.0	27.2
S	1.0	0.0	0.0	0.0	1.0
SW	4.4	0.0	0.0	0.0	4.4
W	6.3	0.0	0.0	0.0	6.3
NW	5.0	0.0	0.0	0.0	5.0
Summary	87.8	0.0	0.0	0.0	87.8

% Icon Classes (ppb) 88 0.0-0.6 0 0.6-1.3 0 1.3-1.9 0 >1.9

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-TRS[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 12.16% Calm Poll Avg: 0.27[ppb]



TRS[ppb] Calibration: LICA COLD LAKE SOUTH Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

TOTAL HYDROCARBON




TOTAL HYDROCARBONS Hourly Averages (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2.0	2.1	2.0	S	2.0	2.1	2.1	2.1	2.2	2.1	2.0	2.1	2.0	2.0	2.0	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.2	2.1	2.0	2.2	2.1	24	
2	2.1	2.1	S	2.1	2.1	2.1	2.2	2.3	2.3	2.2	2.2	2.2	2.1	2.1	2.2	2.3	2.3	2.2	2.2	2.3	2.2	2.2	2.1	2.1	2.1	2.1	2.3	2.2	24
3	2.1	S	2.1	2.2	2.2	2.3	2.4	2.2	2.2	2.1	2.2	2.2	2.2	2.2	2.1	2.1	2.2	2.3	2.4	2.5	2.4	2.5	2.2	2.1	2.1	2.5	2.2	24	
4	S	2.0	2.1	2.2	2.2	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.0	S	2.0	2.2	2.1	24	
5	2.0	2.0	2.0	2.0	2.1	2.0	2.0	2.0	2.0	2.1	2.1	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.2	2.0	S	2.0	2.0	2.2	2.0	24
6	2.0	2.0	2.0	2.0	2.0	2.1	2.1	2.1	2.1	2.2	2.1	2.2	S1	S1	2.1	2.1	Y	2.1	2.1	2.1	2.1	S	2.0	2.0	2.0	2.2	2.1	21	
7	2.0	2.0	2.1	2.1	2.1	2.1	2.2	2.1	C	C	C	C	C	C	2.0	2.0	2.0	2.0	2.0	2.1	2.1	S	2.1	2.1	2.1	2.0	2.2	2.1	24
8	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.0	2.0	2.0	2.0	2.0	S	2.1	2.1	2.0	2.1	2.0	2.1	2.1	24
9	2.1	2.1	2.2	2.2	2.2	2.2	2.3	2.3	2.3	2.2	2.1	2.1	2.2	2.1	2.1	2.1	2.1	2.1	2.1	S	2.1	2.2	2.3	2.3	2.4	2.1	2.4	2.2	24
10	2.4	2.4	2.3	2.3	2.3	2.2	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.0	2.1	2.1	S	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.0	2.4	2.1	24
11	2.1	2.2	2.2	2.2	2.2	2.2	2.1	X	X	2.1	2.1	2.1	2.1	2.1	2.1	2.1	S	2.0	2.1	2.1	2.1	2.0	2.0	2.0	2.0	2.0	2.2	2.1	22
12	2.0	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.2	2.1	2.1	2.1	2.1	2.1	S	2.1	2.1	2.1	2.2	2.2	2.1	2.1	2.2	2.0	2.2	2.1	24	
13	2.2	2.2	2.3	2.2	2.2	2.2	2.1	2.1	2.1	2.1	2.1	2.1	2.1	S	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.2	2.2	2.1	2.3	2.1	24	
14	2.2	2.2	2.2	2.3	2.3	2.3	2.3	2.4	2.4	2.4	2.4	2.4	S	2.3	2.3	2.3	2.3	2.3	2.2	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.4	2.3	24
15	2.1	2.0	2.0	2.1	2.1	2.2	2.2	2.2	2.2	2.3	2.4	2.3	S	2.3	2.4	2.4	2.4	2.3	2.4	2.5	2.5	2.5	2.0	2.0	2.0	2.5	2.2	24	
16	2.0	2.0	2.1	2.1	2.0	2.1	2.0	2.1	2.1	2.0	2.0	S	1.9	1.9	2.0	2.0	2.0	2.0	2.1	2.1	2.1	2.1	2.2	2.2	2.2	1.9	2.2	2.0	24
17	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.2	2.2	S	2.2	2.1	2.1	2.1	2.0	2.0	2.0	2.1	2.1	2.2	2.2	2.2	2.2	2.0	2.2	2.1	24	
18	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	2.1	2.1	2.1	10
19	X	X	X	X	X	X	X	X	S	2.0	2.0	2.0	2.0	1.9	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.9	2.0	2.0	16	
20	2.0	2.0	2.1	2.1	2.1	2.1	2.1	2.1	S	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.2	2.1	2.2	2.1	24	
21	2.3	2.3	2.2	2.1	2.1	2.1	S	2.2	2.2	2.2	2.1	2.1	2.1	2.1	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.3	2.1	24
22	2.0	2.0	2.0	2.0	S	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.1	2.0	2.2	2.2	2.4	2.0	2.0	2.4	2.0	24	
23	2.5	2.5	2.4	2.4	S	2.6	2.6	2.6	2.2	2.1	2.0	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.0	2.6	2.2	24
24	2.2	2.2	2.2	S	2.3	2.4	2.6	2.5	2.5	2.5	2.5	2.4	2.4	2.3	2.2	2.1	2.0	2.0	2.1	2.1	2.1	2.0	2.0	2.0	2.0	2.6	2.2	24	
25	2.0	2.0	S	2.0	2.0	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.0	2.1	2.0	2.0	2.0	2.0	2.0	2.1	2.2	2.3	2.3	2.1	2.0	2.3	2.1	24	
26	2.1	S	2.0	2.1	2.1	2.1	2.1	2.2	2.2	2.2	2.2	2.1	2.0	2.1	2.0	2.0	2.0	2.0	2.0	2.1	2.2	2.4	2.5	2.2	2.0	2.5	2.1	24	
27	S	2.4	2.3	2.2	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.1	2.0	2.0	2.0	2.1	2.0	2.0	1.9	2.0	2.0	2.1	2.1	S	1.9	2.4	2.1	24	
28	2.1	2.1	2.1	2.1	2.1	2.2	2.1	2.2	2.2	2.1	2.1	2.1	2.0	2.1	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	S	2.0	2.0	2.2	2.1	24	
29	2.1	2.1	2.1	2.2	2.3	2.2	2.1	2.2	2.1	2.1	2.2	2.0	2.1	2.1	2.0	2.1	2.1	2.0	2.1	2.0	2.0	S	2.0	2.0	2.0	2.3	2.1	24	
30	2.0	2.1	2.1	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.9	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	S	2.2	2.3	2.3	1.9	2.3	2.0	24
31	2.3	2.3	2.3	2.4	2.1	2.1	2.3	2.2	2.4	2.5	2.5	2.5	2.3	2.1	2.0	2.0	2.0	2.0	2.1	S	2.1	2.1	2.1	2.0	2.0	2.0	2.5	2.2	24
HOURLY MAX	2.47	2.51	2.40	2.43	2.27	2.58	2.64	2.63	2.53	2.49	2.48	2.48	2.48	2.34	2.35	2.39	2.36	2.34	2.38	2.50	2.51	2.51	2.47	2.39					
HOURLY AVG	2.12	2.13	2.13	2.13	2.12	2.15	2.16	2.17	2.17	2.15	2.14	2.12	2.10	2.08	2.06	2.06	2.06	2.05	2.08	2.11	2.13	2.14	2.13	2.12					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	678			
MINIMUM 1-HR AVERAGE:	1.87 ppm	@ HOUR(S)	14	ON DAY(S) 19
MAXIMUM 1-HR AVERAGE:	2.64 ppm	@ HOUR(S)	6	ON DAY(S) 23
MAXIMUM 24-HR AVERAGE:	2.26 ppm			ON DAY(S) 14
				VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	717 hrs	
MONTHLY CALIBRATION TIME:	6 hrs	AMD OPERATION UPTIME:	96.4 %	
STANDARD DEVIATION:	0.12	MONTHLY AVERAGE:	2.12 ppm	

TOTAL HYDROCARBONS Hourly Averages (THC ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake Continuous Monitoring Station - March 2017

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY 1	2.3	2.4	2.4	S	2.4	2.5	2.7	2.4	3.1	2.7	2.4	2.4	2.4	2.4	2.4	2.5	2.5	2.5	2.5	2.5	2.6	2.6	2.6	2.6	2.3	3.1	2.5	24
2	2.5	2.5	S	2.5	2.5	2.8	4.0	2.8	2.8	2.5	2.6	2.6	2.5	2.5	2.5	2.6	2.6	2.5	2.5	2.6	2.5	2.6	2.4	2.3	2.3	4.0	2.6	24
3	2.4	S	2.3	2.4	2.4	3.7	5.4	2.4	2.4	2.3	2.4	2.3	2.3	2.3	2.3	2.3	2.3	2.4	2.5	2.6	2.5	2.8	2.6	2.2	2.2	5.4	2.6	24
4	S	2.1	2.2	2.3	2.3	2.2	2.2	2.2	2.2	2.2	2.6	2.2	2.2	2.2	2.3	2.3	2.3	2.3	2.3	2.3	2.3	2.3	2.3	S	2.1	2.6	2.3	24
5	2.2	2.2	2.2	2.3	2.3	2.3	2.3	2.2	2.3	2.5	2.4	2.2	2.2	2.2	2.2	2.3	2.3	2.2	2.2	2.2	2.6	2.2	S	2.2	2.2	2.6	2.3	24
6	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.6	2.3	2.6	S1	S1	2.8	2.9	X	2.6	2.6	2.2	2.2	S	2.2	2.2	2.2	2.9	24
7	2.2	2.2	2.2	2.2	2.2	2.2	2.3	2.3	C	C	C	C	C	C	2.1	2.6	2.2	2.2	2.2	2.2	S	2.2	2.2	2.3	2.1	2.6	2.2	24
8	2.3	2.3	2.3	2.2	2.3	2.3	2.3	2.3	2.4	2.4	2.3	2.4	2.4	2.3	3.1	2.3	2.2	2.7	2.3	S	2.6	2.4	2.3	2.4	2.2	3.1	2.4	24
9	5.1	2.4	2.5	2.5	2.5	2.6	2.7	2.8	2.7	2.5	2.5	2.5	3.0	4.5	2.7	2.4	2.4	2.5	S	2.6	2.7	2.7	2.8	2.8	2.4	5.1	2.8	24
10	2.8	2.7	2.7	2.7	2.6	3.4	2.5	3.8	2.7	2.5	2.5	2.4	2.6	3.0	2.4	2.5	2.4	S	2.4	2.5	2.5	2.4	2.5	2.4	2.4	3.8	2.6	24
11	2.5	2.5	2.6	2.6	2.6	2.5	2.7	X	X	2.4	2.4	2.4	2.4	2.3	2.3	2.3	S	3.0	2.3	2.3	2.3	2.3	2.3	2.3	2.2	3.0	2.4	22
12	2.3	2.3	2.3	2.3	2.3	2.3	2.3	2.4	3.0	2.5	2.3	2.5	2.5	2.4	S	2.3	2.3	2.3	2.3	2.3	2.3	2.3	2.3	2.3	2.3	3.0	2.4	24
13	2.3	2.4	2.4	2.3	2.2	2.7	2.9	2.2	2.3	2.4	2.2	2.1	2.1	2.1	S	2.3	2.3	2.1	2.1	2.2	3.5	2.2	2.2	2.2	2.1	3.5	2.3	24
14	2.2	2.2	2.2	2.4	2.4	2.4	7.7	2.4	2.4	2.5	2.5	2.4	2.4	S	2.3	2.5	2.3	2.3	2.3	2.2	2.1	2.1	2.0	2.0	2.0	7.7	2.5	24
15	2.0	2.0	1.9	2.0	2.0	2.1	2.6	2.2	2.2	2.3	2.4	2.3	S	2.2	2.5	2.4	2.3	2.4	2.3	2.5	2.5	2.8	1.9	1.9	1.9	2.8	2.2	24
16	1.9	1.9	2.0	2.0	1.9	2.0	2.0	2.1	2.0	2.0	2.0	S	2.0	1.9	2.0	1.9	2.0	2.0	2.0	2.1	2.1	2.1	2.2	2.3	1.9	2.3	2.0	24
17	2.2	2.2	2.1	2.1	2.2	2.4	2.2	2.5	2.4	2.3	S	2.3	2.1	2.1	2.2	2.1	2.0	2.0	2.0	2.1	2.3	2.1	2.1	2.1	2.0	2.5	2.2	24
18	2.0	2.0	2.0	2.0	1.9	1.9	1.9	2.1	1.9	S	X	X	X	X	X	X	X	X	X	X	X	X	X	X	1.9	2.1	2.0	10
19	X	X	X	X	X	X	X	X	S	2.1	2.1	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.1	2.1	2.1	2.1	2.0	2.1	2.0	16
20	2.1	2.1	2.1	2.2	2.2	2.2	2.2	S	2.3	2.2	2.3	2.3	2.2	2.3	2.3	2.6	2.3	2.2	2.2	2.5	2.3	2.7	2.3	2.4	2.1	2.7	2.3	24
21	2.5	2.5	2.5	2.3	2.5	2.3	S	2.7	2.4	2.3	2.6	2.2	2.2	2.1	2.1	2.1	2.1	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.7	2.2	24
22	2.0	2.0	2.0	2.0	1.9	S	2.1	2.0	1.9	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.1	2.1	1.9	2.4	2.0	24
23	2.4	2.5	2.4	2.4	S	2.6	2.6	2.6	2.4	2.1	2.0	2.0	2.0	2.1	2.1	2.1	2.1	2.0	2.1	2.1	2.1	2.1	2.2	2.1	2.0	2.6	2.2	24
24	2.2	2.2	2.5	S	3.8	2.5	3.1	2.8	2.5	2.5	2.5	2.4	2.3	2.3	2.2	2.1	2.0	2.0	2.0	2.0	2.0	1.9	1.9	1.9	1.9	3.8	2.3	24
25	1.9	1.9	S	1.9	1.9	2.0	2.0	2.0	2.0	2.1	2.0	2.3	2.0	2.0	2.1	2.0	2.0	2.0	2.1	2.2	2.3	2.4	2.4	2.3	1.9	2.4	2.1	24
26	2.2	S	2.1	2.1	2.1	2.1	2.2	2.3	3.0	2.3	2.4	2.1	2.0	2.1	2.0	1.9	2.0	2.0	2.0	2.1	2.4	2.6	2.6	2.4	1.9	3.0	2.2	24
27	S	2.5	2.4	2.3	2.1	2.1	2.1	2.2	2.0	2.1	2.1	2.1	2.0	2.0	2.1	2.1	1.9	1.9	1.8	1.9	1.9	2.0	2.0	S	1.8	2.5	2.1	24
28	2.0	1.9	2.0	2.0	2.0	2.6	2.2	2.1	2.1	2.0	2.1	2.0	2.0	2.0	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.0	S	2.0	1.9	2.6	2.0	24
29	2.1	2.1	2.1	2.2	3.6	2.2	2.4	2.1	2.1	2.1	2.1	2.3	2.0	2.1	2.4	2.1	2.0	2.0	1.9	2.0	1.9	1.9	S	1.9	1.9	3.6	2.1	24
30	2.0	2.0	2.0	2.0	2.1	2.0	1.9	2.1	2.5	2.2	2.1	2.0	2.0	2.0	2.0	1.9	2.0	2.0	2.1	2.4	S	2.3	2.6	2.4	1.9	2.6	2.1	24
31	2.4	2.4	2.6	2.6	2.2	2.5	2.4	P	2.5	2.7	2.6	2.5	2.6	2.6	2.3	2.1	2.0	2.0	2.1	S	2.1	2.1	2.1	2.0	2.0	2.7	2.3	23
HOURLY MAX	5.11	2.74	2.68	2.74	3.81	3.69	7.65	3.76	3.13	2.98	2.64	2.59	2.95	4.51	3.08	2.92	2.58	3.04	2.58	2.61	3.45	2.81	2.75	2.81				
HOURLY AVG	2.33	2.23	2.25	2.24	2.33	2.40	2.69	2.37	2.37	2.33	2.31	2.26	2.23	2.30	2.25	2.23	2.17	2.20	2.18	2.23	2.30	2.29	2.25	2.23				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	677
MAXIMUM INSTANTANEOUS VALUE:	7.65 ppm @ HOUR(S) 6 ON DAY(S) 14
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	6 hrs
OPERATIONAL TIME:	716 hrs
STANDARD DEVIATION:	0.39

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)




Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-THC[ppm]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 12.54% Calm Avg: 2.23 [ppm]


Direction	0.0-0.9	0.9-1.8	1.8-2.7	>2.7	Total
N	0.0	0.0	6.1	0.0	6.1
NE	0.0	0.0	17.1	0.0	17.1
E	0.0	0.0	21.8	0.0	21.8
SE	0.0	0.0	27.0	0.0	27.0
S	0.0	0.0	1.0	0.0	1.0
SW	0.0	0.0	4.6	0.0	4.6
W	0.0	0.0	5.9	0.0	5.9
NW	0.0	0.0	4.0	0.0	4.0
Summary	0.0	0.0	87.5	0.0	87.5

% Icon Classes (ppm) 0 0.0-0.9 0 0.9-1.8 87 1.8-2.7 0 >2.7

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-THC[ppm] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 12.54% Calm Poll Avg: 2.23[ppm]



THC[ppm] Calibration: LICA COLD LAKE SOUTH Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

OXIDES OF NITROGEN




OXIDES OF NITROGEN Hourly Averages (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2.7	3.3	2.8	S	3.7	4.5	4.1	3.3	4.3	3.1	3.0	2.9	1.4	1.5	1.7	1.8	1.5	2.5	3.3	4.7	8.1	8.7	9.9	3.4	1.4	9.9	3.7	24	
2	1.5	1.0	S	1.5	1.9	2.0	5.0	15.4	18.0	4.2	4.6	4.4	3.5	3.5	4.0	5.0	5.1	8.9	8.4	3.6	2.9	2.8	2.6	2.7	1.0	18.0	4.9	24	
3	3.8	S	4.9	5.1	10.7	19.2	26.4	11.5	18.8	10.9	5.9	4.1	3.1	4.7	3.7	3.5	3.6	4.6	8.9	21.5	12.2	9.5	3.9	3.4	3.1	26.4	8.9	24	
4	S	3.6	2.5	5.2	3.8	1.9	1.5	1.4	1.4	1.4	1.6	1.4	1.4	1.3	1.7	2.0	2.6	1.7	1.7	1.9	1.7	2.0	1.6	S	1.3	5.2	2.1	24	
5	2.0	2.6	1.8	2.2	1.9	2.3	2.6	3.5	2.7	2.6	2.0	1.4	1.4	1.5	1.1	1.5	1.4	1.6	4.3	5.2	4.1	2.7	S	2.6	1.1	5.2	2.4	24	
6	1.9	2.1	1.6	1.7	1.4	2.1	2.6	5.0	2.5	2.5	C	C	C	C	C	C	C	2.2	2.0	1.5	2.4	S	2.4	1.6	1.4	5.0	2.2	24	
7	1.2	1.5	1.6	1.7	1.5	1.4	1.8	1.9	3.6	2.1	2.2	2.6	2.1	2.0	2.3	2.2	2.5	2.6	2.7	3.0	S	2.7	2.6	2.4	1.2	3.6	2.2	24	
8	2.6	4.8	2.5	2.1	2.1	1.4	4.1	2.5	3.2	1.9	1.4	1.9	1.6	1.6	2.6	1.8	2.5	4.7	4.8	S	12.6	8.6	4.1	6.3	1.4	12.6	3.6	24	
9	9.4	8.9	7.0	10.4	10.6	7.1	11.6	14.5	19.8	8.0	2.3	1.7	2.8	2.0	3.4	2.8	2.9	8.0	S	10.2	14.5	21.6	19.7	20.2	1.7	21.6	9.5	24	
10	22.0	25.8	16.8	13.7	17.9	12.0	6.3	9.8	5.2	3.4	4.2	2.0	2.2	2.1	2.9	3.7	3.6	S	5.1	9.2	7.6	4.0	4.8	3.4	2.0	25.8	8.2	24	
11	2.4	4.2	9.2	9.6	12.1	7.3	2.2	X	X	1.6	2.5	1.6	1.5	1.4	3.2	2.5	S	3.7	4.8	4.1	3.5	3.3	3.2	3.1	1.4	12.1	4.1	22	
12	3.4	2.7	2.8	2.8	4.3	2.9	2.8	2.7	2.5	2.4	2.3	1.7	1.4	1.5	1.2	S	2.7	3.2	1.4	1.7	1.4	1.2	1.2	1.9	1.2	4.3	2.3	24	
13	1.9	1.9	1.7	6.4	2.0	1.9	2.1	2.2	1.9	2.2	3.1	1.9	1.6	1.4	S	1.9	1.8	2.0	2.3	2.3	2.5	2.5	2.8	3.2	1.4	6.4	2.3	24	
14	2.9	6.0	3.8	5.1	3.3	4.9	5.0	9.4	5.8	3.7	5.2	4.0	5.6	S	6.5	6.2	7.4	8.2	4.4	4.0	3.5	3.0	2.7	2.5	2.5	9.4	4.9	24	
15	2.8	2.3	2.2	2.2	2.5	3.9	6.3	5.3	3.9	4.0	3.6	4.4	S	7.8	8.1	8.3	7.9	7.9	12.9	17.2	11.1	7.3	2.0	1.2	1.2	17.2	5.9	24	
16	2.0	3.4	5.1	4.6	4.0	11.9	3.8	11.6	5.4	4.8	2.5	S	1.6	1.3	1.0	1.0	1.5	2.6	2.9	3.1	2.6	2.5	4.3	4.0	1.0	11.9	3.8	24	
17	3.0	2.6	1.0	0.8	2.3	3.4	2.4	2.9	4.2	2.9	S	2.5	1.8	1.9	1.8	1.8	1.7	2.2	2.1	2.1	2.1	1.7	1.7	1.5	0.8	4.2	2.2	24	
18	1.4	1.5	1.4	1.3	1.8	1.7	2.4	3.1	2.8	S	1.8	1.8	1.4	1.3	1.5	1.8	1.7	2.8	3.7	2.8	3.0	2.3	2.1	2.5	1.3	3.7	2.1	24	
19	2.0	2.4	1.5	1.6	1.0	0.9	1.2	1.2	S	1.9	1.7	1.4	1.3	1.2	1.2	1.3	1.8	1.2	1.6	1.3	1.5	2.3	2.5	2.2	0.9	2.5	1.6	24	
20	1.1	1.1	1.4	0.8	0.9	0.9	0.9	S	3.9	1.1	1.7	0.9	0.8	0.9	1.1	1.2	1.4	1.5	1.7	2.4	3.2	3.7	4.8	3.3	0.8	4.8	1.8	24	
21	9.4	10.0	5.4	4.7	4.3	6.8	S	7.2	4.2	4.3	3.4	2.6	2.1	2.3	2.4	1.8	1.5	1.4	1.4	1.3	1.5	1.8	1.8	1.7	1.3	10.0	3.6	24	
22	1.7	1.7	1.6	1.7	3.0	S	3.4	5.5	5.2	4.0	3.9	4.1	3.7	5.3	4.2	3.0	3.0	2.0	1.6	3.2	4.6	7.0	9.0	10.4	1.6	10.4	4.0	24	
23	12.1	11.5	9.1	9.9	S	14.3	18.4	22.4	10.1	4.5	2.4	1.8	1.8	1.6	2.3	2.7	2.8	4.3	5.5	8.0	4.8	3.6	4.2	3.0	1.6	22.4	7.0	24	
24	4.7	4.0	6.2	S	8.5	6.5	11.5	6.5	4.7	4.3	3.9	3.2	2.5	2.9	3.4	4.3	3.2	2.7	2.3	2.1	2.0	2.0	2.4	2.4	2.0	11.5	4.2	24	
25	2.0	2.2	S	2.5	2.7	4.1	2.7	3.0	1.7	2.1	4.6	4.1	3.0	4.3	2.8	2.6	2.4	3.2	6.1	14.3	23.9	20.1	6.7	1.7	23.9	5.4	24		
26	2.6	S	2.3	3.9	6.7	8.5	12.4	9.1	3.0	3.4	3.0	2.5	2.3	2.0	1.5	1.4	1.5	1.6	1.9	3.3	4.6	12.4	20.0	8.0	1.4	20.0	5.1	24	
27	S	8.7	13.4	9.0	4.9	4.5	3.5	3.7	2.0	1.6	1.5	1.4	1.5	1.5	1.7	2.8	2.7	2.3	2.2	3.2	3.3	1.3	1.2	S	1.2	13.4	3.5	24	
28	1.8	1.8	1.1	0.9	1.3	2.4	2.1	3.1	2.2	1.9	2.2	1.8	1.9	1.8	2.1	1.9	1.8	2.1	2.4	2.5	2.1	2.1	S	3.3	0.9	3.3	2.0	24	
29	4.8	2.7	2.6	1.8	5.6	7.9	5.3	4.8	3.4	2.2	2.1	1.9	2.4	4.4	2.9	2.9	3.0	4.1	6.0	3.9	2.3	S	2.0	1.9	1.8	7.9	3.5	24	
30	1.9	2.1	2.2	2.0	2.5	2.8	3.5	3.8	2.6	2.5	2.1	2.1	1.9	1.4	1.4	1.8	1.4	2.0	2.7	6.1	S	6.5	12.0	7.1	1.4	12.0	3.2	24	
31	13.9	8.0	3.4	10.6	1.6	2.4	5.0	4.5	4.7	5.3	5.8	5.2	4.8	5.4	2.6	1.3	1.7	1.5	2.7	S	2.4	3.1	2.1	2.1	1.3	13.9	4.4	24	
HOURLY MAX	22.0	25.8	16.8	13.7	17.9	19.2	26.4	22.4	19.8	10.9	5.9	5.2	5.6	7.8	8.1	8.3	7.9	8.9	12.9	21.5	14.5	23.9	20.1	20.2					
HOURLY AVG	4.3	4.6	4.1	4.3	4.4	5.1	5.4	6.2	5.3	3.4	3.0	2.5	2.2	2.5	2.6	2.6	2.7	3.3	3.7	4.9	4.9	5.4	5.3	4.1					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	702				
MINIMUM 1-HR AVERAGE:	0.8	ppb @ HOUR(S)	3	ON DAY(S) 17	
MAXIMUM 1-HR AVERAGE:	26.4	ppb @ HOUR(S)	6	ON DAY(S) 3	
MAXIMUM 24-HR AVERAGE:	9.5	ppb		ON DAY(S) 9	
				VAR-VARIOUS	
IZS CALIBRATION TIME:	33	hrs	OPERATIONAL TIME:	742	hrs
MONTHLY CALIBRATION TIME:	7	hrs	AMD OPERATION UPTIME:	99.7	%
STANDARD DEVIATION:	3.8		MONTHLY AVERAGE:	4.0	ppb

OXIDES OF NITROGEN Hourly Averages (NO_x ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake Continuous Monitoring Station - March 2017

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	4.2	5.2	4.5	S	8.6	8.9	7.1	5.9	22.2	9.9	12.0	21.3	5.4	4.2	8.1	6.1	4.2	4.4	6.7	10.8	15.3	13.4	17.5	7.6	4.2	22.2	9.3	24
2	3.2	1.7	S	2.7	3.7	5.0	13.9	24.7	50.0	20.1	23.0	9.7	6.9	7.2	8.1	7.6	7.6	13.6	11.9	5.3	3.6	4.2	3.5	3.5	1.7	50.0	10.5	24
3	8.4	S	7.3	9.4	20.1	43.6	38.3	16.5	45.9	20.3	14.9	6.5	6.1	7.0	13.3	6.9	4.5	6.6	13.4	43.4	26.9	14.6	8.3	11.1	4.5	45.9	17.1	24
4	S	8.2	4.5	6.4	5.5	3.2	2.3	2.7	2.7	2.6	9.8	2.4	2.7	2.4	3.1	9.6	7.2	4.6	4.5	3.7	6.4	4.8	3.0	S	2.3	9.8	4.7	24
5	6.9	7.6	4.6	7.1	3.7	5.4	5.9	8.2	12.4	6.4	4.1	4.2	2.8	9.0	2.4	3.2	2.8	3.2	13.2	8.9	8.1	8.9	S	9.1	2.4	13.2	6.4	24
6	5.4	6.8	3.6	5.5	2.8	5.3	6.1	10.6	3.6	8.3	C	C	C	C	C	C	C	4.2	3.2	2.3	3.0	S	3.0	2.2	2.2	10.6	4.7	24
7	1.7	2.2	1.9	2.0	1.9	2.0	2.2	2.3	13.9	2.6	2.7	6.1	3.2	9.7	6.2	3.2	3.4	3.2	5.9	5.7	S	5.4	4.4	4.1	1.7	13.9	4.2	24
8	4.2	8.2	3.5	3.6	6.7	3.1	6.7	4.5	4.9	4.6	4.3	4.7	4.8	3.6	12.1	5.2	6.8	10.2	8.4	S	20.1	12.1	11.6	12.9	3.1	20.1	7.3	24
9	25.2	14.2	11.1	18.5	16.8	11.3	19.6	19.6	34.1	18.8	8.4	4.0	5.9	3.4	6.4	6.1	4.5	110.5	S	17.7	26.6	36.3	26.6	27.3	3.4	110.5	20.6	24
10	29.7	33.4	20.5	17.9	28.3	25.6	11.2	20.4	7.9	5.9	18.6	5.2	5.4	5.1	6.1	6.2	6.1	S	8.0	14.7	13.5	6.9	7.0	5.5	5.1	33.4	13.4	24
11	4.1	7.5	18.7	18.4	25.6	46.0	4.4	X	X	4.1	6.7	5.3	3.1	4.0	21.5	5.7	S	5.9	8.2	6.6	7.1	4.9	9.5	6.3	3.1	46.0	10.6	22
12	7.3	4.5	4.8	4.4	11.9	4.6	6.4	4.5	6.1	4.4	5.1	3.0	2.7	4.2	2.8	S	6.1	8.3	2.5	2.7	2.3	1.7	1.8	3.0	1.7	11.9	4.6	24
13	2.4	2.7	2.6	37.0	3.2	3.1	3.0	3.3	3.9	3.0	9.5	8.7	5.7	2.1	S	5.3	4.6	3.5	8.0	3.0	3.1	3.5	3.9	5.3	2.1	37.0	5.7	24
14	4.7	13.7	6.3	8.2	6.7	14.9	9.6	14.1	10.3	4.9	15.4	5.1	10.7	S	9.4	8.6	15.8	13.9	7.2	4.9	4.9	4.9	4.2	4.5	4.2	15.8	8.8	24
15	4.5	4.0	3.1	3.2	4.1	7.2	12.5	11.2	6.5	5.7	5.9	5.5	S	9.6	12.3	12.9	13.2	12.7	42.4	29.5	24.4	12.6	3.5	2.2	2.2	42.4	10.8	24
16	5.0	7.1	17.8	9.0	8.4	100.9	6.2	27.0	23.3	8.6	6.1	S	5.4	2.4	4.0	2.2	2.3	8.6	8.2	6.7	7.2	5.1	9.4	12.0	2.2	100.9	12.7	24
17	5.9	4.0	2.6	1.3	5.4	6.8	6.3	4.9	7.1	5.5	S	4.2	2.3	3.7	3.9	4.9	2.1	2.8	3.0	4.6	2.8	2.3	2.3	2.8	1.3	7.1	4.0	24
18	1.9	2.1	2.7	2.2	3.1	2.7	4.2	8.3	8.4	S	3.2	2.8	2.3	1.7	3.2	4.3	2.7	4.9	6.2	4.2	4.4	6.8	3.0	3.9	1.7	8.4	3.9	24
19	3.3	3.3	2.0	1.9	1.9	1.7	3.1	1.9	S	2.6	2.4	1.9	1.9	1.7	2.1	1.8	4.2	2.7	3.1	2.0	2.3	3.9	4.2	2.8	1.7	4.2	2.6	24
20	2.3	1.7	1.8	1.3	1.1	1.3	1.8	S	15.7	1.9	8.6	1.4	1.5	2.4	2.4	6.3	3.5	2.7	3.0	4.1	6.3	14.6	14.7	6.3	1.1	15.7	4.6	24
21	59.8	20.0	15.3	6.7	14.3	11.7	S	19.9	20.4	16.1	13.4	14.9	7.9	15.7	9.8	4.6	2.2	2.7	2.8	2.5	2.4	2.9	2.6	2.7	2.2	59.8	11.8	24
22	3.3	3.3	2.4	3.1	7.5	S	5.1	12.6	73.0	16.6	12.6	15.7	6.4	8.7	7.6	6.8	11.1	3.9	3.4	4.8	8.6	9.6	11.1	12.3	2.4	73.0	10.8	24
23	15.3	13.4	10.8	13.0	S	21.2	39.5	48.6	15.4	17.0	4.5	4.0	4.6	2.2	3.6	4.0	5.0	13.4	8.4	15.9	9.2	5.4	8.1	5.8	2.2	48.6	12.5	24
24	8.3	6.2	22.4	S	32.1	9.7	26.2	11.5	10.5	6.1	5.7	15.9	3.5	5.3	4.4	7.9	5.3	4.6	3.3	4.2	3.2	3.2	3.3	4.0	3.2	32.1	9.0	24
25	5.1	4.7	S	3.6	4.8	8.6	5.7	4.1	3.2	4.1	7.5	5.4	5.0	6.0	7.9	8.3	3.6	2.7	13.5	18.8	29.0	33.6	32.9	14.3	2.7	33.6	10.1	24
26	4.6	S	5.0	10.5	10.3	12.9	24.1	20.7	7.2	7.8	4.5	4.2	5.4	5.7	2.1	2.1	3.6	2.3	2.6	6.9	15.0	24.1	27.3	16.4	2.1	27.3	9.8	24
27	S	16.2	36.3	15.1	9.5	7.3	5.4	6.3	3.5	2.7	2.8	2.4	2.2	2.7	3.2	6.9	6.2	4.2	3.6	7.6	9.0	1.9	1.9	S	1.9	36.3	7.1	24
28	2.9	3.0	2.1	1.5	2.6	6.8	3.2	8.7	3.8	4.4	3.7	3.0	2.7	2.2	4.9	4.9	2.6	2.3	3.1	3.9	2.4	3.5	S	6.4	1.5	8.7	3.7	24
29	10.8	4.2	5.9	2.8	11.1	13.9	12.0	8.2	9.6	3.7	3.3	3.5	4.0	12.6	4.5	3.6	4.4	7.6	11.0	9.2	4.1	S	3.0	4.1	2.8	13.9	6.8	24
30	7.9	4.0	3.6	3.5	4.1	6.3	7.8	8.2	3.7	5.7	4.6	4.6	3.6	3.0	5.1	4.9	3.1	6.9	6.9	12.5	S	15.3	20.9	13.4	3.0	20.9	6.9	24
31	22.4	18.2	5.3	20.1	3.2	8.3	11.9	P	6.3	9.4	7.8	5.8	5.4	6.7	5.4	2.8	2.3	2.1	13.9	S	3.6	9.1	3.6	2.3	2.1	22.4	8.0	23
HOURLY MAX	59.8	33.4	36.3	37.0	32.1	100.9	39.5	48.6	73.0	20.3	23.0	21.3	10.7	15.7	21.5	12.9	15.8	110.5	42.4	43.4	29.0	36.3	32.9	27.3				
HOURLY AVG	9.3	8.0	8.0	8.3	9.0	13.6	10.4	12.1	15.0	7.8	8.0	6.3	4.5	5.3	6.4	5.6	5.2	9.3	8.0	9.2	9.5	9.5	8.8	7.4				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY


NUMBER OF NON-ZERO READINGS:	701
MAXIMUM INSTANTANEOUS VALUE:	110.5 ppb @ HOUR(S) 17 ON DAY(S) 9
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	7 hrs
OPERATIONAL TIME:	741 hrs
STANDARD DEVIATION:	9.7

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)




% Icon Classes (ppb) 83 0.0-8.8 4 8.8-17.7 1 17.7-26.5 0 >26.5

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-NOX[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 12.13% Calm Poll Avg: 8.69[ppb]



NOX[ppb] Calibration: LICA COLD LAKE SOUTH Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

NITRIC OXIDES


NITRIC OXIDE Hourly Averages (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	0.3	0.4	0.4	S	0.9	1.0	0.9	0.9	2.2	1.2	1.1	1.1	0.4	0.6	0.6	0.5	0.3	0.4	0.3	0.3	0.8	0.3	0.5	0.1	0.1	2.2	0.7	24	
2	0.0	0.0	S	0.0	0.1	0.0	0.2	2.0	6.8	1.9	1.6	1.9	1.5	1.5	1.6	1.7	1.4	1.3	0.7	0.2	0.1	0.2	0.1	0.2	0.0	6.8	1.1	24	
3	0.4	S	0.4	0.4	1.0	1.8	2.7	1.2	6.1	4.3	2.5	1.6	1.0	1.6	0.9	0.8	0.6	0.4	0.2	1.4	0.3	0.5	0.1	0.2	0.1	6.1	1.3	24	
4	S	0.3	0.3	0.2	0.1	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.4	0.6	0.7	0.3	0.2	0.3	0.2	0.2	0.1	S	0.1	0.7	0.3	24	
5	0.5	0.5	0.3	0.3	0.4	0.6	0.5	0.8	0.7	0.7	0.6	0.4	0.5	0.5	0.3	0.3	0.3	0.2	0.3	0.3	0.5	0.4	S	0.3	0.2	0.8	0.4	24	
6	0.2	0.5	0.4	0.2	0.2	0.4	0.4	0.7	0.2	0.5	C	C	C	C	C	C	C	0.2	0.1	0.0	0.0	S	0.0	0.0	0.0	0.7	0.3	24	
7	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.0	0.4	0.6	0.9	0.8	0.7	0.8	0.6	0.5	0.3	0.1	0.1	S	0.1	0.1	0.1	0.1	0.0	1.0	0.3	24	
8	0.1	0.4	0.3	0.4	0.2	0.0	0.3	0.5	1.1	0.7	0.4	0.8	0.7	0.6	1.2	0.6	0.6	0.8	0.0	S	0.3	0.1	0.5	0.6	0.0	1.2	0.5	24	
9	1.5	1.0	0.7	1.0	0.2	0.0	0.2	3.1	8.9	4.0	1.4	0.9	1.5	1.0	1.7	1.2	1.0	3.1	S	0.2	1.0	2.1	1.3	2.0	0.0	8.9	1.7	24	
10	2.1	4.6	1.4	0.9	2.7	3.0	1.3	3.1	2.3	1.7	2.0	1.1	1.2	0.9	1.0	0.8	0.5	S	0.3	0.1	0.7	0.6	0.5	0.2	0.1	4.6	1.4	24	
11	0.2	0.3	0.5	0.2	1.1	2.0	0.5	X	X	0.5	1.0	0.6	0.6	0.5	1.7	0.8	S	0.7	0.5	0.4	0.5	0.4	0.5	0.3	0.2	2.0	0.7	22	
12	0.4	0.3	0.3	0.3	0.7	0.3	0.3	0.5	0.7	0.7	0.8	0.6	0.4	0.5	0.3	S	0.7	0.6	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.8	0.4	24	
13	0.0	0.1	0.1	2.2	0.1	0.1	0.1	0.3	0.3	0.4	0.8	0.6	0.3	0.2	S	0.2	0.2	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	2.2	0.3	24	
14	0.0	0.2	0.1	0.2	0.1	0.5	0.3	1.4	1.4	0.9	1.7	1.2	1.7	S	1.6	1.1	1.2	0.8	0.2	0.3	0.2	0.1	0.2	0.1	0.0	1.7	0.7	24	
15	0.2	0.1	0.1	0.2	0.1	0.2	0.4	0.8	0.8	1.0	1.0	1.4	S	2.4	2.3	1.6	1.0	0.6	1.0	1.1	0.2	0.1	0.0	0.0	0.0	2.4	0.7	24	
16	0.0	0.1	0.3	0.2	0.1	1.8	0.1	2.3	1.2	1.0	0.6	S	0.3	0.2	0.2	0.1	0.2	0.4	0.2	0.2	0.1	0.2	0.1	0.2	0.0	2.3	0.4	24	
17	0.3	0.2	0.1	0.1	0.2	0.3	0.2	0.3	0.7	0.7	S	0.6	0.3	0.4	0.3	0.4	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.7	0.3	24
18	0.1	0.1	0.1	0.1	0.2	0.2	0.3	0.5	0.6	S	0.4	0.4	0.3	0.2	0.3	0.4	0.3	0.4	0.4	0.2	0.2	0.2	0.1	0.1	0.1	0.6	0.3	24	
19	0.2	0.1	0.1	0.0	0.0	0.0	0.1	0.1	S	0.4	0.4	0.3	0.3	0.2	0.2	0.1	0.2	0.1	0.1	0.1	0.1	0.2	0.1	0.0	0.0	0.4	0.1	24	
20	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	1.8	0.2	0.5	0.1	0.1	0.1	0.2	0.2	0.3	0.2	0.0	0.1	0.1	0.2	0.1	0.2	0.1	1.8	0.2	24	
21	1.4	0.4	0.4	0.3	0.3	0.2	S	3.1	1.2	2.4	1.8	1.0	1.5	0.5	1.0	0.3	0.2	0.1	0.1	0.2	0.2	0.2	0.1	0.2	0.1	3.1	0.7	24	
22	0.1	0.2	0.1	0.2	0.5	S	0.3	1.5	0.8	1.9	1.2	1.0	0.8	1.1	1.2	0.9	0.9	0.2	0.1	0.3	0.4	0.3	0.4	0.3	0.1	1.9	0.6	24	
23	0.5	0.4	0.2	0.1	S	1.2	3.6	9.6	4.0	1.6	0.6	0.4	0.4	0.3	0.5	0.6	0.6	0.8	0.4	0.4	0.1	0.0	0.2	0.1	0.0	9.6	1.2	24	
24	0.1	0.1	0.6	S	1.1	0.2	2.2	1.4	1.3	1.3	1.1	0.8	0.6	0.8	0.9	1.0	0.6	0.3	0.2	0.2	0.2	0.2	0.2	0.3	0.1	2.2	0.7	24	
25	0.2	0.3	S	0.2	0.4	0.4	0.1	0.3	0.2	0.4	1.3	1.3	0.8	1.2	0.8	0.5	0.4	0.2	0.1	0.2	0.5	1.8	2.6	0.3	0.1	2.6	0.6	24	
26	0.1	S	0.2	0.2	0.3	0.2	1.9	2.2	0.7	1.0	0.9	0.7	0.5	0.4	0.2	0.1	0.1	0.1	0.0	0.1	0.4	1.0	0.3	0.0	2.2	0.5	24		
27	S	0.4	1.1	0.4	0.6	0.4	0.4	0.7	0.4	0.3	0.3	0.3	0.3	0.3	0.3	0.6	0.6	0.2	0.1	0.2	0.1	0.1	0.1	S	0.1	1.1	0.4	24	
28	0.1	0.1	0.1	0.0	0.1	0.2	0.2	0.3	0.3	0.3	0.5	0.4	0.3	0.3	0.5	0.4	0.2	0.1	0.1	0.0	0.1	S	0.1	0.0	0.5	0.2	24		
29	0.2	0.1	0.1	0.1	0.7	0.8	0.4	0.6	0.6	0.4	0.4	0.3	0.4	0.9	0.7	0.7	0.6	0.6	0.7	0.3	0.1	S	0.3	0.2	0.1	0.9	0.4	24	
30	0.4	0.4	0.3	0.3	0.4	0.4	0.6	0.6	0.4	0.5	0.5	0.4	0.3	0.3	0.3	0.5	0.3	0.3	0.2	0.3	S	0.5	1.6	0.4	0.2	1.6	0.4	24	
31	2.9	1.2	0.1	3.2	0.2	0.4	0.9	1.2	1.6	2.1	2.1	1.7	1.4	1.4	0.4	0.3	0.2	0.2	0.7	S	0.3	0.5	0.1	0.0	0.0	3.2	1.0	24	
HOURLY MAX	2.9	4.6	1.4	3.2	2.7	3.0	3.6	9.6	8.9	4.3	2.5	1.9	1.7	2.4	2.3	1.7	1.4	3.1	1.0	1.4	1.0	2.1	2.6	2.0					
HOURLY AVG	0.4	0.4	0.3	0.4	0.4	0.6	0.7	1.4	1.7	1.1	1.0	0.8	0.7	0.7	0.8	0.6	0.5	0.5	0.3	0.3	0.3	0.3	0.4	0.2					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	650				
MINIMUM 1-HR AVERAGE:	0.0 ppb	@ HOUR(S)	0	ON DAY(S)	2
MAXIMUM 1-HR AVERAGE:	9.6 ppb	@ HOUR(S)	7	ON DAY(S)	23
MAXIMUM 24-HR AVERAGE:	1.7 ppb			ON DAY(S)	9
				VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	742 hrs		
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	99.7 %		
STANDARD DEVIATION:	0.9	MONTHLY AVERAGE:	0.6 ppb		

NITRIC OXIDE Hourly Averages (NO ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake Continuous Monitoring Station - March 2017

NITRIC OXIDE Instantaneous Maximum (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	1.2	1.1	2.1	S	3.0	3.0	2.7	2.7	27.9	7.0	5.3	5.1	2.3	2.9	2.3	4.3	1.7	1.6	1.7	2.7	4.4	1.4	2.1	1.1	1.1	27.9	3.9	24	
2	0.7	0.5	S	0.7	0.7	0.7	1.4	7.6	25.2	20.8	4.5	6.3	3.4	4.0	3.9	2.7	3.1	3.7	1.8	0.7	0.5	0.9	0.7	0.5	0.5	25.2	4.1	24	
3	1.8	S	1.6	1.6	5.7	18.5	7.5	3.4	24.0	8.7	7.3	3.9	2.2	2.6	2.5	1.8	1.3	1.3	1.3	10.6	1.4	1.8	0.8	1.8	0.8	24.0	4.9	24	
4	S	1.3	1.3	1.4	0.5	1.2	0.5	1.2	0.9	0.8	1.7	0.8	0.9	1.6	1.2	6.7	3.8	1.7	1.3	1.0	2.7	1.3	0.5	S	0.5	6.7	1.6	24	
5	2.7	2.6	1.8	2.1	1.6	2.1	2.7	3.1	3.9	2.9	1.6	2.1	1.6	5.9	0.9	0.9	3.5	0.5	2.3	1.4	3.9	3.4	S	2.3	0.5	5.9	2.4	24	
6	1.7	2.0	1.8	2.0	0.8	1.7	1.8	4.4	0.7	2.8	C	C	C	C	C	C	C	1.3	1.1	0.3	0.3	S	0.4	0.3	0.3	4.4	1.5	24	
7	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.3	6.0	0.8	0.9	2.5	1.3	5.1	3.0	1.1	1.1	0.7	0.5	0.7	S	0.8	0.7	0.7	0.2	6.0	1.2	24	
8	0.9	1.2	1.1	1.2	0.8	0.8	0.9	1.5	2.7	2.4	1.5	2.8	2.6	1.8	8.4	2.3	2.2	3.0	1.3	S	1.7	1.1	2.7	3.9	0.8	8.4	2.1	24	
9	16.1	4.0	2.4	4.9	2.2	0.4	1.3	5.4	16.3	10.7	6.2	2.4	3.1	1.9	3.5	3.3	1.8	49.1	S	1.2	9.7	7.2	4.5	5.7	0.4	49.1	7.1	24	
10	6.0	9.3	4.1	2.3	8.1	9.1	3.2	8.6	4.0	2.8	7.0	3.4	4.8	2.3	2.7	2.4	1.7	S	1.4	0.7	3.0	1.8	1.6	1.0	0.7	9.3	4.0	24	
11	0.9	1.6	3.9	2.2	5.0	26.6	1.7	X	X	1.6	5.1	2.9	1.8	1.7	18.9	2.2	S	2.1	1.4	1.6	2.3	1.2	1.7	2.7	0.9	26.6	4.2	22	
12	2.2	1.0	1.2	1.2	2.9	1.6	1.3	1.4	2.3	1.6	2.5	1.6	1.1	1.6	1.1	S	4.4	3.4	0.4	0.4	0.4	0.3	0.4	0.5	0.3	4.4	1.5	24	
13	0.4	0.5	0.5	20.1	0.5	1.6	0.4	0.7	1.7	1.1	2.7	6.2	1.3	1.1	S	1.2	1.4	1.4	1.4	0.5	0.4	0.4	0.5	0.5	0.4	20.1	2.0	24	
14	0.7	4.2	0.7	0.9	0.8	4.9	2.2	2.6	2.5	1.6	5.5	1.8	4.0	S	3.0	2.0	5.1	2.1	0.5	0.7	0.7	0.5	1.2	0.9	0.5	5.5	2.1	24	
15	1.2	1.1	0.5	0.6	0.6	1.4	1.4	2.2	1.8	1.7	2.5	2.1	S	3.4	3.4	3.0	2.3	1.4	10.9	10.9	1.3	0.9	0.7	0.4	0.4	10.9	2.4	24	
16	0.7	1.1	3.1	1.2	0.9	36.5	0.3	8.1	7.6	2.4	2.6	S	2.1	0.7	2.0	0.5	0.5	5.2	1.7	1.1	0.7	1.3	0.7	2.0	0.3	36.5	3.6	24	
17	2.0	0.8	0.7	0.3	0.8	1.0	0.7	1.0	1.6	1.7	S	1.6	0.7	2.2	1.4	2.9	0.3	0.4	0.5	2.0	0.4	0.5	0.4	0.5	0.3	2.9	1.1	24	
18	0.5	0.4	0.6	0.4	0.9	0.7	0.9	3.7	2.7	S	1.0	1.4	1.2	0.5	1.2	1.6	0.8	1.3	2.1	1.0	0.8	2.0	0.5	0.4	0.4	3.7	1.2	24	
19	1.2	0.7	0.7	0.4	0.4	0.4	0.9	0.5	S	0.8	0.8	0.7	0.8	0.6	0.7	0.5	1.4	0.8	0.6	0.7	0.6	0.8	0.8	0.3	0.3	1.4	0.7	24	
20	0.3	0.1	0.2	0.2	0.1	0.2	0.1	S	10.3	0.8	4.2	0.5	0.5	0.5	1.1	1.2	1.4	0.7	1.0	0.3	0.7	1.6	1.6	0.7	0.1	10.3	1.2	24	
21	31.1	1.4	2.3	1.1	2.1	1.3	S	44.5	8.1	16.8	20.2	8.7	21.9	3.5	4.5	1.7	1.1	0.9	0.3	1.8	1.8	0.7	0.5	0.8	0.3	44.5	7.7	24	
22	0.7	0.9	0.5	0.7	2.0	S	1.4	15.3	9.1	20.6	10.7	9.1	2.3	2.6	4.3	3.3	16.2	0.9	0.5	1.2	2.7	1.4	1.8	1.2	0.5	20.6	4.8	24	
23	2.5	1.3	1.3	0.7	S	4.5	18.4	27.9	7.9	8.3	1.7	1.7	2.1	0.8	1.2	1.8	2.9	7.7	1.4	4.3	0.9	0.4	1.8	0.8	0.4	27.9	4.4	24	
24	0.5	0.4	7.3	S	12.7	1.4	10.4	3.0	4.5	3.0	2.0	3.4	1.3	1.4	1.3	2.9	1.6	0.9	0.6	0.9	0.9	0.8	0.8	0.8	0.4	12.7	2.7	24	
25	2.0	1.4	S	0.8	1.7	2.2	0.4	0.8	0.7	1.4	2.2	2.0	1.6	2.2	3.8	2.9	2.1	0.4	1.4	2.2	3.3	7.1	11.2	0.9	0.4	11.2	2.4	24	
26	0.9	S	1.1	1.2	1.2	8.3	6.4	2.0	3.4	2.3	1.8	2.1	2.6	0.7	0.4	1.4	0.3	0.3	0.4	1.3	3.3	4.3	2.1	0.3	8.3	2.1	24		
27	S	3.7	15.6	2.6	2.1	1.6	1.2	1.4	1.1	0.9	0.9	0.9	0.7	1.1	1.8	1.6	5.5	0.9	0.7	1.2	0.9	0.4	0.4	S	0.4	15.6	2.1	24	
28	0.8	0.7	0.4	0.4	0.7	2.0	2.3	2.5	0.8	0.8	1.7	1.0	0.8	0.6	2.9	4.9	0.7	0.4	0.5	0.7	0.3	0.9	S	0.5	0.3	4.9	1.2	24	
29	2.0	0.3	0.9	0.4	6.7	3.4	2.1	3.0	3.8	1.1	1.8	0.8	0.9	3.7	1.8	1.1	1.3	2.3	4.4	2.5	0.8	S	1.1	1.2	0.3	6.7	2.1	24	
30	3.1	1.4	1.4	1.6	1.2	2.3	3.1	3.0	1.1	1.8	2.3	1.4	1.0	1.3	1.4	2.3	0.8	2.1	0.9	3.7	S	2.9	6.9	2.7	0.8	6.9	2.2	24	
31	8.0	4.2	0.7	8.4	0.8	4.0	4.4	P	2.7	5.2	3.1	2.1	1.7	2.2	1.2	3.9	0.4	0.4	14.4	S	1.3	4.0	0.4	0.3	0.3	14.4	3.4	23	
HOURLY MAX	31.1	9.3	15.6	20.1	12.7	36.5	18.4	44.5	27.9	20.8	20.2	9.1	21.9	5.9	18.9	6.7	16.2	49.1	14.4	10.9	9.7	7.2	11.2	5.7					
HOURLY AVG	3.2	1.7	2.1	2.1	2.3	4.6	2.8	5.9	6.3	4.5	3.9	2.8	2.5	2.2	3.0	2.3	2.5	3.3	2.0	2.0	1.7	1.8	1.8	1.3					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	701
MAXIMUM INSTANTANEOUS VALUE:	49.1 ppb @ HOUR(S) 17 ON DAY(S) 9
VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	7 hrs
STANDARD DEVIATION:	4.7
OPERATIONAL TIME:	741 hrs

NITRIC OXIDE Instantaneous Maximum (NO ppb)




Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-NO[ppb]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 12.13% Calm Avg: 0.78 [ppb]

Direction	0.0-3.2	3.2-6.5	6.5-9.7	>9.7	Total
N	5.7	0.0	0.0	0.0	5.7
NE	15.7	0.6	0.1	0.0	16.4
E	21.5	0.1	0.1	0.0	21.8
SE	27.1	0.0	0.0	0.0	27.1
S	1.0	0.0	0.0	0.0	1.0
SW	4.4	0.0	0.0	0.0	4.4
W	6.3	0.0	0.0	0.0	6.3
NW	5.1	0.0	0.0	0.0	5.1
Summary	86.9	0.7	0.3	0.0	87.9

% Icon Classes (ppb) 87 0.0-3.2 1 3.2-6.5 0 6.5-9.7 0 >9.7

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-NO[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 12.13% Calm Poll Avg: 0.78[ppb]



NITROGEN DIOXIDE

NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	2.4	2.8	2.3	S	2.8	3.5	3.2	2.4	2.0	1.8	1.8	1.8	0.9	0.8	1.1	1.2	1.0	2.0	2.9	4.3	7.2	8.3	9.4	3.2	0.8	9.4	3.0	24
2	1.5	0.9	S	1.4	1.6	1.9	4.8	13.3	11.2	2.4	3.0	2.4	2.1	2.0	2.5	3.4	3.7	7.6	7.8	3.4	2.7	2.7	2.4	2.5	0.9	13.3	3.8	24
3	3.4	S	4.5	4.7	9.8	17.5	23.7	10.4	12.8	6.7	3.5	2.5	2.1	3.1	2.7	2.7	3.1	4.3	8.7	20.1	11.9	9.1	3.8	3.2	2.1	23.7	7.6	24
4	S	3.4	2.2	5.0	3.6	1.7	1.3	1.1	1.1	1.1	1.2	1.0	1.0	1.3	1.3	1.8	1.3	1.4	1.6	1.4	1.7	1.5	S	1.0	5.0	1.7	24	
5	1.5	2.0	1.5	1.7	1.5	1.6	2.0	2.6	2.0	1.7	1.3	0.9	0.8	0.9	0.8	1.0	1.0	1.3	4.0	4.9	3.5	2.2	S	2.2	0.8	4.9	1.9	24
6	1.6	1.5	1.2	1.4	1.2	1.6	2.1	4.2	2.2	2.0	C	C	C	C	C	C	C	1.9	1.9	1.5	2.4	S	2.3	1.6	1.2	4.2	1.9	24
7	1.2	1.5	1.6	1.7	1.5	1.4	1.8	1.8	2.5	1.6	1.5	1.6	1.3	1.2	1.5	1.6	2.0	2.3	2.6	2.9	S	2.5	2.5	2.3	1.2	2.9	1.8	24
8	2.5	4.3	2.2	1.7	1.8	1.3	3.8	2.0	2.1	1.1	0.9	1.0	0.9	0.9	1.3	1.2	1.8	3.9	4.8	S	12.3	8.6	3.7	5.8	0.9	12.3	3.0	24
9	7.9	7.9	6.4	9.5	10.5	7.2	11.5	11.4	10.9	4.0	0.9	0.8	1.3	1.1	1.7	1.6	1.9	4.9	S	10.2	13.6	19.5	18.4	18.2	0.8	19.5	7.9	24
10	19.9	21.2	15.4	12.9	15.1	9.1	5.0	6.7	2.9	1.7	2.2	0.9	1.0	1.2	1.9	3.0	3.1	S	4.9	9.3	7.0	3.5	4.4	3.1	0.9	21.2	6.8	24
11	2.2	4.0	8.8	9.5	11.0	5.3	1.6	X	X	1.1	1.5	1.0	0.9	0.9	1.5	1.7	S	3.0	4.3	3.7	3.0	2.9	2.7	2.7	0.9	11.0	3.5	22
12	3.0	2.4	2.4	2.5	3.5	2.6	2.4	2.2	1.8	1.6	1.4	1.1	0.9	1.0	0.9	S	2.0	2.6	1.3	1.6	1.3	1.2	1.2	1.8	0.9	3.5	1.9	24
13	1.8	1.7	1.6	4.2	1.9	1.8	1.9	1.9	1.5	1.6	2.2	1.2	1.2	1.1	S	1.7	1.6	1.8	2.2	2.2	2.4	2.5	2.7	3.2	1.1	4.2	2.0	24
14	2.8	5.8	3.7	4.9	3.2	4.4	4.7	8.0	4.4	2.7	3.4	2.8	3.9	S	4.9	5.1	6.2	7.5	4.2	3.7	3.3	2.8	2.4	2.3	2.3	8.0	4.2	24
15	2.6	2.1	2.0	2.0	2.4	3.6	6.0	4.6	3.1	3.0	2.5	3.0	S	5.4	5.8	6.7	7.0	7.3	11.9	16.1	11.0	7.2	2.0	1.2	1.2	16.1	5.2	24
16	1.9	3.3	4.9	4.3	3.9	10.2	3.8	9.4	4.2	3.8	1.9	S	1.3	1.1	0.9	0.9	1.2	2.1	2.7	2.8	2.4	2.3	4.2	3.8	0.9	10.2	3.4	24
17	2.6	2.4	0.9	0.7	2.1	3.1	2.1	2.5	3.5	2.2	S	1.8	1.3	1.4	1.4	1.4	1.4	1.9	1.8	1.8	1.8	1.5	1.5	1.3	0.7	3.5	1.8	24
18	1.2	1.3	1.2	1.1	1.4	1.4	2.1	2.5	2.1	S	1.3	1.3	1.0	1.0	1.1	1.3	1.4	2.4	3.3	2.5	2.8	2.0	1.9	2.4	1.0	3.3	1.7	24
19	1.8	2.3	1.3	1.5	0.9	0.8	1.1	1.0	S	1.5	1.3	1.1	1.0	0.9	1.0	1.1	1.5	1.0	1.4	1.2	1.3	2.1	2.3	2.1	0.8	2.3	1.4	24
20	1.1	1.0	1.3	0.8	0.9	0.9	0.9	S	2.1	0.9	1.2	0.7	0.7	0.8	0.9	1.0	1.1	1.2	1.5	2.4	3.2	3.7	4.7	3.3	0.7	4.7	1.6	24
21	8.1	9.8	5.1	4.5	4.0	6.7	S	4.3	3.0	1.9	1.6	1.6	0.5	1.7	1.3	1.5	1.3	1.2	1.3	1.1	1.3	1.6	1.6	1.5	0.5	9.8	2.9	24
22	1.5	1.5	1.4	1.4	2.4	S	3.0	4.0	4.4	2.0	2.7	3.1	2.8	4.2	2.9	2.1	2.0	1.7	1.5	2.9	4.2	6.7	8.6	10.1	1.4	10.1	3.4	24
23	11.7	11.2	8.9	9.8	S	13.1	14.7	12.7	6.1	2.8	1.7	1.3	1.3	1.2	1.7	2.0	2.1	3.5	5.1	7.7	4.8	3.5	4.0	2.8	1.2	14.7	5.8	24
24	4.7	3.9	5.5	S	7.4	6.3	9.3	5.2	3.3	2.9	2.7	2.3	1.8	2.0	2.5	3.2	2.6	2.3	2.0	1.8	1.7	1.8	2.1	2.0	1.7	9.3	3.4	24
25	1.7	1.8	S	2.2	2.2	3.6	2.5	2.6	1.4	1.6	3.2	2.7	2.1	3.0	1.9	2.0	2.0	2.1	3.0	5.9	13.7	19.9	17.5	6.5	1.4	21.9	4.7	24
26	2.4	S	2.0	3.6	6.4	8.3	10.5	6.9	2.3	2.3	2.1	1.8	1.7	1.6	1.3	1.2	1.3	1.5	1.8	3.2	4.5	12.1	19.0	7.8	1.2	19.0	4.6	24
27	S	8.4	12.4	8.6	4.3	4.1	3.1	2.9	1.5	1.2	1.1	1.0	1.2	1.2	1.3	2.2	2.1	2.0	2.0	3.0	3.1	1.1	1.1	S	1.0	12.4	3.1	24
28	1.6	1.6	1.0	0.8	1.2	2.2	1.8	2.7	1.8	1.6	1.7	1.4	1.5	1.4	1.5	1.3	1.5	1.9	2.2	2.3	2.0	1.9	S	3.2	0.8	3.2	1.7	24
29	4.6	2.6	2.4	1.6	4.9	7.1	4.9	4.2	2.8	1.7	1.5	1.6	1.9	3.5	2.1	2.2	2.3	3.5	5.3	3.6	2.1	S	1.7	1.6	1.5	7.1	3.0	24
30	1.5	1.7	1.9	1.7	2.1	2.4	3.0	3.2	2.1	2.0	1.6	1.7	1.5	1.1	1.1	1.3	1.1	1.6	2.4	5.8	S	6.0	10.4	6.7	1.1	10.4	2.8	24
31	11.0	6.8	3.2	7.3	1.4	2.0	4.1	3.3	3.0	3.1	3.7	3.5	3.4	4.1	2.1	1.0	1.4	1.2	2.0	S	2.1	2.6	1.9	1.9	1.0	11.0	3.3	24
HOURLY MAX	19.9	21.2	15.4	12.9	15.1	17.5	23.7	13.3	12.8	6.7	3.7	3.5	3.9	5.4	5.8	6.7	7.0	7.6	11.9	20.1	13.7	21.9	19.0	18.2				
HOURLY AVG	3.9	4.2	3.8	3.9	3.9	4.6	4.8	4.8	3.6	2.2	2.0	1.7	1.5	1.8	1.8	2.0	2.2	2.8	3.4	4.6	4.6	5.0	4.9	3.8				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 159 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0				
NUMBER OF NON-ZERO READINGS:	702				
MINIMUM 1-HR AVERAGE:	0.5 ppb	@ HOUR(S)	12	ON DAY(S)	21
MAXIMUM 1-HR AVERAGE:	23.7 ppb	@ HOUR(S)	6	ON DAY(S)	3
MAXIMUM 24-HR AVERAGE:	7.9 ppb			ON DAY(S)	9
				VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	742 hrs		
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	99.7 %		
STANDARD DEVIATION:	3.4	MONTHLY AVERAGE:	3.4 ppb		

24 HR AVERAGES March 2017



NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake Continuous Monitoring Station - March 2017

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	3.4	4.2	3.3	S	5.5	7.0	4.7	4.0	5.9	3.2	7.9	17.6	3.0	3.3	6.4	3.8	2.4	3.0	5.5	8.1	11.0	12.1	15.3	7.2	2.4	17.6	6.4	24	
2	2.6	1.3	S	2.2	3.1	4.9	12.6	16.9	24.7	14.1	21.2	5.5	3.6	3.8	5.5	5.1	5.5	10.3	10.5	4.6	3.3	3.4	2.9	3.0	1.3	24.7	7.4	24	
3	7.1	S	6.4	7.9	16.8	26.6	30.9	13.8	24.5	12.1	7.5	3.8	4.0	4.4	10.8	5.2	3.6	5.6	12.9	32.6	25.8	12.9	7.9	9.5	3.6	32.6	12.7	24	
4	S	6.8	3.7	6.0	5.2	2.2	1.7	1.7	2.1	1.7	8.4	1.7	1.8	1.6	2.0	4.0	3.8	3.7	3.1	2.7	3.6	3.9	2.5	S	1.6	8.4	3.4	24	
5	4.2	5.2	2.9	5.1	2.5	3.6	3.6	5.3	8.6	3.6	2.5	2.2	1.4	4.8	1.4	2.2	2.2	2.6	11.0	7.2	5.9	5.7	S	7.1	1.4	11.0	4.4	24	
6	3.9	5.0	2.6	3.8	2.3	3.5	4.2	8.4	2.9	5.5	C	C	C	C	C	C	C	3.6	2.9	2.1	2.7	S	2.7	1.9	1.9	8.4	3.6	24	
7	1.4	2.0	1.7	1.9	1.8	1.8	2.0	2.1	8.2	1.9	1.7	3.5	1.8	5.1	3.8	2.2	2.2	2.7	5.4	5.0	S	4.8	3.9	3.6	1.4	8.2	3.1	24	
8	3.6	7.0	2.9	2.5	6.0	2.5	6.2	3.8	3.2	2.4	2.8	2.2	2.5	1.9	6.6	4.1	5.1	7.6	8.3	S	19.3	12.0	9.1	8.9	1.9	19.3	5.7	24	
9	12.3	11.0	8.9	14.2	14.5	11.0	18.4	15.0	17.7	9.1	2.2	2.0	2.7	1.5	3.1	2.9	2.7	63.7	S	17.0	20.1	29.0	23.3	22.7	1.5	63.7	14.1	24	
10	23.5	24.2	18.8	16.3	20.3	16.5	8.0	12.5	4.1	3.1	12.2	2.0	2.6	3.3	3.1	4.6	4.4	S	6.8	14.3	12.6	5.6	5.9	4.7	2.0	24.2	10.0	24	
11	3.3	6.5	14.8	16.1	20.4	24.2	3.3	X	X	2.5	2.9	2.9	1.4	2.3	6.2	3.6	S	4.4	6.6	5.2	5.5	4.0	7.8	4.4	1.4	24.2	7.1	22	
12	5.2	3.6	3.5	3.4	9.4	3.8	5.1	3.3	3.8	2.9	2.7	1.6	1.5	2.5	2.0	S	4.4	5.9	2.1	2.3	2.1	1.4	1.6	2.6	1.4	9.4	3.3	24	
13	2.1	2.1	2.0	18.6	2.7	2.4	2.5	2.7	2.2	2.1	6.6	3.0	4.3	1.3	S	4.9	3.5	2.5	6.8	2.6	2.7	3.1	3.3	4.9	1.3	18.6	3.9	24	
14	4.4	13.0	5.6	7.6	6.4	10.3	8.8	11.6	8.0	3.5	10.1	3.4	6.8	S	7.0	6.6	11.4	12.4	7.0	4.6	4.4	4.3	3.5	3.6	3.4	13.0	7.1	24	
15	3.8	3.1	2.6	2.6	3.5	6.0	11.0	9.9	4.9	3.8	3.8	3.5	S	6.1	8.9	10.0	10.8	12.0	32.9	26.1	23.4	12.0	2.7	1.8	1.8	32.9	8.9	24	
16	4.3	6.1	14.6	7.8	7.9	64.7	5.9	19.3	16.1	6.3	3.6	S	3.3	1.7	2.0	1.7	1.7	3.3	6.4	5.8	6.5	3.9	8.7	10.0	1.7	64.7	9.2	24	
17	4.7	3.4	2.0	1.0	4.9	5.8	5.9	3.9	5.9	3.8	S	2.7	1.7	2.1	2.3	2.3	1.7	2.5	3.0	2.3	2.0	1.9	2.3	1.0	5.9	3.1	24		
18	1.6	1.7	1.9	1.7	2.3	2.1	3.8	5.6	5.9	S	2.1	1.8	1.3	1.2	2.1	2.9	2.2	3.6	5.2	3.2	4.1	5.0	2.6	3.5	1.2	5.9	2.9	24	
19	2.5	2.7	1.7	1.7	1.4	1.3	2.2	1.4	S	1.8	1.7	1.2	1.3	1.3	1.3	1.3	2.9	2.1	2.4	1.7	1.8	3.1	3.5	2.6	1.2	3.5	2.0	24	
20	2.1	1.5	1.6	1.2	1.0	1.2	1.5	S	9.6	1.3	5.2	0.9	1.0	1.9	1.9	5.2	2.4	2.1	2.3	3.9	6.3	13.0	13.3	5.6	0.9	13.3	3.7	24	
21	28.6	18.5	14.4	5.8	12.6	11.6	S	13.5	12.9	7.7	8.4	6.6	3.2	12.1	5.9	3.1	1.8	1.9	2.4	1.4	1.8	2.2	2.0	2.1	1.4	28.6	7.8	24	
22	2.6	2.3	2.0	2.5	5.6	S	4.2	7.6	63.9	6.5	7.0	7.9	4.2	6.6	5.5	4.4	4.3	2.9	2.7	4.4	7.9	8.7	10.3	11.4	2.0	63.9	8.1	24	
23	13.6	12.5	10.1	12.5	S	17.0	22.6	20.9	8.5	8.9	2.9	2.4	2.6	1.4	2.5	2.5	3.5	3.5	9.7	7.2	12.5	9.1	5.0	7.5	4.9	1.4	22.6	8.7	24
24	8.0	6.0	15.1	S	24.4	8.5	15.7	8.4	5.9	4.2	3.6	13.0	2.3	4.2	3.0	5.0	3.9	3.9	2.9	3.2	2.3	2.7	3.0	3.1	2.3	24.4	6.6	24	
25	3.5	3.4	S	3.0	3.2	6.4	5.2	3.5	2.4	3.0	5.4	3.5	3.5	4.0	5.1	5.6	2.5	2.2	12.1	16.5	26.0	27.8	25.8	14.0	2.2	27.8	8.2	24	
26	3.8	S	4.0	9.5	9.2	12.5	17.0	14.2	5.2	5.0	3.1	2.6	3.4	3.3	1.5	2.3	1.9	2.2	6.7	14.4	20.9	23.4	15.6	1.5	23.4	8.0	24		
27	S	14.4	20.9	12.5	7.9	5.9	4.2	4.9	2.7	2.0	1.8	1.5	1.5	1.6	2.2	5.5	3.8	3.4	3.1	6.7	8.0	1.4	1.5	S	1.4	20.9	5.3	24	
28	2.6	2.3	1.7	1.2	2.0	4.8	2.6	6.5	3.1	3.5	2.1	1.9	2.0	1.6	2.6	2.4	1.8	1.9	2.6	3.1	2.1	2.6	S	6.1	1.2	6.5	2.7	24	
29	8.8	3.9	5.0	2.5	8.0	10.5	9.9	6.4	6.0	2.7	2.0	2.7	2.9	9.1	2.9	2.5	3.1	5.4	8.0	8.1	3.4	S	2.2	3.3	2.0	10.5	5.2	24	
30	4.7	2.4	2.6	2.3	3.0	3.9	5.0	5.2	3.0	4.2	2.7	3.1	2.7	1.8	3.9	2.9	2.3	4.8	6.5	10.0	S	12.5	15.5	10.9	1.8	15.5	5.0	24	
31	14.6	14.0	4.4	12.1	2.3	5.0	7.8	P	3.5	4.3	4.7	3.9	3.6	4.9	4.2	1.7	1.8	1.6	7.2	S	2.6	5.7	3.1	2.1	1.6	14.6	5.2	23	
HOURLY MAX	28.6	24.2	20.9	18.6	24.4	64.7	30.9	20.9	63.9	14.1	21.2	17.6	6.8	12.1	10.8	10.0	11.4	63.7	32.9	32.6	26.0	29.0	25.8	22.7					
HOURLY AVG	6.4	6.6	6.3	6.4	7.2	9.6	7.9	8.3	9.5	4.6	5.1	3.8	2.7	3.5	4.0	3.8	3.6	6.4	6.6	7.7	8.3	8.0	7.5	6.3					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	701
MAXIMUM INSTANTANEOUS VALUE:	64.7 ppb @ HOUR(S) 17 ON DAY(S) 9
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	7 hrs
OPERATIONAL TIME:	741 hrs
STANDARD DEVIATION:	6.7

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)




Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-NO2[ppb]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 12.13% Calm Avg: 7.91 [ppb]


Direction	0.0-7.9	7.9-15.9	15.9-23.8	>23.8	Total
N	5.6	0.1	0.0	0.0	5.7
NE	15.4	0.9	0.1	0.0	16.4
E	20.1	1.3	0.4	0.0	21.8
SE	26.5	0.4	0.1	0.0	27.1
S	1.0	0.0	0.0	0.0	1.0
SW	4.3	0.1	0.0	0.0	4.4
W	5.6	0.7	0.0	0.0	6.3
NW	5.1	0.0	0.0	0.0	5.1
Summary	83.6	3.6	0.7	0.0	87.9

% Icon Classes (ppb) 84 0.0-7.9 4 7.9-15.9 1 15.9-23.8 0 >23.8

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-NO2[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 12.13% Calm Poll Avg: 7.91[ppb]



NO2[ppb] Calibration: LICA COLD LAKE SOUTH Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

OZONE

OZONE Hourly Averages (O₃ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	33.7	32.3	32.4	S	31.7	30.6	29.8	30.2	31.1	32.3	33.9	35.0	35.4	34.9	35.0	34.1	33.7	32.7	31.3	28.7	26.8	24.4	31.6	24.4	35.4	31.8	24	
2	33.3	34.5	S	34.2	32.8	30.6	27.0	15.9	23.7	32.3	31.7	31.8	33.9	34.0	33.0	32.0	31.1	27.9	28.1	31.6	32.8	33.3	33.1	32.5	15.9	34.5	30.9	24
3	31.4	S	30.0	29.7	24.9	17.7	10.3	22.8	20.8	26.1	28.6	29.5	31.5	33.6	38.0	37.5	35.3	31.9	24.3	12.3	13.6	18.9	34.8	36.8	10.3	38.0	27.0	24
4	S	36.4	34.2	25.9	24.9	29.5	32.0	34.7	35.9	36.9	38.0	38.7	39.4	39.6	39.3	39.2	38.8	39.2	39.2	39.0	38.8	38.6	39.1	S	24.9	39.6	36.2	24
5	40.2	39.6	39.6	39.3	39.5	39.2	38.6	37.9	38.3	38.5	38.8	39.1	39.1	38.8	38.7	38.5	38.2	37.7	34.0	33.2	34.3	35.3	S	35.2	33.2	40.2	37.9	24
6	36.0	36.4	36.7	36.2	36.0	35.4	34.8	32.1	S1	33.7	35.2	35.3	C	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	32.1	36.7	35.3	15
7	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	40.6	40.2	40.2	40.3	39.6	38.0	S	33.5	31.7	30.0	30.0	40.6	37.1	16
8	29.2	27.2	29.5	32.1	32.5	31.4	29.6	32.3	33.3	35.0	36.3	36.7	37.0	37.1	37.3	38.1	37.3	35.3	32.5	S	22.5	25.4	34.6	32.1	22.5	38.1	32.8	24
9	29.1	28.2	29.7	26.4	23.0	22.3	17.7	16.8	20.5	25.9	22.2	21.2	24.2	29.5	28.0	29.5	27.9	26.0	S	16.8	12.8	7.1	5.7	4.7	4.7	29.7	21.5	24
10	2.7	3.5	8.3	10.6	9.9	15.0	17.0	16.1	19.2	20.5	21.5	22.9	25.0	28.7	29.7	28.4	28.0	S	26.8	21.3	25.1	25.3	23.8	25.9	2.7	29.7	19.8	24
11	29.7	28.3	21.6	17.6	16.2	26.2	32.4	X	X	33.0	32.8	33.7	34.2	35.3	35.9	36.5	S	37.1	34.3	33.7	34.7	34.8	35.1	35.8	16.2	37.1	31.4	22
12	35.9	34.2	34.3	34.6	34.6	34.7	34.3	33.7	34.2	37.7	36.3	37.2	38.1	38.7	38.8	S	38.3	40.0	43.4	43.0	42.8	42.3	40.8	39.7	33.7	43.4	37.7	24
13	39.9	40.0	40.0	38.6	38.7	38.2	37.6	37.0	37.3	37.5	37.2	38.5	39.4	39.7	S	41.0	41.5	41.2	40.5	40.0	39.8	39.6	38.8	38.3	37.0	41.5	39.1	24
14	37.5	33.5	35.7	34.9	35.7	34.1	32.6	29.8	34.0	36.5	36.6	38.7	40.0	S	41.8	42.5	42.6	41.4	44.5	44.7	43.6	43.5	42.7	42.0	29.8	44.7	38.6	24
15	41.1	40.4	39.1	37.9	36.0	33.3	30.0	29.8	31.3	31.8	32.7	34.3	S	37.7	37.3	41.3	39.8	38.2	30.3	20.3	20.4	30.3	43.8	45.3	20.3	45.3	34.9	24
16	45.1	44.1	39.7	42.0	39.3	34.8	37.2	24.7	39.1	42.9	44.0	S	45.2	45.2	45.0	44.2	41.4	38.4	38.7	38.3	38.4	37.4	34.3	33.8	24.7	45.2	39.7	24
17	35.9	35.6	36.4	36.2	35.0	33.8	34.4	33.7	33.3	34.5	S	42.1	46.1	47.7	47.8	48.3	47.6	44.6	43.1	41.9	40.6	39.7	39.1	37.8	33.3	48.3	39.8	24
18	36.4	35.9	35.0	34.1	32.9	32.3	31.8	31.1	31.7	S	32.4	32.8	33.0	33.5	33.4	33.0	32.6	31.9	31.9	32.3	32.0	31.5	28.3	28.8	28.3	36.4	32.5	24
19	27.8	27.2	29.3	30.9	36.9	38.5	37.5	37.8	S	37.5	38.9	40.4	41.5	43.7	44.4	44.0	42.6	43.8	41.8	40.6	39.5	38.6	38.7	36.0	27.2	44.4	38.2	24
20	37.4	39.3	39.8	41.1	40.7	40.4	41.3	S	39.7	40.8	41.4	42.7	43.0	43.2	43.7	44.0	44.1	43.7	41.4	35.7	30.4	26.6	25.3	24.5	24.5	44.1	38.7	24
21	20.4	16.6	31.3	31.0	30.3	28.3	S	30.7	32.6	33.9	36.1	39.3	39.8	40.0	39.7	39.9	39.0	38.7	38.5	38.7	38.2	37.3	36.6	36.1	16.6	40.0	34.5	24
22	35.8	35.8	35.8	35.6	34.6	S	33.0	31.5	32.4	31.7	31.4	31.0	30.8	29.5	30.1	30.9	30.3	30.7	30.4	27.9	25.9	22.2	19.6	16.3	16.3	35.8	30.1	24
23	12.7	13.7	15.4	13.2	S	7.8	7.3	11.3	23.2	30.8	32.6	33.7	35.3	36.9	36.4	36.5	36.5	34.4	31.1	27.3	27.1	27.4	27.5	24.2	7.3	36.9	25.3	24
24	20.6	18.8	12.5	S	10.9	10.0	10.9	22.5	26.9	29.3	30.9	32.4	33.0	32.6	32.4	33.5	35.9	36.5	35.2	34.5	34.7	34.8	34.5	34.7	10.0	36.5	27.7	24
25	34.2	33.9	S	32.9	32.8	31.5	30.7	31.3	32.6	31.2	29.4	30.4	32.5	33.9	38.6	40.2	40.7	41.1	36.1	26.6	16.5	7.1	9.7	23.7	7.1	41.1	30.3	24
26	27.8	S	30.2	23.7	18.3	17.5	16.7	20.8	27.0	27.9	29.9	33.3	35.8	41.1	45.6	47.7	47.7	46.9	44.2	36.1	30.3	18.4	11.8	24.1	11.8	47.7	30.6	24
27	S	15.9	14.2	18.2	28.0	27.6	28.2	29.5	32.6	33.6	35.1	37.8	38.6	38.0	38.6	35.5	36.2	37.4	38.7	38.4	35.6	32.3	31.8	S	14.2	38.7	31.9	24
28	30.8	32.9	34.6	33.4	32.0	29.3	28.3	24.8	25.2	26.3	27.0	28.3	28.7	31.0	33.2	34.0	32.8	30.0	27.8	27.2	26.8	26.7	S	18.6	18.6	34.6	29.1	24
29	15.0	11.9	10.6	8.0	6.1	5.8	15.6	18.1	19.7	19.8	21.3	27.5	28.3	27.5	29.8	31.4	31.4	29.4	27.4	26.0	26.1	S	27.2	25.7	5.8	31.4	21.3	24
30	25.2	24.7	25.4	26.3	24.6	25.5	25.7	26.6	27.0	29.2	30.6	32.9	32.6	33.8	34.5	35.1	35.0	33.3	30.0	20.3	S	8.7	4.5	5.9	4.5	35.1	26.0	24
31	2.3	6.5	6.8	4.0	24.2	23.2	14.2	19.9	20.8	22.3	26.2	31.6	37.4	38.9	42.2	44.9	43.2	40.6	37.0	S	36.6	35.7	37.3	36.7	2.3	44.9	27.5	24
HOURLY MAX	45.1	44.1	40.0	42.0	40.7	40.4	41.3	37.9	39.7	42.9	44.0	42.7	46.1	47.7	47.8	48.3	47.7	46.9	44.5	44.7	43.6	43.5	43.8	45.3				
HOURLY AVG	29.5	28.8	28.9	28.9	29.1	27.7	27.5	27.3	29.7	32.0	32.7	34.1	35.7	36.6	37.5	38.0	37.6	36.9	35.3	32.0	31.0	29.6	29.8	29.9				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 82 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0					
NUMBER OF NON-ZERO READINGS:	683					
MINIMUM 1-HR AVERAGE:	2.3	ppb	@ HOUR(S)	0	ON DAY(S)	2.3
MAXIMUM 1-HR AVERAGE:	48.3	ppb	@ HOUR(S)	15	ON DAY(S)	17
MAXIMUM 24-HR AVERAGE:	39.8	ppb			ON DAY(S)	17
					VAR-VARIOUS	
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	725	hrs	
MONTHLY CALIBRATION TIME:	10	hrs	AMD OPERATION UPTIME:	97.4	%	
STANDARD DEVIATION:	8.7		MONTHLY AVERAGE:	31.9	ppb	

24 HR AVERAGES March 2017



OZONE Hourly Averages (O₃ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake Continuous Monitoring Station - March 2017

OZONE Instantaneous Maximum (O₃ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	34.5	33.2	33.3	S	32.3	31.5	30.7	31.0	31.4	32.5	33.2	34.8	36.1	36.2	35.4	36.2	34.8	34.5	34.1	32.9	31.4	29.7	28.5	33.5	28.5	36.2	33.1	24
2	34.7	35.0	S	35.1	33.6	32.6	31.8	23.2	32.9	33.2	32.7	33.9	34.2	34.2	33.6	33.0	31.9	30.7	31.2	32.3	33.3	33.5	33.5	32.9	23.2	35.1	32.7	24
3	32.4	S	30.8	30.7	29.4	22.4	20.3	24.8	26.1	29.1	29.4	30.4	32.4	37.6	38.5	38.5	36.8	34.1	30.5	20.5	17.7	26.7	37.9	37.6	17.7	38.5	30.2	24
4	S	37.7	35.7	31.0	27.6	31.3	33.3	35.6	36.5	37.7	38.2	39.4	39.8	40.0	39.5	39.6	39.4	39.8	39.4	39.3	39.3	39.0	40.0	S	27.6	40.0	37.2	24
5	40.7	40.3	40.2	40.0	40.1	40.0	39.3	38.8	39.1	39.0	39.4	39.8	39.3	39.1	39.0	39.0	38.5	38.9	37.3	34.8	35.7	36.2	S	36.4	34.8	40.7	38.7	24
6	37.1	37.1	37.3	37.0	36.4	36.0	35.7	S1	S1	35.5	35.6	35.7	C	C	C	C	C	C	Y	Y	Y	Y	Y	Y	35.5	37.3	36.3	16
7	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	41.1	41.1	40.8	40.5	40.4	38.8	S	35.0	32.2	30.8	30.8	41.1	37.9	16
8	29.9	28.8	31.3	33.3	33.3	32.4	31.2	33.2	34.0	36.2	36.6	37.2	37.3	37.5	38.2	38.5	37.9	37.7	35.4	S	29.7	34.5	35.7	34.2	28.8	38.5	34.5	24
9	30.8	29.7	33.0	29.8	24.8	24.4	21.8	18.7	24.8	26.6	25.0	22.2	28.6	32.4	29.4	31.5	29.4	27.6	S	23.3	16.7	10.4	9.2	7.3	7.3	33.0	24.2	24
10	4.2	6.1	11.6	15.6	14.5	18.0	18.0	19.2	19.7	21.2	23.0	24.9	27.5	29.9	31.0	29.8	28.9	S	28.0	26.2	27.6	27.4	24.6	29.8	4.2	31.0	22.0	24
11	30.1	30.1	25.8	21.1	24.3	29.1	33.8	X	X	33.6	33.2	33.9	34.8	36.1	36.4	37.6	S	38.2	35.9	34.2	35.3	35.4	35.6	36.5	21.1	38.2	32.9	22
12	36.8	34.4	34.8	34.8	35.3	35.0	34.1	35.3	43.5	37.3	37.6	38.2	39.3	39.0	S	38.8	43.5	44.1	43.5	43.2	42.7	41.4	40.0	34.1	44.1	44.1	38.6	24
13	40.1	40.1	40.0	39.8	39.3	38.2	38.2	37.3	37.6	37.6	38.1	39.0	39.6	39.9	S	41.3	41.7	41.7	40.7	40.3	40.2	39.6	39.4	39.0	37.3	41.7	39.5	24
14	38.4	38.0	37.0	36.4	36.4	36.0	33.8	31.2	35.3	36.9	37.6	39.4	41.1	S	42.5	43.0	44.2	46.1	45.8	45.2	44.1	43.9	43.5	42.5	31.2	46.1	39.9	24
15	41.9	40.8	39.8	38.4	37.1	35.3	33.2	31.5	31.7	32.6	34.1	36.7	S	41.3	38.0	44.3	42.8	39.9	36.4	30.4	28.5	42.9	44.9	46.0	28.5	46.0	37.8	24
16	47.0	46.4	43.8	45.4	43.9	42.4	42.1	36.3	44.0	44.6	44.6	S	46.0	45.8	45.4	44.8	43.6	39.3	39.6	39.6	39.3	38.4	37.1	37.6	36.3	47.0	42.5	24
17	37.1	36.2	36.8	36.5	36.5	34.7	35.1	34.7	34.8	35.3	S	44.8	46.9	48.9	48.7	48.9	49.1	46.1	43.6	42.4	41.2	40.2	39.6	38.5	34.7	49.1	40.7	24
18	37.0	36.1	35.6	34.8	33.5	32.7	32.6	31.9	32.3	S	32.7	35.1	33.3	33.8	33.6	33.3	33.0	32.6	32.9	32.6	32.6	32.4	29.1	29.4	29.1	37.0	33.2	24
19	28.3	28.5	30.4	33.5	38.3	39.9	38.0	38.2	S	39.1	40.3	41.0	43.2	44.6	44.9	44.8	43.6	44.6	43.0	41.2	39.8	39.3	39.4	37.4	28.3	44.9	39.2	24
20	38.2	39.8	40.5	41.7	41.1	41.0	42.1	S	41.4	41.6	42.4	43.5	43.8	44.4	44.4	44.6	44.6	44.4	42.8	39.9	33.0	30.8	28.3	27.0	27.0	44.6	40.1	24
21	24.0	23.0	33.3	32.3	32.3	31.2	S	32.7	33.5	34.7	38.8	40.1	40.5	40.5	40.6	40.6	39.9	39.0	38.9	38.9	38.7	37.7	37.0	36.5	23.0	40.6	35.9	24
22	36.0	36.0	35.9	35.7	35.3	S	33.5	33.0	33.3	32.4	32.0	31.5	31.3	30.8	31.2	31.2	30.7	30.8	30.7	29.4	28.0	24.0	20.5	18.3	18.3	36.0	30.9	24
23	14.4	14.9	15.8	14.9	S	10.6	11.0	18.0	28.0	33.2	32.7	34.5	36.1	37.3	37.4	36.8	36.8	37.0	33.2	30.5	30.1	29.8	29.1	27.9	10.6	37.4	27.4	24
24	23.4	27.1	15.5	S	14.6	12.5	15.3	26.1	28.0	30.5	31.7	33.2	33.3	33.0	33.6	34.8	37.6	37.9	35.4	35.0	35.0	35.0	35.0	35.1	12.5	37.9	29.5	24
25	34.5	34.2	S	33.2	33.5	32.4	31.3	32.3	33.2	32.2	30.2	31.2	34.4	36.4	39.7	41.0	41.2	41.4	40.5	31.4	23.3	12.5	19.6	29.5	12.5	41.4	32.6	24
26	31.5	S	31.2	30.1	20.0	20.5	23.6	27.3	39.9	30.2	31.5	34.8	36.7	44.1	47.2	48.1	48.1	47.7	46.0	42.1	37.1	27.3	23.6	31.9	20.0	48.1	34.8	24
27	S	19.3	16.8	26.7	28.9	28.3	29.1	32.4	33.3	34.1	36.1	38.4	38.8	38.5	39.7	37.0	37.1	38.2	39.8	39.3	37.6	32.7	31.9	S	16.8	39.8	33.4	24
28	31.3	34.2	35.0	34.1	32.4	30.8	29.2	26.1	26.1	28.5	28.2	29.5	29.9	32.7	35.4	35.0	33.8	31.3	28.5	27.6	27.1	27.1	S	23.0	23.0	35.4	30.3	24
29	18.9	15.8	13.1	12.2	10.1	8.5	18.6	19.9	22.3	22.0	24.4	29.4	28.9	28.6	30.8	32.0	32.2	30.6	30.1	27.6	26.7	S	27.9	26.5	8.5	32.2	23.4	24
30	26.2	25.6	26.8	26.8	25.9	27.0	27.3	28.0	28.6	30.2	31.7	35.3	33.9	34.4	35.9	35.7	35.7	35.1	34.1	25.3	S	13.7	8.2	7.9	7.9	35.9	27.8	24
31	5.8	19.9	9.4	20.8	26.0	25.3	22.3	P	21.0	24.9	28.3	35.9	40.7	40.8	45.8	45.8	45.1	41.7	39.3	S	37.1	37.0	37.9	37.6	5.8	45.8	31.3	23
HOURLY MAX	47.0	46.4	43.8	45.4	43.9	42.4	42.1	38.8	44.0	44.6	44.6	44.8	46.9	48.9	48.7	48.9	49.1	47.7	46.0	45.2	44.1	43.9	44.9	46.0				
HOURLY AVG	30.9	31.0	30.4	31.5	30.9	29.7	29.9	29.8	32.0	33.4	33.8	35.3	36.7	37.8	38.5	38.9	38.6	38.3	37.2	34.4	33.2	32.2	31.8	31.8				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	681
MAXIMUM INSTANTANEOUS VALUE:	49.1 ppb @ HOUR(S) 16 ON DAY(S) 17
VAR-VARIOUS	
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	12 hrs
STANDARD DEVIATION:	7.7
OPERATIONAL TIME:	725 hrs

OZONE Instantaneous Maximum (O₃ ppb)




Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-O3[ppb]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 12.48% Calm Avg: 19.60 [ppb]


Direction	0.0-16.1	16.1-32.3	32.3-48.4	>48.4	Total
N	0.0	1.5	3.5	0.0	5.0
NE	0.3	3.5	13.1	0.0	16.9
E	0.6	12.0	9.8	0.0	22.5
SE	0.4	7.8	19.7	0.0	27.9
S	0.0	0.0	1.0	0.0	1.0
SW	0.0	0.6	4.0	0.0	4.6
W	0.4	2.4	3.7	0.0	6.5
NW	0.0	1.2	2.1	0.0	3.2
Summary	1.8	28.9	56.8	0.0	87.5

% Icon Classes (ppb) 2 0.0-16.1 29 16.1-32.3 57 32.3-48.4 0 >48.4

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-O3[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 12.48% Calm Poll Avg: 19.60[ppb]



O3[ppb] Calibration: LICA COLD LAKE SOUTH Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

PARTICULATE MATTER 2.5

PARTICULATE MATTER < 2.5 MICRONS Hourly Averages (PM_{2.5} µg/m³)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	0.0	1.0	2.6	0.0	0.0	4.4	2.4	6.0	5.1	2.4	0.0	1.9	3.5	1.0	0.0	1.1	2.4	0.0	0.0	0.5	1.0	3.4	2.5	1.9	0.0	6.0	1.8	24	
2	0.0	0.0	1.5	2.1	3.2	0.0	1.2	1.6	8.4	3.6	6.3	5.8	2.5	7.0	6.0	11.8	10.8	13.1	8.1	11.2	2.5	16.0	15.6	9.5	0.0	16.0	6.2	24	
3	9.0	9.9	4.0	7.5	7.4	10.1	8.8	5.9	9.4	11.8	10.5	9.0	1.6	2.9	0.0	0.0	1.2	5.0	5.7	7.7	10.7	1.0	3.3	0.2	0.0	11.8	5.9	24	
4	1.2	0.0	1.7	2.5	5.7	1.0	2.4	0.0	2.5	0.0	1.0	0.2	2.4	0.0	2.2	0.7	X	0.0	7.0	3.1	0.5	3.0	4.0	2.3	0.0	7.0	1.9	23	
5	3.8	0.1	0.0	0.0	4.1	0.0	0.0	0.3	1.4	1.1	0.0	0.8	1.3	2.3	1.0	0.0	0.0	1.6	0.0	0.5	0.0	3.3	2.3	3.4	0.0	4.1	1.1	24	
6	0.0	0.0	1.4	2.1	0.0	0.0	X	3.1	1.4	2.9	0.0	2.9	3.0	1.4	X	0.0	0.0	6.5	4.9	2.5	1.5	1.5	0.0	1.9	0.0	6.5	1.7	22	
7	1.5	1.9	4.0	0.0	2.9	1.3	1.0	2.5	2.8	0.0	C	C	2.9	0.5	1.9	2.5	0.0	0.0	1.4	0.0	1.4	0.0	3.4	6.5	7.5	0.0	7.5	2.0	24
8	6.0	4.0	3.4	2.9	6.5	1.0	3.9	2.5	1.9	1.9	0.4	0.0	0.0	0.4	1.4	0.0	X	0.5	1.4	1.0	3.4	6.5	2.9	5.0	0.0	6.5	2.5	23	
9	1.9	2.9	5.0	8.4	5.4	1.0	7.5	2.9	3.9	7.9	2.4	3.9	1.0	0.0	5.5	3.4	6.0	1.0	7.5	2.9	4.4	8.9	4.9	6.4	0.0	8.9	4.4	24	
10	2.4	1.4	7.5	2.4	3.9	3.4	8.4	6.4	3.4	7.9	1.4	5.5	4.9	2.5	1.4	4.0	7.0	7.0	4.4	3.4	3.4	5.5	0.0	5.5	0.0	8.4	4.3	24	
11	4.0	1.9	3.4	2.5	0.5	1.0	1.4	X	X	2.5	2.5	1.9	3.4	2.9	2.9	1.0	6.0	2.9	4.4	1.9	2.5	3.4	1.4	2.9	0.5	6.0	2.6	22	
12	1.4	2.9	1.9	2.9	1.4	2.5	2.5	4.0	3.4	21.5	4.0	6.0	4.0	4.0	3.4	9.4	0.0	5.0	0.5	1.4	1.4	4.0	1.0	2.9	0.0	21.5	3.8	24	
13	1.9	2.5	1.4	4.0	5.5	9.9	4.4	3.4	0.0	2.9	2.5	4.4	1.4	2.9	9.4	12.4	7.5	9.9	4.4	1.9	5.5	6.0	2.5	9.4	0.0	12.4	4.8	24	
14	4.0	6.0	5.0	4.0	7.0	3.4	3.4	7.0	6.5	5.9	7.9	5.5	4.0	0.0	3.4	7.9	7.0	12.4	6.5	0.0	3.4	7.0	4.0	1.0	0.0	12.4	5.1	24	
15	5.0	4.4	6.5	6.5	9.4	9.0	9.0	7.5	3.4	9.4	4.0	0.0	1.9	1.9	X	9.4	9.4	5.5	9.9	3.4	3.4	2.9	0.0	0.0	0.0	9.9	5.3	23	
16	1.4	1.9	1.0	1.9	0.0	2.9	0.0	2.9	0.0	1.4	3.4	4.0	0.0	X	0.0	X	X	1.9	0.0	0.0	0.0	0.0	0.0	0.5	0.0	4.0	1.1	21	
17	2.9	2.5	4.0	2.9	4.0	2.9	2.9	1.9	6.0	5.5	2.5	1.4	0.0	X	X	X	0.5	5.5	7.9	4.0	2.5	2.9	6.5	1.9	0.0	7.9	3.4	21	
18	3.4	4.0	3.4	1.9	2.5	0.0	0.5	2.9	4.4	1.4	5.0	2.9	1.0	0.0	X	X	X	1.9	1.9	1.9	3.4	1.9	0.0	1.4	0.0	5.0	2.2	21	
19	X	2.9	1.4	2.9	1.4	X	0.0	1.9	1.9	2.9	3.4	3.9	6.5	5.0	5.0	1.9	0.0	1.0	0.5	0.0	0.0	0.0	1.9	3.6	0.0	6.5	2.2	22	
20	2.5	1.3	0.5	1.0	0.2	1.9	0.0	3.5	1.4	1.9	0.0	6.0	0.0	0.5	4.4	1.1	1.5	5.0	0.4	2.5	2.9	6.9	4.1	2.1	0.0	6.9	2.2	24	
21	0.9	1.0	0.0	0.5	0.0	0.6	1.4	4.4	4.7	2.6	0.1	2.1	0.7	5.5	7.3	7.0	0.0	0.0	2.8	3.3	X	3.4	0.0	0.0	0.0	7.3	2.1	23	
22	0.0	2.3	0.0	1.8	1.6	0.0	0.0	4.6	3.4	1.8	0.0	3.4	1.9	5.0	2.2	7.7	3.4	4.0	0.7	7.0	3.4	5.2	4.7	7.2	0.0	7.7	3.0	24	
23	2.2	5.2	5.0	4.4	0.5	6.5	7.7	7.3	7.1	4.2	1.5	4.2	9.9	12.2	6.7	1.4	9.4	3.2	8.6	8.1	7.1	6.7	7.7	6.5	0.5	12.2	6.0	24	
24	6.5	3.2	X	0.0	2.9	3.3	3.4	6.8	8.1	6.5	13.9	6.3	8.3	12.0	9.8	6.3	6.0	4.3	4.8	3.1	3.6	3.2	0.5	3.2	0.0	13.9	5.5	23	
25	1.3	3.7	0.0	1.2	5.6	4.0	0.0	1.1	5.0	6.0	13.0	10.4	11.5	8.9	4.4	6.0	10.7	9.3	4.9	6.4	6.1	7.4	9.0	4.2	0.0	13.0	5.8	24	
26	5.5	5.5	4.0	4.0	1.4	1.0	3.5	8.4	17.9	7.0	6.4	1.5	8.4	1.4	5.5	1.4	0.0	0.0	4.0	2.9	4.0	3.9	7.0	5.0	0.0	17.9	4.6	24	
27	4.4	2.9	4.4	2.9	3.4	2.5	2.9	1.9	4.0	1.9	3.4	3.4	5.5	1.0	C	C	2.9	0.0	2.5	2.5	2.9	4.0	1.9	2.9	0.0	5.5	2.9	24	
28	4.0	0.0	0.0	0.5	2.5	4.0	5.0	7.0	8.4	7.0	9.4	9.0	9.4	9.4	9.0	3.4	6.0	6.0	6.0	5.0	5.5	6.5	6.0	5.0	0.0	9.4	5.6	24	
29	4.4	3.4	4.0	4.0	3.4	7.5	7.0	5.0	7.0	4.4	5.4	5.0	9.4	8.4	12.0	9.4	7.9	8.4	10.9	2.5	2.5	6.5	4.0	2.5	2.5	12.0	6.0	24	
30	0.0	2.9	0.0	1.0	2.9	4.0	0.0	2.9	4.0	4.0	1.4	5.5	4.4	4.4	4.0	5.0	3.4	4.4	4.4	5.0	4.4	3.4	5.0	1.4	0.0	5.5	3.2	24	
31	1.9	4.4	4.0	3.4	0.0	1.4	6.5	0.0	1.4	7.0	8.4	7.5	4.0	5.0	8.4	2.5	5.0	5.0	5.5	4.0	1.0	3.4	4.4	6.0	0.0	8.4	4.2	24	
HOURLY MAX	9.0	9.9	7.5	8.4	9.4	10.1	9.0	8.4	17.9	21.5	13.9	10.4	11.5	12.2	12.0	12.4	10.8	13.1	10.9	11.2	10.7	16.0	15.6	9.5					
HOURLY AVG	2.8	2.8	2.7	2.6	3.1	3.0	3.2	3.9	4.6	4.7	4.0	4.1	3.8	3.7	4.5	4.3	4.2	4.2	4.2	3.3	3.1	4.6	3.7	3.7					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 80 µg/m³ 24-HR 30 µg/m³


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0		
NUMBER OF 24-HR EXCEEDANCES:	0		
NUMBER OF NON-ZERO READINGS:	622		
MINIMUM 1-HR AVERAGE:	0.0 µg/m ³ @ HOUR(S)	0	ON DAY(S) 1
MAXIMUM 1-HR AVERAGE:	21.5 µg/m ³ @ HOUR(S)	9	ON DAY(S) 12
MAXIMUM 24-HR AVERAGE:	6.2 µg/m ³		ON DAY(S) 2
			VAR-VARIOUS
MONTHLY CALIBRATION TIME:	4 hrs	OPERATIONAL TIME:	724 hrs
STANDARD DEVIATION:	3.1	AMD OPERATION UPTIME:	97.3 %
		MONTHLY AVERAGE:	3.7 µg/m ³

24 HR AVERAGES March 2017



PARTICULATE MATTER < 2.5 MICRONS Hourly Averages (PM_{2.5} µg/m³)



Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-PM25[ug/m3(L)]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 12.24%

Calm Avg: 4.55 [ppb]

Direction	0.0-4.3	4.3-8.6	8.6-13.0	13.0-17.3	17.3-21.6	>21.6	Total
N	4.9	1.0	0.3	0.0	0.0	0.0	6.1
NE	12.9	2.6	0.4	0.0	0.0	0.0	16.0
E	15.4	4.9	1.7	0.3	0.1	0.0	22.4
SE	16.3	8.2	2.2	0.3	0.0	0.0	27.0
S	0.3	0.4	0.1	0.0	0.0	0.0	0.8
SW	2.5	1.3	0.4	0.0	0.0	0.0	4.2
W	2.9	2.4	0.8	0.1	0.0	0.0	6.3
NW	4.3	0.6	0.1	0.0	0.0	0.0	5.0
Summary	59.5	21.3	6.1	0.7	0.1	0.0	87.8

% Icon Classes (ug/m3(L)) 60 0.0-4.3 21 4.3-8.6 6 8.6-13.0 1 13.0-17.3 0 17.3-21.6 0 >21.6

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-PM25[ug/m3(L)] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 12.24% Calm Poll Avg: 4.55[ug/m3(L)]



WIND SPEED




WIND SPEED Hourly Averages (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY 1	8.0	7.9	9.4	8.1	6.4	6.2	6.7	8.2	8.5	8.8	9.2	8.7	8.5	7.1	7.2	8.3	7.4	5.5	3.3	2.4	2.5	1.8	1.4	3.9	1.4	9.4	5.7	24
2	6.3	7.0	7.0	5.0	4.4	2.9	2.7	1.3	3.2	5.1	4.3	4.2	6.4	7.6	4.9	4.8	5.9	4.9	6.6	9.5	11.0	11.8	11.2	10.6	1.3	11.8	4.9	24
3	5.6	4.2	4.3	4.5	3.3	2.1	2.6	2.8	1.8	4.0	4.9	6.1	6.6	6.4	8.0	9.7	7.2	4.4	1.7	1.1	1.1	3.7	8.0	5.8	1.1	9.7	2.1	24
4	4.1	7.0	12.1	11.8	13.9	14.4	13.8	16.4	12.7	13.3	14.8	13.0	11.7	10.4	9.6	10.1	9.7	9.6	9.8	11.9	7.4	6.0	6.0	8.1	4.1	16.4	10.6	24
5	8.5	7.1	6.4	4.8	6.5	8.0	6.7	6.0	4.8	5.7	5.6	6.3	6.9	5.7	7.1	5.8	4.0	3.9	0.8	2.9	3.5	4.9	4.2	2.5	0.8	8.5	5.1	24
6	3.7	6.1	6.2	6.1	7.8	7.6	5.9	2.5	3.3	7.0	9.4	8.9	8.7	9.2	7.8	9.0	10.0	10.0	9.6	11.6	12.8	12.8	11.6	12.5	2.5	12.8	7.2	24
7	15.3	14.6	14.7	13.8	14.2	13.7	15.7	13.9	15.2	14.6	13.7	13.1	14.8	14.0	14.6	12.0	12.0	12.7	8.8	7.2	7.1	6.2	5.4	4.3	4.3	15.7	11.9	24
8	4.3	4.5	6.8	8.4	4.4	3.4	3.8	6.9	8.2	8.3	6.9	9.1	7.8	8.3	8.1	7.9	4.6	2.3	1.9	2.4	0.7	1.8	4.9	2.4	0.7	9.1	4.6	24
9	2.2	2.7	3.3	2.4	1.6	0.2	1.2	0.4	1.9	4.5	5.4	5.0	4.2	3.4	3.7	4.7	4.9	3.1	1.0	0.6	0.4	0.9	1.0	0.9	0.2	5.4	2.0	24
10	1.0	2.5	2.0	1.4	2.5	4.7	6.6	4.7	6.9	6.6	7.1	9.8	9.1	7.9	7.6	7.2	5.4	3.7	2.3	0.9	3.9	6.5	4.5	2.7	0.9	9.8	4.7	24
11	3.4	2.6	2.1	0.7	5.7	3.9	6.8	X	X	5.6	5.6	5.7	6.7	7.0	7.2	7.3	7.9	8.7	7.9	8.4	8.4	8.1	7.8	7.1	0.7	8.7	5.6	22
12	6.4	8.1	7.2	7.9	8.1	9.0	8.4	8.0	6.3	6.8	6.0	6.4	5.9	6.0	6.5	7.7	5.2	5.2	4.4	6.7	7.9	8.1	5.9	4.7	4.4	9.0	6.3	24
13	8.2	8.3	7.6	5.9	7.6	9.1	10.0	10.8	10.7	10.9	10.9	13.1	12.3	8.9	9.7	9.1	9.9	8.9	8.1	8.2	6.5	5.1	2.5	1.7	1.7	13.1	8.4	24
14	1.2	1.5	1.4	4.2	5.8	5.0	4.2	5.3	3.2	4.2	5.9	4.9	4.7	4.3	5.3	4.2	3.4	3.9	6.5	6.5	6.5	6.5	6.5	6.0	1.2	6.5	3.9	24
15	5.1	7.2	7.6	6.4	4.7	4.2	3.8	4.9	0.9	3.5	5.2	6.1	4.6	2.2	7.2	5.9	5.4	4.6	0.5	0.7	1.2	3.7	5.2	6.8	0.5	7.6	1.2	24
16	2.6	2.4	1.7	2.6	0.8	2.9	0.3	0.5	1.1	3.3	6.3	7.4	9.2	8.4	7.0	6.5	7.4	5.0	4.3	3.7	4.5	3.4	1.9	3.2	0.3	9.2	3.5	24
17	4.6	4.5	4.9	6.3	3.9	3.8	5.3	4.3	6.2	6.3	7.2	9.3	7.1	8.3	7.4	7.4	5.8	6.9	8.8	8.6	10.8	11.2	11.0	11.5	3.8	11.5	6.3	24
18	12.0	10.8	10.7	10.4	8.9	8.0	6.4	6.9	6.7	8.5	11.3	10.7	12.7	12.4	9.9	8.0	8.1	7.9	7.9	8.4	8.0	3.9	6.1	13.9	3.9	13.9	7.2	24
19	10.7	12.3	15.3	16.1	15.9	14.6	13.8	14.4	13.9	12.9	14.2	15.5	15.0	15.8	12.1	12.7	8.3	12.9	9.1	9.0	8.1	8.2	9.1	10.4	8.1	16.1	12.0	24
20	16.3	17.3	16.3	15.3	12.4	11.8	14.0	15.5	15.1	13.1	10.7	7.4	8.2	7.9	7.3	7.2	9.3	8.2	5.4	2.3	1.3	0.4	0.7	0.6	0.4	17.3	8.0	24
21	0.9	1.2	3.4	2.2	1.8	2.1	2.8	7.2	7.3	9.0	11.1	13.1	13.9	15.4	13.6	12.8	15.9	14.4	12.3	12.8	11.6	8.9	8.2	8.3	0.9	15.9	8.7	24
22	8.1	9.5	10.5	10.5	9.0	8.5	7.9	8.5	8.5	8.3	7.3	7.9	6.4	5.6	5.7	5.6	5.3	4.8	3.5	3.7	4.6	5.9	5.5	4.9	3.5	10.5	4.0	24
23	4.7	5.3	3.4	1.6	2.3	1.2	0.6	1.0	4.8	5.7	5.2	3.8	1.8	5.4	5.1	4.5	4.3	3.5	2.7	1.6	0.8	1.1	2.3	1.0	0.6	5.7	1.4	24
24	0.4	0.6	0.7	0.7	0.8	1.0	0.8	4.6	5.3	6.8	7.5	9.6	8.5	7.7	7.2	6.3	8.3	11.1	10.3	9.6	8.5	10.5	10.0	9.9	0.4	11.1	5.8	24
25	9.7	10.0	7.1	6.0	5.4	3.4	3.3	3.2	4.2	2.6	5.5	5.6	6.3	6.6	6.4	7.3	3.8	2.3	1.0	0.7	0.5	1.3	2.3	3.5	0.5	10.0	0.5	24
26	4.8	5.5	3.8	0.7	1.5	1.9	1.9	1.6	0.4	3.2	3.3	5.3	4.7	2.1	6.7	9.1	6.9	6.1	3.7	1.1	1.1	1.8	2.0	3.1	0.4	9.1	2.0	24
27	3.2	1.4	1.7	2.9	4.9	4.9	7.0	6.4	7.9	8.1	8.4	9.4	8.6	8.4	9.3	6.8	7.2	7.6	7.7	4.0	3.1	9.5	7.0	4.2	1.4	9.5	6.0	24
28	5.3	5.3	7.2	6.5	7.0	5.5	6.1	5.0	7.8	7.0	7.3	8.1	6.8	7.4	6.7	8.9	10.8	6.8	6.2	7.5	9.0	7.3	1.2	1.1	1.1	10.8	6.5	24
29	0.7	1.0	0.9	0.6	0.6	0.7	2.3	2.7	1.7	0.8	2.4	4.6	4.6	4.5	3.7	2.1	3.1	4.2	3.7	4.5	4.8	6.4	5.1	5.3	0.6	6.4	0.8	24
30	5.5	5.7	5.0	6.3	6.2	6.6	5.9	7.5	6.5	7.3	7.0	5.1	5.7	6.4	6.1	4.6	4.8	3.6	2.2	0.4	1.4	0.3	0.8	0.6	0.3	7.5	4.1	24
31	0.6	0.9	0.8	2.3	4.8	4.1	0.5	5.5	4.5	4.1	3.3	2.6	6.2	8.5	7.9	9.6	7.8	6.7	5.4	5.5	8.2	5.2	4.6	6.4	0.5	9.6	3.7	24
HOURLY MAX	16.3	17.3	16.3	16.1	15.9	14.6	15.7	16.4	15.2	14.6	14.8	15.5	15.0	15.8	14.6	12.8	15.9	14.4	12.3	12.8	12.8	12.8	11.6	13.9				
HOURLY AVG	2.2	2.7	2.6	2.3	2.2	2.5	2.5	2.8	2.5	2.8	3.0	3.3	2.7	2.1	2.1	1.4	2.0	1.8	2.1	2.7	2.7	2.8	2.0	1.6				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


LAST CALIBRATION:	April 1, 2015
DECLINATION :	MAGNETIC DECLINATION 14 DEGREE EAST



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	742
MINIMUM 1-HR AVERAGE:	0.2 kph @ HOUR(S) 5 ON DAY(S) 9
MAXIMUM 1-HR AVERAGE:	17.3 kph @ HOUR(S) 1 ON DAY(S) 20
MAXIMUM 24-HR AVERAGE:	12.0 kph ON DAY(S) 19
	VAR-VARIOUS
MONTHLY CALIBRATION TIME:	0 hrs
OPERATIONAL TIME:	742 hrs
AMD OPERATION UPTIME:	99.7 %
STANDARD DEVIATION:	3.7
MONTHLY AVERAGE:	2.3 kph

WIND SPEED Hourly Averages (WS kph)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake Continuous Monitoring Station - March 2017

WIND SPEED Instantaneous Maximum (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	13.2	12.3	14.7	12.7	10.4	10.2	10.8	13.4	15.2	14.3	18.2	14.5	15.0	13.7	13.5	12.1	11.9	8.7	6.0	4.6	4.4	3.1	3.9	6.2	3.1	18.2	11.0	24	
2	10.8	9.7	11.0	7.4	6.1	5.0	5.5	3.8	7.4	9.2	7.6	8.7	9.1	11.7	7.4	8.4	10.5	9.5	11.6	16.3	16.1	18.9	18.2	14.8	3.8	18.9	10.2	24	
3	11.6	5.9	6.6	6.3	5.3	3.4	6.6	7.1	4.6	7.2	7.1	9.5	11.9	14.3	13.4	16.6	15.2	7.9	4.5	3.0	3.9	6.3	12.4	10.1	3.0	16.6	8.4	24	
4	7.8	14.5	18.1	17.4	21.2	20.8	20.2	23.7	17.9	19.2	21.1	19.3	18.8	19.2	17.7	18.2	14.4	13.9	16.1	20.6	12.6	10.4	9.0	13.6	7.8	23.7	16.9	24	
5	12.2	13.9	10.4	8.0	9.1	11.3	10.7	12.2	8.2	9.9	8.6	9.8	10.9	10.5	10.8	10.7	7.2	7.4	5.5	5.3	6.8	8.0	6.1	6.0	5.3	13.9	9.1	24	
6	10.1	10.6	10.7	9.9	11.9	14.3	9.2	5.1	5.6	14.5	14.7	16.5	13.0	13.4	12.6	13.7	16.1	14.2	15.0	17.6	18.7	17.2	15.8	17.8	5.1	18.7	13.3	24	
7	22.6	23.6	19.6	22.4	22.1	18.8	22.0	21.7	22.1	20.6	21.2	18.4	22.7	19.7	23.2	19.4	18.3	17.7	16.0	10.3	11.3	10.2	10.2	7.0	7.0	23.6	18.4	24	
8	7.3	10.2	12.5	12.8	10.3	5.8	7.3	11.1	13.0	12.7	10.9	13.5	12.4	12.7	13.7	12.3	8.2	6.4	5.3	5.9	3.7	8.1	8.8	4.5	3.7	13.7	9.6	24	
9	3.6	4.5	5.1	3.9	3.0	2.0	3.5	2.5	4.4	7.0	9.3	8.4	7.5	7.7	7.6	9.0	8.3	7.3	3.3	2.8	1.9	2.1	3.7	2.8	1.9	9.3	5.1	24	
10	3.6	4.6	3.3	2.9	4.7	8.9	11.0	12.1	11.2	11.9	14.0	14.8	17.3	14.2	12.5	12.3	12.4	7.3	4.6	3.1	8.2	9.1	7.3	5.6	2.9	17.3	9.0	24	
11	6.1	5.3	3.6	2.6	42.9	8.6	12.8	X	X	10.3	8.7	10.1	11.7	12.5	12.0	11.7	13.9	12.4	12.2	11.8	14.7	13.3	14.6	11.9	2.6	42.9	12.0	22	
12	10.0	12.8	11.5	12.3	12.1	14.7	13.2	14.5	10.4	10.5	9.9	10.4	10.3	12.0	10.9	14.1	10.8	10.1	7.7	10.1	13.3	14.3	9.4	7.8	7.7	14.7	11.4	24	
13	11.6	10.7	10.8	9.3	12.3	13.4	14.3	15.4	16.1	14.8	17.9	21.1	18.9	15.8	15.1	15.9	15.2	15.4	12.3	15.5	11.3	8.5	10.7	8.5	8.5	21.1	13.8	24	
14	12.6	6.5	4.2	7.1	12.4	8.8	6.7	7.6	7.5	11.7	10.0	9.8	9.0	7.8	7.4	6.8	6.2	11.6	14.5	11.8	12.1	11.8	9.5	10.9	4.2	14.5	9.3	24	
15	9.1	12.1	12.3	9.7	8.6	7.0	6.6	8.0	17.8	8.1	9.1	11.9	8.2	7.5	11.8	10.6	8.0	8.9	3.2	3.3	3.6	7.9	10.1	11.9	3.2	17.8	9.0	24	
16	6.1	4.7	4.3	4.9	4.5	5.7	3.7	3.0	3.8	6.8	11.1	13.0	17.9	14.8	13.1	11.3	11.4	8.4	6.9	5.8	7.4	5.7	5.6	6.1	3.0	17.9	7.8	24	
17	8.2	7.3	9.9	9.3	9.4	7.1	10.5	7.0	10.9	9.9	10.5	15.7	14.5	20.9	17.6	11.4	13.7	9.9	12.9	12.3	14.7	18.5	15.2	19.0	7.0	20.9	12.3	24	
18	17.0	15.1	19.1	17.2	17.5	12.7	12.0	11.8	11.9	16.0	18.9	20.9	19.6	20.7	17.5	15.7	13.3	13.1	12.1	13.1	11.8	10.5	10.7	20.8	10.5	20.9	15.4	24	
19	19.7	19.5	22.8	25.4	25.7	29.5	19.8	18.5	21.2	18.9	23.5	22.4	21.7	23.3	22.2	20.7	11.8	22.3	12.7	13.3	11.9	11.5	14.8	16.6	11.5	29.5	19.6	24	
20	27.1	25.0	28.7	28.8	20.9	19.9	22.7	21.8	25.2	20.0	21.7	15.7	17.2	13.2	15.1	12.1	16.0	13.0	11.8	5.2	2.9	1.3	1.8	1.8	1.3	28.8	16.2	24	
21	2.3	5.6	6.0	3.7	3.6	5.1	5.1	11.6	10.9	15.2	17.6	21.5	24.2	21.3	24.0	18.9	23.4	27.9	19.8	18.8	18.8	16.9	13.1	13.5	2.3	27.9	14.5	24	
22	12.8	15.0	18.1	19.7	15.5	13.0	13.4	13.6	13.9	13.1	11.0	12.6	11.3	9.8	9.7	8.6	8.9	7.0	7.3	7.1	7.1	8.9	9.4	7.1	7.0	19.7	11.4	24	
23	7.0	8.4	5.8	3.4	3.4	3.6	3.1	5.3	8.4	10.8	8.8	9.8	6.2	10.2	9.9	8.6	7.8	5.9	4.4	4.6	4.4	3.2	4.7	4.7	3.1	10.8	6.4	24	
24	2.0	2.2	1.9	2.5	5.3	2.5	4.4	7.6	8.6	12.9	13.6	19.3	18.5	11.4	12.4	10.5	13.7	17.6	16.0	16.4	16.7	17.5	17.5	17.0	1.9	19.3	11.2	24	
25	15.4	14.5	12.0	9.7	10.5	7.7	5.0	5.3	7.4	6.6	9.6	10.4	12.2	13.4	10.5	13.2	10.9	6.8	2.8	3.6	2.8	2.7	3.7	6.2	2.7	15.4	8.5	24	
26	7.0	7.6	6.0	4.9	6.2	5.1	4.1	4.9	4.0	6.1	7.6	10.5	8.5	12.9	13.8	16.2	11.8	11.0	8.0	2.8	3.2	3.3	3.4	4.4	2.8	16.2	7.2	24	
27	5.1	3.7	3.2	5.1	7.4	8.0	9.9	11.7	15.1	14.8	15.9	18.7	15.0	15.2	16.4	12.8	12.2	12.1	12.4	8.0	9.3	15.5	10.3	7.3	3.2	18.7	11.0	24	
28	9.4	8.5	12.0	9.8	10.5	9.8	10.2	8.4	11.9	10.3	13.7	14.1	14.1	11.8	12.5	13.9	16.2	12.3	9.1	11.0	12.5	14.0	4.7	4.0	4.0	16.2	11.0	24	
29	2.7	3.4	3.3	1.9	1.6	2.6	5.3	4.5	5.0	4.5	5.7	7.7	7.5	7.7	8.0	7.0	7.3	6.1	7.7	7.2	12.1	10.7	9.5	1.6	12.1	6.1	24		
30	10.5	9.8	9.2	10.1	10.2	9.9	11.1	12.9	11.9	11.0	12.4	8.4	9.6	10.3	10.8	12.4	10.2	6.2	5.4	5.3	3.6	2.7	3.6	3.8	2.7	12.9	8.8	24	
31	3.1	3.9	2.5	6.6	6.8	7.1	3.9	7.9	7.7	8.2	10.2	14.7	14.1	16.2	20.1	14.8	12.6	11.1	8.2	10.4	7.6	11.8	10.2	2.5	20.1	9.6	23		
HOURLY MAX	27.1	25.0	28.7	28.8	42.9	29.5	22.7	23.7	25.2	20.6	23.5	22.4	24.2	23.3	24.0	20.7	23.4	27.9	19.8	20.6	18.8	18.9	18.2	20.8					
HOURLY AVG	9.9	10.0	10.3	9.9	11.3	9.8	9.8	10.6	11.3	11.8	12.8	13.8	13.9	13.7	13.5	13.1	12.2	11.4	9.6	9.2	9.3	9.7	9.4	9.4					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

MAXIMUM INSTANTANEOUS VALUE:	42.9	kph	@ HOUR(S)	4	ON DAY(S)	11
					VAR-VARIOUS	
OPERATIONAL TIME:					741	hrs

WIND SPEED Instantaneous Maximum (WS kph)




Wind: LICA COLD LAKE SOUTH
 Monitor: WSP [kph]
 Monthly: 17/03
 Type: WindRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 11.99%

Calm Avg: 0.00 [kph]

Direction	1.8-3.5	3.5-6.9	6.9-10.4	10.4-13.8	13.8-17.3	>17.3	Total
N	0.5	2.4	1.9	0.5	0.9	0.0	6.3
NE	2.8	7.4	4.2	1.2	0.5	0.0	16.2
E	2.8	9.8	8.9	0.3	0.0	0.0	21.8
SE	3.1	10.4	8.8	4.2	0.5	0.0	27.0
S	0.3	0.4	0.3	0.0	0.0	0.0	0.9
SW	0.7	1.8	1.9	0.0	0.0	0.0	4.3
W	0.5	2.8	1.2	0.9	0.8	0.0	6.3
NW	0.5	0.3	0.4	1.6	2.3	0.0	5.1
Summary	11.3	35.3	27.5	8.8	5.1	0.0	88.0

LICA COLD LAKE SOUTH 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 11.99% Calm Wind Avg Speed: 0.97(kph)



WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake Continuous Monitoring Station - March 2017

WIND DIRECTION Hourly Averages (WD)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	24-HOUR AVG	24-HR	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	QUADRANT	RDGS.	
DAY 1	E	E	E	E	ENE	ENE	ENE	ENE	ENE	ENE	E	ESE	ESE	E	NE	NNE	NNE	NE	NE	NE	NE	NE	NE	SE	ENE	24	
2	SE	SE	SE	SE	SE	SE	SE	ENE	E	SE	E	ENE	NNE	NNE	NE	NNE	E	ENE	ESE	SE	SE	SE	SE	SE	ESE	24	
3	SE	ESE	ESE	E	ENE	ENE	E	ESE	NE	ENE	E	ESE	SE	ESE	SE	SW	SW	W	N	E	NNW	N	NNE	NE	E	24	
4	NE	NE	NE	NNE	NE	NE	NE	NE	NNE	NE	NE	NNE	NNE	NE	NE	NE	NE	NE	NNE	ENE	NE	NE	NE	NE	NE	24	
5	NE	ENE	NE	NE	NE	NE	NE	NE	NE	NE	NE	NNE	NNE	NNE	NNE	NNE	N	ENE	NE	NE	NE	NE	NE	ESE	NE	24	
6	NE	ENE	NE	NE	NNE	NE	NE	NNE	NW	N	NNE	NNE	NNE	N	N	N	NNW	NW	NW	NNW	NNW	NNW	NNW	NNW	N	24	
7	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	N	N	NNW	N	N	NNE	N	N	NNW	NNW	NNW	24
8	NNW	NNE	NNE	NE	NNE	NNW	N	NNE	NE	NE	NNE	NE	ENE	ENE	ENE	NE	E	WSW	WSW	NNW	NE	NE	NE	NE	NE	24	
9	NE	NE	NE	NE	NNE	E	NE	N	NE	ENE	ENE	E	ESE	ESE	E	SE	SE	SSE	SW	ENE	ENE	NE	NE	ENE	E	24	
10	ENE	NE	NE	NE	ENE	ENE	E	E	E	E	ESE	E	E	ESE	E	E	ESE	ESE	ESE	ENE	E	E	E	ESE	E	24	
11	ESE	ESE	ESE	E	SW	E	SE	X	X	SE	E	ESE	E	E	E	E	E	E	E	E	E	E	E	E	ESE	22	
12	E	E	E	E	E	E	E	E	ESE	ESE	ESE	ESE	ESE	ESE	E	ESE	SE	SSE	SSE	SE	SE	SSE	SSE	SSE	ESE	24	
13	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SSE	SSE	SSE	SSE	SSE	SSE	SE	SE	SE	SE	SSE	SE	24	
14	S	SE	SSE	SE	SE	SE	SE	SE	SSE	S	SE	ESE	ESE	E	NE	NNE	NE	ESE	SE	SE	SE	ESE	ESE	ESE	ESE	24	
15	ESE	SE	SE	SE	SE	SE	SE	SE	S	WSW	WSW	WSW	W	SW	WSW	WSW	WSW	W	SW	NNW	NNW	NNE	NNE	N	SSW	24	
16	N	NE	E	ENE	ESE	ESE	SSW	WSW	NE	ENE	NE	NE	NNE	NNE	N	N	NNE	NNE	NNE	NE	NE	NE	ENE	ESE	NE	24	
17	E	E	SE	SE	ESE	ESE	SE	SE	SE	SE	SE	S	SSW	SSW	SSW	SSW	SE	SE	SE	SE	SE	SE	SE	SE	SSE	24	
18	SE	SE	SE	SE	SE	ESE	ESE	ESE	ESE	SE	SE	SE	SE	SE	SE	ESE	ESE	E	E	E	E	WNW	NW	ESE	24		
19	WNW	W	WNW	WNW	WNW	WNW	W	W	W	W	W	W	W	WNW	WNW	W	W	W	WSW	WSW	WSW	W	WNW	NW	W	24	
20	NW	NNW	NW	NW	NW	NW	NW	NW	NW	NW	NW	NW	NW	NNW	W	WSW	WSW	WSW	WSW	SW	WSW	SE	ESE	ESE	NW	24	
21	E	E	ESE	ESE	ESE	ESE	ESE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	24	
22	ESE	ESE	SE	ESE	E	E	E	E	E	E	E	E	ENE	E	ENE	NE	NNE	N	NNW	W	W	W	W	WSW	E	24	
23	WSW	WSW	W	WSW	WSW	W	WNW	NE	NNE	NE	NE	NNE	NE	N	N	NE	NE	NE	NE	E	SW	S	SE	WSW	NNE	24	
24	E	SSW	ESE	SSW	W	W	SSE	SE	SE	ESE	ESE	SE	SE	ESE	E	E	ESE	E	ESE	ESE	ESE	E	E	E	ESE	24	
25	E	E	E	E	ENE	ENE	N	NNE	N	W	WSW	W	WSW	SW	SW	WSW	WSW	SSE	SSW	ESE	ESE	E	E	SE	SE	24	
26	SE	SE	SE	ENE	NNE	SE	ENE	ESE	NNW	ESE	ESE	ESE	ESE	SE	SSW	SW	SW	SW	SSE	ESE	ESE	ESE	ESE	ESE	SSE	24	
27	E	NE	E	E	ESE	E	E	ESE	ESE	ESE	ESE	SE	ESE	ESE	SE	E	E	E	ESE	ESE	SE	SE	SE	SE	ESE	24	
28	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	ESE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	E	NNE	SE	24
29	NE	W	NNW	WNW	E	NE	ESE	SE	SSE	SW	NNW	W	W	WSW	W	NNW	ESE	NE	ENE	ESE	SE	ESE	ESE	ESE	SE	24	
30	ESE	E	E	E	E	E	E	E	E	E	E	E	NE	NE	NNE	NE	NE	ENE	NNE	NNW	WSW	N	ESE	S	ENE	24	
31	E	S	W	SE	SE	SE	ESE	SE	SSE	SW	SW	SSW	SW	WSW	SSW	SSW	SSW	S	SE	SE	SE	SE	SSW	SW	S	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION:	April 1, 2015
DECLINATION :	MAGNETIC DECLINATION 14 DEGREE EAST

MONTHLY CALIBRATION TIME:	0	hrs	OPERATIONAL TIME:	742	hrs
STANDARD DEVIATION:	90		AMD OPERATION UPTIME:	99.7	%
			MONTHLY AVERAGE:	89	(E)

WIND DIRECTION Hourly Averages (WD)



STANDARD DEVIATION WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake Continuous Monitoring Station - March 2017

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59		
DAY																										
1	18	18	18	20	19	20	20	19	19	21	22	23	21	24	23	20	19	21	23	29	22	24	43	19	24	
2	12	14	15	13	12	19	24	31	21	22	30	27	19	15	23	21	21	18	18	19	16	17	16	14	24	
3	18	18	18	15	15	30	20	17	29	21	19	23	24	21	19	35	25	20	36	41	41	16	16	22	24	
4	19	18	20	22	21	20	20	18	20	21	20	20	21	22	20	21	18	20	22	18	21	21	18	20	24	
5	19	17	19	20	19	17	18	20	20	20	21	24	22	28	25	26	26	22	56	21	20	21	18	26	24	
6	30	20	22	20	21	20	18	23	15	15	20	18	16	17	20	17	19	19	14	12	14	13	14	17	24	
7	15	15	15	14	14	14	14	14	13	16	16	21	17	16	16	17	17	15	19	17	15	17	17	15	24	
8	20	26	19	19	27	22	27	20	20	19	24	19	21	19	21	19	23	46	23	26	72	32	20	23	24	
9	27	28	22	25	42	65	53	77	49	26	24	25	30	47	42	30	25	24	34	43	71	53	60	52	24	
10	51	29	26	29	38	22	25	32	24	24	25	21	22	24	22	20	24	24	26	42	24	18	22	28	24	
11	24	29	25	42	55	28	21	X	X	25	24	26	25	26	25	23	20	19	18	17	18	20	18	18	22	
12	20	18	21	17	18	17	19	18	22	21	25	26	27	28	25	21	25	19	22	21	14	22	27	24	24	
13	16	12	14	17	16	16	14	15	15	15	17	20	21	33	25	25	24	24	25	23	23	21	49	60	24	
14	69	47	58	17	25	16	22	14	41	32	20	28	22	22	17	19	21	26	18	21	21	22	22	22	24	
15	21	17	16	14	17	16	18	16	51	40	24	24	30	62	19	20	18	38	71	62	47	25	21	17	24	
16	44	24	49	39	49	15	38	45	33	22	19	22	20	21	22	23	20	19	19	19	18	17	38	26	24	
17	16	20	21	17	21	26	36	19	21	14	20	24	41	36	35	39	35	21	15	15	14	14	15	14	24	
18	15	17	19	19	22	22	22	19	22	21	20	21	18	19	22	24	24	18	17	16	19	24	17	16	24	
19	20	20	19	19	19	20	20	19	21	20	21	22	22	22	23	20	19	21	18	18	19	18	21	13	24	
20	17	16	15	15	15	14	15	16	17	18	25	32	31	33	30	29	22	21	17	25	32	34	44	42	24	
21	34	59	19	33	33	25	24	15	18	20	17	18	18	17	20	22	16	18	20	17	18	22	23	23	24	
22	23	22	21	21	20	20	21	18	20	19	19	18	21	22	21	20	17	15	14	17	17	17	16	15	24	
23	13	17	20	35	19	38	56	49	21	24	26	33	61	26	27	25	21	20	19	41	38	25	33	32	24	
24	41	65	48	45	41	42	56	18	16	23	22	21	24	23	22	22	21	19	21	22	23	19	19	18	24	
25	18	19	20	22	21	20	20	19	20	35	21	26	28	24	28	25	43	52	32	54	51	25	16	18	24	
26	16	13	17	58	53	44	42	43	71	31	36	28	33	49	40	25	27	24	18	29	47	21	18	14	24	
27	13	31	16	16	20	19	18	22	24	25	23	23	23	22	21	20	19	19	20	22	24	16	15	19	24	
28	17	19	16	13	15	17	15	20	19	16	22	22	24	18	26	20	17	26	16	15	14	19	64	51	24	
29	66	45	48	60	49	54	27	21	36	53	25	24	25	27	39	43	44	20	16	20	19	22	23	21	24	
30	22	20	22	20	20	20	21	19	20	21	21	24	24	23	27	31	28	24	24	43	45	77	40	43	24	
31	59	57	55	26	14	15	48	19	35	37	55	59	34	23	35	34	37	37	17	11	12	15	39	25	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION: April 1, 2015

CALIBRATION TIME: 0 hrs OPERATIONAL TIME: 742 hrs

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)



RELATIVE HUMIDITY




RELATIVE HUMIDITY Hourly Averages (RH %)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.					
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.						
DAY																																	
1	66	65	59	59	60	62	63	62	62	62	58	56	55	56	58	58	60	62	66	69	72	74	76	75	55	76	63	24					
2	74	75	74	75	77	78	78	76	76	70	65	62	59	58	55	54	54	60	66	69	72	75	76	76	54	78	69	24					
3	77	79	82	83	85	87	86	87	84	76	69	64	56	49	41	41	45	53	66	74	79	78	73	77	41	87	70	24					
4	83	85	82	80	82	79	76	75	78	80	75	75	74	74	76	75	78	76	72	71	72	72	73	71	71	85	76	24					
5	70	69	70	70	71	72	71	72	76	74	71	69	67	64	63	64	63	65	68	72	73	75	77	75	63	77	70	24					
6	74	75	76	76	77	75	81	83	84	84	85	86	86	86	87	88	88	87	86	86	85	84	84	83	74	88	83	24					
7	83	82	81	80	79	79	78	75	70	65	59	51	42	39	40	40	42	44	48	51	52	54	56	59	39	83	60	24					
8	62	62	59	58	59	65	63	59	56	52	45	44	45	44	45	43	45	45	54	61	65	63	51	57	43	65	54	24					
9	61	64	63	66	71	72	72	70	66	58	51	45	42	35	34	32	33	34	48	59	65	69	71	71	32	72	56	24					
10	71	71	71	70	70	63	58	59	54	47	41	39	37	34	35	38	41	48	50	56	53	52	56	63	34	71	53	24					
11	65	70	74	75	75	74	69	X	X	66	65	63	61	60	62	63	65	68	70	71	72	74	75	76	60	76	69	22					
12	77	76	75	75	74	75	74	73	69	63	59	56	52	50	48	48	49	51	54	57	61	63	66	68	48	77	63	24					
13	71	73	74	75	74	75	75	73	71	66	60	54	52	49	48	51	53	58	64	67	68	70	71	73	48	75	65	24					
14	75	79	80	81	83	84	85	83	71	63	60	53	49	47	53	53	54	57	60	64	69	69	70	73	47	85	67	24					
15	74	76	78	81	85	88	90	89	80	73	67	61	53	48	50	53	53	56	63	72	74	58	44	45	44	90	67	24					
16	56	66	72	78	85	87	89	90	80	79	78	74	69	62	55	53	53	53	50	54	53	54	60	57	50	90	67	24					
17	53	56	55	59	60	63	62	62	63	63	60	55	52	53	51	48	54	62	71	76	78	79	81	82	48	82	62	24					
18	83	84	84	84	84	83	83	83	78	71	67	65	62	61	59	58	61	64	68	70	79	80	87	86	58	87	74	24					
19	95	96	93	87	77	70	72	68	63	61	57	54	49	45	42	64	69	55	70	78	80	76	68	59	42	96	69	24					
20	50	49	50	45	49	50	51	52	51	51	48	44	39	36	33	29	31	33	40	49	61	66	69	74	29	74	48	24					
21	77	79	68	69	72	71	71	67	64	58	51	45	43	43	45	47	50	52	56	54	55	58	60	61	43	79	59	24					
22	62	64	70	73	76	77	79	78	78	78	86	91	93	94	94	92	92	94	94	94	95	95	95	94	62	95	85	24					
23	94	94	94	94	93	92	92	92	90	89	88	86	83	77	78	77	77	79	82	84	86	87	87	89	77	94	87	24					
24	89	89	91	92	90	88	87	92	92	93	93	90	81	74	70	67	67	70	71	70	71	72	75	77	67	93	81	24					
25	79	80	82	82	82	82	84	83	79	77	75	70	68	67	62	62	60	61	70	82	88	91	91	91	60	91	77	24					
26	90	88	88	89	92	92	92	88	81	78	68	62	56	48	43	43	44	47	54	68	76	84	85	84	43	92	73	24					
27	83	88	90	91	88	87	87	83	76	71	64	56	55	55	50	53	52	53	54	58	64	77	80	81	50	91	71	24					
28	82	82	82	83	85	86	87	87	87	86	83	80	76	73	67	65	67	73	79	85	90	91	91	92	65	92	82	24					
29	95	97	98	98	99	99	100	100	100	100	100	90	86	83	77	69	67	71	79	87	92	97	99	100	67	100	91	24					
30	100	100	100	100	100	100	100	100	100	100	100	93	87	85	81	73	68	67	73	83	88	92	94	95	67	100	91	24					
31	96	96	95	96	96	96	96	88	82	74	65	57	49	48	46	43	48	54	62	65	64	64	59	60	43	96	71	24					
HOURLY MAX	100	100	100	100	100	100	100	100	100	100	100	93	93	94	94	92	92	94	94	94	94	95	97	99	100								
HOURLY AVG	76	78	78	78	79	79	79	78	75	72	68	64	61	58	56	56	58	60	65	70	73	74	74	75									

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	29	%	@ HOUR(S)	15	ON DAY(S)	20
MAXIMUM 1-HR AVERAGE:	100	%	@ HOUR(S)	6	ON DAY(S)	29
MAXIMUM 24-HR AVERAGE:	91	%			ON DAY(S)	29
					VAR-VARIOUS	
				OPERATIONAL TIME:		742 hrs
				AMD OPERATION UPTIME:		99.7 %
STANDARD DEVIATION:	16			MONTHLY AVERAGE:		70 %

RELATIVE HUMIDITY Hourly Averages (RH %)



AMBIENT TEMPERATURE




AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	-16.0	-17.5	-18.0	-18.3	-18.7	-18.8	-18.6	-18.4	-18.0	-17.7	-17.1	-16.1	-15.0	-14.1	-13.5	-13.0	-13.2	-14.2	-15.8	-16.9	-17.5	-17.6	-17.3	-16.9	-18.8	-13.0	-16.6	24	
2	-17.1	-17.2	-17.5	-18.3	-19.1	-20.3	-21.0	-22.1	-19.5	-16.9	-14.7	-13.1	-12.8	-11.4	-9.0	-7.8	-6.5	-7.6	-8.1	-7.1	-7.8	-8.5	-8.9	-8.9	-22.1	-6.5	-13.4	24	
3	-8.8	-9.8	-10.8	-11.4	-12.0	-12.6	-13.4	-12.4	-11.2	-8.7	-6.0	-2.9	0.1	2.7	5.2	5.2	4.5	2.9	-0.6	-2.8	-4.8	-5.1	-2.3	-2.6	-13.4	5.2	-4.9	24	
4	-4.3	-3.7	-2.8	-3.5	-5.6	-8.8	-10.7	-12.6	-14.1	-15.2	-15.5	-15.5	-15.2	-15.5	-15.4	-15.4	-15.0	-15.5	-16.1	-16.3	-16.1	-16.2	-16.3	-16.3	-16.3	-2.8	-12.5	24	
5	-16.5	-16.6	-16.8	-16.9	-17.1	-17.3	-17.3	-17.1	-16.7	-16.5	-16.1	-15.2	-14.6	-13.7	-13.1	-12.6	-12.3	-12.2	-12.5	-13.0	-13.6	-13.2	-12.8	-12.7	-17.3	-12.2	-14.9	24	
6	-13.0	-13.5	-13.7	-13.9	-14.1	-14.3	-14.6	-14.2	-13.5	-13.0	-12.7	-12.3	-11.9	-11.3	-10.8	-10.4	-10.3	-10.9	-12.0	-12.5	-13.6	-14.0	-14.3	-14.5	-14.6	-10.3	-12.9	24	
7	-14.8	-15.2	-15.9	-16.4	-16.5	-16.5	-16.6	-16.6	-16.0	-14.9	-13.4	-11.8	-11.1	-10.9	-11.5	-11.7	-12.2	-13.7	-15.4	-17.1	-18.6	-20.2	-21.4	-21.4	-21.4	-10.9	-15.2	24	
8	-22.5	-23.0	-23.4	-24.1	-24.7	-26.3	-26.2	-25.4	-24.9	-24.1	-22.2	-20.9	-19.8	-18.7	-17.7	-16.8	-16.5	-16.6	-18.2	-20.2	-21.2	-21.3	-19.9	-21.6	-26.3	-16.5	-21.5	24	
9	-23.0	-24.1	-23.9	-25.1	-27.1	-29.7	-31.2	-31.1	-26.5	-24.6	-23.3	-21.6	-19.8	-18.6	-17.7	-17.2	-17.2	-17.4	-20.1	-23.2	-25.3	-27.0	-28.1	-28.8	-31.2	-17.2	-23.8	24	
10	-29.3	-28.3	-28.1	-28.4	-27.9	-25.4	-24.4	-24.6	-23.3	-21.2	-19.3	-18.7	-17.9	-17.1	-16.4	-16.5	-16.8	-17.1	-17.4	-18.1	-18.0	-18.1	-18.3	-18.6	-29.3	-16.4	-21.2	24	
11	-18.9	-19.7	-21.3	-22.8	-22.4	-20.7	-19.9	X	X	-19.2	-18.2	-17.0	-15.8	-15.0	-14.4	-14.1	-14.1	-14.6	-15.3	-15.5	-15.4	-15.5	-15.6	-22.8	-14.1	-17.3	22		
12	-15.9	-16.1	-16.4	-16.8	-17.0	-17.3	-17.6	-17.6	-16.9	-15.8	-14.6	-13.3	-12.1	-11.0	-10.5	-9.9	-9.9	-10.1	-11.3	-12.4	-13.3	-13.5	-13.8	-14.4	-17.6	-9.9	-14.1	24	
13	-15.1	-15.8	-15.8	-15.7	-15.4	-15.2	-15.1	-14.7	-13.9	-12.4	-10.7	-9.0	-8.1	-6.6	-5.9	-6.5	-6.7	-7.7	-8.9	-9.4	-9.5	-9.6	-9.9	-9.9	-15.8	-5.9	-11.1	24	
14	-10.2	-11.7	-11.2	-11.3	-11.7	-11.8	-12.2	-11.6	-8.6	-6.4	-5.4	-2.8	-0.2	1.7	0.3	-0.2	-0.1	0.7	1.7	0.3	-0.3	-0.1	0.1	0.0	-12.2	1.7	-4.6	24	
15	-0.4	-0.8	-1.3	-1.9	-2.7	-3.4	-4.0	-3.5	-1.6	0.0	1.7	3.7	6.2	7.9	7.5	6.5	6.4	5.7	4.0	2.3	1.7	3.2	4.8	4.9	-4.0	7.9	2.0	24	
16	4.7	3.9	2.9	2.8	2.4	2.1	1.8	1.3	3.6	3.8	3.6	4.2	4.6	5.0	5.5	5.8	5.0	3.9	2.9	2.1	1.7	1.0	-0.1	-0.2	-0.2	5.8	3.1	24	
17	-0.2	-0.6	-0.3	-0.7	-0.8	-1.2	-1.4	-1.4	-0.4	0.7	3.1	5.2	6.7	6.9	7.1	7.0	6.0	4.9	3.8	3.2	2.9	2.7	2.3	1.7	-1.4	7.1	2.4	24	
18	1.3	0.8	0.6	0.5	0.5	0.7	0.4	0.0	1.1	2.7	3.9	4.8	5.5	6.2	6.9	7.5	7.1	6.4	5.0	4.8	4.1	3.8	2.6	2.9	0.0	7.5	3.3	24	
19	1.2	1.7	1.7	1.1	0.7	0.1	-0.6	-0.9	-0.5	0.3	1.1	1.8	2.9	3.6	3.9	1.5	0.7	1.2	0.1	-0.8	-1.5	-1.9	-2.3	-3.2	-3.2	3.9	0.5	24	
20	-4.4	-7.7	-8.6	-9.6	-10.4	-10.8	-11.2	-11.2	-10.7	-10.4	-9.4	-8.3	-7.1	-6.2	-5.0	-4.4	-4.6	-4.9	-6.1	-7.8	-9.6	-10.8	-11.8	-12.8	-12.8	-4.4	-8.5	24	
21	-13.4	-13.9	-13.0	-13.7	-14.6	-14.5	-14.2	-12.6	-10.7	-8.0	-5.0	-2.9	-1.8	-1.3	-1.0	-0.8	-1.1	-1.6	-1.9	-1.9	-2.0	-2.1	-2.2	-2.3	-14.6	-0.8	-6.5	24	
22	-2.3	-2.5	-2.8	-3.1	-3.7	-4.1	-4.5	-4.4	-4.0	-3.7	-4.2	-4.4	-4.2	-4.1	-3.8	-3.6	-3.9	-4.1	-4.1	-4.0	-3.9	-3.7	-3.9	-4.4	-4.5	-2.3	-3.8	24	
23	-5.1	-5.3	-4.6	-4.5	-6.1	-6.8	-6.2	-4.2	-3.0	-3.7	-3.9	-3.0	-2.1	-1.1	-1.4	-1.3	-1.4	-1.6	-1.8	-1.8	-1.8	-2.1	-2.4	-6.8	-1.1	-3.2	24		
24	-2.7	-3.0	-4.9	-7.2	-9.5	-10.8	-10.8	-7.7	-7.1	-6.2	-5.3	-3.7	-1.5	0.6	2.0	3.2	3.5	2.9	2.2	2.0	1.6	1.1	0.4	-0.3	-10.8	3.5	-2.6	24	
25	-0.7	-0.9	-1.0	-0.9	-1.0	-0.9	-1.1	-0.8	-0.1	0.8	1.5	2.8	3.4	3.6	4.5	4.6	5.0	4.7	2.8	-0.3	-1.9	-2.8	-3.2	-2.6	-3.2	5.0	0.6	24	
26	-2.5	-1.6	-1.6	-2.0	-3.1	-3.6	-3.2	-1.7	0.4	1.0	3.2	4.8	6.4	7.9	8.6	8.7	8.5	8.1	6.3	2.6	0.3	-0.9	-1.2	-0.3	-3.6	8.7	1.9	24	
27	0.3	-2.2	-2.8	-2.0	-0.8	-0.9	-1.2	-0.8	0.1	1.1	2.7	4.1	4.7	5.4	6.3	6.1	6.0	5.6	4.9	4.1	3.6	3.3	2.9	2.7	-2.8	6.3	2.2	24	
28	2.6	2.2	1.9	1.8	1.7	1.7	1.6	1.9	2.1	2.7	3.3	3.8	4.8	5.8	7.5	8.0	7.7	6.3	5.1	4.1	3.3	3.0	2.7	2.1	1.6	8.0	3.7	24	
29	1.9	1.8	1.8	1.8	1.7	1.6	1.5	1.5	1.8	2.4	3.3	4.1	4.3	4.6	5.8	7.6	8.0	6.8	4.6	3.1	2.1	1.8	1.6	1.0	1.0	8.0	3.2	24	
30	0.7	0.8	0.6	0.1	0.0	0.0	0.1	0.1	0.4	0.9	1.4	2.2	2.8	3.1	3.6	5.1	6.0	6.3	4.9	2.3	0.5	-0.6	-1.2	-1.8	-1.8	6.3	1.6	24	
31	-1.7	-1.2	-1.3	-1.8	-1.0	-1.3	-1.5	0.7	2.6	5.0	7.8	9.9	11.0	11.6	11.2	11.1	10.5	10.4	8.8	7.5	7.2	7.0	7.7	7.3	-1.8	11.6	5.3	24	
HOURLY MAX	4.7	3.9	2.9	2.8	2.4	2.1	1.8	1.9	3.6	5.0	7.8	9.9	11.0	11.6	11.2	11.1	10.5	10.4	8.8	7.5	7.2	7.0	7.7	7.3					
HOURLY AVG	-7.9	-8.4	-8.7	-9.1	-9.5	-9.9	-10.1	-9.4	-8.3	-7.7	-6.5	-5.2	-4.1	-3.2	-2.6	-2.4	-2.5	-2.9	-4.1	-5.2	-6.0	-6.3	-6.5	-6.8					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	-31.2 °C	@ HOUR(S)	6	ON DAY(S)	9
MAXIMUM 1-HR AVERAGE:	11.6 °C	@ HOUR(S)	13	ON DAY(S)	31
MAXIMUM 24-HR AVERAGE:	5.3 °C			ON DAY(S)	31
				VAR-VARIOUS	
OPERATIONAL TIME:				742	hrs
AMD OPERATION UPTIME:				99.7	%
STANDARD DEVIATION:	9.3			MONTHLY AVERAGE:	-6.4 °C

AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)




APPENDIX II
NON-CONTINUOUS MONITORING DATA RESULTS

PASSIVE RESULTS

Passive Summary Results for February 2017 - March 2017 Lakeland Industry & Community Association

		Sulphur Dioxide ppb																				February - March 2017								
		2016																				Reading	Site							
Mean		2	3	4	5	6	8	9	10	11	12	13	14	15	16	17	18	19	22	23	24	25	26	27	28	29	32	38		
Mean	NA	0.2	0.4	0.4	0.4	0.3	0.3	0.3	0.3	0.3	NA	0.3	1.0	0.3	0.3	0.4	0.3	0.3	0.2	0.1	0.3	NA	0.7	1.0	0.5	0.3	0.3	0.5	0.5	-
Minimum	NA	0.1	0.2	0.3	0.2	0.2	0.2	0.2	0.2	NA	0.2	0.8	0.2	0.2	0.3	0.1	0.1	0.2	0.1	0.2	NA	0.4	0.8	0.3	0.2	0.2	0.3	0.2	#23	
Maximum	NA	0.4	0.6	0.5	0.6	0.5	0.3	0.4	0.3	NA	0.5	1.3	0.4	0.5	0.6	0.4	0.6	0.4	0.2	0.2	0.4	NA	1.3	1.3	0.6	0.4	0.6	0.7	1.3	#14



Lakeland Industry & Community Association SO₂ Passive Bubble Map

FEBRUARY 2017 - MARCH 2017


PASSIVE STATIONS

		DUPLICATE
2 – Sand River	MISSING	NA
3 – Therien	0.4 PPB	NA
4 – Flat Lake	0.7 PPB	NA
5 – Lake Eliza	0.5 PPB	NA
6 – Telegraph Creek	0.4 PPB	NA
8 – Muriel-Kehewin	0.5 PPB	NA
9 – Dupre	0.4 PPB	NA
10 – La Corey	0.5 PPB	NA
11 – Wolf Lake	MISSING	NA
12 – Foster Creek	MISSING	NA
13 – Primrose	0.4 PPB	NA
14 – Maskwa	1.3 PPB	NA
15 – Ardmore	0.4 PPB	NA
16 – Frog Lake	0.4 PPB	0.3 PPB
17 – Clear Range	0.4 PPB	0.4 PPB
18 – Fishing Lake	0.3 PPB	0.3 PPB
19 – Beaverdam	0.4 PPB	NA
22 – Cold Lake South	0.3 PPB	NA
23 – Medley-Martineau	0.2 PPB	NA
24 – Fort George	0.5 PPB	NA
25 – Burnt Lake	MISSING	NA
26 – Mahikan	0.9 PPB	NA
27 – Mahkeses	1.0 PPB	NA
28 – Town of Bonnyville	0.4 PPB	NA
29 – Cold Lake South 2	0.3 PPB	NA
32 – St. Lina	0.5 PPB	NA
38 – Bonnyville	0.5 PPB	NA




Summary

Minimum : 0.2 PPB – Medley-Martineau
 Maximum: 1.3 PPB – Maskwa
 Average: 0.5 PPB *Includes Duplicates



Passive Summary Results for February 2017 - March 2017 Lakeland Industry & Community Association

		Hydrogen Sulphide ppb																	February - March 2017		
		3	5	10	11	12	13	14	16	17	18	22	24	25	26	27	29	32	38	Reading	Site
Mean		0.11	0.30	0.14	0.05	NA	0.09	0.14	0.13	0.22	0.11	0.13	0.13	NA	0.15	0.39	0.13	0.14	1.86	0.13	-
Minimum		0.08	0.13	0.09	0.05	NA	0.08	0.10	0.09	0.14	0.05	0.06	0.09	NA	0.14	0.15	0.08	0.08	0.16	0.08	#18
Maximum		0.17	0.51	0.21	0.06	NA	0.12	0.17	0.16	0.33	0.14	0.20	0.19	NA	0.17	0.67	0.17	0.20	3.61	0.26	#38




Lakeland Industry & Community Association H₂S Passive Bubble Map

FEBRUARY 2017 - MARCH 2017


PASSIVE STATIONS

Station Number	Concentration	Duplicate
3 - Therien	0.12 PPB	NA
5 - Lake Eliza	0.12 PPB	NA
10 - La Corey	0.09 PPB	NA
11 - Wolf Lake	MISSING	NA
12 - Foster Creek	MISSING	NA
13 - Primrose	0.10 PPB	NA
14 - Maskwa	0.13 PPB	NA
16 - Frog Lake	0.11 PPB	NA
17 - Clear Range	0.16 PPB	NA
18 - Fishing Lake	0.08 PPB	NA
22 - Cold Lake South	0.09 PPB	0.10 PPB
24 - Fort George	0.11 PPB	0.12 PPB
25 - Burnt Lake	MISSING	NA
26 - Mahihkan	0.14 PPB	NA
27 - Mahkeses	0.13 PPB	NA
29 - Cold Lake South 2	0.11 PPB	NA
32 - St. Lina	0.12 PPB	NA
38 - Bonnyville	0.26 PPB	NA




Summary

Minimum : 0.08 PPB - Fishing Lake
 Maximum: 0.26 PPB - Bonnyville
 Average: 0.13 PPB *Includes Duplicates



Passive Summary Results for February 2017 - March 2017 Lakeland Industry & Community Association


		Nitrogen Dioxide ppb																				February - March 2017				
		2016																				Reading	Site			
Mean	2	3	4	5	6	8	9	10	11	12	13	14	15	16	17	18	19	22	23	24	28	29	32	38	1.0	-
Minimum	NA	0.5	0.3	0.4	2.1	0.3	0.2	1.2	0.3	NA	0.2	0.5	0.3	0.5	0.6	0.4	0.4	0.5	0.1	1.6	1.4	0.4	0.2	1.3	0.4	Various
Maximum	NA	1.8	1.4	2.2	5.0	1.4	2.2	3.7	1.1	NA	2.6	2.6	1.4	2.3	2.1	1.4	1.5	2.4	0.4	4.6	4.9	2.3	1.1	4.2	3.0	#28



Lakeland Industry & Community Association NO₂ Passive Bubble Map FEBRUARY 2017 – MARCH 2017


PASSIVE STATIONS

		DUPLICATE
2 – Sand River	MISSING	NA
3 – Therien	0.7 PPB	NA
4 – Flat Lake	0.5 PPB	NA
5 – Lake Eliza	0.9 PPB	NA
6 – Telegraph Creek	1.6 PPB	1.8 PPB
8 – Muriel-Kehewin	0.9 PPB	0.8 PPB
9 – Dupre	0.4 PPB	NA
10 – La Corey	1.2 PPB	NA
11 – Wolf Lake	MISSING	NA
12 – Foster Creek	MISSING	NA
13 – Primrose	0.4 PPB	NA
14 – Maskwa	0.7 PPB	NA
15 – Ardmore	0.5 PPB	NA
16 – Frog Lake	0.6 PPB	NA
17 – Clear Range	0.9 PPB	NA
18 – Fishing Lake	0.5 PPB	NA
19 – Beaverdam	0.4 PPB	NA
22 – Cold Lake South	0.9 PPB	NA
23 – Medley-Martineau	<0.1 PPB	NA
24 – Fort George	2.0 PPB	NA
28 – Town of Bonnyville	3.0 PPB	NA
29 – Cold Lake South 2	1.1 PPB	NA
32 – St. Lina	0.6 PPB	NA
38 – Bonnyville	2.6 PPB	NA




Summary

Minimum : 0.4 PPB – Various Stations
Maximum: 3.0 PPB – Town of Bonnyville
Average: 1.0 PPB *Includes Duplicates



Passive Summary Results for February 2017 - March 2017 Lakeland Industry & Community Association

	Ozone ppb																												February - March 2017	
	2	3	4	5	6	8	9	10	11	12	2016	13	14	15	16	17	18	19	22	23	24	28	29	32	38	Reading	Site			
Mean	NA	27.9	29.1	25.0	26.7	31.8	27.8	24.2	16.7	NA	27.6	26.7	25.7	24.8	29.5	26.6	28.8	24.8	22.9	26.0	25.0	26.7	31.8	18.9	34.1	-				
Minimum	NA	14.9	17.0	16.0	14.9	19.3	17.7	13.1	11.8	NA	16.5	16.9	14.9	13.0	19.4	15.1	19.0	14.5	12.2	15.9	16.2	16.1	22.9	16.3	25.9	#38				
Maximum	NA	41.8	47.4	32.8	47.3	43.8	43.3	38.7	21.5	NA	37.7	40.0	39.3	35.3	49.3	41.4	46.5	32.6	33.2	40.8	34.2	36.7	39.1	21.5	43.4	#4				




Lakeland Industry & Community Association O₃ Passive Bubble Map

FEBRUARY 2017 – MARCH 2017


PASSIVE STATIONS

		DUPLICATE
2 – Sand River	MISSING	NA
3 – Therien	36.7 PPB	NA
4 – Flat Lake	43.4 PPB	NA
5 – Lake Eliza	38.0 PPB	NA
6 – Telegraph Creek	34.7 PPB	30.3 PPB
8 – Murlei-Kehewin	35.6 PPB	38.1 PPB
9 – Dupre	35.2 PPB	NA
10 – La Corey	29.8 PPB	NA
11 – Wolf Lake	MISSING	NA
12 – Foster Creek	MISSING	NA
13 – Primrose	33.5 PPB	NA
14 – Maskwa	30.9 PPB	NA
15 – Ardmore	39.0 PPB	NA
16 – Frog Lake	37.0 PPB	NA
17 – Clear Range	35.6 PPB	NA
18 – Fishing Lake	33.7 PPB	NA
19 – Beaverdam	31.8 PPB	NA
22 – Cold Lake South	31.3 PPB	NA
23 – Medley-Martineau	33.3 PPB	NA
24 – Fort George	32.5 PPB	NA
28 – Town of Bonnyville	29.8 PPB	NA
29 – Cold Lake South 2	33.1 PPB	NA
32 – St. Lina	36.8 PPB	NA
38 – Bonnyville	25.9 PPB	NA



Summary

Minimum : 25.9 PPB – Bonnyville
 Maximum: 43.4 PPB – Flat Lake
 Average: 34.1 PPB *Includes Duplicates



VOC RESULTS

Maxxam Analytics

VOC Sample Collection Data Sheet Alberta Air FCD AIR FCD-01320 / 2

Client: LICA Sampler S/N: 6167
 Location: Cold Lake South Canister ID: (AY) 2449 35647
 Station ID: LICA 01 Installation Date/Time (mst): Feb 27, 2017 @ 10:53
 Sample ID: LICA/VOC/CLS/Mar 02, 2017 Removal Date/Time (mst): Mar 06, 2017 @ 16:32

Date and Time Information

Sample Date:	Start Time (mst)	End Time (mst)	Elapsed Time (hours)
<u>Mar 02, 2017</u>	<u>00:00</u>	<u>00:00 Mar 03, 2017</u>	<u>24.0</u>

Canister Pressure/Vacuum	
Initial Vacuum (in. Hg)	Final Pressure (psi)
<u>-27.6</u>	<u>+23.5</u>

Flow Settings		
Flow Reading (sccm)	Pot Set Point	Pump Set (psi)
<u>10.0</u>	<u>6.52</u>	<u>24.0</u>

Deployment/Collection and Maintenance Checklist

Initial leak check deployment vacuum (in. Hg) = - @ - mst
 Final leak check deployment vacuum (in. Hg) = - @ - mst
 Total leak rate = - psi over - minutes
 Timer reset to zero prior to sampling? YES (yes/no)
 Date of last flow calibration: Dec 19, 2016 (due every 3 months)
 Last date of sample line & fitting replacement: Nov 09, 2016 (due every 6 months)

Leak rate must be 0.0 psi over a minimum of 5 minutes or repair is required

Comments: n/a

Deployment Technician Signature: Alex Yakupov

Collection Technician Signature: Alex Yakupov Date: Mar 06, 2017

Sample ID: 17030076-001

Customer ID: LICA
 Cust Samp ID: LICA/VOC/CLS/Mar 2, 2017

Priority: Normal



Volatile Organics Data Results

Date: March 2, 2017
Canister ID: S5647

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.02
1,1,2,2-Tetrachloroethane	< 0.02
1,1,2-Trichloroethane	< 0.02
1,1-Dichloroethane	< 0.02
1,1-Dichloroethylene	< 0.04
1,2,3-Trimethylbenzene	< 0.05
1,2,4-Trichlorobenzene	< 0.8
1,2,4-Trimethylbenzene	< 0.05
1,2-Dibromoethane	< 0.02
1,2-Dichlorobenzene	< 0.03
1,2-Dichloroethane	0.02
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	0.02
1,3-Butadiene	< 0.02
1,3-Dichlorobenzene	< 0.3
1,4-Dichlorobenzene	< 0.4
1,4-Dioxane	< 0.4
1-Butene	0.09
1-Hexene	< 0.02
1-Pentene	0.02
2,2,4-Trimethylpentane	0.07
2,2-Dimethylbutane	0.03
2,3,4-Trimethylpentane	0.03
2,3-Dimethylbutane	0.05
2,3-Dimethylpentane	0.06
2,4-Dimethylpentane	0.04
2-Methylheptane	0.03
2-Methylhexane	0.05
2-Methylpentane	0.2
3-Methylheptane	0.02
3-Methylhexane	0.06
3-Methylpentane	0.07
Acetone	1.6
Acrolein	< 0.3
Benzene	0.22
Benzyl chloride	< 0.4
Bromodichloromethane	< 0.02
Bromoform	< 0.02
Bromomethane	0.01
Carbon disulfide	< 0.01
Carbon tetrachloride	0.13
Chlorobenzene	< 0.02
Chloroethane	< 0.02
Chloroform	0.03
Chloromethane	0.64
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.04
cis-2-Butene	0.03
cis-2-Pentene	< 0.02
Cyclohexane	0.05
Cyclopentane	0.04
Dibromochloromethane	< 0.01
Ethanol	0.8
Ethyl acetate	< 0.4
Ethylbenzene	0.03
Freon-11	0.29
Freon-113	0.13

Volatile Organics Data Results

Date: March 2, 2017
Canister ID: S5647

PARAMETERS	CONCENTRATION (PPB)
Freon-114	0.02
Freon-12	0.61
Hexachloro-1,3-butadiene	< 0.50
Isobutane	0.55
Isopentane	0.45
Isoprene	0.01
Isopropyl alcohol	< 0.4
Isopropylbenzene	< 0.01
m,p-Xylene	0.08
m-Diethylbenzene	< 0.04
m-Ethyltoluene	< 0.08
Methyl butyl ketone	< 0.50
Methyl ethyl ketone	< 0.3
Methyl isobutyl ketone	< 0.4
Methyl methacrylate	< 0.07
Methyl tert butyl ether	< 0.03
Methylcyclohexane	0.06
Methylcyclopentane	0.07
Methylene chloride	< 0.3
n-Butane	1.16
n-Decane	< 0.06
n-Dodecane	< 0.4
n-Heptane	0.05
n-Hexane	0.09
n-Nonane	0.01
n-Octane	0.03
n-Pentane	0.3
n-Propylbenzene	< 0.05
n-Undecane	< 0.5
Naphthalene	< 0.5
o-Ethyltoluene	0.02
o-Xylene	0.04
p-Diethylbenzene	< 0.04
p-Ethyltoluene	< 0.07
Styrene	< 0.04
Tetrachloroethylene	< 0.04
Tetrahydrofuran	< 0.4
Toluene	0.13
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.04
trans-2-Butene	0.03
trans-2-Pentene	0.02
Trichloroethylene	< 0.04
Vinyl acetate	< 0.4
Vinyl chloride	< 0.02

Maxxam Analytics

VOC Sample Collection Data Sheet Alberta Air FCD AIR FCD-01320 / 2

Client: LICA Sampler S/N: 6167
 Location: Cold Lake South Canister ID: H3291
 Station ID: LICA 01 Installation Date/Time (mst): Mar 06, 2017 @ 16:45
 Sample ID: LICA/VOC/CLS/Mar 08, 2017 Removal Date/Time (mst): Mar 09, 2017 @ 10:39

Date and Time Information

Sample Date:	Start Time (mst)	End Time (mst)	Elapsed Time (hours)
<u>Mar 08, 2017</u>	<u>00:00</u>	<u>00:00</u> <u>Mar 09, 2017</u>	<u>24.0</u>

Canister Pressure/Vacuum	
Initial Vacuum (in. Hg)	Final Pressure (psi)
<u>-28.0</u>	<u>+21.0</u>

Flow Settings		
Flow Reading (sccm)	Pot Set Point	Pump Set (psi)
<u>9.3</u>	<u>6.10</u>	<u>24.0</u>

Deployment/Collection and Maintenance Checklist

Initial leak check deployment vacuum (in. Hg) = -28.0 @ 16:45 Mar 06/17 mst
 Final leak check deployment vacuum (in. Hg) = -28.0 @ 08:30 Mar 07/17 mst
 Total leak rate = 0.0 psi over 16 hours -minutes A.Y.
 Timer reset to zero prior to sampling? YES (yes/no)
 Date of last flow calibration: Mar 06, 2017 (due every 3 months)
 Last date of sample line & fitting replacement: Nov 09, 2016 (due every 6 months)

****Leak rate must be 0.0 psi over a minimum of 5 minutes or repair is required****

Comments: The leak check completed on March 07, 2017


Deployment Technician Signature: Alex Yakupov

Collection Technician Signature: Alex Yakupov Date: Mar 09, 2017

Sample ID: 17030114-001

Customer ID: LICA
 Cust Samp ID: LICA/VOC/CLS/Mar 8, 2017

Priority: Normal



Volatile Organics Data Results

Date: March 8, 2017
Canister ID: H3297

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.02
1,1,2,2-Tetrachloroethane	< 0.02
1,1,2-Trichloroethane	< 0.02
1,1-Dichloroethane	< 0.02
1,1-Dichloroethylene	< 0.04
1,2,3-Trimethylbenzene	< 0.05
1,2,4-Trichlorobenzene	< 0.8
1,2,4-Trimethylbenzene	< 0.05
1,2-Dibromoethane	< 0.02
1,2-Dichlorobenzene	< 0.03
1,2-Dichloroethane	0.02
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	0.03
1,3-Butadiene	< 0.02
1,3-Dichlorobenzene	< 0.3
1,4-Dichlorobenzene	< 0.4
1,4-Dioxane	< 0.4
1-Butene	0.07
1-Hexene	0.02
1-Pentene	0.02
2,2,4-Trimethylpentane	0.05
2,2-Dimethylbutane	0.03
2,3,4-Trimethylpentane	0.03
2,3-Dimethylbutane	0.03
2,3-Dimethylpentane	0.04
2,4-Dimethylpentane	0.03
2-Methylheptane	0.02
2-Methylhexane	0.03
2-Methylpentane	0.09
3-Methylheptane	0.02
3-Methylhexane	0.04
3-Methylpentane	0.05
Acetone	1
Acrolein	< 0.3
Benzene	0.25
Benzyl chloride	< 0.4
Bromodichloromethane	< 0.02
Bromoform	< 0.02
Bromomethane	0.01
Carbon disulfide	< 0.01
Carbon tetrachloride	0.13
Chlorobenzene	< 0.02
Chloroethane	< 0.02
Chloroform	0.03
Chloromethane	0.67
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.04
cis-2-Butene	0.03
cis-2-Pentene	< 0.02
Cyclohexane	0.04
Cyclopentane	0.03
Dibromochloromethane	< 0.01
Ethanol	< 0.3
Ethyl acetate	< 0.4
Ethylbenzene	0.02
Freon-11	0.3
Freon-113	0.12

Volatile Organics Data Results

Date: March 8, 2017
Canister ID: H3297

PARAMETERS	CONCENTRATION (PPB)
Freon-114	0.02
Freon-12	0.63
Hexachloro-1,3-butadiene	< 0.50
Isobutane	0.3
Isopentane	0.23
Isoprene	0.02
Isopropyl alcohol	1.6
Isopropylbenzene	0.02
m,p-Xylene	0.07
m-Diethylbenzene	< 0.04
m-Ethyltoluene	< 0.08
Methyl butyl ketone	< 0.50
Methyl ethyl ketone	< 0.3
Methyl isobutyl ketone	< 0.4
Methyl methacrylate	< 0.07
Methyl tert butyl ether	< 0.03
Methylcyclohexane	0.04
Methylcyclopentane	0.04
Methylene chloride	< 0.3
n-Butane	0.66
n-Decane	< 0.06
n-Dodecane	< 0.4
n-Heptane	0.04
n-Hexane	0.06
n-Nonane	0.03
n-Octane	0.03
n-Pentane	0.1
n-Propylbenzene	< 0.05
n-Undecane	< 0.5
Naphthalene	2
o-Ethyltoluene	0.02
o-Xylene	0.03
p-Diethylbenzene	< 0.04
p-Ethyltoluene	< 0.07
Styrene	< 0.04
Tetrachloroethylene	< 0.04
Tetrahydrofuran	< 0.4
Toluene	0.13
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.04
trans-2-Butene	0.03
trans-2-Pentene	0.02
Trichloroethylene	< 0.04
Vinyl acetate	< 0.4
Vinyl chloride	< 0.02

Sample ID: 17030213-001

Customer ID: LICA

Cust Samp ID: LICAVOC/CLS/Mar 14, 2017

Maxxam Analytics

VOC Sample Collection Data Sheet Alberta Air FCD AIR FCD-01320 / 2

Client: LICA Sampler S/N: 6167
 Location: Cold Lake South Canister ID: H 2823
 Station ID: LICA 01 Installation Date/Time (mst): Mar 09, 2017 @ 10:39
 Sample ID: LICA/VOC/CLS/Mar 14, 2017 Removal Date/Time (mst): Mar 16, 2017 @ 10:55

Date and Time Information

Sample Date:	Start Time (mst)	End Time (mst)	Elapsed Time (hours)
<u>Mar 14, 2017</u>	<u>00:00</u>	<u>00:00</u> <u>Mar 15, 2017</u>	<u>24.0</u>

Canister Pressure/Vacuum	
Initial Vacuum (in. Hg)	Final Pressure (psi)
<u>-27.7</u>	<u>24</u>

Flow Settings		
Flow Reading (sccm)	Pot Set Point	Pump Set (psi)
<u>10.0</u>	<u>6.52</u>	<u>24.0</u>

Deployment/Collection and Maintenance Checklist

Initial leak check deployment vacuum (in. Hg) = - @ - mst
 Final leak check deployment vacuum (in. Hg) = - @ - mst
 Total leak rate = - psi over - minutes
 Timer reset to zero prior to sampling? YES (yes/no)
 Date of last flow calibration: Mar 06, 2017 (due every 3 months)
 Last date of sample line & fitting replacement: Nov 09, 2016 (due every 6 months)

Leak rate must be 0.0 psi over a minimum of 5 minutes or repair is required

Comments: _____

The leak check completed on Mar 07, 2017

Deployment Technician Signature: Alex Yakupov

Collection Technician Signature: [Signature] Date: Mar 16, 2017



Volatile Organics Data Results

Date: March 14, 2017
Canister ID: H2823

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.02
1,1,2,2-Tetrachloroethane	< 0.02
1,1,2-Trichloroethane	< 0.02
1,1-Dichloroethane	< 0.02
1,1-Dichloroethylene	< 0.04
1,2,3-Trimethylbenzene	< 0.05
1,2,4-Trichlorobenzene	< 0.8
1,2,4-Trimethylbenzene	< 0.05
1,2-Dibromoethane	< 0.02
1,2-Dichlorobenzene	< 0.03
1,2-Dichloroethane	0.02
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	< 0.02
1,3-Butadiene	< 0.02
1,3-Dichlorobenzene	< 0.3
1,4-Dichlorobenzene	< 0.4
1,4-Dioxane	< 0.4
1-Butene	0.09
1-Hexene	< 0.02
1-Pentene	0.01
2,2,4-Trimethylpentane	0.03
2,2-Dimethylbutane	0.02
2,3,4-Trimethylpentane	< 0.01
2,3-Dimethylbutane	0.04
2,3-Dimethylpentane	0.04
2,4-Dimethylpentane	0.02
2-Methylheptane	0.02
2-Methylhexane	0.04
2-Methylpentane	0.15
3-Methylheptane	< 0.02
3-Methylhexane	0.05
3-Methylpentane	0.09
Acetone	2.7
Acrolein	< 0.3
Benzene	0.16
Benzyl chloride	< 0.4
Bromodichloromethane	< 0.02
Bromoform	< 0.02
Bromomethane	< 0.01
Carbon disulfide	< 0.01
Carbon tetrachloride	0.09
Chlorobenzene	< 0.02
Chloroethane	< 0.02
Chloroform	0.02
Chloromethane	0.67
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.04
cis-2-Butene	< 0.02
cis-2-Pentene	< 0.02
Cyclohexane	0.05
Cyclopentane	0.04
Dibromochloromethane	< 0.01
Ethanol	1.1
Ethyl acetate	< 0.4
Ethylbenzene	0.01
Freon-11	0.28
Freon-113	0.09

Volatile Organics Data Results

Date: March 14, 2017
Canister ID: H2823

PARAMETERS	CONCENTRATION (PPB)
Freon-114	0.02
Freon-12	0.7
Hexachloro-1,3-butadiene	< 0.50
Isobutane	1.54
Isopentane	0.75
Isoprene	< 0.01
Isopropyl alcohol	< 0.4
Isopropylbenzene	< 0.01
m,p-Xylene	0.03
m-Diethylbenzene	< 0.04
m-Ethyltoluene	< 0.08
Methyl butyl ketone	< 0.50
Methyl ethyl ketone	0.6
Methyl isobutyl ketone	< 0.4
Methyl methacrylate	< 0.07
Methyl tert butyl ether	< 0.03
Methylcyclohexane	0.09
Methylcyclopentane	0.1
Methylene chloride	< 0.3
n-Butane	2.75
n-Decane	< 0.06
n-Dodecane	< 0.4
n-Heptane	0.06
n-Hexane	0.17
n-Nonane	0.01
n-Octane	< 0.02
n-Pentane	0.7
n-Propylbenzene	< 0.05
n-Undecane	< 0.5
Naphthalene	< 0.5
o-Ethyltoluene	< 0.01
o-Xylene	0.01
p-Diethylbenzene	< 0.04
p-Ethyltoluene	< 0.07
Styrene	< 0.04
Tetrachloroethylene	< 0.04
Tetrahydrofuran	< 0.4
Toluene	0.08
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.04
trans-2-Butene	0.02
trans-2-Pentene	< 0.02
Trichloroethylene	< 0.04
Vinyl acetate	< 0.4
Vinyl chloride	< 0.02

Maxxam Analytics

VOC Sample Collection Data Sheet Alberta Air FCD AIR FCD-01320 / 2

MAR 16, 2017 CW

Client: LA LICA

Sampler S/N: 6167

Location: COLD LAKE SOUTH

Canister ID: S5627

Station ID: LICA 01

Installation Date/Time (mst): MAR 16, 2017 @ 11:03

Sample ID: LICA/VOC/CLS/MAR 20, 2017

Removal Date/Time (mst): MAR 24, 2017 @ 11:32

Date and Time Information

Sample Date:	Start Time (mst)	End Time (mst)	Elapsed Time (hours)
MAR 20, 2017	00:00	24:00	24:00

Canister Pressure/Vacuum

Initial Vacuum (in. Hg)	Final Pressure (psi)
-28	23.5

Flow Settings

Flow Reading (sccm)	Pot Set Point	Pump Set (psi)
10.0	6.52	24.0

Deployment/Collection and Maintenance Checklist

Initial leak check deployment vacuum (in. Hg) = - @ - mst
 Final leak check deployment vacuum (in. Hg) = - @ - mst
 Total leak rate = - psi over - minutes
 Timer reset to zero prior to sampling? Yes (yes/no)
 Date of last flow calibration: MAR 06, 2017 (due every 3 months)
 Last date of sample line & fitting replacement: NOV 09, 2016 (due every 6 months)

****Leak rate must be 0.0 psi over a minimum of 5 minutes or repair is required****

Comments: Leak check completed MAR 07, 2017

Deployment Technician Signature: [Signature] @ MAR 16, 2017.

Collection Technician Signature: [Signature]

Sample ID: 17030304-001

Customer ID: LICA

Cust Samp ID: LICA/VOC/CLS/Mar 20, 2017



Volatile Organics Data Results

Date: March 20, 2017
Canister ID: S5627

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.02
1,1,2,2-Tetrachloroethane	< 0.02
1,1,2-Trichloroethane	< 0.02
1,1-Dichloroethane	< 0.02
1,1-Dichloroethylene	< 0.04
1,2,3-Trimethylbenzene	< 0.05
1,2,4-Trichlorobenzene	< 0.8
1,2,4-Trimethylbenzene	< 0.05
1,2-Dibromoethane	< 0.02
1,2-Dichlorobenzene	< 0.03
1,2-Dichloroethane	0.02
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	< 0.02
1,3-Butadiene	< 0.02
1,3-Dichlorobenzene	< 0.3
1,4-Dichlorobenzene	< 0.4
1,4-Dioxane	< 0.4
1-Butene	0.05
1-Hexene	< 0.02
1-Pentene	< 0.01
2,2,4-Trimethylpentane	< 0.01
2,2-Dimethylbutane	< 0.01
2,3,4-Trimethylpentane	< 0.01
2,3-Dimethylbutane	< 0.02
2,3-Dimethylpentane	< 0.02
2,4-Dimethylpentane	< 0.01
2-Methylheptane	< 0.01
2-Methylhexane	0.01
2-Methylpentane	0.03
3-Methylheptane	< 0.02
3-Methylhexane	0.02
3-Methylpentane	< 0.01
Acetone	2.7
Acrolein	< 0.3
Benzene	0.17
Benzyl chloride	< 0.4
Bromodichloromethane	< 0.02
Bromoform	< 0.02
Bromomethane	0.02
Carbon disulfide	0.11
Carbon tetrachloride	0.1
Chlorobenzene	< 0.02
Chloroethane	< 0.02
Chloroform	0.02
Chloromethane	0.87
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.04
cis-2-Butene	< 0.02
cis-2-Pentene	< 0.02
Cyclohexane	0.03
Cyclopentane	0.01
Dibromochloromethane	< 0.01
Ethanol	< 0.3
Ethyl acetate	< 0.4
Ethylbenzene	< 0.01
Freon-11	0.28
Freon-113	0.11

Volatile Organics Data Results

Date: March 20, 2017
Canister ID: S5627

PARAMETERS	CONCENTRATION (PPB)
Freon-114	< 0.02
Freon-12	0.44
Hexachloro-1,3-butadiene	< 0.50
Isobutane	0.24
Isopentane	0.28
Isoprene	< 0.01
Isopropyl alcohol	< 0.4
Isopropylbenzene	< 0.01
m,p-Xylene	< 0.03
m-Diethylbenzene	< 0.04
m-Ethyltoluene	< 0.08
Methyl butyl ketone	< 0.50
Methyl ethyl ketone	< 0.3
Methyl isobutyl ketone	< 0.4
Methyl methacrylate	< 0.07
Methyl tert butyl ether	< 0.03
Methylcyclohexane	0.02
Methylcyclopentane	< 0.02
Methylene chloride	< 0.3
n-Butane	0.57
n-Decane	< 0.06
n-Dodecane	< 0.4
n-Heptane	0.02
n-Hexane	< 0.01
n-Nonane	< 0.01
n-Octane	< 0.02
n-Pentane	0.1
n-Propylbenzene	< 0.05
n-Undecane	< 0.5
Naphthalene	< 0.5
o-Ethyltoluene	< 0.01
o-Xylene	0.02
p-Diethylbenzene	< 0.04
p-Ethyltoluene	< 0.07
Styrene	< 0.04
Tetrachloroethylene	< 0.04
Tetrahydrofuran	< 0.4
Toluene	0.14
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.04
trans-2-Butene	< 0.01
trans-2-Pentene	< 0.02
Trichloroethylene	< 0.04
Vinyl acetate	< 0.4
Vinyl chloride	< 0.02

Maxxam Analytics

VOC Sample Collection Data Sheet Alberta Air FCD AIR FCD-01320 / 2

Client: LICA Sampler S/N: 0167
 Location: COLD LAKE SOUTH Canister ID: H3296
 Station ID: LICA 01 Installation Date/Time (mst): MARCH 24, 2017 @ 11:32
 Sample ID: LICA/VOC/CLS/MARCH 26, 2017 Removal Date/Time (mst): Mar 27, 2017 @ 13:14

Date and Time Information

Sample Date:	Start Time (mst)	End Time (mst)	Elapsed Time (hours)
<u>MARCH 26, 2017</u>	<u>00:00</u>	<u>24:00</u>	<u>24:00</u>

Canister Pressure/Vacuum	
Initial Vacuum (in. Hg)	Final Pressure (psi)
<u>-29</u>	<u>24.0</u>

Flow Settings		
Flow Reading (sccm)	Pot Set Point	Pump Set (psi)
<u>10</u>	<u>6.52</u>	<u>24</u>

Deployment/Collection and Maintenance Checklist

Initial leak check deployment vacuum (in. Hg) = — @ — mst
 Final leak check deployment vacuum (in. Hg) = — @ — mst
 Total leak rate = — psi over — minutes
 Timer reset to zero prior to sampling? YES (yes/no)
 Date of last flow calibration: MARCH 06, 2017 (due every 3 months)
 Last date of sample line & fitting replacement: NOV. 09, 2016 (due every 6 months)

****Leak rate must be 0.0 psi over a minimum of 5 minutes or repair is required****

Comments: n/a

Deployment Technician Signature: [Signature]
 Collection Technician Signature: Alex Yakupov Date: Mar 27, 2017

Sample ID: 17030304-003
 Customer ID: LICA
 Cust Samp ID: LICA/VOC/CLS/Mar 26, 2017



Volatile Organics Data Results

Date: March 26, 2017
Canister ID: H3296

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.02
1,1,2,2-Tetrachloroethane	< 0.02
1,1,2-Trichloroethane	< 0.02
1,1-Dichloroethane	< 0.02
1,1-Dichloroethylene	< 0.04
1,2,3-Trimethylbenzene	< 0.05
1,2,4-Trichlorobenzene	< 0.8
1,2,4-Trimethylbenzene	< 0.05
1,2-Dibromoethane	< 0.02
1,2-Dichlorobenzene	< 0.03
1,2-Dichloroethane	0.02
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	< 0.02
1,3-Butadiene	0.02
1,3-Dichlorobenzene	< 0.3
1,4-Dichlorobenzene	< 0.4
1,4-Dioxane	< 0.4
1-Butene	0.08
1-Hexene	< 0.02
1-Pentene	< 0.01
2,2,4-Trimethylpentane	0.04
2,2-Dimethylbutane	< 0.01
2,3,4-Trimethylpentane	< 0.01
2,3-Dimethylbutane	< 0.02
2,3-Dimethylpentane	0.04
2,4-Dimethylpentane	0.02
2-Methylheptane	0.01
2-Methylhexane	0.03
2-Methylpentane	0.07
3-Methylheptane	< 0.02
3-Methylhexane	0.04
3-Methylpentane	0.03
Acetone	3.5
Acrolein	< 0.3
Benzene	0.16
Benzyl chloride	< 0.4
Bromodichloromethane	< 0.02
Bromoform	< 0.02
Bromomethane	0.03
Carbon disulfide	0.12
Carbon tetrachloride	0.09
Chlorobenzene	< 0.02
Chloroethane	0.03
Chloroform	0.02
Chloromethane	0.82
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.04
cis-2-Butene	0.03
cis-2-Pentene	< 0.02
Cyclohexane	0.03
Cyclopentane	0.01
Dibromochloromethane	< 0.01
Ethanol	1.3
Ethyl acetate	< 0.4
Ethylbenzene	< 0.01
Freon-11	0.23
Freon-113	0.1

Volatile Organics Data Results

Date: March 26, 2017
Canister ID: H3296

PARAMETERS	CONCENTRATION (PPB)
Freon-114	< 0.02
Freon-12	0.39
Hexachloro-1,3-butadiene	< 0.50
Isobutane	0.53
Isopentane	0.5
Isoprene	< 0.01
Isopropyl alcohol	< 0.4
Isopropylbenzene	< 0.01
m,p-Xylene	0.06
m-Diethylbenzene	< 0.04
m-Ethyltoluene	< 0.08
Methyl butyl ketone	< 0.50
Methyl ethyl ketone	0.5
Methyl isobutyl ketone	< 0.4
Methyl methacrylate	< 0.07
Methyl tert butyl ether	< 0.03
Methylcyclohexane	0.04
Methylcyclopentane	0.03
Methylene chloride	< 0.3
n-Butane	1.15
n-Decane	< 0.06
n-Dodecane	< 0.4
n-Heptane	0.04
n-Hexane	< 0.01
n-Nonane	0.01
n-Octane	< 0.02
n-Pentane	0.2
n-Propylbenzene	< 0.05
n-Undecane	< 0.5
Naphthalene	< 0.5
o-Ethyltoluene	< 0.01
o-Xylene	0.05
p-Diethylbenzene	< 0.04
p-Ethyltoluene	< 0.07
Styrene	< 0.04
Tetrachloroethylene	< 0.04
Tetrahydrofuran	< 0.4
Toluene	0.17
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.04
trans-2-Butene	0.04
trans-2-Pentene	< 0.02
Trichloroethylene	< 0.04
Vinyl acetate	< 0.4
Vinyl chloride	< 0.02

PAH RESULTS

Sample ID: 17030076-002

Customer ID: LICA

Cust Samp ID: LICA/PUF/CLS/Mar 2, 2017

Priority: Normal

PUF PLUS Sample Collection Data Sheet

Client:	<u>LICA</u>	Puf+ S/N:	<u>TE-03</u>
Location:	<u>Cold Lake South</u>	Motor S/N:	<u>1138/100-1020</u>
Station ID:	<u>LICA 01</u>	Installation Date/Time:	<u>Feb 27, 2017/10:58</u>
Field Sample ID:	<u>LICA/PUF/CLS/Mar 02, 2017</u>	Removal Date/Time:	<u>Mar 07, 2017/09:12</u>


Sample Data Collection Information

Sample Date:	<u>Mar 02, 2017</u>	Average Pressure (mmHg)	<u>719</u>
Start Time (mst):	<u>00:00</u>	Average Flow (Q _{std})	<u>229</u>
End Time (mst):	<u>00:00 Feb Mar 03</u>	Average Temperature (°C)	<u>-12.4°</u>
Elapsed Time (Hours):	<u>24.0</u>	Volume (Vstd m ³)	<u>330.20</u>

Sample Recovery Checklist

(circle one)

Flow Rate 230 slpm +/- 0.2 slpm ?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Average temperature appears correct?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Average pressure appears correct?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Any error messages? (if yes list below)	<input type="radio"/> YES	<input checked="" type="radio"/> NO
Sample duration 24 hours?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Date of last calibration/audit:	<u>Dec 28, 2016</u>	
Other observations?	<u>n/a</u>	



Deployed By: Alex Yakupov

Collected By: Alex Yakupov Date: Mar 07, 2017

Polycyclic Aromatic Hydrocarbons (PAHs) Data Results

Date: March 2, 2017
PUF S/N: TE03

PARAMETERS	CONCENTRATION (µg/puf)
1-Methylnaphthalene	0.18
2-Methylnaphthalene	0.28
3-Methylcholanthrene	< 0.01
7,12-Dimethylbenz(a)anthracene	< 0.01
Acenaphthene	0.14
Acenaphthylene	0.03
Acridine	< 0.01
Anthracene	< 0.01
Benzo(a)anthracene	< 0.01
Benzo(a)pyrene	< 0.01
Benzo(b,j,k)fluoranthene	0.03
Benzo(c)phenanthrene	< 0.01
Benzo(e)pyrene	0.01
Benzo(ghi)perylene	< 0.01
Chrysene	0.01
Dibenzo(a,h)pyrene	< 0.01
Dibenzo(a,i)pyrene	< 0.01
Dibenzo(a,l)pyrene	< 0.01
Dibenzo(ah)anthracene	< 0.01
Fluoranthene	0.06
Fluorene	0.09
Indeno(1,2,3-cd)pyrene	< 0.01
Naphthalene	0.33
Perylene	< 0.01
Phenanthrene	0.24
Pyrene	0.04
Retene	0.02

Sample ID: 17030114-002

Customer ID: LICA

Cust Samp ID: LICA/PUF/CLS/Mar 8, 2017

Priority: Normal

TISCH PUF PLUS Sample Collection Data Sheet

Client:	<u>LICA</u>	Puf+ S/N:	<u>TE-06</u>
Location:	<u>Cold Lake South</u>	Motor S/N:	<u>1138/100-1020</u>
Station ID:	<u>LICA 01</u>	Installation Date/Time:	<u>Mar 07, 2017/09:12</u>
Field Sample ID:	<u>LICA/PUF/CLS/Mar 08, 2017</u>	Removal Date/Time:	<u>Mar 09, 2017/10:49</u>

Sample Data Collection Information

Sample Date:	<u>Mar 08, 2017</u>	Average Pressure (mmHg)	<u>724</u>
Start Time (mst):	<u>00:00</u>	Average Flow (Q _{std})	<u>229</u>
End Time (mst):	<u>00:00 Mar 09, 2017</u>	Average Temperature (°C)	<u>-20.3°</u>
Elapsed Time (Hours):	<u>24.0</u>	Volume (Vstd m ³)	<u>330.16</u>

Sample Recovery Checklist

(circle one)

Flow Rate 230 slpm +/- 0.2 slpm ?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Average temperature appears correct?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Average pressure appears correct?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Any error messages? (if yes list below)	YES	<input checked="" type="radio"/> NO
Sample duration 24 hours?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Date of last calibration/audit:	<u>Dec 28, 2016</u>	
Other observations?	<u>n/a</u>	

Deployed By: Alex Yakupov

Collected By: Alex Yakupov Date: Mar 09, 2017



Polycyclic Aromatic Hydrocarbons (PAHs) Data Results

Date: March 8, 2017
PUF S/N: TE-06

PARAMETERS	CONCENTRATION (µg/puf)
1-Methylnaphthalene	0.4
2-Methylnaphthalene	0.62
3-Methylcholanthrene	< 0.01
7,12-Dimethylbenz(a)anthracene	< 0.01
Acenaphthene	0.08
Acenaphthylene	0.02
Acridine	< 0.01
Anthracene	0.01
Benzo(a)anthracene	0.01
Benzo(a)pyrene	0.01
Benzo(b,j,k)fluoranthene	0.05
Benzo(c)phenanthrene	< 0.01
Benzo(e)pyrene	0.02
Benzo(ghi)perylene	0.02
Chrysene	< 0.01
Dibenzo(a,h)pyrene	< 0.01
Dibenzo(a,i)pyrene	< 0.01
Dibenzo(a,l)pyrene	< 0.01
Dibenzo(ah)anthracene	< 0.01
Fluoranthene	0.1
Fluorene	0.07
Indeno(1,2,3-cd)pyrene	0.03
Naphthalene	1.12
Perylene	< 0.01
Phenanthrene	0.27
Pyrene	0.03
Retene	0.07

Sample ID: 17030213-002

Customer ID: LICA

Cust Samp ID: LICA/PUF/CLS/Mar 14, 2017

TISCH PUF PLUS Sample Collection Data Sheet			
Client:	<u>LICA</u>	Puf+ S/N:	<u>TE-04</u>
Location:	<u>Cold Lake South</u>	Motor S/N:	<u>1138/100-1020</u>
Station ID:	<u>LICA 01</u>	Installation Date/Time:	<u>Mar 09, 2017/10:49</u>
Field Sample ID:	<u>LICA/PUF/CLS/ Mar 14, 2017</u>	Removal Date/Time:	<u>Mar 16, 2017 @ 10:50</u>

Sample Data Collection Information			
Sample Date:	<u>Mar 14, 2017</u>	Average Pressure (mmHg)	<u>712</u>
Start Time (mst):	<u>00:00</u>	Average Flow (Q _{std})	<u>229</u>
End Time (mst):	<u>00:00 Mar 15, 2017</u>	Average Temperature (°C)	<u>-3.5</u>
Elapsed Time (Hours):	<u>24.0</u>	Volume (V _{std} m ³)	<u>330.17</u>

Sample Recovery Checklist		
(circle one)		
Flow Rate 230 slpm +/- 0.2 slpm ?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Average temperature appears correct?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Average pressure appears correct?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Any error messages? (if yes list below)	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Sample duration 24 hours?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Date of last calibration/audit:	<u>Dec 28, 2016</u>	
Other observations?		

Deployed By:	<u>Alex Yakupov</u>
Collected By:	<u>[Signature]</u> Date: <u>Mar 16, 2017</u>



Polycyclic Aromatic Hydrocarbons (PAHs) Data Results

Date: March 14, 2017
PUF S/N: TE-04

PARAMETERS	CONCENTRATION (µg/puf)
1-Methylnaphthalene	0.12
2-Methylnaphthalene	0.15
3-Methylcholanthrene	< 0.01
7,12-Dimethylbenz(a)anthracene	< 0.01
Acenaphthene	0.05
Acenaphthylene	0.01
Acridine	< 0.01
Anthracene	0.02
Benzo(a)anthracene	0.01
Benzo(a)pyrene	< 0.01
Benzo(b,j,k)fluoranthene	0.03
Benzo(c)phenanthrene	< 0.01
Benzo(e)pyrene	0.01
Benzo(ghi)perylene	0.02
Chrysene	< 0.01
Dibenzo(a,h)pyrene	< 0.01
Dibenzo(a,i)pyrene	< 0.01
Dibenzo(a,l)pyrene	< 0.01
Dibenzo(ah)anthracene	< 0.01
Fluoranthene	0.06
Fluorene	0.08
Indeno(1,2,3-cd)pyrene	0.02
Naphthalene	0.15
Perylene	< 0.01
Phenanthrene	0.25
Pyrene	0.05
Retene	0.03

Sample ID: 17030304-002

Customer ID: LICA

Cust Samp ID: LICA/PUF/CLS/Mar 20, 2017

I PUF PLUS Sample Collection Data Sheet

Client:	<u>LICA</u>	Puff S/N:	<u>A13-02</u>
Location:	<u>COLD LAKE SOUTH</u>	Motor S/N:	<u>1138/100-1020</u>
Station ID:	<u>LICA 01</u>	Installation Date/Time:	<u>Mar 16, 2017 @ 10:52</u>
Field Sample ID:	<u>LICA/PUF/CLS/Mar 20, 2017</u>	Removal Date/Time:	<u>Mar 24, 2017 @ 10:54</u>

Sample Data Collection Information

Sample Date:	<u>Mar 20, 2017</u>	Average Pressure (mmHg)	<u>722</u>
Start Time (mst):	<u>00:00</u>	Average Flow (Q _{std})	<u>229</u>
End Time (mst):	<u>24:00</u>	Average Temperature (°C)	<u>-7.7</u>
Elapsed Time (Hours):	<u>24:00</u>	Volume (V _{std} m ³)	<u>330.19</u>

Sample Recovery Checklist

(circle one)

Flow Rate 230 slpm +/- 0.2 slpm ?	YES	NO
Average temperature appears correct?	YES	NO
Average pressure appears correct?	YES	NO
Any error messages? (if yes list below)	YES	NO
Sample duration 24 hours?	YES	NO

Date of last calibration/audit: Dec 28, 2016

Other observations:



Deployed By: [Signature]

Collected By: [Signature]

Polycyclic Aromatic Hydrocarbons (PAHs) Data Results

Date: March 20, 2017
PUF S/N: A13-02

PARAMETERS	CONCENTRATION (µg/puf)
1-Methylnaphthalene	0.08
2-Methylnaphthalene	0.13
3-Methylcholanthrene	< 0.01
7,12-Dimethylbenz(a)anthracene	< 0.01
Acenaphthene	< 0.01
Acenaphthylene	< 0.01
Acridine	< 0.01
Anthracene	< 0.01
Benzo(a)anthracene	< 0.01
Benzo(a)pyrene	< 0.01
Benzo(b,j,k)fluoranthene	< 0.01
Benzo(c)phenanthrene	< 0.01
Benzo(e)pyrene	< 0.01
Benzo(ghi)perylene	< 0.01
Chrysene	< 0.01
Dibenzo(a,h)pyrene	< 0.01
Dibenzo(a,i)pyrene	< 0.01
Dibenzo(a,l)pyrene	< 0.01
Dibenzo(ah)anthracene	< 0.01
Fluoranthene	< 0.01
Fluorene	0.02
Indeno(1,2,3-cd)pyrene	< 0.01
Naphthalene	0.14
Perylene	< 0.01
Phenanthrene	0.03
Pyrene	0.01
Retene	< 0.01

Sample ID: 17030304-004

Customer ID: LICA
Cust Samp ID: LICA/PUF/CLS/Mar 26, 2017

TISCH PUF PLUS Sample Collection Data Sheet

Client:	<u>LICA</u>	Puf+ S/N:	<u>7E 05</u>
Location:	<u>COLD LAKE SOUTH</u>	Motor S/N:	<u>1138 / 100-1020</u>
Station ID:	<u>LICA 01</u>	Installation Date/Time:	<u>MARCH 24, 2017 @ 10:57</u>
Field Sample ID:	<u>LICA/PUF/CLS/Mar 26, 2017</u>	Removal Date/Time:	<u>Mar 27, 2017/14:09</u>

Sample Data Collection Information

Sample Date:	<u>MARCH 26, 2017</u>	Average Pressure (mmHg)	<u>711</u>
Start Time (mst):	<u>00:00</u>	Average Flow (Q _{std})	<u>229</u>
End Time (mst):	<u>24:00</u>	Average Temperature (°C)	<u>+ 3.2°</u>
Elapsed Time (Hours):	<u>24:00</u>	Volume (V _{std} m ³)	<u>330.20</u>

Sample Recovery Checklist

(circle one)

Flow Rate 230 slpm +/- 0.2 slpm ?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Average temperature appears correct?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Average pressure appears correct?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Any error messages? (if yes list below)	<input type="radio"/> YES	<input checked="" type="radio"/> NO
Sample duration 24 hours?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Date of last calibration/audit:	<u>DEC. 28, 2016</u>	
Other observations?	<u>n/a</u>	



Deployed By: [Signature]

Collected By: Alex Yakupov

Polycyclic Aromatic Hydrocarbons (PAHs) Data Results

Date: March 26, 2017
PUF S/N: TE05

PARAMETERS	CONCENTRATION (µg/puf)
1-Methylnaphthalene	0.3
2-Methylnaphthalene	0.57
3-Methylcholanthrene	< 0.01
7,12-Dimethylbenz(a)anthracene	< 0.01
Acenaphthene	0.04
Acenaphthylene	0.07
Acridine	< 0.01
Anthracene	0.13
Benzo(a)anthracene	0.02
Benzo(a)pyrene	0.02
Benzo(b,j,k)fluoranthene	0.07
Benzo(c)phenanthrene	< 0.01
Benzo(e)pyrene	0.02
Benzo(ghi)perylene	0.02
Chrysene	0.05
Dibenzo(a,h)pyrene	< 0.01
Dibenzo(a,i)pyrene	< 0.01
Dibenzo(a,l)pyrene	< 0.01
Dibenzo(ah)anthracene	< 0.01
Fluoranthene	0.28
Fluorene	0.25
Indeno(1,2,3-cd)pyrene	0.03
Naphthalene	0.35
Perylene	< 0.01
Phenanthrene	0.84
Pyrene	0.23
Retene	0.12

PARTISOL RESULTS

Sample ID: 17030081-001


AIR FCD-01318/2

Customer ID: LICA
Cust Samp ID: P6072294

Partisol Sample Data Sheet

Priority: Normal

Date Sampled: Mar 02, 2017
Location: Cold Lake South
Parameter: TSP PM10
Filter #: P607 22 94



Start Time 00:00 Mar 02, 2017
End Time 00:00 Mar 03, 2017
Status OK
Std Vol 26.097
Valid Time 24:00
Total Time 24.0

Comments: Weather Conditions, etc.

n/a

Sample inlet cleaned on Jan 27, 2017
Date of last audit - Jan 27, 2017

Technician Signature: Alex Yakupov
Date: Mar 07, 2017
Time: 09:30

Programming

- 1) Make sure system is in "Stop Mode"
- 2) "ESC" to Time Screen then "Program"
- 3) Enter Beg 1 0:00
- 4) Enter Dur 24:00:00
- 5) Enter Beg D dd-Aug
- 6) Enter End D dd-Aug
- 7) "Stop/Run"

Note: Beginning & End Date should be same date

Sample ID: 17030112-001

Customer ID: LICA

AIR FCD-01318/2

Cust Samp ID: Flt # P6072293

Partisol Sample Data Sheet

Priority: Normal



Date Sampled: Mar 08, 2017

Location: Cold Lake South

Parameter: TSP PM10

PM2.5

Filter #: P 607 22 93

Start Time 00:00 Mar 08, 2017

End Time 00:00 Mar 09, 2017

Status OK

Std Vol 27.057

Valid Time 24:00

Total Time 24.0

Comments: Weather Conditions, etc.

Sample inlet cleaned on Jan 27, 2017
Date of last audit - Jan 27, 2017

Technician Signature:

Alex Yakupov

Date: Mar 09, 2017

Time: 11:15


Programming

- 1) Make sure system is in "Stop Mode"
- 2) "ESC" to Time Screen then "Program"
- 3) Enter Beg 1 0:00
- 4) Enter Dur 24:00:00
- 5) Enter Beg D dd-Aug
- 6) Enter End D dd-Aug
- 7) "Stop/Run"

Note: Beginning & End Date should be same date

Sample ID: 17030214-001

Customer ID: LICA
Cust Samp ID: P6072292



Partisol Sample Data Sheet

Priority: Normal

Date Sampled: Mar 14, 2017

Location: Cold Lake South

Parameter: TSP PM10

PM2.5

Filter #: P607 22 92

Start Time 00:00 Mar 14, 2017

End Time 00:00 Mar 15, 2017

Status OK

Std Vol 25.030

Valid Time 24:00

Total Time 24.0

Comments: Weather Conditions, etc.

-4.9 °C ; 0.932 ATM

Sample inlet cleaned on Jan 27, 2017

Date of last audit - Jan 27, 2017

Technician Signature: Deployed: Alex Yakupov

Date:

Time:

collected: [Signature] Mar 16, 2017.

Programming

- 1) Make sure system is in "Stop Mode"
- 2) "ESC" to Time Screen then "Program"
- 3) Enter Beg 1 0:00
- 4) Enter Dur 24:00:00
- 5) Enter Beg D dd-Aug
- 6) Enter End D dd-Aug
- 7) "Stop/Run"

Note: Beginning & End Date should be same date

Sample ID: 17030306-001

Customer ID: LICA
Cust Samp ID: LICA Flt # P6072291

AIR FCD-01318/2

Partisol Sample Data Sheet

Priority: Normal



Date Sampled: 20-Mar-2017
Location: COLD LAKE SOUTH
Parameter: ISP PM10
Filter #: P6072291

PM2.5

Start Time 00:00
End Time 24:00
Status OK
Std Vol 25.763
Valid Time 24:00
Total Time 24:00

Comments: Weather Conditions, etc.

TEMP - 8.7 | .981 ATM

Technician Signature:

Programming

- 1) Make sure system is in "Stop Mode"
- 2) "ESC" to Time Screen then "Program"
- 3) Enter Beg 1 0:00
- 4) Enter Dur 24:00:00
- 5) Enter Beg D dd-Aug
- 6) Enter End D dd-Aug
- 7) "Stop/Run"

Note: Beginning & End Date should be same date

Sample ID: 17030306-002

Customer ID: LICA

Cust Samp ID: LICA Fit # P6193008

AIR FCD-01318/2

Partisol Sample Data Sheet

Priority: Normal



Date Sampled: MARCH 26, 2017

Location: COLD LAKE SOUTH

Parameter: ~~TSP~~ PM10

PM2.5

Filter #: P6193008

Start Time 00:00

End Time 24:00

Status OK

Std Vol 24.419

Valid Time 24:00

Total Time 24.0

Comments: Weather Conditions, etc.

n/a

Installed: Mike Espiritu

Technician Signature: Alex Yakupov (collected)

Time: 13:41

Date: Mar 27, 2017

Programming

- 1) Make sure system is in "Stop Mode"
- 2) "ESC" to Time Screen then "Program"
- 3) Enter Beg 1 0:00
- 4) Enter Dur 24:00:00
- 5) Enter Beg D dd-Aug
- 6) Enter End D dd-Aug
- 7) "Stop/Run"

Note: Beginning & End Date should be same date

Partisol Sampler Results

Date	Filter NO.	Concentration (mg)
March 2	P6072294	0.110
March 8	P6072293	0.037
March 14	P6072292	0.123
March 20	P6072291	0.033
March 26	P6193008	0.126

APPENDIX III
EQUIPMENT CALIBRATION RESULTS

SULPHUR DIOXIDE



Thermo 43i Sulphur Dioxide Analyzer Calibration

Date: March 6, 2017 Company/Airshed: LICA Location/Station Name: Cold Lake South Parameter: Sulphur Dioxide Start Time 24 hr. (mst): 10:35 End Time 24 hr. (mst): 12:52 Calibration Method: Gas Dilution	Barometric Pressure: 0.929 atm Station Temperature °C: 22 Weather Conditions: Mainly cloudy with snow Calibration Purpose: shut down Performed By/Reviewer: Alex Yakupov / Trina Whitsitt <input type="checkbox"/> Hide Cal Gas Expiry Date: July 18, 2019 Converter Model & s/n (if applicable): n/a
---	--

Analyzer: ID# or Serial Number: 806528242 Last Calibration Date: February 27, 2017 Previous C.F.: 1.001	Range ppb: 500 As Found C.F.: 0.998 New C.F.: n/a
--	--

Calibrator: Flow Meter ID's: n/a Make & Model: API 700 Serial #: 627 Cal Gas Cylinder I.D. #: LL104222 Cal Gas Conc. (ppm): 50.6	Standard Calibration Points for Ranges <table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>380</td></tr> <tr><td>Mid</td><td>180</td></tr> <tr><td>Low</td><td>90</td></tr> </table>	Point	ppb	High	380	Mid	180	Low	90
Point	ppb								
High	380								
Mid	180								
Low	90								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	(ppb)	(ppb)	
as found zero	5000	0.00	5000	0.0	0.0	n/a
as found high	4965	37.50	5003	379.3	380.0	0.998
mid	4985	17.80	5003	180.0	179.0	1.006
low	4991	8.90	5000	90.1	90.0	1.001
Average C.F. =						1.002

Linear Regression/Calibration Results:

Correlation Coefficient = 1.000 Slope = 0.998 b (Intercept as % of full scale) = 0.08% % change in C.F. from last cal = 0.28%	LIMITS > or = 0.995 0.90-1.10 ± 3% F.S. ± 10%
--	--

Thermo 43i Sulphur Dioxide Analyzer Calibration



As found: BKG: 8.0 COEF: 0.964 PMT: -623.5 FLASH: 774 INTERNAL: 30.1 CHAMBER: 45.2 PERM OVEN GAS: 45.00 PERM OVEN HEATER: 44.20 PRESSURE: 674.7 SAMPLE FLOW: 0.471 LAMP INTENSITY: 96 CONVERTER: n/a CONVERTER SET: n/a Expected Value: 465.0	As left: BKG: n/a COEF: n/a PMT: n/a FLASH: n/a INTERNAL: n/a CHAMBER: n/a PERM OVEN GAS: n/a PERM OVEN HEATER: n/a PRESSURE: n/a SAMPLE FLOW: n/a LAMP INTENSITY: n/a CONVERTER: n/a CONVERTER SET: n/a Expected Value: n/a
--	---

Comments:

Shutdown calibration completed for annual maintenance.



Thermo 43i Sulphur Dioxide Analyzer Calibration

Date:	March 6, 2017	Barometric Pressure:	0.929 atm
Company/Airshed:	LICA	Station Temperature °C:	22
Location/Station Name:	Cold Lake South	Weather Conditions:	Mainly cloudy with snow
Parameter:	Sulphur Dioxide	Calibration Purpose:	post repair
Start Time 24 hr. (mst):	15:20	Performed By/Reviewer:	Alex Yakupov Trina Whatsitt
End Time 24 hr. (mst):	18:36	Cal Gas Expiry Date:	July 18, 2019
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	n/a

Analyzer:	ID# or Serial Number:	806528242	Range ppb:	500
	Last Calibration Date:	n/a	As Found C.F.:	n/a
	Previous C.F.:	n/a	New C.F.:	1.001

Calibrator:	Flow Meter ID's:	n/a	Standard Calibration Points for Ranges
	Make & Model:	API 700	
	Serial #:	627	
	Cal Gas Cylinder I.D. #:	LL104222	
	Cal Gas Conc. (ppm):	50.6	

Point	ppb
High	380
Mid	180
Low	90


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	(ppb)	(ppb)	
adjusted zero	5000	0.00	5000	0.0	0.0	n/a
adjusted high	4966	37.50	5004	379.2	379.0	1.001
mid	4986	17.80	5004	180.0	180.0	1.000
low	4991	8.90	5000	90.1	90.0	1.001
calibrator zero	5000	0.00	5000	0.0	0.0	n/a
Average C.F. =						1.000

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS
Slope =	1.001	> or = 0.995
b (Intercept as % of full scale) =	0.00%	± 3% F.S.
% change in C.F. from last cal =	n/a	± 10%

Thermo 43i Sulphur Dioxide Analyzer Calibration




As found:	As left:
BKG: n/a	BKG: 8.0
COEF: n/a	COEF: 0.957
PMT: n/a	PMT: -623.8
FLASH: n/a	FLASH: 776
INTERNAL: n/a	INTERNAL: 29.8
CHAMBER: n/a	CHAMBER: 45.2
PERM OVEN GAS: n/a	PERM OVEN GAS: 45.00
PERM OVEN HEATER: n/a	PERM OVEN HEATER: 44.20
PRESSURE: n/a	PRESSURE: 675.9
SAMPLE FLOW: n/a	SAMPLE FLOW: 0.471
LAMP INTENSITY: n/a	LAMP INTENSITY: 96
CONVERTER: n/a	CONVERTER: n/a
CONVERTER SET: n/a	CONVERTER SET: n/a
Expected Value: n/a	Expected Value: 461.0

Comments:
 The analyzer sample inlet filter was changed. The analyzer cooling fan filter(s) were cleaned.

Annual maintenance completed. Sample pump rebuilt. Leak check performed. No issues found.

SO2[ppb] Station: LICA COLD LAKE SOUTH Daily: 17.03.06 Type: AVG 1 Min. [1 Min.]



— SO2[ppb]

TOTAL REDUCED SULPHUR



Thermo 450i Total Reduced Sulphur Analyzer Calibration

Date:	March 6, 2017	Barometric Pressure:	0.929 atm
Company/Airshed:	LICA	Station Temperature °C:	22
Location/Station Name:	Cold Lake South	Weather Conditions:	Mainly cloudy with snow
Parameter:	Total Reduced Sulphur	Calibration Purpose:	shut down
Start Time 24 hr. (mst):	10:35	Performed By/Reviewer:	Alex Yakupov Trina Whitsitt
End Time 24 hr. (mst):	12:40	Cal Gas Expiry Date:	June 14, 2019
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	CDNova CDN-101 #501

Analyzer:	
ID# or Serial Number:	812728560
Last Calibration Date:	February 27, 2017
Previous C.F.:	1.000
Range ppb:	100
As Found C.F.:	0.982
New C.F.:	n/a

Calibrator:		Standard Calibration Points for Ranges	
Flow Meter ID's:	n/a	Point	ppb
Make & Model:	SABIO 2010 D	High	78
Serial #:	11900613	Mid	38
Cal Gas Cylinder I.D. #:	EY000654	Low	19
Cal Gas Conc. (ppm):	10.2		


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	(ppb)	(ppb)	
as found zero	7500	0.00	7500	0.0	0.0	n/a
as found high	7443	57.40	7500	78.1	79.5	0.982
mid	7472	27.90	7500	37.9	38.3	0.991
low	7485	14.00	7499	19.0	19.1	0.997
Average C.F. =						0.990

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	0.981		0.90-1.10
b (Intercept as % of full scale) =	0.19%		± 3% F.S.
% change in C.F. from last cal =	1.81%		± 10%

Thermo 450i Total Reduced Sulphur Analyzer Calibration



As found:	As left:
BKG: 13.8	BKG: n/a
COEF: 0.937	COEF: n/a
PMT: -650.5	PMT: n/a
FLASH: 741	FLASH: n/a
INTERNAL: 33.1	INTERNAL: n/a
CHAMBER: 44.8	CHAMBER: n/a
CONVERTER TEMP: 825	CONVERTER TEMP: n/a
CONVERTER SET: 825	CONVERTER SET: n/a
PERM OVEN GAS: 45.00	PERM OVEN GAS: n/a
PERM OVEN HTR: 44.37	PERM OVEN HTR: n/a
PRESSURE: 650.8	PRESSURE: n/a
SAMPLE FLOW: 0.508	SAMPLE FLOW: n/a
LAMP INTENSITY: 92	LAMP INTENSITY: n/a
Expected Value: 37.6	Expected Value: n/a

Comments:

Shutdown calibration completed for annual maintenance.



Thermo 450i Total Reduced Sulphur Analyzer Calibration

Date:	March 6, 2017	Barometric Pressure:	0.929 atm
Company/Airshed:	LICA	Station Temperature °C:	22
Location/Station Name:	Cold Lake South	Weather Conditions:	Mainly cloudy with snow
Parameter:	Total Reduced Sulphur	Calibration Purpose:	post repair
Start Time 24 hr. (mst):	15:07	Performed By/Reviewer:	Alex Yakupov Trina Whitsitt
End Time 24 hr. (mst):	18:36	Cal Gas Expiry Date:	June 14, 2019
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	CDNova CDN-101 #501

Analyzer:	
ID# or Serial Number:	812728560
Last Calibration Date:	n/a
Previous C.F.:	n/a
Range ppb:	100
As Found C.F.:	n/a
New C.F.:	0.999


Calibrator:		Standard Calibration Points for Ranges	
Flow Meter ID's:	n/a	Point	ppb
Make & Model:	SABIO 2010 D	High	78
Serial #:	11900613	Mid	38
Cal Gas Cylinder I.D. #:	EY0000654	Low	19
Cal Gas Conc. (ppm):	10.2		

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	(ppb)	(ppb)	
adjusted zero	7500	0.00	7500	0.0	0.0	n/a
adjusted high	7443	57.40	7500	78.1	78.1	0.999
mid	7472	27.90	7500	37.9	38.0	0.997
low	7485	14.00	7499	19.0	19.2	0.992
calibrator zero	7500	0.00	7500	0.0	0.0	n/a
Average C.F. =						0.996

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	1.000		.95-1.05
b (Intercept as % of full scale) =	-0.07%		± 3% F.S.
% change in C.F. from last cal =	n/a		± 10%




As found:		As left:	
BKG:	n/a	BKG:	14.1
COEF:	n/a	COEF:	0.932
PMT:	n/a	PMT:	-650.5
FLASH:	n/a	FLASH:	741
INTERNAL:	n/a	INTERNAL:	32.4
CHAMBER:	n/a	CHAMBER:	44.9
CONVERTER TEMP:	n/a	CONVERTER TEMP:	326.2
CONVERTER SET:	n/a	CONVERTER SET:	325.0
PERM OVEN GAS:	n/a	PERM OVEN GAS:	45.00
PERM OVEN HTR:	n/a	PERM OVEN HTR:	44.38
PRESSURE:	n/a	PRESSURE:	627.6
SAMPLE FLOW:	n/a	SAMPLE FLOW:	0.484
LAMP INTENSITY:	n/a	LAMP INTENSITY:	91
Expected Value:	n/a	Expected Value:	40.2

Comments:
 The analyzer sample inlet filter was changed. The analyzer cooling fan filter(s) were cleaned.

Annual maintenance completed. Leaks check performed. Converter tubing renewed and smaller diameter tubing installed. Sample pump rebuilt. Converter chamber cleaned.

TRS[ppb] Station: LICA COLD LAKE SOUTH Daily: 17.03.06 Type: AVG 1 Min. [1 Min.]



— TRS[ppb]

TOTAL HYDROCARBON



Thermo 51C Total Hydrocarbon Analyzer Calibration

Date:	March 7, 2017	Barometric Pressure:	0.936 atm
Company/Airshed:	LICA	Station Temperature °C:	22
Location/Station Name:	Cold Lake South	Weather Conditions:	Mainly cloudy with snow
Parameter:	Total Hydrocarbon	Calibration Purpose:	shut down
Start/End Time 24 hr. (mst):	8:29 / 10:24	Performed By/Reviewer:	Alex Yakupov / Trina Whitsitt
Calibration Method:	Gas Dilution	Cal Gas Expiry Date:	November 25, 2023

Analyzer:	ID# or Serial Number:	427408718	Range ppm:	50
	Last Calibration Date:	February 23, 2017	As Found C.F.:	1.000
	Previous Cal High Point C.F.:	1.000	New C.F.:	n/a

Calibrator:	Flow Meter ID's:	n/a	Standard Calibration Points for a Range of:	50 ppm
	Make & Model:	API 700		
	Serial #:	627		
	Cal Gas Cylinder I.D. #:	LL165372		
	CH ₄ /C ₃ H ₈ Cylinder Conc. (ppm):	606.0 212.0		
	CH ₄ as propane/total CH ₄ equivalents (ppm):	583.0 1189.0		

Point	Target ppm
High	38
Mid	18
Low	9


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated Concentration (ppm)	Indicated Concentration (ppm)	Correction Factors:
	Diluent	Cal Gas	Total			
as found zero	2000	0.00	2000	0.0	0.10	n/a
as found high	1935	65.00	2000	38.64	38.76	1.000
mid	1972	31.00	2003	18.40	18.26	1.013
low	1983	16.00	1999	9.52	9.40	1.023
Average C.F. =						1.012

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	0.998		.95-1.05
b (Intercept as % of full scale) =	0.08%		± 3% F.S.
% change in C.F. from last cal =	0.05%		± 10%

Thermo 51C Total Hydrocarbon Analyzer Calibration



As found:	H2 cylinder (psi):	900	As left:	H2 cylinder (psi):	n/a
	H2 cylinder reg set (psi):	22		H2 cylinder reg set (psi):	n/a
	Span Cylinder (psi):	1800		Span Cylinder (psi):	n/a
	Span Cylinder Reg Set (psi):	22		Span Cylinder Reg Set (psi):	n/a
	Zero Air Gen Pressure:	35		Zero Air Gen Pressure:	n/a
	measurement alarms:	None		measurement alarms:	n/a
	service alarms:	None		service alarms:	n/a
	cnt:	1441		cnt:	n/a
	rng:	1		rng:	n/a
	try:	3		try:	n/a
	flm:	180.4		flm:	n/a
	det:	125.6		det:	n/a
	Flame:	180		Flame:	n/a
	Filter:	125		Filter:	n/a
	Base:	125		Base:	n/a
	Sample psi:	06.51		Sample psi:	n/a
	Internal Air Pressure:	20		Internal Air Pressure:	n/a
	Internal Fuel Pressure:	12		Internal Fuel Pressure:	n/a
	Measured Flow:	n/a		Measured Flow:	n/a
	Expected Value:	26.81		Expected Value:	n/a

Comments:
The analyzer sample inlet filter was changed.

The analyzer cooling fan filter(s) were cleaned.

Shutdown calibration completed for annual maintenance.



Thermo 51C Total Hydrocarbon Analyzer Calibration

Date:	March 7, 2017	Barometric Pressure:	0.936 atm
Company/Airshed:	LICA	Station Temperature °C:	22
Location/Station Name:	Cold Lake South	Weather Conditions:	Mainly cloudy with snow
Parameter:	Total Hydrocarbon	Calibration Purpose:	post repair
Start/End Time 24 hr. (mst):	10:38 / 13:11	Performed By/Reviewer:	Alex Yakupov / Trina Whitsitt
Calibration Method:	Gas Dilution	Cal Gas Expiry Date:	November 25, 2023

Analyzer:	ID# or Serial Number:	427408718	Range ppm:	50
	Last Calibration Date:	n/a	As Found C.F.:	n/a
	Previous Cal High Point C.F.:	n/a	New C.F.:	1.000

Calibrator:	Flow Meter ID's:	n/a	Standard Calibration Points for a Range of: 50 ppm	
	Make & Model:	API 700		
	Serial #:	627		
	Cal Gas Cylinder I.D. #:	LL165372		
	CH ₄ /C ₂ H ₆ Cylinder Conc. (ppm):	606.0 / 212.0	Point	Target ppm
	CH ₄ as propane/total CH ₄ equivalents (ppm):	583.0 / 1189.0	High	38
			Mid	18
			Low	9


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors:
Point	Diluent	Cal Gas	Total	(ppm)	(ppm)	
adjusted zero	2000	0.00	2000	0.0	0.00	n/a
adjusted high	1935	65.00	2000	38.64	38.66	1.000
mid	1972	31.00	2003	18.40	18.25	1.008
low	1983	16.00	1999	9.52	9.38	1.015
calibrator zero	2000	0.00	2000	0.00	0.00	n/a
Average C.F. =						1.007

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS
Slope =	0.999	> or = 0.995
b (Intercept as % of full scale) =	0.18%	.95-1.05
% change in C.F. from last cal =	n/a	± 3% F.S.
		± 10%

Thermo 51C Total Hydrocarbon Analyzer Calibration




As found:	As left:		
H2 cylinder (psi):	n/a	H2 cylinder (psi):	900
H2 cylinder reg set (psi):	n/a	H2 cylinder reg set (psi):	22
Span Cylinder (psi):	n/a	Span Cylinder (psi):	1800
Span Cylinder Reg Set (psi):	n/a	Span Cylinder Reg Set (psi):	22
Zero Air Gen Pressure:	n/a	Zero Air Gen Pressure:	35
measurement alarms:	n/a	measurement alarms:	None
service alarms:	n/a	service alarms:	None
cnt:	n/a	cnt:	1461
rng:	n/a	rng:	1
try:	n/a	try:	3
flm:	n/a	flm:	180.8
det:	n/a	det:	125.7
Flame:	n/a	Flame:	180
Filter:	n/a	Filter:	125
Base:	n/a	Base:	125
Sample psi:	n/a	Sample psi:	06.51
Internal Air Pressure:	n/a	Internal Air Pressure:	20
Internal Fuel Pressure:	n/a	Internal Fuel Pressure:	13
Measured Flow:	n/a	Measured Flow:	n/a
Expected Value:	n/a	Expected Value:	27.20

Comments:
The analyzer sample inlet filter was changed.

The analyzer cooling fan filter(s) were cleaned.

The LICA's Zero Air generator s/n 4027 was installed back after repair. Maxxam's Zero Air s/n 133 was left in the station as a temporary back-up zero air generator. H2 fuel tubing was renewed. Fuel pressure adjusted to 13 psi.



— THC[ppm]

NITROGEN DIOXIDE



Thermo 42i NO-NO2-NOx Analyzer Calibration

Date: March 6, 2017
Company/Airshed: LICA
Location/Station Name: Cold Lake South
Start/End Time 24 hr. (mst): 10:35 / 16:03
G.P.T. to be used for Ozone? No
Calibration Method: Gas Dilution & Gas Phase Titration

Barometric Pressure: 0.929 atm
Station Temperature °C: 22
Weather Conditions: Mainly cloudy with snow
Calibration Purpose: routine monthly
Performed By/Reviewer: Alex Yakupov | Trina Whitsitt
Cal Gas Expiry Date: July 18, 2019

Analyzer:

ID# or Serial Number: 1505664393
Last Calibration Date: February 27, 2017
Range ppb: 500

Correction Factors:

	Previous C.F.:	As Found C.F.:	New C.F.:
NO =	1.000	1.000	1.000
NO ₂ =	1.000	1.000	1.000
NOx =	1.000	1.000	1.000

Calibrator:

Flow Meter ID's: n/a
Make & Model: API 700
Serial #: 627
Cal Gas Cylinder I.D. #: LL104222
NO/NOx Gas Conc. (ppm): 50.7 | 50.7

Standard Calibration Points for a Range of: 500 ppb

Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?
High	380	250	n/a
Mid	180	145	n/a
Low	90	50	n/a
Extra Point #1	n/a	n/a	n/a
Extra Point #2	n/a	n/a	n/a

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated NO	Calculated NOx	Indicated NO	Indicated NOx	NO C.F.	NOx C.F.
Point	Diluent	Cal Gas	Total Flow	(ppb)	(ppb)	(ppb)	(ppb)		
as found zero	5000	0.0	5000	0	0	0.0	0.0	n/a	n/a
as found high	4965	37.5	5003	380.1	380.1	380.0	380.0	1.000	1.000
adjusted zero	5000	0.00	5000	0.0	0.0	0.0	0.0	n/a	n/a
adjusted high	4965	37.50	5003	380.1	380.1	380.0	380.0	1.000	1.000
mid	4985	17.80	5003	180.4	180.4	180.0	180.0	1.002	1.002
low	4991	8.90	5000	90.2	90.2	90.0	90.0	1.003	1.003
calibrator zero	5000	0.00	5000	0	0	0.0	0.0	n/a	n/a
Average C.F.=								1.002	1.002

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calibrator Setting	Indicated NO	Indicated NOx	Indicated NO ₂	NO drop	NO ₂ gain	NO ₂ C.F.
Point	Diluent	Cal Gas	Total Flow	volts or ppb	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)
NOx reference	4965	37.50	5003	0.0	383.0	383.0	0.0	0.0	0.0	
as found high NO2	4965	37.50	5003	235.0	131.0	382.0	252.0	252.0	252.0	1.000
adjusted high NO2	4965	37.50	5003	235.0	131.0	382.0	252.0	252.0	252.0	1.000
gpt mid	4965	37.50	5003	137.0	238.0	383.0	145.0	145.0	145.0	1.000
gpt low	4965	37.50	5003	45.0	333.0	383.0	50.0	50.0	50.0	1.000
Average NO₂ C.F.=										1.000

Linear Regression/Calibration Results:

	NO	NOx	NO ₂	LIMITS
Correlation Coefficient =	1.000	1.000	1.000	> or = 0.995
Slope =	1.000	1.000	1.000	.95-1.05
b (Intercept as % of full scale)=	-0.03%	-0.03%	0.00%	± 3% F.S.
% change in C.F. from last cal=	-0.02%	-0.02%	0.00%	± 10%
NO2 converter efficiency			1.00	0.96 to 1.04

As found:

NO Bkg: 3.7
 NOx Bkg: 3.8
 NO Coef: 1.009
 NO2 Coef: 0.995
 NOx Coef: 0.997
 PMT: -854.7
 Internal: 27.7
 Chamber: 50.5
 Cooler: -2.9
 NO2 Converter: 323.2
 NO2 Converter Set: 325.0
 Pressure: 178.7
 Flow: 0.780
 Ozonator Flow: OK
 Expected Value NO: 1.9
 Expected Value NO2: 268.0
 Expected Value NOx: 270.0

As left:

NO Bkg: 3.7
 NOx Bkg: 3.8
 NO Coef: 1.009
 NO2 Coef: 0.995
 NOx Coef: 0.997
 PMT: -854.4
 Internal: 27.8
 Chamber: 50.3
 Cooler: -3.1
 NO2 Converter: 326.0
 NO2 Converter Set: 325.0
 Pressure: 178.4
 Flow: 0.780
 Ozonator Flow: OK
 Expected Value NO: 2.0
 Expected Value NO2: 270.0
 Expected Value NOx: 272.0

Comments:

The analyzer sample inlet filter was changed.

No high point NO2 adjustment was required/made. As found values were copied to adjusted high values for linearity calculation purposes.




No zero adjustment was required/made. As found zero values were copied to adjusted zero values for linearity calculation purposes.


The analyzer cooling fan filter(s) were cleaned.

No High Point NOx/NO adjustment made.

Date: March 6, 2017
Company/Airshed: LICA
Location/Station Name: Cold Lake South

Start/End Time 24 hr. (mst): 10:35 / 16:03
Calibration Purpose: routine monthly
Calibration Method: Gas Dilution & Gas Phase Titration





— NOX[ppb] — NO[ppb] — NO2[ppb]

OZONE

Maxxam Thermo 49i Ozone Analyzer Calibration

A Bureau Veritas Group Company

Date: March 6, 2017 Barometric Pressure: 0.929 atm
 Company/Airshed: LICA Station Temperature °C: 22
 Location/Station Name: Cold Lake South Weather Conditions: Mainly cloudy with snow
 Start/End Time 24 hr. (mst): 12:42 / 14:47 Calibration Purpose: shut down
 Ozone Calibration Method: Varying UV Lamp Power Performed By/Reviewer: Alex Yakupov Trina Whitsitt
 G.P.T. Date: n/a-done by Varying UV Lamp Power Cal Gas Expiry Date: n/a

Analyzer:
 ID# or Serial Number: 700419951 Ozone Range ppb: 500
 Last Calibration Date: February 23, 2017 As Found C.F.: 1.000
 Previous Cal High Point C.F.: 1.000 New C.F.: n/a

Calibrator:
 Flow Meter ID's: n/a
 Make & Model: SABIO 2010 D
 Serial #: 11900613
 Cal Gas Cylinder I.D. #: n/a

Point	AMD Required Range of Ozone Calibration Points
High	300-400 ppb
Mid	150-200 ppb
Low	50-75 ppb


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rate (cc/min)		Calculated Concentration:	Corrected Calculated Concentration:	Indicated Concentration:	Correction Factors:
	Total Flow @ Point Start	Total Flow @ Point Finish	(ppb)	(ppb)	(ppb)	
as found zero	5000	5000	0.0	n/a	2.0	n/a
as found high	5000	5000	380.0	380.0	382.0	1.000
mid	5000	5000	180.0	180.0	182.0	1.000
low	5000	5000	60.0	60.0	62.0	1.000
Average C.F. =						1.000

Linear Regression/Calibration Results:

Correlation Coefficient =	<u>1.000</u>	LIMITS	> or = 0.995
Slope =	<u>1.000</u>		0.90-1.10
b (Intercept as % of full scale) =	<u>-0.40%</u>		± 3% F.S.
% change in C.F. from last cal =	<u>0.00%</u>		± 10%

Thermo 49i Ozone Analyzer Calibration




As found:	As left:
O3 Bkg: <u>0.1</u>	O3 Bkg: <u>n/a</u>
O3 Coef: <u>0.997</u>	O3 Coef: <u>n/a</u>
Photo Lamp: <u>9.6</u>	Photo Lamp: <u>n/a</u>
O3 Lamp: <u>9.0</u>	O3 Lamp: <u>n/a</u>
Bench: <u>29.7</u>	Bench: <u>n/a</u>
Bench Lamp: <u>53.5</u>	Bench Lamp: <u>n/a</u>
O3 Lamp: <u>67.4</u>	O3 Lamp: <u>n/a</u>
Pressure: <u>700.5</u>	Pressure: <u>n/a</u>
Cell A lpm: <u>0.711</u>	Cell A lpm: <u>n/a</u>
Cell B lpm: <u>0.751</u>	Cell B lpm: <u>n/a</u>
O3 ppb: <u>1.6</u>	O3 ppb: <u>n/a</u>
Cell A ppb: <u>1.1</u>	Cell A ppb: <u>n/a</u>
Cell B ppb: <u>-1.3</u>	Cell B ppb: <u>n/a</u>
Cell A int: <u>87439</u>	Cell A int: <u>n/a</u>
Expected Value: <u>273.0</u>	Expected Value: <u>n/a</u>

Comments:

Shutdown calibration completed for annual maintenance.

O3[ppb] Station: LICA COLD LAKE SOUTH Daily: 17.03.06 Type: AVG 1 Min. [1 Min.]



— O3[ppb]

Maxxam Thermo 49i Ozone Analyzer Calibration

A Bureau Veritas Group Company

Date:	March 7, 2017	Barometric Pressure:	0.936 atm
Company/Airshed:	LICA	Station Temperature °C:	22
Location/Station Name:	Cold Lake South	Weather Conditions:	Mainly cloudy with snow
Start/End Time 24 hr. (mst):	8:29 / 13:23	Calibration Purpose:	post repair
Ozone Calibration Method:	Varying UV Lamp Power	Performed By/Reviewer:	Alex Yakupov Trina Whitsitt
G.P.T. Date:	n/a-done by Varying UV Lamp Power	Cal Gas Expiry Date:	n/a

Analyzer:	
ID# or Serial Number:	700419951
Last Calibration Date:	n/a
Previous Cal High Point C.F.:	n/a
Ozone Range ppb:	500
As Found C.F.:	n/a
New C.F.:	1.000

Calibrator:	
Flow Meter ID's:	n/a
Make & Model:	SABIO 2010 D
Serial #:	11900613
Cal Gas Cylinder I.D. #:	n/a

Point	AMD Required Range of Ozone Calibration Points
High	300-400 ppb
Mid	150-200 ppb
Low	50-75 ppb


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rate (cc/min)		Calculated Concentration:	Corrected Calculated Concentration:	Indicated Concentration:	Correction Factors:
	Total Flow @ Point Start	Total Flow @ Point Finish	(ppb)	(ppb)	(ppb)	
adjusted zero	5000	5000	0.0	n/a	0.0	n/a
adjusted high	5000	5000	380.0	380.0	380.0	1.000
mid	5000	5000	180.0	180.0	181.0	0.994
low	5000	5000	60.0	60.0	61.0	0.984
calibrator zero	5000	5000	0.0	0.0	0.0	n/a
Average C.F. =						0.993

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	1.000		.95-1.05
b (Intercept as % of full scale) =	0.00%		± 3% F.S.
% change in C.F. from last cal =	n/a		± 10%


Thermo 49i Ozone Analyzer Calibration



As found:	As left:
O3 Bkg: n/a	O3 Bkg: 0.3
O3 Coef: n/a	O3 Coef: 1.002
Photo Lamp: n/a	Photo Lamp: 9.6
O3 Lamp: n/a	O3 Lamp: 9.0
Bench: n/a	Bench: 29.2
Bench Lamp: n/a	Bench Lamp: 53.5
O3 Lamp: n/a	O3 Lamp: 67.4
Pressure: n/a	Pressure: 707.7
Cell A lpm: n/a	Cell A lpm: 0.718
Cell B lpm: n/a	Cell B lpm: 0.757
O3 ppb: n/a	O3 ppb: 0.1
Cell A ppb: n/a	Cell A ppb: 3.8
Cell B ppb: n/a	Cell B ppb: -3.6
Cell A int: n/a	Cell A int: 90405
Expected Value: n/a	Expected Value: 270.0

Comments:
 The analyzer sample inlet filter was changed. The analyzer cooling fan filter(s) were cleaned.

Annual maintenance completed. Sample pump and Zero Air pump were rebuilt. Optical cells A and B were cleaned.



— O3[ppb]

PARTICULATE MATTER



R & P 1405F TEOM PM 2.5 Analyzer Calibration

Date: March 7, 2017
 Company: LICA
 Station Name/Location: Cold Lake South
 Previous Audit Date: February 21, 2017
 Parameter: PM 2.5

Performed By/Reviewer: Alex Yakupov | Trina Whitsitt
 Start Time (mst): 10:16
 End Time (mst): 11:39
 Calibration Purpose: Bi-monthly #1
 Weather Conditions: Mainly cloudy with light snow

1400A Information and Status:

ID# or Serial Number: 1405A201620804 As Found Filter Loading %: 30.36
 Ko Factor: 14578 As Left Filter Loading %: 19.82
 Ambient Temperature °C: -14.17 As Found Noise: 0.004
 Ambient Pressure atm: 0.935 As Left Noise: 0.000
 Main Flow Reading lpm: 3.00 Pump Vacuum: 0.39
 Aux Flow Reading lpm: 13.67 Warnings: none

Reference Standards:

	Flow:	Pressure:	Temperature:
Make:	Dwyer	Fisher Scientific	FLUKE
Model:	475 Mark III	FB 1291	1551A Ex STIK
Serial Number:	#3	ID# 05544	ID# 4295
Calibration Date:	January 1, 2017	December 5, 2016	November 15, 2016

As found leak check:

		Base	Zero	Reference	Zero
PM 2.5 Flow	actual	0.01	0.11	0.01	0.11
	limit	0.15	0.15	0.15	0.15
Bypass Flow	actual	0.00	-0.01	0.00	-0.01
	limit	0.60	0.60	0.60	0.60

As left leak check (same as above if as found passes):

		Base	Zero	Reference	Zero
PM 2.5 Flow	actual	0.01	0.11	0.01	0.11
	limit	0.15	0.15	0.15	0.15
Bypass Flow	actual	0.00	-0.01	0.00	-0.01
	limit	0.60	0.60	0.60	0.60

As found temperature and pressure:

1405F temperature °C: <u>-14.2</u>	tolerance +/- 2.0°C	1405F pressure atm: <u>0.935</u>	tolerance +/- 0.01 atm
reference temperature °C: <u>-13.0</u>		reference pressure: <u>0.936</u>	
difference °C: <u>1.1</u>		difference: <u>-0.001</u>	

As left temperature and pressure (same as above if as found adequate):

1405F temperature °C: <u>-13.0</u>	tolerance +/- 2.0°C	1405F pressure atm: <u>0.936</u>	tolerance +/- 0.01 atm
reference temperature °C: <u>-13.0</u>		reference pressure: <u>0.936</u>	
difference °C: <u>0.0</u>		difference: <u>0.000</u>	

As found flows:

main flow tolerance 3.00 lpm +/- 0.20 lpm	total/aux flow tolerance 16.67/13.67 lpm +/- 1.00 lpm +/- 7%
1405F main flow lpm: <u>3.00</u>	1400A total/aux flow lpm: <u>16.67</u>
reference main flow lpm: <u>3.02</u>	reference total/aux flow lpm: <u>16.56</u>
difference lpm: <u>0.02</u>	difference lpm: <u>-0.11</u>

As left flows (same as above if as found adequate):

main flow tolerance 3.00 lpm +/- 0.20 lpm	total/aux flow tolerance 16.67/13.67 lpm +/- 1.00 lpm +/- 7%
1405F main flow lpm: <u>3.00</u>	1400A total/aux flow lpm: <u>16.67</u>
reference main flow lpm: <u>3.02</u>	reference total/aux flow lpm: <u>16.56</u>
difference lpm: <u>0.02</u>	difference lpm: <u>-0.11</u>

K_o Audit:

Last K_o audit date: February 10, 2017
 1405F K_o factor: 14578
 Measured K_o factor: 14789.9000
 % difference: 1.44

Comments:

The TEOM sample filter was changed. The TEOM intake head and associated sharp cut components were cleaned.
 The 47 mm FDMS filter was changed.

Total and Auxiliary flows were measured and the Main flow was calculated.



R & P 1405F TEOM PM 2.5 Analyzer Calibration

Date: March 27, 2017
 Company: LICA
 Station Name/Location: Cold Lake South
 Previous Audit Date: March 7, 2017
 Parameter: PM 2.5

Performed By/Reviewer: Alex Yakupov | Trina Whitsitt
 Start Time (mst): 14:01
 End Time (mst): 15:28
 Calibration Purpose: Bi-monthly #2
 Weather Conditions: Mix of sun and clouds

1400A Information and Status:

ID# or Serial Number: 1405A201620804 As Found Filter Loading %: 31.29
 Ko Factor: 14578 As Left Filter Loading %: 18.15
 Ambient Temperature °C: 8.21 As Found Noise: 0.005
 Ambient Pressure atm: 0.939 As Left Noise: 0.000
 Main Flow Reading lpm: 3.00 Pump Vacuum: 0.39
 Aux Flow Reading lpm: 13.67 Warnings: none

Reference Standards:

	Flow:	Pressure:	Temperature:
Make:	<u>Dwyer</u>	<u>Fisher Scientific</u>	<u>FLUKE</u>
Model:	<u>475 Mark III</u>	<u>FB 1291</u>	<u>1551A Ex STIK</u>
Serial Number:	<u>#3</u>	<u>ID# 05544</u>	<u>ID# 4295</u>
Calibration Date:	<u>January 1, 2017</u>	<u>December 5, 2016</u>	<u>November 15, 2016</u>

As found leak check:

		Base	Zero	Reference	Zero
PM 2.5 Flow	actual	0.00	0.12	0.00	0.12
	limit	0.15	0.15	0.15	0.15
Bypass Flow	actual	0.00	-0.01	0.00	-0.01
	limit	0.60	0.60	0.60	0.60

As left leak check (same as above if as found passes):

		Base	Zero	Reference	Zero
PM 2.5 Flow	actual	0.00	0.12	0.00	0.12
	limit	0.15	0.15	0.15	0.15
Bypass Flow	actual	0.00	-0.01	0.00	-0.01
	limit	0.60	0.60	0.60	0.60

As found temperature and pressure:

tolerance +/- 2.0°C	tolerance +/- 0.01 atm
1405F temperature °C: <u>8.2</u>	1405F pressure atm: <u>0.939</u>
reference temperature °C: <u>7.2</u>	reference pressure: <u>0.935</u>
difference °C: <u>-1.0</u>	difference: <u>0.004</u>

As left temperature and pressure (same as above if as found adequate):

tolerance +/- 2.0°C	tolerance +/- 0.01 atm
1405F temperature °C: <u>7.2</u>	1405F pressure atm: <u>0.935</u>
reference temperature °C: <u>7.2</u>	reference pressure: <u>0.935</u>
difference °C: <u>0.0</u>	difference: <u>0.000</u>

As found flows:

main flow tolerance 3.00 lpm +/- 0.20 lpm	total/aux flow tolerance 16.67/13.67 lpm +/- 1.00 lpm +/- 7%
1405F main flow lpm: <u>3.00</u>	1400A total/aux flow lpm: <u>16.67</u>
reference main flow lpm: <u>3.02</u>	reference total/aux flow lpm: <u>16.08</u>
difference lpm: <u>0.02</u>	difference lpm: <u>-0.59</u>

As left flows (same as above if as found adequate):

main flow tolerance 3.00 lpm +/- 0.20 lpm	total/aux flow tolerance 16.67/13.67 lpm +/- 1.00 lpm +/- 7%
1405F main flow lpm: <u>3.00</u>	1400A total/aux flow lpm: <u>16.67</u>
reference main flow lpm: <u>3.02</u>	reference total/aux flow lpm: <u>16.08</u>
difference lpm: <u>0.02</u>	difference lpm: <u>-0.59</u>

K_o Audit:

Last K_o audit date: February 10, 2017
 1405F K_o factor: 14578
 Measured K_o factor: 14789.9000
 % difference: 1.44

Comments:

The TEOM sample filter was changed. The TEOM intake head and associated sharp cut components were cleaned.
 The 47 mm FDMS filter was changed.
 Total and auxiliary flows were measured and the main flow was calculated.

WIND SYSTEM



Met One Instruments
1600 NW Washington Blvd.
Grants Pass, Oregon 97526
Telephone 541-471-7111
Facsimile 541-471-7116

Regional Service
3206 Main St. Suite 106
Rowlett, Texas 75088
Telephone 972-412-4715
Facsimile 972-412-4716

Sonic Wind Sensor Certificate of Calibration

Sensor Model No: 50.5H Sonic Sensor Serial No: F1644
 Customer: _____ P.O. No: _____ Sales Order: _____
 Final Calibration By: Kevin Ricks Calibration Date: 04-01-15
 Quality Control Inspected By: AJR Inspection Date: APR 03 2015
 New Unit Repair/Adjust Re-Calibration As Found
 Unit Within Tolerance as Found Unit Within Tolerance as Left

Calibration Equipment

Equipment	Manufacturer	Model No.	Serial No.	Cal. Due
Digital Multimeter 1	Agilent/HP	34401A	MY41039534	4/11/2015
Digital Multimeter 2	Agilent/HP	34401A	US36094551	8/26/2015
Frequency Counter	Agilent/HP	53131A	MY40009285	5/22/2015
Standard Sensor	MOI	010C-1	P22383	7/11/2017
Temperature Probe	MOI	920005/PC8340	E3402	9/03/2015

Test 1: Average Wind Tunnel Speed: 3.08 Meters per Second FirmwareVersion: 3194-01 R2.62

WD Setting (Deg)	WD Output (Volts)	WD Indication (Deg)	WD Error (+/- 3 Deg)	WS Standard (m/s)	WS Output (Volts)	WS Indication (m/s)	WS Error (+/- .20 m/s)	Output Type:
30	.084	30.3	.3	3.06	.059	2.96	-.1	0 to 1 volt <input checked="" type="checkbox"/>
60	.165	59.3	-.7	3.07	.059	2.94	-.13	0 to 2.5 volt <input type="checkbox"/>
120	.334	120.2	.2	3.08	.059	2.94	-.14	0 to 5 volt <input type="checkbox"/>
150	.415	149.5	-.5	3.07	.059	2.94	-.13	RS-232 <input checked="" type="checkbox"/>
210	.583	210	0	3.08	.059	2.95	-.12	SDI-12 <input type="checkbox"/>
240	.668	240.3	.3	3.08	.06	2.98	-.1	RS-422 <input type="checkbox"/>
300	.834	300.4	.4	3.07	.06	3.02	-.04	RS-485 <input type="checkbox"/>
330	.916	329.8	-.2	3.09	.059	2.97	-.12	<input type="checkbox"/>

Test 2: Average Wind Tunnel Speed: 11.85 Meters per Second Output Range: 0-50 m/s

WD Setting (Deg)	WD Output (Volts)	WD Indication (Deg)	WD Error (+/- 3 Deg)	WS Standard (m/s)	WS Output (Volts)	WS Indication (m/s)	WS Error (+/- .24 m/s)	Test Items:
30	.081	29.3	-.7	11.79	.235	11.76	-.04	Array Alignment <input checked="" type="checkbox"/>
60	.165	59.5	-.5	11.85	.237	11.87	.01	Jumper Config <input checked="" type="checkbox"/>
120	.331	119.1	-.9	11.85	.236	11.81	-.03	Firmware Config <input checked="" type="checkbox"/>
150	.415	149.3	-.7	11.88	.236	11.8	-.08	Zero Calibration <input checked="" type="checkbox"/>
210	.582	209.5	-.5	11.81	.236	11.79	-.02	Low Speed Test OK <input checked="" type="checkbox"/>
240	.666	239.9	-.1	11.88	.235	11.73	-.16	High Speed Test OK <input checked="" type="checkbox"/>
300	.833	299.7	-.3	11.87	.235	11.73	-.13	Sensor Function <input checked="" type="checkbox"/>
330	.915	329.6	-.4	11.84	.238	11.9	.06	Physical Inspection <input checked="" type="checkbox"/>

The standards used for this calibration have accuracies equal to or greater than the instruments tested. These standards are on record and traceable to NIST to the extent allowed by the institute's calibration facility. Unless otherwise stated hereon, all instruments are calibrated to meet the manufacturer's published specifications. The calibration system complies with MIL-STD-45662A. Calibration performed by direct comparison to the above standard following test procedure: 50.5-6100 Rev E

VOC SAMPLER

Maxxam Analytics

XONTECK FLOW RATE VERIFICATION/CALIBRATION

Client: LICA	Date: March 6, 2017
Location: Cold Lake South	Last Cal. Date: December 19, 2016
Station ID: LICA 01	Start Time 24 hr. (mst): 15:43
Sampler s/n: 6167	End Time 24 hr. (mst): 16:53
Purpose: Routine Quarterly	Performed By/Reviewer: Limin Li Trina Whatsitt

Pressure Standard:	Flow Standard:
Make/Model: Fisher Scientific /FB1291	BUOS DryCal/DC-Lite
S/N or ID#: ID# 05544	s/n 4425
Certification Date: December 5, 2016	February 3, 2017

The desired flow rate can be calculated using the equation provided by USEPA Method T0-14 Section 9.1.3.1.

$$F = \frac{(P \times V)}{(T \times 60)} = \frac{1.61 \times 6000}{24 \times 60} = \boxed{6.71 \text{ cc/min}} = \text{target flow rate}$$

where;

- F= flow rate in cc/min
- P= final canister in atmosphere absolute
- V= volume of canister in c.c.
- T= sampling period in hours
- bp= barometric pressure in atmospheres

enter:

- bp 0.929 atm
- P= 1.60946 (atm)+.68046
- V= 6000 cubic centimetres
- T= 24 hours

XONTECK QUARTERLY FLOW VERIFICATION/CALIBRATION

FLOW RATE VERIFICATION

Volumetric Flow rate =	n/a (cc/min)	As found pot setting =	6.52
Target Flow Rate (cc/min) =	6.71		
% Difference =	#VALUE!		#VALUE!

FLOW RATE CALIBRATION

Volumetric Flow rate =	n/a (cc/min)	Adjusted pot setting =	6.10
Target Flow Rate (cc/min) =	6.71		
% Difference =	#VALUE!		#VALUE!

XONTECK MAINTENANCE

Item:	Most Recent Date Completed:
1. Replace sample line and fittings from sampler to canister every 6 months.	November 9, 2016
2. Purge line from manifold--> sampler with zero air every 6 months.	December 19, 2016
3. Sample system cleaning every 2 years.	
4. Perform 12 hour leak check procedure every 6 months.	March 6, 2017

COMMENTS:

Flow was measured and adjusted for Flow Reading on the sampler from 10.0 to 9.3 and Pot Set Point from 6.52 to 6.10 for the existing conditions of Temperature and Pressure. A VOC canister was installed for 12 hour leak check at 16:45. The canister checked on March 7, 2017 at 08:30. No leaks detected: the canister keeps the initial vacuum of -28.0 in Hg.

CALIBRATORS

Company Maxxam/SIA Operator: Chris

Calibrator:		Flow Measurement Device:	
Make/Model	<u>API 700</u>	Make/Model	<u>Definer 530</u>
Serial Number	<u>627</u>	Serial Number	<u>H-148944, L-152019</u>
Last Verification Date	<u>February 3, 2016</u>	Temperature (°C)	<u>23.5</u>
NO Cylinder S/N	<u>EY0000597</u>	Barometric Pressure	<u>707.1 mmHg</u>
NO [PPM]	<u>49.0</u>	NOx [PPM]	<u>49.0</u>
Expiry Date	<u>December 8, 2019</u>		

Dilution Flow (sccm)		
Pt. #1	<u>4892</u>	Pt. #3 <u>4951</u>
Pt. #2	<u>4975</u>	
Gas Flow (sccm)		
Pt. #1	<u>79.7</u>	Pt. #3 <u>19.4</u>
Pt. #2	<u>38.8</u>	

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
	0.0	0.0000	0.0000	0.0000	-0.0004	-0.0004	Limit ± 10%	
4972	79.7	0.7855	0.7855	0.7883	0.0004	0.7887	0.4%	0.5%
4936	38.8	0.3822	0.3822	0.3816	0.0005	0.3822	-0.2%	0.1%
4970	19.4	0.1913	0.1913	0.1902	0.0006	0.1913	-0.6%	0.2%
Absolute Average Percent Difference							0.1%	0.3%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO	LIMITS	NOx
Correlation= 1.0000	≥ 0.990	Correlation= 1.0000
m (Slope)= 1.0041	0.90-1.10	m (Slope)= 1.0046
b (Intercept % of FS)= -0.1118	± 3% F.S.	b (Intercept % of FS)= -0.0871

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4972	0	0.0000	0.7867	0.0014	0.7881	NO ₂	% Diff, Limit
4972	500	0.5127	0.2740	0.5104	0.7849	-0.7%	± 10%
4972	275	0.2863	0.5004	0.2860	0.7865	-0.6%	± 10%
4972	90	0.0940	0.6927	0.0954	0.7880	0.0%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO ₂	LIMITS
Correlation= 1.0000	≥ 0.995
m (Slope)= 0.9924	0.90-1.10
b (Intercept % of FS)= 0.1755	± 3% F.S.

AENV Standards	NO _x Analyzer
Audit Calibrator	Make/Model <u>Thermo 42i</u>
Make/Model <u>Thermo 146i</u>	Serial/AMU Number <u>AMU 1868</u>
Serial/AMU Number <u>AMU1809</u>	Last Calibration Date <u>January 25, 2017</u>
SRM Gas Cylinder No. <u>CAL018140</u>	Full Scale (ppm) <u>1.0</u>
Cylinder Conc. (ppm) <u>48.79</u>	Cylinder Gas Expiry Date <u>March 25, 2019</u>

COMMENTS:

Auditor: Shea Beaton
Operator Signature: 

Date: January 27, 2017
Location: McIntyre Center Edmonton

Company <u>Maxxam</u>		Operator: <u>Mike</u>	
Calibrator:		Flow Measurement Device:	
Make/Model	<u>Sabio 2010D</u>	Make/Model	<u>Bios Defender 530</u>
Serial Number	<u>11900613</u>	Serial Number	<u>HI148944 Lo 152019</u>
Last Verification Date	<u>March 31, 2016</u>	Temperature (°C)	<u>23.9</u>
NO Cylinder S/N	<u>EY0000769</u>	Barometric Pressure	<u>698mmHg</u>
NO [PPM]	<u>51.1</u>	NOx [PPM]	<u>51.2</u>
Expiry Date	<u>December 8, 2019</u>		

Dilution Flow (sccm)		
Pt. #1 <u>4879</u>	Pt. #2 <u>4932</u>	Pt. #3 <u>4950</u>
Gas Flow (sccm)		
Pt. #1 <u>74.5</u>	Pt. #2 <u>36.4</u>	Pt. #3 <u>18.2</u>

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
4965	0.0	0.0000	0.0000	0.0001	0.0000	0.0001	Limit ± 10%	
4954	74.5	0.7685	0.7700	0.7915	0.0008	0.7923	3%	3%
4968	36.4	0.3744	0.3751	0.3832	0.0006	0.3838	2%	2%
4968	18.2	0.1872	0.1876	0.1916	0.0002	0.1918	2%	2%
Absolute Average Percent Difference							3%	2%

LINEAR REGRESSION ANALYSIS $y=mx+b$ (where x=calculated concentration, y=indicated concentration)

NO	LIMITS	NOx
Correlation= 1.0000	≥ 0.990	Correlation= 1.0000
m (Slope)= 1.0301	0.90-1.10	m (Slope)= 1.0291
b (Intercept % of FS)= -0.0919	± 3% F.S.	b (Intercept % of FS)= -0.0881

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4954	0.000	0.0000	0.7949	0.0005	0.7954	NO ₂	% Diff. Limit
4954	0.510	0.5104	0.2845	0.5072	0.7917	-1%	± 10%
4954	0.250	0.2516	0.5433	0.2514	0.7944	0%	± 10%
4954	0.100	0.1085	0.6864	0.1087	0.7951	0%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS $y=mx+b$ (where x=calculated concentration, y=indicated concentration)

NO ₂	LIMITS
Correlation= 1.0000	≥ 0.995
m (Slope)= 0.9926	0.90-1.10
b (Intercept % of FS)= 0.0925	± 3% F.S.

AENV Standards		NO_x Analyzer	
Audit Calibrator		Make/Model <u>Thermo 42i</u>	
Make/Model	<u>Thermo 146i</u>	Serial/AMU Number	<u>1868</u>
Serial/AMU Number	<u>1809</u>	Last Calibration Date	<u>March 15, 2017</u>
SRM Gas Cylinder No.	<u>CAL018140</u>	Full Scale (ppm)	<u>1.0</u>
Cylinder Conc. (ppm)	<u>48.79</u>	Cylinder Gas Expiry Date	<u>March 28, 2019</u>

COMMENTS: Gas has ~50ppm SO2

Auditor: Shea Beaton
Operator Signature: [Signature]

Date: March 16, 2017
Location: McIntyre Center Edmonton

CALIBRATION GASES



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2016-335CGA

Company: Maxxam **Operator's Name:** Russell Kirchner

Cylinder #: LL104222 Concentration PPM: 50.6 Tolerance(%) 1 Certified By: Praxair

Expiry Date: July 2019

Reference Calibrator and Gas:	Flow Measurement Device:
Make/Model: <u>R&R MFC 201</u>	Make/Model: <u>Bios DC2</u>
Serial Number: <u>AMU 1690</u>	Serial Number: <u>AMY 1659</u>
Last Verification Date: <u>October 19, 2016</u>	Temp. °C: <u>24.5 C</u>
Gas Type: <u>SO2</u> Conc. <u>98.07</u>	B.P. <u>706 mmhg</u>
Cylinder Number: <u>CA:016625</u>	
Expiry Date: <u>January 2019</u>	

Reference Analyzer:

Make/Model: Teco 43C Serial/AMU Number: 1623

Instrument Settings: Zero: 9.2 Span: 1.024 Range: 1.0

Last Calibration: Date: Oct 19/16 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.000	0.0000	0.0000	0.000
4935	82.0	0.830	0.01662	60.183	50.0
4968	40.8	0.412	0.00821	121.765	50.2
4955	20.2	0.203	0.00408	245.297	49.8
Average Cylinder Concentration:					50.0

Previous Stated Concentration PPM: 50.6

Percent variance from Stated: 1

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____

< =5% Outside Manufacturer Tolerance. Use manufacturers concentration _____

> 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark

Operator Signature: *Al Clark*

Date: October 19, 2016

Location: McIntyre Center Edmonton



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2016-334CGA

Company: Maxxam **Operator's Name:** Russell Kirchner
Cylinder #: EY0000654 **Concentration PPM:** 10.2 **Tolerance(%)** 2 **Certified By:** Praxair
Expiry Date: June 2019

Reference Calibrator and Gas:

Make/Model: R&R MFC 201
Serial Number: AMU 1690
Last Verification Date: October 19, 2016
Gas Type: H2S **Conc.** 20.43
Cylinder Number: CAL015584
Expiry Date: January 2019

Flow Measurement Device:

Make/Model: Bios DC2
Serial Number: AMU 1659
Temp. °C: 24.0 C
B.P. 706 mmhg

Reference Analyzer:

Make/Model: Teco 450i **Serial/AMU Number:** 1980
Instrument Settings: **Zero:** 16.6 **Span:** 1.231 **Range:** 0.1
Last Calibration: **Date:** Oct 19/16 **C.F.** 1.000 **Done By:** Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.0000	 	 	
5050	38.0	0.0764	0.00752	132.895	10.2
5050	17.8	0.0355	0.00352	283.708	10.1
5023	9.1	0.0182	0.00181	551.978	10.0
Average Cylinder Concentration:					10.1

Previous Stated Concentration PPM: 10.2

Percent variance from Stated: 1

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration _____
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark
Operator Signature: *Al Clark*

Date: October 19, 2016
Location: McIntyre Center Edmonton



Calibration Gas Audit

CH4 / C3H8 Cylinder Gas

File No. 2015-092CGA

Company: Maxxam Operators name: Chris Wesson
Cylinder #: LL165372 Conc CH4 (PPM) 606/212 Tolerance (%) 0.5 Certified By: Praxair

Reference Calibrator and Gas:

Make/Model R&R MFC 201
Serial Number AMU 1698
Last Verification Date January 18, 2016
Gas Type CH4 Conc. 999.2
Cylinder Number D751932
Gas Type C3H8 Conc. 246.5
Cylinder Number XF0037998

Flow Measurement Device:

Make/Model Bios DC-2
Serial Number Blos D
Temp. °C 24.5
B.P. 688mmHg

Reference Analyzer:

Make/Model Thermo 55C Serial/AMU Number: 1643
Instrument Settings Zero: NA Span: NA Range: 20.0
Last Calibration: Date: 18-Jan-16 C.F. 1.000 Done By: SB

Calibrator Flows (scem)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	CH4	C3H8			CH4	C3H8
2568	0.00	0.00	0.00	0.02140	46.722	607	214
2630	56.29	12.99	12.62	0.02140	46.722	607	214
2588	19.73	4.62	4.50	0.00762	131.171	606	215
2580	9.69	2.29	2.24	0.00376	266.254	610	217
Average Cylinder Concentration:						608	215

CH4	C3H8
Previous Stated Concentration PPM: <u>606</u>	<u>212</u>
Percent variance from Stated: <u>0.3</u>	<u>1.6</u>

Cylinder gas tolerances based on CH4 only

Meets Manufacturer Tolerance. Use manufacturers stated concentration COMMENTS: _____
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration C3H8 manufacturers tolerance 1.1%
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Shea Beaton Date: January 19, 2016
Operator Signature: _____ Location: McIntyre Center Edmonton



Calibration Gas Audit

NO Cylinder Gas

File No. 2016-336CGA

Company: Maxxam **Operators name:** Russell Kirchner

Cylinder #: LL104222 Conc (PPM) 50.7/50.9 Tolerance (%) 1 Certified By: Praxair

Expiry Date: July 2019

Reference Calibrator and Gas:				Flow Measurement Device:	
Make/Model	<u>Teco 146i</u>			Make/Model	<u>Bios DC2</u>
Serial Number	<u>AMU 1809</u>			Serial Number	<u>AMU 1659</u>
Last Verification Date	<u>October 19, 2019</u>			Temp. °C	<u>24.5 C</u>
Gas Type	<u>NO</u>	Conc.	<u>48.79</u>	B.P.	<u>706 mmhg</u>
Cylinder Number	<u>CAL018188</u>				
Expiry Date	<u>March 2019</u>				

Reference Analyzer:

Make/Model Teco 42i Serial/AMU Number: 1868

Instrument Settings Zero: 4.4 Span: 1.080 Range: 1.0

Last Calibration: Date: Oct 18/16 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	NO	NOX			NO	NOX
5000	0.0	0.000	0.000				
4935	82.0	0.838	0.837	0.017	60.183	50.4	50.4
4968	40.8	0.417	0.417	0.008	121.765	50.8	50.8
4955	20.2	0.207	0.207	0.004	245.297	50.8	50.8
Average Cylinder Concentration:						50.7	50.6

<u>NO</u>	<u>NOx</u>
Previous Stated Concentration PPM: <u>50.7</u>	<u>50.9</u>
Percent variance from Stated: <u>0</u>	<u>1</u>

Cylinder gas tolerances based on NO only

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:**

< =5% Outside Manufacturer Tolerance. Use manufacturers concentration Contains 50.6 ppm SO2.

> 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark Date: October 19, 2016

Operator Signature: *Al Clark* Location: McIntyre Center Edmonton

APPENDIX IV
ANALYTICAL RESULTS

PASSIVE SAMPLES

Your Project #: 2017/01/30 - 2017/03/30
Site Location: LICA

Attention: MICHAEL BISAGA

LAKELAND INDUSTRY AND COMMUNITY ASSOCIATION
PO BOX 8237
5107W- 50TH STREET
BONNYVILLE, AB
CANADA T9N 2J5

Report Date: 2017/04/13
Report #: R2369247
Version: 1 - Final

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: B724492

Received: 2017/04/04, 11:19

Sample Matrix: Air
Samples Received: 33

Analyses	Quantity	Date	Date	Laboratory Method	Analytical Method
		Extracted	Analyzed		
H2S Passive Analysis (1)	20	2017/04/11	2017/04/13	PTC SOP-00150	Tang.Passive H2S in
NO2 Passive Analysis (1)	5	2017/04/10	2017/04/13	PTC SOP-00148	Passive NO2 in ATM
NO2 Passive Analysis (1)	20	2017/04/11	2017/04/13	PTC SOP-00148	Passive NO2 in ATM
O3 Passive Analysis (1)	25	2017/04/07	2017/04/13	PTC SOP-00197	EPA 300 R2.1
SO2 Passive Analysis (1)	29	2017/04/05	2017/04/13	PTC SOP-00149	Passive SO2 in Air

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) The detection limit is based on a 30 day sampling period.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

Levi Manchak, Project Manager

Email: LManchak@maxxam.ca

Phone# (780)468-3536

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. For Service Group specific validation please refer to the Validation Signature Page.

Maxxam Job #: B724492
Report Date: 2017/04/13

LAKELAND INDUSTRY AND COMMUNITY ASSOCIATION
Client Project #: 2017/01/30 - 2017/03/30
Site Location: LICA
Sampler Initials: AY

RESULTS OF CHEMICAL ANALYSES OF AIR

Maxxam ID		QV2133	QV2134	QV2137	QV2145		QV2151	QV2153		
Sampling Date		2017/01/30 16:47	2017/01/31 11:18	2017/01/31 13:51	2017/01/31 15:11		2017/01/31 11:31	2017/01/31 09:58		
	UNITS	3	4	5	6	QC Batch	8	9	RDL	QC Batch

Passive Monitoring										
Calculated H2S	ppb	0.12		0.12		8601650			0.02	8601650
Calculated NO2	ppb	0.7	0.5	0.9	1.6	8600439	0.9	0.4	0.1	8600786
Calculated O3	ppb	36.7	43.4	38.0	34.7	8598275	35.6	35.2	0.1	8598275
Calculated SO2	ppb	0.4	0.7	0.5	0.4	8596513	0.5	0.4	0.1	8596513
RDL = Reportable Detection Limit										

Maxxam ID		QV2154	QV2155	QV2156	QV2157	QV2161		QV2162		
Sampling Date		2017/01/30 15:37	2017/01/30	2017/01/30	2017/01/30 12:35	2017/01/30 11:34		2017/01/31 08:35		
	UNITS	10	11	12	13	14	QC Batch	15	RDL	QC Batch

Passive Monitoring										
Calculated H2S	ppb	0.09	MISSING	MISSING	0.10	0.13	8601648		0.02	8601648
Calculated NO2	ppb	1.2	MISSING	MISSING	0.4	0.7	8600786	0.5	0.1	8600786
Calculated O3	ppb	29.8	MISSING	MISSING	33.5	30.9	8598275	39.0	0.1	8598820
Calculated SO2	ppb	0.5	MISSING	MISSING	0.4	1.3	8596513	0.4	0.1	8596513
RDL = Reportable Detection Limit										

Maxxam ID		QV2163	QV2164	QV2165	QV2166	QV2167	QV2168	QV2169		
Sampling Date		2017/01/31 18:46	2017/01/31 16:15	2017/01/31 17:46	2017/01/31 19:23	2017/02/01 09:09	2017/01/30 14:26	2017/01/31 14:33		
	UNITS	16	17	18	19	22	23	24	RDL	QC Batch

Passive Monitoring										
Calculated H2S	ppb	0.11	0.16	0.08		0.09		0.11	0.02	8601648
Calculated NO2	ppb	0.6	0.9	0.5	0.4	0.9	<0.1	2.0	0.1	8600786
Calculated O3	ppb	37.0	35.6	33.7	31.8	31.3	33.3	32.5	0.1	8598820
Calculated SO2	ppb	0.4	0.4	0.3	0.4	0.3	0.2	0.5	0.1	8596513
RDL = Reportable Detection Limit										

Maxxam ID		QV2170	QV2171	QV2172	QV2173	QV2174	QV2175	QV2176		
Sampling Date		2017/01/30	2017/01/30 11:56	2017/01/30 13:24	2017/01/31 09:32	2017/02/01 09:16	2017/01/30 17:44	2017/01/30 10:28		
	UNITS	25	26	27	28	29	32	38	RDL	QC Batch

Passive Monitoring										
Calculated H2S	ppb	MISSING	0.14	0.13		0.11	0.12	0.26	0.02	8601648
Calculated NO2	ppb				3.0	1.1	0.6	2.6	0.1	8600786
Calculated O3	ppb				29.8	33.1	36.8	25.9	0.1	8598820
Calculated SO2	ppb	MISSING	0.9	1.0	0.4	0.3	0.5	0.5	0.1	8596518
RDL = Reportable Detection Limit										

Maxxam Job #: B724492
Report Date: 2017/04/13

LAKELAND INDUSTRY AND COMMUNITY ASSOCIATION
Client Project #: 2017/01/30 - 2017/03/30
Site Location: LICA
Sampler Initials: AY

RESULTS OF CHEMICAL ANALYSES OF AIR

Maxxam ID		QV2179		QV2180	QV2181	QV2182	QV2183	QV2184		
Sampling Date		2017/01/31 15:11		2017/01/31 11:31	2017/01/31 18:46	2017/01/31 16:15	2017/01/31 17:46	2017/02/01 09:09		
	UNITS	6 DUP	QC Batch	8 DUP	16 DUP	17 DUP	18 DUP	22 DUP	RDL	QC Batch

Passive Monitoring										
Calculated H2S	ppb		8601648					0.10	0.02	8601648
Calculated NO2	ppb	1.8	8600439	0.8					0.1	8600786
Calculated O3	ppb	30.3	8598275	38.1					0.1	8598275
Calculated SO2	ppb				0.3	0.4	0.3		0.1	8596518

RDL = Reportable Detection Limit

Maxxam ID		QV2185		
Sampling Date		2017/01/31 14:33		
	UNITS	24 DUP	RDL	QC Batch

Passive Monitoring				
Calculated H2S	ppb	0.12	0.02	8601648

RDL = Reportable Detection Limit

Maxxam Job #: B724492
Report Date: 2017/04/13

LAKELAND INDUSTRY AND COMMUNITY ASSOCIATION
Client Project #: 2017/01/30 - 2017/03/30
Site Location: LICA
Sampler Initials: AY

GENERAL COMMENTS

Sample QV2155 [11] : Site inaccessible for sample retrieval.

Results relate only to the items tested.

Maxxam Job #: B724492
Report Date: 2017/04/13

LAKELAND INDUSTRY AND COMMUNITY ASSOCIATION
Client Project #: 2017/01/30 - 2017/03/30
Site Location: LICA
Sampler Initials: AY

QUALITY ASSURANCE REPORT

QA/QC Batch	Init	QC Type	Parameter	Date Analyzed	Value	Recovery	UNITS	QC Limits
8596513	OZ	Spiked Blank	Calculated SO2	2017/04/05		99	%	90 - 110
8596513	OZ	Method Blank	Calculated SO2	2017/04/05	<0.1		ppb	
8596518	OZ	Spiked Blank	Calculated SO2	2017/04/05		98	%	90 - 110
8596518	OZ	Method Blank	Calculated SO2	2017/04/05	<0.1		ppb	
8598275	SS6	Spiked Blank	Calculated O3	2017/04/07		104	%	90 - 110
8598275	SS6	Method Blank	Calculated O3	2017/04/07	<0.1		ppb	
8598820	SS6	Spiked Blank	Calculated O3	2017/04/07		101	%	90 - 110
8598820	SS6	Method Blank	Calculated O3	2017/04/07	<0.1		ppb	
8600439	IK2	Spiked Blank	Calculated NO2	2017/04/10		102	%	90 - 110
8600439	IK2	Method Blank	Calculated NO2	2017/04/10	<0.1		ppb	
8600786	IK2	Spiked Blank	Calculated NO2	2017/04/11		101	%	90 - 110
8600786	IK2	Method Blank	Calculated NO2	2017/04/11	<0.1		ppb	
8601648	LCH	Spiked Blank	Calculated H2S	2017/04/11		99	%	90 - 110
8601650	LCH	Spiked Blank	Calculated H2S	2017/04/11		99	%	90 - 110

Spiked Blank: A blank matrix sample to which a known amount of the analyte, usually from a second source, has been added. Used to evaluate method accuracy.

Method Blank: A blank matrix containing all reagents used in the analytical procedure. Used to identify laboratory contamination.

Maxxam Job #: B724492
Report Date: 2017/04/13

LAKELAND INDUSTRY AND COMMUNITY ASSOCIATION
Client Project #: 2017/01/30 - 2017/03/30
Site Location: LICA
Sampler Initials: AY

VALIDATION SIGNATURE PAGE

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).



Linda Lin, Supervisor, Centre for Passive Sampling Technology

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. For Service Group specific validation please refer to the Validation Signature Page.

VOCS SAMPLES

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 2, 2017	S5647	Ambient Air	02-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030076	REPORT CREATED:	10-Apr-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030076-001	1,1,1-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	1,1,2,2-Tetrachloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	1,1,2-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	1,1-Dichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	1,1-Dichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030076-001	1,2,3-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030076-001	1,2,4-Trichlorobenzene	K, T, U	< 0.8	ppbv	0.8	AC-058	15-Mar-17
17030076-001	1,2,4-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030076-001	1,2-Dibromoethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	1,2-Dichlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030076-001	1,2-Dichloroethane	I	0.02	ppbv	0.01	AC-058	15-Mar-17
17030076-001	1,2-Dichloropropane	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030076-001	1,3,5-Trimethylbenzene		0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	1,3-Butadiene	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	1,3-Dichlorobenzene	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030076-001	1,4-Dichlorobenzene	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030076-001	1,4-Dioxane	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030076-001	1-Butene		0.09	ppbv	0.02	AC-058	15-Mar-17
17030076-001	1-Hexene	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	1-Pentene		0.02	ppbv	0.01	AC-058	15-Mar-17
17030076-001	2,2,4-Trimethylpentane		0.07	ppbv	0.01	AC-058	15-Mar-17
17030076-001	2,2-Dimethylbutane		0.03	ppbv	0.01	AC-058	15-Mar-17
17030076-001	2,3,4-Trimethylpentane		0.03	ppbv	0.01	AC-058	15-Mar-17
17030076-001	2,3-Dimethylbutane		0.05	ppbv	0.02	AC-058	15-Mar-17
17030076-001	2,3-Dimethylpentane		0.06	ppbv	0.02	AC-058	15-Mar-17

Report certified by:	Graham Knox, Team Lead	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services	
Date:	April-10-17	Inquiries:	(780) 632 8455	E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
LICA/VOC/CLS/Mar 2, 2017	S5647	Ambient Air	02-Mar-17 0:00
DESCRIPTION:	Cold Lake South		
REPORT NUMBER:	17030076	REPORT CREATED:	10-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030076-001	2,4-Dimethylpentane		0.04	ppbv	0.01	AC-058	15-Mar-17
17030076-001	2-Methylheptane		0.03	ppbv	0.01	AC-058	15-Mar-17
17030076-001	2-Methylhexane		0.05	ppbv	0.01	AC-058	15-Mar-17
17030076-001	2-Methylpentane		0.20	ppbv	0.01	AC-058	15-Mar-17
17030076-001	3-Methylheptane		0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	3-Methylhexane		0.06	ppbv	0.02	AC-058	15-Mar-17
17030076-001	3-Methylpentane		0.07	ppbv	0.01	AC-058	15-Mar-17
17030076-001	Acetone		1.6	ppbv	0.4	AC-058	15-Mar-17
17030076-001	Acrolein	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030076-001	Benzene		0.22	ppbv	0.01	AC-058	15-Mar-17
17030076-001	Benzyl chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030076-001	Bromodichloromethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	Bromoform	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	Bromomethane	I	0.01	ppbv	0.01	AC-058	15-Mar-17
17030076-001	Carbon disulfide	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030076-001	Carbon tetrachloride	I	0.13	ppbv	0.01	AC-058	15-Mar-17
17030076-001	Chlorobenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	Chloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	Chloroform	I	0.03	ppbv	0.02	AC-058	15-Mar-17
17030076-001	Chloromethane		0.64	ppbv	0.02	AC-058	15-Mar-17
17030076-001	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030076-001	cis-1,3-Dichloropropene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030076-001	cis-2-Butene		0.03	ppbv	0.02	AC-058	15-Mar-17
17030076-001	cis-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	Cyclohexane		0.05	ppbv	0.02	AC-058	15-Mar-17

Report certified by:	Graham Knox, Team Lead	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services
Date:	April-10-17	Inquiries:	(780) 632 8455
		E-mail:	EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 2, 2017	S5647	Ambient Air	02-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030076	REPORT CREATED:	10-Apr-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030076-001	Cyclopentane		0.04	ppbv	0.01	AC-058	15-Mar-17
17030076-001	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030076-001	Ethanol		0.8	ppbv	0.3	AC-058	15-Mar-17
17030076-001	Ethyl acetate	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030076-001	Ethylbenzene		0.03	ppbv	0.01	AC-058	15-Mar-17
17030076-001	Freon-11	I	0.29	ppbv	0.02	AC-058	15-Mar-17
17030076-001	Freon-113	I	0.13	ppbv	0.01	AC-058	15-Mar-17
17030076-001	Freon-114	I	0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	Freon-12		0.61	ppbv	0.02	AC-058	15-Mar-17
17030076-001	Hexachloro-1,3-butadiene	K, T, U	< 0.50	ppbv	0.50	AC-058	15-Mar-17
17030076-001	Isobutane		0.55	ppbv	0.02	AC-058	15-Mar-17
17030076-001	Isopentane		0.45	ppbv	0.03	AC-058	15-Mar-17
17030076-001	Isoprene		0.01	ppbv	0.01	AC-058	15-Mar-17
17030076-001	Isopropyl alcohol	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030076-001	Isopropylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030076-001	m,p-Xylene		0.08	ppbv	0.03	AC-058	15-Mar-17
17030076-001	m-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030076-001	m-Ethyltoluene	K, T, U	< 0.08	ppbv	0.08	AC-058	15-Mar-17
17030076-001	Methyl butyl ketone	K, T, U	< 0.50	ppbv	0.50	AC-058	15-Mar-17
17030076-001	Methyl ethyl ketone	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030076-001	Methyl isobutyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030076-001	Methyl methacrylate	K, T, U	< 0.07	ppbv	0.07	AC-058	15-Mar-17
17030076-001	Methyl tert butyl ether	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030076-001	Methylcyclohexane		0.06	ppbv	0.01	AC-058	15-Mar-17
17030076-001	Methylcyclopentane		0.07	ppbv	0.02	AC-058	15-Mar-17

Report certified by: Graham Knox, Team Lead

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-10-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 2, 2017	S5647	Ambient Air	02-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030076	REPORT CREATED:	10-Apr-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030076-001	Methylene chloride	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030076-001	n-Butane		1.16	ppbv	0.03	AC-058	15-Mar-17
17030076-001	n-Decane	K, T, U	< 0.06	ppbv	0.06	AC-058	15-Mar-17
17030076-001	n-Dodecane	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030076-001	n-Heptane		0.05	ppbv	0.01	AC-058	15-Mar-17
17030076-001	n-Hexane		0.09	ppbv	0.01	AC-058	15-Mar-17
17030076-001	n-Octane		0.03	ppbv	0.02	AC-058	15-Mar-17
17030076-001	n-Pentane		0.3	ppbv	0.1	AC-058	15-Mar-17
17030076-001	n-Propylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030076-001	n-Undecane	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030076-001	Naphthalene	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030076-001	n-Nonane		0.01	ppbv	0.01	AC-058	15-Mar-17
17030076-001	o-Ethyltoluene	I	0.02	ppbv	0.01	AC-058	15-Mar-17
17030076-001	o-Xylene		0.04	ppbv	0.01	AC-058	15-Mar-17
17030076-001	p-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030076-001	p-Ethyltoluene	K, T, U	< 0.07	ppbv	0.07	AC-058	15-Mar-17
17030076-001	Styrene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030076-001	Tetrachloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030076-001	Tetrahydrofuran	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030076-001	Toluene		0.13	ppbv	0.01	AC-058	15-Mar-17
17030076-001	trans-1,2-Dichloroethylene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030076-001	trans-1,3-Dichloropropylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030076-001	trans-2-Butene		0.03	ppbv	0.01	AC-058	15-Mar-17
17030076-001	trans-2-Pentene		0.02	ppbv	0.02	AC-058	15-Mar-17
17030076-001	Trichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17

Report certified by: Graham Knox, Team Lead

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-10-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
LICA/VOC/CLS/Mar 2, 2017	S5647	Ambient Air	02-Mar-17 0:00
DESCRIPTION:	Cold Lake South		
REPORT NUMBER:	17030076	REPORT CREATED:	10-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030076-001	Vinyl acetate	K, T, U	< 0.4 ppbv	0.4	AC-058	15-Mar-17
17030076-001	Vinyl chloride	K, T, U	< 0.02 ppbv	0.02	AC-058	15-Mar-17

Report certified by:	Graham Knox, Team Lead	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services		
Date:	April-10-17	Inquiries:	(780) 632 8455	E-mail:	EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 8, 2017	H3297	Ambient Air	08-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030114	REPORT CREATED:	21-Apr-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030114-001	1,1,1-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	1,1,2,2-Tetrachloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	1,1,2-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	1,1-Dichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	1,1-Dichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030114-001	1,2,3-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030114-001	1,2,4-Trichlorobenzene	K, T, U	< 0.8	ppbv	0.8	AC-058	15-Mar-17
17030114-001	1,2,4-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030114-001	1,2-Dibromoethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	1,2-Dichlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030114-001	1,2-Dichloroethane	I	0.02	ppbv	0.01	AC-058	15-Mar-17
17030114-001	1,2-Dichloropropane	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030114-001	1,3,5-Trimethylbenzene		0.03	ppbv	0.02	AC-058	15-Mar-17
17030114-001	1,3-Butadiene	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	1,3-Dichlorobenzene	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030114-001	1,4-Dichlorobenzene	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030114-001	1,4-Dioxane	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030114-001	1-Butene		0.07	ppbv	0.02	AC-058	15-Mar-17
17030114-001	1-Hexene	I	0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	1-Pentene		0.02	ppbv	0.01	AC-058	15-Mar-17
17030114-001	2,2,4-Trimethylpentane		0.05	ppbv	0.01	AC-058	15-Mar-17
17030114-001	2,2-Dimethylbutane		0.03	ppbv	0.01	AC-058	15-Mar-17
17030114-001	2,3,4-Trimethylpentane		0.03	ppbv	0.01	AC-058	15-Mar-17
17030114-001	2,3-Dimethylbutane		0.03	ppbv	0.02	AC-058	15-Mar-17
17030114-001	2,3-Dimethylpentane		0.04	ppbv	0.02	AC-058	15-Mar-17

Report certified by:	Colleen McGerrigle, Account Coordinator	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services	
Date:	April-21-17	Inquiries:	(780) 632 8455	E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 8, 2017	H3297	Ambient Air	08-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030114	REPORT CREATED:	21-Apr-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030114-001	2,4-Dimethylpentane		0.03	ppbv	0.01	AC-058	15-Mar-17
17030114-001	2-Methylheptane		0.02	ppbv	0.01	AC-058	15-Mar-17
17030114-001	2-Methylhexane		0.03	ppbv	0.01	AC-058	15-Mar-17
17030114-001	2-Methylpentane		0.09	ppbv	0.01	AC-058	15-Mar-17
17030114-001	3-Methylheptane		0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	3-Methylhexane		0.04	ppbv	0.02	AC-058	15-Mar-17
17030114-001	3-Methylpentane		0.05	ppbv	0.01	AC-058	15-Mar-17
17030114-001	Acetone		1.0	ppbv	0.4	AC-058	15-Mar-17
17030114-001	Acrolein	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030114-001	Benzene		0.25	ppbv	0.01	AC-058	15-Mar-17
17030114-001	Benzyl chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030114-001	Bromodichloromethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	Bromoform	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	Bromomethane	I	0.01	ppbv	0.01	AC-058	15-Mar-17
17030114-001	Carbon disulfide	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030114-001	Carbon tetrachloride	I	0.13	ppbv	0.01	AC-058	15-Mar-17
17030114-001	Chlorobenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	Chloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	Chloroform	I	0.03	ppbv	0.02	AC-058	15-Mar-17
17030114-001	Chloromethane		0.67	ppbv	0.02	AC-058	15-Mar-17
17030114-001	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030114-001	cis-1,3-Dichloropropene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030114-001	cis-2-Butene		0.03	ppbv	0.02	AC-058	15-Mar-17
17030114-001	cis-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	Cyclohexane		0.04	ppbv	0.02	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-21-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 8, 2017	H3297	Ambient Air	08-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030114	REPORT CREATED:	21-Apr-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030114-001	Cyclopentane		0.03	ppbv	0.01	AC-058	15-Mar-17
17030114-001	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030114-001	Ethanol	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030114-001	Ethyl acetate	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030114-001	Ethylbenzene		0.02	ppbv	0.01	AC-058	15-Mar-17
17030114-001	Freon-11		0.30	ppbv	0.02	AC-058	15-Mar-17
17030114-001	Freon-113	I	0.12	ppbv	0.01	AC-058	15-Mar-17
17030114-001	Freon-114	I	0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	Freon-12		0.63	ppbv	0.02	AC-058	15-Mar-17
17030114-001	Hexachloro-1,3-butadiene	K, T, U	< 0.50	ppbv	0.50	AC-058	15-Mar-17
17030114-001	Isobutane		0.30	ppbv	0.02	AC-058	15-Mar-17
17030114-001	Isopentane		0.23	ppbv	0.03	AC-058	15-Mar-17
17030114-001	Isoprene		0.02	ppbv	0.01	AC-058	15-Mar-17
17030114-001	Isopropyl alcohol		1.6	ppbv	0.4	AC-058	15-Mar-17
17030114-001	Isopropylbenzene		0.02	ppbv	0.01	AC-058	15-Mar-17
17030114-001	m,p-Xylene		0.07	ppbv	0.03	AC-058	15-Mar-17
17030114-001	m-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030114-001	m-Ethyltoluene	K, T, U	< 0.08	ppbv	0.08	AC-058	15-Mar-17
17030114-001	Methyl butyl ketone	K, T, U	< 0.50	ppbv	0.50	AC-058	15-Mar-17
17030114-001	Methyl ethyl ketone	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030114-001	Methyl isobutyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030114-001	Methyl methacrylate	K, T, U	< 0.07	ppbv	0.07	AC-058	15-Mar-17
17030114-001	Methyl tert butyl ether	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030114-001	Methylcyclohexane		0.04	ppbv	0.01	AC-058	15-Mar-17
17030114-001	Methylcyclopentane		0.04	ppbv	0.02	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-21-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
LICA/VOC/CLS/Mar 8, 2017	H3297	Ambient Air	08-Mar-17 0:00
DESCRIPTION:	Cold Lake South		
REPORT NUMBER:	17030114	REPORT CREATED:	21-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030114-001	Methylene chloride	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030114-001	n-Butane		0.66	ppbv	0.03	AC-058	15-Mar-17
17030114-001	n-Decane	K, T, U	< 0.06	ppbv	0.06	AC-058	15-Mar-17
17030114-001	n-Dodecane	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030114-001	n-Heptane		0.04	ppbv	0.01	AC-058	15-Mar-17
17030114-001	n-Hexane		0.06	ppbv	0.01	AC-058	15-Mar-17
17030114-001	n-Octane		0.03	ppbv	0.02	AC-058	15-Mar-17
17030114-001	n-Pentane		0.1	ppbv	0.1	AC-058	15-Mar-17
17030114-001	n-Propylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030114-001	n-Undecane	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030114-001	Naphthalene		2.0	ppbv	0.5	AC-058	15-Mar-17
17030114-001	n-Nonane		0.03	ppbv	0.01	AC-058	15-Mar-17
17030114-001	o-Ethyltoluene	I	0.02	ppbv	0.01	AC-058	15-Mar-17
17030114-001	o-Xylene		0.03	ppbv	0.01	AC-058	15-Mar-17
17030114-001	p-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030114-001	p-Ethyltoluene	K, T, U	< 0.07	ppbv	0.07	AC-058	15-Mar-17
17030114-001	Styrene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030114-001	Tetrachloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030114-001	Tetrahydrofuran	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030114-001	Toluene		0.13	ppbv	0.01	AC-058	15-Mar-17
17030114-001	trans-1,2-Dichloroethylene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030114-001	trans-1,3-Dichloropropylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030114-001	trans-2-Butene		0.03	ppbv	0.01	AC-058	15-Mar-17
17030114-001	trans-2-Pentene		0.02	ppbv	0.02	AC-058	15-Mar-17
17030114-001	Trichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-21-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 8, 2017	H3297	Ambient Air	08-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030114	REPORT CREATED:	21-Apr-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030114-001	Vinyl acetate	K, T, U	< 0.4 ppbv	0.4	AC-058	15-Mar-17
17030114-001	Vinyl chloride	K, T, U	< 0.02 ppbv	0.02	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-21-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 14, 2017	H2823	Ambient Air	14-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030213	REPORT CREATED:	02-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030213-001	1,1,1-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	1,1,2,2-Tetrachloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	1,1,2-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	1,1-Dichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	1,1-Dichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	22-Mar-17
17030213-001	1,2,3-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	22-Mar-17
17030213-001	1,2,4-Trichlorobenzene	K, T, U	< 0.8	ppbv	0.8	AC-058	22-Mar-17
17030213-001	1,2,4-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	22-Mar-17
17030213-001	1,2-Dibromoethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	1,2-Dichlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	22-Mar-17
17030213-001	1,2-Dichloroethane	I	0.02	ppbv	0.01	AC-058	22-Mar-17
17030213-001	1,2-Dichloropropane	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	1,3,5-Trimethylbenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	1,3-Butadiene	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	1,3-Dichlorobenzene	K, T, U	< 0.3	ppbv	0.3	AC-058	22-Mar-17
17030213-001	1,4-Dichlorobenzene	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030213-001	1,4-Dioxane	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030213-001	1-Butene		0.09	ppbv	0.02	AC-058	22-Mar-17
17030213-001	1-Hexene	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	1-Pentene		0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	2,2,4-Trimethylpentane		0.03	ppbv	0.01	AC-058	22-Mar-17
17030213-001	2,2-Dimethylbutane		0.02	ppbv	0.01	AC-058	22-Mar-17
17030213-001	2,3,4-Trimethylpentane	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	2,3-Dimethylbutane		0.04	ppbv	0.02	AC-058	22-Mar-17
17030213-001	2,3-Dimethylpentane		0.04	ppbv	0.02	AC-058	22-Mar-17

Report certified by:	Rebecca Holgate, Account Coordinator	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services
Date:	Tuesday, May 02, 2017	Inquiries:	(780) 632 8455
		E-mail:	EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 14, 2017	H2823	Ambient Air	14-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030213	REPORT CREATED:	02-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030213-001	2,4-Dimethylpentane		0.02	ppbv	0.01	AC-058	22-Mar-17
17030213-001	2-Methylheptane		0.02	ppbv	0.01	AC-058	22-Mar-17
17030213-001	2-Methylhexane		0.04	ppbv	0.01	AC-058	22-Mar-17
17030213-001	2-Methylpentane		0.15	ppbv	0.01	AC-058	22-Mar-17
17030213-001	3-Methylheptane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	3-Methylhexane		0.05	ppbv	0.02	AC-058	22-Mar-17
17030213-001	3-Methylpentane		0.09	ppbv	0.01	AC-058	22-Mar-17
17030213-001	Acetone		2.7	ppbv	0.4	AC-058	22-Mar-17
17030213-001	Acrolein	K, T, U	< 0.3	ppbv	0.3	AC-058	22-Mar-17
17030213-001	Benzene		0.16	ppbv	0.01	AC-058	22-Mar-17
17030213-001	Benzyl chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030213-001	Bromodichloromethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	Bromoform	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	Bromomethane	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	Carbon disulfide	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	Carbon tetrachloride	I	0.09	ppbv	0.01	AC-058	22-Mar-17
17030213-001	Chlorobenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	Chloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	Chloroform	I	0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	Chloromethane		0.67	ppbv	0.02	AC-058	22-Mar-17
17030213-001	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	cis-1,3-Dichloropropene	K, T, U	< 0.04	ppbv	0.04	AC-058	22-Mar-17
17030213-001	cis-2-Butene	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	cis-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	Cyclohexane		0.05	ppbv	0.02	AC-058	22-Mar-17

Report certified by:	Rebecca Holgate, Account Coordinator	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services
Date:	Tuesday, May 02, 2017	Inquiries:	(780) 632 8455
		E-mail:	EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 14, 2017	H2823	Ambient Air	14-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030213	REPORT CREATED:	02-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030213-001	Cyclopentane		0.04	ppbv	0.01	AC-058	22-Mar-17
17030213-001	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	Ethanol		1.1	ppbv	0.3	AC-058	22-Mar-17
17030213-001	Ethyl acetate	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030213-001	Ethylbenzene		0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	Freon-11	I	0.28	ppbv	0.02	AC-058	22-Mar-17
17030213-001	Freon-113	I	0.09	ppbv	0.01	AC-058	22-Mar-17
17030213-001	Freon-114	I	0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	Freon-12		0.70	ppbv	0.02	AC-058	22-Mar-17
17030213-001	Hexachloro-1,3-butadiene	K, T, U	< 0.50	ppbv	0.50	AC-058	22-Mar-17
17030213-001	Isobutane		1.54	ppbv	0.02	AC-058	22-Mar-17
17030213-001	Isopentane		0.75	ppbv	0.03	AC-058	22-Mar-17
17030213-001	Isoprene	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	Isopropyl alcohol	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030213-001	Isopropylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	m,p-Xylene		0.03	ppbv	0.03	AC-058	22-Mar-17
17030213-001	m-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	22-Mar-17
17030213-001	m-Ethyltoluene	K, T, U	< 0.08	ppbv	0.08	AC-058	22-Mar-17
17030213-001	Methyl butyl ketone	K, T, U	< 0.50	ppbv	0.50	AC-058	22-Mar-17
17030213-001	Methyl ethyl ketone		0.6	ppbv	0.3	AC-058	22-Mar-17
17030213-001	Methyl isobutyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030213-001	Methyl methacrylate	K, T, U	< 0.07	ppbv	0.07	AC-058	22-Mar-17
17030213-001	Methyl tert butyl ether	K, T, U	< 0.03	ppbv	0.03	AC-058	22-Mar-17
17030213-001	Methylcyclohexane		0.09	ppbv	0.01	AC-058	22-Mar-17
17030213-001	Methylcyclopentane		0.10	ppbv	0.02	AC-058	22-Mar-17

Report certified by: Rebecca Holgate, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Tuesday, May 02, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 14, 2017	H2823	Ambient Air	14-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030213	REPORT CREATED:	02-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030213-001	Methylene chloride	K, T, U	< 0.3	ppbv	0.3	AC-058	22-Mar-17
17030213-001	n-Butane		2.75	ppbv	0.03	AC-058	22-Mar-17
17030213-001	n-Decane	K, T, U	< 0.06	ppbv	0.06	AC-058	22-Mar-17
17030213-001	n-Dodecane	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030213-001	n-Heptane		0.06	ppbv	0.01	AC-058	22-Mar-17
17030213-001	n-Hexane		0.17	ppbv	0.01	AC-058	22-Mar-17
17030213-001	n-Octane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	n-Pentane		0.7	ppbv	0.1	AC-058	22-Mar-17
17030213-001	n-Propylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	22-Mar-17
17030213-001	n-Undecane	K, T, U	< 0.5	ppbv	0.5	AC-058	22-Mar-17
17030213-001	Naphthalene	K, T, U	< 0.5	ppbv	0.5	AC-058	22-Mar-17
17030213-001	n-Nonane		0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	o-Ethyltoluene	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	o-Xylene		0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	p-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	22-Mar-17
17030213-001	p-Ethyltoluene	K, T, U	< 0.07	ppbv	0.07	AC-058	22-Mar-17
17030213-001	Styrene	K, T, U	< 0.04	ppbv	0.04	AC-058	22-Mar-17
17030213-001	Tetrachloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	22-Mar-17
17030213-001	Tetrahydrofuran	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030213-001	Toluene		0.08	ppbv	0.01	AC-058	22-Mar-17
17030213-001	trans-1,2-Dichloroethylene	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030213-001	trans-1,3-Dichloropropylene	K, T, U	< 0.04	ppbv	0.04	AC-058	22-Mar-17
17030213-001	trans-2-Butene		0.02	ppbv	0.01	AC-058	22-Mar-17
17030213-001	trans-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030213-001	Trichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	22-Mar-17

Report certified by:	Rebecca Holgate, Account Coordinator	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services
Date:	Tuesday, May 02, 2017	Inquiries:	(780) 632 8455
		E-mail:	EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 14, 2017	H2823	Ambient Air	14-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030213	REPORT CREATED:	02-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030213-001	Vinyl acetate	K, T, U	< 0.4 ppbv	0.4	AC-058	22-Mar-17
17030213-001	Vinyl chloride	K, T, U	< 0.02 ppbv	0.02	AC-058	22-Mar-17

Report certified by: Rebecca Holgate, Account Coordinator **On behalf of:** PJ Pretorius, Manager, Analysis and Testing Services
Date: Tuesday, May 02, 2017 **Inquiries:** (780) 632 8455 **E-mail:** EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 20, 2017	S5627	Ambient Air	20-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030304-001	1,1,1-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	1,1,2,2-Tetrachloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	1,1,2-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	1,1-Dichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	1,1-Dichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-001	1,2,3-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	04-Apr-17
17030304-001	1,2,4-Trichlorobenzene	K, T, U	< 0.8	ppbv	0.8	AC-058	04-Apr-17
17030304-001	1,2,4-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	04-Apr-17
17030304-001	1,2-Dibromoethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	1,2-Dichlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030304-001	1,2-Dichloroethane	I	0.02	ppbv	0.01	AC-058	04-Apr-17
17030304-001	1,2-Dichloropropane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	1,3,5-Trimethylbenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	1,3-Butadiene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	1,3-Dichlorobenzene	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030304-001	1,4-Dichlorobenzene	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-001	1,4-Dioxane	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-001	1-Butene		0.05	ppbv	0.02	AC-058	04-Apr-17
17030304-001	1-Hexene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	1-Pentene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	2,2,4-Trimethylpentane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	2,2-Dimethylbutane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	2,3,4-Trimethylpentane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	2,3-Dimethylbutane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	2,3-Dimethylpentane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17

Report certified by:	Rebecca Holgate, Account Coordinator	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services
Date:	Monday, May 01, 2017	Inquiries:	(780) 632 8455
		E-mail:	EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 20, 2017	S5627	Ambient Air	20-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030304-001	2,4-Dimethylpentane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	2-Methylheptane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	2-Methylhexane		0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	2-Methylpentane		0.03	ppbv	0.01	AC-058	04-Apr-17
17030304-001	3-Methylheptane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	3-Methylhexane		0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	3-Methylpentane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	Acetone		2.7	ppbv	0.4	AC-058	04-Apr-17
17030304-001	Acrolein	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030304-001	Benzene		0.17	ppbv	0.01	AC-058	04-Apr-17
17030304-001	Benzyl chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-001	Bromodichloromethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	Bromoform	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	Bromomethane	I	0.02	ppbv	0.01	AC-058	04-Apr-17
17030304-001	Carbon disulfide	I	0.11	ppbv	0.01	AC-058	04-Apr-17
17030304-001	Carbon tetrachloride	I	0.10	ppbv	0.01	AC-058	04-Apr-17
17030304-001	Chlorobenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	Chloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	Chloroform	I	0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	Chloromethane		0.87	ppbv	0.02	AC-058	04-Apr-17
17030304-001	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	cis-1,3-Dichloropropene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-001	cis-2-Butene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	cis-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	Cyclohexane		0.03	ppbv	0.02	AC-058	04-Apr-17

Report certified by:	Rebecca Holgate, Account Coordinator	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services
Date:	Monday, May 01, 2017	Inquiries:	(780) 632 8455
		E-mail:	EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 20, 2017	S5627	Ambient Air	20-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030304-001	Cyclopentane		0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	Ethanol	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030304-001	Ethyl acetate	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-001	Ethylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	Freon-11	I	0.28	ppbv	0.02	AC-058	04-Apr-17
17030304-001	Freon-113	I	0.11	ppbv	0.01	AC-058	04-Apr-17
17030304-001	Freon-114	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	Freon-12		0.44	ppbv	0.02	AC-058	04-Apr-17
17030304-001	Hexachloro-1,3-butadiene	K, T, U	< 0.50	ppbv	0.50	AC-058	04-Apr-17
17030304-001	Isobutane		0.24	ppbv	0.02	AC-058	04-Apr-17
17030304-001	Isopentane		0.28	ppbv	0.03	AC-058	04-Apr-17
17030304-001	Isoprene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	Isopropyl alcohol	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-001	Isopropylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	m,p-Xylene	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030304-001	m-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-001	m-Ethyltoluene	K, T, U	< 0.08	ppbv	0.08	AC-058	04-Apr-17
17030304-001	Methyl butyl ketone	K, T, U	< 0.50	ppbv	0.50	AC-058	04-Apr-17
17030304-001	Methyl ethyl ketone	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030304-001	Methyl isobutyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-001	Methyl methacrylate	K, T, U	< 0.07	ppbv	0.07	AC-058	04-Apr-17
17030304-001	Methyl tert butyl ether	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030304-001	Methylcyclohexane		0.02	ppbv	0.01	AC-058	04-Apr-17
17030304-001	Methylcyclopentane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17

Report certified by: Rebecca Holgate, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Monday, May 01, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 20, 2017	S5627	Ambient Air	20-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030304-001	Methylene chloride	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030304-001	n-Butane		0.57	ppbv	0.03	AC-058	04-Apr-17
17030304-001	n-Decane	K, T, U	< 0.06	ppbv	0.06	AC-058	04-Apr-17
17030304-001	n-Dodecane	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-001	n-Heptane		0.02	ppbv	0.01	AC-058	04-Apr-17
17030304-001	n-Hexane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	n-Octane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	n-Pentane		0.1	ppbv	0.1	AC-058	04-Apr-17
17030304-001	n-Propylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	04-Apr-17
17030304-001	n-Undecane	K, T, U	< 0.5	ppbv	0.5	AC-058	04-Apr-17
17030304-001	Naphthalene	K, T, U	< 0.5	ppbv	0.5	AC-058	04-Apr-17
17030304-001	n-Nonane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	o-Ethyltoluene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	o-Xylene		0.02	ppbv	0.01	AC-058	04-Apr-17
17030304-001	p-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-001	p-Ethyltoluene	K, T, U	< 0.07	ppbv	0.07	AC-058	04-Apr-17
17030304-001	Styrene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-001	Tetrachloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-001	Tetrahydrofuran	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-001	Toluene		0.14	ppbv	0.01	AC-058	04-Apr-17
17030304-001	trans-1,2-Dichloroethylene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	trans-1,3-Dichloropropylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-001	trans-2-Butene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-001	trans-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-001	Trichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17

Report certified by:	Rebecca Holgate, Account Coordinator	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services
Date:	Monday, May 01, 2017	Inquiries:	(780) 632 8455
		E-mail:	EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 20, 2017	S5627	Ambient Air	20-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030304-001	Vinyl acetate	K, T, U	< 0.4 ppbv	0.4	AC-058	04-Apr-17
17030304-001	Vinyl chloride	K, T, U	< 0.02 ppbv	0.02	AC-058	04-Apr-17

Report certified by: Rebecca Holgate, Account Coordinator	On behalf of: PJ Pretorius, Manager, Analysis and Testing Services
Date: Monday, May 01, 2017	Inquiries: (780) 632 8455 E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 26, 2017	H3296	Ambient Air	26-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030304-003	1,1,1-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	1,1,2,2-Tetrachloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	1,1,2-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	1,1-Dichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	1,1-Dichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-003	1,2,3-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	04-Apr-17
17030304-003	1,2,4-Trichlorobenzene	K, T, U	< 0.8	ppbv	0.8	AC-058	04-Apr-17
17030304-003	1,2,4-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	04-Apr-17
17030304-003	1,2-Dibromoethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	1,2-Dichlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030304-003	1,2-Dichloroethane	I	0.02	ppbv	0.01	AC-058	04-Apr-17
17030304-003	1,2-Dichloropropane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	1,3,5-Trimethylbenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	1,3-Butadiene	I	0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	1,3-Dichlorobenzene	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030304-003	1,4-Dichlorobenzene	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-003	1,4-Dioxane	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-003	1-Butene		0.08	ppbv	0.02	AC-058	04-Apr-17
17030304-003	1-Hexene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	1-Pentene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	2,2,4-Trimethylpentane		0.04	ppbv	0.01	AC-058	04-Apr-17
17030304-003	2,2-Dimethylbutane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	2,3,4-Trimethylpentane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	2,3-Dimethylbutane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	2,3-Dimethylpentane		0.04	ppbv	0.02	AC-058	04-Apr-17

Report certified by:	Rebecca Holgate, Account Coordinator	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services
Date:	Monday, May 01, 2017	Inquiries:	(780) 632 8455
		E-mail:	EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 26, 2017	H3296	Ambient Air	26-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030304-003	2,4-Dimethylpentane		0.02	ppbv	0.01	AC-058	04-Apr-17
17030304-003	2-Methylheptane		0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	2-Methylhexane		0.03	ppbv	0.01	AC-058	04-Apr-17
17030304-003	2-Methylpentane		0.07	ppbv	0.01	AC-058	04-Apr-17
17030304-003	3-Methylheptane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	3-Methylhexane		0.04	ppbv	0.02	AC-058	04-Apr-17
17030304-003	3-Methylpentane		0.03	ppbv	0.01	AC-058	04-Apr-17
17030304-003	Acetone		3.5	ppbv	0.4	AC-058	04-Apr-17
17030304-003	Acrolein	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030304-003	Benzene		0.16	ppbv	0.01	AC-058	04-Apr-17
17030304-003	Benzyl chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-003	Bromodichloromethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	Bromoform	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	Bromomethane	I	0.03	ppbv	0.01	AC-058	04-Apr-17
17030304-003	Carbon disulfide	I	0.12	ppbv	0.01	AC-058	04-Apr-17
17030304-003	Carbon tetrachloride	I	0.09	ppbv	0.01	AC-058	04-Apr-17
17030304-003	Chlorobenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	Chloroethane	I	0.03	ppbv	0.02	AC-058	04-Apr-17
17030304-003	Chloroform	I	0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	Chloromethane		0.82	ppbv	0.02	AC-058	04-Apr-17
17030304-003	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	cis-1,3-Dichloropropene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-003	cis-2-Butene		0.03	ppbv	0.02	AC-058	04-Apr-17
17030304-003	cis-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	Cyclohexane		0.03	ppbv	0.02	AC-058	04-Apr-17

Report certified by: Rebecca Holgate, Account Coordinator	On behalf of: PJ Pretorius, Manager, Analysis and Testing Services
Date: Monday, May 01, 2017	Inquiries: (780) 632 8455 E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 26, 2017	H3296	Ambient Air	26-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030304-003	Cyclopentane		0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	Ethanol		1.3	ppbv	0.3	AC-058	04-Apr-17
17030304-003	Ethyl acetate	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-003	Ethylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	Freon-11	I	0.23	ppbv	0.02	AC-058	04-Apr-17
17030304-003	Freon-113	I	0.10	ppbv	0.01	AC-058	04-Apr-17
17030304-003	Freon-114	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	Freon-12		0.39	ppbv	0.02	AC-058	04-Apr-17
17030304-003	Hexachloro-1,3-butadiene	K, T, U	< 0.50	ppbv	0.50	AC-058	04-Apr-17
17030304-003	Isobutane		0.53	ppbv	0.02	AC-058	04-Apr-17
17030304-003	Isopentane		0.50	ppbv	0.03	AC-058	04-Apr-17
17030304-003	Isoprene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	Isopropyl alcohol	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-003	Isopropylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	m,p-Xylene		0.06	ppbv	0.03	AC-058	04-Apr-17
17030304-003	m-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-003	m-Ethyltoluene	K, T, U	< 0.08	ppbv	0.08	AC-058	04-Apr-17
17030304-003	Methyl butyl ketone	K, T, U	< 0.50	ppbv	0.50	AC-058	04-Apr-17
17030304-003	Methyl ethyl ketone		0.5	ppbv	0.3	AC-058	04-Apr-17
17030304-003	Methyl isobutyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-003	Methyl methacrylate	K, T, U	< 0.07	ppbv	0.07	AC-058	04-Apr-17
17030304-003	Methyl tert butyl ether	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030304-003	Methylcyclohexane		0.04	ppbv	0.01	AC-058	04-Apr-17
17030304-003	Methylcyclopentane		0.03	ppbv	0.02	AC-058	04-Apr-17

Report certified by: Rebecca Holgate, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Monday, May 01, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 26, 2017	H3296	Ambient Air	26-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030304-003	Methylene chloride	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030304-003	n-Butane		1.15	ppbv	0.03	AC-058	04-Apr-17
17030304-003	n-Decane	K, T, U	< 0.06	ppbv	0.06	AC-058	04-Apr-17
17030304-003	n-Dodecane	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-003	n-Heptane		0.04	ppbv	0.01	AC-058	04-Apr-17
17030304-003	n-Hexane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	n-Octane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	n-Pentane		0.2	ppbv	0.1	AC-058	04-Apr-17
17030304-003	n-Propylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	04-Apr-17
17030304-003	n-Undecane	K, T, U	< 0.5	ppbv	0.5	AC-058	04-Apr-17
17030304-003	Naphthalene	K, T, U	< 0.5	ppbv	0.5	AC-058	04-Apr-17
17030304-003	n-Nonane		0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	o-Ethyltoluene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	o-Xylene		0.05	ppbv	0.01	AC-058	04-Apr-17
17030304-003	p-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-003	p-Ethyltoluene	K, T, U	< 0.07	ppbv	0.07	AC-058	04-Apr-17
17030304-003	Styrene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-003	Tetrachloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-003	Tetrahydrofuran	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030304-003	Toluene		0.17	ppbv	0.01	AC-058	04-Apr-17
17030304-003	trans-1,2-Dichloroethylene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030304-003	trans-1,3-Dichloropropylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030304-003	trans-2-Butene		0.04	ppbv	0.01	AC-058	04-Apr-17
17030304-003	trans-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030304-003	Trichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17

Report certified by: Rebecca Holgate, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Monday, May 01, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/CLS/Mar 26, 2017	H3296	Ambient Air	26-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030304-003	Vinyl acetate	K, T, U	< 0.4 ppbv	0.4	AC-058	04-Apr-17
17030304-003	Vinyl chloride	K, T, U	< 0.02 ppbv	0.02	AC-058	04-Apr-17

Report certified by: Rebecca Holgate, Account Coordinator **On behalf of:** PJ Pretorius, Manager, Analysis and Testing Services
Date: Monday, May 01, 2017 **Inquiries:** (780) 632 8455 **E-mail:** EAS.Results@innotechalberta.ca

PAHS SAMPLES

RESULTS:	Adewunmi Adekanmbi Lakeland Industry and Community Assn 4000, 19 St NE		CLIENT SAMPLE ID	CANISTER ID	Matrix	Priority
	Calgary AB T2E 6P8		LICA/PUF/CLS/Mar 2, 2017	TE03	Air Filter	Normal
INVOICE:	Arianna Cook 780 812 2182 PO Box 8237 5107W-50 St Bonnyville AB T9N 2J5		DESCRIPTION:	Cold Lake South		
			DATE SAMPLED:	02-Mar-17 0:00	DATE RECEIVED:	09-Mar-17
			REPORT CREATED:	10-Apr-17	REPORT NUMBER:	17030076
					VERSION:	Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030076-002	1-Methylnaphthalene		0.18 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	2-Methylnaphthalene		0.28 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	3-Methylcholanthrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	7,12-Dimethylbenz(a)anthracene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Acenaphthene		0.14 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Acenaphthylene		0.03 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Acridine	K, T, U	< 0.01 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Anthracene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Benzo(a)anthracene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Benzo(a)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Benzo(b,j,k)fluoranthene		0.03 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Benzo(c)phenanthrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Benzo(e)pyrene		0.01 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Benzo(ghi)perylene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Chrysene		0.01 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Dibenzo(a,h)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Dibenzo(a,i)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Dibenzo(a,l)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	16-Mar-17

Report certified by: Rebecca Holgate, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-10-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/PUF/CLS/Mar 2, 2017	TE03	Air Filter	02-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030076	REPORT CREATED:	10-Apr-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030076-002	Dibenzo(ah)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Fluoranthene		0.06	ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Fluorene		0.09	ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Indeno(1,2,3-cd)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Naphthalene		0.33	ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Perylene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Phenanthrene		0.24	ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Pyrene		0.04	ug/Filter	0.01	NA-017	16-Mar-17
17030076-002	Retene		0.02	ug/Filter	0.01	NA-017	16-Mar-17

Report certified by: Rebecca Holgate, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-10-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

<p>RESULTS: Adewunmi Adekanmbi Lakeland Industry and Community Assn 4000, 19 St NE</p> <p>Calgary AB T2E 6P8</p> <p>INVOICE: Arianna Cook 780 812 2182 PO Box 8237 5107W-50 St Bonnyville AB T9N 2J5</p>	<p>CLIENT SAMPLE ID LICA/PUF/CLS/Mar 8, 2017</p> <p>CANISTER ID TE06</p> <p>Matrix Air Filter</p> <p>Priority Normal</p> <p>DESCRIPTION: Cold Lake South</p> <p>DATE SAMPLED: 08-Mar-17 0:00</p> <p>REPORT CREATED: 21-Apr-17</p> <p>DATE RECEIVED: 13-Mar-17</p> <p>REPORT NUMBER: 17030114</p> <p>VERSION: Version 01</p>
--	---

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030114-002	1-Methylnaphthalene		0.40	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	2-Methylnaphthalene		0.62	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	3-Methylcholanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	7,12-Dimethylbenz(a)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Acenaphthene		0.08	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Acenaphthylene		0.02	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Acridine	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Anthracene		0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Benzo(a)anthracene		0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Benzo(a)pyrene		0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Benzo(b,j,k)fluoranthene		0.05	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Benzo(c)phenanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Benzo(e)pyrene		0.02	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Benzo(ghi)perylene		0.02	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Chrysene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Dibenzo(a,h)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Dibenzo(a,i)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Dibenzo(a,l)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17

Report certified by: Alberto dos Santos Pereira, Supervisor

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-21-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
LICA/PUF/CLS/Mar 8, 2017	TE06	Air Filter	08-Mar-17 0:00
DESCRIPTION:	Cold Lake South		
REPORT NUMBER:	17030114	REPORT CREATED:	21-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030114-002	Dibenzo(ah)anthracene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Fluoranthene		0.10 ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Fluorene		0.07 ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Indeno(1,2,3-cd)pyrene		0.03 ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Naphthalene		1.12 ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Perylene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Phenanthrene		0.27 ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Pyrene		0.03 ug/Filter	0.01	NA-017	22-Mar-17
17030114-002	Retene		0.07 ug/Filter	0.01	NA-017	22-Mar-17

Report certified by: Alberto dos Santos Pereira, Supervisor

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-21-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
Vegreville, Alberta
Canada T9C 1T4
(780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

<p>RESULTS: Adewunmi Adekanmbi Lakeland Industry and Community Assn 4000, 19 St NE</p> <p>Calgary AB T2E 6P8</p> <p>INVOICE: Arianna Cook 780 812 2182 PO Box 8237 5107W-50 St Bonnyville AB T9N 2J5</p>	<p>CLIENT SAMPLE ID LICA/PUF/CLS/Mar 14, 2017</p> <p>CANISTER ID TE-04</p> <p>Matrix Air Filter</p> <p>Priority Normal</p> <p>DESCRIPTION: Cold Lake South</p> <p>DATE SAMPLED: 14-Mar-17 0:00</p> <p>REPORT CREATED: 02-May-17</p> <p>DATE RECEIVED: 20-Mar-17</p> <p>REPORT NUMBER: 17030213</p> <p>VERSION: Version 01</p>
--	---

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030213-002	1-Methylnaphthalene		0.12	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	2-Methylnaphthalene		0.15	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	3-Methylcholanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	7,12-Dimethylbenz(a)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Acenaphthene		0.05	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Acenaphthylene		0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Acridine	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Anthracene		0.02	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Benzo(a)anthracene		0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Benzo(a)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Benzo(b,j,k)fluoranthene		0.03	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Benzo(c)phenanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Benzo(e)pyrene		0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Benzo(ghi)perylene		0.02	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Chrysene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Dibenzo(a,h)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Dibenzo(a,i)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Dibenzo(a,l)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17

Report certified by: Alberto dos Santos Pereira, Supervisor

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Tuesday, May 02, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/PUF/CLS/Mar 14, 2017	TE-04	Air Filter	14-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030213	REPORT CREATED:	02-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030213-002	Dibenzo(ah)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Fluoranthene		0.06	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Fluorene		0.08	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Indeno(1,2,3-cd)pyrene		0.02	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Naphthalene		0.15	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Perylene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Phenanthrene		0.25	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Pyrene		0.05	ug/Filter	0.01	NA-017	22-Mar-17
17030213-002	Retene		0.03	ug/Filter	0.01	NA-017	22-Mar-17

Report certified by: Alberto dos Santos Pereira, Supervisor	On behalf of: PJ Pretorius, Manager, Analysis and Testing Services
Date: Tuesday, May 02, 2017	Inquiries: (780) 632 8455 E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

RESULTS: Adewunmi Adekanmbi Lakeland Industry and Community Assn 4000, 19 St NE Calgary AB T2E 6P8	CLIENT SAMPLE ID LICA/PUF/CLS/Mar 20, 2017	CANISTER ID A13-02	Matrix Air Filter	Priority Normal
	DESCRIPTION: Cold Lake South			
INVOICE: Arianna Cook PO Box 8237 5107W-50 St Bonnyville AB T9N 2J5	DATE SAMPLED: 20-Mar-17	0:00	DATE RECEIVED: 30-Mar-17	
	REPORT CREATED: 01-May-17		REPORT NUMBER: 17030304	
			VERSION: Version 01	

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030304-002	1-Methylnaphthalene		0.08	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	2-Methylnaphthalene		0.13	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	3-Methylcholanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	7,12-Dimethylbenz(a)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Acenaphthene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Acenaphthylene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Acridine	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Benzo(a)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Benzo(a)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Benzo(b,j,k)fluoranthene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Benzo(c)phenanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Benzo(e)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Benzo(ghi)perylene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Chrysene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Dibenzo(a,h)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Dibenzo(a,i)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Dibenzo(a,l)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17

Report certified by: Rebecca Holgate, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Monday, May 01, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/PUF/CLS/Mar 20, 2017	A13-02	Air Filter	20-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030304-002	Dibenzo(ah)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Fluoranthene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Fluorene		0.02	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Indeno(1,2,3-cd)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Naphthalene		0.14	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Perylene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Phenanthrene		0.03	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Pyrene		0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-002	Retene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17

Report certified by:	Rebecca Holgate, Account Coordinator	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services	
Date:	Monday, May 01, 2017	Inquiries:	(780) 632 8455	E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/PUF/CLS/Mar 26, 2017	TE05	Air Filter	26-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030304-004	1-Methylnaphthalene		0.30	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	2-Methylnaphthalene		0.57	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	3-Methylcholanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	7,12-Dimethylbenz(a)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Acenaphthene		0.04	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Acenaphthylene		0.07	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Acridine	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Anthracene		0.13	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Benzo(a)anthracene		0.02	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Benzo(a)pyrene		0.02	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Benzo(b,j,k)fluoranthene		0.07	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Benzo(c)phenanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Benzo(e)pyrene		0.02	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Benzo(ghi)perylene		0.02	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Chrysene		0.05	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Dibenzo(a,h)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Dibenzo(a,i)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Dibenzo(a,l)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Dibenzo(ah)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Fluoranthene		0.28	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Fluorene		0.25	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Indeno(1,2,3-cd)pyrene		0.03	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Naphthalene		0.35	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Perylene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Phenanthrene		0.84	ug/Filter	0.01	NA-017	13-Apr-17

Report certified by: Rebecca Holgate, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Monday, May 01, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/PUF/CLS/Mar 26, 2017	TE05	Air Filter	26-Mar-17	0:00
DESCRIPTION:	Cold Lake South			
REPORT NUMBER:	17030304	REPORT CREATED:	01-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030304-004	Pyrene		0.23	ug/Filter	0.01	NA-017	13-Apr-17
17030304-004	Retene		0.12	ug/Filter	0.01	NA-017	13-Apr-17

Report certified by: Rebecca Holgate, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Monday, May 01, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

PARTISOL SAMPLES



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

<p>RESULTS: Adewunmi Adekanmbi Lakeland Industry and Community Assn 4000, 19 St NE</p> <p>Calgary AB T2E 6P8</p> <p>INVOICE: Arianna Cook 780 812 2182 PO Box 8237 5107W-50 St Bonnyville AB T9N 2J5</p>	<p>CLIENT SAMPLE ID P6072294</p> <p>CANISTER ID</p> <p>Matrix Air Filter</p> <p>Priority Normal</p> <p>DESCRIPTION: Cold Lake South</p> <p>DATE SAMPLED: 02-Mar-17 0:00</p> <p>REPORT CREATED: 21-Mar-17</p> <p>DATE RECEIVED: 09-Mar-17</p> <p>REPORT NUMBER: 17030081</p> <p>VERSION: Version 01</p>
---	--

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030081-001	Particulate Weight		0.110	mg	0.004	AC-029	10-Mar-17

Report certified by: Krista Gegolick, Account Coordinator **On behalf of:** PJ Pretorius, Manager, Analysis and Testing Services

Date: Tuesday, March 21, 2017 **Inquiries:** (780) 632 8455 **E-mail:** EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

<p>RESULTS: Adewunmi Adekanmbi Lakeland Industry and Community Assn 4000, 19 St NE</p> <p>Calgary AB T2E 6P8</p> <p>INVOICE: Arianna Cook 780 812 2182 PO Box 8237 5107W-50 St Bonnyville AB T9N 2J5</p>	<p>CLIENT SAMPLE ID Flter # P6072293</p> <p>CANISTER ID</p> <p>Matrix Air Filter</p> <p>Priority Normal</p> <p>DESCRIPTION: Cold Lake South</p> <p>DATE SAMPLED: 08-Mar-17 0:00</p> <p>REPORT CREATED: 02-May-17</p> <p>DATE RECEIVED: 13-Mar-17</p> <p>REPORT NUMBER: 17030112</p> <p>VERSION: Version 01</p>
---	---

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030112-001	Particulate Weight		0.037 mg	0.004	AC-029	15-Mar-17

Report certified by: Rebecca Holgate, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Tuesday, May 02, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

<p>RESULTS: Adewunmi Adekanmbi Lakeland Industry and Community Assn 4000, 19 St NE</p> <p>Calgary AB T2E 6P8</p> <p>INVOICE: Arianna Cook 780 812 2182 PO Box 8237 5107W-50 St Bonnyville AB T9N 2J5</p>	<p>CLIENT SAMPLE ID P6072292</p> <p>CANISTER ID</p> <p>Matrix Air Filter</p> <p>Priority Normal</p> <p>DESCRIPTION: Cold Lake South</p> <p>DATE SAMPLED: 14-Mar-17 0:00</p> <p>REPORT CREATED: 18-Apr-17</p> <p>DATE RECEIVED: 20-Mar-17</p> <p>REPORT NUMBER: 17030214</p> <p>VERSION: Version 01</p>
---	---

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030214-001	Particulate Weight		0.123	mg	0.004	AC-029	22-Mar-17

Report certified by: Krista Gegolick, Account Coordinator **On behalf of:** PJ Pretorius, Manager, Analysis and Testing Services

Date: April-18-17 **Inquiries:** (780) 632 8455 **E-mail:** EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

<p>RESULTS: Adewunmi Adekanmbi Lakeland Industry and Community Assn 4000, 19 St NE</p> <p>Calgary AB T2E 6P8</p> <p>INVOICE: Arianna Cook 780 812 2182 PO Box 8237 5107W-50 St Bonnyville AB T9N 2J5</p>	<p>CLIENT SAMPLE ID LICA Flt # P6072291</p> <p>CANISTER ID</p> <p>Matrix Air Filter</p> <p>Priority Normal</p> <p>DESCRIPTION: Cold Lake South P6072291</p> <p>DATE SAMPLED: 20-Mar-17 0:00</p> <p>REPORT CREATED: 13-Apr-17</p> <p>DATE RECEIVED: 30-Mar-17</p> <p>REPORT NUMBER: 17030306</p> <p>VERSION: Version 01</p>
---	---

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030306-001	Particulate Weight		0.033 mg	0.004	AC-029	31-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
Vegreville, Alberta
Canada T9C 1T4
(780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID LICA Flt # P6193008	CANISTER ID	Matrix Air Filter	DATE SAMPLED 26-Mar-17 0:00
DESCRIPTION: Cold Lake South P6193008			
REPORT NUMBER: 17030306	REPORT CREATED: 13-Apr-17		VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030306-002	Particulate Weight		0.126 mg	0.004	AC-029	31-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

APPENDIX V
REPORT CERTIFICATION FORM

Report Certification Form

Alberta Airshed (if applicable)	EPA Approval or Code of Practice Registration # (if applicable)
YES	NA
Company Name (if applicable)	Industrial Operation Name (if applicable)
Lakeland Industry & Community Association	Cold Lake Continuous Monitoring Station
Name of the Representative of the Person Responsible (Last, First, Middle)	Position / Title of the Representative of the Person Responsible
Maram Ghaleb	Project Manager, Customer Service, Air Services
Is an External Party Certifying the Report? (If 'Yes', fill in the fields below for the external person.)	
<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	
Name of External Person Certifying the Report (Last, First, Middle)	Position / Title of External Person Certifying the Report
NA	NA
Company Name for the External Person Certifying the Report	Identification of Qualifications / Professional Designations of the External Person Certifying the Report
NA	NA

I certify that I have reviewed and verified the submitted report. I also certify that the report presented with this certification form is complete, accurate and representative of the monitoring results and timeframe.

Maram Ghaleb

Signature of the Representative of the Person Responsible / External Person Certifying the Report

May 10, 2017

Report Issued Date (dd-mm-yyyy)

APPENDIX VI
DATA VALIDATION CERTIFICATION FORM



Validation Certificate Form

Client: <u>Lakeland Industry & Community Association</u>	Project #: <u>2833-2017-03-1-C</u>
Site: <u>Cold Lake Continuous Monitoring Station</u>	Contact: <u>Mike Bisaga</u>

Level 0 Preliminary Verification	<u>Maram Ghaleb</u>	Date <u>April 30, 2017</u>
Level 1 Primary Validation	<u>Maram Ghaleb</u>	Date <u>May 1, 2017</u>
Level 2 Final Validation	<u>Maram Ghaleb</u>	Date <u>May 9, 2017</u>
Level 3 Independent Data Review	<u>CSA-LMBQ</u>	Date <u>May 10, 2017</u>
Post-Final Validation	<u>NA</u>	Date <u>NA</u>

Notes
The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. This validation is performed on an annual basis.



Alberta Environment and Parks (AEP)
Air.Reporting@gov.ab.ca

February 22, 2018

Subject: Monthly Report Submission for the LICA Maskwa station

Lakeland Industry & Community Association (LICA) is pleased to submit the ambient air monitoring monthly report for the LICA Maskwa AQM Station in the month of March 2017.

The air monitoring program consists of continuous air monitoring results for Sulphur Dioxide (SO₂), Hydrogen Sulphide (H₂S), Total Hydrocarbon (THC), Oxides of Nitrogen (NO_x), Nitric Oxides (NO), Nitrogen Dioxide (NO₂), Relative Humidity (RH), Barometric Pressure (BP), Precipitation, Ambient Temperature (AmbTPX), Wind Speed (WS), Wind Direction (WD) and Standard Deviation Wind Direction (STDWD).

Sampling Program	Monitoring Activities Conducted By	Sample Analysis Conducted By	Data/Report Review and Prepared By	Electronic Submission Conducted By
Continuous ambient air	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics

All data collected in March 2017 was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement systems.

As the LICA Environmental Program Manager and Data & Reporting Specialist, we certify that we have reviewed and verified this report and that the information is complete, accurate and representative of the monitoring results, reporting timeframe and the specified analysis, summarization and reporting requirements. We also certify all air data that are required by the AMD to be electronically submitted to AEP and Alberta's Ambient Air Quality Data Warehouse have been submitted by the time of this report submission.

Should you have any questions, please don't hesitate to contact me.

Respectfully,



Lakeland Industry & Community Association
5107 50 St
Bonnyville, AB T9N 2J7

A handwritten signature in blue ink that reads 'Michael Bisaga'.

Michael Bisaga
Technical Program Manager
Lakeland Industry & Community Association
780-266-7068
mbisaga@otonabee.ca

A handwritten signature in blue ink that reads 'Lily Lin'.

Lily Lin
Data & Reporting Specialist
587-225-2248
rebbacaa@gmail.com



MAXXAM ANALYTICS
#1 2080 39 Ave. NE, Calgary, AB
T2E 6P7

maxxam.ca
Toll Free 800-386-7247
Fax 403-219-3673

AMBIENT AIR MONITORING MONTHLY DATA REPORT
LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
MASKWA CONTINUOUS MONITORING STATION

JOB #: 2833-2017-03-30-C

March 2017

Prepared for:

LAKELAND INDUSTRY & COMMUNITY ASSOCIATION

402 - 19 ST NW
CALGARY, ALBERTA
T2N 2J1

Attention: MIKE BISAGA

DATE: **May 9, 2017**

Prepared by:

A handwritten signature in blue ink, appearing to read "Bim Adeniji".

Bim Adeniji, M.Sc.
Project Manager Assistant, Customer Service, Air Services

Reviewed by:

A handwritten signature in blue ink, appearing to read "Wunmi Adekanmbi".

Wunmi Adekanmbi, M.Sc., EPT.
Project Manager, Customer Service, Air Services

SUMMARY

In March 2017, Maxxam Analytics was contracted to manage the ambient air quality monitoring and maintenance activities at the Maskwa Continuous Monitoring Station, near Bonnyville, Alberta. The monitoring station provides continuous meteorological measurements and air quality data for non-compliance parameters, as requested by Lakeland Industry & Community Association.

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

Annual Maintenance: A scheduled annual station maintenance was completed on March 5 for the SO₂ and THC analyzers.

All Parameters: On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

THC: Three hours of downtime were recorded on March 5, due to the annual maintenance event.

SO₂: Two hours of downtime were recorded on March 5, due to the annual maintenance event.

NO_x/NO/NO₂: The NO_x gas concentration 50.7 ppm labelled as "Calculated NO_x" on the calibration record is not the actual concentration on the certificate of analysis, which is 50.9 ppm. A sample of affected calculations has been rerun and the error has no significant effect on the calibration. The NO_x calibration still meets the AMD calibration criteria.




The summary of results is presented on the following pages.

Any deviations or modifications made to the sampling or analytical methods are outlined in Section 1.0, Discussion. On this basis, Maxxam Analytics is issuing this completed report to Lakeland Industry & Community Association, Maskwa Continuous Monitoring Station.

Should you have any questions concerning the results or if we can be of further assistance, please contact us at 403-219-3677 or toll-free at 1-800-386-7247.

March 2017 Monthly Report Summary

Pollutants		Monthly Records		1-Hour Records					24-Hour Records			
Name	Unit	Avg. Conc.	Uptime	Maximum			AAAQO Objective	Exceed. Hours	Maximum		AAAQO Objective	Exceed. Days
				Conc.	Date	Hour			Conc.	Date		
SO ₂	ppb	0.6	99.5%	12.8	March 18	22	172	0	3.0	March 18	48	0
H ₂ S	ppb	0.1	99.7%	0.8	March 19	2	10	0	0.2	March 15	3	0
THC	ppm	2.15	99.3%	2.55	March 15	17	-	-	2.36	March 14	-	-
NO _x	ppb	2.2	99.7%	24.2	March 19	2	-	-	6.3	March 18	-	-
NO	ppb	0.5	99.7%	9.1	March 19	2	-	-	1.5	March 9	-	-
NO ₂	ppb	1.7	99.7%	15.1	March 19	2	159	0	5.0	March 18	-	-
WS	kph	2.2	99.7%	19.7	March 4	6	-	-	12.6	March 4	-	-
WD	degree	84 (E)	99.7%	-	-	-	-	-	-	-	-	-
RH	%	67	99.7%	91	March 29	6	-	-	83	March 30	-	-
BP	mbar	941	99.7%	963	March 9	6	-	-	962	March 9	-	-
PRECIP	mm	0.0	99.7%	2.8	March 22	14	-	-	0.8	March 22	-	-
AmbTPX	°C	-6.5	99.7%	14.4	March 15	14	-	-	3.9	March 31	-	-



Monthly Update

- * All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.
- * All data collected this month were within the objectives outlined in the AMD 2016 and AAAQO 2016.
- * The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above 90%.

Operational Issues

All Parameters: On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

THC: Three hours of downtime were recorded on March 5, due to an annual maintenance event.

SO₂: Two hours of downtime were recorded on March 5, due to an annual maintenance event.

Annual Maintenance: A scheduled annual station maintenance was completed on March 5 for the SO₂ and THC analyzers.

Monthly Continuous Data Summary

Lakeland Industry & Community Association Maskwa Continuous Monitoring Station						MAXIMUM VALUES							OPERATIONAL TIME (%)
PARAMETER	OBJECTIVES		EXCEEDANCES		MONTHLY AVERAGE	READING	DAY	1-HOUR			24-HOUR		
	1-hr	24-hr	1-hr	24-hr				HOUR	WIND SPEED (kph)	WIND DIRECTION (sector)	READING	DAY	
SO ₂ (ppb)	172	48	0	0	0.6	12.8	18	22	8.5	WNW	3.0	18	99.5
H ₂ S (ppb)	10	3	0	0	0.1	0.8	19	2	13.9	NW	0.2	15	99.7
THC (ppm)	-	-	-	-	2.15	2.55	15	17	2.9	SW	2.36	14	99.3
NO ₂ (ppb)	159	-	0	-	1.7	15.1	19	2	13.9	NW	5.0	18	99.7
NO (ppb)	-	-	-	-	0.5	9.1	19	2	13.9	NW	1.5	9	99.7
NO _x (ppb)	-	-	-	-	2.2	24.2	19	2	13.9	NW	6.3	18	99.7
RELATIVE HUMIDITY (%)	-	-	-	-	67	91	29	6	2	SSW	83	30	99.7
BAROMETRIC PRESSURE (millibar)	-	-	-	-	941	963	9	6	0.6	SSE	962	9	99.7
AMBIENT TEMPERATURE (°C)	-	-	-	-	-6.5	14.4	15	14	4.7	NNW	3.9	31	99.7
PRECIPITATION (mm)	-	-	-	-	0.0	2.8	22	14	5	E	0.8	22	99.7
VECTOR WS (kph)	-	-	-	-	2.2	19.7	4	6	-	NNE	12.6	4	99.7
VECTOR WD (sec)	-	-	-	-	84 (E)	-	-	-	-	-	-	-	99.7

Exceedance Summary Report

SO₂ 1-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 1-hour AAAQO of 172 ppb.

SO₂ 24-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 24-hour AAAQO of 48.0 ppb.

H₂S 1-Hour Exceedances

Measured concentrations of hydrogen sulphide were below the 1-hour AAAQO of 10 ppb.

H₂S 24-Hour Exceedances

Measured concentrations of hydrogen sulphide were below the 24-hour AAAQO of 3 ppb.

NO₂ 1-Hour Exceedances

Measured concentrations of nitrogen dioxide were below the 1-hour AAAQO of 159 ppb.

In accordance with EPEA and the Substance Release Regulation.

In accordance with A Guide to Release Reporting and the Alberta Ambient Air Quality Objectives and Guidelines Summary.

TABLE OF CONTENTS

<u>Title</u>	<u>Page</u>
SUMMARY	1
MONTHLY DASHBOARD	2
MONTHLY CONTINUOUS DATA SUMMARY REPORT	3
EXCEEDANCE SUMMARY REPORT	4
TABLE OF CONTENTS	5
<hr style="border-top: 1px dashed black;"/>	
1.0 Discussion	7
<hr style="border-top: 1px dashed black;"/>	
2.0 Project Personnel	10
<hr style="border-top: 1px dashed black;"/>	
3.0 Plant Monthly Required AMD Summary	10
<hr style="border-top: 1px dashed black;"/>	
4.0 Calculations and Results	10
<hr style="border-top: 1px dashed black;"/>	
5.0 Methods and Procedures	11
<hr style="border-top: 1px dashed black;"/>	
Appendix I	Continuous Monitoring Data Results 14
	Sulphur Dioxide 15
	Hydrogen Sulphide 23
	Total Hydrocarbon 31
	Oxides of Nitrogen 39
	Nitric Oxides 47
	Nitrogen Dioxide 54
	Wind Speed 62
	Wind Direction 69
	Standard Deviation Wind Direction 72
	Relative Humidity 75
	Barometric Pressure 78
	Ambient Temperature 81
	Precipitation 84
Appendix II	Equipment Calibration Results 87
	Sulphur Dioxide 88
	Hydrogen Sulphide 92
	Total Hydrocarbon 95
	Nitrogen Dioxide 99

	Wind System	103
	Meteorological System Check	105
	Calibrators	107
	Calibration Gases	110
Appendix III	Report Certification Form	115
Appendix IV	Data Validation Certification Form	117

1.0 Discussion

This monthly report consists of continuous monitoring results for the following parameters: Sulphur Dioxide (SO₂), Hydrogen Sulphide (H₂S), Total Hydrocarbon (THC), Oxides of Nitrogen (NO_x), Nitric Oxides (NO), Nitrogen Dioxide (NO₂), Relative Humidity (RH), Barometric Pressure (BP), Precipitation, Ambient Temperature (AmbTPX), Wind Speed (WS), Wind Direction (WD) and Standard Deviation Wind Direction (STDWD).

Sample filters for all continuous air monitors are changed before the calibration begins. The sample manifold is cleaned during the site visit each month.

Control checks, consisting of a zero and span, are conducted daily on all continuous air monitors. In place of the air sample, zero air (from scrubbed air or gas cylinders) is used for zero checks, and a known concentration of the pollutant being analyzed is used for span checks. These checks are controlled by automatic timers and valves. The total zero span cycle is completed within an hour, the commencement of the zero span cycle is at the beginning of the hour.

Multipoint calibrations are done a minimum of once a month for each continuous air monitor. An additional calibration is required under the following conditions: 1) within three days after the initial start-up and stabilization of a newly installed instrument, 2) prior to shut-down or moving of an instrument which has been working to specification, and 3) when major repair has been done on the instrument.

Time during the first multi-point calibration is not considered downtime (Data is flagged as C). If more than one calibration is performed during the month, the time during the additional calibration is considered as downtime (Data is flagged as C1).

Only one zero/span check is run per day. Time during the zero/span check is not considered as downtime (Data is flagged as S). If an extra zero/span check is performed, the time during the additional check is considered as downtime (Data is flagged as S1).

The AMD requires each instrument and accompanying data recording system to be operational 90% of the time, at a minimum, for each monthly monitoring period.

All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.

Data contained in this monthly report has undergone the verification and validation based on the requirements of the AMD Chapter 6: Ambient Data Quality (December, 2016). The descriptions of the data verification and validation process can be found in Section 5 of this report. Instantaneous data, where applicable, is provided for reference purposes and has not undergone zero correction. The minimum and maximum statistics are highlighted in the data table and are for reference only. The highlighted cells are based on the software's interpretation of the exact position of the minimum or maximum value. The visual presentation of these statistics may not be the obvious choice in a data range due to rounding, truncating or analyzer specifications.

Hourly/minute data have been reviewed based on daily zero/span results and multi-point calibration results. Data may be considered invalid if a zero-corrected span check in excess of +/- 10% of the span concentration (established by the previous multi-point calibration) is encountered and/or significant differences in the calibration factor occurs (greater than 10%).

SULPHUR DIOXIDE (SO₂)

- Operational time, for the monitoring period was 99.5%, equivalent to four hours of downtime.
- A shut-down calibration was performed on March 5, prior to completing a scheduled annual maintenance on the analyzer. The sample valve was cleaned, the reaction cell was inspected and the high voltage power supply was adjusted. The sample flow rate and pressure were then calibrated. A successful post-repair calibration was completed afterwards. Two hours of downtime were attributed to the maintenance event.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

HYDROGEN SULPHIDE (H₂S)

- Operational time, for the monitoring period was 99.7%, equivalent to two hours of downtime.
- The routine monthly calibration was performed on March 5.
- One hour of maximum instantaneous data on March 5 at hour 16:00, was invalidated due to an anomalous spike.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- One hour of maximum instantaneous data was invalidated on March 5 at hour 09:00, due to a brief power interference that occurred during monthly visit activities.

TOTAL HYDROCARBONS (THC)

- Operational time, for the monitoring period was 99.3%, equivalent to five hours of downtime.
- A shut-down calibration was performed on March 5, prior to completing a scheduled annual maintenance on the analyzer. The zero air pressure sensor was adjusted and the pump was rebuilt. A successful post-repair calibration was completed afterwards. Three hours of downtime were attributed to the maintenance event.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

OXIDES OF NITROGEN (NO_x), NITRIC OXIDE (NO) and NITROGEN DIOXIDE (NO₂)

- Operational time, for the monitoring period was 99.7%, equivalent to two hours of downtime.
- The routine monthly calibration was performed on March 5.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- The NO_x gas concentration 50.7 ppm labelled as "Calculated NO_x" on the calibration record is not the actual concentration on the certificate of analysis, which is 50.9 ppm. A sample of affected calculations has been rerun and the error has no significant effect on the calibration. The NO_x calibration still meets the AMD calibration criteria.

WIND SPEED (WS), WIND DIRECTION (WD) and STANDARD DEVIATION WIND DIRECTION (STDWD)

- Operational time, for the monitoring period was 99.7%, equivalent to two hours of downtime.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- One hour of maximum instantaneous data was invalidated on March 5 at 09:00, due to a brief power interference that occurred during monthly visit activities.
- Wind data is reported as vector wind speed and vector wind direction. Wind direction is defined as the direction from which the wind is blowing from and is measured in degrees from true north.

RELATIVE HUMIDITY (RH)

- Operational time, for the monitoring period was 99.7%, equivalent to two hours of downtime.
- A function check was performed on the relative humidity sensor on March 5. The sensor was determined to operate properly.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

BAROMETRIC PRESSURE (BP)

- Operational time, for the monitoring period was 99.7%, equivalent to two hours of downtime.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

PRECIPITATION (PRECIP)

- Operational time, for the monitoring period was 99.7%, equivalent to two hours of downtime.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

AMBIENT TEMPERATURE (AmbTPX)

- Operational time, for the monitoring period was 99.7%, equivalent to two hours of downtime.
- A function check was performed on the temperature sensor on March 5. The sensor was determined to operate properly.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

2.0 Project Personnel

Mike Bisaga was the contact for Lakeland Industry & Community Association and the Maxxam field technician was Alexander Yakupov.

3.0 Plant Monthly Required AMD Summary

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

4.0 Calculations and Results

All calculations and reporting of results follow the methods described in the AMD, 2016.

5.0 Methods and Procedures

The following methods and procedures were used to complete the monitoring program:

- Maxxam AIR SOP-00209: Ambient Sulphur Monitoring
- Maxxam AIR SOP-00213: Ambient NO/NO₂/NO_x Monitoring
- Maxxam AIR SOP-00214: Ambient Hydrocarbon (THC) Monitoring
- Maxxam AIR SOP-00242: Precipitation Collector Installation/Maintenance

There were no deviations from the prescribed methods.

The following instruments were used to perform the test program:

- Sulphur Dioxide - API 100E UV Fluorescent Analyzer
- Hydrogen Sulphide - API 101A UV Fluorescent Analyzer
- Total Hydrocarbons - Thermo 51C FID Analyzer
- Oxides of Nitrogen - API 200A Chemiluminescent Analyzer
- Wind System - Met One Unit
- Relative Humidity - Met One Unit
- Barometric Pressure - Met One Unit
- Ambient Temperature - Met One Unit
- Precipitation - Met One Unit
- Datalogger - ESC 8832

The following steps were used to complete the data verification and validation process:

Level 0 Preliminary Verification

Level 0 data are raw data obtained directly from the data acquisition system (DAS). Under the step of Level 0, these data undergo a certain amount of manual or automated screening and flagging. It included a) identification of periods of missing data; b) verification of time stamps against reference time; c) verification that instrument diagnostics/datalogger flags indicate normal operation; d) comparison of data to upper and lower limits; e) rate of change flagging indicating that data changed too rapidly or not at all; and f) verification that zero, span and multipoint performance checks are within specifications. This level of verification is performed on a daily basis.

Level 1 Primary Validation

Validation actions under the step of Level 1 include a) review of all screening flags assigned during preliminary verification; b) review of all supporting site information and documentation; c) review of operational acceptance limits for each parameter/analyzer; d) review of daily zero/span and monthly calibration results for all gaseous parameters; and e) application of any necessary adjustments to data (e.g. baseline adjustments, below zero adjustments). This level of validation is performed on a monthly basis.

Level 2 Final Validation


The purpose of Level 2 validation is to verify that there are no inconsistencies among related data, or among regional data measured at nearby sites.

Level 3 Independent Data Review

Level 3 validation is the last step of data review, and it is completed by an individual that is independent of both field operations and primary data validation. A final independent QA review and endorsement is performed during this step before data is submitted to Alberta Environment.

Post-Final Validation

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. Any data issues or patterns which were not clear on a monthly basis are highlighted during this step. This validation is performed on an annual basis.



Source: Air Monitoring Directive (December 2016), Chapter 6, Ambient Data Quality; Figure 1 Data Collection and Management Process Flow Chart

APPENDIX I
CONTINUOUS MONITORING DATA RESULTS

SULPHUR DIOXIDE

SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
2	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.0	1.4	0.0	0.0	0.9	0.4	0.0	0.0	0.0	0.0	1.4	0.1	24
3	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.1	0.4	0.2	0.8	0.3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.8	0.1	24
4	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
5	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	C	C	C	C	C	Y	Y	C	C	C	C	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	22
6	0.0	S	0.0	0.0	0.0	0.0	0.0	0.1	0.3	0.3	0.0	0.0	0.1	0.0	0.1	0.1	0.1	0.0	0.1	0.1	0.2	0.1	0.1	0.2	0.0	0.3	0.1	24	
7	S	0.3	0.4	0.4	0.4	0.5	0.5	0.8	0.6	0.3	0.6	0.4	0.3	0.4	0.5	0.5	0.2	0.2	0.3	0.2	0.1	0.2	0.2	S	0.1	0.8	0.4	24	
8	0.0	0.0	0.2	0.1	0.1	0.0	0.0	0.1	0.2	0.0	0.0	0.0	0.4	0.4	0.3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.4	0.1	24	
9	0.0	0.2	0.3	0.2	0.0	0.0	0.0	0.0	0.0	0.1	0.0	3.5	3.6	2.5	2.9	1.7	0.9	0.8	0.4	0.3	0.2	S	0.0	0.0	0.0	3.6	0.8	24	
10	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.8	2.7	2.1	0.5	0.3	1.3	2.0	0.0	0.0	0.0	0.0	S	0.0	0.3	0.0	0.0	2.7	0.4	24		
11	0.1	0.2	0.1	0.0	0.0	0.0	0.0	X	X	0.0	0.0	1.6	1.8	1.6	1.5	0.3	0.6	0.0	0.0	S	0.4	0.1	0.0	0.1	0.0	1.8	0.4	22	
12	0.2	0.0	0.1	0.1	0.1	0.0	0.0	0.1	0.4	2.8	2.2	2.8	1.6	0.4	0.4	0.9	0.5	0.4	S	0.1	1.6	2.4	0.5	0.3	0.0	2.8	0.8	24	
13	1.3	2.1	0.2	0.2	0.1	0.1	0.2	0.3	0.3	0.7	1.8	1.7	1.1	1.2	1.0	0.8	0.5	S	0.6	0.6	0.5	0.3	0.2	0.3	0.1	2.1	0.7	24	
14	0.1	0.1	0.1	0.1	1.0	0.8	0.6	0.3	0.3	0.4	0.3	0.4	0.9	1.7	2.9	1.2	S	3.0	1.9	0.7	0.5	2.3	4.4	4.1	0.1	4.4	1.2	24	
15	0.8	0.5	0.5	0.4	0.6	0.9	0.5	1.2	1.1	0.7	1.1	4.6	3.6	1.1	0.6	S	0.1	0.2	0.5	0.6	0.4	0.2	0.5	1.2	0.1	4.6	1.0	24	
16	0.3	0.3	0.2	0.3	0.4	0.5	0.4	0.3	0.3	0.3	0.2	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.2	24	
17	0.0	0.8	1.3	0.0	0.0	0.0	0.1	0.0	0.0	0.2	0.5	0.6	0.0	S	0.2	0.9	0.2	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	1.3	0.2	24	
18	0.0	2.4	3.8	4.0	1.1	3.5	3.2	4.0	3.4	1.4	1.5	4.2	S	0.5	3.0	2.6	3.9	1.3	0.8	1.1	1.2	1.2	12.8	7.1	0.0	12.8	3.0	24	
19	5.0	2.4	9.4	6.4	4.4	2.2	0.6	0.6	0.7	0.1	0.1	S	0.2	0.2	0.3	0.4	0.2	0.4	0.4	0.3	0.3	0.5	0.7	11.8	0.1	11.8	2.1	24	
20	3.0	0.9	3.7	5.8	5.0	5.1	9.3	3.7	5.1	3.3	S	2.8	1.2	1.5	1.1	0.8	0.5	0.3	0.1	0.1	0.0	0.0	0.0	0.0	0.0	9.3	2.3	24	
21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.5	S	0.2	0.8	0.9	0.2	0.3	1.5	3.2	3.1	0.9	0.2	0.4	1.4	3.1	4.3	0.0	4.3	0.9	24	
22	3.7	5.2	5.0	4.7	3.0	1.1	0.4	0.2	S	0.1	0.5	0.3	0.2	0.4	0.9	0.6	0.5	0.4	0.4	0.4	1.1	2.0	2.1	6.8	0.1	6.8	1.7	24	
23	4.1	3.9	1.7	0.6	0.4	0.6	0.3	S	0.1	0.4	0.8	0.3	1.4	1.2	2.3	2.7	0.6	0.4	0.1	0.3	0.2	0.0	0.6	0.1	0.0	4.1	1.0	24	
24	0.0	0.0	0.0	0.0	0.1	0.0	S	0.0	0.7	0.0	0.1	0.7	0.2	1.5	0.3	0.1	0.1	0.1	0.3	0.5	2.2	0.4	0.0	0.1	0.0	2.2	0.3	24	
25	0.0	0.1	0.0	0.0	0.0	S	0.0	0.0	0.0	1.0	1.1	1.4	1.4	0.2	1.5	0.4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.5	0.3	24	
26	0.0	0.1	0.9	0.1	S	0.0	0.0	0.4	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.9	0.1	24	
27	0.0	0.0	0.0	S	0.0	0.3	0.0	0.0	0.5	2.6	1.9	1.0	1.6	2.2	1.9	1.7	2.5	2.3	3.9	3.7	1.5	0.0	0.0	0.3	0.0	3.9	1.2	24	
28	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.3	0.0	0.0	0.0	0.1	0.5	0.6	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.6	0.1	24	
29	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.5	0.7	0.3	0.0	0.7	0.1	24	
30	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.1	0.0	24	
31	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	1.3	0.0	S	0.2	0.0	1.3	0.1	24	
HOURLY MAX	5.0	5.2	9.4	6.4	5.0	5.1	9.3	4.0	5.1	3.3	2.7	4.6	3.6	2.5	3.0	2.7	3.9	3.1	3.9	3.7	2.2	2.4	12.8	11.8					
HOURLY AVG	0.6	0.7	1.0	0.8	0.6	0.5	0.6	0.4	0.5	0.5	0.6	1.0	0.8	0.6	0.8	0.7	0.6	0.4	0.4	0.3	0.4	0.4	0.9	1.3					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT:	1-HR	172	ppb	24-HR	48	ppb
----------------------	------	-----	-----	-------	----	-----


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0		
NUMBER OF 24-HR EXCEEDANCES:	0		
NUMBER OF NON-ZERO READINGS:	379		
MINIMUM 1-HR AVERAGE:	0.0 ppb @ HOUR(S) 0 ON DAY(S) 1		
MAXIMUM 1-HR AVERAGE:	12.8 ppb @ HOUR(S) 22 ON DAY(S) 18		
MAXIMUM 24-HR AVERAGE:	3.0 ppb ON DAY(S) 18		
	VAR-VARIOUS		
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	740 hrs
MONTHLY CALIBRATION TIME:	8 hrs	AMD OPERATION UPTIME:	99.5 %
STANDARD DEVIATION:	1.3	MONTHLY AVERAGE:	0.6 ppb

24 HR AVERAGES March 2017



SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - March 2017

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	0.6	0.6	0.6	0.5	0.2	0.3	S	0.3	0.4	0.3	0.2	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.6	0.2	24
2	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.3	2.2	0.6	0.9	1.7	4.5	2.6	5.3	1.6	1.2	4.4	4.0	1.7	1.4	2.1	0.0	5.3	1.5	24
3	1.1	1.6	1.3	1.1	S	1.3	1.4	1.4	1.9	3.0	3.3	3.3	3.7	3.7	3.5	2.9	2.4	2.1	2.0	2.2	2.4	2.2	1.9	2.2	1.1	3.7	2.3	24	
4	1.8	1.8	2.4	S	3.5	1.6	1.7	2.1	1.6	1.9	1.6	1.9	1.8	1.6	1.5	1.7	1.4	1.6	1.2	1.3	1.0	0.8	1.1	1.1	0.8	3.5	1.7	24	
5	1.0	1.0	S	1.0	1.1	1.2	1.2	1.2	1.2	C	C	C	C	C	Y	Y	C	C	C	C	1.6	1.5	1.4	1.5	1.0	1.6	1.2	22	
6	1.6	S	1.5	1.5	1.6	1.5	1.5	1.8	1.8	1.8	1.5	1.3	1.4	1.6	1.6	1.4	1.4	1.3	1.5	1.2	1.2	1.2	1.3	1.3	1.2	1.8	1.5	24	
7	S	1.4	1.4	1.4	1.3	1.2	1.3	1.6	1.4	1.0	1.4	1.0	0.9	1.0	1.1	1.0	0.7	0.5	0.6	0.4	0.4	0.4	0.3	S	0.3	1.6	1.0	24	
8	0.0	0.0	0.2	0.2	0.1	0.0	0.0	0.1	0.1	0.0	0.0	0.0	0.4	1.5	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	1.5	0.1	24
9	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	5.0	8.0	3.8	4.4	3.1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	8.0	1.1	24	
10	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.3	4.5	4.3	1.2	0.9	2.2	2.5	0.7	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	4.5	0.8	24	
11	0.4	0.4	0.0	0.0	0.0	0.0	0.0	X	X	0.0	0.6	2.6	3.5	2.8	3.2	1.5	2.1	0.0	0.0	S	0.7	0.0	0.0	0.0	0.0	3.5	0.8	22	
12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.1	3.5	2.8	4.2	2.4	0.3	1.0	1.0	1.0	0.0	S	0.0	3.8	2.2	1.4	0.0	0.0	4.2	1.1	24	
13	2.1	2.4	0.2	0.0	0.0	0.2	0.2	0.2	0.5	1.3	2.4	2.5	1.8	1.8	1.9	1.9	1.2	S	1.3	1.3	1.2	1.1	0.8	1.0	0.0	2.5	1.2	24	
14	0.9	0.7	0.9	0.9	2.7	2.2	1.8	1.2	1.2	1.3	1.1	1.3	2.9	3.9	8.1	2.6	S	5.2	5.5	1.8	1.6	8.4	7.6	7.4	0.7	8.4	3.1	24	
15	2.2	1.8	1.5	1.6	2.0	2.4	1.9	3.0	3.2	2.2	3.0	19.1	11.5	7.1	5.8	S	1.6	1.6	2.0	1.9	1.7	1.4	1.6	11.1	1.4	19.1	4.0	24	
16	1.5	1.5	1.5	1.5	1.5	1.4	1.5	1.4	1.3	1.3	1.2	1.0	1.0	1.0	S	0.8	0.8	0.8	0.7	0.4	0.4	0.4	0.4	1.4	0.4	1.5	1.1	24	
17	0.9	4.1	4.1	2.2	1.1	0.4	1.3	0.5	0.7	1.0	1.2	1.7	0.5	S	0.9	3.0	1.0	0.7	1.0	1.2	1.0	0.8	0.9	0.9	0.4	4.1	1.4	24	
18	1.6	4.3	6.0	6.3	3.4	7.2	5.8	6.6	7.2	3.4	5.5	9.8	S	2.8	7.0	6.4	8.5	3.5	2.7	2.8	2.9	2.8	27.3	24.4	1.6	27.3	6.9	24	
19	15.4	7.5	16.3	21.3	9.5	5.4	2.2	2.1	2.5	1.3	1.2	S	1.4	1.2	1.7	1.4	1.2	1.4	1.6	0.9	0.8	3.9	4.2	23.6	0.8	23.6	5.6	24	
20	14.3	5.5	10.8	15.7	8.9	15.5	16.2	10.7	15.3	12.3	S	7.6	2.6	2.7	1.8	2.4	1.6	0.4	0.1	0.1	0.0	0.0	0.0	0.0	0.0	16.2	6.3	24	
21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.6	S	0.2	3.2	2.5	0.4	1.1	5.2	6.2	6.4	2.2	0.8	1.2	3.7	6.2	6.8	0.0	6.8	2.0	24	
22	6.2	7.8	7.4	7.5	6.4	3.3	1.9	1.6	S	1.7	2.8	1.5	1.6	1.7	3.5	2.2	1.8	1.7	1.9	1.9	3.1	3.7	3.7	16.5	1.5	16.5	4.0	24	
23	16.1	17.7	5.0	2.1	2.0	2.5	1.8	S	1.5	2.1	2.2	1.8	3.0	2.7	4.1	5.1	1.9	1.7	1.6	1.6	1.3	1.0	2.1	1.5	1.0	17.7	3.6	24	
24	1.0	1.1	0.9	0.9	1.0	0.8	S	0.8	2.6	0.8	2.1	3.2	3.0	3.5	2.0	1.8	1.4	1.4	2.8	2.8	4.8	2.5	1.7	1.7	0.8	4.8	1.9	24	
25	1.7	1.7	1.8	1.7	1.6	S	1.6	1.9	1.8	3.7	3.2	4.4	5.7	2.2	10.0	3.4	1.7	1.7	1.6	1.5	1.5	1.6	1.4	1.6	1.4	10.0	2.6	24	
26	1.4	2.3	2.7	1.9	S	1.3	1.4	2.4	1.6	1.6	1.4	1.4	1.4	1.4	1.5	1.5	1.4	1.4	1.4	1.4	1.4	1.4	1.4	1.4	1.4	1.3	2.7	1.6	24
27	1.3	1.4	1.1	S	1.3	1.9	1.3	1.1	4.3	5.9	5.4	3.8	4.9	6.9	7.0	5.5	8.1	5.8	7.8	6.4	5.6	1.3	1.1	1.6	1.1	8.1	3.9	24	
28	1.1	1.2	S	1.2	1.0	1.0	1.2	1.2	1.1	1.9	2.0	1.2	1.1	1.4	1.4	4.2	2.7	1.6	1.4	1.3	1.3	1.4	1.4	1.3	1.0	4.2	1.5	24	
29	1.3	S	1.2	1.2	1.5	1.3	1.2	1.2	1.3	1.3	1.3	1.7	1.7	1.3	1.5	1.7	1.6	1.6	1.6	1.8	1.6	3.0	3.0	2.4	1.2	3.0	1.6	24	
30	S	1.6	1.6	1.6	1.6	1.3	1.4	1.3	1.3	1.4	1.3	1.3	1.3	1.3	1.2	1.1	1.1	1.1	1.1	1.0	0.9	0.8	0.9	S	0.8	1.6	1.3	24	
31	0.5	0.6	0.6	0.7	0.6	0.6	0.5	1.0	1.6	1.2	1.2	1.2	1.4	1.2	1.5	1.0	1.2	1.3	1.2	2.9	3.6	2.1	S	2.1	0.5	3.6	1.3	24	
HOURLY MAX	16.1	17.7	16.3	21.3	9.5	15.5	16.2	10.7	15.3	12.3	S	7.6	2.6	2.7	1.8	2.4	1.6	0.4	0.1	0.1	0.0	0.0	0.0	0.0	0.0	16.2	6.3	24	
HOURLY AVG	2.6	2.4	2.4	2.6	1.9	1.9	1.7	1.6	2.0	2.0	2.0	3.2	2.5	2.2	2.9	2.4	2.1	1.6	1.6	1.5	1.7	1.7	2.6	4.0					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	597
MAXIMUM INSTANTANEOUS VALUE:	27.3 ppb @ HOUR(S) 22 ON DAY(S) 18
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	9 hrs
OPERATIONAL TIME:	740 hrs
STANDARD DEVIATION:	3.1

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)



Wind: LICA MASKWA
 Poll.: LICA MASKWA-SO2[ppb]
 Monthly: 2017/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 7.15%

Calm Avg: 0.40 [ppb]

Direction	0.0-2.6	2.6-5.2	5.2-7.7	7.7-10.3	10.3-12.9	>12.9	Total
N	10.4	0.1	0.0	0.0	0.0	0.0	10.6
NE	21.0	0.0	0.0	0.0	0.0	0.0	21.0
E	14.5	2.7	0.1	0.0	0.0	0.0	17.3
SE	19.9	1.3	0.0	0.0	0.0	0.0	21.2
S	11.6	0.0	0.0	0.0	0.0	0.0	11.6
SW	2.7	0.1	0.0	0.0	0.0	0.0	2.9
W	3.6	0.3	0.0	0.0	0.0	0.0	3.9
NW	1.7	1.6	0.6	0.3	0.3	0.0	4.4
Summary	85.4	6.2	0.7	0.3	0.3	0.0	92.9

%	Icon	Classes (ppb)	85		0.0-2.6	6		2.6-5.2	1		5.2-7.7	0		7.7-10.3	0		10.3-12.9	0		>12.9
---	------	---------------	----	---	---------	---	---	---------	---	---	---------	---	---	----------	---	---	-----------	---	---	-------

LICA MASKWA Poll.: LICA MASKWA-SO₂[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 7.15% Calm Poll Avg: 0.40[ppb]



SO2[ppb] Calibration: LICA MASKWA Monthly: 2017/03 Type: Span



Span Meas Span Ref Span Low Span High

HYDROGEN SULPHIDE



HYDROGEN SULPHIDE Hourly Averages (H₂S ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	0.0	0.0	0.0	0.1	0.1	0.0	S	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	24
2	0.0	0.0	0.1	0.0	0.0	S	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.2	0.1	0.1	0.0	0.0	0.2	0.2	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.1	24	
3	0.0	0.0	0.0	0.0	S	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.1	0.1	0.1	0.1	0.0	0.2	0.1	0.1	0.1	0.1	0.1	0.0	0.0	0.2	0.1	0.1	24	
4	0.2	0.1	0.1	S	0.4	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.4	0.0	24	
5	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	C	C	C	C	C	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	24		
6	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
7	S	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.1	S	0.0	0.0	0.2	0.0	0.0	24		
8	0.2	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	S	0.1	0.0	0.2	0.0	0.0	24		
9	0.0	0.1	0.1	0.1	0.0	0.0	0.0	0.1	0.3	0.0	0.0	0.0	0.1	0.0	0.1	0.1	0.1	0.1	0.1	0.1	0.0	0.1	S	0.3	0.1	0.0	0.3	0.1	0.0	0.3	0.1	24
10	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.1	0.0	0.0	0.0	0.1	0.0	0.1	0.0	24	
11	0.0	0.3	0.5	0.2	0.1	0.0	0.0	X	X	0.0	0.0	0.0	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.2	0.0	0.0	0.0	0.0	0.0	0.5	0.1	0.1	22	
12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.1	S	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	0.0	24	
13	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	S	0.2	0.1	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.2	0.0	0.0	24	
14	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.1	0.0	0.0	0.0	0.1	0.1	0.1	S	0.3	0.3	0.1	0.1	0.2	0.4	0.4	0.4	0.0	0.4	0.1	0.1	0.1	24	
15	0.2	0.2	0.1	0.1	0.1	0.2	0.2	0.2	0.5	0.2	0.2	0.3	0.2	0.1	0.1	S	0.2	0.1	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.5	0.2	24	
16	0.1	0.1	0.2	0.1	0.4	0.4	0.2	0.1	0.3	0.1	0.0	0.0	0.0	0.0	S	0.2	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.4	0.1	0.1	24	
17	0.0	0.0	0.2	0.2	0.1	0.0	0.0	0.1	0.1	0.1	0.1	0.2	0.1	0.1	S	0.3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.3	0.1	0.0	0.1	24	
18	0.1	0.3	0.3	0.3	0.4	0.4	0.2	0.2	0.2	0.2	0.2	0.1	S	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.6	0.3	0.1	0.6	0.2	0.1	0.1	24	
19	0.2	0.2	0.8	0.5	0.3	0.1	0.0	0.0	0.0	0.0	0.0	S	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.4	0.0	0.8	0.1	0.1	0.1	24	
20	0.1	0.0	0.0	0.1	0.0	0.0	0.2	0.0	0.0	0.0	0.0	S	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.3	0.0	0.3	0.0	0.0	0.0	24	
21	0.2	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.2	0.1	0.0	0.0	0.0	0.0	0.1	0.2	0.0	0.0	0.0	0.1	0.2	0.2	0.0	0.2	0.1	0.1	0.1	24	
22	0.1	0.3	0.2	0.2	0.2	0.1	0.0	0.0	S	0.2	0.1	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.5	0.1	0.1	0.1	0.2	0.1	0.6	0.0	0.6	0.1	0.1	0.1	24	
23	0.5	0.3	0.2	0.0	0.0	0.3	0.0	S	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.3	0.5	0.1	0.0	0.1	0.0	0.0	0.5	0.1	0.1	0.1	24	
24	0.1	0.0	0.1	0.1	0.2	0.0	S	0.2	0.3	0.0	0.0	0.1	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.1	0.0	0.0	0.0	0.3	0.1	0.1	0.1	24	
25	0.1	0.0	0.0	0.0	0.0	S	0.2	0.0	0.0	0.0	0.1	0.1	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.3	0.0	0.0	0.3	0.0	0.0	0.0	24	
26	0.0	0.0	0.0	0.0	S	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.1	0.0	0.2	0.0	0.0	0.0	24	
27	0.1	0.0	0.0	S	0.2	0.2	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.1	0.0	0.0	0.1	0.1	0.0	0.1	0.0	0.0	0.1	0.0	0.0	0.2	0.0	0.0	0.0	24	
28	0.0	0.1	S	0.2	0.0	0.0	0.0	0.1	0.0	0.1	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.1	0.0	0.1	0.0	0.0	0.1	0.2	0.2	0.0	0.2	0.1	0.1	0.1	24	
29	0.1	S	0.3	0.2	0.1	0.1	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.1	0.1	0.3	0.1	0.1	0.1	24	
30	S	0.3	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	S	0.0	0.3	0.0	0.0	0.0	24	
31	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.2	0.0	0.2	0.0	0.0	0.0	0.0	24
HOURLY MAX	0.5	0.3	0.8	0.5	0.4	0.4	0.2	0.2	0.5	0.2	0.2	0.3	0.2	0.2	0.3	0.2	0.2	0.5	0.3	0.5	0.2	0.2	0.6	0.6								
HOURLY AVG	0.1	0.1	0.1	0.1	0.1	0.1	0.0	0.1	0.1	0.0	0.0	0.1	0.1	0.0	0.0	0.0	0.0	0.1	0.1	0.1	0.1	0.1	0.1	0.1								

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 10 ppb 24-HR 3 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0
NUMBER OF 24-HR EXCEEDANCES:	0
NUMBER OF NON-ZERO READINGS:	288
MINIMUM 1-HR AVERAGE:	0.0 ppb @ HOUR(S) 0 ON DAY(S) 1
MAXIMUM 1-HR AVERAGE:	0.8 ppb @ HOUR(S) 2 ON DAY(S) 19
MAXIMUM 24-HR AVERAGE:	0.2 ppb ON DAY(S) 15
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs OPERATIONAL TIME: 742 hrs
MONTHLY CALIBRATION TIME:	5 hrs AMD OPERATION UPTIME: 99.7 %
STANDARD DEVIATION:	0.1 MONTHLY AVERAGE: 0.1 ppb

24 HR AVERAGES March 2017



HYDROGEN SULPHIDE Hourly Averages (H₂S ppb)





HYDROGEN SULPHIDE Instantaneous Maximum (H₂S ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
2	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
3	0.0	0.0	0.0	0.0	S	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	24
4	0.0	0.0	0.0	S	0.3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.3	0.0	24
5	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	P	0.0	C	C	C	C	C	X	0.3	0.3	0.2	0.3	0.3	0.3	0.3	0.3	0.0	0.3	0.1	22
6	0.3	S	0.6	0.2	0.3	0.2	0.3	0.2	0.2	0.2	0.2	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.2	0.2	0.1	0.6	0.2	24
7	S	0.6	0.2	0.1	0.2	0.2	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.0	0.0	0.1	0.0	0.0	0.2	0.0	0.0	0.0	S	0.0	0.6	0.1	24	
8	0.4	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.2	0.0	0.4	0.0	24	
9	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.6	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	S	0.3	0.0	0.6	0.1	24	
10	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.3	0.0	0.0	0.3	0.0	24	
11	0.0	0.2	0.4	0.1	0.0	0.0	0.0	X	X	0.0	0.0	0.0	1.2	0.0	0.0	0.0	0.0	0.0	0.0	S	0.2	0.0	0.0	0.0	0.0	1.2	0.1	22	
12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.3	0.0	0.0	0.0	0.0	0.0	0.0	0.3	0.0	24
13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.2	0.0	0.0	0.0	0.0	0.1	S	0.4	0.2	0.1	0.2	0.1	0.0	0.0	0.4	0.1	24	
14	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.0	0.2	0.1	0.2	0.2	0.3	0.3	S	0.6	0.5	0.2	0.2	0.5	0.4	0.5	0.0	0.6	0.3	24	
15	0.4	0.3	0.2	0.2	0.2	0.2	0.2	0.3	1.5	0.2	0.3	0.7	0.3	0.2	0.2	S	0.7	0.2	0.3	0.3	0.2	0.1	0.2	0.3	0.1	1.5	0.3	24	
16	0.2	0.2	0.2	0.2	0.7	0.4	0.3	0.2	0.5	0.2	0.0	0.0	0.0	0.0	S	0.5	0.1	0.1	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.7	0.2	24	
17	0.0	0.1	0.4	0.4	0.0	0.0	0.0	0.1	0.2	0.0	0.1	S	0.5	0.1	0.1	0.1	0.1	0.0	0.0	0.1	0.1	0.1	0.1	0.2	0.0	0.5	0.1	24	
18	0.6	0.6	0.5	0.5	0.6	0.7	0.6	0.4	0.4	0.4	0.4	0.5	S	0.7	0.4	0.4	0.4	0.3	0.3	0.3	0.3	0.4	1.4	1.1	0.3	1.4	0.5	24	
19	0.6	1.2	1.2	1.2	0.7	0.6	0.2	0.2	0.2	0.3	0.2	S	0.7	0.2	0.2	0.1	0.1	0.1	0.1	0.2	0.0	0.2	0.2	1.0	0.0	1.2	0.4	24	
20	0.6	0.0	0.0	0.2	0.0	0.2	0.3	0.0	0.2	0.2	S	0.4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.3	0.2	0.0	0.6	0.1	24	
21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.2	0.0	0.0	0.0	0.0	0.0	0.1	0.2	0.0	0.0	0.0	0.3	0.2	0.3	0.0	0.3	0.1	24	
22	0.3	0.4	0.4	0.4	0.3	0.2	0.2	0.2	S	0.5	0.3	0.3	0.2	0.3	0.3	0.3	0.2	1.3	0.4	0.3	0.3	0.3	0.3	1.3	0.2	1.3	0.4	24	
23	1.5	1.1	1.5	0.2	0.2	1.3	0.2	S	0.6	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.5	1.1	0.3	0.1	0.2	0.1	0.1	1.5	0.5	24	
24	0.2	0.1	0.1	0.1	0.2	0.1	S	0.6	0.4	0.0	0.1	0.2	0.4	0.4	0.2	0.1	0.1	0.1	0.2	0.2	0.3	0.2	0.1	0.2	0.0	0.6	0.2	24	
25	0.2	0.2	0.2	0.2	0.3	S	0.5	0.2	0.2	0.3	0.4	0.3	0.2	0.2	0.3	0.1	0.1	0.2	0.2	0.2	0.2	0.4	0.6	0.4	0.1	0.6	0.3	24	
26	0.1	0.0	0.1	0.2	S	0.8	0.2	0.2	0.2	0.1	0.1	0.1	0.0	0.0	0.0	0.1	0.1	0.2	0.0	0.2	0.1	0.2	0.2	0.2	0.0	0.8	0.1	24	
27	0.3	0.2	0.2	S	0.5	0.4	0.4	0.2	0.1	0.2	0.2	0.1	0.0	0.2	0.2	0.2	0.3	0.2	0.2	0.2	0.2	0.3	0.2	0.2	0.3	0.0	0.5	0.2	24
28	0.0	0.2	S	0.4	0.2	0.0	0.0	0.2	0.2	0.2	0.2	0.3	0.2	0.1	0.0	0.2	0.2	0.0	0.3	0.2	0.2	0.2	0.3	0.3	0.0	0.4	0.2	24	
29	0.4	S	0.6	0.3	0.2	0.3	0.3	0.2	0.4	0.3	0.2	0.2	0.2	0.3	0.2	0.2	0.3	0.2	0.2	0.3	0.3	0.4	0.3	0.3	0.2	0.6	0.3	24	
30	S	0.5	0.2	0.3	0.2	0.1	0.1	0.1	0.0	0.0	0.1	0.1	0.1	0.1	0.0	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.5	0.1	24	
31	0.3	0.0	0.1	0.1	0.0	0.0	0.0	0.9	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.2	0.1	0.2	0.2	S	0.6	0.0	0.9	0.1	24	
HOURLY MAX	1.5	1.2	1.5	1.2	0.7	1.3	0.6	0.9	1.5	0.5	0.4	0.7	1.2	0.7	0.5	0.5	0.7	1.3	0.5	1.1	0.3	0.5	1.4	1.3					
HOURLY AVG	0.2	0.2	0.3	0.2	0.2	0.2	0.1	0.2	0.2	0.1	0.1	0.1	0.2	0.1	0.1	0.1	0.1	0.2	0.1	0.2	0.1	0.2	0.2	0.3					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	399
MAXIMUM INSTANTANEOUS VALUE:	1.5 ppb @ HOUR(S) 8 ON DAY(S) 15
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	5 hrs
OPERATIONAL TIME:	740 hrs
STANDARD DEVIATION:	0.2

HYDROGEN SULPHIDE Instantaneous Maximum (H₂S ppb)




Wind: LICA MASKWA
 Poll.: LICA MASKWA-H2S[ppb]
 Monthly: 2017/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 7.10% Calm Avg: 0.11 [ppb]

Direction	0.0-0.7	0.7-1.3	1.3-2.0	>2.0	Total
N	10.5	0.0	0.0	0.0	10.5
NE	21.5	0.0	0.0	0.0	21.5
E	17.3	0.0	0.0	0.0	17.3
SE	21.0	0.0	0.0	0.0	21.0
S	11.5	0.0	0.0	0.0	11.5
SW	2.8	0.0	0.0	0.0	2.8
W	3.8	0.0	0.0	0.0	3.8
NW	4.3	0.1	0.0	0.0	4.4
Summary	92.8	0.1	0.0	0.0	92.9

% Icon Classes (ppb) 93 0.0-0.7 0 0.7-1.3 0 1.3-2.0 0 >2.0

LICA MASKWA Poll.: LICA MASKWA-H2S[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 7.10% Calm Poll
Avg: 0.11[ppb]



H2S[ppb] Calibration: LICA MASKWA Monthly: 2017/03 Type: Span



Span Meas Span Ref Span Low Span High

TOTAL HYDROCARBON


TOTAL HYDROCARBONS Hourly Averages (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2.11	2.11	2.11	2.10	2.11	2.11	S	2.12	2.12	2.12	2.12	2.12	2.12	2.11	2.11	2.11	2.11	2.12	2.13	2.13	2.13	2.13	2.14	2.13	2.10	2.14	2.12	24	
2	2.13	2.13	2.12	2.12	2.11	S	2.13	2.17	2.21	2.18	2.28	2.21	2.21	2.28	2.37	2.44	2.42	2.36	2.28	2.31	2.31	2.27	2.25	2.26	2.11	2.44	2.24	24	
3	2.24	2.23	2.20	2.20	S	2.18	2.19	2.19	2.20	2.21	2.24	2.21	2.20	2.19	2.18	2.24	2.32	2.34	2.36	2.37	2.33	2.21	2.13	2.11	2.11	2.37	2.23	24	
4	2.09	2.10	2.17	S	2.17	2.13	2.13	2.12	2.12	2.12	2.13	2.12	2.11	2.10	2.11	2.11	2.11	2.13	2.12	2.11	2.11	2.10	2.09	2.09	2.09	2.17	2.12	24	
5	2.09	2.09	S	2.09	2.10	2.10	2.11	2.11	2.11	C	C	C	Y	Y	Y	C	C	C	C	2.10	2.08	2.09	2.10	2.11	2.08	2.11	2.10	21	
6	2.09	S	2.08	2.08	2.09	2.09	2.09	2.09	2.09	2.09	2.09	2.09	2.09	2.09	2.09	2.09	2.09	2.10	2.12	2.12	2.11	2.11	2.11	2.10	2.08	2.12	2.10	24	
7	S	2.09	2.11	2.10	2.10	2.11	2.11	2.11	2.10	2.09	2.09	2.10	2.09	2.10	2.11	2.11	2.11	2.12	2.12	2.12	2.11	2.12	2.12	S	2.09	2.12	2.11	24	
8	2.13	2.14	2.13	2.13	2.13	2.13	2.14	2.13	2.12	2.12	2.11	2.11	2.11	2.10	2.10	2.10	2.10	2.10	2.11	2.12	2.12	2.13	2.14	S	2.15	2.10	2.15	2.12	24
9	2.16	2.17	2.18	2.18	2.18	2.19	2.20	2.21	2.25	2.20	2.18	2.19	2.18	2.19	2.21	2.24	2.25	2.25	2.25	2.25	2.24	S	2.26	2.26	2.16	2.26	2.21	24	
10	2.25	2.24	2.21	2.20	2.20	2.21	2.21	2.19	2.18	2.18	2.15	2.17	2.15	2.14	2.15	2.16	2.14	2.14	2.14	2.15	S	2.17	2.18	2.18	2.14	2.25	2.18	24	
11	2.18	2.17	2.18	2.17	2.16	2.17	2.16	X	X	2.16	2.16	2.16	2.16	2.15	2.15	2.16	2.15	2.16	2.14	2.13	S	2.14	2.13	2.12	2.12	2.18	2.16	22	
12	2.12	2.14	2.15	2.15	2.15	2.16	2.16	2.15	2.16	2.16	2.16	2.16	2.15	2.15	2.15	2.15	2.15	2.15	2.15	S	2.15	2.21	2.24	2.20	2.19	2.12	2.24	2.16	24
13	2.20	2.24	2.26	2.26	2.22	2.19	2.19	2.20	2.19	2.20	2.19	2.17	2.18	2.18	2.19	2.21	S	2.22	2.22	2.23	2.23	2.23	2.23	2.25	2.17	2.26	2.21	24	
14	2.26	2.27	2.31	2.35	2.38	2.42	2.41	2.43	2.48	2.48	2.46	2.47	2.49	2.45	2.41	2.36	S	2.33	2.30	2.28	2.28	2.25	2.24	2.17	2.17	2.49	2.36	24	
15	2.13	2.13	2.13	2.14	2.17	2.18	2.18	2.21	2.30	2.27	2.29	2.28	2.23	2.15	2.12	S	2.54	2.55	2.43	2.15	2.13	2.09	2.07	2.05	2.05	2.55	2.21	24	
16	2.04	2.03	2.03	2.04	2.08	2.08	2.05	2.04	2.10	2.04	2.04	2.06	2.08	S	2.09	2.11	2.12	2.12	2.12	2.12	2.12	2.12	2.13	2.17	2.03	2.17	2.08	24	
17	2.16	2.17	2.17	2.16	2.16	2.16	2.17	2.17	2.21	2.21	2.27	2.29	2.28	S	2.24	2.19	2.19	2.19	2.19	2.24	2.22	2.24	2.25	2.22	2.16	2.29	2.21	24	
18	2.18	2.20	2.21	2.17	2.11	2.13	2.14	2.12	2.09	2.06	2.06	2.07	S	2.02	2.02	1.98	1.98	1.99	1.99	1.97	1.95	1.94	2.02	1.97	1.94	2.21	2.06	24	
19	1.99	2.01	2.11	2.15	2.09	2.07	2.06	2.06	2.08	2.08	S	2.08	2.06	2.07	2.05	2.06	2.05	2.06	2.07	2.07	2.07	2.09	2.15	1.99	2.15	2.07	24		
20	2.13	2.12	2.13	2.14	2.13	2.15	2.16	2.16	2.17	2.16	S	2.14	2.14	2.14	2.13	2.14	2.13	2.14	2.15	2.15	2.17	2.22	2.22	2.25	2.12	2.25	2.16	24	
21	2.28	2.29	2.29	2.24	2.21	2.22	2.20	2.20	2.29	S	2.28	2.25	2.19	2.17	2.17	2.16	2.17	2.14	2.12	2.12	2.12	2.13	2.14	2.15	2.12	2.29	2.20	24	
22	2.14	2.18	2.15	2.16	2.15	2.12	2.11	2.11	S	2.10	2.11	2.09	2.09	2.08	2.10	2.08	2.08	2.16	2.10	2.09	2.09	2.11	2.14	2.24	2.08	2.24	2.12	24	
23	2.18	2.16	2.17	2.11	2.11	2.21	2.17	S	2.10	2.10	2.09	2.09	2.10	2.09	2.11	2.12	2.10	2.11	2.10	2.11	2.14	2.18	2.20	2.09	2.21	2.13	24		
24	2.19	2.20	2.20	2.23	2.24	2.21	S	2.24	2.32	2.24	2.28	2.42	2.41	2.41	2.31	2.25	2.12	2.07	2.09	2.09	2.14	2.07	2.05	2.05	2.05	2.42	2.21	24	
25	2.05	2.05	2.06	2.05	2.05	S	2.08	2.08	2.07	2.09	2.08	2.08	2.08	2.06	2.14	2.14	2.13	2.11	2.12	2.18	2.20	2.19	2.05	2.20	2.10	2.10	24		
26	2.13	2.08	2.08	2.10	S	2.12	2.11	2.16	2.20	2.22	2.23	2.25	2.27	2.25	2.25	2.18	2.13	2.16	2.19	2.17	2.17	2.21	2.24	2.28	2.08	2.28	2.18	24	
27	2.32	2.23	2.22	S	2.22	2.13	2.13	2.12	2.08	2.09	2.06	2.04	2.06	2.06	2.05	2.09	2.07	2.09	2.13	2.09	2.06	2.09	2.11	2.12	2.04	2.32	2.12	24	
28	2.11	2.10	S	2.06	2.06	2.08	2.09	2.09	2.10	2.12	2.12	2.10	2.09	2.09	2.09	2.07	2.13	2.13	2.14	2.11	2.08	2.07	2.06	2.06	2.14	2.09	24		
29	2.05	S	2.07	2.08	2.10	2.09	2.09	2.09	2.09	2.10	2.12	2.10	2.11	2.16	2.23	2.24	2.18	2.11	2.10	2.13	2.11	2.07	2.10	2.06	2.07	2.05	2.24	2.11	24
30	S	2.04	2.04	2.03	2.02	2.03	2.03	2.03	2.03	2.03	2.02	2.02	2.02	2.02	2.02	2.01	2.01	2.02	2.04	2.04	2.04	2.07	2.09	S	2.01	2.09	2.03	24	
31	2.09	2.12	2.11	2.14	2.19	2.20	2.16	2.23	2.19	2.23	2.23	2.28	2.23	2.14	2.08	2.06	2.09	2.11	2.12	2.11	2.12	2.12	S	2.19	2.06	2.28	2.15	24	
HOURLY MAX	2.32	2.29	2.31	2.35	2.38	2.42	2.41	2.43	2.48	2.48	2.46	2.47	2.49	2.45	2.41	2.44	2.54	2.55	2.43	2.37	2.33	2.27	2.26	2.28					
HOURLY AVG	2.15	2.15	2.15	2.14	2.15	2.15	2.14	2.15	2.16	2.15	2.16	2.17	2.16	2.15	2.15	2.14	2.15	2.16	2.16	2.14	2.14	2.14	2.15	2.15					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	699			
MINIMUM 1-HR AVERAGE:	1.94 ppm	@ HOUR(S)	21	ON DAY(S) 18
MAXIMUM 1-HR AVERAGE:	2.55 ppm	@ HOUR(S)	17	ON DAY(S) 15
MAXIMUM 24-HR AVERAGE:	2.36 ppm			ON DAY(S) 14
				VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	739 hrs	
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	99.3 %	
STANDARD DEVIATION:	0.09	MONTHLY AVERAGE:	2.15 ppm	

TOTAL HYDROCARBONS Hourly Averages (THC ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - March 2017

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY 1	2.25	2.25	2.25	2.25	2.25	2.26	S	2.26	2.28	2.28	2.28	2.28	2.28	2.28	2.26	2.29	2.29	2.31	2.31	2.31	2.32	2.32	2.32	2.32	2.25	2.32	2.28	24	
2	2.32	2.32	2.32	2.32	2.32	S	2.32	2.40	2.53	2.38	2.56	2.47	2.40	2.45	2.60	2.60	2.60	2.53	2.47	2.50	2.47	2.41	2.38	2.38	2.32	2.60	2.44	24	
3	2.35	2.35	2.31	2.29	S	2.29	2.29	2.29	2.29	2.31	2.32	2.31	2.29	2.29	2.26	2.38	2.42	2.44	2.44	2.50	2.50	2.35	2.20	2.20	2.20	2.20	2.50	2.33	24
4	2.20	2.20	2.28	S	2.26	2.22	2.20	2.20	2.20	2.20	2.20	2.20	2.20	2.20	2.20	2.20	2.20	2.20	2.31	2.22	2.20	2.20	2.20	2.20	2.20	2.20	2.31	2.21	24
5	2.20	2.20	S	2.20	2.20	2.20	2.20	2.20	2.20	C	C	C	Y	Y	Y	C	C	C	C	2.23	2.13	2.13	2.14	2.16	2.13	2.23	2.18	21	
6	2.13	S	2.11	2.11	2.10	2.11	2.12	2.12	2.12	2.13	2.13	2.12	2.13	2.12	2.11	2.12	2.13	2.13	2.14	2.14	2.14	2.16	2.15	2.14	2.10	2.16	2.13	24	
7	S	2.14	2.15	2.14	2.17	2.17	2.17	2.20	2.17	2.17	2.17	2.17	2.17	2.17	2.18	2.20	2.17	2.20	2.20	2.21	2.20	2.20	2.23	S	2.14	2.12	2.18	24	
8	2.23	2.23	2.23	2.23	2.21	2.23	2.23	2.23	2.23	2.23	2.21	2.21	2.21	2.23	2.20	2.20	2.20	2.21	2.23	2.23	2.24	2.25	S	2.28	2.20	2.28	2.23	24	
9	2.29	2.31	2.32	2.31	2.32	2.32	2.32	2.44	2.44	2.37	2.32	2.35	2.32	2.34	2.38	2.41	2.38	2.41	2.41	2.41	2.41	S	2.44	2.43	2.29	2.44	2.37	24	
10	2.43	2.41	2.38	2.38	2.36	2.37	2.38	2.34	2.34	2.39	2.33	2.38	2.35	2.29	2.32	2.31	2.29	2.29	2.29	2.28	S	2.32	2.32	2.31	2.28	2.43	2.34	24	
11	2.32	2.31	2.31	2.31	2.31	2.29	2.26	X	X	2.28	2.28	2.29	2.32	2.36	2.32	2.26	2.32	2.23	2.23	S	2.26	2.23	2.23	2.23	2.23	2.36	2.28	22	
12	2.23	2.23	2.25	2.25	2.26	2.26	2.26	2.26	2.29	2.29	2.29	2.29	2.28	2.23	2.27	2.26	2.23	2.25	S	2.26	2.33	2.34	2.29	2.28	2.23	2.34	2.27	24	
13	2.29	2.31	2.33	2.32	2.29	2.25	2.24	2.26	2.25	2.25	2.24	2.23	2.20	2.20	2.20	2.23	2.23	S	2.24	2.24	2.26	2.26	2.28	2.29	2.20	2.33	2.26	24	
14	2.29	2.32	2.36	2.38	2.44	2.45	2.44	2.51	2.52	2.51	2.57	2.51	2.51	2.48	2.57	2.38	S	2.66	2.60	2.29	2.31	2.32	2.28	2.23	2.23	2.66	2.43	24	
15	2.14	2.13	2.13	2.15	2.17	2.17	2.20	2.20	2.58	2.26	2.29	2.33	2.23	2.20	2.10	S	2.54	2.54	2.54	2.18	2.13	2.10	2.08	2.04	2.04	2.58	2.24	24	
16	2.04	2.04	2.02	2.05	2.26	2.11	2.10	2.05	2.29	2.07	2.05	2.07	2.08	2.10	S	2.11	2.14	2.13	2.14	2.15	2.14	2.15	2.17	2.31	2.02	2.31	2.12	24	
17	2.21	2.20	2.21	2.20	2.20	2.20	2.20	2.20	2.26	2.28	2.32	2.32	2.31	S	2.38	2.21	2.20	2.20	2.20	2.23	2.23	2.24	2.26	2.23	2.20	2.38	2.24	24	
18	2.17	2.23	2.23	2.20	2.10	2.18	2.18	2.13	2.10	2.04	2.05	2.08	S	1.98	2.04	1.98	1.98	1.98	1.95	1.95	1.93	1.92	2.12	2.01	1.92	2.23	2.07	24	
19	2.07	2.12	2.18	2.20	2.14	2.15	2.07	2.07	2.10	2.07	2.10	S	2.14	2.08	2.15	2.12	2.11	2.08	2.10	2.10	2.13	2.23	2.15	2.26	2.07	2.26	2.13	24	
20	2.29	2.19	2.20	2.23	2.24	2.28	2.29	2.26	2.31	2.31	S	2.26	2.23	2.24	2.21	2.21	2.20	2.23	2.23	2.23	2.29	2.34	2.32	2.35	2.19	2.35	2.26	24	
21	2.38	2.38	2.38	2.35	2.31	2.31	2.29	2.35	2.38	S	2.35	2.32	2.27	2.21	2.20	2.26	2.23	2.21	2.15	2.15	2.15	2.18	2.20	2.33	2.15	2.38	2.28	24	
22	2.18	2.23	2.21	2.20	2.20	2.17	2.13	2.10	S	2.10	2.13	2.07	2.07	2.07	2.17	2.08	2.08	2.26	2.11	2.10	2.17	2.26	2.21	2.44	2.07	2.44	2.16	24	
23	2.46	2.35	2.66	2.13	2.13	2.49	2.25	S	2.13	2.12	2.10	2.12	2.13	2.11	2.14	2.15	2.15	2.13	2.14	2.13	2.13	2.17	2.20	2.23	2.10	2.66	2.21	24	
24	2.20	2.23	2.24	2.28	2.33	2.29	S	2.35	2.45	2.28	2.44	2.47	2.48	2.46	2.36	2.29	2.17	2.07	2.14	2.13	2.18	2.10	2.04	2.04	2.04	2.48	2.26	24	
25	2.04	2.04	2.04	2.04	2.02	S	2.07	2.08	2.07	2.11	2.10	2.10	2.10	2.13	2.10	2.07	2.13	2.13	2.13	2.13	2.14	2.26	2.20	2.26	2.02	2.26	2.11	24	
26	2.23	2.08	2.10	2.11	S	2.14	2.14	2.20	2.21	2.23	2.24	2.26	2.29	2.28	2.30	2.23	2.14	2.18	2.23	2.21	2.20	2.26	2.26	2.35	2.08	2.35	2.21	24	
27	2.36	2.29	2.27	S	2.41	2.20	2.20	2.20	2.17	2.20	2.12	2.07	2.10	2.12	2.12	2.30	2.15	2.15	2.20	2.17	2.10	2.13	2.13	2.14	2.07	2.41	2.19	24	
28	2.14	2.13	S	2.08	2.10	2.10	2.11	2.10	2.14	2.15	2.14	2.14	2.11	2.10	2.12	2.11	2.10	2.15	2.14	2.15	2.14	2.10	2.08	2.07	2.07	2.15	2.12	24	
29	2.07	S	2.07	2.10	2.10	2.10	2.10	2.10	2.08	2.08	2.10	2.11	2.20	2.23	2.24	2.20	2.12	2.10	2.13	2.11	2.10	2.13	2.08	2.15	2.07	2.24	2.12	24	
30	S	2.04	2.02	2.01	2.01	2.01	2.01	2.01	2.02	2.02	2.02	2.02	2.02	2.02	2.02	2.01	2.01	2.04	2.05	2.07	2.07	2.13	2.13	S	2.01	2.13	2.03	24	
31	2.12	2.17	2.13	2.31	2.29	2.26	2.23	2.72	2.21	2.26	2.26	2.31	2.29	2.18	2.12	2.07	2.10	2.12	2.13	2.13	2.13	2.13	S	2.20	2.07	2.72	2.21	24	
HOURLY MAX	2.46	2.41	2.66	2.38	2.44	2.49	2.44	2.72	2.58	2.51	2.57	2.51	2.51	2.48	2.60	2.60	2.60	2.66	2.60	2.50	2.50	2.41	2.44	2.44					
HOURLY AVG	2.23	2.22	2.23	2.21	2.22	2.23	2.21	2.24	2.25	2.22	2.23	2.24	2.23	2.21	2.23	2.21	2.21	2.23	2.22	2.20	2.20	2.21	2.21	2.24					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	699
MAXIMUM INSTANTANEOUS VALUE:	2.72 ppm @ HOUR(S) 7 ON DAY(S) 31
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	7 hrs
OPERATIONAL TIME:	739 hrs
STANDARD DEVIATION:	0.12

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)




Wind: LICA MASKWA
 Poll.: LICA MASKWA-THC[ppm]
 Monthly: 2017/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 7.15% Calm Avg: 2.17 [ppm]

Direction	0.0-0.9	0.9-1.7	1.7-2.6	>2.6	Total
N	0.0	0.0	10.6	0.0	10.6
NE	0.0	0.0	21.0	0.0	21.0
E	0.0	0.0	17.3	0.0	17.3
SE	0.0	0.0	21.2	0.0	21.2
S	0.0	0.0	11.6	0.0	11.6
SW	0.0	0.0	2.9	0.0	2.9
W	0.0	0.0	3.9	0.0	3.9
NW	0.0	0.0	4.4	0.0	4.4
Summary	0.0	0.0	92.8	0.0	92.8

% Icon	Classes (ppm)	0	0.0-0.9	0	0.9-1.7	93	1.7-2.6	0	>2.6
--------	---------------	---	---------	---	---------	----	---------	---	------

LICA MASKWA Poll.: LICA MASKWA-THC[ppm] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 7.15% Calm Poll Avg: 2.17[ppm]



THC[ppm] Calibration: LICA MASKWA Monthly: 2017/03 Type: Span



■ Span Meas
 — Span Ref
 — Span Low
 — Span High

OXIDES OF NITROGEN




OXIDES OF NITROGEN Hourly Averages (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.0	24
2	0.0	0.0	0.0	0.0	0.0	S	0.6	0.9	0.2	0.6	1.4	1.1	1.0	1.5	3.7	2.4	6.4	0.9	0.4	5.8	4.6	1.8	1.6	2.8	0.0	6.4	1.6	24	
3	0.8	1.5	0.6	0.4	S	1.8	1.0	0.8	0.9	2.2	3.0	2.6	3.5	2.7	1.9	2.3	2.9	3.1	2.8	2.5	3.3	2.5	1.1	1.1	0.4	3.5	2.0	24	
4	0.9	0.8	1.6	S	4.0	1.2	0.3	0.4	0.1	0.2	0.0	0.2	0.5	0.3	0.0	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	4.0	0.5	24	
5	0.0	0.0	S	0.4	0.0	0.0	0.0	0.0	0.0	C	C	C	C	C	C	C	C	2.4	0.2	0.0	0.0	0.0	0.0	0.0	0.0	2.4	0.2	24	
6	0.0	S	1.0	0.4	0.5	0.4	0.3	1.1	1.4	1.2	0.5	0.1	0.0	0.1	0.2	0.1	0.2	0.3	1.5	1.6	1.6	1.6	1.6	1.5	0.0	1.6	0.7	24	
7	S	2.8	2.1	1.6	1.8	2.6	2.8	3.0	2.1	1.2	1.5	1.2	1.2	1.6	1.5	1.6	0.7	1.0	1.8	1.7	0.5	0.7	0.8	S	0.5	3.0	1.6	24	
8	1.7	1.0	1.5	0.6	2.4	0.7	0.6	1.1	0.6	0.0	0.0	0.0	0.8	0.6	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.2	S	2.6	0.0	2.6	0.6	24	
9	2.3	2.3	2.3	2.1	1.8	1.7	3.3	8.8	8.0	3.1	0.1	6.5	10.2	8.8	8.9	7.3	5.9	5.6	4.9	3.5	3.1	S	2.9	2.1	0.1	10.2	4.6	24	
10	1.8	2.9	2.7	1.8	1.3	0.6	0.2	0.0	0.0	1.7	4.9	3.7	0.1	0.0	1.7	3.2	0.0	0.0	0.0	0.0	S	1.3	0.9	0.0	0.0	4.9	1.3	24	
11	0.5	0.6	0.4	0.0	0.2	0.0	0.0	X	X	0.0	0.0	3.9	5.4	4.5	3.2	0.6	0.9	0.0	0.0	S	1.5	0.0	0.0	0.0	0.0	5.4	1.0	22	
12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	5.9	4.2	5.3	3.8	0.7	0.0	1.5	0.4	0.3	S	0.9	2.6	2.8	0.6	0.1	0.0	5.9	1.3	24	
13	1.9	2.5	1.0	1.0	0.5	0.0	0.4	0.5	0.5	0.9	2.2	1.3	0.7	0.9	0.8	0.8	0.4	S	1.9	1.3	1.1	0.8	0.6	0.9	0.0	2.5	1.0	24	
14	0.9	0.9	1.0	1.6	3.5	4.4	4.4	4.1	3.7	3.7	3.4	3.2	4.2	4.6	5.6	3.4	S	6.6	4.0	1.8	1.6	4.2	8.4	8.3	0.9	8.4	3.8	24	
15	1.2	1.1	1.1	1.1	1.7	2.2	2.2	3.8	4.4	3.7	3.7	8.7	4.6	3.4	1.1	S	8.9	7.0	5.5	2.6	1.3	0.3	1.7	1.7	0.3	8.9	3.2	24	
16	1.1	0.5	0.5	0.3	1.5	1.9	1.6	1.1	1.9	0.5	0.2	0.1	0.1	0.2	S	1.6	0.6	0.6	0.5	0.4	0.1	0.0	0.0	0.7	0.0	1.9	0.7	24	
17	0.6	2.2	3.2	0.8	0.8	0.5	0.8	0.9	1.5	2.0	2.5	2.4	1.4	S	3.1	2.4	1.8	1.8	2.0	2.2	1.8	1.8	1.8	1.4	0.5	3.2	1.7	24	
18	1.2	7.5	12.7	11.7	2.4	8.7	9.7	10.3	8.0	2.8	3.5	8.3	S	2.7	6.6	5.4	6.8	2.1	0.6	0.6	0.6	0.6	19.7	11.5	0.6	19.7	6.3	24	
19	10.2	4.8	24.2	14.4	9.2	3.6	0.6	1.0	1.2	0.3	0.2	S	2.0	1.1	0.8	0.7	0.3	0.2	0.4	0.1	0.3	0.5	0.7	19.7	0.1	24.2	4.2	24	
20	6.4	0.5	3.4	5.6	5.1	7.9	11.8	3.8	5.9	5.5	S	5.3	2.4	2.4	1.9	1.3	0.4	0.3	0.0	0.0	0.4	0.4	0.7	0.0	11.8	3.1	24		
21	0.9	1.0	0.6	0.1	0.0	0.0	1.5	0.9	1.7	S	3.3	2.9	2.6	0.9	1.1	3.2	7.0	6.8	1.9	0.4	1.0	3.4	7.6	9.8	0.0	9.8	2.5	24	
22	8.6	11.3	11.0	10.8	6.7	2.4	1.1	0.6	S	2.5	2.2	1.1	1.0	1.1	2.2	1.0	0.6	0.8	1.4	1.5	3.4	7.6	7.6	9.9	0.6	11.3	4.2	24	
23	5.9	5.4	4.0	1.4	1.4	3.1	6.6	S	3.6	2.8	3.5	1.9	4.3	5.0	7.1	7.4	2.6	2.2	1.6	1.6	1.9	1.7	3.3	3.7	1.4	7.4	3.6	24	
24	2.1	1.7	1.3	1.6	2.3	1.6	S	4.4	5.7	2.1	2.5	4.5	2.5	5.5	1.9	1.7	1.2	1.2	1.8	1.7	5.9	1.5	0.6	0.6	0.6	5.9	2.4	24	
25	0.6	0.6	0.6	0.6	0.6	S	2.6	1.3	1.3	4.0	7.6	5.7	5.5	2.3	3.5	1.9	2.4	2.4	2.2	1.3	1.2	1.7	1.6	1.1	0.6	7.6	2.3	24	
26	1.1	1.5	2.5	2.4	S	3.1	3.8	4.8	2.7	2.6	2.5	2.7	2.5	2.7	2.1	1.8	1.2	1.4	1.6	1.6	1.5	1.1	1.2	0.7	0.7	4.8	2.1	24	
27	0.6	0.6	0.5	S	2.9	3.2	2.1	1.3	2.4	8.1	5.4	2.6	4.0	5.8	4.4	4.0	5.8	5.1	9.2	8.3	3.4	0.8	1.0	2.3	0.5	9.2	3.6	24	
28	1.2	0.9	S	2.9	1.8	1.4	1.7	2.6	1.9	3.0	3.0	1.8	1.8	1.4	1.6	2.2	2.9	1.7	1.8	1.8	1.6	1.7	1.6	1.5	0.9	3.0	1.9	24	
29	1.2	S	3.6	2.5	2.1	2.0	3.4	2.8	2.1	2.3	2.4	2.8	3.0	3.6	4.5	4.1	3.2	2.9	3.0	2.5	1.4	3.9	5.6	2.4	1.2	5.6	2.9	24	
30	S	3.3	1.7	1.2	0.7	0.6	0.6	0.6	0.6	0.6	0.6	0.6	0.6	0.5	0.5	0.5	0.6	0.6	0.6	0.4	0.2	0.1	0.0	S	0.0	3.3	0.7	24	
31	2.7	1.1	0.6	0.6	0.4	0.3	0.7	3.8	3.9	3.8	4.3	4.3	2.8	2.0	1.2	1.7	2.1	1.7	2.4	3.5	1.8	S	5.2	0.3	5.2	2.2	24		
HOURLY MAX	10.2	11.3	24.2	14.4	9.2	8.7	11.8	10.3	8.0	8.1	7.6	8.7	10.2	8.8	8.9	7.4	8.9	7.0	9.2	8.3	5.9	7.6	19.7	19.7					
HOURLY AVG	1.9	2.0	3.0	2.3	1.9	1.9	2.2	2.1	2.3	2.3	2.3	2.9	2.6	2.3	2.5	2.2	2.3	2.0	1.8	1.6	1.6	1.5	2.5	3.2					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	605		
MINIMUM 1-HR AVERAGE:	0.0 ppb	@ HOUR(S)	0 ON DAY(S) 1
MAXIMUM 1-HR AVERAGE:	24.2 ppb	@ HOUR(S)	2 ON DAY(S) 19
MAXIMUM 24-HR AVERAGE:	6.3 ppb		ON DAY(S) 18
			VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	742 hrs
MONTHLY CALIBRATION TIME:	8 hrs	AMD OPERATION UPTIME:	99.7 %
STANDARD DEVIATION:	2.7	MONTHLY AVERAGE:	2.2 ppb

OXIDES OF NITROGEN Hourly Averages (NO_x ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - March 2017

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	0.6	0.0	0.0	0.1	0.0	0.0	S	3.6	1.2	0.6	0.6	0.6	0.6	0.6	0.6	0.6	0.6	0.0	0.6	0.6	0.0	0.0	0.6	0.0	0.0	0.0	3.6	0.5	24
2	0.6	1.2	1.2	0.6	0.6	S	3.6	6.5	2.4	2.4	4.7	4.7	3.0	4.7	8.2	7.1	11.8	5.2	3.0	11.2	11.2	4.1	4.1	6.5	0.6	11.8	4.7	24	
3	3.6	4.1	3.6	2.4	S	5.9	3.0	3.0	3.0	4.7	5.3	4.7	5.3	4.7	4.1	4.1	5.3	4.7	4.1	4.7	4.7	5.9	2.4	2.4	2.4	2.4	5.9	4.2	24
4	2.4	2.4	3.5	S	9.4	2.4	1.8	2.4	1.8	1.8	1.2	1.8	1.8	1.8	1.2	1.8	1.8	1.2	0.6	0.6	0.0	0.0	0.6	0.0	0.0	0.0	9.4	1.8	24
5	0.0	0.0	S	3.6	0.6	0.6	1.2	1.2	0.6	C	C	C	C	C	C	C	C	C	0.6	0.1	0.1	0.0	0.1	0.1	0.1	0.0	3.6	0.6	24
6	0.0	S	3.0	0.6	1.2	0.6	1.3	3.0	2.4	1.8	1.8	1.2	0.0	1.2	0.6	0.6	0.6	1.2	1.8	1.8	1.8	1.8	1.8	1.8	1.8	0.0	3.0	1.4	24
7	S	4.7	2.4	2.4	3.0	3.0	3.6	3.6	3.0	1.8	1.8	1.8	2.4	2.4	1.8	2.4	1.2	1.8	2.4	2.4	1.2	1.2	1.2	S	1.2	4.7	2.3	24	
8	3.6	1.8	2.4	1.3	22.3	2.4	3.0	3.0	3.0	0.0	0.0	0.6	1.8	3.0	1.2	0.6	0.6	0.6	0.6	0.6	0.6	1.8	S	4.7	0.0	22.3	2.6	24	
9	3.0	3.0	3.0	3.0	2.4	2.4	10.6	16.4	14.7	10.0	1.8	12.4	16.4	17.5	11.2	19.4	7.1	7.1	6.5	5.3	4.1	S	5.3	3.6	1.8	19.4	8.1	24	
10	3.6	4.1	4.1	3.0	2.4	1.8	1.2	0.6	0.6	5.3	8.2	8.8	2.4	2.4	4.7	6.5	4.7	0.1	0.0	0.0	S	4.7	3.0	0.6	0.0	8.8	3.2	24	
11	3.0	3.0	1.8	0.6	1.3	0.6	0.6	X	X	0.6	3.0	7.1	9.4	7.6	6.4	3.6	4.1	1.8	0.6	S	5.3	0.6	0.6	0.6	0.6	9.4	3.0	22	
12	0.6	0.0	0.6	0.0	0.1	0.1	0.0	0.0	4.1	10.0	6.4	9.4	8.2	3.0	2.4	4.1	2.4	2.4	S	3.6	5.9	4.7	3.0	1.2	0.0	10.0	3.1	24	
13	3.6	4.1	1.8	1.8	1.2	1.2	1.3	1.8	1.2	2.4	3.6	3.0	1.8	1.8	1.8	1.8	1.8	S	4.7	2.4	2.4	1.8	1.8	1.8	1.2	4.7	2.2	24	
14	1.8	1.8	2.4	2.4	5.9	5.9	6.5	7.1	5.9	5.3	5.3	4.7	7.1	7.1	11.2	5.3	S	11.2	10.0	10.0	3.0	2.4	12.4	13.0	13.0	1.8	13.0	6.5	24
15	2.4	1.8	1.8	1.8	3.0	3.0	3.5	5.9	6.4	5.3	5.9	20.6	8.8	7.1	3.0	S	11.2	8.2	7.1	3.5	2.4	0.6	2.4	5.9	0.6	20.6	5.3	24	
16	1.8	1.3	1.2	1.2	3.0	4.1	4.1	1.2	4.7	1.2	0.6	0.6	0.6	0.6	S	3.5	0.6	0.6	0.6	0.6	0.6	0.0	0.0	2.4	0.0	4.7	1.5	24	
17	1.8	8.2	8.2	4.1	2.4	0.6	1.2	1.3	1.8	3.5	3.0	3.5	1.8	S	5.3	3.5	2.4	1.8	3.0	2.4	1.8	1.8	1.8	1.8	0.6	8.2	2.9	24	
18	1.8	10.6	14.1	13.5	5.9	15.2	14.6	14.2	12.4	4.7	7.6	12.4	S	4.1	11.2	8.2	11.8	4.7	0.6	0.6	0.6	0.6	47.6	42.2	0.6	47.6	11.3	24	
19	25.8	22.9	33.4	56.3	14.2	8.2	1.3	1.8	1.8	1.2	0.6	S	4.1	1.2	1.8	1.8	0.6	1.2	1.2	0.0	0.6	4.1	4.1	33.5	0.0	56.3	9.6	24	
20	26.4	3.0	10.6	17.0	30.5	30.5	20.0	8.2	33.4	25.9	S	11.2	4.1	4.7	3.5	3.0	1.2	1.2	0.0	0.6	0.6	0.6	0.6	1.3	0.0	33.4	10.4	24	
21	1.2	1.3	1.2	0.6	0.6	0.6	4.7	3.0	2.4	S	5.8	5.9	1.2	1.8	10.0	11.2	11.8	4.1	0.6	1.8	7.1	11.8	13.5	0.6	13.5	4.7	24		
22	13.0	14.1	14.7	14.7	11.2	4.1	1.8	1.2	S	5.9	4.1	1.3	1.3	1.2	5.3	3.5	0.6	1.2	1.8	1.8	7.1	8.8	8.2	16.4	0.6	16.4	6.2	24	
23	15.2	17.0	6.4	1.8	1.8	4.7	44.5	S	5.9	4.1	4.1	3.5	5.9	5.9	8.2	9.4	3.5	2.4	2.4	2.4	2.4	2.4	1.8	4.7	4.7	1.8	44.5	7.1	24
24	3.0	1.8	1.8	2.4	3.0	3.0	S	7.0	7.7	3.0	4.1	7.1	6.4	7.6	3.0	1.3	1.3	4.7	3.5	8.2	3.0	0.6	0.6	0.6	0.6	8.2	3.8	24	
25	0.6	0.6	0.6	0.6	0.6	S	4.7	1.8	1.8	8.2	10.0	7.7	8.2	5.3	9.4	4.1	3.0	2.4	3.0	1.8	1.3	2.4	1.8	1.3	0.6	10.0	3.5	24	
26	1.2	2.4	3.0	3.0	S	5.9	10.6	8.2	3.0	3.5	3.0	4.1	3.0	14.6	2.4	2.4	1.8	1.8	1.8	1.8	1.8	1.2	1.2	1.2	1.2	14.6	3.6	24	
27	0.6	0.6	0.6	S	5.3	4.1	3.5	1.8	8.2	12.9	10.0	4.1	7.1	9.4	11.2	7.1	14.7	10.0	13.5	11.8	10.0	1.2	1.3	3.0	0.6	14.7	6.6	24	
28	1.8	1.3	S	5.2	1.8	1.8	1.8	3.5	2.4	3.5	3.5	2.4	1.8	1.8	1.8	4.7	4.7	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.3	5.2	2.5	24	
29	1.8	S	5.3	3.0	2.4	2.4	7.6	3.5	3.0	3.0	3.0	3.5	3.5	4.1	5.3	4.7	4.1	3.0	3.0	3.0	2.4	9.4	9.4	5.9	1.8	9.4	4.2	24	
30	S	5.9	2.4	1.3	1.2	0.6	0.6	0.6	0.6	0.6	0.6	0.6	0.6	0.6	0.6	0.6	0.6	1.2	0.6	0.6	0.6	0.6	0.0	S	0.0	5.9	1.0	24	
31	5.3	1.8	0.6	0.6	0.6	0.6	0.6	2.4	4.1	4.1	4.1	4.7	4.7	3.5	3.0	1.3	2.4	2.4	1.8	3.5	4.7	3.0	S	7.1	0.6	7.1	2.9	24	
HOURLY MAX	26.4	22.9	33.4	56.3	30.5	30.5	44.5	16.4	33.4	25.9	10.0	20.6	16.4	17.5	11.2	19.4	14.7	11.8	13.5	11.8	11.2	12.4	47.6	42.2					
HOURLY AVG	4.5	4.3	4.7	5.1	4.8	4.0	5.6	4.1	4.9	4.7	3.9	5.3	4.4	4.5	4.6	4.4	4.1	3.2	2.9	2.6	2.9	2.9	4.6	6.2					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	671
MAXIMUM INSTANTANEOUS VALUE:	56.3 ppb @ HOUR(S) 3 ON DAY(S) 19
VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	9 hrs
STANDARD DEVIATION:	5.7
OPERATIONAL TIME:	742 hrs

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)



Wind: LICA MASKWA
 Poll.: LICA MASKWA-NOX[ppb]
 Monthly: 2017/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 7.13% Calm Avg: 2.54 [ppb]

Direction	0.0-8.1	8.1-16.2	16.2-24.3	>24.3	Total
N	10.6	0.0	0.0	0.0	10.6
NE	21.3	0.0	0.0	0.0	21.3
E	15.3	2.0	0.0	0.0	17.3
SE	20.8	0.3	0.0	0.0	21.1
S	11.4	0.1	0.0	0.0	11.6
SW	2.6	0.3	0.0	0.0	2.9
W	3.7	0.1	0.0	0.0	3.9
NW	3.1	0.9	0.4	0.0	4.4
Summary	88.7	3.7	0.4	0.0	92.9

% Icon Classes (ppb)


89 0.0-8.1

4 8.1-16.2

0 16.2-24.3

0 >24.3

LICA MASKWA Poll.: LICA MASKWA-NOX[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 7.13% Calm Poll
Avg: 2.54[ppb]



NOX[ppb] Calibration: LICA MASKWA Monthly: 2017/03 Type: Span



■ Span Meas
 — Span Ref
 — Span Low
 — Span High

NITRIC OXIDES


NITRIC OXIDE Hourly Averages (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
2	0.0	0.0	0.0	0.0	0.0	0.0	S	0.1	0.4	0.2	0.6	1.1	1.1	1.0	1.3	2.0	1.3	1.9	0.1	0.0	0.2	0.2	0.0	0.0	0.0	0.0	2.0	0.5	24
3	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.3	0.8	1.3	1.2	1.7	1.0	0.7	0.6	0.7	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.7	0.4	24	
4	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.2	0.6	0.3	0.0	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.6	0.1	24	
5	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	C	C	C	C	C	C	C	C	C	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
6	0.0	S	0.1	0.2	0.1	0.2	0.0	0.4	0.6	0.6	0.4	0.1	0.0	0.1	0.2	0.1	0.2	0.3	0.1	0.2	0.2	0.2	0.1	0.3	0.0	0.6	0.2	24	
7	S	0.3	0.3	0.1	0.4	0.3	0.2	0.5	0.6	0.5	0.7	0.7	0.8	1.0	1.0	0.7	0.4	0.4	0.2	0.4	0.0	0.0	0.3	S	0.0	1.0	0.4	24	
8	0.2	0.2	0.3	0.2	0.7	0.3	0.3	0.2	0.2	0.0	0.0	0.0	0.4	0.4	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.7	0.2	24	
9	0.0	0.0	0.0	0.0	0.0	0.0	0.1	1.8	3.0	1.7	0.1	4.2	6.4	5.4	5.1	3.6	2.4	1.1	0.0	0.0	0.0	S	0.0	0.0	0.0	6.4	1.5	24	
10	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.0	2.9	2.2	0.1	0.0	0.3	0.3	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	2.9	0.3	24	
11	0.0	0.0	0.0	0.0	0.0	0.0	0.0	X	X	0.0	0.0	1.5	2.5	1.8	1.0	0.2	0.2	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	2.5	0.3	22	
12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	2.4	2.0	2.8	2.2	0.7	0.0	1.0	0.4	0.3	S	0.3	0.3	0.3	0.2	0.1	0.0	2.8	0.6	24	
13	0.2	0.3	0.1	0.3	0.2	0.0	0.4	0.6	0.5	0.6	1.0	0.8	0.6	0.7	0.6	0.6	0.4	S	0.2	0.2	0.1	0.1	0.1	0.4	0.0	1.0	0.4	24	
14	0.2	0.1	0.2	0.0	0.0	0.2	0.4	0.8	1.2	1.5	1.5	1.2	1.6	1.5	1.8	0.6	S	0.4	0.3	0.0	0.0	0.2	0.1	0.4	0.0	1.8	0.6	24	
15	0.0	0.0	0.0	0.0	0.2	0.0	0.0	0.5	1.1	1.2	1.5	3.8	1.8	1.3	0.5	S	1.1	0.6	0.3	0.0	0.0	0.0	0.0	0.0	0.0	3.8	0.6	24	
16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.5	0.0	24	
17	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.3	0.5	0.8	0.9	0.6	S	0.6	0.6	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.9	0.2	24	
18	0.0	0.1	0.6	0.6	0.0	0.6	0.6	1.4	1.9	0.8	1.2	3.2	S	0.5	2.3	1.8	1.8	0.3	0.0	0.0	0.0	0.0	7.4	3.8	0.0	7.4	1.3	24	
19	1.8	1.0	9.1	4.0	1.3	0.1	0.1	0.4	0.7	0.3	0.2	S	0.0	0.4	0.1	0.2	0.0	0.0	0.1	0.0	0.0	0.1	0.1	5.9	0.0	9.1	1.1	24	
20	1.7	0.2	0.7	1.3	1.1	2.2	2.8	1.3	2.4	2.5	S	2.0	1.0	1.1	0.9	0.7	0.3	0.3	0.0	0.0	0.2	0.3	0.2	0.5	0.0	2.8	1.0	24	
21	0.2	0.1	0.1	0.1	0.0	0.0	0.6	0.6	0.6	S	1.2	1.1	1.0	0.5	0.6	1.0	1.2	0.9	0.1	0.0	0.1	0.3	0.4	0.8	0.0	1.2	0.5	24	
22	0.5	0.8	0.7	0.9	0.7	0.3	0.0	0.0	S	0.3	0.4	0.1	0.1	0.0	0.3	0.2	0.0	0.0	0.0	0.0	0.0	0.2	0.3	1.3	0.0	1.3	0.3	24	
23	0.9	0.5	0.3	0.1	0.0	0.0	1.9	S	0.6	0.7	1.2	0.7	1.6	1.7	2.2	2.2	0.6	0.3	0.2	0.0	0.0	0.0	0.0	0.0	0.0	2.2	0.7	24	
24	0.0	0.0	0.0	0.0	0.0	0.0	S	0.6	1.4	0.6	0.8	1.5	0.8	2.1	0.5	0.3	0.0	0.0	0.2	0.2	0.6	0.3	0.1	0.0	0.0	2.1	0.4	24	
25	0.0	0.0	0.0	0.0	0.0	S	0.1	0.0	0.4	1.1	2.5	2.0	1.9	0.9	1.4	0.6	0.6	0.6	0.1	0.0	0.0	0.0	0.1	0.0	0.0	2.5	0.5	24	
26	0.1	0.0	0.0	0.1	S	0.0	0.7	1.0	0.6	0.8	0.6	0.8	0.7	0.8	0.6	0.4	0.3	0.1	0.0	0.0	0.0	0.1	0.0	0.0	0.0	1.0	0.3	24	
27	0.0	0.0	0.0	S	0.1	0.0	0.2	0.5	0.6	2.8	1.9	0.9	1.4	1.9	1.4	1.2	1.5	0.8	0.6	0.4	0.2	0.1	0.2	0.2	0.0	2.8	0.7	24	
28	0.0	0.1	S	0.0	0.0	0.1	0.1	0.5	0.3	0.7	0.7	0.6	0.6	0.6	0.7	0.6	0.3	0.1	0.1	0.0	0.0	0.0	0.0	0.1	0.0	0.7	0.3	24	
29	0.0	S	0.1	0.0	0.0	0.0	0.4	0.2	0.2	0.5	0.6	0.6	1.0	1.2	1.4	1.3	0.7	0.5	0.2	0.0	0.0	0.2	0.5	0.2	0.0	1.4	0.4	24	
30	S	0.1	0.0	0.0	0.0	0.3	0.0	0.0	0.1	0.0	0.1	0.2	0.2	0.0	0.0	0.2	0.3	0.0	0.0	0.0	0.0	0.1	0.0	S	0.0	0.3	0.1	24	
31	0.0	0.0	0.1	0.0	0.0	0.0	0.1	0.3	1.2	1.7	1.5	1.5	1.3	0.7	0.5	0.2	0.4	0.4	0.3	0.0	0.0	0.0	S	0.2	0.0	1.7	0.5	24	
HOURLY MAX	1.8	1.0	9.1	4.0	1.3	2.2	2.8	1.8	3.0	2.8	2.9	4.2	6.4	5.4	5.1	3.6	2.4	1.1	0.6	0.4	0.6	0.3	7.4	5.9					
HOURLY AVG	0.2	0.1	0.4	0.3	0.2	0.2	0.3	0.4	0.7	0.8	0.9	1.2	1.1	1.0	0.9	0.7	0.6	0.3	0.1	0.1	0.1	0.1	0.3	0.5					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	408			
MINIMUM 1-HR AVERAGE:	0.0 ppb	@ HOUR(S)	0	ON DAY(S) 1
MAXIMUM 1-HR AVERAGE:	9.1 ppb	@ HOUR(S)	2	ON DAY(S) 19
MAXIMUM 24-HR AVERAGE:	1.5 ppb			ON DAY(S) 9
				VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	742 hrs	
MONTHLY CALIBRATION TIME:	8 hrs	AMD OPERATION UPTIME:	99.7 %	
STANDARD DEVIATION:	0.9	MONTHLY AVERAGE:	0.5 ppb	

NITRIC OXIDE Hourly Averages (NO ppb)





NITRIC OXIDE Instantaneous Maximum (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.			
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.				
DAY																															
1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.0	0.0	24
2	0.0	0.0	0.0	0.0	0.0	0.0	S	0.5	1.7	0.5	1.1	2.3	2.3	1.1	1.7	2.8	2.3	2.8	0.5	0.0	0.5	0.5	0.0	0.0	0.0	0.0	0.0	0.0	2.8	0.9	24
3	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.5	1.1	1.7	1.7	1.7	1.1	1.1	1.1	1.1	0.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.7	0.5	24	
4	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.2	24	
5	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	C	C	C	C	C	C	C	C	C	0.0	0.5	0.0	0.5	0.5	0.0	0.0	0.0	0.5	0.1	24		
6	0.0	S	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	1.1	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.0	1.1	0.5	24	
7	S	0.5	0.5	0.5	0.5	0.5	0.5	0.5	1.1	0.5	1.1	1.1	1.1	1.1	1.1	1.1	1.1	0.5	0.5	0.5	0.5	0.5	0.5	0.5	S	0.5	1.1	0.7	24		
8	0.5	0.5	0.5	0.5	11.0	0.5	1.1	0.5	0.5	0.5	0.5	0.5	1.1	1.7	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	S	0.5	0.5	11.0	1.1	24		
9	0.5	0.5	0.5	0.5	0.5	0.5	2.3	5.1	5.7	3.4	1.7	8.1	11.1	12.8	6.3	8.7	3.4	2.3	0.5	0.5	0.5	S	0.5	0.5	0.5	0.5	12.8	3.3	24		
10	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	3.4	5.1	5.1	1.7	1.1	1.1	1.1	1.1	0.5	0.5	0.5	0.5	S	0.5	0.5	0.5	0.5	5.1	1.2	24		
11	0.5	0.5	0.5	0.5	0.5	0.5	0.5	X	X	0.5	1.7	3.4	4.6	2.8	2.3	1.1	1.7	0.5	0.5	S	0.5	0.5	0.5	0.5	0.5	0.5	4.6	1.2	22		
12	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	1.7	3.4	2.8	4.6	3.9	1.7	1.1	1.7	1.1	0.5	S	0.5	0.5	0.5	0.5	0.5	0.5	0.5	4.6	1.3	24		
13	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	1.1	1.1	1.1	1.1	1.1	0.5	0.5	0.5	S	0.5	0.5	0.5	0.5	0.5	0.5	0.5	1.1	0.6	24		
14	0.5	0.5	0.5	0.5	0.5	0.0	0.5	0.5	1.7	1.7	1.7	1.7	2.3	1.7	3.4	1.1	S	0.5	0.5	0.0	0.0	0.5	0.5	0.5	0.5	0.0	3.4	1.0	24		
15	0.0	0.5	0.0	0.0	0.5	0.0	0.5	1.1	1.7	1.7	2.3	9.3	3.4	2.3	1.1	S	1.7	0.5	0.5	0.0	0.0	0.0	0.0	0.5	0.0	0.0	9.3	1.2	24		
16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.1	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.5	0.5	0.5	0.0	0.5	0.5	0.0	1.1	0.1	24		
17	0.5	0.5	0.5	0.5	0.5	0.0	0.0	0.5	0.5	0.5	1.1	1.1	0.5	1.1	0.5	S	0.5	0.5	0.5	0.5	0.5	0.0	0.0	0.0	0.0	0.0	1.1	0.4	24		
18	0.0	0.5	0.5	0.5	0.0	1.1	1.1	2.3	3.3	1.1	2.8	4.6	S	0.5	3.9	2.3	3.4	0.5	0.0	0.0	0.0	0.0	22.8	19.2	0.0	22.8	3.1	24			
19	7.5	7.5	14.5	28.0	3.4	0.5	0.5	1.1	1.1	0.5	S	0.5	0.5	0.5	0.5	0.5	0.5	0.0	0.5	0.5	0.0	0.5	0.5	11.7	0.0	28.0	3.6	24			
20	8.6	0.5	2.3	3.4	9.3	8.7	5.7	2.8	15.6	10.5	S	3.9	1.7	1.7	1.7	1.1	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	15.6	3.5	24			
21	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.5	1.1	S	1.1	2.3	2.3	0.5	0.5	2.3	2.8	1.7	0.5	0.0	0.5	0.5	0.5	1.1	0.0	2.8	0.9	24			
22	0.5	1.1	1.1	1.1	1.1	0.5	0.5	0.5	S	0.5	0.5	0.5	0.5	0.5	0.5	0.0	0.0	0.0	0.0	0.0	0.5	0.5	0.5	2.8	0.0	2.8	0.6	24			
23	2.8	2.3	0.5	0.5	0.0	0.0	26.2	S	0.5	1.1	1.1	1.6	1.7	2.3	2.3	2.8	0.5	0.5	0.5	0.5	0.5	0.0	0.5	0.5	0.0	26.2	2.1	24			
24	0.0	0.0	0.0	0.0	0.0	0.5	S	0.5	2.3	1.1	1.7	2.3	2.3	2.8	1.1	0.5	0.0	0.0	0.5	0.5	0.5	0.5	0.5	0.5	0.0	2.8	0.8	24			
25	0.0	0.0	0.0	0.0	0.0	S	0.5	0.0	0.5	2.3	3.4	2.3	2.8	1.7	3.4	1.1	0.5	0.5	0.5	0.5	0.0	0.0	0.5	0.5	0.0	3.4	0.9	24			
26	0.5	0.0	0.0	0.5	S	0.5	3.4	1.7	0.5	1.7	0.5	1.7	1.1	8.6	0.5	0.5	0.5	0.5	0.0	0.0	0.0	0.5	0.5	0.0	0.0	8.6	1.0	24			
27	0.0	0.0	0.0	S	0.5	0.0	0.5	0.5	2.3	4.5	3.4	1.1	2.3	3.4	3.4	2.3	3.4	1.7	0.5	0.5	0.5	0.5	0.5	0.5	0.0	4.5	1.4	24			
28	0.5	0.5	S	0.0	0.0	0.5	0.5	0.5	0.5	1.1	1.1	0.5	0.5	0.5	0.5	1.1	1.1	0.5	0.5	0.0	0.5	0.5	0.5	0.5	0.0	1.1	0.5	24			
29	0.5	S	0.5	0.0	0.0	0.5	1.7	0.5	0.5	0.5	0.5	1.1	1.1	1.7	1.7	1.1	0.5	0.5	0.0	0.5	0.5	0.5	0.5	0.5	0.0	1.7	0.7	24			
30	S	0.5	0.5	0.0	0.0	0.5	0.5	0.0	0.5	0.0	0.5	0.5	0.5	0.0	0.5	0.0	0.5	0.5	0.0	0.5	0.5	0.5	0.5	0.0	S	0.0	0.5	0.3	24		
31	0.5	0.0	0.5	0.5	0.0	0.5	0.5	1.1	1.1	1.7	1.7	1.7	1.7	1.1	0.5	0.5	0.5	0.5	0.5	0.0	0.0	0.0	S	0.5	0.0	1.7	0.7	24			
HOURLY MAX	8.6	7.5	14.5	28.0	11.0	8.7	26.2	5.1	15.6	10.5	5.1	9.3	11.1	12.8	6.3	8.7	3.4	2.3	0.5	0.5	0.5	0.5	22.8	19.2							
HOURLY AVG	0.9	0.6	0.9	1.4	1.0	0.6	1.7	0.9	1.6	1.6	1.5	2.2	1.8	1.9	1.5	1.3	1.1	0.5	0.4	0.3	0.3	0.3	1.1	1.5							


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	531
MAXIMUM INSTANTANEOUS VALUE:	28.0 ppb @ HOUR(S) 3 ON DAY(S) 19
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	9 hrs
STANDARD DEVIATION:	2.5
OPERATIONAL TIME:	742 hrs

NITRIC OXIDE Instantaneous Maximum (NO ppb)



Wind: LICA MASKWA
 Poll.: LICA MASKWA-NO[ppb]
 Monthly: 2017/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 7.13% Calm Avg: 0.53 [ppb]

Direction	0.0-3.1	3.1-6.1	6.1-9.2	>9.2	Total
N	10.6	0.0	0.0	0.0	10.6
NE	21.3	0.0	0.0	0.0	21.3
E	17.1	0.1	0.0	0.0	17.3
SE	21.1	0.0	0.0	0.0	21.1
S	11.6	0.0	0.0	0.0	11.6
SW	2.4	0.4	0.0	0.0	2.9
W	3.9	0.0	0.0	0.0	3.9
NW	3.6	0.6	0.3	0.0	4.4
Summary	91.5	1.1	0.3	0.0	92.9

% Icon Classes (ppb)


91 0.0-3.1

1 3.1-6.1

0 6.1-9.2

0 >9.2

LICA MASKWA Poll.: LICA MASKWA-NO[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 7.13% Calm Poll
Avg: 0.53[ppb]



NITROGEN DIOXIDE

NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.0	24
2	0.0	0.0	0.0	0.0	0.0	S	0.5	0.5	0.0	0.0	0.3	0.0	0.0	0.2	1.7	1.1	4.5	0.8	0.4	5.6	4.4	1.7	1.6	2.8	0.0	5.6	1.1	24	
3	0.8	1.5	0.6	0.4	S	1.8	1.0	0.8	0.6	1.4	1.7	1.4	1.8	1.7	1.2	1.7	2.3	3.0	2.8	2.5	3.3	2.5	1.1	1.1	0.4	3.3	1.6	24	
4	0.9	0.8	1.6	S	4.0	1.2	0.3	0.4	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	4.0	0.4	24	
5	0.0	0.0	S	0.4	0.0	0.0	0.0	0.0	0.0	C	C	C	C	C	C	C	C	2.4	0.2	0.0	0.0	0.0	0.0	0.0	0.0	2.4	0.2	24	
6	0.0	S	0.8	0.2	0.4	0.2	0.3	0.7	0.9	0.6	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.4	1.5	1.4	1.5	1.5	1.2	0.0	1.5	0.6	24	
7	S	2.5	1.8	1.5	1.4	2.3	2.6	2.5	1.5	0.7	0.8	0.6	0.4	0.6	0.5	0.9	0.3	0.6	1.5	1.3	0.5	0.7	0.6	S	0.3	2.6	1.2	24	
8	1.4	0.8	1.2	0.4	1.7	0.4	0.4	0.8	0.3	0.0	0.0	0.0	0.4	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.2	S	2.6	0.0	2.6	0.5	24	
9	2.2	2.3	2.3	2.1	1.7	1.7	3.2	7.0	4.9	1.4	0.0	2.3	3.7	3.5	3.8	3.6	3.5	4.5	4.9	3.5	3.1	S	2.9	2.1	0.0	7.0	3.1	24	
10	1.8	2.9	2.7	1.8	1.3	0.6	0.2	0.0	0.0	0.6	2.0	1.5	0.0	0.0	1.3	2.9	0.0	0.0	0.0	0.0	S	1.3	0.9	0.0	0.0	2.9	0.9	24	
11	0.5	0.6	0.4	0.0	0.2	0.0	0.0	X	X	0.0	0.0	2.3	2.9	2.7	2.2	0.4	0.7	0.0	0.0	S	1.5	0.0	0.0	0.0	0.0	2.9	0.7	22	
12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	3.5	2.2	2.5	1.6	0.0	0.0	0.6	0.0	0.0	S	0.7	2.3	2.5	0.4	0.0	0.0	3.5	0.7	24	
13	1.8	2.2	0.8	0.6	0.3	0.0	0.0	0.0	0.0	0.3	1.2	0.5	0.2	0.2	0.2	0.2	0.0	S	1.7	1.1	1.0	0.6	0.5	0.6	0.0	2.2	0.6	24	
14	0.7	0.9	0.8	1.5	3.5	4.2	4.0	3.3	2.5	2.2	1.9	2.0	2.7	3.1	3.8	2.8	S	6.2	3.6	1.8	1.6	4.1	8.3	7.9	0.7	8.3	3.2	24	
15	1.2	1.0	1.1	1.1	1.5	2.1	2.2	3.2	3.2	2.5	2.2	4.9	2.8	2.2	0.6	S	7.8	6.4	5.2	2.6	1.3	0.3	1.7	1.6	0.3	7.8	2.6	24	
16	1.1	0.5	0.5	0.3	1.5	1.9	1.6	1.1	1.3	0.5	0.2	0.1	0.1	0.2	S	1.6	0.6	0.6	0.5	0.3	0.1	0.0	0.0	0.6	0.0	1.9	0.7	24	
17	0.5	2.0	3.2	0.8	0.8	0.5	0.8	0.9	1.2	1.4	1.6	1.5	0.8	S	2.6	1.8	1.6	1.7	2.0	2.2	1.8	1.8	1.8	1.4	0.5	3.2	1.5	24	
18	1.2	7.4	12.1	11.1	2.4	8.2	9.1	8.9	6.0	1.9	2.3	5.2	S	2.1	4.2	3.6	5.1	1.8	0.6	0.6	0.6	0.6	12.3	7.7	0.6	12.3	5.0	24	
19	8.4	3.8	15.1	10.4	7.9	3.5	0.5	0.5	0.6	0.1	0.0	S	1.9	0.7	0.6	0.5	0.3	0.2	0.4	0.1	0.3	0.4	0.7	13.8	0.0	15.1	3.1	24	
20	4.7	0.3	2.7	4.3	4.0	5.8	8.9	2.5	3.5	3.0	S	3.3	1.3	1.3	1.0	0.5	0.1	0.0	0.0	0.0	0.3	0.1	0.2	0.2	0.0	8.9	2.1	24	
21	0.7	0.9	0.5	0.0	0.0	0.0	0.9	0.3	1.1	S	2.1	1.8	1.5	0.4	0.5	2.2	5.8	5.9	1.8	0.4	0.9	3.1	7.2	9.0	0.0	9.0	2.0	24	
22	8.1	10.5	10.3	9.9	6.0	2.1	1.1	0.6	S	2.3	1.7	1.1	0.9	1.1	1.9	0.8	0.6	0.8	1.4	1.5	3.3	7.4	7.4	8.7	0.6	10.5	3.9	24	
23	5.0	4.9	3.7	1.3	1.4	3.1	4.8	S	3.0	2.1	2.3	1.3	2.7	3.3	4.9	5.3	2.0	2.0	1.4	1.6	1.9	1.7	3.3	3.7	1.3	5.3	2.9	24	
24	2.1	1.7	1.3	1.6	2.3	1.6	S	3.8	4.4	1.5	1.7	3.0	1.7	3.5	1.4	1.3	1.2	1.2	1.6	1.6	5.3	1.2	0.5	0.6	0.5	5.3	2.0	24	
25	0.6	0.6	0.6	0.6	0.6	S	2.5	1.3	0.9	2.9	5.1	3.7	3.7	1.4	2.1	1.3	1.8	1.8	2.1	1.3	1.2	1.6	1.5	1.1	0.6	5.1	1.8	24	
26	1.0	1.5	2.5	2.4	S	3.1	3.0	3.8	2.1	1.8	1.9	1.9	1.9	1.9	1.5	1.4	1.0	1.2	1.6	1.6	1.5	1.1	1.1	0.7	0.7	3.8	1.8	24	
27	0.6	0.6	0.5	S	2.7	3.2	1.9	0.8	1.8	5.3	3.5	1.7	2.6	3.9	3.0	2.8	4.3	4.3	8.6	8.0	3.2	0.7	0.7	2.1	0.5	8.6	2.9	24	
28	1.2	0.8	S	2.8	1.7	1.3	1.7	2.1	1.5	2.3	2.2	1.2	1.2	0.8	1.0	1.5	2.3	1.3	1.7	1.8	1.6	1.7	1.6	1.4	0.8	2.8	1.6	24	
29	1.2	S	3.6	2.5	2.1	2.0	3.0	2.5	1.8	1.8	1.8	2.2	2.1	2.4	3.1	2.8	2.5	2.5	2.8	2.5	1.4	3.7	5.1	2.2	1.2	5.1	2.5	24	
30	S	3.2	1.6	1.2	0.6	0.3	0.6	0.6	0.5	0.6	0.5	0.4	0.4	0.5	0.4	0.5	0.4	0.3	0.6	0.4	0.2	0.0	0.0	S	0.0	3.2	0.6	24	
31	2.7	1.1	0.4	0.6	0.6	0.4	0.2	0.4	2.6	2.2	2.4	2.8	3.0	2.1	1.6	1.0	1.3	1.7	1.4	2.4	3.5	1.8	S	5.1	0.2	5.1	1.8	24	
HOURLY MAX	8.4	10.5	15.1	11.1	7.9	8.2	9.1	8.9	6.0	5.3	5.1	5.2	3.7	3.9	4.9	5.3	7.8	6.4	8.6	8.0	5.3	7.4	12.3	13.8					
HOURLY AVG	1.7	1.9	2.5	2.1	1.7	1.8	1.9	1.7	1.6	1.5	1.4	1.7	1.5	1.4	1.6	1.5	1.7	1.7	1.7	1.6	1.6	1.4	2.2	2.7					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 159 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0				
NUMBER OF NON-ZERO READINGS:	576				
MINIMUM 1-HR AVERAGE:	0.0 ppb	@ HOUR(S)	0	ON DAY(S)	1
MAXIMUM 1-HR AVERAGE:	15.1 ppb	@ HOUR(S)	2	ON DAY(S)	19
MAXIMUM 24-HR AVERAGE:	5.0 ppb			ON DAY(S)	18
				VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	742 hrs		
MONTHLY CALIBRATION TIME:	8 hrs	AMD OPERATION UPTIME:	99.7 %		
STANDARD DEVIATION:	2.1	MONTHLY AVERAGE:	1.7 ppb		

24 HR AVERAGES March 2017



NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - March 2017

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	0.3	0.3	0.3	0.3	0.3	0.3	S	3.3	0.9	0.9	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	3.3	0.5	24
2	0.3	0.9	0.9	0.3	0.9	S	3.3	5.0	2.1	2.1	2.7	2.7	2.1	2.7	5.0	5.0	9.1	4.4	3.3	10.9	10.3	3.8	3.8	6.2	0.3	10.9	3.8	24	
3	3.3	4.4	3.8	2.7	S	6.2	3.3	2.7	2.7	3.3	3.8	3.8	3.8	3.8	3.3	3.8	4.4	4.4	4.4	4.9	5.0	5.6	2.7	2.7	2.7	6.2	3.9	24	
4	2.1	2.1	3.8	S	9.1	2.7	2.1	2.1	1.5	1.5	0.9	1.5	1.5	1.5	0.9	1.5	1.5	1.5	0.9	0.3	0.3	0.3	0.3	0.3	0.3	0.3	9.1	1.7	24
5	0.3	0.3	S	3.3	0.9	0.9	0.9	0.9	0.3	C	C	C	C	C	C	C	C	C	C	0.9	0.3	0.3	0.3	0.3	0.3	0.3	3.3	0.7	24
6	0.3	S	2.7	0.3	0.9	0.9	0.9	2.7	1.5	1.5	1.5	0.9	0.3	0.9	0.3	0.3	0.3	0.9	2.1	2.1	2.1	1.5	1.5	1.5	0.3	2.7	1.2	24	
7	S	4.4	2.1	2.1	2.7	3.3	3.3	3.3	2.1	1.5	0.9	0.9	0.9	0.9	0.9	1.5	0.9	1.5	2.1	2.1	1.5	0.9	0.9	S	0.9	4.4	1.9	24	
8	3.3	1.5	2.1	1.5	12.1	2.1	2.1	2.1	2.1	0.0	0.0	0.0	0.9	1.5	0.3	0.3	0.3	0.3	0.3	0.3	0.3	1.5	S	4.4	0.0	12.1	1.7	24	
9	2.7	2.7	2.7	2.7	2.1	2.1	8.5	12.1	9.1	6.8	0.9	4.4	6.2	6.2	4.4	10.9	4.4	5.0	6.2	4.4	3.9	S	5.0	3.3	0.9	12.1	5.1	24	
10	3.3	3.9	3.9	2.7	2.1	1.5	0.9	0.3	0.3	2.1	3.3	3.3	0.9	0.9	3.3	5.6	3.8	0.0	0.0	0.0	S	4.4	2.7	0.3	0.0	5.6	2.2	24	
11	2.7	2.7	1.5	0.9	0.9	0.3	0.3	X	X	0.3	1.5	3.8	5.0	4.4	3.8	2.1	2.7	0.9	0.3	S	5.0	0.9	0.3	0.3	0.3	5.0	1.9	22	
12	0.3	0.3	0.3	0.0	0.3	0.3	0.3	0.0	2.7	6.2	3.8	5.0	4.4	1.5	1.5	2.7	2.1	1.5	S	3.3	5.6	4.4	2.7	0.9	0.0	6.2	2.2	24	
13	3.8	3.8	1.5	1.5	1.5	0.9	0.9	1.5	1.5	2.1	2.7	1.5	1.5	0.9	1.5	1.5	1.5	S	4.4	2.1	2.1	1.5	1.5	1.5	0.9	4.4	1.9	24	
14	1.5	2.1	2.1	2.1	5.6	5.6	6.2	5.6	4.4	3.8	3.3	3.3	5.0	5.0	7.9	4.4	S	9.7	9.7	2.7	2.7	12.7	12.7	12.7	1.5	12.7	5.7	24	
15	2.7	2.1	1.5	2.1	2.7	3.3	3.3	5.0	5.0	3.8	3.8	10.9	5.6	4.4	2.1	S	9.7	7.9	6.8	3.8	2.7	0.9	2.7	5.6	0.9	10.9	4.3	24	
16	1.5	0.9	0.9	0.9	2.7	3.8	3.8	1.5	3.8	0.9	0.3	0.3	0.3	0.3	S	3.3	0.9	0.3	0.3	0.3	0.3	0.3	0.3	0.3	2.7	0.3	3.8	1.3	24
17	1.5	8.0	7.9	4.4	2.7	0.9	1.5	0.9	2.1	2.7	2.1	2.1	1.5	S	4.4	2.7	2.1	1.5	2.7	2.7	2.1	2.1	2.1	1.5	0.9	8.0	2.7	24	
18	1.5	10.3	13.9	13.3	5.6	14.4	13.9	12.1	9.1	3.3	5.0	7.9	S	3.8	7.3	5.6	7.9	3.8	0.9	0.9	0.9	0.9	26.7	23.2	0.9	26.7	8.4	24	
19	18.5	15.6	19.7	28.4	12.1	8.5	0.9	1.5	1.5	0.9	0.3	S	3.8	0.9	1.5	1.5	0.3	0.9	1.5	0.3	0.3	3.8	3.8	22.0	0.3	28.4	6.5	24	
20	17.9	2.7	8.5	13.9	22.0	22.0	15.0	6.2	18.5	15.6	S	7.3	2.7	2.7	2.1	2.1	0.9	0.9	0.3	0.3	0.3	0.3	0.3	0.9	0.3	22.0	7.1	24	
21	0.9	1.5	0.9	0.9	0.3	0.3	4.4	2.7	1.5	S	4.9	3.8	3.8	0.9	1.5	7.9	9.1	9.7	3.8	0.9	1.5	6.8	10.9	12.7	0.3	12.7	4.0	24	
22	12.1	13.3	13.9	13.3	10.3	3.8	2.1	0.9	S	5.6	3.8	1.5	0.9	0.9	5.0	3.3	0.9	0.9	2.1	2.1	6.8	8.5	8.0	13.9	0.9	13.9	5.8	24	
23	12.7	15.0	6.2	1.5	2.1	4.4	18.5	S	5.6	3.3	3.3	2.7	3.8	3.8	5.6	6.8	2.7	2.1	2.1	2.1	2.1	2.1	4.4	4.4	1.5	18.5	5.1	24	
24	2.7	2.1	2.1	2.1	3.3	3.3	S	6.7	5.6	2.7	3.3	4.4	3.8	5.0	2.1	2.1	1.5	0.9	3.8	3.3	7.9	2.7	0.3	0.3	0.3	7.9	3.1	24	
25	0.3	0.3	0.3	0.3	0.3	S	5.0	1.5	1.5	6.2	6.8	5.0	5.0	3.8	6.2	2.7	2.1	2.1	2.7	1.5	1.5	2.1	2.1	1.5	0.3	6.8	2.6	24	
26	0.9	2.1	2.7	3.3	S	5.6	7.9	6.2	2.7	2.7	2.1	2.7	2.1	7.3	2.1	1.5	1.5	1.5	1.5	1.5	1.5	1.5	0.9	0.9	0.9	7.9	2.7	24	
27	0.9	0.9	0.3	S	5.0	3.8	3.3	1.5	6.2	8.5	6.7	3.3	4.4	6.8	7.9	5.0	10.9	7.9	12.7	11.5	9.7	0.9	0.9	2.7	0.3	12.7	5.3	24	
28	1.5	0.9	S	5.0	2.1	1.5	2.1	3.3	2.1	2.7	2.7	1.5	1.5	1.5	1.5	3.8	3.8	2.1	2.1	2.1	2.1	2.1	1.5	1.5	0.9	5.0	2.2	24	
29	1.5	S	5.0	2.7	2.1	2.1	5.6	3.3	2.7	2.7	2.7	2.7	2.7	2.7	3.3	3.3	3.2	3.2	3.3	3.3	2.1	9.1	8.5	5.6	1.5	9.1	3.6	24	
30	S	5.6	2.1	1.5	0.9	0.9	0.9	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	S	0.3	5.6	0.8	24	
31	5.0	1.5	0.9	0.9	0.3	0.3	0.3	1.5	3.3	2.7	2.7	3.3	3.3	2.7	2.1	1.5	2.1	2.1	1.5	3.8	4.4	3.3	S	6.7	0.3	6.7	2.4	24	
HOURLY MAX	18.5	15.6	19.7	28.4	22.0	22.0	18.5	12.1	18.5	15.6	6.8	10.9	6.2	7.3	7.9	10.9	10.9	9.7	12.7	11.5	10.3	12.7	26.7	23.2					
HOURLY AVG	3.7	3.9	3.9	4.0	3.9	3.7	4.2	3.3	3.5	3.3	2.6	3.1	2.7	2.7	3.0	3.2	3.1	2.7	2.8	2.5	2.9	2.9	3.7	4.9					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	692
MAXIMUM INSTANTANEOUS VALUE:	28.4 ppb @ HOUR(S) 3 ON DAY(S) 19
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	9 hrs
STANDARD DEVIATION:	3.8
OPERATIONAL TIME:	742 hrs

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)




Wind: LICA MASKWA
 Poll.: LICA MASKWA-NO2[ppb]
 Monthly: 2017/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 7.13% Calm Avg: 2.01 [ppb]

Direction	0.0-5.1	5.1-10.1	10.1-15.2	>15.2	Total
N	10.4	0.1	0.0	0.0	10.6
NE	21.3	0.0	0.0	0.0	21.3
E	14.3	2.4	0.6	0.0	17.3
SE	20.4	0.7	0.0	0.0	21.1
S	11.3	0.3	0.0	0.0	11.6
SW	2.7	0.1	0.0	0.0	2.9
W	3.3	0.6	0.0	0.0	3.9
NW	3.1	0.7	0.6	0.0	4.4
Summary	86.7	5.0	1.1	0.0	92.9

% Icon Classes (ppb) 87 0.0-5.1 5 5.1-10.1 1 10.1-15.2 0 >15.2

LICA MASKWA Poll.: LICA MASKWA-NO2[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 7.13% Calm Poll Avg: 2.01[ppb]



NO2[ppb] Calibration: LICA MASKWA Monthly: 2017/03 Type: Span



Span Meas Span Ref Span Low Span High

WIND SPEED




WIND SPEED Hourly Averages (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY 1	6.1	6.4	5.1	5.6	7.4	6.3	6.1	8.1	8.5	8.8	9.3	9.4	10.3	9.7	7.9	8.0	6.5	4.8	5.2	3.4	2.7	3.8	4.1	2.2	2.2	10.3	6.4	24
2	2.5	2.8	2.9	2.9	2.4	1.9	1.5	2.6	3.0	4.3	4.1	4.1	4.3	4.5	6.9	7.7	5.4	5.2	5.0	5.6	6.7	7.0	4.1	3.7	1.5	7.7	3.8	24
3	3.8	4.7	5.2	4.7	4.7	5.0	4.7	4.4	1.2	5.5	6.5	5.7	6.3	6.3	5.5	8.8	8.1	3.7	0.3	1.4	4.9	5.9	7.7	7.0	0.3	8.8	2.7	24
4	7.1	8.8	13.5	13.1	17.6	16.5	19.7	16.5	15.2	15.8	16.7	14.7	15.1	12.7	12.3	11.6	11.2	12.9	10.7	9.2	7.8	9.3	7.3	8.3	7.1	19.7	12.6	24
5	6.4	7.8	8.4	5.9	6.3	5.8	4.6	6.5	6.3	4.9	4.0	4.6	5.1	3.4	5.1	4.4	5.1	4.8	5.2	5.1	3.7	2.3	1.8	3.8	1.8	8.4	4.9	24
6	8.1	6.8	6.4	7.4	8.6	9.6	7.7	5.7	5.4	11.0	10.0	9.3	8.4	8.0	7.4	7.3	7.0	7.5	8.5	8.3	8.6	7.9	7.8	7.3	5.4	11.0	7.7	24
7	9.2	9.7	8.7	9.6	8.7	8.8	8.4	7.8	9.5	10.5	9.6	9.4	9.6	9.6	8.4	9.3	9.4	9.6	7.3	7.6	8.5	8.4	4.0	3.3	3.3	10.5	8.4	24
8	4.2	5.7	4.7	8.8	3.4	3.0	3.6	8.3	10.3	6.9	6.0	7.3	2.9	3.3	7.3	6.8	5.9	4.4	3.8	5.2	2.2	4.9	3.8	3.6	2.2	10.3	5.1	24
9	1.5	1.3	1.2	1.3	0.9	0.5	0.6	0.5	2.7	4.8	4.4	1.6	0.7	1.9	3.9	6.3	7.3	5.7	2.8	0.9	1.2	1.1	1.0	1.0	0.5	7.3	0.5	24
10	2.1	3.7	6.3	4.9	4.4	4.4	5.6	6.2	6.0	5.7	8.1	8.6	8.8	8.9	5.2	4.1	3.0	5.4	4.6	3.8	3.1	3.1	2.6	2.5	2.1	8.9	4.4	24
11	3.0	4.0	3.8	3.0	3.4	4.4	5.3	X	X	5.5	6.6	6.1	6.4	5.7	6.1	7.1	7.3	5.5	4.6	6.1	5.0	4.5	5.1	6.1	3.0	7.3	4.9	22
12	6.2	6.1	5.4	4.7	5.1	5.1	4.7	5.0	4.5	7.7	7.4	7.7	6.9	6.0	5.6	6.1	6.6	5.0	4.5	3.7	6.1	7.9	6.7	6.9	3.7	7.9	4.6	24
13	7.0	7.4	6.1	6.5	6.5	6.4	7.9	7.6	8.7	7.0	9.0	9.6	9.7	10.0	11.2	8.3	8.8	10.5	10.1	8.3	7.5	7.7	6.5	5.3	5.3	11.2	7.7	24
14	3.7	5.4	4.1	4.5	3.7	3.1	3.2	4.4	5.3	6.2	6.5	4.9	5.7	5.7	4.8	4.8	4.8	3.9	5.5	5.2	4.6	5.2	5.6	6.9	3.1	6.9	4.0	24
15	6.0	5.6	5.6	5.6	5.0	3.6	2.3	2.2	1.2	3.4	4.2	3.6	3.2	4.9	4.7	4.3	7.5	2.9	2.1	2.8	3.8	5.2	4.4	3.1	1.2	7.5	0.5	24
16	2.4	2.6	3.2	1.2	1.7	2.8	3.9	1.6	2.4	4.6	7.6	10.3	9.3	6.5	5.2	7.0	6.6	5.7	5.0	4.4	5.7	6.7	4.9	3.2	1.2	10.3	4.2	24
17	4.4	3.8	4.6	4.0	5.3	5.7	4.6	4.1	4.0	5.0	5.8	7.9	11.5	11.9	13.7	11.0	10.1	7.9	6.9	6.8	8.6	8.0	7.8	8.1	3.8	13.7	6.4	24
18	7.9	6.2	6.6	7.8	9.1	8.3	7.5	8.4	9.0	8.9	8.4	9.9	10.8	11.1	10.1	10.5	9.3	7.4	6.1	6.0	8.6	7.8	8.5	8.1	6.0	11.1	6.5	24
19	7.8	10.9	13.9	13.1	12.2	11.1	10.4	10.2	11.6	11.3	8.9	12.2	12.0	10.4	11.2	10.1	12.6	8.8	7.7	9.7	5.6	8.3	7.0	10.3	5.6	13.9	10.1	24
20	8.9	10.7	8.1	9.4	8.3	8.1	9.1	9.4	7.9	7.3	6.5	5.9	4.0	3.5	3.9	2.9	4.6	4.2	2.2	0.8	2.4	3.2	1.1	1.3	0.8	10.7	4.8	24
21	1.2	1.5	2.2	2.7	1.5	2.4	4.2	4.0	5.0	9.2	8.7	9.1	8.2	10.6	11.9	9.6	10.2	10.0	9.2	8.9	9.1	8.5	8.3	8.4	1.2	11.9	6.2	24
22	7.6	6.7	8.0	7.2	8.0	7.2	7.0	7.0	6.3	7.0	6.7	6.0	6.5	4.7	5.0	4.1	2.8	1.3	1.8	2.9	4.5	4.1	3.8	3.7	1.3	8.0	3.5	24
23	3.0	2.6	2.6	2.6	1.5	0.9	1.5	3.9	5.5	4.0	3.3	1.0	2.2	1.2	0.2	1.4	3.5	3.0	3.1	3.1	2.9	2.5	2.2	4.3	0.2	5.5	0.8	24
24	4.4	2.6	4.1	3.8	3.8	4.0	3.1	5.4	3.3	6.0	7.1	6.5	7.7	6.3	6.7	5.5	6.4	7.1	6.2	6.2	6.1	6.5	6.1	6.3	2.6	7.7	4.7	24
25	6.5	5.8	5.6	5.3	3.6	1.9	1.8	2.7	3.3	2.6	3.5	3.6	3.8	2.8	1.5	3.0	7.7	6.7	5.9	2.6	1.7	2.7	1.4	2.1	1.4	7.7	0.4	24
26	3.7	6.9	4.7	2.0	3.2	3.5	1.6	3.3	6.9	6.4	7.7	7.5	8.5	10.1	9.7	9.2	6.9	8.0	3.8	3.5	0.7	1.7	2.5	2.1	0.7	10.1	4.3	24
27	2.3	2.6	3.5	4.5	3.0	3.8	3.2	4.0	5.3	5.9	7.3	8.3	7.0	7.5	7.1	6.9	8.5	6.4	4.5	5.7	5.4	6.4	6.4	5.0	2.3	8.5	4.7	24
28	5.3	4.8	5.8	4.6	4.0	4.0	4.4	4.6	5.7	5.7	5.7	6.3	6.9	6.2	6.7	4.8	5.5	6.6	4.2	2.3	6.2	6.2	4.0	2.2	2.2	6.9	4.9	24
29	0.7	2.0	1.3	1.3	0.9	2.0	2.0	2.3	3.8	3.6	5.0	3.6	5.7	8.0	8.4	6.2	6.6	5.9	4.7	5.5	4.0	4.7	5.6	5.7	0.7	8.4	2.9	24
30	4.8	5.8	6.5	6.7	8.2	7.2	6.7	6.6	7.8	7.6	8.5	8.3	7.3	7.1	5.6	6.7	6.1	4.3	3.7	0.9	1.0	3.1	4.5	3.5	0.9	8.5	5.5	24
31	3.3	2.5	3.1	2.7	2.6	1.8	2.6	1.7	6.5	7.6	10.1	11.5	11.3	14.1	14.5	14.9	14.7	12.0	4.8	6.6	8.1	7.0	9.5	9.1	1.7	14.9	6.0	24
HOURLY MAX	9.2	10.9	13.9	13.1	17.6	16.5	19.7	16.5	15.2	15.8	16.7	14.7	15.1	14.1	14.5	14.9	14.7	12.9	10.7	9.7	9.1	9.3	9.5	10.3				
HOURLY AVG	2.5	2.6	2.8	2.6	2.7	2.6	2.5	2.8	2.4	2.4	2.7	2.6	2.4	2.4	2.5	2.5	2.7	1.7	1.7	2.2	2.3	2.5	1.9	1.7				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


LAST CALIBRATION:	March 30, 2016
DECLINATION :	MAGNETIC DECLINATION 19 DEGREE EAST



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	742
MINIMUM 1-HR AVERAGE:	0.2 kph @ HOUR(S) 14 ON DAY(S) 23
MAXIMUM 1-HR AVERAGE:	19.7 kph @ HOUR(S) 6 ON DAY(S) 4
MAXIMUM 24-HR AVERAGE:	12.6 kph ON DAY(S) 4
	VAR-VARIOUS
MONTHLY CALIBRATION TIME:	0 hrs
OPERATIONAL TIME:	742 hrs
AMD OPERATION UPTIME:	99.7 %
STANDARD DEVIATION:	3.0
MONTHLY AVERAGE:	2.2 kph

WIND SPEED Hourly Averages (WS kph)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - March 2017

WIND SPEED Instantaneous Maximum (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	22.6	20.6	18.2	22.8	23.0	22.8	19.9	25.6	24.5	27.6	28.5	28.9	24.7	25.2	21.5	20.1	19.5	19.9	16.9	14.4	30.9	14.9	27.4	14.7	14.4	30.9	22.3	24
2	34.6	16.4	15.5	59.2	15.8	41.7	18.2	17.8	12.9	15.8	16.6	21.4	17.1	20.7	23.8	25.6	18.1	21.0	17.3	23.4	25.3	25.0	15.3	15.1	12.9	59.2	22.2	24
3	15.1	17.3	17.2	16.6	17.9	12.9	16.0	14.0	13.4	14.7	19.5	19.9	18.8	20.1	16.8	24.9	22.1	15.5	9.6	12.5	14.2	19.5	14.6	16.2	9.6	24.9	16.6	24
4	21.2	27.6	30.9	42.5	44.8	39.1	49.1	35.3	33.3	44.3	47.4	40.5	35.1	31.1	30.9	29.2	25.4	28.7	28.2	29.1	23.6	19.7	19.7	20.1	19.7	49.1	32.4	24
5	16.5	18.8	18.8	20.2	21.0	19.6	15.3	13.8	14.4																			23
6	16.0	17.1	17.1	16.6	18.4	18.9	18.8	18.2	15.7	27.6	25.9	25.2	23.2	21.2	21.7	25.4	23.2	28.1	26.7	31.8	29.4	23.1	22.8	27.8	15.7	31.8	22.5	24
7	30.2	32.0	35.8	39.5	34.0	37.0	32.4	27.9	30.9	34.2	35.9	46.5	33.3	37.5	32.6	26.9	32.7	26.0	21.8	23.0	24.8	26.7	39.5	33.6	21.8	46.5	32.3	24
8	44.5	14.7	16.2	18.0	72.7	38.0	51.4	20.0	27.7	24.5	19.7	21.4	15.1	17.7	21.2	23.2	19.1	18.0	14.7	17.5	74.0	14.9	45.9	32.1	14.7	74.0	28.4	24
9	29.4	56.5	34.5	44.4	28.4	28.1	27.7	40.2	28.6	17.3	16.9	54.8	16.4	14.9	14.9	18.0	17.3	17.3	16.8	20.9	15.9	35.1	37.0	28.6	14.9	56.5	27.5	24
10	41.5	24.2	18.6	29.3	36.7	14.3	16.2	15.2	17.6	21.6	27.0	25.9	30.3	29.8	19.3	15.6	14.2	16.5	13.6	14.5	39.1	20.3	14.4	40.2	13.6	41.5	23.2	24
11	16.1	15.0	22.7	19.2	42.7	43.4	16.5																					22
12	22.2	15.4	16.3	16.3	15.0	15.9	16.4	18.7	20.4	26.5	24.5	26.4	27.1	22.8	21.4	21.4	22.3	17.5	14.5	12.1	18.2	26.3	20.9	21.2	12.1	27.1	20.0	24
13	19.6	21.3	21.5	18.5	22.2	19.1	22.4	23.7	26.0	21.9	26.5	28.5	32.5	28.2	29.1	25.0	25.0	29.2	29.3	25.2	27.2	21.5	22.2	22.1	18.5	32.5	24.5	24
14	16.4	18.4	12.1	14.7	9.1	10.8	10.8	14.7	13.4	13.1	13.3	16.7	17.3	14.8	14.2	19.7	14.4	12.2	16.2	18.4	16.4	24.9	26.0	23.3	9.1	26.0	15.9	24
15	20.2	18.1	23.4	18.4	15.1	10.5	11.4	9.4	28.0	11.2	15.5	13.1	12.6	18.8	15.3	16.1	15.3	12.9	11.3	10.9	18.1	14.0	15.5	19.2	9.4	28.0	15.6	24
16	9.6	10.5	13.1	5.7	9.8	13.3	12.9	9.8	8.5	12.7	24.7	28.9	23.4	21.0	17.1	23.9	18.1	16.0	14.1	11.1	20.6	20.6	15.7	12.7	5.7	28.9	15.6	24
17	13.8	14.0	13.1	14.7	12.2	16.7	16.6	12.4	13.3	15.2	20.5	29.3	34.3	36.1	36.8	29.7	29.1	32.0	18.9	18.6	21.8	20.2	21.3	22.2	12.2	36.8	21.4	24
18	24.1	22.8	25.5	26.1	26.8	26.8	26.3	28.9	28.7	31.1	23.0	30.0	34.9	32.7	35.2	29.9	31.3	21.9	14.2	15.3	27.6	17.8	33.6	30.6	14.2	35.2	26.9	24
19	30.8	37.0	56.0	47.7	46.4	36.3	37.4	41.4	38.2	38.3	44.6	39.4	40.3	37.4	46.2	53.0	42.4	34.6	34.6	51.7	27.6	39.6	28.4	37.9	27.6	56.0	40.3	24
20	41.2	42.7	31.4	41.5	32.6	46.3	44.5	39.1	36.6	29.4	23.0	24.5	26.0	19.7	21.0	17.7	19.3	14.4	12.5	8.8	10.6	10.8	21.8	9.6	8.8	46.3	26.0	24
21	23.7	15.1	13.8	30.7	51.5	67.9	14.7	12.5	22.5	26.0	29.2	27.4	28.9	31.7	36.5	38.5	36.6	35.0	31.0	28.3	33.7	29.8	32.0	28.7	12.5	67.9	30.2	24
22	31.1	25.1	26.9	25.7	29.6	27.0	25.6	27.2	21.1	27.0	21.4	20.3	20.9	17.6	19.0	19.4	11.8	9.3	10.6	12.6	17.4	14.6	13.2	13.9	9.3	31.1	20.3	24
23	14.6	12.0	11.8	12.0	11.4	9.9	10.7	14.2	14.9	13.3	13.1	11.3	11.8	10.2	10.2	14.9	15.4	13.1	12.2	11.6	8.7	7.2	10.5	10.2	7.2	15.4	11.9	24
24	11.8	7.8	11.4	8.1	10.7	14.0	12.0	15.1	16.6	20.4	24.8	27.6	24.7	24.7	23.6	14.8	19.9	30.4	24.9	20.8	21.2	21.7	18.0	18.6	7.8	30.4	18.5	24
25	22.3	17.7	22.5	14.6	11.8	4.6	5.6	10.4	11.0	10.9	11.8	13.1	12.2	10.7	9.6	15.7	16.8	13.3	13.7	9.4	10.9	12.1	12.1	10.9	4.6	22.5	12.7	24
26	14.6	23.6	30.2	10.5	13.1	13.1	10.7	14.9	18.5	16.8	18.8	20.1	27.5	28.4	28.4	22.7	22.3	19.2	12.2	6.9	7.8	10.7	10.9	11.1	6.9	30.2	17.2	24
27	10.9	11.2	11.4	7.4	12.9	17.1	10.7	9.7	18.5	20.8	25.1	22.3	26.7	20.1	26.9	25.1	25.1	18.1	17.3	17.7	15.1	23.8	18.1	14.3	7.4	26.9	17.8	24
28	17.7	14.0	15.3	13.5	12.9	17.3	11.1	12.4	16.6	15.7	17.7	17.9	20.4	18.0	18.1	14.2	14.4	13.5	13.1	14.9	21.2	21.2	13.7	12.2	11.1	21.2	15.7	24
29	4.2	10.3	10.2	4.0	9.9	8.1	7.9	7.3	11.3	10.6	11.2	11.2	17.1	19.8	22.7	18.1	16.7	15.0	9.5	10.5	11.2	14.5	15.6	14.5	4.0	22.7	12.1	24
30	13.9	15.4	22.1	17.1	19.9	24.9	17.6	17.0	23.8	19.3	19.5	17.2	20.9	17.2	15.9	18.8	15.7	13.7	7.9	7.8	3.7	12.6	12.3	12.8	3.7	24.9	16.1	24
31	12.5	11.8	12.7	11.7	12.6	11.6	12.2	10.2	14.8	16.7	26.1	27.2	26.6	35.7	37.1	39.9	36.6	33.1	11.9	25.3	24.8	16.7	20.7	20.0	10.2	39.9	21.2	24
HOURLY MAX	44.5	56.5	56.0	59.2	72.7	67.9	51.4	41.4	38.2	44.3	47.4	54.8	40.3	37.5	46.2	53.0	42.4	35.0	34.6	51.7	74.0	39.6	45.9	40.2				
HOURLY AVG	21.4	20.1	20.8	22.2	23.9	23.5	20.0	19.2	20.7	21.4	22.8	24.9	24.1	23.4	23.4	23.5	22.0	20.4	17.0	18.1	21.8	19.8	21.1	20.3				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

MAXIMUM INSTANTANEOUS VALUE:	74.0	kph	@ HOUR(S)	20	ON DAY(S)	8
					VAR-VARIOUS	
OPERATIONAL TIME:					741	hrs

WIND SPEED Instantaneous Maximum (WS kph)



Wind: LICA MASKWA
 Monitor: WSP [kph]
 Monthly: 2017/03
 Type: WindRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 6.87%

Calm Avg: 0.00 [kph]

Direction	1.8-4.0	4.0-7.9	7.9-11.9	11.9-15.8	15.8-19.8	>19.8	Total
N	2.6	3.4	3.9	0.7	0.1	0.0	10.6
NE	5.3	11.7	3.8	0.4	0.7	0.0	21.8
E	3.8	11.9	1.9	0.0	0.0	0.0	17.5
SE	3.5	13.2	4.0	0.1	0.0	0.0	20.9
S	1.8	5.9	2.8	0.8	0.0	0.0	11.3
SW	1.6	0.7	0.4	0.1	0.0	0.0	2.8
W	0.8	0.8	1.8	0.4	0.0	0.0	3.8
NW	1.8	0.8	1.4	0.4	0.0	0.0	4.3
Summary	21.0	48.4	19.9	2.9	0.8	0.0	93.1

% Icon Classes (kph)	21	1.8-4.0	48	4.0-7.9	20	7.9-11.9	3	11.9-15.8	1	15.8-19.8	0	>19.8
----------------------	----	---------	----	---------	----	----------	---	-----------	---	-----------	---	-------

LICA MASKWA 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 6.87% Calm Wind Avg Speed: 1.17(kph)



WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - March 2017

WIND DIRECTION Hourly Averages (WD)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	24-HOUR AVG	24-HR	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	QUADRANT	RDGS.	
DAY 1	ENE	ENE	ENE	ENE	ENE	ENE	NE	NE	NE	NE	NE	NE	NE	NE	NE	NE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	NE	NE	24	
2	NE	ENE	ESE	SE	SE	ESE	ESE	ENE	ENE	SE	SE	NE	ENE	ESE	ESE	SE	ESE	ENE	ENE	ESE	ESE	E	ENE	E	ENE	24	
3	ENE	E	ESE	ESE	ESE	ESE	ESE	ESE	ESE	SSE	SSE	SE	SSE	SE	SE	SSW	SSW	SSW	SE	NNW	N	N	NNE	NNE	ESE	24	
4	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NE	NE	NNE	NNE	NE	NNE	24	
5	NE	NE	NNE	ENE	ENE	ENE	NE	NNE	NNE	NNE	NE	NE	ENE	NE	ENE	E	NE	NNE	NE	NE	ENE	NE	NNE	NE	NNE	24	
6	NNE	NE	NE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	N	N	N	N	N	N	N	N	N	N	N	N	N	N	NNE	24
7	N	N	N	N	NNW	NNW	NNW	NNW	N	N	N	N	N	NNW	N	N	N	NNE	N	NNE	N	NNE	N	N	N	N	24
8	NNE	NNE	NE	NNE	N	N	NNE	NNE	NE	NE	ENE	NE	N	ENE	NNE	NNE	NNE	N	N	NNE	NNE	NNE	NNE	NNE	NNE	NNE	24
9	NE	E	NE	NE	NE	ESE	SSE	WNW	NNE	NNE	NNE	NNW	W	SW	SW	SSW	SSW	SSW	SSW	SSE	SE	E	E	NE	S	24	
10	ENE	ENE	NE	NE	NE	NE	NE	NE	E	ESE	E	E	ENE	E	ESE	ESE	SE	SE	SE	SE	E	ENE	ENE	NE	E	24	
11	E	ESE	ESE	ESE	E	ESE	SE	X	X	ESE	ESE	ESE	ESE	ESE	ESE	E	E	ENE	ENE	E	E	ENE	ENE	E	ENE	22	
12	E	NE	NE	ENE	NE	NE	ENE	ENE	ENE	ESE	ESE	ESE	ESE	ESE	SE	ESE	ESE	ESE	ESE	ESE	SSE	SSE	S	SSE	ESE	24	
13	SSE	SSE	SE	SE	SE	SE	SE	SE	SE	SSE	SSE	SSE	SSE	SE	SSE	SSE	S	S	S	S	S	S	S	S	S	SSE	24
14	SSW	S	S	S	SSE	SSE	S	S	SSW	SSW	SSW	S	SSE	SSE	SE	SE	ESE	E	ESE	ESE	ESE	E	ESE	ESE	ESE	SSE	24
15	ESE	ESE	SE	SE	SE	SE	ESE	ESE	WSW	SSW	SW	WNW	NW	N	NNW	SSW	SSW	SW	N	N	N	NNE	N	NNW	SE	24	
16	N	ENE	E	NNE	WSW	SE	SE	ESE	NE	NE	NE	NE	NE	ENE	E	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	E	ENE	24	
17	ENE	E	ESE	SE	ESE	SE	SE	ESE	SE	ESE	SSE	SSE	S	S	S	SSE	SSE	SSE	SE	SE	SE	SE	SE	SE	SSE	24	
18	ESE	E	E	ESE	ESE	ESE	E	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ESE	ENE	NE	NNE	WNW	NW	E	24		
19	WNW	WNW	NW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	W	W	WNW	W	W	W	W	W	W	WNW	NNW	NW	WNW	24
20	NNW	NNW	ENE	NW	NW	NW	NW	NW	NW	NW	NW	NW	NNW	WNW	NW	W	NNW	W	W	SSE	SSW	SSW	SE	ENE	NW	24	
21	E	E	ENE	NE	NNE	NE	NE	NE	SE	SE	SE	SE	ESE	SE	SE	ESE	ESE	ESE	ESE	SE	ESE	ESE	ESE	ESE	ESE	ESE	24
22	ESE	ESE	ESE	ESE	E	E	E	E	E	ENE	E	ENE	ENE	ENE	E	E	NE	WNW	WNW	W	WNW	WNW	WNW	WNW	E	24	
23	NNW	W	NW	N	NW	NNW	N	NNE	NNE	NNE	NE	E	E	WSW	NW	E	E	E	ESE	SE	SSW	SSE	SSE	S	ENE	24	
24	SSE	S	SSE	S	SE	SE	ESE	SE	ESE	ENE	E	ESE	ESE	E	ENE	ENE	ENE	ENE	ENE	E	E	E	E	E	ENE	E	24
25	ENE	ENE	ENE	ENE	NE	NNE	N	N	N	WNW	WNW	WNW	WNW	WNW	W	SSW	S	SSW	SSW	SSE	SE	ESE	E	SE	E	24	
26	SSE	SSE	S	NE	ESE	ESE	E	S	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	S	SSW	S	S	SSE	ENE	ESE	ESE	S	24	
27	NNE	NE	NE	NNE	E	ENE	NE	NE	ENE	E	ESE	ESE	ESE	ESE	ESE	ESE	ESE	E	E	ESE	ESE	SE	SE	SSE	ESE	24	
28	SE	SE	SE	SE	SE	SE	SE	SE	SE	SSE	SSE	SE	SE	SE	SSE	SE	ESE	S	SSW	SE	SE	SE	SE	SE	SE	24	
29	E	ENE	NE	NNE	ESE	S	SSW	SSW	S	SW	SSW	SSW	SSW	S	S	SSW	S	SSW	S	SE	E	E	E	E	SSE	24	
30	ESE	E	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	NE	NE	NE	NE	ENE	ENE	ENE	NNE	NE	ESE	WSW	ENE	ENE	ENE	ENE	24	
31	NE	NE	ENE	NNE	NNE	NNE	ENE	NNE	SSW	S	SSW	SSW	SSW	SSW	SSW	SSW	S	S	SSE	SSE	S	S	SSW	S	24		


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION:	March 30, 2016
DECLINATION :	MAGNETIC DECLINATION 19 DEGREE EAST

MONTHLY CALIBRATION TIME:	0 hrs	OPERATIONAL TIME:	742 hrs
STANDARD DEVIATION:	92	AMD OPERATION UPTIME:	99.7 %
		MONTHLY AVERAGE:	84 (E)

WIND DIRECTION Hourly Averages (WD)



STANDARD DEVIATION WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - March 2017

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00		
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	RDGS.	
DAY																										
1	23	23	22	23	22	25	22	23	24	26	26	26	21	22	26	23	25	22	19	19	17	22	25	29	24	
2	19	27	32	26	33	24	30	19	21	29	43	39	45	50	30	27	29	21	19	26	35	31	29	23	24	
3	21	21	20	17	17	20	18	17	70	34	30	35	34	33	30	24	20	26	40	60	25	19	12	12	24	
4	12	16	13	18	16	15	15	14	16	17	18	17	17	17	16	18	15	17	16	22	18	13	16	17	24	
5	20	18	16	26	24	26	26	15	16	21	32	35	36	49	37	38	29	23	15	14	19	23	36	21	24	
6	12	17	19	15	15	13	17	19	21	16	20	21	22	24	26	28	33	27	26	27	27	27	27	31	24	
7	29	30	34	33	35	33	33	36	30	28	29	31	32	33	33	30	24	25	19	15	19	18	28	30	24	
8	18	11	16	12	26	25	29	16	18	33	33	26	64	55	31	29	28	29	24	18	24	13	21	16	24	
9	36	31	32	50	64	46	65	55	24	26	40	71	60	63	40	31	22	18	15	33	34	38	42	66	24	
10	43	27	21	24	25	22	17	16	19	32	34	31	29	29	34	32	27	20	25	21	23	22	28	18	24	
11	26	16	13	21	24	25	32	X	X	31	28	38	35	38	36	31	29	27	23	27	27	24	26	25	22	
12	26	22	20	22	18	19	23	21	33	33	35	35	40	41	40	38	29	27	20	22	20	23	25	24	24	
13	25	24	25	25	27	27	24	26	24	29	28	29	30	29	27	30	25	23	22	23	26	23	36	38	24	
14	35	26	26	27	24	26	26	24	21	21	23	37	34	32	31	33	30	25	23	31	33	32	32	30	24	
15	30	25	26	25	23	20	28	33	71	34	30	43	50	41	39	39	17	40	34	23	19	29	49	24		
16	29	27	37	56	32	24	19	29	38	28	27	23	23	36	37	28	24	22	22	21	20	24	23	35	24	
17	23	26	31	41	26	28	34	36	36	30	34	33	29	25	23	26	27	22	22	21	21	22	20	23	24	
18	28	31	29	29	30	27	29	28	29	31	30	31	28	29	30	30	28	26	22	19	24	27	31	36	24	
19	30	25	35	27	28	27	29	27	26	28	34	30	30	32	33	36	30	29	32	28	27	30	39	34	24	
20	32	37	38	39	38	38	36	37	38	38	37	43	56	50	52	51	46	36	40	49	15	11	45	53	24	
21	29	36	29	19	49	20	13	16	31	26	31	32	38	28	26	29	31	30	28	28	27	31	30	31	24	
22	30	30	28	29	28	30	28	26	29	28	29	27	26	26	29	26	19	45	32	37	30	27	28	29	24	
23	30	27	41	38	45	44	46	46	24	31	48	58	52	59	68	52	32	31	24	29	17	38	35	18	24	
24	17	21	12	14	18	18	19	28	37	26	27	33	30	34	28	28	25	27	34	28	28	27	24	26	24	
25	25	29	25	23	19	18	27	31	28	37	40	37	36	45	66	61	18	16	15	17	40	35	51	52	24	
26	30	24	41	44	28	24	40	26	20	23	25	28	29	23	24	25	28	20	18	17	60	30	13	38	24	
27	21	18	16	8	28	26	24	16	27	36	31	27	39	29	30	31	27	28	31	27	22	25	25	26	24	
28	26	28	24	28	25	25	22	27	29	29	34	33	32	34	29	32	27	19	20	21	25	28	32	41	24	
29	63	32	41	53	47	39	52	36	28	39	24	45	42	30	30	39	28	23	19	17	28	30	28	29	24	
30	30	29	29	27	27	26	26	26	26	30	24	22	33	37	48	37	34	29	21	51	71	34	22	28	24	
31	25	31	24	31	29	58	35	50	27	26	20	20	21	20	22	20	20	21	17	19	22	22	19	17	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION: March 30, 2016

CALIBRATION TIME: 0 hrs OPERATIONAL TIME: 742 hrs

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)



RELATIVE HUMIDITY




RELATIVE HUMIDITY Hourly Averages (RH %)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	71	70	70	70	70	70	70	69	66	63	59	56	54	53	51	54	58	62	66	70	71	70	71	71	51	71	65	24
2	71	73	72	72	73	73	73	72	69	64	56	52	48	46	45	48	53	62	69	71	72	75	77	79	45	79	65	24
3	80	81	81	81	82	82	83	83	80	72	63	53	48	44	42	44	49	58	72	78	72	74	75	78	42	83	69	24
4	82	81	78	77	76	72	69	69	72	73	71	72	72	71	C	C	74	73	72	73	73	73	73	72	69	82	74	24
5	72	71	71	71	71	71	71	71	71	68	64	61	59	57	57	56	61	65	68	70	73	77	76	75	56	77	68	24
6	74	73	73	73	73	72	72	76	77	76	76	76	76	76	76	76	76	77	76	76	75	75	74	74	72	77	75	24
7	74	73	72	70	70	67	65	62	55	50	45	40	37	36	36	38	41	44	50	54	53	54	57	60	36	74	54	24
8	62	62	62	58	61	64	60	55	48	41	38	37	33	35	36	38	39	41	45	48	54	55	56	62	33	64	50	24
9	65	68	67	67	67	65	63	63	61	48	37	29	26	26	26	30	36	41	49	57	62	65	66	65	26	68	52	24
10	64	64	64	63	64	63	61	59	53	43	33	32	28	26	30	33	36	39	44	48	52	55	57	62	26	64	49	24
11	67	71	73	72	71	69	65	X	X	59	55	54	52	51	56	57	59	64	69	69	70	72	73	74	51	74	65	22
12	74	73	73	73	73	72	72	71	65	57	52	46	42	41	40	41	46	51	59	63	63	62	65	67	40	74	60	24
13	70	71	73	72	72	72	72	71	68	63	59	50	44	40	43	48	51	56	60	64	66	68	69	71	40	73	62	24
14	72	74	76	78	80	81	81	79	69	61	54	45	44	44	41	44	48	53	61	64	67	69	70	71	41	81	64	24
15	73	77	79	81	83	86	87	86	75	69	61	50	38	30	24	32	48	53	55	42	39	40	47	66	24	87	59	24
16	86	80	75	74	78	81	83	84	77	72	68	72	62	52	50	50	50	47	48	49	50	51	55	57	47	86	65	24
17	57	57	57	63	64	66	68	63	60	59	57	57	53	53	53	51	56	67	73	76	79	81	82	83	51	83	64	24
18	84	85	86	85	84	83	82	80	74	67	65	61	57	53	53	55	59	65	69	72	72	73	80	87	53	87	72	24
19	90	90	87	86	80	78	78	71	66	61	54	53	52	49	46	57	49	53	55	56	62	62	58	53	46	90	64	24
20	50	50	51	51	53	54	55	56	54	49	43	35	30	27	23	22	23	28	37	47	53	63	70	72	22	72	46	24
21	74	75	75	76	76	76	75	68	56	45	43	41	39	38	41	43	50	54	56	58	57	59	61	63	38	76	58	24
22	64	65	67	70	74	76	77	77	73	71	73	80	83	85	85	84	84	85	86	87	87	87	87	87	64	87	79	24
23	86	86	85	85	86	86	86	85	80	77	74	71	68	68	68	70	76	78	81	83	83	84	84	84	68	86	80	24
24	83	83	84	87	87	87	87	86	85	82	78	75	75	72	73	75	77	79	80	80	79	80	82	82	72	87	81	24
25	83	83	84	84	84	84	85	83	78	72	70	67	64	60	56	58	65	67	74	84	88	89	89	89	56	89	77	24
26	89	86	82	84	87	87	87	85	79	72	61	56	53	49	47	49	48	53	62	73	81	85	86	86	47	89	72	24
27	85	88	87	87	88	88	87	84	76	61	54	52	50	51	51	51	53	54	56	59	62	68	77	79	50	88	69	24
28	80	81	81	81	82	83	84	83	82	81	76	73	71	68	67	66	64	72	77	82	83	87	88	88	64	88	78	24
29	88	88	88	88	90	90	91	90	89	85	81	77	72	68	64	59	60	64	73	83	87	89	90	90	59	91	81	24
30	91	91	91	91	91	91	90	90	90	90	89	87	82	76	65	61	61	65	73	82	87	88	88	89	61	91	83	24
31	89	89	89	89	88	88	88	86	80	72	65	58	54	52	51	49	51	59	69	71	68	67	70	49	89	71	24	
HOURLY MAX	91	91	91	91	91	91	91	90	90	90	89	87	83	85	85	84	84	84	85	86	88	89	90	90				
HOURLY AVG	76	76	76	76	77	77	76	75	71	65	60	57	54	52	50	51	55	59	64	67	69	71	73	74				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY


MINIMUM 1-HR AVERAGE:	22	%	@ HOUR(S)	15	ON DAY(S)	20
MAXIMUM 1-HR AVERAGE:	91	%	@ HOUR(S)	6	ON DAY(S)	29
MAXIMUM 24-HR AVERAGE:	83	%			ON DAY(S)	30
					VAR-VARIOUS	
OPERATIONAL TIME:						742 hrs
AMD OPERATION UPTIME:						99.7 %
STANDARD DEVIATION:	15					MONTHLY AVERAGE: 67 %

RELATIVE HUMIDITY Hourly Averages (RH %)



BAROMETRIC PRESSURE

BAROMETRIC PRESSURE Hourly Averages (BP mbar)



AMBIENT TEMPERATURE




AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	-19.0	-18.9	-18.7	-18.3	-18.5	-19.0	-19.3	-19.3	-19.0	-18.1	-16.8	-15.5	-14.6	-13.5	-12.4	-12.8	-13.7	-15.4	-17.0	-18.9	-19.4	-18.5	-18.3	-17.7	-19.4	-12.4	-17.2	24	
2	-17.4	-17.6	-17.5	-18.1	-18.7	-20.9	-23.2	-23.9	-19.6	-15.1	-11.3	-9.4	-7.2	-5.5	-4.7	-5.2	-6.4	-9.1	-10.6	-9.6	-9.0	-9.3	-9.8	-11.4	-23.9	-4.7	-12.9	24	
3	-13.2	-12.7	-11.0	-11.3	-11.6	-11.2	-11.2	-11.1	-9.9	-5.1	-2.3	0.8	2.9	5.0	6.0	5.0	3.1	0.5	-3.5	-5.0	-3.1	-3.2	-2.7	-3.6	-13.2	6.0	-4.5	24	
4	-4.0	-2.7	-3.2	-5.2	-7.3	-10.0	-11.8	-13.1	-14.3	-14.7	-14.2	-14.1	-14.4	-14.3	-14.4	-14.2	-14.2	-15.3	-15.9	-16.1	-16.4	-16.8	-16.8	-17.2	-17.2	-2.7	-12.5	24	
5	-17.5	-17.9	-18.1	-18.0	-18.1	-18.2	-18.1	-18.1	-17.5	-15.8	-14.5	-12.8	-12.0	-10.4	C	C	-11.7	-12.4	-13.1	-13.6	-14.9	-17.3	-15.4	-14.4	-18.2	-10.4	-15.4	24	
6	-14.2	-14.3	-14.3	-14.4	-14.3	-14.1	-13.9	-13.6	-13.0	-12.7	-12.5	-12.2	-11.9	-11.4	-10.9	-10.7	-11.4	-12.6	-13.3	-14.1	-14.5	-14.8	-15.2	-15.4	-15.4	-10.7	-13.3	24	
7	-15.9	-16.6	-17.2	-17.3	-17.1	-17.3	-17.5	-17.6	-17.0	-15.3	-13.6	-11.5	-10.4	-10.0	-10.3	-10.8	-11.4	-12.6	-15.1	-17.0	-18.4	-20.1	-21.9	-23.3	-23.3	-10.0	-15.6	24	
8	-25.0	-25.1	-25.7	-24.9	-26.3	-27.6	-26.2	-25.3	-23.6	-20.9	-18.7	-17.6	-14.4	-13.8	-14.5	-14.3	-14.5	-16.2	-18.4	-19.9	-21.9	-22.2	-23.1	-25.5	-27.6	-13.8	-21.1	24	
9	-27.2	-27.8	-28.8	-30.5	-31.4	-32.6	-33.4	-32.8	-26.2	-21.6	-18.2	-14.4	-13.2	-12.7	-12.7	-13.5	-15.8	-18.2	-21.7	-25.0	-27.2	-29.2	-30.0	-31.0	-33.4	-12.7	-24.0	24	
10	-31.1	-31.3	-29.5	-29.6	-29.8	-29.1	-28.4	-27.7	-24.7	-19.6	-16.0	-15.8	-14.9	-14.4	-14.8	-15.7	-17.1	-17.9	-18.1	-18.6	-19.1	-19.4	-19.3	-19.6	-31.3	-14.4	-21.7	24	
11	-20.2	-21.9	-23.2	-23.1	-22.3	-20.8	-19.1	X	X	-17.5	-15.8	-14.5	-13.1	-12.8	-13.6	-13.4	-13.9	-15.2	-16.6	-16.4	-16.2	-16.5	-16.4	-16.2	-23.2	-12.8	-17.2	22	
12	-16.4	-17.2	-18.3	-18.6	-18.7	-19.3	-19.6	-19.3	-16.5	-13.7	-11.9	-9.9	-8.6	-7.8	-7.5	-7.5	-9.6	-11.2	-13.8	-15.1	-14.4	-13.8	-14.3	-15.1	-19.6	-7.5	-14.1	24	
13	-15.8	-16.2	-16.6	-16.4	-16.0	-15.8	-15.6	-15.0	-13.5	-12.0	-10.5	-7.3	-4.9	-3.1	-3.8	-5.4	-6.2	-7.3	-8.4	-9.3	-9.6	-10.1	-10.1	-10.0	-16.6	-3.1	-10.8	24	
14	-10.4	-10.8	-11.0	-11.4	-12.1	-12.4	-12.4	-10.9	-8.0	-5.4	-2.7	0.8	2.4	3.0	4.8	4.5	3.5	1.9	0.5	-0.2	-0.5	-0.5	-0.4	-0.4	-12.4	4.8	-3.7	24	
15	-0.7	-1.8	-2.2	-2.6	-3.3	-4.3	-5.5	-4.1	-0.1	1.4	4.4	8.3	12.4	13.8	14.4	12.2	7.5	6.3	4.1	5.0	4.6	4.4	4.5	3.9	-5.5	14.4	3.4	24	
16	2.3	2.6	2.8	2.3	1.1	0.2	1.2	1.4	3.6	4.5	5.1	3.7	5.3	6.1	6.3	5.6	4.8	3.7	2.5	1.6	1.0	0.3	-0.6	-0.9	-0.9	6.3	2.8	24	
17	-1.1	-1.0	-1.0	-1.3	-1.4	-1.5	-1.5	-1.0	-0.1	1.1	3.5	6.2	7.6	7.9	7.1	7.4	5.8	3.8	2.5	1.7	1.8	1.5	1.2	0.9	-1.5	7.9	2.1	24	
18	0.4	-0.3	-0.8	-0.5	-0.4	-0.1	-0.2	0.0	1.5	3.3	4.2	5.4	7.0	8.4	8.7	8.3	7.0	5.1	3.7	3.0	3.4	3.5	3.1	1.2	-0.8	8.7	3.1	24	
19	0.6	0.9	0.8	0.3	-0.6	-1.4	-2.2	-1.9	-0.9	0.3	2.5	3.0	2.6	2.8	3.9	3.0	2.5	1.0	-0.2	-1.1	-2.3	-2.5	-3.7	-4.7	-4.7	3.9	0.1	24	
20	-7.2	-9.1	-10.2	-11.2	-11.7	-12.1	-12.6	-12.1	-10.9	-9.5	-7.5	-5.3	-4.0	-3.1	-2.3	-2.2	-2.8	-4.8	-7.3	-10.3	-11.9	-13.8	-14.8	-16.0	-16.0	-2.2	-8.9	24	
21	-16.9	-17.9	-18.5	-19.5	-19.6	-19.8	-18.7	-15.2	-9.6	-5.6	-3.4	-1.1	-0.1	0.6	0.1	0.0	-1.3	-2.1	-2.6	-2.8	-2.9	-3.0	-3.1	-3.2	-19.8	0.6	-7.8	24	
22	-3.2	-3.5	-3.6	-3.9	-4.5	-5.1	-5.4	-5.1	-4.2	-3.5	-3.3	-3.7	-4.3	-4.1	-4.0	-3.7	-3.6	-3.4	-3.9	-4.0	-4.0	-3.9	-4.2	-4.4	-5.4	-3.2	-4.0	24	
23	-4.0	-4.0	-4.0	-4.1	-5.5	-7.0	-6.9	-5.2	-3.5	-2.5	-1.4	-0.5	0.5	0.6	0.8	0.7	-0.4	-1.0	-1.8	-2.4	-2.6	-3.0	-2.8	-2.7	-7.0	0.8	-2.6	24	
24	-2.5	-2.6	-3.4	-4.8	-4.6	-5.8	-7.5	-5.8	-4.8	-3.6	-3.0	-1.9	-1.5	0.3	1.1	1.2	1.3	1.0	0.6	0.3	0.2	-0.4	-1.3	-1.7	-7.5	1.3	-2.1	24	
25	-2.0	-2.0	-1.9	-1.7	-1.7	-1.6	-1.7	-1.1	0.4	2.3	3.4	4.4	5.4	6.9	8.2	7.4	5.1	4.3	2.0	-1.4	-2.6	-3.3	-4.2	-4.0	-4.2	8.2	0.9	24	
26	-2.8	-1.3	-0.5	-1.3	-2.6	-2.4	-2.3	-1.0	0.8	2.9	6.0	7.3	8.2	8.5	8.9	8.6	8.7	7.0	4.3	1.0	-1.1	-2.7	-3.7	-3.9	-3.9	8.9	1.9	24	
27	-4.4	-5.0	-5.5	-4.5	-4.0	-2.7	-3.1	-1.9	0.4	3.6	4.7	5.2	6.5	6.5	6.6	6.6	5.7	5.2	4.1	3.3	2.8	2.6	2.4	2.2	-5.5	6.6	1.6	24	
28	2.0	1.9	1.6	1.3	1.2	1.1	1.2	1.7	2.1	2.5	4.0	5.0	5.6	6.4	7.0	7.3	7.7	5.7	4.0	2.5	3.2	2.5	2.2	2.1	1.1	7.7	3.4	24	
29	1.7	1.8	1.6	1.4	1.4	1.3	1.7	2.0	2.3	3.2	3.7	4.7	5.6	6.6	7.6	8.5	8.2	7.0	4.5	1.7	0.7	0.3	0.7	0.3	0.3	8.5	3.3	24	
30	0.1	-0.3	-0.5	-0.6	-0.7	-1.3	-1.3	-1.0	-0.6	-0.1	0.5	1.4	2.5	3.9	6.8	7.7	7.6	6.2	3.5	0.8	-0.6	-1.5	-1.7	-2.7	-2.7	7.7	1.2	24	
31	-3.1	-3.6	-3.3	-3.5	-4.4	-4.9	-4.1	-0.3	2.6	5.1	7.4	9.3	10.5	10.7	10.5	10.1	9.3	9.1	6.7	6.3	6.8	6.3	6.0	5.0	-4.9	10.7	3.9	24	
HOURLY MAX	2.3	2.6	2.8	2.3	1.4	1.3	1.7	2.0	3.6	5.1	7.4	9.3	12.4	13.8	14.4	12.2	9.3	9.1	6.7	6.3	6.8	6.3	6.0	5.0					
HOURLY AVG	-9.3	-9.6	-9.7	-10.0	-10.4	-10.8	-10.9	-9.9	-8.1	-6.5	-4.8	-3.3	-2.1	-1.3	-0.6	-0.8	-2.1	-3.4	-5.1	-6.2	-6.7	-7.2	-7.5	-8.0					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY


MINIMUM 1-HR AVERAGE:	-33.4 °C	@ HOUR(S)	6	ON DAY(S)	9
MAXIMUM 1-HR AVERAGE:	14.4 °C	@ HOUR(S)	14	ON DAY(S)	15
MAXIMUM 24-HR AVERAGE:	3.9 °C			ON DAY(S)	31
				VAR-VARIOUS	
OPERATIONAL TIME:				742	hrs
AMD OPERATION UPTIME:				99.7	%
STANDARD DEVIATION:	9.7			MONTHLY AVERAGE:	-6.5 °C

AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)



PRECIPITATION

PRECIPITATION Hourly Averages (mm)



APPENDIX II
EQUIPMENT CALIBRATION RESULTS

SULPHUR DIOXIDE



API 100E Sulphur Dioxide Analyzer Calibration

Date: <u>March 5, 2017</u>	Barometric Pressure: <u>0.920 atm</u>
Company/Airshed: <u>LICA</u>	Station Temperature °C: <u>22</u>
Location/Station Name: <u>Maskwa</u>	Weather Conditions: <u>Mainly cloudy with snow</u>
Parameter: <u>Sulphur Dioxide</u>	Calibration Purpose: <u>shut down</u>
Start Time 24 hr. (mst): <u>9:22</u>	Performed By/Reviewer: <u>Alex Yakupov</u> <u>Trina Whitsitt</u>
End Time 24 hr. (mst): <u>13:38</u>	Cal Gas Expiry Date: <u>July 18, 2019</u>
Calibration Method: <u>Gas Dilution</u>	Converter Model & s/n (if applicable): <u>n/a</u>

Analyzer:	
ID# or Serial Number: <u>508</u>	Range ppb: <u>1000</u>
Last Calibration Date: <u>February 14, 2017</u>	As Found C.F.: <u>1.000</u>
Previous C.F.: <u>1.000</u>	New C.F.: <u>n/a</u>

Calibrator: Flow Meter ID's: <u>n/a</u> Make & Model: <u>API 700</u> Serial #: <u>627</u> Cal Gas Cylinder I.D. #: <u>LL104222</u> Cal Gas Conc. (ppm): <u>50.6</u>	Standard Calibration Points for Ranges <table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>780</td></tr> <tr><td>Mid</td><td>380</td></tr> <tr><td>Low</td><td>190</td></tr> </table>	Point	ppb	High	780	Mid	380	Low	190
Point	ppb								
High	780								
Mid	380								
Low	190								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated Concentration: (ppb)	Indicated Concentration: (ppb)	Correction Factors (C.F.):
	Diluent	Cal Gas	Total			
as found zero	5000	0.00	5000	0.0	0.0	n/a
as found high	4923	76.90	5000	778.2	778.0	1.000
mid	4964	37.50	5002	379.4	378.0	1.004
low	4982	18.70	5001	189.2	188.0	1.006
Average C.F.=						1.003

Linear Regression/Calibration Results:

Correlation Coefficient =	<u>1.000</u>	LIMITS
		> or = 0.995
Slope =	<u>1.000</u>	0.90-1.10
b (Intercept as % of full scale) =	<u>0.07%</u>	± 3% F.S.
% change in C.F. from last cal =	<u>-0.03%</u>	± 10%

API 100E Sulphur Dioxide Analyzer Calibration



As found: SLOPE: <u>0.980</u> OFFSET: <u>126.3</u> HVPS: <u>479</u> RCELL TEMP: <u>50.0</u> BOX TEMP: <u>30.6</u> PMT TEMP: <u>7.7</u> IZS TEMP: <u>45.0</u> PRES: <u>24.5</u> SAMP FL: <u>616</u> NORM PMT: <u>126.9</u> UV LAMP: <u>2764.7</u> LAMP RATIO: <u>93.0</u> STR. LGT: <u>61.9</u> DRK PMT: <u>10.1</u> DRK LMP: <u>-0.6</u> Expected Value: <u>365.0</u>	As left: SLOPE: <u>n/a</u> OFFSET: <u>n/a</u> HVPS: <u>n/a</u> RCELL TEMP: <u>n/a</u> BOX TEMP: <u>n/a</u> PMT TEMP: <u>n/a</u> IZS TEMP: <u>n/a</u> PRES: <u>n/a</u> SAMP FL: <u>n/a</u> NORM PMT: <u>n/a</u> UV LAMP: <u>n/a</u> LAMP RATIO: <u>n/a</u> STR. LGT: <u>n/a</u> DRK PMT: <u>n/a</u> DRK LMP: <u>n/a</u> Expected Value: <u>n/a</u>
--	--

Comments:

NOx and SO2 calibrations performed with same gas and calibration equipment. 'SO2 as-found high was extended while obtaining points for Nox.

Shutdown calibration completed for annual maintenance. Clean sample valve, Check reaction cell is clean. Move zero. Check charcoal inside analyzer. HVPS adjusted. Sample flow rate and pressure calibrated.



API 100E Sulphur Dioxide Analyzer Calibration

Date: March 5, 2017	Barometric Pressure: 0.920 atm
Company/Airshed: LICA	Station Temperature °C: 22
Location/Station Name: Maskwa	Weather Conditions: Mainly cloudy with snow
Parameter: Sulphur Dioxide	Calibration Purpose: post repair
Start Time 24 hr. (mst): 16:27	Performed By/Reviewer: Alex Yakupov Trina Whitsitt
End Time 24 hr. (mst): 19:13	Cal Gas Expiry Date: July 18, 2019
Calibration Method: Gas Dilution	Converter Model & s/n (if applicable): n/a

Analyzer:	
ID# or Serial Number: 508	Range ppb: 1000
Last Calibration Date: n/a	As Found C.F.: n/a
Previous C.F.: n/a	New C.F.: 1.000


Calibrator:	Standard Calibration Points for Ranges								
Flow Meter ID's: n/a	<table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>780</td></tr> <tr><td>Mid</td><td>380</td></tr> <tr><td>Low</td><td>190</td></tr> </table>	Point	ppb	High	780	Mid	380	Low	190
Point	ppb								
High	780								
Mid	380								
Low	190								
Make & Model: API 700									
Serial #: 627									
Cal Gas Cylinder I.D. #: LL104222									
Cal Gas Conc. (ppm): 50.6									

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	(ppb)	(ppb)	
adjusted zero	5000	0.00	5000	0.0	0.0	n/a
adjusted high	4924	77.10	5001	780.1	780.0	1.000
mid	4965	37.50	5003	379.3	378.0	1.003
low	4982	18.80	5001	190.2	188.0	1.012
calibrator zero	5000	0.00	5000	0.0	0.0	n/a
Average C.F. =						1.005

Linear Regression/Calibration Results:


Correlation Coefficient = 1.000	LIMITS
Slope = 0.999	> or = 0.995
b (Intercept as % of full scale) = 0.11%	.95-1.05
% change in C.F. from last cal = n/a	± 3% F.S.
	± 10%



As found:	As left:
SLOPE: n/a	SLOPE: 0.973
OFFSET: n/a	OFFSET: 127.2
HVPS: n/a	HVPS: 483
RCELL TEMP: n/a	RCELL TEMP: 50.0
BOX TEMP: n/a	BOX TEMP: 31.2
PMT TEMP: n/a	PMT TEMP: 7.7
IZS TEMP: n/a	IZS TEMP: 45.0
PRES: n/a	PRES: 24.5
SAMP FL: n/a	SAMP FL: 579
NORM PMT: n/a	NORM PMT: 127.4
UV LAMP: n/a	UV LAMP: 2745.9
LAMP RATIO: n/a	LAMP RATIO: 100.3
STR. LGT: n/a	STR. LGT: 61.9
DRK PMT: n/a	DRK PMT: 11.3
DRK LMP: n/a	DRK LMP: -0.5
Expected Value: n/a	Expected Value: 346.0

Comments:
The analyzer sample inlet filter was changed.

Post-repair calibration completed after annual maintenance. Zero Air scrubber repositioned inside. Valves tested for leaks. Factory calibration completed.



— SO2[ppb]

HYDROGEN SULPHIDE



API 101A Hydrogen Sulphide Analyzer Calibration

Date: <u>March 5, 2017</u>	Barometric Pressure: <u>0.920 atm</u>
Company/Airshed: <u>LICA</u>	Station Temperature °C: <u>22</u>
Location/Station Name: <u>Maskwa</u>	Weather Conditions: <u>Mainly cloudy with snow</u>
Parameter: <u>Hydrogen Sulphide</u>	Calibration Purpose: <u>routine monthly</u>
Start Time 24 hr. (mst): <u>11:12</u>	Performed By/Reviewer: <u>Alex Yakupov</u> / <u>Trina Whitsitt</u>
End Time 24 hr. (mst): <u>15:58</u>	Cal Gas Expiry Date: <u>June 14, 2019</u>
Calibration Method: <u>Gas Dilution</u>	Converter Model & s/n (if applicable): <u>n/a</u>

Analyzer:	
ID# or Serial Number: <u>324</u>	Range ppb: <u>100</u>
Last Calibration Date: <u>February 15, 2017</u>	As Found C.F.: <u>1.014</u>
Previous C.F.: <u>0.999</u>	New C.F.: <u>0.999</u>

Calibrator:	Standard Calibration Points for Ranges	SO ₂ Scrubber Check (10 mins.)								
Flow Meter ID's: <u>n/a</u>	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19	Start/End Time 24 hr.: <u>12:25 / 12:35</u>
Point	ppb									
High	78									
Mid	38									
Low	19									
Make & Model: <u>SABIO 2010 D</u>		Target Concentration (ppb): <u>780</u>								
Serial #: <u>11900613</u>		Result (ppb): <u>0</u>								
Cal Gas Cylinder I.D. #: <u>EY0000654</u>		Zero Corrected Result (ppb): <u>0</u>								
Cal Gas Conc. (ppm): <u>10.2</u>										


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	(ppb)	(ppb)	
as found zero	7500	0.00	7500	0.0	-0.6	n/a
as found high	7443	57.40	7500	78.1	76.4	1.014
adjusted zero	7500	0.00	7500	0.0	0.0	n/a
adjusted high	7443	57.40	7500	78.1	78.1	0.999
mid	7471	27.90	7499	37.9	38.1	0.996
low	7485	14.00	7499	19.0	19.1	0.997
calibrator zero	7500	0.00	7500	0.0	0.0	n/a
Average C.F.=						0.998

Linear Regression/Calibration Results:

Correlation Coefficient = <u>1.000</u>	LIMITS
Slope = <u>1.000</u>	> or = 0.995
b (Intercept as % of full scale) = <u>-0.05%</u>	.95-1.05
% change in C.F. from last cal = <u>-1.48%</u>	± 3% F.S.
	± 10%


API 101A Hydrogen Sulphide Analyzer Calibration



<p style="text-align: center;">As found:</p> SLOPE: <u>0.944</u> OFFSET: <u>25.3</u> HVPS: <u>676</u> DCPS: <u>2582</u> RCCELL TEMP: <u>50.1</u> BOX TEMP: <u>28.7</u> PMT TEMP: <u>6.8</u> IZS TEMP: <u>50.0</u> Converter Temp: <u>324.2</u> PRES: <u>23.7</u> SAMP FL: <u>566</u> UV LAMP: <u>3558.3</u> LAMP RATIO: <u>109.0</u> STR. LGT: <u>12.0</u> DRK PMT: <u>51.3</u> Expected Value: <u>47.7</u>	<p style="text-align: center;">As left:</p> SLOPE: <u>0.953</u> OFFSET: <u>34.0</u> HVPS: <u>677</u> DCPS: <u>2578</u> RCCELL TEMP: <u>50.6</u> BOX TEMP: <u>30.1</u> PMT TEMP: <u>6.8</u> IZS TEMP: <u>50.1</u> Converter Temp: <u>324.0</u> PRES: <u>23.7</u> SAMP FL: <u>566</u> UV LAMP: <u>3843.5</u> LAMP RATIO: <u>108.9</u> STR. LGT: <u>11.4</u> DRK PMT: <u>36.0</u> Expected Value: <u>46.6</u>
---	--

Comments:

The analyzer sample inlet filter was changed.



— H2S[ppb]

TOTAL HYDROCARBON



Thermo 51C Total Hydrocarbon Analyzer Calibration

Date:	March 5, 2017	Barometric Pressure:	0.920 atm
Company/Airshed:	LICA	Station Temperature °C:	22
Location/Station Name:	Maskwa	Weather Conditions:	Mainly cloudy with snow
Parameter:	Total Hydrocarbon	Calibration Purpose:	shut down
Start/End Time 24 hr. (mst):	9:22 / 11:14	Performed By/Reviewer:	Alex Yakupov / Trina Whitsitt
Calibration Method:	Gas Dilution	Cal Gas Expiry Date:	November 25, 2023

Analyzer:	ID# or Serial Number:	436609738	Range ppm:	50
	Last Calibration Date:	February 15, 2017	As Found C.F.:	1.003
	Previous Cal High Point C.F.:	1.000	New C.F.:	n/a


Calibrator:	Flow Meter ID's:	n/a	Standard Calibration Points for a Range of: 50 ppm	
	Make & Model:	SABIO 2010 D		
	Serial #:	11900613		
	Cal Gas Cylinder I.D. #:	LL165372		
	CH ₄ /C ₃ H ₈ Cylinder Conc. (ppm):	606.0 212.0	Point	Target ppm
	CH ₄ as propane/total CH ₄ equivalents (ppm):	583.0 1189.0	High	38
			Mid	18
			Low	9

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors:
Point	Diluent	Cal Gas	Total	(ppm)	(ppm)	
as found zero	1999	0.00	1999	0.0	0.00	n/a
as found high	1937	63.90	2001	37.97	37.86	1.003
mid	1970	30.30	2000	18.01	17.74	1.015
low	1985	15.10	2000	8.98	8.81	1.019
Average C.F. =						1.012

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	> or = 0.995
Slope =	1.002	.95-1.05
b (Intercept as % of full scale) =	0.21%	± 3% F.S.
% change in C.F. from last cal =	-0.29%	± 10%



As found:	As left:
H2 cylinder (psi): 800	H2 cylinder (psi): n/a
H2 cylinder reg set (psi): 22	H2 cylinder reg set (psi): n/a
Span Cylinder (psi): 700	Span Cylinder (psi): n/a
Span Cylinder Reg Set (psi): 22	Span Cylinder Reg Set (psi): n/a
Zero Air Gen Pressure: 38	Zero Air Gen Pressure: n/a
measurement alarms: None	measurement alarms: n/a
service alarms: None	service alarms: n/a
cnt: 1079	cnt: n/a
rng: 1	rng: n/a
try: 2	try: n/a
flm: 186.5	flm: n/a
det: 125.9	det: n/a
Flame: 186	Flame: n/a
Filter: 125	Filter: n/a
Base: 125	Base: n/a
Sample psi: 07.49	Sample psi: n/a
Internal Air Pressure: 20	Internal Air Pressure: n/a
Internal Fuel Pressure: 12	Internal Fuel Pressure: n/a
Measured Flow: n/a	Measured Flow: n/a
Expected Value: 26.51	Expected Value: n/a

Comments:

Shutdown calibration completed for annual maintenance. Pump rebuilt. Zero air pressure controller sensor adjusted.



Thermo 51C Total Hydrocarbon Analyzer Calibration

Date:	March 5, 2017	Barometric Pressure:	0.920 atm
Company/Airshed:	LICA	Station Temperature °C:	22
Location/Station Name:	Maskwa	Weather Conditions:	Mainly cloudy with snow
Parameter:	Total Hydrocarbon	Calibration Purpose:	post repair
Start/End Time 24 hr. (mst):	15:19 / 18:33	Performed By/Reviewer:	Alex Yakupov Trina Whitsitt
Calibration Method:	Gas Dilution	Cal Gas Expiry Date:	November 25, 2023

Analyzer:	ID# or Serial Number:	436609738	Range ppm:	50
	Last Calibration Date:	n/a	As Found C.F.:	n/a
	Previous Cal High Point C.F.:	n/a	New C.F.:	1.000

Calibrator:	Flow Meter ID's:	n/a	Standard Calibration Points for a Range of:	50 ppm
	Make & Model:	SABIO 2010 D		
	Serial #:	11900613		
	Cal Gas Cylinder I.D. #:	LL165372		
	CH ₄ /C ₂ H ₆ Cylinder Conc. (ppm):	606.0 212.0		
	CH ₄ as propane/total CH ₄ equivalents (ppm):	583.0 1189.0		


Point	Target ppm
High	38
Mid	18
Low	9

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated Concentration:	Indicated Concentration:	Correction Factors:
	Diluent	Cal Gas	Total	(ppm)	(ppm)	
adjusted zero	1999	0.00	1999	0.0	0.00	n/a
adjusted high	1937	63.90	2001	37.97	37.99	1.000
mid	1971	30.30	2001	18.00	17.74	1.015
low	1985	15.10	2000	8.98	8.78	1.022
calibrator zero	1999	0.00	1999	0.00	0.00	n/a
Average C.F.=						1.012

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	0.998		.95-1.05
b (Intercept as % of full scale)=	0.28%		± 3% F.S.
% change in C.F. from last cal=	n/a		± 10%




<p>As found:</p> <p>H2 cylinder (psi): n/a</p> <p>H2 cylinder reg set (psi): n/a</p> <p>Span Cylinder (psi): n/a</p> <p>Span Cylinder Reg Set (psi): n/a</p> <p>Zero Air Gen Pressure: n/a</p> <p>measurement alarms: n/a</p> <p>service alarms: n/a</p> <p>cnt: n/a</p> <p>rng: n/a</p> <p>try: n/a</p> <p>flm: n/a</p> <p>det: n/a</p> <p>Flame: n/a</p> <p>Filter: n/a</p> <p>Base: n/a</p> <p>Sample psi: n/a</p> <p>Internal Air Pressure: n/a</p> <p>Internal Fuel Pressure: n/a</p> <p>Measured Flow: n/a</p> <p>Expected Value: n/a</p>	<p>As left:</p> <p>H2 cylinder (psi): 800</p> <p>H2 cylinder reg set (psi): 22</p> <p>Span Cylinder (psi): 700</p> <p>Span Cylinder Reg Set (psi): 22</p> <p>Zero Air Gen Pressure: 36</p> <p>measurement alarms: None</p> <p>service alarms: None</p> <p>cnt: 1041</p> <p>rng: 1</p> <p>try: 3</p> <p>flm: 184.5</p> <p>det: 125.9</p> <p>Flame: 184</p> <p>Filter: 125</p> <p>Base: 125</p> <p>Sample psi: 07.51</p> <p>Internal Air: 20</p> <p>Internal Fuel: 12</p> <p>Measured Flow: n/a</p> <p>Expected Value: 26.80</p>
--	---

Comments:
 The analyzer sample inlet filter was changed.

The analyzer cooling fan filter(s) were cleaned.

Post-repair calibration completed after annual maintenance. Sample pump rebuilt. Sample inlet tubing purged. Zero Air svciced, cut-in/out valves adjusted.



— THC[ppm]

NITROGEN DIOXIDE



API 200A NO-NO2-NOx Analyzer Calibration

Date: March 5, 2017 Company/Airshed: LICA Location/Station Name: Maskwa Start/End Time 24 hr. (mst): 9:22 / 17:09 G.P.T. to be used for Ozone? No Calibration Method: Gas Dilution & Gas Phase Titration	Barometric Pressure: 0.920 atm Station Temperature °C: 22 Weather Conditions: Mainly cloudy with snow Calibration Purpose: routine monthly Performed By/Reviewer: Alex Yakupov Trina Whitsitt Cal Gas Expiry Date: July 18, 2019
---	---

Analyzer: ID# or Serial Number: 1899 Last Calibration Date: February 14, 2017 Range ppb: 1000	Correction Factors: <table border="1" style="width: 100%; text-align: center;"> <tr> <td></td> <td>Previous C.F.:</td> <td>As Found C.F.:</td> <td>New C.F.:</td> </tr> <tr> <td>NO =</td> <td>0.998</td> <td>0.992</td> <td>1.001</td> </tr> <tr> <td>NO₂ =</td> <td>1.000</td> <td>0.998</td> <td>0.998</td> </tr> <tr> <td>NOx =</td> <td>0.998</td> <td>0.993</td> <td>1.001</td> </tr> </table>		Previous C.F.:	As Found C.F.:	New C.F.:	NO =	0.998	0.992	1.001	NO ₂ =	1.000	0.998	0.998	NOx =	0.998	0.993	1.001
	Previous C.F.:	As Found C.F.:	New C.F.:														
NO =	0.998	0.992	1.001														
NO ₂ =	1.000	0.998	0.998														
NOx =	0.998	0.993	1.001														

Calibrator: Flow Meter ID's: n/a Make & Model: API 700 Serial #: 627 Cal Gas Cylinder I.D. #: LL104222 NO/NOx Gas Conc. (ppm): 50.7 50.7	Standard Calibration Points for a Range of: 1000 ppb <table border="1" style="width: 100%; text-align: center;"> <tr> <th>Point</th> <th>Target NO (ppb)</th> <th>Target NO₂ (ppb)</th> <th>Cc Ozone ?</th> </tr> <tr> <td>High</td> <td>780</td> <td>500</td> <td>n/a</td> </tr> <tr> <td>Mid</td> <td>380</td> <td>275</td> <td>n/a</td> </tr> <tr> <td>Low</td> <td>190</td> <td>100</td> <td>n/a</td> </tr> <tr> <td>Extra Point #1</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> <tr> <td>Extra Point #2</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> </table>	Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?	High	780	500	n/a	Mid	380	275	n/a	Low	190	100	n/a	Extra Point #1	n/a	n/a	n/a	Extra Point #2	n/a	n/a	n/a
Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?																						
High	780	500	n/a																						
Mid	380	275	n/a																						
Low	190	100	n/a																						
Extra Point #1	n/a	n/a	n/a																						
Extra Point #2	n/a	n/a	n/a																						

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)			Calculated NO	Calculated NOx	Indicated NO	Indicated NOx	NO C.F.	NOx C.F.
Point	Diluent	Cal Gas	Total Flow	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)
as found zero	5000	0.0	5000	0	0	0.0	1.0	n/a
as found high	4923	76.9	5000	779.8	779.8	786.0	786.0	0.992
adjusted zero	5000	0.00	5000	0.0	0.0	0.0	0.0	n/a
adjusted high	4923	76.90	5000	779.8	779.8	779.0	779.0	1.001
mid	4964	37.50	5002	380.1	380.1	372.0	372.0	1.022
low	4980	18.70	4999	189.7	189.7	183.0	183.0	1.036
calibrator zero	5000	0.00	5000	0	0	0.0	0.0	n/a
							Average C.F.=	1.020

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calibrator Setting	Indicated NO	Indicated NOx	Indicated NO ₂	NO drop	NO ₂ gain	NO ₂ C.F.
Point	Diluent	Cal Gas	Total Flow	volts or ppb	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)
NOx reference	4923	76.90	5000	0.0	779.0	780.0	780.0	1.0	0.0	1.0
as found high NO ₂	4923	76.90	5000	480.0	280.0	781.0	501.0	499.0	500.0	0.998
adjusted high NO ₂	4923	76.90	5000	480.0	280.0	781.0	501.0	499.0	500.0	0.998
gpt mid	4923	76.90	5000	260.0	504.0	783.0	279.0	275.0	278.0	0.989
gpt low	4923	76.90	5000	90.0	680.0	780.0	100.0	99.0	99.0	1.000
									Average NO ₂ C.F.=	0.996

Linear Regression/Calibration Results:

	NO	NOx	NO ₂	LIMITS
Correlation Coefficient =	1.000	1.000	1.000	> or = 0.995
Slope =	0.999	0.999	0.998	.95-1.05
b (Intercept as % of full scale)=	-0.42%	-0.42%	0.09%	± 3% F.S.
% change in C.F. from last cal=	0.59%	0.47%	0.20%	± 10%
NO ₂ converter efficiency	0.99		0.99	0.96 to 1.04




As found: NOx SLOPE: 0.913 NOx OFFS: -0.1 NO SLOPE: -0.931 NO OFFS: -0.7 SAMP FLW: 549 OZONE FL: 78 NORM PMT: 2.5 AZERO: 21.1 HVPS: 686 DCPS: 2582 RCELL: 50.7 BOX TEMP: 28.5 IZS TEMP: 48.0 MOLY TEMP: 314.9 RCEL: 5.5 SAMP: 25.7 Expected Value NO: 4.5 Expected Value NO ₂ : 490.0 Expected Value NOx: 494.0	As left: NOx SLOPE: 0.908 NOx OFFS: 0.3 NO SLOPE: 0.924 NO OFFS: -1.1 SAMP FLW: 546 OZONE FL: 77 NORM PMT: -1.9 AZERO: 21.6 HVPS: 686 DCPS: 2573 RCELL: 49.9 BOX TEMP: 31.4 IZS TEMP: 48.4 MOLY TEMP: 314.0 RCEL: 5.5 SAMP: 25.6 Expected Value NO: 4.5 Expected Value NO ₂ : 487.0 Expected Value NOx: 491.0
--	--


Comments:
 The analyzer sample inlet filter was changed. No high point NO₂ adjustment was required/made. As found values were copied to adjusted high values for linearity calculation purposes.

At 10:53, after "As Found - High Point" in AUTO mode, the GPT mode with O3=0.0 ppb was tested. No High Point adjustment was made so far. "As Found High" started again at 11:59.

Date: March 5, 2017
Company/Airshed: LICA
Location/Station Name: Maskwa

Start/End Time 24 hr. (mst): 9:22 / 17:09
Calibration Purpose: routine monthly
Calibration Method: Gas Dilution & Gas Phase Titration





— NOX[ppb] — NO[ppb] — NO2[ppb]

WIND SYSTEM



Met One Instruments

Sonic Wind Sensor Certificate of Calibration

Sensor Model No.: 50.5H Sensor Serial No.: H10703
 Sensor Output Swing 0V - 1.0V Sensor Output Range: 0 - 50.0 MPS
 Customer: Maxxam Analytics Sales Order No.: 115035
 Tested per PO: 35-62828 Calibration Date: 03/30/2016
 Calibrated by: David Frith D7

QC Inspection Byron Dawson

Instrument Condition Within Tolerance: As Found As Left X
 Corrective Action: No Adjustment Adjust X Repair
 Preventative Maintenance

As Found Test Date: N/A As Left Test Date 03/30/2016

Quality Control Manual Revision: September 16, 2013 MP42201 Rev. G.

All Work Performed per Customer Purchase Order Requirements.

Calibration Document No. 50.5-6100

Test Equipment Used for Calibration of Instruments

Description	Manufacturer	Model No.	Serial No.	Cal Date	Cal Due	Voltage Accuracy	Time Base Accuracy
Data Acquisition	Campbell Scientific	CR1000	6569	4/06/2015	4/06/2018	+/- 3mV	< 6 ppm
NIST Cupset	Met One Instruments	170-41	3309	4/24/2012	4/24/2017	Accuracy < 0.15 mph or 1% WS	

Environmental Data: Temperature 65 to 80 Deg F Vibration none

Humidity 20 to 70%

Radiation none

The standards used for calibration have accuracies equal to or greater than the instruments tested. These standards are on record and are traceable to NIST to the extent allowed by the institute's calibration facility. Unless otherwise stated heron, all instruments are calibrated to meet the manufacturer's published specifications. The calibration system complies with MIL-STD-45662A (8/1/88). Instrument's accuracy meets the requirements of Regulatory Guide 1.23 (2/72). Compliant with IS) 9001:2008 requirements

METEOROLOGICAL SYSTEM CHECK



**Calibration complies with ISO/IEC
17025, ANSI/NCSL Z540-1, and 9001**



Cert. No.: 4085-7568459

Traceable® Certificate of Calibration for Digital Hygrometer

Manufactured for and distributed by: Fisher Scientific, 300 Industry Drive, Pittsburgh, PA 15275-1001

Instrument Identification:

Model Numbers: 11-661-7B, 11755843 S/N: 160348895 Manufacturer: Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Digital Thermometer	122044330	11/20/16	4000-7226968
Chilled Mirror Hygrometer	31874/H2048MCR	10/12/16	13366

Certificate Information:

Technician: 353 Procedure: CAL-18 Cal Date: 4/08/16 Due Date: 4/08/18
 Test Conditions: 26.4°C 37.0 %RH 1021 mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
%RH		N.A.		20.21	20.21	Y	18.71	21.71	0.46	3.3:1
%RH		N.A.		50.59	50.52	Y	49.09	52.09	0.46	3.3:1
%RH		N.A.		75.06	75.10	Y	73.56	76.56	0.46	3.3:1
°C		N.A.		27.490	27.29	Y	27.09	27.89	0.059	>4:1

This Instrument was calibrated using Instruments Traceable to National Institute of Standards and Technology.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Min = As Left Nominal(Rounded) - Tolerance; Max = As Left Nominal(Rounded) + Tolerance; Date=MM/DD/YY

Nicol Rodriguez
Nicol Rodriguez, Quality Manager

Aaron Justice
Aaron Justice, Technical Manager

Maintaining Accuracy:

In our opinion once calibrated your Digital Hygrometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Hygrometers change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 4455 Rex Road Friendswood, TX 77546 USA
Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com

Control Company is an ISO 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
 Control Company is ISO 9001:2008 Quality Certified by (DNV) Det Norske Veritas, Certificate No. CERT-01805-2006-AQ-HOU-RvA.
 International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA).

CALIBRATORS

Company Maxxam/SIA **Operator:** Chris

Calibrator:		Flow Measurement Device:	
Make/Model	<u>API 700</u>	Make/Model	<u>Definer 530</u>
Serial Number	<u>627</u>	Serial Number	<u>H-148944, L-152019</u>
Last Verification Date	<u>February 3, 2016</u>	Temperature (°C)	<u>23.5</u>
NO Cylinder S/N	<u>EY0000597</u>	Barometric Pressure	<u>707.1 mmHg</u>
NO [PPM]	<u>49.0</u>	NOx [PPM]	<u>49.0</u>
Expiry Date	<u>December 8, 2019</u>		

Dilution Flow (sccm)		
Pt. #1	<u>4892</u>	Pt. #3 <u>4951</u>
Pt. #2	<u>4975</u>	
Gas Flow (sccm)		
Pt. #1	<u>79.7</u>	Pt. #3 <u>19.4</u>
Pt. #2	<u>38.8</u>	

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
	0.0	0.0000	0.0000	0.0000	-0.0004	-0.0004	Limit ± 10%	
4972	79.7	0.7855	0.7855	0.7883	0.0004	0.7887	0.4%	0.5%
4936	38.8	0.3822	0.3822	0.3816	0.0005	0.3822	-0.2%	0.1%
4970	19.4	0.1913	0.1913	0.1902	0.0006	0.1913	-0.6%	0.2%
Absolute Average Percent Difference							0.1%	0.3%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO	LIMITS	NOx
Correlation= 1.0000	≥ 0.990	Correlation= 1.0000
m (Slope)= 1.0041	0.90-1.10	m (Slope)= 1.0046
b (Intercept % of FS)= -0.1118	± 3% F.S.	b (Intercept % of FS)= -0.0871

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4972	0	0.0000	0.7867	0.0014	0.7881	NO ₂	% Diff, Limit
4972	500	0.5127	0.2740	0.5104	0.7849	-0.7%	± 10%
4972	275	0.2863	0.5004	0.2860	0.7865	-0.6%	± 10%
4972	90	0.0940	0.6927	0.0954	0.7880	0.0%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO₂	LIMITS
Correlation= 1.0000	≥ 0.995
m (Slope)= 0.9924	0.90-1.10
b (Intercept % of FS)= 0.1755	± 3% F.S.

AENV Standards	NO_x Analyzer
Audit Calibrator	Make/Model <u>Thermo 42i</u>
Make/Model <u>Thermo 146i</u>	Serial/AMU Number <u>AMU 1868</u>
Serial/AMU Number <u>AMU1809</u>	Last Calibration Date <u>January 25, 2017</u>
SRM Gas Cylinder No. <u>CAL018140</u>	Full Scale (ppm) <u>1.0</u>
Cylinder Conc. (ppm) <u>48.79</u>	Cylinder Gas Expiry Date <u>March 25, 2019</u>

COMMENTS:

Auditor: Shea Beaton
Operator Signature: 

Date: January 27, 2017
Location: McIntyre Center Edmonton

Company <u>Maxxam</u>		Operator: <u>Mike</u>	
Calibrator:		Flow Measurement Device:	
Make/Model	<u>Sabio 2010D</u>	Make/Model	<u>Bios Defender 530</u>
Serial Number	<u>11900613</u>	Serial Number	<u>HI148944 Lo 152019</u>
Last Verification Date	<u>March 31, 2016</u>	Temperature (°C)	<u>23.9</u>
NO Cylinder S/N	<u>EY0000769</u>	Barometric Pressure	<u>698mmHg</u>
NO [PPM]	<u>51.1</u> NOx [PPM]		<u>51.2</u>
Expiry Date	<u>December 8, 2019</u>		

Dilution Flow (sccm)		
Pt. #1 <u>4879</u>	Pt. #2 <u>4932</u>	Pt. #3 <u>4950</u>
Gas Flow (sccm)		
Pt. #1 <u>74.5</u>	Pt. #2 <u>36.4</u>	Pt. #3 <u>18.2</u>

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
4965	0.0	0.0000	0.0000	0.0001	0.0000	0.0001	Limit ± 10%	
4954	74.5	0.7685	0.7700	0.7915	0.0008	0.7923	3%	3%
4968	36.4	0.3744	0.3751	0.3832	0.0006	0.3838	2%	2%
4968	18.2	0.1872	0.1876	0.1916	0.0002	0.1918	2%	2%
Absolute Average Percent Difference							3%	2%


LINEAR REGRESSION ANALYSIS			<i>y=mx+b (where x=calculated concentration, y=indicated concentration)</i>		
NO		LIMITS		NOx	
Correlation=	1.0000	≥ 0.990		Correlation=	1.0000
m (Slope)=	1.0301	0.90-1.10		m (Slope)=	1.0291
b (Intercept % of FS)=	-0.0919	± 3% F.S.		b (Intercept % of FS)=	-0.0881

Flow	O ₂ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4954	0.000	0.0000	0.7949	0.0005	0.7954	NO ₂	% Diff. Limit
4954	0.510	0.5104	0.2845	0.5072	0.7917	-1%	± 10%
4954	0.250	0.2516	0.5433	0.2514	0.7944	0%	± 10%
4954	0.100	0.1085	0.6864	0.1087	0.7951	0%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS			<i>y=mx+b (where x=calculated concentration, y=indicated concentration)</i>		
NO₂		LIMITS			
Correlation=	1.0000	≥ 0.995			
m (Slope)=	0.9926	0.90-1.10			
b (Intercept % of FS)=	0.0925	± 3% F.S.			

AENV Standards		NO_x Analyzer	
Audit Calibrator			
Make/Model	<u>Thermo 146i</u>	Make/Model	<u>Thermo 42i</u>
Serial/AMU Number	<u>1809</u>	Serial/AMU Number	<u>1868</u>
SRM Gas Cylinder No.	<u>CAL018140</u>	Last Calibration Date	<u>March 15, 2017</u>
Cylinder Conc. (ppm)	<u>48.79</u>	Full Scale (ppm)	<u>1.0</u>
		Cylinder Gas Expiry Date	<u>March 28, 2019</u>

COMMENTS: Gas has ~50ppm SO2

Auditor: Shea Beaton
Operator Signature: 

Date: March 16, 2017
Location: McIntyre Center Edmonton

CALIBRATION GASES



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2016-335CGA

Company: Maxxam **Operator's Name:** Russell Kirchner

Cylinder #: LL104222 Concentration PPM: 50.6 Tolerance(%) 1 Certified By: Praxair

Expiry Date: July 2019

Reference Calibrator and Gas:	Flow Measurement Device:
Make/Model: <u>R&R MFC 201</u>	Make/Model: <u>Bios DC2</u>
Serial Number: <u>AMU 1690</u>	Serial Number: <u>AMY 1659</u>
Last Verification Date: <u>October 19, 2016</u>	Temp. °C: <u>24.5 C</u>
Gas Type: <u>SO2</u> Conc. <u>98.07</u>	B.P. <u>706 mmhg</u>
Cylinder Number: <u>CA:016625</u>	
Expiry Date: <u>January 2019</u>	

Reference Analyzer:

Make/Model: Teco 43C Serial/AMU Number: 1623

Instrument Settings: Zero: 9.2 Span: 1.024 Range: 1.0

Last Calibration: Date: Oct 19/16 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.000	0.000	0.000	0.000
4935	82.0	0.830	0.01662	60.183	50.0
4968	40.8	0.412	0.00821	121.765	50.2
4955	20.2	0.203	0.00408	245.297	49.8
Average Cylinder Concentration:					50.0

Previous Stated Concentration PPM: 50.6

Percent variance from Stated: 1

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____

< =5% Outside Manufacturer Tolerance. Use manufacturers concentration _____

> 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark Date: October 19, 2016

Operator Signature: *Al Clark* Location: McIntyre Center Edmonton



Calibration Gas Audit

CH4 / C3H8 Cylinder Gas

File No. 2015-092CGA

Company: Maxxam Operators name: Chris Wesson
Cylinder #: LL165372 Conc CH4 (PPM) 606/212 Tolerance (%) 0.5 Certified By: Praxair

Reference Calibrator and Gas:

Make/Model R&R MFC 201
Serial Number AMU 1698
Last Verification Date January 18, 2016
Gas Type CH4 Conc. 999.2
Cylinder Number D751932
Gas Type C3H8 Conc. 246.5
Cylinder Number XF0037998

Flow Measurement Device:

Make/Model Bios DC-2
Serial Number Blos D
Temp. °C 24.5
B.P. 688mmHg

Reference Analyzer:

Make/Model Thermo 55C Serial/AMU Number: 1643
Instrument Settings Zero: NA Span: NA Range: 20.0
Last Calibration: Date: 18-Jan-16 C.F. 1.000 Done By: SB

Calibrator Flows (scem)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	CH4	C3H8			CH4	C3H8
2568	0.00	0.00	0.00	0.02140	46.722	607	214
2630	56.29	12.99	12.62	0.02140	46.722	607	214
2588	19.73	4.62	4.50	0.00762	131.171	606	215
2580	9.69	2.29	2.24	0.00376	266.254	610	217
Average Cylinder Concentration:						608	215

<u>CH4</u>	<u>C3H8</u>
Previous Stated Concentration PPM: <u>606</u>	<u>212</u>
Percent variance from Stated: <u>0.3</u>	<u>1.6</u>

Cylinder gas tolerances based on CH4 only

Meets Manufacturer Tolerance. Use manufacturers stated concentration COMMENTS: _____
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration C3H8 manufacturers tolerance 1.1%
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Shea Beaton Date: January 19, 2016
Operator Signature: _____ Location: McIntyre Center Edmonton



Calibration Gas Audit

NO Cylinder Gas

File No. 2016-336CGA

Company: Maxxam **Operators name:** Russell Kirchner

Cylinder #: LL104222 Conc (PPM) 50.7/50.9 Tolerance (%) 1 Certified By: Praxair

Expiry Date: July 2019

Reference Calibrator and Gas:				Flow Measurement Device:	
Make/Model	<u>Teco 146i</u>			Make/Model	<u>Bios DC2</u>
Serial Number	<u>AMU 1809</u>			Serial Number	<u>AMU 1659</u>
Last Verification Date	<u>October 19, 2019</u>			Temp. °C	<u>24.5 C</u>
Gas Type	<u>NO</u>	Conc.	<u>48.79</u>	B.P.	<u>706 mmhg</u>
Cylinder Number	<u>CAL018188</u>				
Expiry Date	<u>March 2019</u>				

Reference Analyzer:

Make/Model Teco 42i Serial/AMU Number: 1868

Instrument Settings Zero: 4.4 Span: 1.080 Range: 1.0

Last Calibration: Date: Oct 18/16 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	NO	NOX			NO	NOX
5000	0.0	0.000	0.000				
4935	82.0	0.838	0.837	0.017	60.183	50.4	50.4
4968	40.8	0.417	0.417	0.008	121.765	50.8	50.8
4955	20.2	0.207	0.207	0.004	245.297	50.8	50.8
Average Cylinder Concentration:						50.7	50.6

<u>NO</u>	<u>NOx</u>
Previous Stated Concentration PPM: <u>50.7</u>	<u>50.9</u>
Percent variance from Stated: <u>0</u>	<u>1</u>

Cylinder gas tolerances based on NO only

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:**

< =5% Outside Manufacturer Tolerance. Use manufacturers concentration Contains 50.6 ppm SO2.

> 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark Date: October 19, 2016

Operator Signature: *Al Clark* Location: McIntyre Center Edmonton

***APPENDIX III
REPORT CERTIFICATION FORM***

Report Certification Form

Alberta Airshed (if applicable)	EPA Approval or Code of Practice Registration # (if applicable)
YES	NA
Company Name (if applicable)	Industrial Operation Name (if applicable)
Lakeland Industry & Community Association	Maskwa Continuous Monitoring Station
Name of the Representative of the Person Responsible (Last, First, Middle)	Position / Title of the Representative of the Person Responsible
Bim Adeniji	Project Manager Assistant, Customer Service, Air Services
Is an External Party Certifying the Report? (If 'Yes', fill in the fields below for the external person.)	
<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	
Name of External Person Certifying the Report (Last, First, Middle)	Position / Title of External Person Certifying the Report
NA	NA
Company Name for the External Person Certifying the Report	Identification of Qualifications / Professional Designations of the External Person Certifying the Report
NA	NA

I certify that I have reviewed and verified the submitted report. I also certify that the report presented with this certification form is complete, accurate and representative of the monitoring results and timeframe.



Signature of the Representative of the Person Responsible / External Person Certifying the Report

09-05-2017





Report Issued Date (dd-mm-yyyy)

***APPENDIX IV
DATA VALIDATION CERTIFICATION FORM***



Validation Certificate Form

Client: <u>Lakeland Industry & Community Association</u>	Project #: <u>2833-2017-03-30-C</u>
Site: <u>Maskwa Continuous Monitoring Station</u>	Contact: <u>Mike Bisaga</u>

Level 0 Preliminary Verification	 _____	Date <u>29-Apr-2017</u>
Level 1 Primary Validation	 _____	Date <u>29-Apr-2017</u>
Level 2 Final Validation	 _____	Date <u>05-May-2017</u>
Level 3 Independent Data Review	 _____	Date <u>09-May-2017</u>
Post-Final Validation	NA _____	Date NA _____

Notes
The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. This validation is performed on an annual basis.



Alberta Environment and Parks (AEP)
Air.Reporting@gov.ab.ca

February 22, 2018

Subject: Monthly Report Submission for the LICA St. Lina station

Lakeland Industry & Community Association (LICA) is pleased to submit the ambient air monitoring monthly report for the LICA St. Lina AQM Station in the month of March 2017.

The air monitoring program consists of continuous air monitoring results for Sulphur Dioxide (SO₂), Hydrogen Sulphide (H₂S), Total Hydrocarbon (THC), Oxides of Nitrogen (NO_x), Nitric Oxides (NO), Nitrogen Dioxide (NO₂), Ozone (O₃), Particulate Matter 2.5 (PM_{2.5}), Relative Humidity (RH), Barometric Pressure (BP), Precipitation, Ambient Temperature (AmbTPX), Wind Speed (WS), Wind Direction (WD) and Standard Deviation Wind Direction (STDWD).

Sampling Program	Monitoring Activities Conducted By	Sample Analysis Conducted By	Data/Report Review and Prepared By	Electronic Submission Conducted By
Continuous ambient air	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics

All data collected in March 2017 was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement systems.

As the LICA Environmental Program Manager and Data & Reporting Specialist, we certify that we have reviewed and verified this report and that the information is complete, accurate and representative of the monitoring results, reporting timeframe and the specified analysis, summarization and reporting requirements. We also certify all air data that are required by the AMD to be electronically submitted to AEP and Alberta's Ambient Air Quality Data Warehouse have been submitted by the time of this report submission.

Should you have any questions, please don't hesitate to contact us.

Respectfully,



Lakeland Industry & Community Association
5107 50 St
Bonnyville, AB T9N 2J7

A handwritten signature in blue ink that reads "Michael Bisaga".

Michael Bisaga
Technical Program Manager
Lakeland Industry & Community Association
780-266-7068
mbisaga@otonabee.ca

A handwritten signature in blue ink that reads "Lily Lin".

Lily Lin
Data & Reporting Specialist
587-225-2248
rebbacaa@gmail.com



MAXXAM ANALYTICS
#1 2080 39 Ave. NE, Calgary, AB
T2E 6P7

maxxam.ca
Toll Free 800-386-7247
Fax 403-219-3673

**AMBIENT AIR MONITORING MONTHLY DATA REPORT
LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
ST. LINA CONTINUOUS MONITORING STATION**

JOB #: 2833-2017-03-31-C

March 2017

Prepared for:

LAKELAND INDUSTRY & COMMUNITY ASSOCIATION

402 - 19 ST NW
CALGARY, ALBERTA
T2N 2J1

Attention: MIKE BISAGA

DATE: **May 12, 2017**

Prepared by: *Maram Ghaleb*

Maram Ghaleb, B.Sc.
Project Manager, Customer Service, Air Services

Reviewed by: *Wunmi Adekanmbi*

Wunmi Adekanmbi, M.Sc., EPT.
Project Manager, Customer Service, Air Services

SUMMARY

In March 2017, Maxxam Analytics was contracted to manage the ambient air quality monitoring and maintenance activities at the St. Lina Continuous Monitoring Station, near Bonnyville, Alberta. The monitoring station provides continuous meteorological measurements and air quality data for non-compliance parameters, as requested by the Lakeland Industry and Community Association.

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

Annual Maintenance: A scheduled annual maintenance was completed on all gas analyzers between March 1 and March 2. Twenty-two, twenty, one, nineteen and twenty hours of downtime were recorded for SO₂, H₂S, THC, NO₂ and O₃ respectively, due this maintenance event.

All parameters: On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

SO₂: A "dark cal" alert was triggered on the analyzer on March 6. The analyzer was reset onsite and no further action was required. Two hours of anomalous data were, however, invalidated at 10:00-11:00 due to this event.

H₂S:

- Seven hours of downtime were recorded between March 6 and March 7 due to the additional quality checks performed to verify the stability of span response after the annual maintenance.
- Eight more hours of downtime were incurred between March 19 and March 23 due to quality check and maintenance performed to address a biased high span response.
- The zero air filter was re-installed on March 28 to address a zero drift. One hour of downtime was recorded due to this event.

NO_x/NO/NO₂: Seven hours of downtime were recorded on March 7 due to a repeat calibration, performed for quality assurance and diligence purposes.

NO_x calibration concentrations were calculated using a NO_x gas concentration of 50.7 rather than 50.9 ppm. This yielded incorrect values for Calculated NO_x that were presented on the Analyzer Calibration Form. The calibration concentrations were recalculated using the correct NO_x gas value (50.9 ppm) and the outcome was insignificant. The calibration is still deemed AMD complaint.

PM_{2.5}: Ten hours of data were recorded at concentrations lower than -3 µg/m³ this month, rendering the data invalid.

WS/WD: One hour of downtime due to maintenance was recorded on March 2 at hour 17:00.

The summary of results is presented on the following pages.

Any deviations or modifications made to the sampling or analytical methods are outlined in Section 1.0, Discussion. On this basis, Maxxam Analytics is issuing this completed report to Lakeland Industry & Community Association, St. Lina Continuous Monitoring Station.




Should you have any questions concerning the results or if we can be of further assistance, please contact us at 403-219-3677 or toll-free at 1-800-386-7247.

Monthly Continuous Data Summary

Lakeland Industry & Community Association						MAXIMUM VALUES							OPERATIONAL TIME (%)
St. Lina Continuous Monitoring Station						1-HOUR				24-HOUR			
PARAMETER	OBJECTIVES		EXCEEDANCES		MONTHLY AVERAGE	READING	DAY	HOUR	WIND SPEED (kph)	WIND DIRECTION (sector)	READING	DAY	
	1-hr	24-hr	1-hr	24-hr									
SO ₂ (ppb)	172	48	0	0	0.3	3.5	6	23	19.9	NE	1.0	7	96.5
H ₂ S (ppb)	10	3	0	0	0.1	0.5	18	20	10.2	NNE	0.3	22	94.9
THC (ppm)	-	-	-	-	2.13	2.74	2	15	17.7	NNE	2.41	2	99.6
NO ₂ (ppb)	159	-	0	-	2.0	12.4	10	6	13.5	W	4.7	9	96.2
NO (ppb)	-	-	-	-	0.2	3.2	31	8	22.1	ESE	0.8	31	96.2
NO _x (ppb)	-	-	-	-	2.2	12.6	10	6	13.5	W	5.4	9	96.2
O ₃ (ppb)	82	-	0	-	34.5	49.6	15	16	20	ENE	45.2	16	97.0
PM _{2.5} (µg/m ³)	80	30	0	0	4.5	19.5	26	0	15.1	N	13.2	26	98.4
RELATIVE HUMIDITY (%)	-	-	-	-	69	90	28	21	20.5	NNE	87	28	99.7
BAROMETRIC PRESSURE (millibar)	-	-	-	-	927	948	9	13	4.9	NW	945	9	99.7
AMBIENT TEMPERATURE (°C)	-	-	-	-	-7.3	9.9	15	14	9	NNE	3.2	31	99.7
PRECIPITATION (mm)	-	-	-	-	0.1	3.3	22	15	22.2	W	0.6	22	99.7
VECTOR WS (kph)	-	-	-	-	2.8	25.1	16	23	-	W	20.3	29	99.6
VECTOR WD (sec)	-	-	-	-	322 (NW)	-	-	-	-	-	-	-	99.6

March 2017 Monthly Report Summary

Pollutants		Monthly Records		1-Hour Records					24-Hour Records			
Name	Unit	Avg. Conc.	Uptime	Maximum			AAAQO Objective	Exceed. Hours	Maximum		AAAQO Objective	Exceed. Days
				Conc.	Date	Hour			Conc.	Date		
SO ₂	ppb	0.3	96.5%	3.5	March 6	23	172	0	1.0	March 7	48	0
H ₂ S	ppb	0.1	94.9%	0.5	March 18	20	10	0	0.3	March 22	3	0
THC	ppm	2.13	99.6%	2.74	March 2	15	-	-	2.41	March 2	-	-
NO _x	ppb	2.2	96.2%	12.6	March 10	6	-	-	5.4	March 9	-	-
NO	ppb	0.2	96.2%	3.2	March 31	8	-	-	0.8	March 31	-	-
NO ₂	ppb	2.0	96.2%	12.4	March 10	6	159	0	4.7	March 9	-	-
O ₃	ppb	34.5	97.0%	49.6	March 15	16	82	0	45.2	March 16	-	-
PM _{2.5}	µg/m ³	4.5	98.4%	19.5	March 26	0	80	0	13.2	March 26	30	0
WS	kph	2.8	99.6%	25.1	March 16	23	-	-	20.3	March 29	-	-
WD	degree	322 (NW)	99.6%	-	-	-	-	-	-	-	-	-
RH	%	69	99.7%	90	March 28	21	-	-	87	March 28	-	-
BP	mbar	927	99.7%	948	March 9	13	-	-	945	March 9	-	-
PRECIP	mm	0.1	99.7%	3.3	March 22	15	-	-	0.6	March 22	-	-
AmbTPX	°C	-7.3	99.7%	9.9	March 15	14	-	-	3.2	March 31	-	-



Monthly Update

- * All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.
- * All data collected this month were within the objectives outlined in the AMD 2016 and AAAQO 2016.
- * The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above 90%.

Operational Issues

Annual Maintenance: A scheduled annual maintenance was completed on all gas analyzers between March 1 and March 2. Twenty-two, twenty-one, nineteen and twenty hours of downtime were recorded for SO₂, H₂S, THC, NO₂ and O₃ respectively, due to this maintenance event.

All parameters: On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

SO₂: A "dark cal" alert was triggered on the analyzer on March 6. The analyzer was reset onsite and no further action was required. Two hours of anomalous data were, however, invalidated at 10:00-11:00 due to this event.

H₂S:

- Seven hours of downtime were recorded between March 6 and March 7 due to the additional quality checks performed to verify the stability of span response after the annual maintenance.
- Eight more hours of downtime were incurred between March 19 and March 23 due to quality check and maintenance performed to address a biased high span response.
- The zero air filter was re-installed on March 28 to address a zero drift. One hour of downtime was recorded due to this event.

NO_x: Seven hours of downtime were recorded on March 7 due to a repeat calibration, performed for quality assurance and diligence purposes.

NO_x calibration concentrations were calculated using a NO_x gas concentration of 50.7 rather than 50.9 ppm. This yielded incorrect values for Calculated NO_x that were presented on the Analyzer Calibration Form. The calibration concentrations were recalculated using the correct NO_x gas value (50.9 ppm) and the outcome was insignificant. The calibration is still deemed AMD compliant.

PM_{2.5}: Ten hours of data were recorded at concentrations lower than -3 µg/m³ this month, rendering the data invalid.

WS/WD: One hour of downtime due to maintenance was recorded on March 2 at hour 17:00.

Exceedance Summary Report

SO₂ 1-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 1-hour AAAQO of 172 ppb.

SO₂ 24-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 24-hour AAAQO of 48.0 ppb.

H₂S 1-Hour Exceedances

Measured concentrations of hydrogen sulphide were below the 1-hour AAAQO of 10 ppb.

H₂S 24-Hour Exceedances

Measured concentrations of hydrogen sulphide were below the 24-hour AAAQO of 3 ppb.

NO₂ 1-Hour Exceedances

Measured concentrations of nitrogen dioxide were below the 1-hour AAAQO of 159 ppb.

PM_{2.5} 1-Hour Exceedances

Measured concentrations of fine particulate matter were below the 1-hour AAAQO of 80 µg/m³.

PM_{2.5} 24-Hour Exceedances

Measured concentrations of fine particulate matter were below the 24-hour AAAQO of 30 µg/m³.

O₃ 1-Hour Exceedances

Measured concentrations of ozone were below the 1-hour AAAQO of 82 ppb.

In accordance with EPEA and the Substance Release Regulation.

In accordance with A Guide to Release Reporting and the Alberta Ambient Air Quality Objectives and Guidelines Summary.

TABLE OF CONTENTS

<u>Title</u>	<u>Page</u>
SUMMARY	1
MONTHLY CONTINUOUS DATA SUMMARY REPORT	2
DASHBOARD	3
EXCEEDANCE SUMMARY REPORT	4
TABLE OF CONTENTS	5
1.0 Discussion	6
2.0 Project Personnel	10
3.0 Plant Monthly Required AMD Summary	10
4.0 Calculations and Results	10
5.0 Methods and Procedures	11
Appendix I	Continuous Monitoring Data Results 14
	Sulphur Dioxide 15
	Hydrogen Sulphide 23
	Total Hydrocarbon 31
	Oxides of Nitrogen 39
	Nitric Oxides 47
	Nitrogen Dioxide 54
	Ozone 62
	Particulate Matter 2.5 70
	Wind Speed 75
	Wind Direction 82
	Standard Deviation Wind Direction 85
	Relative Humidity 88
	Barometric Pressure 91
	Ambient Temperature 94
	Precipitation 97
Appendix II	Equipment Calibration Results 100
	Sulphur Dioxide 101
	Hydrogen Sulphide 106
	Total Hydrocarbon 116
	Nitrogen Dioxide 120
	Ozone 131
	Particulate Matter 136
	Wind System 139
	Calibrators 141
	Calibration Gases 147
Appendix III	Report Certification Form 157
Appendix IV	Data Validation Certification Form 159

1.0 Discussion

This monthly report consists of continuous monitoring results for the following parameters: Sulphur Dioxide (SO₂), Hydrogen Sulphide (H₂S), Total Hydrocarbon (THC), Oxides of Nitrogen (NO_x), Nitric Oxides (NO), Nitrogen Dioxide (NO₂), Ozone (O₃), Particulate Matter 2.5 (PM_{2.5}), Relative Humidity (RH), Barometric Pressure (BP), Precipitation, Ambient Temperature (AmbTPX), Wind Speed (WS), Wind Direction (WD) and Standard Deviation Wind Direction (STDWD).

Sample filters for all continuous air monitors are changed before the calibration begins. The sample manifold is cleaned during the site visit each month.

Control checks, consisting of a zero and span, are conducted daily on all continuous air monitors. In place of the air sample, zero air (from scrubbed air or gas cylinders) is used for zero checks, and a known concentration of the pollutant being analyzed is used for span checks. These checks are controlled by automatic timers and valves. The total zero span cycle is completed within an hour, the commencement of the zero span cycle is at the beginning of the hour.

Multipoint calibrations are done a minimum of once a month for each continuous air monitor. An additional calibration is required under the following conditions: 1) within three days after the initial start-up and stabilization of a newly installed instrument, 2) prior to shut-down or moving of an instrument which has been working to specification, and 3) when major repair has been done on the instrument.

Time during the first multi-point calibration is not considered downtime (Data is flagged as C). If more than one calibration is performed during the month, the time during the additional calibration is considered as downtime (Data is flagged as C1).

Only one zero/span check is run per day. Time during the zero/span check is not considered as downtime (Data is flagged as S). If an extra zero/span check is performed, the time during the additional check is considered as downtime (Data is flagged as S1).

The AMD requires each instrument and accompanying data recording system to be operational 90% of the time, at a minimum, for each monthly monitoring period.

All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.

Data contained in this monthly report has undergone the verification and validation based on the requirements of the AMD Chapter 6: Ambient Data Quality (December, 2016). The descriptions of the data verification and validation process can be found in Section 5 of this report. Instantaneous data, where applicable, is provided for reference purposes and has not undergone zero correction. The minimum and maximum statistics are highlighted in the data table and are for reference only. The highlighted cells are based on the software's interpretation of the exact position of the minimum or maximum value. The visual presentation of these statistics may not be the obvious choice in a data range due to rounding, truncating or analyzer specifications.

Hourly/minute data have been reviewed based on daily zero/span results and multi-point calibration results. Data may be considered invalid if a zero-corrected span check in excess of +/- 10% of the span concentration (established by the previous multi-point calibration) is encountered and/or significant differences in the calibration factor occurs (greater than 10%).

SULPHUR DIOXIDE (SO₂)

- Operational time, for the monitoring period, was 96.5% equivalent to twenty-six hours of downtime.
- A shut-down calibration was performed on March 1, prior to completing a scheduled annual maintenance on the analyzer. The sample valve was changed, the reaction cell was inspected and cleaned, and the analogue output was calibrated. A successful post-repair calibration was performed on March 2. Twenty-two hours of downtime were recorded due to this maintenance event.
- The Ozone and SO₂ span programs are designed to run concurrently. An additional quality check was recorded on the SO₂ channel, on March 2 at 19:00, during the monthly calibration of the Ozone analyzer.
- A "dark cal" alert was triggered on the analyzer on March 6. The analyzer was reset onsite and no further action was required. Two hours of anomalous data were, however, invalidated at 10:00-11:00 due to this event.
- Two maximum instantaneous data points were invalidated on March 7, due to interference from station activities.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

HYDROGEN SULPHIDE (H₂S)

- Operational time, for the monitoring period, was 94.9% equivalent to thirty-eight hours of downtime.
- A shut-down calibration was performed on March 1, prior to completing a scheduled annual maintenance on the analyzer. The reaction cell and sample valve were cleaned, and the analogue output calibrated. A successful post-repair calibration was performed on March 2. Twenty hours of downtime were recorded due to this maintenance event.
- The zero/span system required time to stabilize after the annual maintenance. An additional zero/span check was performed on March 6 to verify the stability of span response. A repeat calibration was completed on March 7, after which the expected span value was updated. Eight hours of downtime was recorded due to the additional quality checks.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- The span response exceeded the upper acceptance limit on March 18. A repeat span check was performed on March 19 and the result was within acceptance limits. Maintenance of the zero/span system was conducted on March 23, following a successful shut-down calibration. The valves were purged, the tubings were changed, the zero air scrubber material was renewed and a leak check was completed. A successful post-repair calibration was subsequently performed. As both calibrations met AMD requirements, no data was discarded due to this event. Seven hours of downtime were, however, recorded due to the additional quality checks.
- The zero air filter was re-positioned inside the analyzer on March 28 to address a zero drift, after it had been removed to the outside of the analyzer on March 23. One hour of downtime was recorded due to this event.

TOTAL HYDROCARBONS (THC)

- Operational time, for the monitoring period, was 99.6% equivalent to three hours of downtime.
- A shut-down calibration was performed on March 1, prior to completing a scheduled annual maintenance on the analyzer. The sample pump was rebuilt. A successful post-repair calibration was subsequently completed. One hour of downtime was recorded due to this maintenance event.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

OXIDES OF NITROGEN (NO_x), NITRIC OXIDE (NO) and NITROGEN DIOXIDE (NO₂)

- Operational time, for the monitoring period, was 96.2% equivalent to twenty-eight hours of downtime.
- A shut-down calibration was performed on March 1, prior to completing a scheduled annual maintenance on the analyzer. The reaction cell and sample valve were cleaned, and the analogue output calibrated. A successful post-repair calibration was performed on March 2. Nineteen hours of downtime were recorded due to this maintenance event.
- There were concerns about the GPT reference points from the March 2 calibration. A repeat calibration was performed on March 7 for quality assurance and diligence reasons. The calibration met AMD requirements, no issues were identified and no further action was taken. Seven hours of downtime were recorded due this event.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- NO_x calibration concentrations were calculated using a NO_x gas concentration of 50.7 rather than 50.9 ppm. This yielded incorrect values for Calculated NO_x that were presented on the Analyzer Calibration Form. The calibration concentrations were recalculated using the correct NO_x gas value (50.9 ppm) and the outcome was insignificant. The calibration is still deemed AMD compliant.

OZONE (O₃)

- Operational time, for the monitoring period was 97.0% equivalent to twenty-two hours of downtime.
- A shut-down calibration was performed on March 1, prior to completing a scheduled annual maintenance on the analyzer. The sample pump was rebuilt and absorption cells A and B cleaned. Zero air scrubber container was changed due to a leak. A successful post-repair calibration was performed on March 2. Twenty hours of downtime were recorded due to this maintenance event.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

PARTICULATE MATTER < 2.5 MICRONS (PM_{2.5})

- Operational time, for the monitoring period, was 98.4% equivalent to twelve hours of downtime.
- Two routine TEOM audits were performed this month. The first was completed on March 1, and the second on March 28.
- Data was corrected in accordance with AMD (2016), Chapter 6, Table 2, Zero Adjustment Criteria. Data recorded between 0 and -3 µg/m³ was corrected to 0 µg/m³. Data recorded below -3 µg/m³ was invalidated. Ten hours of data were invalidated as the data was below -3 µg/m³ this month.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

WIND SPEED (WS), WIND DIRECTION (WD) and STANDARD DEVIATION WIND DIRECTION (STDWD)

- Operational time, for the monitoring period, was 99.6% equivalent to three hours of downtime.
- One hour of downtime due to maintenance was recorded on March 2 at hour 17:00 to check the alignment of the instrument.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

RELATIVE HUMIDITY (RH)

- Operational time, for the monitoring period, was 99.7% equivalent to two hours of downtime.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

BAROMETRIC PRESSURE (BP)

- Operational time, for the monitoring period, was 99.7% equivalent to two hours of downtime.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

2.0 Project Personnel

Mike Bisaga was the contact for Lakeland Industry & Community Association and the Maxxam field technicians were Limin Li, Michael Espiritu, and Alexander Yakupov.

3.0 Plant Monthly Required AMD Summary

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

4.0 Calculations and Results

All calculations and reporting of results follow the methods described in the AMD, 2016.

5.0 Methods and Procedures

The following methods and procedures were used to complete the monitoring program:

- Maxxam AIR SOP-00209: Ambient Sulphur Monitoring
- Maxxam AIR SOP-00212: Ambient O₃ Monitoring
- Maxxam AIR SOP-00213: Ambient NO/NO₂/NO_x Monitoring
- Maxxam AIR SOP-00214: Ambient Hydrocarbon (THC) Monitoring
- Maxxam AIR SOP-00215: TEOM Operation
- Maxxam AIR SOP-00242: Precipitation Collector Installation/Maintenance

There were no deviations from the prescribed methods.

The following instruments were used to perform the test program:

- Sulphur Dioxide - API 100E UV Fluorescent Analyzer
- Hydrogen Sulphide - API 101E UV Fluorescent Analyzer
- Total Hydrocarbons - Thermo 51C FID Analyzer
- Oxides of Nitrogen - API 200E Chemiluminescent Analyzer
- Ozone - Thermo 49i Photometric Analyzer
- Particulate Matter (PM_{2.5}) - R&P 1405F TEOM Unit
- Wind System - Met One Unit
- Relative Humidity - Met One Unit
- Barometric Pressure - Met One Unit
- Ambient Temperature - Met One Unit
- Precipitation - Met One Unit
- Datalogger - ESC 8832

The following steps were used to complete the data verification and validation process:

Level 0 Preliminary Verification

Level 0 data are raw data obtained directly from the data acquisition system (DAS). Under the step of Level 0, these data undergo a certain amount of manual or automated screening and flagging. It included a) identification of periods of missing data; b) verification of time stamps against reference time; c) verification that instrument diagnostics/datalogger flags indicate normal operation; d) comparison of data to upper and lower limits; e) rate of change flagging indicating that data changed too rapidly or not at all; and f) verification that zero, span and multipoint performance checks are within specifications. This level of verification is performed on a daily basis.

Level 1 Primary Validation

Validation actions under the step of Level 1 include a) review of all screening flags assigned during preliminary verification; b) review of all supporting site information and documentation; c) review of operational acceptance limits for each parameter/analyzer; d) review of daily zero/span and monthly calibration results for all gaseous parameters; and e) application of any necessary adjustments to data (e.g. baseline adjustments, below zero adjustments). This level of validation is performed on a monthly basis.

Level 2 Final Validation


The purpose of Level 2 validation is to verify that there are no inconsistencies among related data, or among regional data measured at nearby sites.

Level 3 Independent Data Review

Level 3 validation is the last step of data review, and it is completed by an individual that is independent of both field operations and primary data validation. A final independent QA review and endorsement is performed during this step before data is submitted to Alberta Environment.

Post-Final Validation

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. Any data issues or patterns which were not clear on a monthly basis are highlighted during this step. This validation is performed on an annual basis.



Source: Air Monitoring Directive (December 2016), Chapter 6, Ambient Data Quality; Figure 1 Data Collection and Management Process Flow Chart

APPENDIX I
CONTINUOUS MONITORING DATA RESULTS

SULPHUR DIOXIDE

SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	0.6	0.5	0.5	0.4	0.3	0.3	S	0.1	0.1	C	C	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	0.1	0.6	0.4	13			
2	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	Q	1.0	0.5	0.5	0.4	0.4	0.4	2.0	1.0	13						
3	0.3	0.3	0.2	0.3	S	0.2	0.3	0.4	0.3	0.4	0.5	0.4	0.4	0.7	0.6	0.4	0.5	0.2	0.7	0.8	0.5	0.2	0.1	0.0	0.0	0.0	0.8	0.4	24			
4	0.1	0.0	0.1	S	0.0	0.4	0.8	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.1	0.0	0.0	0.0	0.0	0.1	0.0	0.8	0.1	24				
5	0.4	0.5	S	0.3	0.4	0.3	0.3	0.2	0.3	0.1	0.2	0.2	0.2	0.3	0.7	0.7	0.5	0.5	0.4	0.3	0.5	0.4	0.4	0.7	0.1	0.7	0.4	24				
6	0.8	S	0.7	0.9	0.6	0.4	0.3	0.3	0.2	0.3	X	X	0.5	0.4	0.4	0.3	0.3	0.8	0.8	0.9	1.2	2.0	2.3	3.5	0.2	3.5	0.9	22				
7	S	1.8	2.9	2.7	2.4	1.3	1.8	2.8	2.0	1.0	0.6	0.2	0.1	0.0	0.0	0.1	0.1	0.0	0.1	0.1	0.3	0.7	1.1	S	0.0	2.9	1.0	24				
8	0.4	0.4	0.5	0.4	0.3	0.3	0.4	0.4	0.5	0.4	0.4	0.7	0.8	0.7	0.6	0.5	0.4	0.5	0.4	0.3	0.4	0.4	S	0.1	0.1	0.8	0.4	24				
9	0.3	0.6	0.6	0.6	0.6	0.6	0.6	0.7	0.7	0.8	1.1	0.7	0.2	0.4	0.4	0.4	0.4	0.4	0.2	0.2	0.3	S	0.2	0.3	0.2	1.1	0.5	24				
10	0.4	0.3	0.4	0.5	0.5	0.5	0.4	0.4	0.3	0.7	0.7	0.4	0.4	0.4	0.3	0.4	0.3	0.3	0.4	0.3	S	0.1	0.2	0.2	0.1	0.7	0.4	24				
11	0.2	0.1	0.2	0.3	0.2	0.3	0.3	X	X	0.2	0.2	0.2	0.1	0.0	0.3	0.4	0.3	0.4	0.3	0.4	S	0.3	0.3	0.3	0.0	0.4	0.2	22				
12	0.3	0.3	0.4	0.4	0.2	0.3	0.2	0.2	0.3	0.1	0.1	0.1	0.0	0.3	0.4	0.3	0.1	0.0	S	0.1	0.0	0.0	0.2	0.3	0.0	0.4	0.2	24				
13	0.3	0.4	0.3	0.4	0.4	0.4	0.4	0.4	0.6	0.7	0.8	0.9	0.9	1.1	1.1	0.8	0.7	S	0.4	0.6	0.5	0.5	0.4	0.4	0.3	1.1	0.6	24				
14	0.3	0.3	0.4	0.5	0.3	0.1	0.2	0.2	0.2	0.1	0.2	0.5	0.2	0.3	0.5	0.6	S	0.6	0.6	0.8	0.6	0.7	0.8	0.7	0.1	0.8	0.4	24				
15	0.7	0.7	0.6	0.5	0.6	0.4	0.4	0.4	0.2	0.2	0.4	0.3	0.2	0.2	0.1	S	0.0	0.0	0.1	0.2	0.2	0.3	0.0	0.2	0.0	0.7	0.3	24				
16	0.2	0.1	0.3	0.4	0.4	0.3	0.3	0.3	0.4	0.3	0.3	0.2	0.1	0.1	S	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	0.0	0.0	0.4	0.2	24				
17	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.1	0.2	0.1	0.2	0.1	0.2	0.9	S	1.5	0.9	0.4	0.2	0.3	0.4	0.1	0.0	0.0	1.5	0.2	24				
18	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.1	0.0	0.1	0.2	0.2	S	0.5	0.8	0.9	1.0	1.0	1.2	1.2	1.4	1.3	1.3	1.2	0.0	1.4	0.5	24				
19	1.0	0.8	0.7	0.6	0.5	0.4	0.4	0.3	0.3	0.3	0.3	S	0.2	0.3	0.4	0.4	0.4	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.2	1.0	0.5	24				
20	0.5	0.3	0.3	0.3	0.3	0.4	0.4	0.3	0.4	1.2	S	0.5	0.4	0.5	0.1	0.1	0.2	0.1	0.2	0.3	0.3	0.3	0.2	0.0	0.0	1.2	0.3	24				
21	0.1	0.0	0.0	0.0	0.0	0.1	0.2	0.2	S	0.0	0.1	0.2	0.1	0.2	0.2	0.1	0.0	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.0	0.2	0.1	24				
22	0.2	0.2	0.4	0.3	0.5	0.4	0.3	0.4	S	0.2	0.4	0.5	0.4	0.3	0.3	0.3	0.2	0.4	0.3	0.2	0.2	0.3	0.3	0.3	0.2	0.5	0.3	24				
23	0.2	0.2	0.1	0.1	0.1	0.1	0.0	S	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.2	0.1	0.1	0.1	0.1	0.1	0.0	0.2	0.1	24				
24	0.1	0.0	0.1	0.1	0.2	0.1	S	0.0	0.1	0.1	0.2	0.2	0.3	0.2	0.3	0.3	0.5	0.5	0.6	0.6	0.6	0.5	0.6	0.6	0.0	0.6	0.3	24				
25	0.6	0.5	0.5	0.5	0.2	S	0.2	0.3	0.2	0.3	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.0	0.1	0.1	0.1	0.1	0.0	0.0	0.0	0.6	0.2	24				
26	0.1	0.0	0.1	0.2	S	0.2	0.3	0.2	0.4	0.3	0.1	0.4	0.4	0.4	0.6	0.6	0.3	0.1	0.0	0.0	0.1	0.1	0.0	0.1	0.0	0.6	0.2	24				
27	0.2	0.1	0.2	S	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.1	0.0	0.1	0.1	0.0	0.2	0.2	24				
28	0.2	0.2	S	0.1	0.1	0.1	0.1	0.1	0.0	0.0	0.1	0.0	0.1	0.2	0.2	0.1	0.2	0.2	0.1	0.2	0.1	0.1	0.1	0.0	0.0	0.2	0.1	24				
29	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24				
30	S	0.0	0.1	0.2	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.1	0.2	0.2	0.1	0.2	0.1	0.2	0.1	0.2	0.2	0.2	0.2	S	0.0	0.3	24				
31	0.0	0.0	0.0	0.1	0.1	0.2	0.1	0.2	0.4	0.9	1.8	1.3	0.3	0.0	0.2	0.2	0.1	0.1	0.2	0.2	0.2	0.1	S	0.0	0.0	1.8	0.3	24				
HOURLY MAX	1.0	1.8	2.9	2.7	2.4	1.3	1.8	2.8	2.0	1.2	1.8	1.3	0.9	1.1	1.5	0.9	1.0	2.0	1.6	1.2	1.4	2.0	2.3	3.5								
HOURLY AVG	0.3	0.3	0.4	0.4	0.3	0.3	0.3	0.3	0.3	0.3	0.4	0.3	0.3	0.3	0.4	0.3	0.3	0.3	0.4	0.3	0.3	0.3	0.4	0.4								

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT:	1-HR	172	ppb	24-HR	48	ppb
-----------------------------	------	-----	-----	-------	----	-----


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0		
NUMBER OF 24-HR EXCEEDANCES:	0		
NUMBER OF NON-ZERO READINGS:	580		
MINIMUM 1-HR AVERAGE:	0.0 ppb @ HOUR(S) 23 ON DAY(S) 3		
MAXIMUM 1-HR AVERAGE:	3.5 ppb @ HOUR(S) 23 ON DAY(S) 6		
MAXIMUM 24-HR AVERAGE:	1.0 ppb ON DAY(S) 7		
	VAR-VARIOUS		
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	718 hrs
MONTHLY CALIBRATION TIME:	9 hrs	AMD OPERATION UPTIME:	96.5 %
STANDARD DEVIATION:	0.4	MONTHLY AVERAGE:	0.3 ppb

24 HR AVERAGES March 2017



SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - March 2017

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	4.0	3.8	3.6	3.7	3.6	3.5	S	3.3	3.3	C	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	3.3	4.0	3.6	13	
2	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	4.5	Q	Q	3.9	3.1	3.3	3.1	3.1	4.5	3.6	13	
3	2.8	3.0	2.9	2.9	S	3.0	3.1	3.1	3.2	3.3	3.5	3.3	3.3	3.7	3.5	3.5	3.5	3.4	3.7	4.1	3.8	3.2	3.3	3.3	2.8	4.1	3.3	24	
4	3.3	3.1	3.3	S	3.1	3.5	3.8	3.4	3.3	3.1	3.2	3.0	2.9	2.8	3.1	2.9	2.8	3.0	2.7	2.5	2.6	2.5	2.8	2.7	2.5	3.8	3.0	24	
5	2.9	2.9	S	2.7	2.9	2.9	2.8	2.8	2.8	2.7	2.8	2.7	2.8	2.7	2.9	3.0	3.3	3.1	3.1	3.1	2.9	3.2	2.9	3.0	3.3	2.7	3.3	2.9	24
6	3.6	S	3.6	3.7	3.1	3.3	2.8	2.7	2.7	2.7	X	X	3.1	3.1	2.8	2.8	2.8	3.3	3.2	3.3	4.0	4.5	5.2	5.8	2.7	5.8	3.4	22	
7	S	4.7	5.4	5.1	4.8	3.5	4.0	5.1	4.4	3.3	2.8	X	X	1.9	1.7	1.9	1.9	1.7	1.7	1.7	1.9	2.4	2.9	S	1.7	5.4	3.1	22	
8	1.8	1.8	2.0	1.9	1.9	1.7	1.6	1.6	1.8	1.8	1.7	1.9	1.9	1.8	1.7	1.6	1.4	1.4	1.3	1.3	1.3	1.5	S	1.0	1.0	2.0	1.6	24	
9	1.2	1.3	1.6	1.5	1.3	1.5	1.4	1.4	1.4	1.5	1.7	1.7	0.8	1.0	1.0	1.0	1.2	0.9	0.8	0.9	0.9	S	0.8	0.8	0.8	1.7	1.2	24	
10	0.8	0.9	0.9	1.1	1.1	1.1	1.0	1.1	0.8	1.6	1.5	1.0	1.0	1.2	0.8	0.8	0.8	0.8	0.9	0.9	S	0.9	0.8	1.0	0.8	1.6	1.0	24	
11	1.0	0.8	0.8	0.9	0.9	1.2	1.2	X	X	1.1	1.2	1.1	1.0	0.9	1.3	1.4	1.5	1.3	1.6	S	1.4	1.4	1.3	1.3	0.8	1.6	1.2	22	
12	1.3	1.4	1.5	1.5	1.3	1.4	1.4	1.4	1.4	1.2	1.2	1.2	1.2	1.5	1.5	1.4	1.4	1.3	S	1.5	1.5	1.5	1.7	1.7	1.2	1.7	1.4	24	
13	2.0	2.1	1.9	1.9	2.2	2.1	2.3	2.3	2.5	2.8	2.9	2.8	3.1	3.3	3.5	3.1	3.2	S	2.8	2.9	2.9	2.8	2.8	2.8	1.9	3.5	2.7	24	
14	2.9	2.9	3.1	3.0	2.7	2.7	2.9	2.7	2.9	2.9	2.9	3.1	2.9	3.0	3.1	3.6	S	3.4	3.3	3.7	3.5	3.5	3.5	3.7	2.7	3.7	3.1	24	
15	3.6	3.5	3.4	3.5	3.5	3.4	3.5	3.5	3.3	3.6	3.5	3.4	3.3	3.3	S	3.1	3.2	3.4	3.3	3.3	3.4	2.9	3.0	2.9	3.6	3.4	24		
16	3.2	3.1	3.2	3.2	3.3	3.3	3.2	2.9	3.2	2.9	3.0	3.1	2.7	2.7	S	2.6	2.8	2.9	2.7	2.6	2.6	2.8	2.7	2.4	2.4	3.3	2.9	24	
17	2.6	2.5	2.4	2.5	2.3	2.4	2.5	2.5	2.5	2.6	2.4	2.7	3.5	S	4.3	3.5	3.0	2.9	3.3	3.4	2.9	3.0	3.0	3.2	2.3	4.3	2.9	24	
18	3.1	3.3	3.2	3.4	3.5	3.6	3.8	3.7	3.7	3.9	4.0	4.2	S	4.3	4.6	4.7	4.8	4.7	4.9	4.8	4.9	4.9	4.7	4.6	3.1	4.9	4.1	24	
19	4.4	4.1	3.9	3.9	3.6	3.6	3.6	3.4	3.2	3.2	3.3	S	3.1	3.1	3.1	3.3	3.0	2.9	2.9	2.9	2.9	2.9	2.8	2.8	2.8	4.4	3.3	24	
20	2.8	2.5	2.4	2.4	2.3	2.2	2.3	2.1	2.4	3.0	S	2.4	2.3	2.5	2.3	2.3	2.0	1.9	2.2	2.3	2.2	2.3	2.3	2.0	1.9	3.0	2.3	24	
21	2.3	2.1	2.2	2.0	2.1	2.1	2.4	2.3	2.6	S	2.5	2.5	2.7	2.6	2.7	2.9	3.0	2.9	2.9	2.9	3.0	3.1	3.1	3.3	2.0	3.3	2.6	24	
22	3.3	3.3	3.6	3.6	4.1	3.8	3.9	3.9	S	3.9	4.0	3.9	4.0	3.9	3.8	4.0	4.1	3.9	3.8	3.9	3.7	3.7	3.6	3.7	3.3	4.1	3.8	24	
23	3.7	3.9	3.6	3.5	3.7	3.6	3.5	S	3.5	3.6	3.5	3.5	3.5	3.5	3.5	3.4	3.3	3.4	3.3	3.2	3.2	3.2	3.2	3.2	3.2	3.9	3.5	24	
24	3.2	2.9	3.0	3.2	3.0	3.1	S	3.0	3.3	3.1	3.3	3.1	3.4	3.2	3.5	3.5	3.8	3.9	4.1	3.8	4.1	3.9	4.1	4.1	2.9	4.1	3.5	24	
25	4.0	4.2	3.9	4.1	3.8	S	3.8	4.0	4.1	3.7	4.0	3.8	3.8	4.0	3.7	3.8	3.7	3.7	3.8	3.7	3.7	3.9	3.7	3.7	3.7	4.2	3.9	24	
26	3.9	3.8	3.8	3.9	S	3.9	3.9	3.8	4.1	4.1	3.9	4.1	4.1	4.1	4.4	4.3	4.1	3.6	3.8	3.5	3.8	3.5	3.5	3.9	3.5	4.4	3.9	24	
27	3.9	3.7	3.7	S	3.8	3.9	3.7	3.9	3.8	3.8	3.7	3.6	3.7	3.6	3.6	3.7	3.6	3.5	3.8	3.5	3.3	3.4	3.3	3.7	3.3	3.9	3.7	24	
28	3.6	3.4	S	3.4	3.4	3.5	3.6	3.7	3.5	3.5	3.6	3.7	3.6	4.1	3.9	3.7	3.9	3.9	4.1	3.7	3.8	3.9	3.9	3.9	3.4	4.1	3.7	24	
29	4.1	S	3.9	3.8	3.7	3.7	3.9	3.9	3.7	4.1	4.0	3.9	4.1	3.9	4.1	4.1	4.2	4.0	4.2	4.3	4.3	4.2	4.3	4.4	3.7	4.4	4.0	24	
30	S	4.2	4.3	4.5	4.3	4.3	4.1	4.4	4.3	4.3	4.0	3.9	3.9	3.8	3.8	3.8	3.8	3.7	3.5	3.5	3.5	3.5	3.2	S	3.2	4.5	3.9	24	
31	3.3	3.4	3.2	3.5	3.5	3.6	3.8	3.7	4.1	5.2	5.6	5.2	4.5	3.8	4.1	4.2	4.3	3.8	4.0	4.0	4.1	4.0	S	4.0	3.2	5.6	4.0	24	
HOURLY MAX	4.4	4.7	5.4	5.1	4.8	4.3	4.1	5.1	4.4	5.2	5.6	5.2	4.5	4.3	4.6	4.7	4.8	4.7	4.9	4.8	4.9	4.9	5.2	5.8					
HOURLY AVG	3.0	3.0	3.0	3.0	3.0	2.9	3.0	3.1	3.1	3.1	3.1	3.0	3.0	3.0	3.1	3.0	3.0	3.0	3.1	3.0	3.1	3.1	3.1	3.1					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	672
MAXIMUM INSTANTANEOUS VALUE:	5.8 ppb @ HOUR(S) 23 ON DAY(S) 6
	VAR-VARIOUS
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	10 hrs
OPERATIONAL TIME:	716 hrs
STANDARD DEVIATION:	1.0

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)



Wind: LICA ST. LINA
 Poll.: LICA ST. LINA-SO2[ppb]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 0.15%

Calm Avg: 0.12 [ppb]

Direction	0.0-0.7	0.7-1.4	1.4-2.2	2.2-2.9	2.9-3.6	>3.6	Total
N	18.4	0.3	0.0	0.0	0.0	0.0	18.7
NE	12.7	1.9	0.7	0.6	0.3	0.0	16.3
E	8.0	0.2	0.0	0.0	0.0	0.0	8.2
SE	8.0	0.3	0.2	0.0	0.0	0.0	8.5
S	6.2	0.2	0.0	0.0	0.0	0.0	6.4
SW	15.4	2.4	0.2	0.0	0.0	0.0	17.9
W	19.1	1.8	0.0	0.0	0.0	0.0	20.9
NW	3.0	0.2	0.0	0.0	0.0	0.0	3.1
Summary	90.8	7.1	1.0	0.6	0.3	0.0	100.0

% Icon Classes (ppb)

91 0.0-0.7

7 0.7-1.4


1 1.4-2.2

1 2.2-2.9

0 2.9-3.6

0 >3.6

LICA ST. LINA Poll.: LICA ST. LINA-SO2[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 0.15% Calm Poll Avg: 0.12[ppb]



SO2[ppb] Calibration: LICA ST. LINA Monthly; 17/03 Type: Span



Span Meas Span Ref Span Low Span High

HYDROGEN SULPHIDE



HYDROGEN SULPHIDE Hourly Averages (H₂S ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.		
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.			
DAY																														
1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	C	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	0.0	0.0	0.0	13		
2	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	C	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	15	
3	0.1	0.0	0.0	0.0	S	0.0	0.1	0.0	0.0	0.1	0.1	0.1	0.2	0.2	0.1	0.2	0.2	0.1	0.0	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.2	0.1	24	
4	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
5	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
6	0.0	S	0.0	0.1	0.0	0.1	0.0	0.0	S1	0.0	0.0	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.1	0.1	0.0	0.2	0.1	23	
7	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	C1	C1	C1	C1	C1	C1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	18	
8	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	S	0.1	0.0	0.1	0.1	24	
9	0.1	0.1	0.1	0.2	0.2	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.1	0.0	0.2	0.0	24	
10	0.1	0.0	0.1	0.1	0.3	0.3	0.3	0.2	0.3	0.2	0.2	0.2	0.2	0.2	0.3	0.2	0.2	0.2	0.2	0.2	S	0.1	0.2	0.2	0.0	0.3	0.2	0.2	24	
11	0.2	0.2	0.1	0.1	0.1	0.0	0.0	X	X	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.2	0.0	0.0	22	
12	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.1	0.1	0.0	0.0	0.0	0.0	0.1	0.0	24	
13	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	24	
14	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.2	0.1	0.2	S	0.2	0.2	0.3	0.2	0.2	0.2	0.3	0.0	0.3	0.0	0.3	0.1	24
15	0.3	0.3	0.3	0.3	0.3	0.2	0.3	0.2	0.1	0.1	0.1	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.3	0.1	0.1	24
16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
17	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.2	0.2	0.3	0.3	0.4	0.5	0.4	0.4	0.5	0.0	0.5	0.1	0.1	24	
19	0.4	0.3	0.2	0.1	0.1	0.1	S1	0.0	0.0	0.0	0.0	S	0.0	0.1	0.1	0.1	0.0	0.0	0.2	0.1	0.2	0.1	0.1	0.1	0.0	0.4	0.1	0.1	23	
20	0.1	0.0	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.1	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	24
21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.1	0.0	0.2	0.0	0.0	24
22	0.1	0.2	0.3	0.3	0.3	0.3	0.3	S	0.3	S	0.3	0.4	0.3	0.3	0.4	0.4	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.4	0.3	0.1	24
23	0.1	0.1	0.0	0.0	0.0	0.0	0.0	S	0.0	C1	C1	Y	C1	C1	C1	C1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	17
24	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
25	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
26	0.0	0.0	0.0	0.0	S	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	24
27	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
28	0.0	0.0	S	0.0	0.1	0.0	0.1	0.1	0.1	Y	0.1	0.2	0.2	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.2	0.2	0.0	0.3	0.2	0.2	23
29	0.3	S	0.3	0.2	0.2	0.2	0.3	0.2	0.2	0.2	0.1	0.2	0.2	0.2	0.3	0.3	0.2	0.1	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.1	0.3	0.2	0.2	24
30	S	0.3	0.3	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.1	0.0	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.0	S	0.0	0.3	0.2	0.2	24	
31	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.1	0.1	0.1	0.0	0.0	0.0	0.1	0.1	0.1	0.2	0.3	0.3	0.2	0.2	S	0.2	0.0	0.3	0.1	0.1	24	
HOURLY MAX	0.4	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.4	0.4	0.3	0.3	0.4	0.4	0.3	0.3	0.3	0.4	0.5	0.4	0.4	0.5						
HOURLY AVG	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.0	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1						

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT:	1-HR	10	ppb	24-HR	3	ppb
----------------------	------	----	-----	-------	---	-----


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0					
NUMBER OF 24-HR EXCEEDANCES:	0					
NUMBER OF NON-ZERO READINGS:	256					
MINIMUM 1-HR AVERAGE:	0.0	ppb	@ HOUR(S)	0	ON DAY(S)	1
MAXIMUM 1-HR AVERAGE:	0.5	ppb	@ HOUR(S)	20	ON DAY(S)	18
MAXIMUM 24-HR AVERAGE:	0.3	ppb			ON DAY(S)	22
					VAR-VARIOUS	
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	706	hrs	
MONTHLY CALIBRATION TIME:	11	hrs	AMD OPERATION UPTIME:	94.9	%	
STANDARD DEVIATION:	0.1		MONTHLY AVERAGE:	0.1	ppb	

24 HR AVERAGES March 2017



HYDROGEN SULPHIDE Hourly Averages (H₂S ppb)





HYDROGEN SULPHIDE Instantaneous Maximum (H₂S ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	1.0	0.9	1.0	1.0	0.9	0.9	S	0.8	0.9	C	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	0.8	1.0	0.9	13	
2	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	C	C	0.8	0.9	0.8	0.9	0.9	1.0	1.1	0.8	1.1	0.9	15	
3	1.0	1.0	1.0	1.2	S	1.2	1.1	1.1	1.2	1.2	1.2	1.2	1.3	1.2	1.3	1.3	1.3	1.2	1.2	1.2	1.2	1.2	1.2	1.1	1.0	1.3	1.2	24	
4	1.1	1.2	1.1	S	1.2	1.1	1.2	1.1	1.0	1.1	1.1	1.0	0.9	1.0	0.9	0.9	0.8	1.0	0.8	0.7	0.8	0.8	0.7	0.8	0.7	1.2	1.0	24	
5	0.8	0.8	S	0.8	0.8	0.9	0.9	0.9	0.8	0.8	0.9	0.8	0.9	0.9	1.0	1.0	1.0	1.0	1.0	1.0	0.9	1.0	1.0	1.0	0.8	1.0	0.9	24	
6	1.0	S	1.0	1.1	1.0	1.1	0.9	S1	S1	0.9	0.9	1.0	1.0	0.9	0.9	0.8	0.9	1.0	0.9	0.8	0.9	0.8	0.8	0.8	0.8	1.1	0.9	22	
7	S	0.7	0.8	0.8	0.7	0.7	0.6	0.6	0.5	C1	C1	C1	C1	C1	C1	0.9	0.7	0.8	0.6	0.7	0.7	0.7	0.6	S	0.5	0.9	0.7	18	
8	0.7	0.7	0.6	0.6	0.6	0.7	0.6	0.7	0.6	0.6	0.5	0.5	0.5	1.1	0.5	0.5	0.4	0.4	0.5	0.5	0.4	0.3	S	0.4	0.3	1.1	0.6	24	
9	0.4	0.4	0.5	0.4	0.5	0.5	0.3	0.2	0.2	0.1	0.2	0.2	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.1	0.0	S	0.0	0.1	0.0	0.5	0.2	24	
10	0.1	0.0	0.1	0.1	0.3	0.3	0.3	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.0	0.0	0.1	0.1	S	0.0	0.1	0.1	0.0	0.3	0.1	24	
11	0.1	0.3	0.1	0.1	0.2	0.1	0.2	X	X	0.2	0.2	0.2	0.1	0.2	0.3	0.3	0.3	0.3	0.3	S	0.4	0.4	0.4	0.3	0.1	0.4	0.2	22	
12	0.4	0.5	0.3	0.4	0.3	0.3	0.2	0.2	0.3	0.2	0.4	0.2	0.2	0.2	0.3	0.2	0.2	0.2	S	0.4	0.5	0.5	0.4	0.4	0.2	0.5	0.3	24	
13	0.5	0.6	0.6	0.6	0.7	0.7	0.7	0.7	1.4	0.7	0.7	0.8	0.9	0.9	0.8	0.9	0.8	S	1.0	0.9	0.9	0.9	0.9	0.9	0.5	1.4	0.8	24	
14	0.9	0.9	1.1	0.9	0.9	1.1	0.8	0.9	0.9	0.9	0.9	1.0	1.1	1.2	1.1	1.2	S	1.1	1.2	1.3	1.2	1.2	1.3	1.2	0.8	1.3	1.1	24	
15	1.3	1.5	1.4	1.3	1.3	1.3	1.4	1.3	1.2	1.2	1.2	1.1	1.2	1.2	1.1	S	1.1	1.1	1.1	1.0	1.1	1.0	1.1	1.0	1.0	1.5	1.2	24	
16	1.1	1.0	1.0	1.0	1.1	1.2	1.1	1.1	1.0	1.0	1.1	1.0	0.9	0.9	S	0.9	0.8	0.9	0.8	0.7	0.7	0.7	0.7	0.7	0.7	1.2	0.9	24	
17	0.7	0.5	0.6	0.6	0.6	0.6	0.7	0.7	0.6	0.8	0.6	0.7	0.7	S	0.7	0.7	0.8	0.8	0.8	0.8	1.6	1.0	1.0	1.0	0.5	1.6	0.8	24	
18	1.1	1.1	1.1	1.1	1.3	1.3	1.2	1.3	1.4	1.4	1.5	1.5	S	1.6	1.6	1.7	1.8	1.8	1.9	1.9	1.9	1.9	1.8	1.7	1.1	1.9	1.5	24	
19	1.7	1.5	1.4	1.2	1.2	1.2	S1	S1	1.0	0.9	0.9	S	0.9	0.9	0.9	1.0	0.7	0.7	0.9	0.7	0.8	0.7	0.6	0.7	0.6	1.7	1.0	22	
20	0.7	0.5	0.5	0.5	0.4	0.4	0.3	0.3	0.3	0.3	S	0.3	0.2	0.3	0.3	0.2	0.2	0.3	0.3	0.2	0.2	0.2	0.1	0.2	0.1	0.7	0.3	24	
21	0.3	0.3	0.4	0.2	0.4	0.5	0.5	0.4	0.5	S	0.4	0.5	0.5	0.7	0.6	0.7	0.7	0.8	0.7	0.8	0.8	0.8	0.9	1.0	0.9	0.2	1.0	0.6	24
22	0.9	1.0	1.0	1.1	1.1	1.2	1.3	1.2	S	1.2	1.3	1.3	1.3	1.3	1.3	1.3	1.3	1.3	1.2	1.2	1.2	1.2	1.2	1.2	0.9	1.3	1.2	24	
23	1.1	1.1	1.1	1.2	1.2	1.0	1.0	S	1.0	C1	C1	Y	C1	C1	C1	C1	0.4	0.5	0.6	0.6	0.5	0.6	0.6	0.4	0.4	1.2	0.8	17	
24	0.3	0.4	0.4	0.4	0.4	0.5	S	0.5	0.6	0.6	0.6	0.6	0.6	0.6	0.6	0.8	0.8	0.8	0.8	0.8	0.8	0.8	0.9	0.9	0.3	0.9	0.6	24	
25	0.9	0.9	0.8	0.9	0.8	S	0.8	0.8	0.7	0.8	0.8	0.7	0.8	0.7	0.7	0.7	0.7	0.8	0.7	0.7	0.7	0.7	0.7	0.7	0.7	0.9	0.8	24	
26	0.7	0.7	0.7	0.7	S	0.8	0.9	0.8	0.7	0.8	0.7	0.8	0.7	0.8	0.7	0.6	0.6	0.7	0.8	0.8	0.7	0.7	0.7	0.7	0.6	0.9	0.7	24	
27	0.7	0.7	0.7	S	0.7	0.6	0.7	0.6	0.5	0.9	0.6	0.6	0.5	0.5	0.4	0.5	0.5	0.5	0.5	0.7	0.5	0.7	0.6	0.6	0.4	0.9	0.6	24	
28	1.5	0.6	S	0.6	0.6	0.6	0.7	0.7	0.7	Y	0.7	0.7	1.3	0.7	0.8	0.8	0.8	0.8	0.8	0.8	0.8	0.8	0.9	0.9	0.6	1.5	0.8	23	
29	0.9	S	0.9	1.0	0.9	0.9	0.9	0.8	0.9	0.8	0.8	0.8	0.9	1.0	1.0	1.0	0.9	0.9	1.0	1.0	1.0	1.1	1.0	0.9	0.8	1.1	0.9	24	
30	S	1.0	0.9	0.9	0.9	1.0	0.9	1.0	0.9	0.8	0.7	0.8	0.7	0.6	0.6	0.7	0.6	0.6	0.5	0.6	0.6	0.4	0.4	S	0.4	1.0	0.7	24	
31	0.4	0.5	0.4	0.4	0.5	0.4	0.3	0.7	0.6	0.6	0.6	0.5	0.5	0.6	0.7	0.7	0.7	0.8	0.8	1.0	0.8	0.7	S	0.9	0.3	1.0	0.6	24	
HOURLY MAX	1.7	1.5	1.4	1.3	1.3	1.3	1.4	1.3	1.4	1.4	1.5	1.5	1.3	1.6	1.6	1.7	1.8	1.8	1.9	1.9	1.9	1.9	1.8	1.7					
HOURLY AVG	0.8	0.8	0.8	0.8	0.8	0.8	0.8	0.8	0.8	0.8	0.8	0.7	0.7	0.8	0.7	0.8	0.7	0.8	0.8	0.8	0.8	0.8	0.8	0.8					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY


NUMBER OF NON-ZERO READINGS:	648
MAXIMUM INSTANTANEOUS VALUE:	1.9 ppb @ HOUR(S) 18 ON DAY(S) 18
	VAR-VARIOUS
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	12 hrs
OPERATIONAL TIME:	704 hrs
STANDARD DEVIATION:	0.4

HYDROGEN SULPHIDE Instantaneous Maximum (H₂S ppb)




% Icon Classes (ppb) 100 0.0-0.6 0 0.6-1.3 0 1.3-1.9 0 >1.9

LICA ST. LINA Poll.: LICA ST. LINA-H2S[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 0.15% Calm Poll Avg: 0.00[ppb]



H2S[ppb] Calibration: LICA ST. LINA Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

TOTAL HYDROCARBON


TOTAL HYDROCARBONS Hourly Averages (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	2.06	2.05	2.05	2.02	2.02	2.03	S	2.05	2.10	2.09	2.13	2.09	C	C	C	C	Y	C	C	C	C	2.15	2.12	2.12	2.02	2.15	2.08	23
2	2.09	2.12	2.14	2.28	2.28	S	2.28	2.48	2.37	2.43	2.40	2.40	2.42	2.59	2.67	2.74	2.57	2.45	2.41	2.47	2.56	2.47	2.42	2.38	2.09	2.74	2.41	24
3	2.32	2.36	2.36	2.33	S	2.36	2.39	2.39	2.38	2.29	2.24	2.25	2.21	2.20	2.13	2.05	2.02	2.22	2.12	2.09	2.07	2.00	1.99	1.98	1.98	2.39	2.21	24
4	1.98	1.98	2.02	S	2.02	2.04	2.06	2.07	2.05	2.06	2.06	2.05	2.06	2.05	2.03	2.03	2.03	2.05	2.05	2.06	2.05	2.04	2.09	2.12	1.98	2.12	2.05	24
5	2.05	2.05	S	2.08	2.05	2.06	2.05	2.07	2.09	2.08	2.08	2.08	2.09	2.11	2.11	2.11	2.12	2.12	2.12	2.13	2.15	2.29	2.21	2.13	2.05	2.29	2.11	24
6	2.11	S	2.33	2.16	2.14	2.12	2.04	2.05	2.04	2.03	2.04	2.03	2.03	2.03	2.03	2.03	2.04	2.07	2.07	2.06	2.04	2.04	2.07	2.07	2.03	2.33	2.07	24
7	S	2.05	2.06	2.05	2.07	2.07	2.07	2.07	2.07	2.10	2.06	2.02	2.02	2.00	2.00	2.00	2.03	2.05	2.04	2.03	2.03	2.03	2.04	S	2.00	2.10	2.10	24
8	2.03	2.05	2.17	2.19	2.16	2.16	2.11	2.26	2.22	2.11	2.10	2.07	2.06	2.06	2.06	2.05	2.00	2.02	2.03	2.05	2.06	2.06	S	2.18	2.00	2.26	2.10	24
9	2.27	2.22	2.23	2.30	2.38	2.37	2.33	2.32	2.24	2.20	2.21	2.16	2.12	2.13	2.12	2.15	2.18	2.17	2.26	2.18	2.15	S	2.30	2.31	2.12	2.38	2.23	24
10	2.29	2.19	2.19	2.27	2.35	2.47	2.58	2.50	2.32	2.23	2.18	2.11	2.09	2.07	2.07	2.07	2.08	2.09	2.10	2.11	S	2.12	2.15	2.19	2.07	2.58	2.21	24
11	2.20	2.24	2.32	2.38	2.38	2.41	2.44	X	X	2.39	2.38	2.37	2.25	2.28	2.22	2.17	2.14	2.10	2.07	S	2.12	2.13	2.06	2.03	2.44	2.24	2.24	22
12	2.02	2.03	2.02	2.03	2.05	2.05	2.07	2.07	2.08	2.09	2.08	2.08	2.13	2.11	2.10	2.16	2.19	2.20	S	2.21	2.16	2.19	2.14	2.14	2.02	2.21	2.10	24
13	2.16	2.15	2.14	2.14	2.15	2.15	2.22	2.34	2.25	2.14	2.15	2.15	2.13	2.14	2.12	2.12	2.14	S	2.17	2.20	2.21	2.17	2.19	2.21	2.12	2.34	2.17	24
14	2.23	2.23	2.24	2.23	2.26	2.30	2.30	2.30	2.31	2.30	2.34	2.39	2.67	2.69	2.72	2.65	S	2.58	2.56	2.55	2.47	2.39	2.35	2.34	2.23	2.72	2.41	24
15	2.30	2.28	2.28	2.27	2.32	2.33	2.35	2.38	2.31	2.29	2.26	2.19	2.08	2.02	2.02	S	1.99	2.01	2.00	2.01	2.00	2.01	1.99	2.01	1.99	2.38	2.16	24
16	2.01	2.00	2.00	2.00	1.99	2.00	2.03	2.04	2.05	2.04	2.01	1.98	1.96	1.98	S	1.97	1.99	2.09	2.05	2.06	2.05	2.04	2.07	2.11	1.96	2.11	2.02	24
17	2.06	2.11	2.07	2.06	2.04	2.05	2.03	2.02	2.01	1.97	1.98	1.98	1.99	S	1.98	2.02	2.03	2.08	2.15	2.11	2.13	2.05	2.04	2.05	1.97	2.15	2.04	24
18	2.13	2.16	2.14	2.13	2.13	2.15	2.12	2.11	2.08	2.08	2.08	2.10	S	2.04	2.02	2.00	1.97	1.92	1.91	1.85	1.88	1.90	1.85	1.87	1.85	2.16	2.03	24
19	1.86	1.88	1.88	1.89	1.92	1.93	1.95	1.95	1.96	1.96	1.96	S	1.93	1.92	1.93	1.90	1.89	1.90	1.91	1.89	1.90	1.91	1.92	1.94	1.86	1.96	1.92	24
20	1.96	1.98	1.99	2.01	2.01	2.02	2.03	2.03	2.03	2.04	S	2.03	2.05	2.03	2.02	2.02	2.03	2.03	2.05	2.12	2.12	2.11	2.17	2.15	1.96	2.17	2.04	24
21	2.14	2.13	2.11	2.14	2.41	2.67	2.66	2.42	2.24	S	2.16	2.07	2.08	2.07	2.05	2.03	2.03	2.05	2.06	2.05	2.04	2.04	2.06	2.06	2.03	2.67	2.16	24
22	2.06	2.06	2.06	2.06	2.07	2.08	2.09	2.08	S	2.07	2.07	2.06	2.07	2.04	2.01	2.03	2.05	2.05	2.07	2.09	2.07	2.04	2.02	2.01	2.01	2.09	2.06	24
23	2.03	2.02	2.02	2.02	2.02	2.03	2.02	S	2.00	2.01	2.04	2.05	2.04	2.01	1.98	2.01	2.05	2.05	2.11	2.09	2.09	2.12	2.11	2.14	1.98	2.14	2.05	24
24	2.13	2.13	2.14	2.17	2.21	2.16	S	2.24	2.45	2.61	2.70	2.45	2.39	2.33	2.39	2.30	2.18	2.08	2.01	1.99	1.98	1.99	1.97	1.96	1.96	2.70	2.22	24
25	1.97	1.95	1.97	1.98	2.01	S	2.03	2.07	2.09	2.05	2.04	2.06	2.03	2.00	1.98	1.99	2.01	2.02	2.06	2.10	2.17	2.10	2.09	2.14	1.95	2.17	2.04	24
26	2.21	2.18	2.20	2.26	S	2.43	2.52	2.46	2.51	2.57	2.56	2.51	2.33	2.26	2.24	2.33	2.38	2.50	2.52	2.36	2.29	2.27	2.18	2.22	2.18	2.57	2.36	24
27	2.29	2.33	2.33	S	2.23	2.18	2.12	2.09	2.09	2.17	2.18	2.14	2.12	2.08	2.05	2.06	2.09	2.09	2.08	2.08	2.07	2.08	2.09	2.08	2.05	2.33	2.14	24
28	2.10	2.11	S	2.16	2.17	2.20	2.23	2.24	2.23	2.18	2.16	2.17	2.20	2.23	2.16	2.12	2.08	2.06	2.06	2.07	2.06	2.05	2.05	2.05	2.05	2.24	2.14	24
29	2.09	S	2.16	2.16	2.10	2.10	2.08	2.08	2.07	2.07	2.08	2.08	2.11	2.10	2.07	2.04	2.05	2.08	2.12	2.11	2.17	2.17	2.36	2.34	2.04	2.36	2.12	24
30	S	2.09	2.08	2.00	1.95	1.95	1.95	1.95	1.95	1.95	1.95	1.93	1.94	1.95	1.97	1.96	1.96	1.99	1.99	2.00	1.96	1.96	1.99	S	1.93	2.09	1.97	24
31	2.03	2.03	2.04	2.05	2.09	2.12	2.17	2.20	2.19	2.16	2.06	2.02	2.01	1.97	1.96	1.96	1.96	1.94	1.92	1.97	1.96	1.98	S	2.01	1.92	2.20	2.03	24
HOURLY MAX	2.32	2.36	2.36	2.38	2.41	2.67	2.66	2.50	2.51	2.61	2.70	2.51	2.67	2.69	2.72	2.74	2.57	2.58	2.56	2.55	2.56	2.47	2.42	2.38				
HOURLY AVG	2.11	2.11	2.13	2.13	2.14	2.17	2.18	2.18	2.16	2.16	2.16	2.14	2.12	2.12	2.11	2.11	2.08	2.11	2.11	2.11	2.10	2.10	2.11	2.12				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	700		
MINIMUM 1-HR AVERAGE:	1.85 ppm	@ HOUR(S)	19 ON DAY(S)
MAXIMUM 1-HR AVERAGE:	2.74 ppm	@ HOUR(S)	15 ON DAY(S)
MAXIMUM 24-HR AVERAGE:	2.41 ppm		2 ON DAY(S)
			VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	741 hrs
MONTHLY CALIBRATION TIME:	8 hrs	AMD OPERATION UPTIME:	99.6 %
STANDARD DEVIATION:	0.15	MONTHLY AVERAGE:	2.13 ppm

TOTAL HYDROCARBONS Hourly Averages (THC ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - March 2017

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	2.41	2.41	2.41	2.38	2.38	2.39	S	2.38	2.44	2.51	2.58	C	C	C	C	C	Y	C	C	C	C	2.60	2.51	2.49	2.38	2.60	2.45	23
2	2.17	2.23	2.30	2.61	2.51	S	2.49	2.58	2.45	2.47	2.42	2.39	2.41	2.66	2.64	2.73	2.61	2.39	2.38	2.59	2.50	2.41	2.32	2.26	2.17	2.73	2.46	24
3	2.20	2.22	2.23	2.17	S	2.20	2.25	2.23	2.23	2.17	2.08	2.10	2.08	2.07	1.95	1.89	1.91	2.14	2.11	1.89	1.89	1.82	1.79	1.78	1.78	2.25	2.06	24
4	1.78	1.78	1.85	S	1.82	1.83	1.86	1.88	1.86	1.88	1.89	1.89	1.89	1.88	1.86	1.86	1.86	1.89	1.89	1.91	1.89	1.89	1.95	1.98	1.78	1.98	1.87	24
5	1.92	1.91	S	1.95	1.89	1.91	1.91	1.92	1.94	1.92	1.92	1.92	1.92	1.95	1.94	1.94	1.95	1.94	1.94	1.95	1.97	2.47	2.14	1.94	1.89	2.47	1.96	24
6	1.92	S	2.22	2.08	2.10	2.11	1.85	1.89	1.86	1.86	1.86	1.86	1.86	1.86	1.86	1.88	1.91	1.92	1.89	1.89	1.91	1.92	1.94	1.85	2.22	1.93	24	
7	S	1.94	1.94	1.94	1.95	1.98	1.98	1.98	1.98	2.16	2.04	1.98	1.98	2.01	1.98	2.04	2.22	2.02	2.02	2.02	2.04	2.04	2.07	S	1.94	2.22	2.01	24
8	2.05	2.11	2.23	2.23	2.22	2.26	2.22	2.35	2.35	2.17	2.16	2.13	2.13	2.11	2.11	2.11	2.08	2.10	2.13	2.14	2.14	2.15	S	2.32	2.05	2.35	2.17	24
9	2.39	2.35	2.35	2.47	2.51	2.51	2.47	2.44	2.39	2.32	2.34	2.33	2.25	2.26	2.26	2.32	2.33	2.35	2.81	2.36	2.35	S	2.58	2.63	2.25	2.81	2.41	24
10	2.49	2.35	2.38	2.44	2.56	2.66	2.72	2.69	2.53	2.39	2.32	2.25	2.22	2.19	2.18	2.18	2.20	2.22	2.23	2.26	S	2.32	2.32	2.35	2.18	2.72	2.37	24
11	2.36	2.38	2.45	2.54	2.54	2.53	2.53	X	X	2.51	2.49	2.49	2.35	2.35	2.29	2.25	2.23	2.17	2.13	S	2.20	2.20	2.15	2.11	2.11	2.54	2.35	22
12	2.08	2.09	2.10	2.10	2.11	2.13	2.14	2.14	2.15	2.16	2.14	2.17	2.22	2.20	2.18	2.25	2.29	2.29	S	2.29	2.23	2.23	2.19	2.17	2.08	2.29	2.18	24
13	2.18	2.17	2.14	2.11	2.14	2.14	2.32	2.32	2.23	2.10	2.08	2.11	2.05	2.04	2.02	2.02	2.04	S	2.05	2.08	2.10	2.04	2.07	2.09	2.02	2.32	2.11	24
14	2.11	2.11	2.11	2.11	2.14	2.17	2.16	2.17	2.16	2.14	2.25	2.32	2.54	2.54	2.57	2.57	S	2.44	2.39	2.39	2.32	2.22	2.17	2.17	2.11	2.57	2.27	24
15	2.11	2.10	2.10	2.08	2.13	2.14	2.16	2.17	2.14	2.11	2.04	2.01	1.86	1.81	1.78	S	1.78	1.81	1.81	1.79	1.79	1.81	1.78	1.81	1.78	2.17	1.96	24
16	1.83	1.81	1.83	1.81	1.81	1.84	1.88	1.88	1.89	1.89	1.84	1.83	1.83	1.83	S	1.83	1.88	2.04	1.98	1.95	1.95	1.95	2.08	2.08	1.81	2.08	1.89	24
17	2.02	2.11	2.08	2.05	2.01	2.01	2.01	2.01	2.01	1.98	1.98	2.01	2.01	S	1.98	2.01	2.01	2.11	2.14	2.08	2.10	2.04	1.98	1.98	1.98	2.14	2.03	24
18	2.05	2.08	2.04	2.01	2.01	2.01	1.97	1.95	1.92	1.92	1.89	1.89	S	1.85	1.83	1.82	1.81	1.76	1.75	1.69	1.78	1.78	1.72	1.75	1.69	2.08	1.88	24
19	1.76	1.78	1.78	1.79	1.83	1.83	1.85	1.86	1.88	1.89	1.89	S	1.86	1.86	1.88	1.86	1.86	1.88	1.89	1.88	1.89	1.92	1.92	1.97	1.76	1.97	1.86	24
20	2.01	2.04	2.04	2.07	2.08	2.10	2.11	2.11	2.11	2.14	S	2.11	2.13	2.11	2.10	2.11	2.11	2.11	2.11	2.29	2.26	2.29	2.28	2.23	2.01	2.29	2.13	24
21	2.19	2.19	2.16	2.25	2.60	2.76	2.73	2.60	2.32	S	2.25	2.07	2.07	2.07	2.01	1.99	1.98	1.98	1.98	1.97	1.94	1.92	1.94	1.92	1.92	2.76	2.17	24
22	1.92	1.92	1.91	1.89	1.89	1.89	1.91	1.89	S	1.86	1.86	1.86	1.86	1.83	1.85	1.86	1.88	1.89	1.92	1.89	1.89	1.84	1.85	1.83	1.92	1.92	1.88	24
23	1.85	1.85	1.86	1.86	1.86	1.86	1.86	S	1.85	1.91	1.91	1.91	1.91	1.89	1.85	1.95	1.92	1.95	2.08	1.99	2.04	2.04	2.04	2.11	1.85	2.11	1.93	24
24	2.05	2.06	2.08	2.28	2.26	2.14	S	2.35	2.45	2.81	2.78	2.49	2.38	2.29	2.32	2.26	2.11	2.04	1.92	1.91	1.89	1.89	1.86	1.85	1.85	2.81	2.19	24
25	1.85	1.83	1.85	1.85	1.92	S	1.89	1.95	1.98	1.94	1.92	1.92	1.92	1.86	1.85	1.88	1.88	1.91	1.95	1.98	2.22	2.04	1.99	2.04	1.83	2.22	1.93	24
26	2.16	2.08	2.11	2.23	S	2.38	2.49	2.38	2.45	2.51	2.52	2.49	2.38	2.20	2.20	2.33	2.38	2.42	2.47	2.36	2.29	2.23	2.19	2.23	2.08	2.52	2.33	24
27	2.26	2.26	2.31	S	2.20	2.13	2.10	2.04	2.04	2.11	2.13	2.08	2.08	2.04	2.01	2.01	2.04	2.04	2.05	2.04	2.04	2.04	2.05	2.05	2.01	2.31	2.09	24
28	2.10	2.08	S	2.16	2.13	2.17	2.18	2.20	2.19	2.14	2.11	2.19	2.17	2.17	2.13	2.05	2.05	2.01	1.98	1.99	1.98	1.99	1.98	1.97	1.97	2.20	2.09	24
29	2.04	S	2.11	2.11	2.11	2.01	1.99	1.99	1.98	1.97	1.98	1.98	2.01	1.99	1.95	1.95	1.95	2.07	2.05	1.98	2.08	2.11	2.41	2.29	1.95	2.41	2.05	24
30	S	1.97	1.97	1.95	1.83	1.86	1.86	1.86	1.94	1.89	1.91	1.86	1.89	1.89	1.92	1.91	1.94	2.01	1.97	1.98	1.95	1.98	2.05	S	1.83	2.05	1.93	24
31	2.05	2.04	2.01	2.04	2.08	2.08	2.14	2.17	2.14	2.14	2.04	1.97	1.95	1.92	1.89	1.88	1.86	1.85	1.83	1.89	1.86	1.86	S	1.89	1.83	2.17	1.98	24
HOURLY MAX	2.49	2.41	2.45	2.61	2.60	2.76	2.73	2.69	2.53	2.81	2.78	2.49	2.54	2.66	2.64	2.73	2.61	2.44	2.81	2.59	2.50	2.60	2.58	2.63				
HOURLY AVG	2.08	2.08	2.10	2.12	2.12	2.14	2.14	2.15	2.13	2.13	2.12	2.09	2.08	2.06	2.05	2.06	2.04	2.06	2.06	2.05	2.05	2.07	2.08	2.08				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY


NUMBER OF NON-ZERO READINGS:	699
MAXIMUM INSTANTANEOUS VALUE:	2.81 ppm @ HOUR(S) 18 ON DAY(S) 9
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	9 hrs
OPERATIONAL TIME:	741 hrs
STANDARD DEVIATION:	0.21

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)




% Icon Classes (ppm) 0 0.0-0.9 0 0.9-1.8 100 1.8-2.7 0 >2.7

LICA ST. LINA Poll.: LICA ST. LINA-THC[ppm] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 0.14% Calm Poll Avg: 2.05[ppm]



THC[ppm] Calibration: LICA ST. LINA Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

OXIDES OF NITROGEN




OXIDES OF NITROGEN Hourly Averages (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	2.5	1.8	1.7	1.0	0.9	0.9	S	1.0	1.1	C	C	C	C	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	0.9	2.5	1.4	16				
2	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	C	C	C	C	C	C	C	4.5	4.6	4.5	3.7	3.7	4.6	4.3	13
3	3.3	3.4	3.3	3.0	S	3.4	3.4	3.2	3.5	3.4	3.3	3.6	3.8	4.2	3.3	3.1	3.4	4.8	4.5	3.6	2.3	1.0	0.8	0.7	0.7	4.8	3.1	2.4	24			
4	0.4	0.3	0.6	S	1.0	1.7	2.8	1.7	0.4	0.8	0.7	0.0	0.4	0.2	0.2	0.1	0.4	0.8	0.6	0.6	0.4	0.3	1.6	2.8	0.0	2.8	0.8	2.4	24			
5	1.5	1.5	S	3.1	2.3	2.3	1.2	1.6	2.8	2.1	1.9	2.3	2.6	3.7	4.1	3.8	4.1	3.9	3.7	3.6	3.9	4.1	4.4	3.5	1.2	4.4	3.0	2.4	24			
6	3.9	S	11.2	7.2	4.8	4.5	1.4	0.9	0.8	1.7	1.1	1.2	1.1	1.0	1.2	1.1	2.0	3.7	4.1	2.6	2.3	2.2	3.2	3.7	0.8	11.2	2.9	2.4	24			
7	S	3.1	3.4	3.1	3.0	2.1	2.4	2.7	2.1	1.5	C1	C1	C1	C1	C1	C1	C1	0.8	0.6	1.1	1.3	1.5	2.0	S	0.6	3.4	2.0	17	24			
8	1.2	2.3	5.1	6.1	6.5	5.3	2.4	7.1	6.8	4.2	4.2	3.6	3.6	3.2	2.7	1.9	0.7	0.2	0.8	0.8	0.7	1.1	S	5.0	0.2	7.1	3.3	2.4	24			
9	7.5	6.2	6.4	8.8	12.2	12.3	11.5	10.9	7.4	5.7	6.0	3.5	1.3	2.0	1.8	2.8	3.2	3.1	3.5	2.1	1.6	S	1.4	3.7	1.3	12.3	5.4	2.4	24			
10	4.1	2.9	3.9	7.5	11.7	12.1	12.6	10.3	6.4	3.9	3.7	2.6	1.5	0.9	0.5	0.6	0.8	1.3	0.8	0.3	S	1.0	1.0	1.3	0.3	12.6	4.0	2.4	24			
11	1.4	2.3	3.5	3.8	3.7	4.4	4.5	X	X	3.8	3.8	4.1	2.4	2.7	2.6	2.2	2.0	1.4	1.1	S	2.0	2.0	1.1	1.2	1.1	4.5	2.7	2.2	24			
12	1.0	0.7	0.6	0.7	0.4	0.7	0.8	0.7	0.6	0.6	0.4	0.5	1.2	1.1	0.8	0.9	1.0	1.1	S	1.9	1.3	1.2	1.0	1.0	0.4	1.9	0.9	2.4	24			
13	1.3	1.1	1.3	1.2	1.4	1.3	2.3	3.7	2.5	1.6	1.6	1.6	1.9	1.8	1.5	1.6	1.8	S	2.1	1.8	1.9	1.5	1.6	1.6	1.1	3.7	1.7	2.4	24			
14	1.5	1.7	1.8	1.9	1.9	2.1	2.4	2.4	3.7	3.4	4.2	4.4	7.1	6.9	7.3	6.8	S	6.8	6.2	5.5	5.0	4.3	3.6	3.4	1.5	7.3	4.1	2.4	24			
15	3.3	3.1	2.6	2.6	2.5	2.5	2.9	3.7	2.9	2.9	3.0	2.5	1.9	1.1	1.1	S	1.0	0.8	0.3	0.2	0.0	0.1	0.1	0.0	0.0	3.7	1.8	2.4	24			
16	0.2	0.2	1.0	0.2	0.1	0.2	0.4	0.4	0.7	0.8	0.7	0.4	0.5	0.6	S	0.9	0.7	0.9	1.0	1.2	1.4	2.9	2.2	1.8	0.1	2.9	0.8	2.4	24			
17	1.4	1.1	1.1	1.1	1.2	1.4	1.9	2.0	1.7	1.0	1.3	1.1	1.7	S	2.4	2.3	2.2	2.2	2.9	2.2	2.2	1.6	1.4	1.4	1.0	2.9	1.7	2.4	24			
18	1.3	1.6	1.5	1.7	2.0	1.8	1.8	2.0	2.1	2.0	2.0	1.9	S	1.8	1.9	1.9	1.6	1.4	1.3	1.1	1.7	2.3	1.1	1.0	1.0	2.3	1.7	2.4	24			
19	0.7	0.6	0.2	0.0	0.2	0.1	0.2	0.1	0.2	0.4	0.2	S	0.1	0.0	0.1	0.1	0.2	0.2	0.2	0.2	0.3	0.2	0.1	0.2	0.0	0.7	0.2	2.4	24			
20	0.1	0.2	0.3	0.3	0.3	0.3	0.9	0.7	0.9	0.5	S	0.7	0.6	0.4	0.3	0.3	0.3	0.3	0.3	0.3	0.6	0.7	0.4	1.1	0.1	1.1	0.5	2.4	24			
21	0.8	0.7	0.6	1.0	3.2	5.5	5.7	3.7	2.3	S	1.9	1.0	1.2	1.3	1.0	0.7	0.5	0.8	0.9	0.8	0.9	0.8	0.8	0.7	0.5	5.7	1.6	2.4	24			
22	0.6	0.6	0.7	1.1	1.1	1.2	1.2	1.3	S	1.7	1.6	1.6	1.7	1.6	1.4	1.5	2.2	2.0	2.4	2.9	2.3	1.4	0.9	0.9	0.6	2.9	1.5	2.4	24			
23	0.8	0.6	0.7	0.6	0.6	0.7	0.7	S	0.7	0.8	1.4	1.9	2.0	1.0	0.8	0.9	1.0	0.7	1.2	1.0	1.1	1.6	1.6	1.2	0.6	2.0	1.0	2.4	24			
24	1.3	1.2	1.1	1.4	1.9	2.2	S	3.1	5.3	6.3	7.9	4.4	3.6	2.7	3.3	3.3	2.7	2.2	2.0	1.8	1.2	1.3	1.1	1.2	1.1	7.9	2.7	2.4	24			
25	1.6	0.9	1.0	1.0	1.8	S	1.1	2.4	3.5	2.5	1.9	1.7	1.2	1.4	1.4	1.5	1.9	2.0	2.2	2.0	1.5	1.6	1.7	2.1	0.9	3.5	1.8	2.4	24			
26	2.5	2.2	2.5	2.7	S	3.3	4.3	3.9	4.0	4.4	4.2	3.8	3.3	2.7	2.7	3.3	3.6	3.6	3.5	2.9	2.2	2.0	2.2	2.6	2.0	4.4	3.1	2.4	24			
27	3.2	3.2	2.6	S	2.5	1.8	1.4	1.4	1.8	2.5	2.8	2.5	2.3	1.9	1.4	1.0	1.5	2.0	1.6	1.5	1.6	1.2	1.5	1.4	1.0	3.2	1.9	2.4	24			
28	1.5	1.8	S	2.8	2.3	2.2	2.7	3.4	3.1	2.9	2.7	2.4	3.0	4.0	3.7	3.1	2.7	2.6	2.5	2.4	2.4	1.9	1.8	1.8	1.5	4.0	2.6	2.4	24			
29	2.0	S	1.8	1.6	1.7	1.9	1.7	1.6	2.2	2.1	2.3	1.9	1.8	1.7	2.0	1.7	1.5	1.9	1.9	2.3	2.9	2.8	5.6	5.3	1.5	5.6	2.3	2.4	24			
30	S	1.9	2.0	1.0	0.5	0.1	0.3	0.6	0.8	0.8	0.7	0.3	0.8	1.1	1.0	1.1	1.0	1.3	1.3	1.1	1.1	1.1	S	0.1	2.0	1.0	2.4	24				
31	1.2	1.4	1.4	1.9	2.2	2.8	4.6	6.1	7.2	6.9	4.9	3.7	2.4	2.1	2.7	2.1	1.3	1.4	1.7	1.6	1.6	1.4	S	2.2	1.2	7.2	2.8	2.4	24			
HOURLY MAX	7.5	6.2	11.2	8.8	12.2	12.3	12.6	10.9	7.4	6.9	7.9	4.4	7.1	6.9	7.3	6.8	4.1	6.8	6.2	5.5	5.0	4.6	5.6	5.3								
HOURLY AVG	1.9	1.7	2.3	2.4	2.6	2.8	2.9	3.0	2.8	2.5	2.6	2.2	2.0	2.0	2.0	1.9	1.7	1.9	2.0	1.8	1.8	1.7	1.8	2.0								

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	663			
MINIMUM 1-HR AVERAGE:	0.0 ppb	@ HOUR(S)	11	ON DAY(S) 4
MAXIMUM 1-HR AVERAGE:	12.6 ppb	@ HOUR(S)	6	ON DAY(S) 10
MAXIMUM 24-HR AVERAGE:	5.4 ppb			ON DAY(S) 9
				VAR-VARIOUS
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	716 hrs	
MONTHLY CALIBRATION TIME:	16 hrs	AMD OPERATION UPTIME:	96.2 %	
STANDARD DEVIATION:	1.9	MONTHLY AVERAGE:	2.2 ppb	

OXIDES OF NITROGEN Hourly Averages (NO_x ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - March 2017

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	3.4	2.4	2.6	1.9	1.4	1.6	S	1.5	1.6	C	C	C	C	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	1.4	3.4	2.1	16	
2	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	C	C	C	C	6.0	6.3	6.1	5.2	5.2	6.3	5.9	13
3	5.0	5.1	5.2	4.7	S	5.1	5.1	4.7	4.9	5.3	5.3	5.5	5.7	6.8	5.6	4.8	4.9	7.3	7.1	5.2	5.1	2.6	2.3	2.1	2.1	2.1	7.3	5.0	24
4	1.8	1.9	2.2	S	2.6	3.3	4.7	3.6	2.4	2.2	2.8	1.5	2.3	1.7	1.7	1.5	1.9	2.2	2.1	2.2	2.2	2.0	4.5	4.5	1.5	4.7	2.5	24	
5	3.2	2.9	S	5.3	3.9	4.1	2.7	3.6	4.3	3.7	3.5	3.8	4.1	5.7	5.5	5.1	6.6	5.3	5.0	4.9	5.1	5.7	6.0	4.9	2.7	6.6	4.6	24	
6	5.1	S	15.2	13.6	11.0	12.1	2.6	2.1	1.9	6.3	2.3	2.3	2.4	2.4	2.9	2.6	3.6	5.4	5.5	4.5	4.1	3.6	4.8	5.1	1.9	15.2	5.3	24	
7	S	4.5	5.0	4.4	5.1	3.6	4.1	4.6	3.8	2.9	C1	C1	C1	C1	C1	C1	C1	3.7	2.4	7.8	3.3	3.4	3.6	S	2.4	7.8	4.1	17	
8	2.3	4.1	6.3	7.1	8.7	8.8	5.1	9.1	8.8	5.5	5.1	4.9	4.6	4.0	3.7	3.3	1.7	1.0	3.4	2.6	1.3	2.0	S	7.6	1.0	9.1	4.8	24	
9	8.6	8.1	7.6	11.8	14.3	15.3	13.1	11.7	11.0	6.9	6.8	6.8	2.7	2.9	2.6	4.3	4.2	4.0	4.3	3.2	2.4	S	2.9	4.7	2.4	15.3	7.0	24	
10	5.4	4.3	5.7	10.9	13.5	13.8	14.0	13.2	9.2	5.6	5.1	3.9	2.8	2.1	1.5	1.7	2.4	2.4	1.9	1.6	S	2.1	2.1	2.6	1.5	14.0	5.6	24	
11	2.6	4.0	4.7	5.0	4.7	5.6	5.7	X	X	5.1	5.5	5.4	3.3	3.8	3.4	3.2	3.0	2.6	2.0	S	2.9	3.0	2.5	2.1	2.0	5.7	3.8	22	
12	1.9	1.9	1.7	1.6	1.3	1.8	2.0	1.8	1.7	1.9	1.7	1.6	2.7	2.7	2.0	2.1	2.4	2.5	S	3.4	2.6	2.3	2.2	2.2	1.3	3.4	2.1	24	
13	2.5	2.2	2.3	2.3	2.5	2.2	5.0	5.0	3.5	2.6	2.6	2.9	2.9	2.9	2.6	2.7	2.8	S	3.0	2.9	2.9	2.5	2.7	2.7	2.2	5.0	2.9	24	
14	2.7	2.9	3.0	3.2	3.2	3.8	3.9	4.0	36.6	5.4	5.8	7.1	9.1	8.8	9.2	9.1	S	9.1	8.0	7.7	7.0	6.4	5.3	5.3	2.7	36.6	7.2	24	
15	5.1	5.1	4.4	4.2	4.4	4.3	4.8	5.6	5.1	4.8	4.6	4.5	3.9	2.8	2.8	S	3.4	2.6	1.9	2.0	1.7	1.5	1.6	1.7	1.5	5.6	3.6	24	
16	1.7	1.6	2.9	1.6	1.6	2.7	3.0	1.9	3.7	3.6	4.0	1.8	2.7	2.8	S	2.2	2.2	2.2	2.3	5.7	3.9	4.5	3.6	3.2	1.6	5.7	2.8	24	
17	3.0	2.7	2.4	2.4	2.5	3.0	4.3	3.7	3.6	2.5	4.3	2.6	3.2	S	3.7	3.8	3.5	4.0	4.3	3.8	3.6	3.2	2.7	2.8	2.4	4.3	3.3	24	
18	2.7	3.0	3.0	3.2	3.5	3.2	3.2	3.5	3.4	3.4	3.4	S	3.2	3.1	3.1	3.3	2.8	2.8	2.5	3.9	4.0	2.6	2.3	2.3	2.3	4.0	3.1	24	
19	2.1	1.9	1.5	1.4	1.7	1.5	1.6	1.5	1.6	1.7	1.6	S	1.5	1.5	1.5	1.7	1.7	2.3	1.7	1.5	1.5	1.6	1.2	1.2	1.2	2.3	1.6	24	
20	1.3	1.5	1.4	1.3	1.3	1.4	14.2	2.8	4.3	1.7	S	1.6	1.6	1.6	1.4	1.9	1.5	1.4	1.3	1.8	1.8	1.5	1.5	2.1	1.3	14.2	2.3	24	
21	2.0	1.9	1.7	2.2	5.3	7.1	7.4	5.6	3.8	S	3.8	2.4	2.4	2.6	2.7	2.1	1.9	2.1	2.3	2.3	2.2	2.3	2.3	2.3	1.7	7.4	3.1	24	
22	2.2	2.3	2.3	2.7	2.6	3.0	2.7	3.0	S	3.2	3.3	3.3	3.1	3.2	3.2	4.1	4.5	4.3	4.1	5.2	3.9	3.4	2.6	2.4	2.2	5.2	3.2	24	
23	2.3	2.0	2.2	2.0	2.0	2.2	2.1	S	2.1	2.4	2.9	3.6	4.1	2.9	3.4	3.5	3.6	3.4	2.9	3.0	3.2	24.2	4.8	2.8	2.0	24.2	3.8	24	
24	2.8	2.6	2.6	3.1	4.9	5.3	S	6.6	6.7	11.1	11.1	7.0	5.4	4.1	4.6	4.5	4.1	3.7	3.3	3.0	2.6	2.6	2.4	2.8	2.4	11.1	4.6	24	
25	2.8	2.0	2.0	2.0	3.5	S	3.9	4.5	4.9	4.3	3.5	3.2	2.8	2.9	3.9	3.2	3.9	3.6	4.5	4.7	3.3	3.3	3.5	4.5	2.0	4.9	3.5	24	
26	4.5	4.1	4.5	4.8	S	5.7	6.8	6.4	6.3	7.5	7.5	5.9	15.0	4.9	4.6	5.9	5.9	5.7	5.5	5.1	4.3	4.2	4.3	4.7	4.1	15.0	5.8	24	
27	5.3	5.4	4.9	S	5.1	3.7	3.4	3.6	3.7	4.6	4.8	4.7	4.3	4.1	3.3	2.8	3.4	4.1	3.2	3.3	3.3	3.0	3.3	3.0	2.8	5.4	3.9	24	
28	3.2	3.5	S	4.8	3.9	3.9	4.3	5.1	4.8	4.4	4.3	4.2	4.9	5.9	5.5	4.7	4.4	4.2	4.3	4.0	4.1	3.7	3.5	3.4	3.2	5.9	4.3	24	
29	3.5	S	3.4	3.4	3.4	3.7	3.4	3.5	5.1	4.8	4.5	3.9	3.7	3.5	3.8	4.0	3.5	3.9	3.9	4.3	5.1	5.8	8.1	7.8	3.4	8.1	4.3	24	
30	S	4.3	4.1	3.1	2.3	1.8	2.0	2.3	2.6	2.5	2.4	1.9	2.5	2.7	2.4	2.7	2.3	2.4	2.8	2.6	2.4	2.6	2.2	S	1.8	4.3	2.6	24	
31	2.4	2.8	3.0	4.0	3.7	4.5	7.6	9.3	9.1	9.1	6.8	5.8	4.7	5.5	5.0	4.9	3.1	3.3	4.1	3.8	5.4	3.3	S	7.8	2.4	9.3	5.2	24	
HOURLY MAX	8.6	8.1	15.2	13.6	14.3	15.3	14.2	13.2	36.6	11.1	11.1	7.1	15.0	8.8	9.2	9.1	6.6	9.1	8.0	7.8	7.0	24.2	8.1	7.8					
HOURLY AVG	3.3	3.3	3.9	4.2	4.4	4.8	5.1	4.8	5.7	4.5	4.4	3.9	4.0	3.6	3.5	3.5	3.3	3.6	3.6	3.7	3.5	4.0	3.4	3.7					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY


NUMBER OF NON-ZERO READINGS:	668
MAXIMUM INSTANTANEOUS VALUE:	36.6 ppb @ HOUR(S) 8 ON DAY(S) 14
VAR-VARIOUS	
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	16 hrs
OPERATIONAL TIME:	716 hrs
STANDARD DEVIATION:	2.7

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)




% Icon Classes (ppb) 90 0.0-4.2 8 4.2-8.5 2 8.5-12.7 0 >12.7

LICA ST. LINA Poll.: LICA ST. LINA-NOX[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 0.15% Calm Poll Avg: 0.75[ppb]



NOX[ppb] Calibration: LICA ST. LINA Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

NITRIC OXIDES


NITRIC OXIDE Hourly Averages (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.1	C	C	C	C	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	0.0	0.1	0.0	0.0	16			
2	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	C	C	C	C	C	C	0.0	0.0	0.0	0.0	13			
3	0.0	0.0	0.0	0.1	S	0.0	0.1	0.0	0.7	1.0	1.1	1.2	1.4	1.4	1.0	0.7	0.3	0.2	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	1.4	0.4	24			
4	0.0	0.0	0.0	S	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	24			
5	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.1	0.2	0.4	0.6	0.8	1.0	1.1	0.8	0.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.1	0.2	24			
6	0.1	S	0.5	0.3	0.2	0.1	0.0	0.0	0.0	0.4	0.2	0.2	0.1	0.1	0.1	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.1	24			
7	S	0.1	0.0	0.0	0.0	0.0	0.0	0.1	0.2	0.4	C1	C1	C1	C1	C1	C1	C1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.4	0.1	17			
8	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.7	1.4	1.3	1.4	1.3	1.4	1.2	1.0	0.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	1.4	0.4	24			
9	0.0	0.0	0.0	0.0	0.1	0.2	0.3	1.4	2.3	2.5	3.0	1.7	0.5	0.8	0.6	1.1	0.9	0.4	0.2	0.0	0.1	S	0.1	0.0	0.0	0.0	3.0	0.7	24			
10	0.0	0.0	0.0	0.0	0.0	0.1	0.3	1.5	2.2	1.7	1.7	1.3	0.6	0.2	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	2.2	0.4	24			
11	0.0	0.0	0.0	0.0	0.0	0.3	0.9	X	X	1.2	1.4	1.7	0.9	0.7	0.6	0.4	0.2	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.7	0.4	22			
12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.1	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	24			
13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.2	0.2	0.1	0.3	0.3	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.3	0.1	24			
14	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.8	0.7	1.0	1.1	2.3	1.9	1.8	1.3	S	0.5	0.1	0.0	0.0	0.2	0.0	0.0	0.0	0.0	2.3	0.5	24			
15	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.5	0.6	0.6	0.5	0.4	0.3	0.0	0.1	S	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.6	0.1	24			
16	0.1	0.0	0.3	0.1	0.1	0.0	0.3	0.1	0.2	0.3	0.3	0.2	0.1	0.0	S	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.3	0.1	24			
17	0.0	0.0	0.0	0.1	0.1	0.0	0.0	0.0	0.1	0.0	0.2	0.0	0.3	S	0.5	0.3	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.5	0.1	24			
18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.3	0.5	0.6	0.5	S	0.5	0.4	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.6	0.1	24			
19	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	24			
20	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	S	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	24			
21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.2	S	0.5	0.1	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.1	24			
22	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.2	0.2	0.2	0.2	0.0	0.0	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	24			
23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.1	0.0	0.3	0.5	0.5	0.2	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.1	24			
24	0.0	0.0	0.0	0.0	0.0	0.0	S	0.6	1.6	2.2	2.9	1.4	1.0	0.5	0.7	0.6	0.3	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	2.9	0.5	24			
25	0.0	0.0	0.0	0.0	S	0.0	0.1	0.4	0.4	0.3	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.4	0.1	24			
26	0.0	0.0	0.0	0.0	S	0.0	0.0	0.2	0.6	1.1	1.1	0.9	0.6	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.1	0.2	24			
27	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.2	0.6	0.8	0.7	0.6	0.6	0.4	0.1	0.1	0.2	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.8	0.2	24			
28	0.0	0.0	S	0.1	0.0	0.0	0.0	0.3	0.5	0.8	0.8	0.6	0.9	1.1	1.0	0.6	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.1	0.3	24			
29	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.2	0.3	0.3	0.2	0.3	0.4	0.3	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.4	0.1	24			
30	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.1	0.1	0.0	0.2	0.4	0.4	0.4	0.2	0.1	0.1	0.1	0.1	0.0	S	0.0	0.0	0.4	0.1	24			
31	0.1	0.2	0.2	0.3	0.2	0.1	1.0	2.4	3.2	3.0	1.8	1.3	0.6	0.5	0.8	0.6	0.3	0.3	0.4	0.3	0.2	0.3	S	0.5	0.1	0.1	3.2	0.8	24			
HOURLY MAX	0.1	0.2	0.5	0.3	0.2	0.3	1.0	2.4	3.2	3.0	3.0	1.7	2.3	1.9	1.8	1.3	0.9	0.5	0.4	0.3	0.2	0.3	0.1	0.5								
HOURLY AVG	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.3	0.6	0.7	0.8	0.6	0.5	0.4	0.4	0.3	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0							

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	257			
MINIMUM 1-HR AVERAGE:	0.0 ppb	@ HOUR(S)	0	ON DAY(S) 1
MAXIMUM 1-HR AVERAGE:	3.2 ppb	@ HOUR(S)	8	ON DAY(S) 31
MAXIMUM 24-HR AVERAGE:	0.8 ppb			ON DAY(S) 31
				VAR-VARIOUS
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	716 hrs	
MONTHLY CALIBRATION TIME:	16 hrs	AMD OPERATION UPTIME:	96.2 %	
STANDARD DEVIATION:	0.5	MONTHLY AVERAGE:	0.2 ppb	

NITRIC OXIDE Hourly Averages (NO ppb)





NITRIC OXIDE Instantaneous Maximum (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	0.9	0.7	0.9	0.8	0.8	0.8	S	1.1	1.3	C	C	C	C	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	0.7	1.3	0.9	16				
2	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	C	C	C	C	C	C	1.4	1.1	0.7	0.7	0.7	1.4	1.0	13
3	0.6	0.7	0.8	1.1	S	0.9	0.9	1.0	1.8	2.2	2.2	2.5	2.4	2.8	2.4	1.7	1.1	2.1	0.7	0.8	0.9	0.8	1.0	1.1	0.6	2.8	1.4	2.4	2.4	2.4	2.4	24
4	1.1	1.0	1.0	S	1.3	1.3	1.3	1.3	1.0	1.3	1.5	1.1	1.5	1.2	1.2	1.1	1.2	1.2	1.3	1.2	1.1	1.2	1.0	1.3	1.0	1.5	1.2	1.0	1.2	1.0	1.2	24
5	1.3	1.5	S	1.5	1.4	1.3	1.2	1.3	1.7	1.9	2.3	2.1	2.4	2.7	2.6	2.2	2.2	1.7	1.2	1.4	1.5	1.5	1.5	1.5	1.2	2.7	1.7	2.4	2.4	2.4	2.4	24
6	1.6	S	2.0	1.8	1.6	1.8	1.5	1.3	1.5	3.1	1.9	1.7	1.7	1.8	1.7	1.5	1.7	1.8	1.7	1.7	1.9	1.5	1.5	1.5	1.3	3.1	1.7	2.4	2.4	2.4	2.4	24
7	S	1.9	1.8	1.8	1.8	1.8	1.9	2.4	2.2	2.3	C1	C1	C1	C1	C1	C1	C1	2.0	0.8	2.5	0.8	0.5	0.7	S	0.5	2.5	1.7	17	17	17	17	17
8	0.7	0.7	0.7	0.7	0.8	0.7	0.9	2.2	2.2	2.1	2.1	2.4	2.5	2.1	1.8	1.4	0.8	0.7	1.4	1.2	0.7	0.7	S	0.8	0.7	2.5	1.3	24	24	24	24	24
9	0.7	0.8	0.6	0.7	0.9	0.8	1.1	3.2	3.2	3.5	3.9	4.0	1.5	1.6	1.3	1.9	1.9	1.6	0.9	0.9	0.6	S	1.0	0.7	0.6	4.0	1.6	24	24	24	24	24
10	0.7	1.0	0.7	0.7	0.8	1.1	1.1	3.0	3.2	2.8	2.9	2.1	1.5	1.1	0.9	0.8	0.8	0.6	0.7	0.7	S	0.9	0.8	0.7	0.6	3.2	1.3	24	24	24	24	24
11	0.9	0.7	0.8	0.7	1.0	1.6	1.9	X	X	2.3	2.6	2.8	1.8	1.7	1.5	1.5	1.3	0.9	0.7	S	0.9	0.8	0.8	0.8	0.7	2.8	1.3	22	22	22	22	22
12	0.9	0.7	0.7	0.7	0.5	0.8	0.7	0.7	0.9	1.0	0.9	1.0	1.5	1.2	1.1	0.9	1.1	1.0	S	1.1	0.8	0.7	0.8	0.8	0.5	1.5	0.9	24	24	24	24	24
13	0.8	0.6	0.6	0.6	0.8	0.6	1.0	0.9	1.3	1.1	1.0	1.0	1.2	1.2	0.9	1.0	0.9	S	0.9	0.8	0.8	0.8	0.8	0.7	0.6	1.3	0.9	24	24	24	24	24
14	0.7	0.7	0.6	0.7	0.7	0.8	0.9	17.4	1.7	2.0	2.2	3.2	2.8	2.9	2.2	S	1.4	0.9	0.8	0.7	0.9	0.6	0.6	0.6	17.4	2.0	24	24	24	24	24	24
15	0.7	0.8	0.7	0.7	0.7	0.7	0.8	1.2	1.4	1.2	1.2	1.2	1.1	0.7	0.6	S	1.1	0.6	0.5	0.5	0.6	0.7	0.7	0.6	0.5	1.4	0.8	24	24	24	24	24
16	0.8	0.7	1.0	0.7	0.6	0.8	1.1	0.7	1.2	1.5	1.7	0.9	0.9	0.8	S	0.7	0.8	0.6	0.4	1.1	0.7	0.8	0.8	0.8	0.4	1.7	0.9	24	24	24	24	24
17	0.6	0.8	0.7	0.8	0.8	0.6	0.7	0.8	1.3	0.9	1.7	0.7	1.3	S	1.5	1.2	1.1	0.7	0.9	0.7	0.7	0.8	0.7	0.8	0.6	1.7	0.9	24	24	24	24	24
18	0.7	0.7	0.7	1.1	0.8	0.9	0.8	1.0	1.3	1.4	1.5	1.3	S	1.5	1.2	1.2	1.0	0.6	0.7	0.6	0.9	0.7	0.7	0.7	0.6	1.5	1.0	24	24	24	24	24
19	0.8	0.8	0.7	0.7	0.8	0.8	1.3	0.7	0.9	1.1	0.8	S	0.9	1.0	0.9	0.9	0.8	1.3	1.0	0.7	0.9	1.0	0.7	0.7	0.7	1.3	0.9	24	24	24	24	24
20	0.7	0.9	0.8	0.8	0.8	0.9	7.9	1.5	2.4	1.2	S	1.2	1.1	1.1	0.8	1.0	1.0	0.7	0.8	0.7	1.0	0.6	0.7	1.0	0.6	7.9	1.3	24	24	24	24	24
21	0.7	0.7	1.0	1.1	0.9	0.8	0.9	1.3	1.3	S	1.8	1.3	1.3	1.2	1.0	1.0	0.7	0.7	0.8	0.8	1.0	0.7	0.9	0.7	0.7	1.8	1.0	24	24	24	24	24
22	0.7	0.6	0.7	0.7	0.8	0.7	0.6	1.0	S	1.0	1.2	1.3	1.0	0.9	0.9	1.0	0.9	0.9	1.2	0.8	0.7	0.8	0.9	0.6	1.3	0.9	24	24	24	24	24	
23	0.7	0.7	0.8	0.9	0.7	0.7	0.9	S	1.0	1.2	1.4	1.6	1.7	1.4	1.7	1.8	1.9	1.4	1.0	0.8	0.8	9.8	1.5	0.7	0.7	9.8	1.5	24	24	24	24	24
24	0.8	0.8	0.7	0.8	1.4	1.5	S	2.0	2.5	4.5	4.5	2.6	2.1	1.6	1.6	1.3	1.1	1.1	1.0	0.9	0.7	0.9	0.7	0.9	0.7	4.5	1.6	24	24	24	24	24
25	0.7	0.7	0.9	0.8	0.7	S	1.1	1.2	1.4	1.6	1.3	1.3	1.1	1.0	1.3	0.9	1.0	0.9	0.9	0.7	0.9	1.0	1.1	0.6	0.6	1.6	1.0	24	24	24	24	24
26	0.9	0.7	0.8	0.9	S	0.9	1.1	1.7	1.7	2.6	2.9	1.9	12.2	1.3	1.0	1.2	1.2	0.9	0.9	1.0	1.1	1.1	1.1	0.7	0.7	12.2	1.7	24	24	24	24	24
27	1.1	1.0	0.8	S	1.2	1.0	1.1	1.0	1.3	1.8	1.8	1.7	1.6	1.6	1.4	1.1	1.2	1.2	0.9	0.9	0.9	0.8	0.9	0.9	0.8	1.8	1.2	24	24	24	24	24
28	0.7	0.8	S	1.1	1.0	0.7	0.9	1.5	1.5	1.7	1.8	1.6	2.1	2.2	2.1	1.6	1.3	1.1	1.1	0.8	1.0	0.8	0.7	0.8	0.7	2.2	1.3	24	24	24	24	24
29	0.8	S	0.9	0.7	0.7	0.9	0.7	1.0	1.7	1.7	1.7	1.5	1.3	1.5	1.5	1.8	1.1	1.2	0.7	1.0	0.9	0.9	0.9	0.8	0.7	1.8	1.1	24	24	24	24	24
30	S	1.0	0.9	0.9	0.6	0.6	0.9	0.9	0.9	0.9	1.1	1.0	1.2	1.2	1.1	1.3	1.2	1.0	1.1	0.8	0.9	0.9	0.7	S	0.6	1.3	1.0	24	24	24	24	24
31	0.7	0.9	1.0	1.0	0.9	0.8	2.7	4.3	4.2	4.1	2.5	2.1	1.6	1.4	1.5	1.5	0.9	1.0	0.9	0.7	1.0	0.7	S	1.9	0.7	4.3	1.7	24	24	24	24	24
HOURLY MAX	1.6	1.9	2.0	1.8	1.8	1.8	7.9	4.3	17.4	4.5	4.5	4.0	12.2	2.8	2.9	2.2	2.2	2.1	1.7	2.5	1.9	9.8	1.5	1.9								
HOURLY AVG	0.8	0.8	0.9	0.9	0.9	0.9	1.4	1.5	2.3	1.9	1.9	1.7	2.0	1.5	1.4	1.3	1.2	1.1	0.9	1.0	0.9	1.2	0.9	0.9								


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	668
MAXIMUM INSTANTANEOUS VALUE:	17.4 ppb @ HOUR(S) 8 ON DAY(S) 14
VAR-VARIOUS	
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	16 hrs
STANDARD DEVIATION:	1.1
OPERATIONAL TIME:	716 hrs

NITRIC OXIDE Instantaneous Maximum (NO ppb)



% Icon Classes (ppb)


94 0.0-1.1

4 1.1-2.2

1 2.2-3.3

0 >3.3

LICA ST. LINA Poll.: LICA ST. LINA-NO[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 0.15% Calm Poll Avg: 0.00[ppb]



NITROGEN DIOXIDE

NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	2.5	1.8	1.7	1.0	0.9	0.9	S	1.0	0.9	C	C	C	C	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	0.9	2.5	1.3	16
2	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	C	C	C	C	C	C	4.5	4.6	4.5	3.7
3	3.3	3.4	3.3	2.9	S	3.4	3.3	3.2	2.8	2.5	2.2	2.3	2.4	2.8	2.3	2.4	3.1	4.6	4.5	3.6	2.3	1.0	0.8	0.6	0.6	4.6	2.7	24
4	0.4	0.3	0.6	S	0.9	1.7	2.8	1.7	0.4	0.8	0.7	0.0	0.4	0.2	0.2	0.1	0.4	0.8	0.6	0.6	0.4	0.3	1.6	2.8	0.0	2.8	0.8	24
5	1.5	1.5	S	3.1	2.3	2.3	1.2	1.6	2.7	1.9	1.5	1.7	1.9	2.7	3.0	3.1	3.6	3.8	3.7	3.6	3.9	4.1	4.4	3.5	1.2	4.4	2.7	24
6	3.8	S	10.7	6.9	4.6	4.4	1.4	0.9	0.8	1.3	0.9	1.0	1.0	0.9	1.1	1.1	1.9	3.6	4.1	2.6	2.3	2.2	3.2	3.7	0.8	10.7	2.8	24
7	S	2.9	3.4	3.1	3.0	2.1	2.4	2.7	1.9	1.1	C1	C1	C1	C1	C1	C1	C1	0.8	0.6	1.1	1.3	1.5	2.0	S	0.6	3.4	2.0	17
8	1.2	2.3	5.1	6.1	6.5	5.3	2.4	6.4	5.4	2.9	2.8	2.3	2.2	2.0	1.8	1.4	0.7	0.2	0.8	0.8	0.7	1.1	S	5.0	0.2	6.5	2.8	24
9	7.5	6.2	6.4	8.8	12.1	12.2	11.2	9.5	5.1	3.2	3.0	1.8	0.8	1.2	1.2	1.7	2.3	2.7	3.3	2.0	1.5	S	1.4	3.7	0.8	12.2	4.7	24
10	4.1	2.9	3.9	7.5	11.7	12.0	12.4	8.7	4.2	2.2	2.0	1.3	0.9	0.7	0.5	0.6	0.8	1.3	0.8	0.3	S	1.0	1.0	1.3	0.3	12.4	3.6	24
11	1.4	2.3	3.5	3.8	3.7	4.1	3.6	X	X	2.6	2.4	2.4	1.6	2.0	2.0	1.8	1.4	1.1	S	2.0	2.0	1.1	1.2	1.1	4.1	2.3	22	
12	1.0	0.7	0.6	0.7	0.4	0.7	0.8	0.7	0.6	0.6	0.4	0.5	1.0	1.0	0.8	0.9	1.0	1.1	S	1.9	1.3	1.2	1.0	1.0	0.4	1.9	0.9	24
13	1.3	1.1	1.3	1.2	1.4	1.3	2.3	3.7	2.3	1.3	1.4	1.5	1.6	1.5	1.5	1.8	S	2.1	1.8	1.9	1.5	1.6	1.6	1.1	3.7	1.7	24	
14	1.5	1.7	1.8	1.9	1.9	2.1	2.3	2.4	3.0	2.7	3.1	3.3	4.8	4.9	5.6	5.5	S	6.4	6.1	5.5	5.0	4.1	3.6	3.4	1.5	6.4	3.6	24
15	3.3	3.1	2.6	2.5	2.5	2.5	2.9	3.3	2.3	2.3	2.5	2.1	1.6	1.1	1.0	S	0.9	0.8	0.3	0.2	0.0	0.1	0.1	0.0	0.0	3.3	1.7	24
16	0.1	0.2	0.8	0.2	0.1	0.1	0.1	0.3	0.5	0.5	0.4	0.2	0.4	0.6	S	0.8	0.7	0.9	1.0	1.2	1.4	2.9	2.2	1.7	0.1	2.9	0.8	24
17	1.4	1.1	1.1	1.0	1.2	1.4	1.9	1.9	1.6	1.0	1.1	1.1	1.4	S	2.0	2.0	1.9	2.2	2.8	2.2	2.2	1.6	1.4	1.3	1.0	2.8	1.6	24
18	1.3	1.6	1.5	1.7	2.0	1.8	1.8	1.8	1.8	1.5	1.4	1.3	S	1.3	1.5	1.6	1.6	1.4	1.3	1.1	1.7	2.3	1.1	1.0	1.0	2.3	1.5	24
19	0.7	0.6	0.2	0.0	0.2	0.1	0.2	0.1	0.2	0.3	0.2	S	0.1	0.0	0.1	0.1	0.2	0.2	0.2	0.2	0.3	0.2	0.1	0.2	0.0	0.7	0.2	24
20	0.1	0.2	0.3	0.3	0.3	0.3	0.9	0.7	0.7	0.5	S	0.6	0.6	0.4	0.3	0.3	0.3	0.3	0.3	0.2	0.6	0.7	0.4	1.1	0.1	1.1	0.5	24
21	0.8	0.7	0.6	1.0	3.2	5.5	5.7	3.6	2.1	S	1.4	0.9	1.0	1.2	1.0	0.7	0.5	0.8	0.9	0.8	0.9	0.8	0.8	0.7	0.5	5.7	1.5	24
22	0.6	0.6	0.7	1.1	1.1	1.2	1.2	1.3	S	1.4	1.4	1.4	1.5	1.6	1.4	1.5	2.1	2.0	2.4	2.9	2.3	1.4	0.9	0.9	0.6	2.9	1.4	24
23	0.8	0.6	0.7	0.6	0.6	0.7	0.7	S	0.7	0.8	1.1	1.4	1.5	0.8	0.8	0.8	0.8	0.7	1.2	1.0	1.1	1.6	1.6	1.2	0.6	1.6	0.9	24
24	1.3	1.2	1.1	1.4	1.9	2.1	S	2.5	3.7	4.1	5.0	3.0	2.5	2.2	2.6	2.7	2.4	2.2	1.9	1.8	1.2	1.3	1.1	1.2	1.1	5.0	2.2	24
25	1.6	0.9	1.0	1.0	1.8	S	1.1	2.4	3.1	2.1	1.6	1.5	1.2	1.4	1.4	1.5	1.9	2.0	2.2	2.0	1.5	1.6	1.7	2.1	0.9	3.1	1.7	24
26	2.5	2.2	2.5	2.7	S	3.3	4.3	3.7	3.3	3.3	3.1	3.0	2.8	2.6	2.7	3.3	3.6	3.6	3.5	2.9	2.2	2.0	2.2	2.6	2.0	4.3	3.0	24
27	3.2	3.2	2.6	S	2.5	1.8	1.4	1.4	1.6	1.9	2.1	1.8	1.7	1.4	1.0	1.4	1.8	1.6	1.5	1.6	1.2	1.5	1.3	1.0	3.2	1.8	24	
28	1.5	1.8	S	2.7	2.3	2.2	2.7	3.0	2.6	2.1	1.9	1.7	2.1	2.8	2.7	2.5	2.6	2.5	2.4	2.4	1.9	1.8	1.8	1.5	3.0	2.3	24	
29	2.0	S	1.8	1.6	1.7	1.9	1.7	1.6	2.0	1.9	1.9	1.6	1.5	1.3	1.6	1.4	1.4	1.9	1.9	2.3	2.9	2.8	5.6	5.3	1.3	5.6	2.2	24
30	S	1.9	2.0	1.0	0.5	0.1	0.3	0.6	0.8	0.7	0.6	0.3	0.5	0.6	0.7	0.7	0.6	0.8	1.1	1.2	1.0	0.9	1.0	S	0.1	2.0	0.8	24
31	1.0	1.2	1.2	1.6	2.0	2.7	3.6	3.7	3.9	3.8	3.1	2.4	1.8	1.6	1.8	1.5	0.9	1.1	1.3	1.2	1.4	1.1	S	1.7	0.9	3.9	2.0	24
HOURLY MAX	7.5	6.2	10.7	8.8	12.1	12.2	12.4	9.5	5.4	4.1	5.0	3.3	4.8	4.9	5.6	5.5	3.6	6.4	6.1	5.5	5.0	4.6	5.6	5.3				
HOURLY AVG	1.8	1.7	2.3	2.4	2.6	2.8	2.8	2.7	2.2	1.8	1.8	1.6	1.5	1.5	1.6	1.6	1.5	1.9	1.9	1.8	1.8	1.7	1.8	2.0				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 159 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0				
NUMBER OF NON-ZERO READINGS:	663				
MINIMUM 1-HR AVERAGE:	0.0 ppb	@ HOUR(S)	11	ON DAY(S)	4
MAXIMUM 1-HR AVERAGE:	12.4 ppb	@ HOUR(S)	6	ON DAY(S)	10
MAXIMUM 24-HR AVERAGE:	4.7 ppb			ON DAY(S)	9
				VAR-VARIOUS	
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	716 hrs		
MONTHLY CALIBRATION TIME:	16 hrs	AMD OPERATION UPTIME:	96.2 %		
STANDARD DEVIATION:	1.7	MONTHLY AVERAGE:	2.0 ppb		

24 HR AVERAGES March 2017



NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - March 2017

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	3.4	2.5	2.6	1.7	1.7	1.7	S	1.1	1.2	C	C	C	C	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	1.1	3.4	2.0	16
DAY 2	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	C	C	C	C	C	C	C	5.6	6.0	6.0	5.2
DAY 3	4.8	4.7	4.9	4.2	S	4.5	4.4	4.3	4.3	3.6	3.5	3.4	3.7	4.6	3.7	3.5	3.7	6.6	6.4	4.7	4.3	2.1	2.0	1.7	1.7	6.6	4.1	24
DAY 4	1.4	1.5	2.0	S	1.7	2.7	3.8	3.0	2.0	1.5	1.7	0.8	0.9	0.8	0.8	0.8	1.0	1.3	1.4	1.6	1.7	0.9	3.4	3.4	0.8	3.8	1.7	24
DAY 5	2.2	2.1	S	3.9	2.6	2.8	1.7	2.3	3.0	2.2	1.7	2.0	2.1	3.3	3.3	3.2	4.6	4.0	4.1	3.8	4.1	4.5	5.3	3.7	1.7	5.3	3.2	24
DAY 6	3.8	S	13.5	12.1	9.7	10.6	1.5	0.9	0.9	3.5	0.9	1.1	1.3	0.9	1.4	1.4	1.9	4.3	4.3	3.3	2.7	2.6	3.8	4.1	0.9	13.5	3.9	24
DAY 7	S	3.0	3.6	3.3	3.5	2.3	2.6	2.8	2.4	1.5	C1	C1	C1	C1	C1	C1	C1	1.8	1.9	5.4	3.0	3.3	3.3	S	1.5	5.4	2.9	17
DAY 8	2.2	4.5	6.2	6.8	8.8	8.8	5.1	7.6	7.0	4.3	3.4	2.8	2.6	2.5	2.3	2.2	1.4	0.9	2.3	1.9	1.5	1.9	S	7.6	0.9	8.8	4.1	24
DAY 9	8.5	7.9	7.7	11.8	13.6	15.0	12.8	10.7	8.1	3.9	3.7	3.4	1.7	1.8	1.8	2.5	3.0	3.4	3.9	2.9	2.4	S	2.8	4.4	1.7	15.0	6.0	24
DAY 10	5.2	3.6	5.8	10.6	13.3	13.3	13.4	12.3	6.4	3.2	2.3	1.9	1.7	1.3	1.2	1.3	2.0	2.0	1.8	1.5	S	1.6	2.1	2.0	1.2	13.4	4.8	24
DAY 11	2.2	3.3	4.4	4.7	4.5	4.7	4.5	X	X	3.3	3.1	3.0	2.3	2.4	2.2	2.3	2.3	2.0	1.8	S	2.3	2.5	2.2	1.8	1.8	4.7	2.9	22
DAY 12	1.5	1.5	1.5	1.5	1.5	1.4	1.5	1.6	1.5	1.5	1.4	1.4	1.7	1.9	1.7	1.8	1.8	2.0	S	2.9	2.0	2.3	2.0	2.0	1.4	2.9	1.7	24
DAY 13	2.1	2.1	2.3	2.3	2.3	2.3	4.7	4.9	3.4	2.0	2.2	2.2	2.2	2.2	2.0	2.2	2.4	S	2.5	2.8	2.7	2.6	2.6	2.8	2.0	4.9	2.6	24
DAY 14	2.8	2.8	3.1	3.2	3.1	3.6	3.7	3.7	20.5	4.2	4.4	5.3	6.4	6.8	7.5	7.4	S	8.3	7.8	7.3	6.8	5.9	5.3	5.3	2.8	20.5	5.9	24
DAY 15	5.1	4.8	4.4	4.1	4.2	4.2	4.8	5.2	4.4	3.9	3.9	3.8	3.0	2.4	2.3	S	2.8	2.1	1.6	1.7	1.4	1.5	1.3	1.5	1.3	5.2	3.2	24
DAY 16	1.7	1.7	2.8	1.7	1.5	2.4	2.3	1.7	3.1	2.4	2.6	1.5	1.9	2.4	S	1.5	1.9	2.1	2.1	5.4	3.7	4.1	3.7	3.0	1.5	5.4	2.5	24
DAY 17	2.8	2.3	2.3	2.3	2.4	2.9	4.1	3.7	3.2	2.3	2.8	2.2	2.7	S	3.2	3.2	3.1	3.9	4.1	3.7	3.6	3.0	2.6	2.6	2.2	4.1	3.0	24
DAY 18	2.6	2.9	2.7	3.0	3.3	3.2	3.0	2.8	2.8	2.6	2.5	2.5	S	2.4	2.4	2.4	2.4	2.4	2.6	2.4	3.6	3.9	2.4	2.2	2.2	3.9	2.7	24
DAY 19	2.1	1.8	1.7	1.4	1.4	1.4	1.4	1.5	1.3	1.5	1.4	S	1.2	1.1	1.2	1.4	1.4	2.0	1.3	1.2	1.2	1.5	0.9	0.9	0.9	2.1	1.4	24
DAY 20	0.9	1.2	1.2	1.2	1.1	1.2	9.3	1.9	2.3	1.2	S	1.3	1.1	1.1	1.1	1.3	0.9	1.1	1.2	1.5	1.5	1.4	1.4	1.9	0.9	9.3	1.7	24
DAY 21	2.0	1.7	1.4	2.3	5.3	7.2	7.2	5.3	3.1	S	2.4	1.7	1.8	2.1	2.0	1.7	1.9	1.9	1.9	2.2	2.2	2.0	2.0	2.0	1.4	7.2	2.7	24
DAY 22	2.0	2.1	2.1	2.2	2.5	2.8	2.8	2.8	S	2.7	2.7	2.6	2.5	2.8	2.7	3.5	4.1	3.5	3.7	4.6	3.6	3.2	2.2	2.2	2.0	4.6	2.9	24
DAY 23	2.1	2.0	1.9	1.7	1.7	1.9	1.7	S	1.7	1.7	2.2	2.8	2.9	2.2	2.2	2.0	2.4	2.5	2.5	3.0	2.8	16.7	4.1	2.7	1.7	16.7	2.9	24
DAY 24	2.7	2.5	2.4	2.9	3.8	4.1	S	5.1	5.1	6.9	7.2	4.9	3.9	3.2	3.9	3.6	3.4	3.4	3.0	2.8	2.5	2.5	2.2	2.4	2.2	7.2	3.7	24
DAY 25	2.6	1.8	1.8	1.9	3.3	S	3.1	3.7	4.0	3.3	2.8	2.6	2.3	2.5	3.0	2.7	3.0	3.1	4.4	4.6	3.1	3.1	3.1	4.4	1.8	4.6	3.1	24
DAY 26	4.5	3.9	4.4	4.7	S	5.4	6.4	5.3	4.9	5.3	5.2	4.8	7.0	4.2	4.3	5.1	5.2	5.3	5.3	4.8	4.1	3.8	4.0	4.7	3.8	7.0	4.9	24
DAY 27	5.1	5.3	4.7	S	4.4	3.4	3.2	2.9	3.0	3.5	3.4	3.3	3.3	3.2	2.6	2.4	3.1	3.4	3.1	3.1	2.9	3.1	2.9	2.4	2.4	5.3	3.4	24
DAY 28	3.1	3.4	S	4.1	3.4	3.6	4.1	4.4	3.9	3.3	3.1	3.0	3.3	4.5	4.0	3.7	3.9	3.7	3.9	3.7	3.9	3.6	3.4	3.1	3.0	4.5	3.7	24
DAY 29	3.1	S	3.1	3.0	3.2	3.1	3.2	3.0	3.8	3.4	3.2	3.0	2.8	2.8	2.7	2.8	2.8	3.5	3.5	3.9	4.7	5.4	8.3	7.5	2.7	8.3	3.7	24
DAY 30	S	3.6	3.7	2.8	2.2	1.9	1.9	1.9	2.0	1.9	1.8	1.4	1.9	1.9	1.9	1.8	1.9	2.3	2.5	2.3	2.1	2.2	2.4	S	1.4	3.7	2.2	24
DAY 31	2.2	2.4	2.6	3.6	3.6	4.1	5.4	5.3	5.5	5.7	4.7	4.1	3.6	4.4	3.7	3.5	2.6	2.7	3.7	3.1	4.5	2.7	S	6.4	2.2	6.4	3.9	24
HOURLY MAX	8.5	7.9	13.5	12.1	13.6	15.0	13.4	12.3	20.5	6.9	7.2	5.3	7.0	6.8	7.5	7.4	5.2	8.3	7.8	7.3	6.8	16.7	8.3	7.6				
HOURLY AVG	3.0	3.0	3.6	3.9	4.1	4.4	4.4	4.0	4.1	3.1	3.0	2.7	2.7	2.7	2.6	2.6	2.6	3.1	3.2	3.3	3.1	3.4	3.1	3.4				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY


NUMBER OF NON-ZERO READINGS:	668
MAXIMUM INSTANTANEOUS VALUE:	20.5 ppb @ HOUR(S) 8 ON DAY(S) 14
	VAR-VARIOUS
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	16 hrs
OPERATIONAL TIME:	716 hrs
STANDARD DEVIATION:	2.2

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)



% Icon	Classes (ppb)	93		0.0-4.2	5		4.2-8.3	2		8.3-12.5	0		>12.5
--------	---------------	----	---	---------	---	---	---------	---	--	----------	---	---	-------

LICA ST. LINA Poll.: LICA ST. LINA-NO2[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 0.15% Calm Poll Avg: 0.75[ppb]



NO2[ppb] Calibration: LICA ST. LINA Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

OZONE

OZONE Hourly Averages (O₃ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	28.4	28.4	28.4	29.8	30.2	29.5	S	28.5	27.2	27.1	27.5	28.6	C	C	C	C	C	C	C	Y	Y	Y	Y	Y	27.1	30.2	28.5	19				
2	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	29.2	28.7	29.4	30.6	28.7	30.6	29.5	9
3	31.5	30.7	30.3	30.7	S	28.7	27.6	26.6	26.6	27.9	30.8	31.8	33.7	33.7	34.1	35.4	35.7	32.6	32.9	33.1	37.4	43.3	43.3	43.8	26.6	43.8	33.1	24				
4	45.2	44.6	38.0	S	30.4	30.1	27.8	31.4	34.8	36.3	36.6	37.5	38.5	39.7	40.2	40.7	40.6	40.1	39.7	39.6	39.8	39.6	37.5	36.6	27.8	45.2	37.6	24				
5	38.5	38.6	S	37.7	38.2	38.4	39.6	39.2	38.2	39.3	39.5	39.5	39.3	38.4	38.0	37.9	37.3	36.9	36.5	36.1	35.8	34.9	34.5	34.6	34.5	39.6	37.7	24				
6	34.1	S	25.9	30.7	32.0	32.3	35.5	36.4	36.3	36.7	37.5	37.6	38.3	38.1	38.1	37.7	36.8	33.4	32.8	34.8	35.4	35.1	33.6	32.8	25.9	38.3	34.9	24				
7	S	33.2	31.5	31.0	30.2	30.8	29.8	29.0	29.3	30.5	31.1	31.3	32.2	33.3	34.9	35.5	38.5	38.1	37.7	37.3	37.9	38.0	37.6	S	29.0	38.5	33.6	24				
8	36.3	32.7	28.9	27.1	25.5	26.6	28.1	24.9	28.9	33.2	35.1	36.6	37.1	37.6	37.7	38.2	39.4	39.6	38.4	38.2	38.2	38.0	S	33.2	24.9	39.6	33.9	24				
9	29.6	30.3	29.6	26.6	23.2	22.4	23.0	24.6	28.0	29.6	29.8	31.9	34.8	34.1	34.2	33.7	33.6	33.8	33.9	35.3	35.9	S	36.8	32.7	22.4	36.8	30.8	24				
10	31.6	33.4	32.4	26.2	21.0	19.8	18.3	20.8	23.8	26.4	25.9	25.7	25.9	26.5	28.3	28.7	29.0	28.8	28.3	27.8	S	27.3	26.6	26.3	18.3	33.4	26.5	24				
11	26.1	25.7	25.0	24.6	24.5	26.1	28.1	X	X	30.1	30.7	31.3	33.7	34.1	34.3	34.8	35.6	36.4	36.5	S	38.4	37.6	37.9	36.1	24.5	38.4	31.8	22				
12	35.8	36.1	36.2	36.0	35.6	35.8	35.9	36.0	36.0	36.6	37.5	38.9	40.1	40.4	40.6	41.1	42.0	41.7	S	41.6	43.4	42.8	41.5	40.6	35.6	43.4	38.8	24				
13	39.9	40.0	40.4	40.4	40.3	40.2	38.2	35.8	37.3	38.5	38.9	39.5	40.0	40.6	41.9	42.8	42.8	S	42.2	41.7	41.3	41.6	41.8	41.8	35.8	42.8	40.3	24				
14	42.0	41.7	41.2	41.0	40.2	39.5	38.7	38.3	37.9	36.1	38.6	39.7	38.0	39.2	39.7	41.6	S	45.5	47.8	45.6	44.0	43.3	42.5	41.4	36.1	47.8	41.0	24				
15	39.9	38.6	37.8	37.2	35.8	34.5	32.8	31.4	32.4	32.9	35.4	39.8	45.0	47.8	49.5	S	49.6	48.5	47.5	46.7	47.2	48.3	49.4	48.9	31.4	49.6	41.6	24				
16	48.3	48.8	48.0	48.0	46.5	45.0	44.2	43.5	42.5	43.1	44.5	47.0	48.2	49.0	S	49.6	49.4	48.2	46.9	46.0	42.3	37.9	37.1	35.2	35.2	49.6	45.2	24				
17	36.9	37.8	38.8	36.7	34.8	33.1	32.1	31.9	35.3	42.8	40.8	40.7	41.2	S	41.6	42.7	42.7	42.2	40.4	39.6	37.9	36.5	37.2	37.7	31.9	42.8	38.3	24				
18	36.5	34.6	33.4	32.3	30.2	29.2	28.7	28.6	29.1	29.4	29.8	30.7	S	33.1	33.4	33.6	32.7	31.5	31.1	30.7	29.5	29.0	31.4	31.5	28.6	36.5	31.3	24				
19	33.1	32.7	38.2	42.6	38.9	38.3	39.3	39.7	39.7	41.8	43.6	S	46.8	47.8	47.8	47.0	45.7	45.0	43.9	43.3	42.9	43.0	43.7	41.3	32.7	47.8	42.0	24				
20	39.3	38.5	40.1	40.5	41.5	41.9	41.0	40.7	40.8	40.9	S	41.3	41.5	42.4	43.8	43.5	43.4	42.9	42.9	41.9	42.1	42.2	41.6	39.1	38.5	43.8	41.5	24				
21	37.1	38.6	39.5	37.8	33.7	30.0	29.7	33.3	37.3	S	39.2	41.0	41.0	40.7	41.1	40.9	39.9	38.9	37.6	37.1	36.6	36.3	35.6	35.2	29.7	41.1	37.3	24				
22	34.8	34.5	33.9	33.0	32.2	31.6	31.1	30.6	S	29.5	29.2	29.5	29.4	29.4	30.5	30.8	31.5	31.0	29.7	28.1	27.7	29.5	30.3	30.1	27.7	34.8	30.8	24				
23	29.5	30.3	30.7	31.1	31.0	31.5	32.3	S	34.1	33.7	32.6	32.5	33.3	35.3	36.8	38.0	38.7	38.2	38.8	36.3	34.9	33.5	32.5	32.2	29.5	38.8	33.8	24				
24	31.6	31.1	30.7	31.9	31.1	30.5	S	31.4	30.3	30.0	29.6	33.7	34.8	34.1	33.7	34.3	34.5	36.1	38.0	37.7	36.4	35.6	35.4	34.7	29.6	38.0	33.4	24				
25	33.4	34.8	34.9	34.4	S	31.7	31.3	30.7	32.2	34.7	36.8	38.4	39.5	40.7	40.8	40.3	39.5	39.0	39.3	40.9	39.5	38.2	37.1	30.7	30.7	40.9	36.5	24				
26	36.6	34.8	34.5	34.5	S	31.4	28.7	29.1	30.1	30.0	31.5	33.3	38.3	42.1	46.1	46.9	45.8	41.3	38.7	38.4	35.7	33.6	34.0	33.3	28.7	46.9	36.0	24				
27	30.7	29.2	28.2	S	28.3	29.7	31.3	32.1	32.0	30.4	30.2	31.0	31.8	33.5	35.8	36.7	34.4	31.8	30.5	29.7	29.5	28.9	28.2	27.1	27.1	36.7	30.9	24				
28	25.8	25.3	S	26.8	26.4	24.5	22.4	20.8	20.8	21.6	21.8	21.4	21.0	20.7	22.3	24.4	25.5	25.7	24.3	22.0	23.0	23.0	23.1	23.7	20.7	26.8	23.3	24				
29	22.1	S	21.9	23.0	23.3	22.9	23.5	24.7	24.7	26.0	26.6	26.8	25.8	26.5	27.7	28.3	30.6	32.7	31.1	31.8	30.2	27.3	22.1	19.5	19.5	32.7	26.0	24				
30	S	22.6	21.9	22.9	24.7	23.7	24.5	25.3	25.4	26.1	26.9	27.5	28.6	27.8	28.4	29.6	30.8	31.3	30.7	28.1	33.0	31.4	30.2	S	21.9	33.0	27.3	24				
31	26.1	24.1	22.6	20.8	17.8	15.8	13.6	13.4	16.9	22.1	32.8	35.9	38.4	39.8	32.3	30.7	33.9	36.0	36.3	36.7	36.8	35.1	S	31.7	13.4	39.8	28.2	24				
HOURLY MAX	48.3	48.8	48.0	48.0	46.5	45.0	44.2	43.5	42.5	43.1	44.5	47.0	48.2	49.0	49.5	49.6	49.6	48.5	47.8	46.7	47.2	48.3	49.4	48.9								
HOURLY AVG	34.3	34.0	33.0	32.7	31.4	30.8	30.6	30.7	31.7	32.4	33.4	34.4	36.3	36.6	36.9	37.4	37.9	37.4	36.9	36.6	36.7	35.9	35.5	34.6								

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 82 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0				
NUMBER OF NON-ZERO READINGS:	678				
MINIMUM 1-HR AVERAGE:	13.4 ppb	@ HOUR(S)	7	ON DAY(S)	31
MAXIMUM 1-HR AVERAGE:	49.6 ppb	@ HOUR(S)	16	ON DAY(S)	15
MAXIMUM 24-HR AVERAGE:	45.2 ppb			ON DAY(S)	16
				VAR-VARIOUS	
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	722 hrs		
MONTHLY CALIBRATION TIME:	12 hrs	AMD OPERATION UPTIME:	97.0 %		
STANDARD DEVIATION:	6.6	MONTHLY AVERAGE:	34.5 ppb		

24 HR AVERAGES March 2017



OZONE Hourly Averages (O₃ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - March 2017

OZONE Instantaneous Maximum (O₃ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	28.9	29.1	29.0	30.5	30.5	29.7	S	28.9	28.1	27.4	27.8	29.4	C	C	C	C	C	C	C	Y	Y	Y	Y	Y	27.4	30.5	29.0	19
2	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	C	C	C	C	C	30.6	32.9	31.6	9
3	33.2	32.9	32.4	32.5	S	30.8	29.8	28.5	28.8	30.9	33.9	34.0	36.2	36.1	36.1	37.9	38.1	36.5	34.8	35.0	43.2	46.0	46.0	45.8	28.5	46.0	35.6	24
4	49.2	47.5	45.1	S	32.5	32.5	30.6	33.9	37.5	38.4	38.8	39.6	41.0	41.7	42.3	42.6	42.5	41.9	41.5	41.5	41.7	41.4	40.9	38.9	30.6	49.2	40.2	24
5	40.5	40.3	S	40.2	41.8	41.0	41.4	41.3	40.3	41.3	41.4	41.4	41.3	40.6	39.9	39.8	39.4	38.9	38.4	38.0	37.8	37.5	40.1	37.3	37.3	41.8	40.0	24
6	36.2	S	31.9	35.1	35.4	37.6	38.8	39.6	38.2	39.0	39.2	39.6	40.6	40.1	40.3	39.7	38.8	37.6	35.6	36.9	37.2	36.9	36.6	35.1	31.9	40.6	37.7	24
7	S	35.4	33.6	33.4	32.5	32.7	31.9	31.5	31.5	32.5	33.0	33.4	34.2	35.8	36.8	39.4	40.7	40.5	39.8	39.6	40.2	40.5	41.5	S	31.5	41.5	35.9	24
8	38.3	36.6	31.7	30.0	27.7	30.8	30.8	28.0	33.1	36.1	38.0	38.4	39.2	39.4	39.4	40.3	41.4	41.5	40.8	40.1	39.9	39.8	S	36.3	27.7	41.5	36.4	24
9	32.6	32.4	31.8	30.1	25.7	25.9	25.4	27.4	31.3	31.7	31.5	35.8	37.3	36.5	36.1	35.8	36.0	36.0	36.6	37.2	38.0	S	39.7	36.0	25.4	39.7	33.3	24
10	34.7	35.6	35.4	31.4	24.1	22.6	21.3	23.9	26.4	29.2	28.6	27.4	27.9	28.5	30.0	30.8	30.8	30.6	30.5	29.4	S	29.2	28.5	28.1	21.3	35.6	28.9	24
11	28.0	27.9	26.9	26.2	26.2	28.5	30.9	X	X	32.3	32.9	34.7	35.6	36.1	36.1	36.6	37.7	38.1	38.4	S	40.2	39.4	40.1	38.4	26.2	40.2	33.9	22
12	37.6	37.9	38.0	37.9	37.2	37.6	37.6	37.7	37.9	38.5	39.8	41.5	41.9	42.5	42.5	43.6	44.0	43.4	S	44.2	45.4	45.1	43.8	42.6	37.2	45.4	40.8	24
13	41.8	41.9	42.0	42.0	42.2	42.1	41.5	38.4	39.7	40.2	40.9	41.5	41.9	42.7	44.0	44.7	44.6	S	44.2	43.7	43.2	43.5	43.5	43.7	38.4	44.7	42.3	24
14	43.8	43.6	43.2	43.0	42.3	41.7	40.9	40.1	40.1	40.5	40.5	42.2	40.7	41.1	41.9	44.6	S	49.8	50.2	48.1	46.4	45.2	44.7	43.5	40.1	50.2	43.4	24
15	42.6	40.8	39.7	39.4	38.1	36.8	35.3	33.5	34.6	36.0	38.7	45.0	48.6	50.3	51.7	S	52.0	50.9	49.5	48.6	49.9	51.2	51.5	51.1	33.5	52.0	44.2	24
16	50.3	51.3	50.7	51.1	48.5	47.9	46.5	45.8	44.7	45.4	48.4	49.1	50.6	51.2	S	51.7	51.5	50.7	48.7	48.6	46.7	40.3	39.9	39.7	39.7	51.7	47.8	24
17	40.3	41.0	43.0	40.7	37.2	36.1	35.2	35.6	43.5	45.7	43.5	43.0	43.2	S	44.2	45.0	44.7	44.6	42.3	41.7	40.5	38.6	39.0	39.8	35.2	45.7	41.2	24
18	39.1	37.0	35.3	34.4	32.9	31.1	30.7	30.5	31.2	31.2	31.8	33.1	S	35.0	35.4	35.4	35.2	33.8	33.1	32.9	32.2	32.2	33.5	33.6	30.5	39.1	33.5	24
19	35.8	38.5	43.1	44.7	42.8	41.1	41.4	41.7	42.0	44.7	46.0	S	48.9	49.8	50.3	49.0	48.3	47.2	46.1	45.2	44.9	45.0	45.6	44.4	35.8	50.3	44.6	24
20	42.0	41.4	42.3	42.2	43.7	43.7	43.2	42.5	42.7	42.7	S	43.2	43.5	45.8	45.6	45.5	45.2	44.9	45.0	44.2	44.2	44.2	44.0	42.2	41.4	45.8	43.6	24
21	39.7	40.9	41.8	40.2	37.7	32.7	32.5	37.2	40.3	S	42.6	43.0	43.0	43.1	43.0	43.2	41.9	41.0	40.0	38.9	38.5	38.3	37.5	37.0	32.5	43.2	39.7	24
22	36.6	36.3	35.8	35.1	34.2	33.5	33.1	32.5	S	31.4	31.1	31.4	31.2	31.5	33.1	33.0	34.7	33.2	31.9	30.6	29.7	32.1	32.1	32.1	29.7	36.6	32.9	24
23	32.2	32.4	33.1	33.4	33.4	34.2	34.6	S	36.2	36.3	34.8	34.6	35.6	37.6	39.3	40.2	40.8	40.2	40.7	40.2	37.0	36.1	34.8	34.4	32.2	40.8	36.2	24
24	33.5	33.4	32.7	35.1	35.1	32.9	S	34.0	32.4	32.7	32.5	36.7	37.0	36.3	35.8	36.5	36.7	39.0	40.0	39.9	38.6	37.5	37.2	37.0	32.4	40.0	35.8	24
25	36.1	36.9	36.8	36.6	35.1	S	33.6	33.8	33.0	34.8	37.3	40.0	40.6	41.9	42.7	43.0	42.6	41.7	41.3	42.2	43.0	42.6	40.1	40.0	33.0	43.0	38.9	24
26	38.5	38.0	37.6	37.1	S	34.6	31.5	31.4	32.4	32.7	34.4	37.0	43.7	46.3	51.5	50.5	51.7	45.2	41.4	40.5	39.2	35.3	36.1	35.8	31.4	51.7	39.2	24
27	33.8	31.0	30.8	S	31.1	31.8	34.2	34.2	34.2	33.1	32.9	32.9	33.9	37.0	38.1	38.9	37.8	33.9	32.5	31.7	31.7	30.8	30.4	29.3	29.3	38.9	33.3	24
28	27.7	27.3	S	28.9	28.8	27.1	25.0	23.1	22.9	23.4	24.6	23.1	23.0	22.7	25.4	26.9	27.6	27.6	27.2	24.3	25.4	25.2	25.2	26.0	22.7	28.9	25.6	24
29	24.6	S	25.2	25.6	26.4	26.1	26.4	27.4	27.0	28.5	28.9	29.6	28.4	29.3	30.8	30.5	36.9	35.2	33.6	34.0	34.0	29.5	27.0	22.5	22.5	36.9	29.0	24
30	S	24.7	24.1	25.6	26.8	26.0	26.8	27.7	27.6	28.8	29.1	29.8	30.6	30.4	30.6	32.5	33.1	33.2	32.7	32.4	37.9	35.4	34.2	S	24.1	37.9	30.0	24
31	28.5	26.6	25.1	23.9	21.3	18.1	16.8	16.0	20.7	29.6	38.8	38.4	43.2	43.2	39.3	33.2	37.2	38.0	38.4	38.9	39.0	37.6	S	34.4	16.0	43.2	31.6	24
HOURLY MAX	50.3	51.3	50.7	51.1	48.5	47.9	46.5	45.8	44.7	45.7	48.4	49.1	50.6	51.2	51.7	51.7	52.0	50.9	50.2	48.6	49.9	51.2	51.5	51.1				
HOURLY AVG	36.6	36.4	35.6	35.2	34.0	33.4	33.1	33.1	34.2	35.0	35.9	36.9	38.7	39.0	39.4	39.7	40.4	39.8	39.1	38.9	39.2	38.2	38.0	37.1				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY


NUMBER OF NON-ZERO READINGS:	678
MAXIMUM INSTANTANEOUS VALUE:	52.0 ppb @ HOUR(S) 16 ON DAY(S) 15
VAR-VARIOUS	
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	12 hrs
OPERATIONAL TIME:	722 hrs
STANDARD DEVIATION:	6.6

OZONE Instantaneous Maximum (O₃ ppb)




% Icon Classes (ppb) 0 0.0-16.5 41 16.5-33.1 58 33.1-49.6 0 >49.6

LICA ST. LINA Poll.: LICA ST. LINA-O3[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 0.15% Calm Poll Avg: 38.22[ppb]



O3[ppb] Calibration: LICA ST. LINA Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

PARTICULATE MATTER 2.5

PARTICULATE MATTER < 2.5 MICRONS Hourly Averages (PM_{2.5} µg/m³)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	X	X	6.4	11.9	1.4	7.5	5.4	0.9	1.9	10.4	3.0	C	C	C	0.0	0.0	0.0	0.0	1.4	0.0	2.9	4.0	5.9	3.4	0.0	11.9	3.5	22
2	3.9	0.4	1.9	3.4	4.4	2.9	4.4	2.4	5.4	4.4	10.9	7.9	7.9	9.9	9.4	15.5	11.9	15.5	12.4	12.4	7.5	9.9	9.4	3.4	0.4	15.5	7.4	24
3	6.4	10.9	5.9	7.5	5.4	6.4	4.4	5.4	5.9	8.4	6.9	5.4	4.4	4.4	9.9	X	0.0	7.5	9.5	0.9	6.4	1.9	0.9	2.9	0.0	10.9	5.5	23
4	2.9	0.9	0.0	7.5	0.0	0.4	3.4	1.4	2.4	3.4	4.0	6.4	5.9	0.9	0.0	X	0.0	0.0	0.0	0.0	0.4	1.9	0.9	5.0	0.0	7.5	2.1	23
5	0.0	0.0	0.0	3.4	1.9	0.9	1.4	1.4	2.4	0.0	0.9	1.4	0.0	1.4	0.0	0.0	1.9	1.9	2.4	0.9	0.0	0.4	1.9	5.9	0.0	5.9	1.3	24
6	1.9	1.9	0.4	0.9	1.4	0.9	3.9	6.4	6.4	X	3.4	1.4	3.4	6.9	4.4	0.9	0.0	0.0	0.9	4.0	1.9	0.0	3.5	5.0	0.0	6.9	2.6	23
7	1.9	3.5	4.0	2.4	5.9	2.4	4.4	7.5	0.0	2.4	2.4	8.4	3.9	3.9	0.0	0.0	0.0	1.4	1.9	5.4	0.0	2.9	6.9	3.4	0.0	8.4	3.1	24
8	3.4	3.9	4.9	2.4	1.9	5.9	0.9	3.4	3.5	2.4	5.0	1.4	0.0	0.0	2.4	0.4	0.0	1.4	0.0	3.4	5.4	5.4	5.0	0.0	5.9	2.6	24	
9	1.4	1.4	5.9	3.4	2.0	4.4	6.4	5.0	5.4	1.0	5.4	5.5	3.4	6.3	1.0	2.0	0.0	0.0	0.4	1.4	0.4	4.4	0.0	4.4	0.0	6.4	3.0	24
10	3.4	2.5	4.0	3.4	10.4	2.4	0.0	8.4	2.4	9.1	5.5	2.9	2.9	4.4	2.4	3.9	0.9	5.0	2.4	4.4	5.9	4.0	2.9	3.4	0.0	10.4	4.0	24
11	7.9	0.9	0.0	6.4	1.9	3.9	3.5	X	X	4.4	3.4	0.9	0.0	2.9	1.9	4.4	4.0	5.9	3.9	2.4	3.9	1.4	2.9	0.0	0.0	7.9	3.0	22
12	0.0	2.9	0.4	2.4	2.9	3.9	1.9	2.4	3.4	2.5	3.4	2.9	6.4	6.4	4.0	5.9	1.9	3.5	2.9	3.4	2.9	0.0	1.4	3.4	0.0	6.4	3.0	24
13	0.0	0.0	6.9	3.9	11.5	2.4	0.0	5.4	3.9	4.4	4.5	1.3	9.0	2.8	3.5	2.9	3.5	9.5	6.9	2.4	4.5	2.4	4.4	6.5	0.0	11.5	4.3	24
14	0.4	5.4	8.4	5.9	4.6	6.4	8.9	8.0	5.9	2.4	2.4	11.3	10.5	8.5	8.9	8.0	8.8	11.4	10.4	10.4	12.8	12.1	8.3	6.0	0.4	12.8	7.8	24
15	2.3	5.9	7.3	8.5	4.9	3.9	9.1	5.3	11.6	9.0	13.6	5.9	2.5	4.4	2.5	0.4	0.0	1.4	2.9	0.0	2.1	0.0	1.2	1.6	0.0	13.6	4.4	24
16	0.0	0.9	3.0	0.0	0.0	0.0	0.4	0.0	0.4	0.8	1.7	0.0	0.5	0.8	1.1	2.4	3.8	1.5	1.9	2.8	3.3	5.0	4.4	3.5	0.0	5.0	1.6	24
17	1.9	3.8	5.3	2.3	4.7	5.4	7.9	6.3	5.9	1.5	6.6	0.1	7.9	4.3	5.4	9.5	13.1	9.1	9.2	5.1	6.3	8.2	3.3	5.3	0.1	13.1	5.8	24
18	4.6	2.4	4.8	4.3	3.8	2.6	0.0	6.4	0.0	3.9	5.3	7.6	0.0	4.9	7.9	2.8	4.6	1.6	0.0	0.0	1.9	1.4	6.8	3.6	0.0	7.9	3.4	24
19	0.0	1.2	1.0	3.7	0.3	3.2	2.6	0.6	4.2	0.0	0.8	0.3	X	0.0	3.3	0.0	X	0.0	4.7	0.0	0.0	2.9	0.0	1.9	0.0	4.7	1.4	22
20	7.3	0.0	0.4	0.0	1.1	2.1	0.0	2.4	1.1	0.0	1.3	0.0	0.4	3.8	0.0	X	0.0	X	0.0	0.0	0.4	1.4	1.4	2.7	0.0	7.3	1.2	22
21	5.2	0.7	5.4	0.4	0.9	0.2	5.2	1.0	0.0	2.7	4.4	2.7	3.2	4.4	4.9	2.9	0.4	2.0	0.2	3.2	3.1	1.4	2.1	1.2	0.0	5.4	2.4	24
22	0.9	4.7	5.0	5.8	4.9	5.0	6.7	4.0	4.2	5.4	3.4	5.4	5.4	5.2	7.2	1.4	7.8	1.6	3.7	4.4	2.2	5.9	3.7	7.7	0.9	7.8	4.7	24
23	3.5	6.4	5.2	6.4	3.7	6.6	0.0	7.1	0.0	0.4	3.2	9.7	2.9	6.2	5.0	0.2	6.0	5.2	7.4	5.7	4.1	5.7	4.2	9.6	0.0	9.7	4.8	24
24	10.8	7.9	2.9	2.8	0.9	3.6	4.2	6.1	6.2	5.4	4.1	5.1	7.1	8.1	3.6	11.7	7.6	7.9	2.3	6.4	7.3	9.4	10.8	6.3	0.9	11.7	6.2	24
25	4.6	9.9	7.2	0.0	9.1	0.1	5.5	0.3	9.5	5.4	8.6	6.7	9.0	5.7	8.3	6.2	12.3	12.1	13.4	12.5	11.4	15.1	10.3	11.6	0.0	15.1	8.1	24
26	19.5	9.0	15.3	10.7	13.1	13.1	12.9	14.9	14.5	13.6	14.9	16.6	17.9	15.9	15.9	14.3	15.4	14.9	10.9	9.9	9.9	11.7	5.6	6.7	5.6	19.5	13.2	24
27	7.1	5.7	0.0	2.3	1.1	8.6	4.0	2.9	12.9	3.7	2.5	5.4	6.4	14.5	7.9	7.3	6.9	4.5	2.7	2.0	4.6	2.4	3.9	7.7	0.0	14.5	5.3	24
28	4.4	7.9	6.9	8.0	10.4	4.0	4.3	6.2	3.4	6.9	C	C	7.9	9.9	8.4	9.0	9.0	9.9	10.9	11.9	5.0	4.4	5.0	2.4	2.4	11.9	7.1	24
29	9.0	1.9	3.4	2.9	4.4	0.9	5.9	5.0	3.4	5.4	8.4	9.0	5.9	2.4	8.4	3.9	6.9	1.9	5.0	11.4	13.4	12.4	12.9	9.0	0.9	13.4	6.4	24
30	7.5	7.5	4.9	2.4	0.0	1.4	X	0.0	0.0	0.0	0.0	1.4	2.9	0.0	2.9	4.4	0.0	3.9	1.4	3.4	5.9	4.0	5.0	4.9	0.0	7.5	2.8	23
31	5.0	2.4	1.9	2.4	3.4	11.9	12.9	9.9	16.4	11.9	12.4	7.5	11.9	5.4	9.0	7.9	5.9	5.9	8.4	12.9	10.4	5.9	3.9	12.9	1.9	16.4	8.3	24
HOURLY MAX	19.5	10.9	15.3	11.9	13.1	13.1	12.9	14.9	16.4	13.6	14.9	16.6	17.9	15.9	15.9	15.5	15.4	15.5	13.4	12.9	13.4	15.1	12.9	12.9				
HOURLY AVG	4.2	3.8	4.2	4.1	3.9	4.0	4.4	4.5	4.8	4.4	5.0	5.0	5.2	5.2	4.7	4.7	4.4	4.8	4.6	4.5	4.7	4.8	4.5	5.0				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 80 µg/m³ 24-HR 30 µg/m³


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0		
NUMBER OF 24-HR EXCEEDANCES:	0		
NUMBER OF NON-ZERO READINGS:	639		
MINIMUM 1-HR AVERAGE:	0.0 µg/m ³	@ HOUR(S)	14 ON DAY(S)
MAXIMUM 1-HR AVERAGE:	19.5 µg/m ³	@ HOUR(S)	0 ON DAY(S)
MAXIMUM 24-HR AVERAGE:	13.2 µg/m ³		26 ON DAY(S)
			26 ON DAY(S)
			VAR-VARIOUS
MONTHLY CALIBRATION TIME:	5 hrs	OPERATIONAL TIME:	732 hrs
		AMD OPERATION UPTIME:	98.4 %
STANDARD DEVIATION:	3.8	MONTHLY AVERAGE:	4.5 µg/m ³

24 HR AVERAGES March 2017



PARTICULATE MATTER < 2.5 MICRONS Hourly Averages (PM_{2.5} µg/m³)



Wind: LICA ST. LINA
 Poll.: LICA ST. LINA-PM25[ug/m3(L)]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 0.14%

Calm Avg: 5.21 [ug/m3]

Direction	0.0-3.9	3.9-7.8	7.8-11.8	11.8-15.7	15.7-19.6	>19.6	Total
N	5.0	7.6	4.1	2.8	0.7	0.0	20.1
NE	9.2	4.1	1.8	1.0	0.0	0.0	16.1
E	4.4	2.2	0.7	0.4	0.1	0.0	7.9
SE	3.9	2.6	1.1	0.4	0.0	0.0	8.0
S	3.4	1.8	0.8	0.1	0.0	0.0	6.2
SW	8.5	6.5	2.3	0.4	0.0	0.0	17.8
W	13.2	5.9	1.2	0.1	0.0	0.0	20.5
NW	1.8	1.1	0.4	0.0	0.0	0.0	3.3
Summary	49.5	31.8	12.5	5.2	0.8	0.0	100.0

% Icon Classes (ug/m3(L)) 49 0.0-3.9 32 3.9-7.8 13 7.8-11.8 5 11.8-15.7 1 15.7-19.6 0 >19.6

LICA ST. LINA Poll.: LICA ST. LINA-PM25[ug/m3(L)] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 0.14% Calm Poll Avg: 5.21[ug/m3(L)]



WIND SPEED




WIND SPEED Hourly Averages (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	9.8	10.3	10.3	10.4	11.6	11.2	8.9	6.5	8.9	7.3	8.4	8.5	14.3	15.4	18.5	18.2	20.5	21.4	17.1	18.9	21.3	14.5	18.8	19.5	6.5	21.4	10.7	24
2	10.5	24.4	17.9	13.2	12.9	17.4	19.2	20.0	19.3	19.2	20.2	19.4	20.0	18.5	18.9	17.7	20.2	Y	22.1	22.7	21.9	18.4	12.1	10.8	10.5	24.4	17.4	23
3	9.0	7.9	8.0	6.3	8.4	6.4	6.5	7.7	8.6	11.7	11.4	15.2	15.2	12.8	14.8	18.1	20.9	9.8	12.7	12.7	11.2	11.4	11.8	13.6	6.3	20.9	4.5	24
4	13.8	15.3	12.5	7.9	8.3	10.6	12.0	11.8	14.6	13.0	10.1	9.1	8.1	8.7	10.2	11.9	12.7	14.4	14.6	14.3	14.1	17.2	19.3	20.2	7.9	20.2	12.2	24
5	21.5	20.5	21.3	19.2	21.8	21.1	23.6	22.3	21.0	20.5	10.5	11.1	17.8	19.1	17.0	20.1	6.0	20.2	20.2	22.0	9.2	11.8	15.1	11.7	6.0	23.6	14.8	24
6	18.5	19.1	18.9	19.1	19.4	15.7	20.8	20.4	17.7	16.2	16.5	13.8	18.3	19.2	18.6	17.7	17.4	20.2	20.2	20.6	21.3	20.7	19.7	19.9	13.8	21.3	11.9	24
7	19.4	20.0	20.7	19.1	19.5	20.9	20.0	18.8	20.1	21.9	19.2	18.4	18.4	18.8	18.7	18.2	18.2	12.1	9.3	10.0	15.6	14.1	15.6	20.2	9.3	21.9	14.9	24
8	18.1	17.5	20.3	18.3	19.5	17.5	17.5	16.9	17.7	16.2	16.3	18.5	17.5	19.7	8.1	6.9	9.7	14.5	20.0	19.6	19.3	11.2	19.8	17.0	6.9	20.3	10.8	24
9	16.4	15.7	18.1	17.6	19.1	18.1	17.9	17.4	16.8	21.1	23.5	4.9	4.8	4.9	15.6	20.7	21.5	20.3	20.2	20.5	18.8	17.4	18.7	18.3	4.8	23.5	15.8	24
10	17.8	16.3	15.6	15.2	14.7	14.2	13.5	13.6	14.5	14.6	14.7	14.4	14.4	14.2	11.9	12.8	14.5	16.4	17.2	16.9	18.3	20.6	20.8	19.7	11.9	20.8	15.4	24
11	19.1	18.8	20.1	22.6	21.8	19.8	18.9	X	X	18.7	18.4	17.9	15.4	14.6	15.5	15.6	17.1	18.1	18.1	19.2	22.8	22.9	19.0	17.4	14.6	22.9	18.6	22
12	15.9	15.5	15.4	14.3	13.0	14.8	17.0	15.0	14.1	12.4	15.5	15.8	14.7	17.6	15.1	17.8	17.9	17.5	19.3	17.7	18.8	19.7	20.4	19.8	12.4	20.4	15.5	24
13	19.9	20.1	19.6	20.0	18.9	19.0	17.4	18.3	18.6	18.7	19.1	19.3	18.2	18.9	20.2	19.8	21.3	19.8	19.1	19.6	18.3	14.3	15.9	13.8	13.8	21.3	18.2	24
14	18.5	19.1	21.0	21.4	20.9	20.2	17.8	17.2	15.4	19.8	8.0	21.1	17.9	19.1	15.9	15.0	14.9	15.8	16.1	16.5	15.4	17.0	16.9	19.2	8.0	21.4	9.1	24
15	19.7	20.8	21.0	21.4	21.1	17.1	11.4	16.9	20.5	19.7	17.9	20.1	12.4	10.3	9.0	10.8	20.0	20.1	21.1	19.6	22.2	15.1	17.9	20.4	9.0	22.2	3.4	24
16	21.0	21.2	19.0	10.7	21.5	23.2	21.6	20.8	22.4	23.9	22.0	20.5	9.3	5.0	14.6	16.3	20.3	21.5	20.4	15.5	13.9	4.8	14.5	25.1	4.8	25.1	2.0	24
17	22.2	22.4	22.2	14.0	17.9	19.7	5.2	5.8	9.3	11.3	15.5	15.3	19.5	18.7	21.5	19.6	21.1	21.0	21.0	19.7	18.0	19.3	18.7	17.5	5.2	22.4	12.4	24
18	19.7	18.5	16.8	15.0	15.4	15.0	15.1	16.0	14.1	13.4	13.0	12.5	11.8	11.6	11.9	12.8	13.6	13.0	13.3	14.4	10.2	15.1	15.7	10.9	10.2	19.7	9.5	24
19	9.0	14.9	14.6	11.9	8.9	9.7	10.9	12.0	11.1	8.2	7.8	7.5	5.9	4.2	5.1	17.0	18.3	14.2	12.3	12.2	12.4	12.7	13.1	15.4	4.2	18.3	10.0	24
20	15.0	16.5	20.7	19.3	17.8	15.4	15.7	17.1	17.3	17.6	16.9	16.1	15.5	14.7	17.3	18.0	18.5	14.8	23.4	20.7	20.6	21.0	20.6	20.4	14.7	23.4	7.6	24
21	18.0	21.3	20.1	19.3	17.3	18.2	17.4	20.0	18.8	19.1	20.0	13.9	15.8	12.3	9.3	10.2	14.2	14.4	15.1	12.1	10.6	10.1	9.6	10.0	9.3	21.3	14.9	24
22	11.0	11.0	10.0	9.8	10.0	11.2	10.0	10.5	10.4	13.3	14.1	16.2	17.0	18.7	20.6	22.2	7.0	22.7	23.1	21.6	19.6	19.2	19.8	20.3	7.0	23.1	2.7	24
23	21.3	23.9	22.8	16.1	6.2	18.4	23.9	24.2	21.7	19.5	22.2	5.4	23.6	23.2	6.2	24.3	20.5	1.2	17.0	18.9	17.4	19.3	19.3	17.2	1.2	24.3	4.2	24
24	18.9	19.1	20.6	20.4	15.8	18.4	20.2	17.3	18.9	19.7	20.3	21.2	21.0	22.5	22.5	23.5	17.4	22.9	24.6	24.4	21.2	12.6	13.3	19.7	12.6	24.6	16.6	24
25	22.5	23.8	23.1	19.9	19.2	19.8	20.3	22.0	6.1	6.8	20.2	9.5	21.1	20.8	20.5	18.1	17.3	18.4	20.7	19.7	20.1	18.3	16.3	15.0	6.1	23.8	1.9	24
26	15.1	16.3	19.8	15.7	17.3	17.7	16.9	20.1	18.4	17.5	18.5	19.8	16.5	16.8	17.8	20.4	18.1	18.3	17.1	16.4	17.9	18.6	15.9	20.2	15.1	20.4	17.2	24
27	20.5	20.1	20.6	22.9	21.6	18.4	18.8	21.3	20.7	19.8	20.8	19.8	20.4	17.7	17.8	18.7	17.7	15.7	16.8	16.0	16.4	17.5	20.1	21.8	15.7	22.9	19.1	24
28	20.4	22.9	22.2	20.4	20.0	19.6	20.0	18.7	19.4	20.0	19.4	20.1	19.1	21.2	20.8	21.4	20.9	21.5	18.6	20.0	20.5	20.5	21.2	20.3	18.6	22.9	20.3	24
29	21.2	22.1	23.0	24.1	24.7	24.5	23.5	22.6	21.5	20.8	22.7	21.3	20.7	20.8	20.8	21.4	22.8	22.7	5.5	19.2	22.2	10.2	21.4	20.7	5.5	24.7	20.3	24
30	17.2	11.0	17.7	19.6	10.5	12.9	8.6	8.4	11.0	15.0	16.5	17.0	15.5	20.7	21.3	23.8	22.7	7.2	9.7	22.1	21.6	13.1	13.1	19.6	7.2	23.8	5.5	24
31	21.9	20.8	21.4	21.8	21.1	22.1	21.7	22.3	22.1	22.4	20.2	20.7	22.0	20.2	22.3	18.0	16.2	12.4	8.8	10.9	22.4	22.9	22.7	22.8	8.8	22.9	18.2	24
HOURLY MAX	22.5	24.4	23.1	24.1	24.7	24.5	23.9	24.2	22.4	23.9	23.5	21.3	23.6	23.2	22.5	24.3	22.8	22.9	24.6	24.4	22.8	22.9	22.7	25.1				
HOURLY AVG	2.6	3.5	3.4	1.7	2.0	2.4	2.4	2.3	3.0	2.6	3.1	3.9	3.3	3.7	4.0	3.5	3.5	2.1	2.2	2.5	3.9	2.5	3.2	2.3				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


LAST CALIBRATION:	September 12, 2016
DECLINATION :	MAGNETIC DECLINATION 14 DEGREE EAST



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	741
MINIMUM 1-HR AVERAGE:	1.2 kph @ HOUR(S) 17 ON DAY(S) 23
MAXIMUM 1-HR AVERAGE:	25.1 kph @ HOUR(S) 23 ON DAY(S) 16
MAXIMUM 24-HR AVERAGE:	20.3 kph ON DAY(S) 29
	VAR-VARIOUS
MONTHLY CALIBRATION TIME:	0 hrs
OPERATIONAL TIME:	741 hrs
AMD OPERATION UPTIME:	99.6 %
STANDARD DEVIATION:	4.4
MONTHLY AVERAGE:	2.8 kph

WIND SPEED Hourly Averages (WS kph)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - March 2017

WIND SPEED Instantaneous Maximum (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	23.5	25.5	25.0	25.7	27.0	25.9	23.2	35.0	26.3	36.0	27.0	26.8	33.4	30.7	33.1	31.2	34.4	35.1	42.6	31.5	32.1	48.2	44.7	33.1	23.2	23.2	48.2	31.5	24
2	33.8	31.2	23.9	24.8	19.6	24.3	27.8	30.7	27.8	32.3	31.9	33.0	35.1	33.5	31.5	29.7	29.6	Y	Y	33.8	36.8	30.2	25.8	25.0	19.6	19.6	36.8	29.6	22
3	24.3	20.6	22.0	18.8	19.9	21.4	19.0	21.8	23.2	28.5	27.2	29.0	27.0	27.0	29.0	29.2	29.9	30.3	25.7	25.4	24.8	25.7	24.4	22.2	18.8	18.8	30.3	24.8	24
4	23.3	24.0	24.6	25.6	26.3	37.5	37.2	35.5	39.8	40.0	30.8	32.0	25.1	25.8	29.4	28.4	28.5	28.9	27.3	29.5	29.1	26.9	27.6	29.1	23.3	23.3	40.0	29.7	24
5	31.8	33.1	32.1	32.8	32.1	32.2	32.1	53.3	30.8	32.0	32.4	33.1	33.3	32.0	32.5	35.7	58.6	72.2	31.1	31.7	44.3	41.3	42.7	33.8	30.8	30.8	72.2	37.4	24
6	29.8	28.2	27.8	27.1	28.0	31.3	30.6	33.2	40.7	38.7	41.8	40.3	37.0	41.6	40.1	37.9	40.7	34.9	36.3	40.1	40.2	40.6	33.9	34.3	27.1	27.1	41.8	35.6	24
7	34.6	38.9	43.1	40.0	36.6	41.1	38.1	35.0	36.7	40.0	36.3	34.0	35.9	37.1	37.7	34.9	34.2	34.8	31.1	29.6	30.8	31.8	30.8	29.6	29.6	29.6	43.1	35.5	24
8	29.3	28.0	33.8	31.9	33.6	24.4	27.3	29.2	31.8	31.9	28.2	31.9	30.8	36.1	29.8	29.6	30.0	31.5	33.7	32.4	27.1	28.1	24.0	25.3	24.0	24.0	36.1	30.0	24
9	27.8	23.6	27.1	24.7	23.1	23.1	23.1	22.0	30.4	35.8	43.8	34.5	34.8	34.8	32.2	37.5	33.5	33.7	25.8	24.7	22.2	24.7	24.0	26.7	22.0	22.0	43.8	28.9	24
10	29.3	29.7	29.7	27.9	28.4	29.7	24.6	29.0	28.0	31.9	28.0	31.3	28.0	30.0	24.5	29.0	27.8	30.0	30.8	30.6	31.5	32.4	32.3	31.9	24.5	24.5	32.4	29.4	24
11	30.6	31.7	33.4	37.8	35.4	35.4	29.1	X	X	28.8	35.5	33.7	30.9	29.8	30.2	31.3	30.9	30.2	31.9	35.2	35.6	32.8	30.4	28.8	28.8	37.8	32.3	22	
12	29.6	29.3	32.0	30.2	29.1	28.5	29.9	28.0	31.9	27.8	31.1	32.6	29.8	34.6	30.5	34.8	28.9	27.8	30.5	29.5	33.1	34.6	36.4	37.7	27.8	27.8	37.7	31.2	24
13	35.0	36.1	34.3	32.9	29.5	31.5	31.3	30.1	32.0	38.0	33.5	32.2	31.1	37.0	36.4	38.7	38.0	38.3	35.7	30.7	31.4	33.7	37.9	35.3	29.5	29.5	38.7	34.2	24
14	42.1	36.4	34.9	35.3	35.5	34.2	32.2	32.0	28.7	30.3	32.3	32.5	28.1	31.2	28.8	25.3	24.6	25.7	25.0	25.6	26.3	29.2	26.6	29.0	24.6	24.6	42.1	30.5	24
15	34.6	30.3	27.9	28.1	31.0	34.4	32.6	32.1	32.1	29.5	28.7	30.2	33.0	34.5	33.9	30.7	30.0	29.0	33.0	37.5	37.6	33.4	32.7	27.0	27.0	27.0	37.6	31.8	24
16	28.8	26.5	26.8	31.0	27.7	33.8	27.2	29.2	30.4	35.0	30.2	31.1	30.8	30.6	26.9	27.0	28.1	27.2	27.4	26.8	27.7	27.7	38.6	34.9	26.5	26.5	38.6	29.6	24
17	31.4	27.9	31.2	28.8	29.6	29.9	28.7	28.7	33.1	35.1	37.5	29.9	32.1	36.8	42.2	38.4	37.1	29.2	29.4	29.6	28.3	30.5	32.8	27.2	27.2	27.2	42.2	31.9	24
18	38.9	34.9	27.7	27.2	28.1	25.2	25.5	27.2	24.8	26.6	26.1	28.2	29.7	27.1	26.2	26.9	24.5	23.1	24.4	24.6	34.1	33.6	39.1	32.2	23.1	23.1	39.1	28.6	24
19	39.0	65.3	50.8	46.3	33.8	40.5	27.4	27.0	28.1	28.3	28.1	28.8	28.3	42.6	44.5	31.8	32.0	29.2	26.5	23.9	26.6	26.6	27.0	29.8	23.9	23.9	65.3	33.8	24
20	36.6	44.7	43.6	32.9	31.6	29.1	29.8	30.9	31.1	37.5	32.3	28.5	31.1	34.3	32.3	30.6	30.3	32.1	39.9	29.6	25.4	24.8	24.7	27.6	24.7	24.7	44.7	32.1	24
21	30.5	30.0	26.1	27.4	27.1	29.5	26.9	32.0	31.8	33.9	39.5	30.5	36.6	40.3	30.1	37.2	41.4	38.3	40.3	32.7	27.6	25.3	22.6	23.9	22.6	22.6	41.4	31.7	24
22	25.3	23.1	28.1	23.5	23.0	26.5	23.2	25.0	26.1	26.3	27.7	28.3	27.9	29.8	29.6	36.8	28.5	34.8	33.3	32.7	31.0	28.7	32.9	34.1	23.0	23.0	36.8	28.6	24
23	32.9	36.2	34.7	29.2	37.4	36.6	39.4	34.5	33.8	31.1	33.1	37.5	36.3	35.7	32.3	34.6	31.8	35.3	30.2	26.0	25.3	24.5	33.5	24.5	24.5	24.5	39.4	32.8	24
24	24.0	24.5	23.3	32.1	24.9	32.1	30.5	27.4	28.5	29.9	32.1	38.7	40.7	44.1	45.0	47.5	43.1	46.4	51.6	45.0	45.2	42.4	41.3	36.3	23.3	23.3	51.6	36.5	24
25	33.4	36.2	35.1	26.9	24.4	27.1	35.0	34.7	31.6	29.5	30.2	30.9	29.1	29.8	29.6	27.5	27.9	27.9	27.7	24.4	25.8	27.1	22.7	21.0	21.0	21.0	36.2	29.0	24
26	22.9	31.3	28.4	23.8	25.5	26.6	35.5	26.6	26.6	28.9	28.9	27.3	27.0	28.5	28.9	30.5	27.5	28.7	25.6	25.1	25.5	25.8	26.8	27.8	22.9	22.9	35.5	27.5	24
27	27.3	27.5	27.9	30.7	35.3	38.0	33.8	29.2	29.4	29.0	40.2	31.7	32.7	31.5	31.0	33.4	32.1	28.2	27.2	26.2	26.6	27.9	30.0	29.3	26.2	26.2	40.2	30.7	24
28	28.6	29.0	28.2	27.9	27.3	28.8	28.4	28.1	27.3	28.3	28.9	28.8	28.6	30.4	29.5	31.2	29.0	28.6	28.8	28.3	29.7	29.2	28.4	30.3	27.3	27.3	31.2	28.8	24
29	29.5	29.6	30.4	28.8	30.8	30.8	29.0	28.8	29.5	28.8	29.7	29.0	29.3	29.9	28.0	29.7	29.1	29.5	29.7	28.8	28.3	28.9	28.2	30.6	28.0	28.0	30.8	29.4	24
30	39.8	34.9	37.6	38.4	40.4	26.6	27.0	19.8	24.2	27.2	27.7	27.7	28.6	29.7	29.9	31.0	31.0	31.6	35.1	28.8	33.0	32.1	27.2	26.9	19.8	19.8	40.4	30.7	24
31	26.0	23.8	25.9	25.9	25.5	25.5	26.2	26.8	27.6	28.9	29.5	37.9	41.6	42.2	39.4	37.8	39.6	31.3	28.2	30.3	31.0	30.2	31.5	30.0	23.8	23.8	42.2	30.9	24
HOURLY MAX	42.1	65.3	50.8	46.3	40.4	41.1	39.4	53.3	40.7	40.0	43.8	40.3	41.6	44.1	45.0	47.5	58.6	72.2	51.6	45.0	45.2	48.2	44.7	37.7					
HOURLY AVG	30.8	31.4	30.9	29.8	29.3	30.4	29.4	30.1	30.1	31.8	31.9	31.7	31.7	33.5	32.4	32.7	32.7	32.8	31.5	29.9	30.8	31.0	31.1	29.6					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

MAXIMUM INSTANTANEOUS VALUE:	72.2	kph	@ HOUR(S)	17	ON DAY(S)	5
					VAR-VARIOUS	
OPERATIONAL TIME:					740	hrs

WIND SPEED Instantaneous Maximum (WS kph)




Wind: LICA ST. LINA
 Monitor: WSP [kph]
 Monthly: 17/03
 Type: WindRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 0.13%

Calm Avg: 0.00 [kph]

Direction	1.8-5.0	5.0-10.1	10.1-15.1	15.1-20.2	20.2-25.2	>25.2	Total
N	0.3	1.5	2.3	10.0	6.1	0.0	20.1
NE	0.1	1.4	2.0	8.5	4.2	0.0	16.2
E	0.0	0.4	2.0	1.4	4.1	0.0	7.8
SE	0.0	0.7	1.9	2.7	2.7	0.0	8.0
S	0.1	1.6	2.3	0.8	1.2	0.0	6.1
SW	0.0	1.4	3.8	9.7	2.6	0.0	17.4
W	0.0	0.8	5.0	9.9	5.1	0.0	20.8
NW	0.1	1.1	0.3	1.5	0.5	0.0	3.5
Summary	0.7	8.8	19.6	44.4	26.4	0.0	100.0

LICA ST. LINA 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 0.13% Calm Wind Avg Speed: 1.20(kph)



WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - March 2017

WIND DIRECTION Hourly Averages (WD)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	24-HOUR AVG	24-HR	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	QUADRANT	RDGS.	
DAY 1	W	W	W	W	W	WNW	WNW	WNW	W	WNW	WNW	WNW	W	W	W	W	WSW	NNE	N	NNW	NNW	N	NNE	WNW	24		
2	ESE	NNE	N	N	N	N	N	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	Y	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	23	
3	NNE	NNE	N	N	N	N	NNE	N	NNW	N	N	N	N	N	N	N	S	S	S	S	SSE	SSE	SSE	N	24		
4	SSE	SSE	S	SSE	SSE	ESE	ESE	ESE	SE	ESE	SE	SSE	SSE	SSE	SSE	SSE	SE	SSE	SSE	SSE	SSE	SSE	SSE	SE	24		
5	SE	SSE	SSE	SE	SSE	SSE	SSE	SSE	SSE	SSE	SE	ESE	SSE	SSE	SSE	SSE	S	SSE	SSE	W	WSW	W	NE	SSE	24		
6	WNW	WNW	WNW	WNW	WNW	W	WNW	WNW	NW	NNW	NNE	NNE	NNE	NNE	NNE	NNE	NE	NE	NE	NE	NE	NE	NE	N	24		
7	NE	NNE	NNE	NE	NE	NNE	NNE	NE	NE	NNE	NE	NNE	NE	NNE	NE	NE	NE	NNW	NNW	NNW	NW	NW	WNW	WNW	NNE	24	
8	WNW	WNW	W	W	W	WNW	WNW	WNW	W	W	WNW	W	W	W	NW	NNW	NNE	NE	ENE	ENE	ENE	NW	W	WNW	WNW	24	
9	WNW	WNW	W	WNW	W	WNW	WNW	W	W	W	NNW	N	NW	W	WSW	WSW	WSW	WSW	WSW	WSW	WSW	SW	WSW	W	24		
10	WSW	W	W	W	W	W	W	W	W	W	WNW	WNW	WNW	WNW	W	W	W	W	WSW	WSW	WSW	WSW	WSW	W	24		
11	WSW	WSW	WSW	WSW	W	WSW	WSW	X	X	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	W	WSW	WSW	WSW	W	W	W	WSW	22	
12	W	W	W	W	W	W	W	W	W	W	WSW	WSW	WSW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	WSW	24	
13	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SSW	SW	SW	SW	SSW	SSW	SSE	SW	24		
14	SE	ESE	SE	SE	SE	ESE	ESE	ESE	ESE	ESE	SW	WSW	WSW	WSW	WSW	WSW	SW	SW	SW	SW	SW	SW	SW	S	24		
15	SW	SW	SW	SW	SW	SW	S	ESE	E	E	ENE	ENE	NE	N	NNE	NE	ENE	NE	NE	NE	NE	NE	WNW	W	24		
16	W	W	W	ENE	ENE	E	E	ENE	ENE	E	E	E	ESE	SSW	WSW	WSW	WSW	WSW	WSW	ESE	SW	S	WSW	W	S	24	
17	WSW	WSW	WSW	ESE	E	E	SSE	SSE	S	SSE	SE	SW	SW	SW	SSW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	24	
18	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	WSW	W	W	WNW	NNE	NE	NE	WSW	24		
19	NNE	N	NNE	NNE	NNE	NE	ENE	ENE	ENE	ENE	ENE	ENE	E	NE	ENE	NE	NE	ENE	E	E	E	ENE	NE	NE	NE	24	
20	NE	NNE	NE	NE	NE	NE	NE	NE	NE	NE	NE	ENE	ENE	ENE	NE	ENE	ENE	ENE	W	WSW	WSW	WSW	WSW	SW	NE	24	
21	SW	SW	SW	SW	SW	SW	SW	SW	SW	SW	SSW	SSW	SSW	SSW	SSW	S	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SW	24	
22	SW	WSW	SW	SW	SW	SW	SW	SW	WSW	WSW	WSW	W	W	W	W	W	WNW	ENE	ENE	ENE	ENE	ENE	ENE	ENE	WSW	24	
23	E	E	ENE	ENE	NE	W	W	W	W	W	WSW	WSW	E	E	ESE	E	E	N	W	NNW	N	NNW	NNW	N	N	24	
24	NNW	NNW	NNW	NNW	N	N	NNE	N	NNE	NNE	NNE	NNE	NE	NE	NE	NE	NE	NE	NE	NE	NE	NE	ENE	E	SE	NNE	24
25	SE	SE	SE	SSE	SSE	SSE	S	S	SSW	SW	S	WNW	NNW	NNW	NNW	NNW	NNW	N	NNW	N	N	NNE	N	N	NNE	24	
26	N	NNW	NNW	N	N	N	NNW	NNW	N	N	NNW	NNW	NNW	NNW	NNW	NNW	N	NNE	N	N	NNE	NNE	NNE	NNE	N	NNE	24
27	NNE	NNE	NNE	NNE	NNE	NNE	NE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	N	N	NNE	24
28	N	N	NNE	NNE	N	N	NNE	NNE	N	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NNE	24
29	N	N	N	N	N	N	N	N	N	N	NNW	NNW	NNW	NNW	NNW	NNW	N	NE	NNE	NNE	NE	NNE	NNE	N	NNE	24	
30	NNE	E	SE	SE	E	W	WNW	ENE	WNW	W	W	W	W	W	W	W	WSW	WSW	SW	W	W	E	ESE	ESE	WSW	24	
31	E	E	ESE	ESE	ESE	E	ESE	ESE	ESE	ESE	ESE	ESE	SE	SE	SE	SE	SSE	SSW	SSW	SSW	ESE	ESE	ESE	ESE	ESE	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION:	September 12, 2016
DECLINATION :	MAGNETIC DECLINATION 14 DEGREE EAST

MONTHLY CALIBRATION TIME:	0	hrs	OPERATIONAL TIME:	741	hrs
STANDARD DEVIATION:	111		AMD OPERATION UPTIME:	99.6	%
			MONTHLY AVERAGE:	322	(NW)

WIND DIRECTION Hourly Averages (WD)



STANDARD DEVIATION WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - March 2017

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00		
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	RDGS.	
DAY																										
1	28	26	26	25	20	21	34	49	42	49	41	40	20	14	10	11	7	22	22	2	2	50	24	37	24	
2	64	7	3	4	4	3	4	7	9	32	25	27	35	25	13	15	11	Y	29	22	16	11	17	19	23	
3	24	25	21	36	23	34	30	28	22	17	17	13	16	21	16	12	10	46	16	15	20	19	17	12	24	
4	11	10	24	43	38	33	28	30	27	26	28	34	36	33	28	18	18	12	9	10	13	11	10	12	24	
5	22	29	17	43	23	28	16	7	8	29	71	64	35	21	42	31	73	12	5	8	41	41	47	62	24	
6	31	10	9	12	22	47	13	16	38	51	51	55	39	36	37	41	43	20	23	23	27	21	14	14	24	
7	15	16	17	18	19	18	18	17	16	16	18	17	19	19	22	21	33	58	69	54	48	53	42	9	24	
8	11	12	7	5	4	9	11	11	10	12	12	10	12	20	70	74	67	47	5	5	5	41	4	9	24	
9	8	9	6	5	4	5	5	4	6	9	13	62	80	77	51	8	6	9	4	5	3	5	4	5	24	
10	5	4	9	9	10	9	7	9	10	14	12	14	14	17	18	18	12	9	8	8	6	4	3	4	24	
11	4	6	5	12	15	8	13	X	X	15	23	11	15	16	13	13	10	8	8	6	6	5	7	8	22	
12	11	12	13	15	19	13	8	11	19	26	15	14	18	20	17	20	14	13	20	15	20	22	26	20	24	
13	22	28	28	22	15	16	16	16	26	32	21	20	16	21	24	22	20	30	25	21	33	52	51	53	24	
14	37	34	28	24	26	17	37	43	46	25	60	7	10	7	11	12	12	11	10	11	13	11	11	13	24	
15	12	16	18	18	23	45	51	21	9	8	12	21	60	71	75	66	28	11	18	16	19	52	41	8	24	
16	3	5	18	28	4	3	2	3	6	15	4	13	54	44	13	10	17	15	30	48	57	69	59	20	24	
17	4	13	20	42	44	35	66	72	64	68	49	41	33	35	32	34	22	15	16	14	13	14	14	13	24	
18	19	15	13	15	15	13	13	13	17	18	19	23	26	27	25	20	16	14	12	11	48	25	31	34	24	
19	44	41	34	36	46	38	29	24	30	46	50	52	51	72	43	20	16	18	13	13	16	23	25	21	24	
20	22	27	22	16	17	19	17	15	17	19	28	21	22	28	30	21	20	51	9	4	6	3	4	23	24	
21	39	21	9	9	10	11	11	16	15	20	21	25	23	31	38	38	28	27	25	27	28	28	30	26	24	
22	24	23	29	30	27	23	30	28	30	20	19	13	10	9	6	27	48	18	4	6	9	29	22	25	24	
23	6	6	14	47	63	45	19	7	7	9	12	47	18	8	54	13	6	22	22	5	6	5	5	7	24	
24	4	10	3	4	7	13	29	9	14	26	18	21	30	25	29	31	52	32	21	25	40	55	65	42	24	
25	26	19	17	6	5	5	3	5	53	43	12	48	22	13	7	10	14	8	4	3	3	5	7	4	24	
26	7	39	6	12	8	9	29	4	13	10	9	6	10	10	13	10	11	10	8	6	6	7	11	8	24	
27	7	7	8	17	29	45	45	28	24	19	22	30	15	16	17	14	14	15	10	11	13	10	11	13	24	
28	7	9	7	8	7	8	6	8	9	11	16	14	15	21	25	20	15	22	43	36	11	10	19	35	24	
29	9	7	8	8	8	8	8	10	9	7	10	9	8	13	6	7	7	11	36	38	12	64	5	30	24	
30	50	71	50	44	34	23	41	40	29	13	10	11	14	7	8	7	13	45	57	2	34	63	50	36	24	
31	3	2	13	4	3	2	21	15	14	13	30	32	27	40	23	45	56	62	63	52	24	13	26	20	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION: September 12, 2016

CALIBRATION TIME: 0 hrs OPERATIONAL TIME: 741 hrs

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)



RELATIVE HUMIDITY




RELATIVE HUMIDITY Hourly Averages (RH %)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	74	76	75	73	71	69	67	65	64	63	62	59	59	60	59	58	57	64	69	69	69	70	66	69	57	76	66	24
2	75	75	74	74	74	75	74	73	69	63	59	55	54	54	56	57	63	68	70	73	74	75	73	73	54	75	68	24
3	73	75	77	78	80	81	82	82	75	64	56	50	45	45	45	44	45	52	59	60	64	66	66	69	44	82	64	24
4	71	73	74	73	74	72	73	74	77	75	74	73	72	71	70	68	67	68	71	72	71	70	71	72	67	77	72	24
5	71	71	72	71	72	72	72	72	70	65	60	57	56	54	55	56	58	61	66	69	71	71	70	70	54	72	66	24
6	71	72	73	73	73	74	76	76	75	73	74	73	71	71	72	73	75	76	76	76	75	72	73	72	71	76	74	24
7	72	72	72	71	70	69	68	66	65	62	56	53	48	45	41	41	39	42	47	50	51	54	55	56	39	72	57	24
8	53	54	58	61	61	60	59	57	52	46	42	39	37	35	35	35	39	52	58	61	60	55	54	52	35	61	51	24
9	54	55	56	58	60	62	64	60	50	43	37	33	31	31	31	36	36	40	46	50	49	47	46	52	31	64	47	24
10	54	55	51	56	58	60	62	60	52	45	42	37	38	37	54	62	66	68	69	70	71	71	72	72	37	72	58	24
11	72	71	71	71	71	70	69	X	X	62	58	54	53	56	59	63	67	70	73	73	74	74	74	74	53	74	67	22
12	74	74	74	74	74	74	74	73	71	68	66	63	60	53	51	52	56	60	64	66	67	68	67	69	51	74	66	24
13	70	71	71	72	73	74	74	73	68	61	56	57	52	53	55	58	62	64	69	72	74	76	76	76	52	76	67	24
14	75	75	75	74	76	76	77	72	68	54	48	46	50	50	55	59	63	68	72	74	76	77	79	80	46	80	67	24
15	81	83	84	85	86	86	86	84	77	64	58	48	40	36	31	32	31	36	39	38	39	47	64	72	31	86	59	24
16	74	73	70	68	69	72	74	49	45	29	29	26	34	43	63	64	66	72	78	82	86	88	88	89	26	89	64	24
17	89	89	89	89	89	89	89	89	88	83	80	71	66	65	63	65	67	73	77	78	81	83	83	81	63	89	80	24
18	80	81	82	82	84	85	85	83	80	76	74	69	66	64	63	65	68	71	73	75	76	78	85	87	63	87	76	24
19	87	88	88	84	79	75	76	67	59	54	51	44	39	35	42	60	80	83	81	78	80	72	66	67	35	88	68	24
20	63	57	55	55	53	52	54	50	45	41	38	37	36	34	32	31	30	31	42	49	50	50	51	56	30	63	46	24
21	59	55	54	57	61	65	67	64	56	50	49	48	51	53	54	57	57	59	62	63	65	66	68	69	48	69	59	24
22	70	70	71	72	73	73	75	75	75	75	78	81	82	83	83	83	83	84	85	86	86	86	86	86	70	86	79	24
23	86	86	86	86	86	86	86	85	79	74	73	69	65	63	61	53	66	68	77	78	81	83	85	86	53	86	77	24
24	85	83	83	83	81	81	82	80	79	79	80	80	80	79	78	75	73	74	75	76	76	77	80	80	73	85	79	24
25	81	80	81	82	84	84	83	81	74	68	64	58	52	52	53	55	59	66	73	78	78	80	81	83	52	84	72	24
26	84	85	87	86	87	87	87	87	78	70	65	61	62	59	56	54	56	62	70	77	81	82	80	80	54	87	74	24
27	83	85	87	86	85	85	84	81	79	78	78	76	76	75	72	72	76	78	80	82	83	85	85	85	72	87	81	24
28	87	87	86	86	88	89	89	89	89	88	88	86	84	83	83	83	83	84	85	88	89	90	90	83	90	87	87	24
29	90	90	90	90	90	90	90	90	90	90	90	89	87	79	72	69	71	71	78	82	84	87	88	89	69	90	85	24
30	89	89	89	89	89	89	89	89	89	89	88	88	87	84	80	74	73	75	78	84	85	82	86	88	73	89	85	24
31	88	89	89	89	89	89	89	86	77	70	63	58	57	58	66	70	68	68	71	72	70	72	73	74	57	89	75	24
HOURLY MAX	90	90	90	90	90	90	90	90	90	90	90	89	87	84	83	83	83	84	85	88	89	90	90	90	57	90	85	24
HOURLY AVG	75	75	76	76	76	76	77	74	71	65	62	59	58	57	58	59	61	65	69	71	72	73	74	75	57	89	75	24

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	26	%	@ HOUR(S)	11	ON DAY(S)	16
MAXIMUM 1-HR AVERAGE:	90	%	@ HOUR(S)	21	ON DAY(S)	28
MAXIMUM 24-HR AVERAGE:	87	%			ON DAY(S)	28
					VAR-VARIOUS	
				OPERATIONAL TIME:		742 hrs
				AMD OPERATION UPTIME:		99.7 %
STANDARD DEVIATION:	14			MONTHLY AVERAGE:		69 %

RELATIVE HUMIDITY Hourly Averages (RH %)



BAROMETRIC PRESSURE




BAROMETRIC PRESSURE Hourly Averages (BP mbar)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.					
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.						
DAY																																	
1	922	922	922	922	922	923	923	924	925	926	927	928	929	931	931	932	933	934	934	935	935	935	936	936	922	936	929	24					
2	936	935	935	935	935	934	934	934	933	933	933	932	931	931	930	929	928	927	925	924	923	922	921	921	921	921	936	930	24				
3	920	920	919	918	917	917	916	915	915	915	915	915	915	915	915	915	916	915	914	914	915	914	914	914	914	914	920	916	24				
4	914	914	915	914	914	914	914	916	916	916	916	917	918	918	918	918	919	919	919	920	920	920	920	919	914	920	917	24					
5	919	919	918	919	918	918	918	918	918	918	918	918	917	917	917	917	917	916	916	916	916	916	916	916	916	916	919	917	24				
6	916	916	916	916	916	916	917	917	917	918	918	918	918	918	918	919	919	919	919	920	920	920	920	920	916	920	918	24					
7	920	921	921	922	922	923	924	924	925	926	927	927	928	929	930	931	931	931	931	931	931	931	931	931	920	931	927	24					
8	931	931	931	931	932	932	932	933	933	934	935	935	935	936	936	936	936	936	937	937	938	938	939	939	931	939	935	24					
9	939	940	941	941	942	943	944	944	945	946	946	947	947	948	947	947	947	946	946	946	945	945	945	945	939	948	945	24					
10	944	944	943	943	943	942	942	942	942	942	942	943	943	943	943	943	943	943	943	943	943	943	943	943	942	944	943	24					
11	943	943	943	943	942	942	942	X	X	942	942	942	942	942	941	940	940	939	938	937	937	936	936	936	936	943	940	22					
12	936	936	936	936	937	937	937	938	938	939	940	940	941	941	940	941	940	940	939	938	937	937	936	935	935	941	938	24					
13	934	933	932	931	931	930	929	928	927	928	927	927	926	926	926	925	924	924	924	924	924	924	923	923	923	934	927	24					
14	924	924	924	923	923	924	924	924	924	925	925	926	926	926	925	924	923	923	922	921	921	920	920	918	918	918	926	923	24				
15	918	917	917	917	917	916	916	917	918	920	921	922	922	923	923	923	923	922	922	922	922	921	922	922	922	916	923	920	24				
16	922	921	921	922	921	922	922	922	923	924	925	926	927	928	928	928	929	929	930	930	931	931	931	932	921	932	926	24					
17	932	932	933	933	933	933	933	933	933	933	933	932	932	932	932	931	930	930	929	928	927	927	926	925	925	933	931	24					
18	924	922	921	920	919	918	917	917	916	915	914	914	913	912	911	910	909	907	906	905	904	905	906	908	904	924	913	24					
19	910	912	915	916	918	919	920	921	923	924	924	925	926	926	926	926	927	927	927	927	928	928	929	930	910	930	923	24					
20	931	932	933	934	935	935	936	937	938	939	940	940	941	941	941	941	941	941	940	940	939	939	939	939	931	941	938	24					
21	939	939	938	938	937	937	936	935	935	936	935	934	934	933	931	930	929	928	928	927	926	926	925	924	924	939	933	24					
22	923	922	921	920	920	920	919	918	918	918	917	917	917	916	916	917	917	917	917	917	917	918	918	918	919	916	923	918	24				
23	919	920	920	920	920	921	921	921	922	922	923	924	925	925	926	926	927	927	927	927	928	928	928	928	919	928	924	24					
24	928	928	928	927	927	927	927	926	926	926	926	926	926	926	925	924	923	923	922	921	921	921	920	919	919	919	928	924	24				
25	918	918	918	918	918	919	919	920	920	921	922	922	923	923	923	924	924	924	924	923	922	922	922	923	918	924	921	24					
26	923	923	923	923	923	923	923	924	924	924	925	925	926	926	926	927	927	927	926	926	925	924	924	924	923	927	925	24					
27	924	924	924	924	924	924	924	924	925	926	926	926	927	927	927	928	928	928	928	928	928	928	928	928	924	928	926	24					
28	928	928	928	928	928	927	927	927	927	928	928	928	928	928	927	927	926	926	926	926	926	926	926	926	926	926	928	927	24				
29	926	925	925	925	925	925	925	925	925	925	925	925	925	925	925	924	924	924	923	922	922	921	921	921	921	921	926	924	24				
30	921	921	921	921	921	921	922	923	923	924	925	926	927	928	929	930	931	931	932	931	931	932	932	933	921	933	927	24					
31	933	933	933	933	933	933	933	933	934	934	935	935	934	934	933	932	930	929	929	927	927	926	925	925	925	935	931	24					
HOURLY MAX	944	944	943	943	943	943	944	944	945	946	946	947	947	948	947	947	947	946	946	946	945	945	945	945									
HOURLY AVG	926	926	926	926	926	926	926	926	926	927	928	928	928	928	928	928	928	927	927	927	927	927	926	927									

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	904 mbar	@ HOUR(S)	20	ON DAY(S)	18
MAXIMUM 1-HR AVERAGE:	948 mbar	@ HOUR(S)	13	ON DAY(S)	9
MAXIMUM 24-HR AVERAGE:	945 mbar			ON DAY(S)	9
				VAR-VARIOUS	
		OPERATIONAL TIME:		742	hrs
		AMD OPERATION UPTIME:		99.7	%
STANDARD DEVIATION:	8	MONTHLY AVERAGE:		927	mbar

BAROMETRIC PRESSURE Hourly Averages (BP mbar)



AMBIENT TEMPERATURE




AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	-16.4	-17.4	-18.0	-19.1	-19.8	-20.5	-20.7	-20.4	-19.5	-18.5	-16.7	-15.3	-13.9	-13.9	-13.7	-13.2	-13.4	-16.3	-17.4	-17.5	-17.2	-17.5	-16.9	-17.7	-20.7	-13.2	-17.1	24	
2	-19.6	-20.1	-19.8	-19.7	-20.1	-19.6	-20.2	-20.3	-15.9	-12.8	-11.7	-10.0	-8.9	-7.5	-6.9	-6.7	-7.0	-7.9	-8.3	-8.4	-8.0	-8.3	-7.7	-7.6	-20.3	-6.7	-12.6	24	
3	-7.6	-7.9	-8.5	-8.6	-9.1	-9.7	-10.2	-10.4	-8.6	-5.3	-2.6	0.2	2.2	2.9	3.9	5.0	5.1	1.9	-1.7	-2.5	-2.9	-3.0	-3.5	-4.1	-10.4	5.1	-3.5	24	
4	-4.4	-4.3	-4.4	-5.9	-7.4	-8.5	-10.4	-12.3	-13.9	-14.5	-14.5	-14.2	-14.4	-13.6	-13.6	-14.1	-15.3	-17.0	-17.5	-17.4	-17.5	-17.6	-17.6	-17.6	-17.6	-4.3	-12.7	24	
5	-17.9	-18.2	-18.4	-18.7	-18.9	-19.1	-19.2	-19.1	-18.5	-16.9	-14.8	-13.8	-13.2	-12.6	-12.5	-12.7	-13.2	-14.0	-15.2	-16.1	-16.4	-16.3	-15.6	-15.1	-19.2	-12.5	-16.1	24	
6	-15.0	-15.1	-15.2	-15.3	-15.3	-15.1	-15.2	-15.1	-14.6	-13.9	-13.5	-12.6	-11.2	-11.3	-11.6	-12.0	-12.5	-13.2	-13.9	-14.5	-14.9	-15.7	-16.9	-16.9	-16.9	-11.2	-14.2	24	
7	-17.1	-17.7	-18.3	-19.0	-19.7	-20.3	-20.3	-20.2	-19.8	-18.8	-16.8	-15.7	-14.3	-13.1	-11.2	-11.4	-11.7	-13.4	-15.7	-17.1	-18.2	-19.3	-20.3	-21.4	-21.4	-11.2	-17.1	24	
8	-22.1	-23.1	-24.5	-25.6	-26.2	-25.8	-25.5	-25.2	-24.2	-22.8	-21.5	-19.6	-18.3	-16.3	-15.2	-14.9	-15.2	-17.4	-19.3	-20.3	-20.9	-21.3	-21.6	-22.1	-26.2	-14.9	-21.2	24	
9	-23.6	-24.4	-24.8	-25.4	-25.9	-26.7	-27.4	-26.1	-22.7	-19.8	-17.5	-15.6	-14.6	-14.0	-14.1	-16.0	-16.6	-18.9	-21.5	-22.1	-21.9	-21.9	-22.1	-22.7	-27.4	-14.0	-21.1	24	
10	-23.3	-25.0	-25.5	-26.4	-26.9	-27.0	-28.3	-28.1	-25.3	-21.6	-19.9	-17.1	-16.5	-15.8	-18.2	-19.0	-19.8	-20.1	-20.3	-20.4	-20.6	-20.9	-21.6	-21.9	-28.3	-15.8	-22.1	24	
11	-22.5	-22.0	-21.8	-21.7	-21.8	-21.3	-20.1	X	X	-18.0	-16.7	-15.5	-14.8	-15.5	-15.6	-16.1	-16.8	-17.3	-17.6	-17.9	-17.8	-17.7	-18.1	-22.5	-14.8	-18.3	22		
12	-18.1	-17.9	-17.9	-18.2	-18.4	-18.6	-19.1	-19.0	-17.4	-15.5	-13.6	-12.0	-11.3	-11.1	-11.9	-12.1	-13.3	-14.3	-15.4	-15.7	-15.8	-15.7	-15.7	-16.1	-19.1	-11.1	-15.6	24	
13	-16.6	-16.8	-17.3	-17.4	-17.0	-16.7	-16.7	-16.4	-14.2	-12.4	-11.0	-10.9	-9.0	-8.8	-9.0	-9.2	-9.8	-10.2	-11.6	-12.3	-12.6	-12.7	-12.5	-12.0	-17.4	-8.8	-13.0	24	
14	-11.2	-11.1	-11.1	-11.3	-12.1	-12.5	-13.0	-12.0	-10.6	-5.9	-2.7	-0.6	-2.2	-1.1	-1.9	-2.3	-2.4	-2.6	-2.5	-2.1	-1.9	-1.6	-2.3	-2.9	-13.0	-0.6	-5.8	24	
15	-3.3	-4.1	-4.7	-5.0	-5.4	-5.4	-5.7	-5.3	-2.7	1.1	3.4	7.0	8.9	9.2	9.9	8.7	7.6	5.8	3.9	3.3	3.6	3.6	2.7	1.8	-5.7	9.9	1.6	24	
16	1.4	1.5	1.7	1.6	0.7	-0.1	-0.8	0.2	1.2	7.2	8.1	8.9	7.7	6.3	4.2	3.7	3.3	2.5	1.3	0.6	-0.3	-0.8	-1.0	-1.1	-1.1	8.9	2.4	24	
17	-1.2	-1.3	-1.1	-1.2	-1.3	-1.4	-1.6	-1.1	-0.2	0.4	0.4	2.8	3.8	4.2	4.5	4.1	3.8	2.4	1.3	1.0	0.7	0.7	0.4	-1.6	4.5	0.9	24		
18	0.2	-0.4	-0.7	-0.6	-1.0	-1.1	-1.0	-0.6	0.0	1.3	1.8	3.4	4.6	5.2	5.6	5.2	4.6	3.9	3.3	2.7	2.2	1.9	1.0	0.6	-1.1	5.6	1.8	24	
19	0.5	-0.9	-1.8	-2.2	-2.5	-3.1	-3.6	-2.1	-0.5	0.7	2.0	3.2	3.4	3.3	2.6	0.1	-1.9	-2.9	-3.1	-3.4	-3.7	-4.0	-4.6	-5.7	-5.7	3.4	-1.3	24	
20	-7.1	-8.5	-10.6	-11.6	-12.3	-12.9	-13.3	-11.8	-10.1	-9.0	-7.3	-6.5	-5.5	-4.9	-4.3	-3.7	-3.9	-4.9	-7.9	-9.4	-10.1	-10.3	-10.5	-11.1	-13.3	-3.7	-8.6	24	
21	-11.3	-11.0	-11.0	-11.2	-11.6	-12.0	-12.1	-10.6	-7.7	-5.3	-3.6	-2.8	-2.8	-2.7	-3.0	-3.3	-3.2	-3.4	-3.5	-3.5	-3.7	-3.9	-4.0	-4.1	-12.1	-2.7	-6.3	24	
22	-3.9	-3.9	-3.9	-4.0	-4.2	-4.4	-4.5	-4.4	-4.2	-4.1	-3.9	-3.9	-4.1	-4.3	-4.1	-4.2	-4.6	-4.8	-5.0	-5.1	-5.5	-5.3	-5.5	-5.2	-5.5	-3.9	-4.5	24	
23	-5.2	-5.3	-5.4	-5.6	-5.6	-5.9	-6.0	-4.2	-3.5	-2.7	-2.5	-1.4	-0.6	0.2	0.9	3.6	-0.3	-0.9	-3.2	-3.6	-4.5	-5.2	-5.8	-6.5	-6.5	3.6	-3.3	24	
24	-7.3	-8.9	-10.0	-9.3	-11.2	-11.9	-11.2	-11.2	-10.2	-8.9	-7.2	-5.3	-4.0	-2.8	-1.8	-0.5	0.3	0.1	-0.4	-0.6	-0.6	-1.2	-2.1	-2.2	-11.9	0.3	-5.4	24	
25	-2.8	-2.9	-3.2	-3.7	-4.1	-3.9	-3.7	-2.9	-0.8	1.0	2.6	4.3	6.3	6.5	6.1	5.6	4.8	2.8	0.9	-0.5	-1.0	-1.2	-1.6	-2.1	-4.1	6.5	0.3	24	
26	-2.4	-3.4	-3.3	-2.8	-3.4	-4.1	-4.6	-3.8	-0.7	1.4	2.9	4.4	4.9	5.9	6.4	7.0	6.2	4.4	2.1	0.6	0.1	0.3	0.5	-0.2	-4.6	7.0	0.8	24	
27	-1.0	-1.5	-1.7	-2.7	-3.0	-3.5	-3.8	-3.2	-2.5	-1.7	-1.2	-0.4	-0.3	0.2	1.0	1.1	0.7	0.3	-0.1	-0.3	-0.3	-0.4	-0.4	-0.4	-3.8	1.1	-1.0	24	
28	-0.4	-0.3	-0.3	-0.5	-1.0	-1.4	-1.6	-1.5	-0.8	0.1	0.8	1.1	1.6	1.7	1.8	1.7	2.0	2.0	1.6	1.3	1.0	0.9	0.5	0.3	-1.6	2.0	0.4	24	
29	0.3	0.3	0.3	0.1	0.1	0.0	-0.1	0.0	0.1	0.4	1.0	1.5	1.9	2.9	3.9	4.5	3.9	3.9	2.2	1.2	0.7	-0.4	-0.4	-2.1	-2.1	4.5	1.1	24	
30	-2.2	-1.8	-1.8	-1.7	-1.8	-1.6	-1.6	-1.5	-1.3	-0.8	-0.2	0.4	1.1	1.9	2.1	3.1	3.4	2.6	1.7	-0.1	0.0	0.5	-0.6	-0.8	-2.2	3.4	0.0	24	
31	-1.3	-1.4	-1.7	-1.9	-2.3	-2.5	-2.4	-0.7	2.1	4.8	6.8	7.8	8.2	8.1	7.0	6.0	6.3	6.9	6.2	5.5	5.1	4.1	3.7	3.2	-2.5	8.2	3.2	24	
HOURLY MAX	1.4	1.5	1.7	1.6	0.7	0.0	-0.1	0.2	2.1	7.2	8.1	8.9	8.9	9.2	9.9	8.7	7.6	6.9	6.2	5.5	5.1	4.1	3.7	3.2					
HOURLY AVG	-9.1	-9.5	-9.8	-10.1	-10.6	-10.9	-11.1	-10.3	-8.9	-7.4	-6.1	-4.8	-4.0	-3.6	-3.5	-3.6	-4.1	-5.1	-6.3	-6.9	-7.2	-7.4	-7.7	-8.1					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	-28.3 °C	@ HOUR(S)	6	ON DAY(S)	10
MAXIMUM 1-HR AVERAGE:	9.9 °C	@ HOUR(S)	14	ON DAY(S)	15
MAXIMUM 24-HR AVERAGE:	3.2 °C			ON DAY(S)	31
				VAR-VARIOUS	
		OPERATIONAL TIME:		742	hrs
		AMD OPERATION UPTIME:		99.7	%
STANDARD DEVIATION:	8.8	MONTHLY AVERAGE:		-7.3	°C

AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)



PRECIPITATION




PRECIPITATION Hourly Averages (mm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.3	1.0	0.3	0.5	0.6	0.1	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.0	0.1	24
5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.8	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.8	0.1	24
6	0.0	0.0	0.0	0.0	0.0	0.1	0.6	0.6	0.6	0.4	0.5	0.4	0.0	0.3	0.1	0.1	0.2	0.1	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.6	0.2	24
7	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
8	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
9	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.8	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.8	0.0	24
10	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.1	0.0	24
11	0.0	0.0	0.0	0.0	0.0	0.0	0.0	X	X	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.1	0.0	0.1	0.0	0.1	0.1	0.1	0.1	0.0	0.2	0.0	22
12	0.0	0.1	0.2	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	24
13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
14	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
15	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.1	0.0	24
16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.5	0.8	2.4	0.2	0.0	2.4	0.2	24
17	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.8	0.0	0.0	1.8	0.1	24
19	0.6	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.4	0.0	0.0	0.1	0.0	0.0	0.1	0.0	0.0	0.0	0.6	0.1	24
20	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
22	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.9	1.4	2.6	2.5	3.3	2.1	1.7	0.5	0.1	0.0	0.0	0.1	0.1	0.0	3.3	0.6	24
23	0.1	0.1	0.3	0.2	2.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	2.5	0.1	24
24	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
25	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
26	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
27	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
28	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
29	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	24
30	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
31	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
HOURLY MAX	0.6	0.2	0.3	0.2	2.5	0.1	0.6	0.6	1.0	0.4	0.5	0.9	1.8	2.6	2.5	3.3	2.1	1.7	0.5	0.2	0.5	0.8	2.4	0.2	0.0	0.0	0.0	24
HOURLY AVG	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.1	0.0	0.0	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.2	0.0	0.0	0.0	0.0	24

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	0.0	mm	@ HOUR(S)	0	ON DAY(S)	1
MAXIMUM 1-HR AVERAGE:	3.3	mm	@ HOUR(S)	15	ON DAY(S)	22
MAXIMUM 24-HR AVERAGE:	0.6	mm			ON DAY(S)	22
MONTHLY TOTAL	37.5	mm			VAR-VARIOUS	
OPERATIONAL TIME:						742 hrs
AMD OPERATION UPTIME:						99.7 %
STANDARD DEVIATION:	0.3					MONTHLY AVERAGE: 0.1 mm

PRECIPITATION Hourly Averages (mm)



APPENDIX II
EQUIPMENT CALIBRATION RESULTS

SULPHUR DIOXIDE



API 100E Sulphur Dioxide Analyzer Calibration

Date: <u>March 1, 2017</u>	Barometric Pressure: <u>27.6 inHg</u>
Company/Airshed: <u>LICA</u>	Station Temperature °C: <u>19</u>
Location/Station Name: <u>St. Lina</u>	Weather Conditions: <u>Mainly sunny</u>
Parameter: <u>Sulphur Dioxide</u>	Calibration Purpose: <u>shut down</u>
Start Time 24 hr. (mst): <u>9:40</u>	Performed By/Reviewer: <u>Limin Li</u> <u>Trina Whitsitt</u>
End Time 24 hr. (mst): <u>12:15</u>	Cal Gas Expiry Date: <u>December 25, 2018</u>
Calibration Method: <u>Gas Dilution</u>	Converter Model & s/n (if applicable): <u>n/a</u>

Analyzer:		
ID# or Serial Number: <u>468</u>	Range ppb: <u>1000</u>	Station SO2 Analyzer Range? <u>1000</u> ppb
Last Calibration Date: <u>February 2, 2017</u>	As Found C.F.: <u>0.995</u>	
Previous C.F.: <u>1.000</u>	New C.F.: <u>n/a</u>	

Calibrator: Flow Meter ID's: <u>n/a</u> Make & Model: <u>Sabio 2010</u> Serial #: <u>17200415</u> Cal Gas Cylinder I.D. #: <u>BLM002756T</u> Cal Gas Conc. (ppm): <u>49.9</u>	Standard Calibration Points for Ranges <table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>780</td></tr> <tr><td>Mid</td><td>380</td></tr> <tr><td>Low</td><td>190</td></tr> </table>	Point	ppb	High	780	Mid	380	Low	190
Point	ppb								
High	780								
Mid	380								
Low	190								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated Concentration:		Indicated Concentration:		Correction Factors (C.F.):
	Diluent	Cal Gas	Total	(ppb)	(ppb)	(ppb)	(ppb)	
as found zero	5028	0.00	5028	0.0	1.9			n/a
as found high	4952	77.10	5029	765.0	771.0			0.995
mid	4991	37.70	5029	374.1	379.0			0.992
low	5008	19.10	5027	189.6	190.0			1.008
Average C.F.=								0.998

Linear Regression/Calibration Results:

Correlation Coefficient = <u>1.000</u>	LIMITS
Slope = <u>0.993</u>	> or = 0.995
b (Intercept as % of full scale) = <u>-0.11%</u>	0.90-1.10
% change in C.F. from last cal = <u>0.53%</u>	± 3% F.S.
	± 10%


API 100E Sulphur Dioxide Analyzer Calibration



As found: SLOPE: <u>0.987</u> OFFSET: <u>117.8</u> HVPS: <u>651</u> RCELL TEMP: <u>50</u> BOX TEMP: <u>29.3</u> PMT TEMP: <u>7.9</u> IZS TEMP: <u>50</u> PRES: <u>23.8</u> SAMP FL: <u>614</u> NORM PMT: <u>121</u> UV LAMP: <u>3154</u> LAMP RATIO: <u>96.7</u> STR. LGT: <u>58.1</u> DRK PMT: <u>5.1</u> DRK LMP: <u>6.7</u> Expected Value: <u>395.0</u>	As left: SLOPE: <u>n/a</u> OFFSET: <u>n/a</u> HVPS: <u>n/a</u> RCELL TEMP: <u>n/a</u> BOX TEMP: <u>n/a</u> PMT TEMP: <u>n/a</u> IZS TEMP: <u>n/a</u> PRES: <u>n/a</u> SAMP FL: <u>n/a</u> NORM PMT: <u>n/a</u> UV LAMP: <u>n/a</u> LAMP RATIO: <u>n/a</u> STR. LGT: <u>n/a</u> DRK PMT: <u>n/a</u> DRK LMP: <u>n/a</u> Expected Value: <u>n/a</u>
--	--

Comments:

As found high point was repeated due to operator error. After shut-down calibration, the reaction cell was cleaned. The sinter filter and o-rings were changed. The sample valve (flow rate lower than zero check. Dirty inside) was changed. The analogue output was calibrated. The sample pump was rebuilt. Valves were tested for leaks. ZERO Air scrubber was repositioned inside. Factory calibration completed.



— SO2[ppb]



API 100E Sulphur Dioxide Analyzer Calibration

Date: March 2, 2017	Barometric Pressure: 27.78 inHg
Company/Airshed: LICA	Station Temperature °C: 22
Location/Station Name: St. Lina	Weather Conditions: Mainly sunny
Parameter: Sulphur Dioxide	Calibration Purpose: post repair
Start Time 24 hr. (mst): 10:59	Performed By/Reviewer: Alex Yakupov Trina Whitsitt
End Time 24 hr. (mst): 15:35	Cal Gas Expiry Date: July 18, 2019
Calibration Method: Gas Dilution	Converter Model & s/n (if applicable): n/a

Analyzer:	Range ppb: 1000	Station SO2 Analyzer Range?
ID# or Serial Number: 468	As Found C.F.: n/a	1000 ppb
Last Calibration Date: n/a	New C.F.: 1.000	
Previous C.F.: n/a		

Calibrator:	Standard Calibration Points for Ranges								
Flow Meter ID's: n/a	<table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>780</td></tr> <tr><td>Mid</td><td>380</td></tr> <tr><td>Low</td><td>190</td></tr> </table>	Point	ppb	High	780	Mid	380	Low	190
Point	ppb								
High	780								
Mid	380								
Low	190								
Make & Model: API 700									
Serial #: 627									
Cal Gas Cylinder I.D. #: LL104222									
Cal Gas Conc. (ppm): 50.6									


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	(ppb)	(ppb)	
adjusted zero	5000	0.00	5000	0.0	0.0	n/a
adjusted high	4924	76.90	5001	778.1	778.0	1.000
mid	4965	37.50	5003	379.3	377.0	1.006
low	4980	18.70	4999	189.3	188.0	1.007
calibrator zero	5000	0.00	5000	0.0	0.0	n/a
Average C.F.=						1.004

Linear Regression/Calibration Results:

Correlation Coefficient = 1.000	LIMITS > or = 0.995
Slope = 1.000	.95-1.05
b (Intercept as % of full scale)= 0.10%	± 3% F.S.
% change in C.F. from last cal= n/a	± 10%

API 100E Sulphur Dioxide Analyzer Calibration




As found:	As left:
SLOPE: n/a	SLOPE: 1.010
OFFSET: n/a	OFFSET: 116.4
HVPS: n/a	HVPS: 651
RCELL TEMP: n/a	RCELL TEMP: 50.0
BOX TEMP: n/a	BOX TEMP: 29.1
PMT TEMP: n/a	PMT TEMP: 7.8
IZS TEMP: n/a	IZS TEMP: 53.0
PRES: n/a	PRES: 24.3
SAMP FL: n/a	SAMP FL: 599
NORM PMT: n/a	NORM PMT: 116.7
UV LAMP: n/a	UV LAMP: 3151.5
LAMP RATIO: n/a	LAMP RATIO: 100.1
STR. LGT: n/a	STR. LGT: 58.8
DRK PMT: n/a	DRK PMT: 5.6
DRK LMP: n/a	DRK LMP: 6.8
Expected Value: n/a	Expected Value: 442.0

Comments:
The analyzer sample inlet filter was changed.

Sample pump was rebuilt. Reaction cell cleaned. Output voltage calibrated. Manifold rebuilt. Valves tested for leaks. New Sample valve installed. ZERO Air scrubber was repositioned inside. Factory calibration completed.

SO2[ppb] Station: LICA ST. LINA Daily: 2017/03/02 Type: AVG 1 Min. [1 Min.]



— SO2[ppb]

HYDROGEN SULPHIDE



API 101E Hydrogen Sulphide Analyzer Calibration

Date: <u>March 1, 2017</u>	Barometric Pressure: <u>27.60 inHg</u>
Company/Airshed: <u>LICA</u>	Station Temperature °C: <u>19</u>
Location/Station Name: <u>St. Lina</u>	Weather Conditions: <u>Mainly sunny</u>
Parameter: <u>Hydrogen Sulphide</u>	Calibration Purpose: <u>shut down</u>
Start Time 24 hr. (mst): <u>9:20</u>	Performed By/Reviewer: <u>Limin Li</u> <u>Trina Whitsitt</u>
End Time 24 hr. (mst): <u>12:15</u>	Cal Gas Expiry Date: <u>January 6, 2018</u>
Calibration Method: <u>Gas Dilution</u>	Converter Model & s/n (if applicable): <u>Internal</u>

Analyzer:	
ID# or Serial Number: <u>509</u>	Range ppb: <u>100</u>
Last Calibration Date: <u>February 22, 2017</u>	As Found C.F.: <u>1.031</u>
Previous C.F.: <u>0.999</u>	New C.F.: <u>n/a</u>

Calibrator: Flow Meter ID's: <u>n/a</u> Make & Model: <u>Enviroics 2000</u> Serial #: <u>1991</u> Cal Gas Cylinder I.D. #: <u>BLM002508</u> Cal Gas Conc. (ppm): <u>10.2</u>	Standard Calibration Points for Ranges <table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19
Point	ppb								
High	78								
Mid	38								
Low	19								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated Concentration:		Indicated Concentration:		Correction Factors (C.F.):
	Diluent	Cal Gas	Total	(ppb)	(ppb)			
as found zero	7500	0.00	7500	0.0	0.2			n/a
as found high	7443	57.34	7500	78.0	75.8			1.031
mid	7473	27.92	7501	38.0	36.8			1.037
low	7488	13.92	7502	18.9	18.2			1.052
Average C.F.=								1.040

Linear Regression/Calibration Results:

Correlation Coefficient = <u>1.000</u>	LIMITS > or = 0.995
Slope = <u>1.030</u>	0.90-1.10
b (Intercept as % of full scale) = <u>-0.02%</u>	± 3% F.S.
% change in C.F. from last cal = <u>-3.25%</u>	± 10%


API 101E Hydrogen Sulphide Analyzer Calibration



As found: SLOPE: <u>0.924</u> OFFSET: <u>52.9</u> HVPS: <u>675</u> RCELL TEMP: <u>50</u> BOX TEMP: <u>29.1</u> PMT TEMP: <u>8.0</u> IZS TEMP: <u>48</u> Converter Temp: <u>314</u> PRES: <u>20.4</u> SAMP FL: <u>559</u> UV LAMP: <u>3452</u> LAMP RATIO: <u>92.4</u> STR. LGT: <u>24.5</u> DRK PMT: <u>0.3</u> DRK LMP: <u>0.5</u> Expected Value: <u>47.6</u>	As left: SLOPE: <u>n/a</u> OFFSET: <u>n/a</u> HVPS: <u>n/a</u> RCELL TEMP: <u>n/a</u> BOX TEMP: <u>n/a</u> PMT TEMP: <u>n/a</u> IZS TEMP: <u>n/a</u> Converter Temp: <u>n/a</u> PRES: <u>n/a</u> SAMP FL: <u>n/a</u> UV LAMP: <u>n/a</u> LAMP RATIO: <u>n/a</u> STR. LGT: <u>n/a</u> DRK PMT: <u>n/a</u> DRK LMP: <u>n/a</u> Expected Value: <u>n/a</u>
--	--

Comments:

As found high point repeated due to operator error. After shut-down calibration, the reaction cell was cleaned. The sinter filter and o-rings were changed. The sample valve was cleaned. The analogue output was calibrated. The valves were tested for leaks. The sample pump was rebuilt. Factory calibration completed. Zero Air scrubber was repositioned inside.



— H2S[ppb]



API 101E Hydrogen Sulphide Analyzer Calibration

Date: March 2, 2017	Barometric Pressure: 27.78 inHg
Company/Airshed: LICA	Station Temperature °C: 20
Location/Station Name: St. Lina	Weather Conditions: Mainly sunny
Parameter: Hydrogen Sulphide	Calibration Purpose: post repair
Start Time 24 hr. (mst): 12:00	Performed By/Reviewer: Alex Yakupov Trina Whitsitt
End Time 24 hr. (mst): 15:00	Cal Gas Expiry Date: June 14, 2019
Calibration Method: Gas Dilution	Converter Model & s/n (if applicable): n/a

Analyzer:		
ID# or Serial Number: 509	Range ppb: 100	Station SO2 Analyzer Range?
Last Calibration Date: n/a	As Found C.F.: n/a	1000 ppb
Previous C.F.: n/a	New C.F.: 0.999	

Calibrator:	Standard Calibration Points for Ranges								
Flow Meter ID's: n/a	<table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19
Point	ppb								
High	78								
Mid	38								
Low	19								
Make & Model: SABIO 2010 D									
Serial #: 11900613									
Cal Gas Cylinder I.D. #: EY0000654									
Cal Gas Conc. (ppm): 10.2									


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	(ppb)	(ppb)	
adjusted zero	7500	0.00	7500	0.0	0.0	n/a
adjusted high	7443	57.40	7500	78.1	78.1	0.999
mid	7471	27.90	7499	37.9	37.9	1.001
low	7486	14.00	7500	19.0	19.0	1.002
calibrator zero	7500	0.00	7500	0.0	0.0	n/a
Average C.F.=						1.001

Linear Regression/Calibration Results:

Correlation Coefficient = 1.000	LIMITS
Slope = 0.999	> or = 0.995
b (Intercept as % of full scale)= 0.03%	.95-1.05
% change in C.F. from last cal= n/a	± 3% F.S.
	± 10%


API 101E Hydrogen Sulphide Analyzer Calibration



As found:	As left:
SLOPE: n/a	SLOPE: 1.030
OFFSET: n/a	OFFSET: 53.0
HVPS: n/a	HVPS: 679
RCELL TEMP: n/a	RCELL TEMP: 50.0
BOX TEMP: n/a	BOX TEMP: 29.4
PMT TEMP: n/a	PMT TEMP: 8.0
IZS TEMP: n/a	IZS TEMP: 48.0
Converter Temp: n/a	Converter Temp: 315.4
PRES: n/a	PRES: 20.7
SAMP FL: n/a	SAMP FL: 536
UV LAMP: n/a	UV LAMP: 3077.7
LAMP RATIO: n/a	LAMP RATIO: 100.9
STR. LGT: n/a	STR. LGT: 27.3
DRK PMT: n/a	DRK PMT: 0.6
DRK LMP: n/a	DRK LMP: 0.3
Expected Value: n/a	Expected Value: 52.8

Comments:
The analyzer sample inlet filter was changed.

Valves tested for leaks. Sample pump was rebuilt. Reaction cell cleaned. Manifold rebuilt. Output voltage calibrated. Factory calibration completed. Zero Air scrubber was repositioned inside.



— H2S[ppb]



API 101E Hydrogen Sulphide Analyzer Calibration

Date: <u>March 7, 2017</u>	Barometric Pressure: <u>27.60 inHg</u>
Company/Airshed: <u>LICA</u>	Station Temperature °C: <u>21</u>
Location/Station Name: <u>St. Lina</u>	Weather Conditions: <u>Mainly sunny</u>
Parameter: <u>Hydrogen Sulphide</u>	Calibration Purpose: <u>repeat</u>
Start Time 24 hr. (mst): <u>9:38</u>	Performed By/Reviewer: <u>Limin Li</u> <u>Trina Whitsitt</u>
End Time 24 hr. (mst): <u>15:00</u>	Cal Gas Expiry Date: <u>January 6, 2018</u>
Calibration Method: <u>Gas Dilution</u>	Converter Model & s/n (if applicable): <u>Internal</u>

Analyzer:	
ID# or Serial Number: <u>509</u>	Range ppb: <u>100</u>
Last Calibration Date: <u>March 2, 2017</u>	As Found C.F.: <u>1.043</u>
Previous C.F.: <u>0.999</u>	New C.F.: <u>1.000</u>

Calibrator:	Standard Calibration Points for Ranges								
Flow Meter ID's: <u>n/a</u>	<table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19
Point	ppb								
High	78								
Mid	38								
Low	19								
Make & Model: <u>Enviroics 2000</u>									
Serial #: <u>1991</u>									
Cal Gas Cylinder I.D. #: <u>BLM002508</u>									
Cal Gas Conc. (ppm): <u>10.2</u>									


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated Concentration: (ppb)	Indicated Concentration: (ppb)	Correction Factors (C.F.):
	Diluent	Cal Gas	Total			
as found zero	7504	0.00	7504	0.0	-0.3	n/a
as found high	7448	57.38	7505	78.0	74.5	1.043
adjusted zero	7504	0.00	7504	0.0	0.0	n/a
adjusted high	7448	57.38	7505	78.0	78.0	1.000
mid	7477	27.93	7505	38.0	38.0	0.999
low	7491	13.95	7505	19.0	19.0	0.998
calibrator zero	7504	0.00	7504	0.0	0.2	n/a
Average C.F. =						0.999

Linear Regression/Calibration Results:


Correlation Coefficient = <u>1.000</u>	LIMITS
Slope = <u>1.000</u>	> or = 0.995
b (Intercept as % of full scale) = <u>-0.02%</u>	.95-1.05
% change in C.F. from last cal = <u>-4.36%</u>	± 3% F.S.
	± 10%

API 101E Hydrogen Sulphide Analyzer Calibration



<p style="text-align: center;">As found:</p> SLOPE: <u>1.03</u> OFFSET: <u>53</u> HVPS: <u>679</u> RCELL TEMP: <u>50</u> BOX TEMP: <u>30</u> PMT TEMP: <u>8.0</u> IZS TEMP: <u>48</u> Converter Temp: <u>314</u> PRES: <u>20.6</u> SAMP FL: <u>535</u> UV LAMP: <u>3092</u> LAMP RATIO: <u>101.1</u> STR. LGT: <u>27.3</u> DRK PMT: <u>0.5</u> DRK LMP: <u>0.3</u> Expected Value: <u>52.8</u>	<p style="text-align: center;">As left:</p> SLOPE: <u>1.066</u> OFFSET: <u>51.2</u> HVPS: <u>679</u> RCELL TEMP: <u>50</u> BOX TEMP: <u>30</u> PMT TEMP: <u>8</u> IZS TEMP: <u>48</u> Converter Temp: <u>314</u> PRES: <u>20.6</u> SAMP FL: <u>535</u> UV LAMP: <u>3092</u> LAMP RATIO: <u>101.1</u> STR. LGT: <u>27.3</u> DRK PMT: <u>0.5</u> DRK LMP: <u>0.3</u> Expected Value: <u>56.7</u>
--	--

Comments:



— H2S[ppb]



API 101E Hydrogen Sulphide Analyzer Calibration

Date: <u>March 23, 2017</u>	Barometric Pressure: <u>27.37 inHg</u>
Company/Airshed: <u>LICA</u>	Station Temperature °C: <u>21</u>
Location/Station Name: <u>St. Lina, Ab</u>	Weather Conditions: <u>A few clouds</u>
Parameter: <u>Hydrogen Sulphide</u>	Calibration Purpose: <u>shut down</u>
Start Time 24 hr. (mst): <u>9:40</u>	Performed By/Reviewer: <u>Michael Espiritu</u> <u>Trina Whitsitt</u>
End Time 24 hr. (mst): <u>10:56</u>	Cal Gas Expiry Date: <u>July 15, 2017</u>
Calibration Method: <u>Gas Dilution</u>	Converter Model & s/n (if applicable): <u>n/a</u>

Analyzer:	
ID# or Serial Number: <u>509</u>	Range ppb: <u>100</u>
Last Calibration Date: <u>March 7, 2017</u>	As Found C.F.: <u>0.983</u>
Previous C.F.: <u>1.000</u>	New C.F.: <u>n/a</u>

Calibrator: Flow Meter ID's: <u>n/a</u> Make & Model: <u>Sabio 2010</u> Serial #: <u>042531101(0911)</u> Cal Gas Cylinder I.D. #: <u>LL74267</u> Cal Gas Conc. (ppm): <u>9.88</u>	Standard Calibration Points for Ranges <table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19
Point	ppb								
High	78								
Mid	38								
Low	19								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Concentration		Correction Factors (C.F.):
	Diluent	Cal Gas	Total	Calculated (ppb)	Indicated (ppb)	
as found zero	7499	0.00	7499	0.0	0.4	n/a
as found high	7441	59.20	7500	78.0	79.7	0.983
mid	7471	28.80	7500	37.9	38.8	0.988
low	7484	14.40	7498	19.0	19.2	1.009
Average C.F. =						0.994

Linear Regression/Calibration Results:

Correlation Coefficient = <u>1.000</u>	LIMITS > or = 0.995
Slope = <u>0.981</u>	0.90-1.10
b (Intercept as % of full scale) = <u>-0.16%</u>	± 3% F.S.
% change in C.F. from last cal = <u>1.66%</u>	± 10%

API 101E Hydrogen Sulphide Analyzer Calibration



As found: SLOPE: <u>1.066</u> OFFSET: <u>51.2</u> HVPS: <u>679</u> RCELL TEMP: <u>50</u> BOX TEMP: <u>30.1</u> PMT TEMP: <u>8</u> IZS TEMP: <u>48</u> Converter Temp: <u>314</u> PRES: <u>21.6</u> SAMP FL: <u>535</u> UV LAMP: <u>3150</u> LAMP RATIO: <u>103</u> STR. LGT: <u>27.3</u> DRK PMT: <u>0.4</u> DRK LMP: <u>0.3</u> Expected Value: <u>56.7</u>	As left: SLOPE: <u>n/a</u> OFFSET: <u>n/a</u> HVPS: <u>n/a</u> RCELL TEMP: <u>n/a</u> BOX TEMP: <u>n/a</u> PMT TEMP: <u>n/a</u> IZS TEMP: <u>n/a</u> Converter Temp: <u>n/a</u> PRES: <u>n/a</u> SAMP FL: <u>n/a</u> UV LAMP: <u>n/a</u> LAMP RATIO: <u>n/a</u> STR. LGT: <u>n/a</u> DRK PMT: <u>n/a</u> DRK LMP: <u>n/a</u> Expected Value: <u>n/a</u>
---	--

Comments:

Shutdown calibration before checking the IZS system.



API 101E Hydrogen Sulphide Analyzer Calibration

Date: <u>March 23, 2017</u>	Barometric Pressure: <u>27.37 inHg</u>
Company/Airshed: <u>LICA</u>	Station Temperature °C: <u>21</u>
Location/Station Name: <u>St. Lina, Ab</u>	Weather Conditions: <u>A few clouds</u>
Parameter: <u>Hydrogen Sulphide</u>	Calibration Purpose: <u>post repair</u>
Start Time 24 hr. (mst): <u>12:50</u>	Performed By/Reviewer: <u>Michael Espiritu</u> <u>Trina Whitsitt</u>
End Time 24 hr. (mst): <u>15:25</u>	Cal Gas Expiry Date: <u>July 15, 2017</u>
Calibration Method: <u>Gas Dilution</u>	Converter Model & s/n (if applicable): <u>n/a</u>

Analyzer:	
ID# or Serial Number: <u>509</u>	Range ppb: <u>100</u>
Last Calibration Date: <u>March 7, 2017</u>	As Found C.F.: <u>n/a</u>
Previous C.F.: <u>n/a</u>	New C.F.: <u>0.997</u>


Calibrator:	Standard Calibration Points for Ranges								
Flow Meter ID's: <u>n/a</u>	<table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19
Point	ppb								
High	78								
Mid	38								
Low	19								
Make & Model: <u>Sabio 2010</u>									
Serial #: <u>042531101(0911)</u>									
Cal Gas Cylinder I.D. #: <u>LL74267</u>									
Cal Gas Conc. (ppm): <u>9.88</u>									

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	(ppb)	(ppb)	
adjusted zero	7499	0.00	7499	0.0	0.0	n/a
adjusted high	7441	59.20	7500	78.0	78.2	0.997
mid	7471	28.80	7500	37.9	37.8	1.004
low	7484	14.40	7498	19.0	18.5	1.026
calibrator zero	7499	0.00	7499	0.0	0.1	n/a
Average C.F. =						1.009

Linear Regression/Calibration Results:


Correlation Coefficient = <u>1.000</u>	LIMITS > or = 0.995
Slope = <u>0.995</u>	.95-1.05
b (Intercept as % of full scale) = <u>0.26%</u>	± 3% F.S.
% change in C.F. from last cal = <u>n/a</u>	± 10%



<p style="text-align: center;">As found:</p> SLOPE: <u>1.066</u> OFFSET: <u>51.2</u> HVPS: <u>679</u> RCELL TEMP: <u>50</u> BOX TEMP: <u>30.1</u> PMT TEMP: <u>8</u> IZS TEMP: <u>48</u> Converter Temp: <u>314</u> PRES: <u>21.6</u> SAMP FL: <u>535</u> UV LAMP: <u>3150</u> LAMP RATIO: <u>103</u> STR. LGT: <u>27.3</u> DRK PMT: <u>0.4</u> DRK LMP: <u>0.3</u> Expected Value: <u>56.7</u>	<p style="text-align: center;">As left:</p> SLOPE: <u>1.047</u> OFFSET: <u>51.8</u> HVPS: <u>679</u> RCELL TEMP: <u>50</u> BOX TEMP: <u>30.9</u> PMT TEMP: <u>8</u> IZS TEMP: <u>48</u> Converter Temp: <u>314</u> PRES: <u>20.5</u> SAMP FL: <u>535</u> UV LAMP: <u>3151</u> LAMP RATIO: <u>103</u> STR. LGT: <u>27.1</u> DRK PMT: <u>0.4</u> DRK LMP: <u>0.4</u> Expected Value: <u>61.5</u>
---	--

Comments:

Replaced a few IZS tubings, purged valves and restored original zero scrubber setup.



— H2S[ppb]

TOTAL HYDROCARBON



Thermo 51C Total Hydrocarbon Analyzer Calibration

Date:	March 1, 2017	Barometric Pressure:	27.61 inHg
Company/Airshed:	LICA	Station Temperature °C:	21
Location/Station Name:	St. Lina	Weather Conditions:	Mainly sunny
Parameter:	Total Hydrocarbon	Calibration Purpose:	shut down
Start/End Time 24 hr. (mst):	12:22/15:00	Performed By/Reviewer:	Limin Li Trina Whitsitt
Calibration Method:	Gas Dilution	Cal Gas Expiry Date:	July 7, 2022

Analyzer:	ID# or Serial Number:	51CLT-77021-384	Range ppm:	50
	Last Calibration Date:	February 3, 2017	As Found C.F.:	1.003
	Previous Cal High Point C.F.:	1.000	New C.F.:	n/a

Calibrator:	Flow Meter ID's:	n/a	Standard Calibration Points for a Range of:	50 ppm
	Make & Model:	EnviroNics 2000		
	Serial #:	1991		
	Cal Gas Cylinder I.D. #:	LL83638		
	CH ₄ /C ₃ H ₈ Cylinder Conc. (ppm):	582.0 203.0		
	CH ₄ as propane/total CH ₄ equivalents (ppm):	558.3 1140.3		


Point	Target ppm
High	38
Mid	18
Low	9

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated Concentration (ppm)	Indicated Concentration (ppm)	Correction Factors:
	Diluent	Cal Gas	Total			
as found zero	2005	0.00	2005	0.0	0.30	n/a
as found high	2005	69.12	2074	38.00	38.20	1.003
mid	2005	32.20	2037	18.02	18.10	1.013
low	2005	16.00	2021	9.03	9.13	1.022
Average C.F.=						1.013

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	1.001		.95-1.05
b (Intercept as % of full scale) =	-0.39%		± 3% F.S.
% change in C.F. from last cal =	-0.27%		± 10%



As found:	H2 cylinder (psi):	1800	As left:	H2 cylinder (psi):	n/a
	H2 cylinder reg set (psi):	24		H2 cylinder reg set (psi):	n/a
	Span Cylinder (psi):	1900		Span Cylinder (psi):	n/a
	Span Cylinder Reg Set (psi):	20		Span Cylinder Reg Set (psi):	n/a
	Zero Air Gen Pressure:	46		Zero Air Gen Pressure:	n/a
	measurement alarms:	None		measurement alarms:	n/a
	service alarms:	None		service alarms:	n/a
	cnt:	3292		cnt:	n/a
	rng:	1		rng:	n/a
	try:	0		try:	n/a
	flm:	194.5		flm:	n/a
	det:	125.5		det:	n/a
	Flame:	194		Flame:	n/a
	Filter:	125		Filter:	n/a
	Base:	125		Base:	n/a
	Sample psi:	6.9		Sample psi:	n/a
	Internal Air Pressure:	20		Internal Air Pressure:	n/a
	Internal Fuel Pressure:	12		Internal Fuel Pressure:	n/a
	Measured Flow:	0.8988		Measured Flow:	n/a
	Expected Value:	26.94		Expected Value:	n/a

Comments:

After shut-down calibration, the pump was rebuilt.



Thermo 51C Total Hydrocarbon Analyzer Calibration

Date:	March 1, 2017	Barometric Pressure:	27.61 inHg
Company/Airshed:	UICA	Station Temperature °C:	21
Location/Station Name:	St. Lina	Weather Conditions:	Mainly sunny
Parameter:	Total Hydrocarbon	Calibration Purpose:	post repair
Start/End Time 24 hr. (mst):	17:20 / 20:20	Performed By/Reviewer:	Alex Yakupov / Trina Whitsitt
Calibration Method:	Gas Dilution	Cal Gas Expiry Date:	July 7, 2022

Analyzer:	
ID# or Serial Number:	51CLT-77021-384
Last Calibration Date:	n/a
Previous Cal High Point C.F.:	n/a
Range ppm:	50
As Found C.F.:	n/a
New C.F.:	1.000

Calibrator:	
Flow Meter ID's:	n/a
Make & Model:	API 700
Serial #:	627
Cal Gas Cylinder I.D. #:	LL83638
CH ₄ /C ₂ H ₆ Cylinder Conc. (ppm):	582.0 203.0
CH ₄ as propane/total CH ₄ equivalents (ppm):	558.3 1140.3
Standard Calibration Points for a Range of: 50 ppm	
Point	Target ppm
High	38
Mid	18
Low	9


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors:
Point	Diluent	Cal Gas	Total	(ppm)	(ppm)	
adjusted zero	2005	0.00	2005	0.0	0.00	n/a
adjusted high	2002	69.10	2071	38.04	38.04	1.000
mid	2003	32.20	2035	18.04	18.00	1.002
low	2003	16.00	2019	9.04	9.00	1.004
calibrator zero	2005	0.00	2005	0.00	0.00	n/a
Average C.F. =						1.002

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS
Slope =	1.000	> or = 0.995
b (Intercept as % of full scale) =	0.04%	.95-1.05
% change in C.F. from last cal =	n/a	± 3% F.S.
		± 10%

Thermo 51C Total Hydrocarbon Analyzer Calibration



As found:

H2 cylinder (psi):	n/a
H2 cylinder reg set (psi):	n/a
Span Cylinder (psi):	n/a
Span Cylinder Reg Set (psi):	n/a
Zero Air Gen Pressure:	n/a
measurement alarms:	n/a
service alarms:	n/a
cnt:	n/a
rng:	n/a
try:	n/a
flm:	n/a
det:	n/a
Flame:	n/a
Filter:	n/a
Base:	n/a
Sample psi:	n/a
Internal Air Pressure:	n/a
Internal Fuel Pressure:	n/a
Measured Flow:	n/a
Expected Value:	n/a

As left:


H2 cylinder (psi):	1800
H2 cylinder reg set (psi):	24
Span Cylinder (psi):	1900
Span Cylinder Reg Set (psi):	20
Zero Air Gen Pressure:	46
measurement alarms:	None
service alarms:	None
cnt:	1954
rng:	1
try:	1
flm:	193.3
det:	125.7
Flame:	193
Filter:	125
Base:	125
Sample psi:	06.91
Internal Air Pressure:	20
Internal Fuel Pressure:	13
Measured Flow:	n/a
Expected Value:	30.70

Comments:

The analyzer sample inlet filter was changed.

A new span gas cylinder was installed.
The analyzer cooling fan filter(s) were cleaned.

Sample pump was rebuilt.



— THC[ppm]

NITROGEN DIOXIDE



API 200E NO-NO2-NOx Analyzer Calibration

Date: March 1, 2017 Company/Airshed: LICA Location/Station Name: St. Lina Start/End Time 24 hr. (mst): 9:40 / 15:00 G.P.T. to be used for Ozone? Yes with 1000 ppb NOx full scale Calibration Method: Gas Dilution & Varying UV Lamp Power	Barometric Pressure: 27.6 inHg Station Temperature °C: 19 Weather Conditions: Mainly sunny Calibration Purpose: shut down Performed By/Reviewer: Limin Li Trina Whatsitt Cal Gas Expiry Date: December 25, 2018
---	--

Analyzer: ID# or Serial Number: 594 Last Calibration Date: February 2, 2017 Range ppb: 1000	Correction Factors: <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th>Previous C.F.:</th> <th>As Found C.F.:</th> <th>New C.F.:</th> </tr> </thead> <tbody> <tr> <td>NO =</td> <td>1.000</td> <td>1.012</td> <td>n/a</td> </tr> <tr> <td>NO₂ =</td> <td>1.012</td> <td>1.014</td> <td>n/a</td> </tr> <tr> <td>NOx =</td> <td>1.000</td> <td>1.009</td> <td>n/a</td> </tr> </tbody> </table>		Previous C.F.:	As Found C.F.:	New C.F.:	NO =	1.000	1.012	n/a	NO ₂ =	1.012	1.014	n/a	NOx =	1.000	1.009	n/a
	Previous C.F.:	As Found C.F.:	New C.F.:														
NO =	1.000	1.012	n/a														
NO ₂ =	1.012	1.014	n/a														
NOx =	1.000	1.009	n/a														

Calibrator: Flow Meter ID's: n/a Make & Model: Sabio 2010 Serial #: 17200415 Cal Gas Cylinder I.D. #: BLM002756T NO/NOx Gas Conc. (ppm): 50.7 50.7	Standard Calibration Points for a Range of: 1000 ppb <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Point</th> <th>Target NO (ppb)</th> <th>Target NO₂ (ppb)</th> <th>Cc Ozone ?</th> </tr> </thead> <tbody> <tr> <td>High</td> <td>610</td> <td>375</td> <td><-high ozone</td> </tr> <tr> <td>Mid</td> <td>380</td> <td>190</td> <td><-mid ozone</td> </tr> <tr> <td>Low</td> <td>190</td> <td>70</td> <td><-low ozone</td> </tr> <tr> <td>Extra Point #1</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> <tr> <td>Extra Point #2</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> </tbody> </table>	Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?	High	610	375	<-high ozone	Mid	380	190	<-mid ozone	Low	190	70	<-low ozone	Extra Point #1	n/a	n/a	n/a	Extra Point #2	n/a	n/a	n/a
Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?																						
High	610	375	<-high ozone																						
Mid	380	190	<-mid ozone																						
Low	190	70	<-low ozone																						
Extra Point #1	n/a	n/a	n/a																						
Extra Point #2	n/a	n/a	n/a																						

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Diluent	Cal Gas	Total Flow	Calculated NO (ppb)	Calculated NOx (ppb)	Indicated NO (ppb)	Indicated NOx (ppb)	NO C.F.	NOx C.F.
as found zero	5028	0.0	5028	0	0	0.1	0.5	n/a	n/a
as found high	4951	77.1	5028	777.4	777.4	768.0	771.0	1.012	1.009
mid	4991	37.70	5029	380.1	380.1	377.0	377.0	1.008	1.010
low	5008	19.10	5027	192.6	192.6	190.0	190.0	1.014	1.017
Average C.F.=								1.012	1.012

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Diluent	Cal Gas	Total Flow	Calibrator Setting (volts or ppb)	Indicated NO (ppb)	Indicated NOx (ppb)	Indicated NO ₂ (ppb)	NO drop (ppb)	NO ₂ gain (ppb)	NO ₂ C.F.
NOx reference	4951	77.10	5028	0.0	768.0	769.0	1.0	0.1	1.0	n/a
as found high NO2	4951	77.10	5028	510.0	268.0	758.0	494.0	500.0	493.0	1.014
gpt mid	4951	77.10	5028	275.0	500.0	762.0	261.0	268.0	260.0	1.031
gpt low	4951	77.10	5028	100.0	673.0	768.0	95.0	95.0	94.0	1.011
Average NO₂ C.F.=										1.019

Linear Regression/Calibration Results:

	NO	NOx	NO ₂	LIMITS
Correlation Coefficient =	1.000	1.000	1.000	> or = 0.995
Slope =	1.012	1.008	1.017	0.90-1.10
b (Intercept as % of full scale)=	0.03%	-0.01%	-0.01%	± 3% F.S.
% change in C.F. from last cal=	-1.24%	-0.22%	-0.90%	± 10%
NO2 converter efficiency			0.99	0.96 to 1.04




As found: NOx SLOPE: 0.998 NOx OFFS: 2.2 NO SLOPE: 1.008 NO OFFS: 0.3 SAMP FLW: 483 OZONE FL: 78 PMT: 19.1 NORM PMT: 1 AZERO: 15.9 HVPS: 767 RCELL TEMP: 50 BOX TEMP: 29.8 PMT TEMP: 6.7 IZS TEMP: 40 MOLY TEMP: 315.2 RCEL: 5.1 SAMP: 26.2 Expected Value NO: 422.4 Expected Value NO2: 6.0 Expected Value NOx: 427.5	As left: NOx SLOPE: n/a NOx OFFS: n/a NO SLOPE: n/a NO OFFS: n/a SAMP FLW: n/a OZONE FL: n/a PMT: n/a NORM PMT: n/a AZERO: n/a HVPS: n/a RCELL TEMP: n/a BOX TEMP: n/a PMT TEMP: n/a IZS TEMP: n/a MOLY TEMP: n/a RCEL: n/a SAMP: n/a Expected Value NO: n/a Expected Value NO2: n/a Expected Value NOx: n/a
---	---


Comments:

As found high point repeated due to operator error. GPT low point starts at 14:40. After shut-down calibration, the reaction cell was cleaned. The sinter filter and o-rings were changed. The sample valve was cleaned. The NO/NO2 valve (leak) was changed. The O3 O-ring was changed. Output voltage calibrated.

Date: March 1, 2017
Company/Airshed: LICA
Location/Station Name: St. Lina

Start/End Time 24 hr. (mst): 9:40 / 15:00
Calibration Purpose: shut down
Calibration Method: Gas Dilution & Varying UV Lamp Power





— NOX[ppb] — NO[ppb] — NO2[ppb]



API 200E NO-NO2-NOx Analyzer Calibration

Date: March 2, 2017
Company/Airshed: LICA
Location/Station Name: St. Lina
Start/End Time 24 hr. (mst): 10:59 / 19:50
G.P.T. to be used for Ozone? No
Calibration Method: Gas Dilution & Gas Phase Titration

Barometric Pressure: 27.78 inHg
Station Temperature °C: 20
Weather Conditions: Mainly sunny
Calibration Purpose: post repair
Performed By/Reviewer: Alex Yakupov | Trina Whitsitt
Cal Gas Expiry Date: July 18, 2019

Analyzer:

ID# or Serial Number: 594
Last Calibration Date: n/a
Range ppb: 1000

Correction Factors:

	Previous C.F.:	As Found C.F.:	New C.F.:
NO =	n/a	n/a	1.000
NO ₂ =	n/a	n/a	1.010
NOx =	n/a	n/a	1.000

Calibrator:

Flow Meter ID's: n/a
Make & Model: API 700
Serial #: 627
Cal Gas Cylinder I.D. #: LL104222
NO/NOx Gas Conc. (ppm): 50.7 | 50.7

Standard Calibration Points for a Range of: 1000 ppb

Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?
High	780	500	n/a
Mid	380	275	n/a
Low	190	100	n/a
Extra Point #1	n/a	n/a	n/a
Extra Point #2	n/a	n/a	n/a

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated NO	Calculated NOx	Indicated NO	Indicated NOx	NO C.F.	NOx C.F.
	Diluent	Cal Gas	Total Flow	(ppb)	(ppb)	(ppb)	(ppb)		
adjusted zero	5000	0.0	5000	0	0	0.0	0.0	n/a	n/a
adjusted high	4924	76.9	5001	779.6	779.6	780.0	780.0	1.000	1.000
mid	4965	37.50	5003	380.1	380.1	379.0	379.0	1.003	1.003
low	4980	18.70	4999	189.7	189.7	189.0	189.0	1.004	1.004
calibrator zero	5000	0.00	5000	0.0	0.0	0.0	0.0	n/a	n/a
Average C.F.=								1.002	1.002

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calibrator Setting	Indicated NO	Indicated NOx	Indicated NO ₂	NO drop	NO ₂ gain	NO ₂ C.F.
	Diluent	Cal Gas	Total Flow	volts or ppb	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)
NOx reference	4924	76.90	5001	0.0	783.0	781.0	-2.0	0.0	-2.0	
adjusted high NO2	4924	76.90	5001	500.0	271.0	276.0	505.0	512.0	507.0	1.010
gpt mid	4924	76.90	5001	275.0	500.0	776.0	276.0	283.0	278.0	1.018
gpt low	4924	76.90	5001	105.0	676.0	778.0	102.0	107.0	104.0	1.029
Average NO₂ C.F.=									1.019	

Linear Regression/Calibration Results:

	NO	NOx	NO ₂	LIMITS
Correlation Coefficient =	1.000	1.000	1.000	> or = 0.995
Slope =	0.999	0.999	1.006	.95-1.05
b (Intercept as % of full scale)=	-0.06%	-0.06%	-0.24%	± 3% F.S.
% change in C.F. from last cal=	n/a	n/a	n/a	± 10%
NO2 converter efficiency			0.96	0.96 to 1.04

As found:

NOx SLOPE: n/a
 NOx OFFS: n/a
 NO SLOPE: n/a
 NO OFFS: n/a
 SAMP FLW: n/a
 OZONE FL: n/a
 PMT: n/a
 NORM PMT: n/a
 AZERO: n/a
 HVPS: n/a
 RCELL TEMP: n/a
 BOX TEMP: n/a
 PMT TEMP: n/a
 IZS TEMP: n/a
 MOLY TEMP: n/a
 RCEL: n/a
 SAMP: n/a
 Expected Value NO: n/a
 Expected Value NO2: n/a
 Expected Value NOx: n/a

As left:

NOx SLOPE: 0.967
 NOx OFFS: -1.6
 NO SLOPE: 0.964
 NO OFFS: -2.7
 SAMP FLW: 485
 OZONE FL: 78
 PMT: 20.0
 NORM PMT: -2.3
 AZERO: 21.1
 HVPS: 767
 RCELL TEMP: 50.0
 BOX TEMP: 29.7
 PMT TEMP: 6.6
 IZS TEMP: 40.0
 MOLY TEMP: 316.3
 RCEL: 5.2
 SAMP: 26.2
 Expected Value NO: 7.1
 Expected Value NO2: 420.0
 Expected Value NOx: 426.0




Comments:

The analyzer sample inlet filter was changed.

Reaction cell cleaned. Valves rebuilt and checked for leaks. Factory calibration completed. Output voltage calibrated.

Date: March 2, 2017
Company/Airshed: LICA
Location/Station Name: St. Lina

Start/End Time 24 hr. (mst): 10:59 / 19:50
Calibration Purpose: post repair
Calibration Method: Gas Dilution & Gas Phase Titration





API 200E NO-NO2-NOx Analyzer Calibration

Date: <u>March 2, 2017</u>	Barometric Pressure: <u>27.78 inHg</u>
Company/Airshed: <u>LICA</u>	Station Temperature °C: <u>20</u>
Location/Station Name: <u>St. Lina</u>	Weather Conditions: <u>Mainly sunny</u>
Start/End Time 24 hr. (mst): <u>10:59 / 19:50</u>	Calibration Purpose: <u>post repair</u>
G.P.T. to be used for Ozone? <u>No</u>	Performed By/Reviewer: <u>Alex Yakupov</u> <u>Trina Whitsitt</u>
Calibration Method: <u>Gas Dilution & Gas Phase Titration</u>	Cal Gas Expiry Date: <u>July 18, 2019</u>

Analyzer:		Correction Factors:		
ID# or Serial Number: <u>594</u>	Previous C.F.:	As Found C.F.:	New C.F.:	
Last Calibration Date: <u>n/a</u>	NO = <u>n/a</u>	<u>n/a</u>	<u>1.000</u>	
Range ppb: <u>1000</u>	NO ₂ = <u>n/a</u>	<u>n/a</u>	<u>1.004</u>	
	NOx = <u>n/a</u>	<u>n/a</u>	<u>1.000</u>	

Calibrator: Flow Meter ID's: <u>n/a</u> Make & Model: <u>API 700</u> Serial #: <u>627</u> Cal Gas Cylinder I.D. #: <u>LL104222</u> NO/NOx Gas Conc. (ppm): <u>50.7</u> <u>50.7</u>	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th colspan="4" style="text-align: center;">Standard Calibration Points for a Range of: 1000 ppb</th> </tr> <tr> <th>Point</th> <th>Target NO (ppb)</th> <th>Target NO₂ (ppb)</th> <th>Cc Ozone ?</th> </tr> <tr> <td>High</td> <td>780</td> <td>500</td> <td>n/a</td> </tr> <tr> <td>Mid</td> <td>380</td> <td>275</td> <td>n/a</td> </tr> <tr> <td>Low</td> <td>190</td> <td>100</td> <td>n/a</td> </tr> <tr> <td>Extra Point #1</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> <tr> <td>Extra Point #2</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> </table>	Standard Calibration Points for a Range of: 1000 ppb				Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?	High	780	500	n/a	Mid	380	275	n/a	Low	190	100	n/a	Extra Point #1	n/a	n/a	n/a	Extra Point #2	n/a	n/a	n/a
Standard Calibration Points for a Range of: 1000 ppb																													
Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?																										
High	780	500	n/a																										
Mid	380	275	n/a																										
Low	190	100	n/a																										
Extra Point #1	n/a	n/a	n/a																										
Extra Point #2	n/a	n/a	n/a																										

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015


Calibrator Flow Rates (cc/min)			Calibrator Setting	Indicated NO	Indicated NOx	Indicated NO ₂	NO drop	NO ₂ gain	NO ₂ C.F.
Point	Diluent	Cal Gas	Total Flow	volts or ppb	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)
NOx reference	4924	76.90	5001	0.0	783.0	778.0	-5.0	0.0	-5.0
adjusted high NO2	4924	76.90	5001	500.0	270.0	776.0	506.0	513.0	511.0
gpt mid	4924	76.90	5001	275.0	501.0	776.0	276.0	282.0	281.0
gpt low	4924	76.90	5001	105.0	677.0	777.0	101.0	106.0	106.0
									Average NO ₂ C.F.= <u>1.002</u>

Linear Regression/Calibration Results:				LIMITS	
Correlation Coefficient =	NO	NOx	NO ₂	> or =	0.995
Slope =	1.000	1.000	1.000	.95-1.05	
b (Intercept as % of full scale)=	0.999	0.999	0.997	± 3% F.S.	
% change in C.F. from last cal=	-0.06%	-0.06%	-0.27%	± 10%	
NO2 converter efficiency	n/a	n/a	n/a	0.96 to 1.04	
	n/a	n/a	0.96		

<p style="text-align: center;">As found:</p> NOx SLOPE: <u>n/a</u> NOx OFFS: <u>n/a</u> NO SLOPE: <u>n/a</u> NO OFFS: <u>n/a</u> SAMP FLW: <u>n/a</u> OZONE FL: <u>n/a</u> PMT: <u>n/a</u> NORM PMT: <u>n/a</u> AZERO: <u>n/a</u> HVPS: <u>n/a</u> RCELL TEMP: <u>n/a</u> BOX TEMP: <u>n/a</u> PMT TEMP: <u>n/a</u> IZS TEMP: <u>n/a</u> MOLY TEMP: <u>n/a</u> RCEL: <u>n/a</u> SAMP: <u>n/a</u> Expected Value NO: <u>n/a</u> Expected Value NO ₂ : <u>n/a</u> Expected Value NOx: <u>n/a</u>	<p style="text-align: center;">As left:</p> NOx SLOPE: <u>0.967</u> NOx OFFS: <u>-1.6</u> NO SLOPE: <u>0.964</u> NO OFFS: <u>-2.7</u> SAMP FLW: <u>485</u> OZONE FL: <u>78</u> PMT: <u>20.0</u> NORM PMT: <u>-2.3</u> AZERO: <u>21.1</u> HVPS: <u>767</u> RCELL TEMP: <u>50.0</u> BOX TEMP: <u>29.7</u> PMT TEMP: <u>6.6</u> IZS TEMP: <u>40.0</u> MOLY TEMP: <u>316.3</u> RCEL: <u>5.2</u> SAMP: <u>26.2</u> Expected Value NO: <u>6.1</u> Expected Value NO ₂ : <u>387.0</u> Expected Value NOx: <u>392.0</u>
--	---

Comments:
 The analyzer sample inlet filter was changed.

GPT calibration repeated with reference point in GPT mode.



— NOX[ppb] — NO[ppb] — NO2[ppb]



API 200E NO-NO2-NOx Analyzer Calibration

Date: March 7, 2017	Barometric Pressure: 27.6 inHg
Company/Airshed: LICA	Station Temperature °C: 21
Location/Station Name: St. Lina	Weather Conditions: Mainly sunny
Start/End Time 24 hr. (mst): 10:00 / 16:50	Calibration Purpose: repeat
G.P.T. to be used for Ozone? No	Performed By/Reviewer: Limin Li Trina Whitsitt
Calibration Method: Gas Dilution & Gas Phase Titration	Cal Gas Expiry Date: December 25, 2018

Analyzer: ID# or Serial Number: 594 Last Calibration Date: March 2, 2017 Range ppb: 1000	Correction Factors: <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th>Previous C.F.:</th> <th>As Found C.F.:</th> <th>New C.F.:</th> </tr> </thead> <tbody> <tr> <td>NO =</td> <td>1.000</td> <td>1.001</td> <td>1.000</td> </tr> <tr> <td>NO₂ =</td> <td>1.010</td> <td>0.988</td> <td>1.002</td> </tr> <tr> <td>NOx =</td> <td>1.000</td> <td>0.999</td> <td>1.000</td> </tr> </tbody> </table>		Previous C.F.:	As Found C.F.:	New C.F.:	NO =	1.000	1.001	1.000	NO ₂ =	1.010	0.988	1.002	NOx =	1.000	0.999	1.000
	Previous C.F.:	As Found C.F.:	New C.F.:														
NO =	1.000	1.001	1.000														
NO ₂ =	1.010	0.988	1.002														
NOx =	1.000	0.999	1.000														

Calibrator: Flow Meter ID's: n/a Make & Model: Sabio 2010 Serial #: 17200415 Cal Gas Cylinder I.D. #: BLM002756T NO/NOx Gas Conc. (ppm): 50.7 50.7	Standard Calibration Points for a Range of: 1000 ppb <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Point</th> <th>Target NO (ppb)</th> <th>Target NO₂ (ppb)</th> <th>Cc Ozone ?</th> </tr> </thead> <tbody> <tr> <td>High</td> <td>780</td> <td>500</td> <td>n/a</td> </tr> <tr> <td>Mid</td> <td>380</td> <td>275</td> <td>n/a</td> </tr> <tr> <td>Low</td> <td>190</td> <td>100</td> <td>n/a</td> </tr> <tr> <td>Extra Point #1</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> <tr> <td>Extra Point #2</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> </tbody> </table>	Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?	High	780	500	n/a	Mid	380	275	n/a	Low	190	100	n/a	Extra Point #1	n/a	n/a	n/a	Extra Point #2	n/a	n/a	n/a
Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?																						
High	780	500	n/a																						
Mid	380	275	n/a																						
Low	190	100	n/a																						
Extra Point #1	n/a	n/a	n/a																						
Extra Point #2	n/a	n/a	n/a																						

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated NO	Calculated NOx	Indicated NO	Indicated NOx	NO C.F.	NOx C.F.
Point	Diluent	Cal Gas	Total Flow	(ppb)	(ppb)	(ppb)	(ppb)		
as found zero	5028	0.0	5028	0	0	1.4	1.0	n/a	n/a
as found high	4951	77.1	5028	777.4	777.4	778.0	779.0	1.001	0.999
adjusted zero	5028	0.00	5028	0.0	0.0	0.0	0.0	n/a	n/a
adjusted high	4951	77.10	5028	777.4	777.4	777.4	777.4	1.000	1.000
mid	4991	37.70	5029	380.1	380.1	380.0	380.0	1.000	1.000
low	5009	19.00	5028	191.6	191.6	192.0	192.0	0.998	0.998
calibrator zero	1	0.00		0	0	0.3	0.3	n/a	n/a
Average C.F.=								0.999	0.999

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calibrator Setting	Indicated NO	Indicated NOx	Indicated NO ₂	NO drop	NO ₂ gain	NO ₂ C.F.
Point	Diluent	Cal Gas	Total Flow	volts or ppb	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)
NOx reference	4951	77.10	5028	0.0	777.0	775.0	-2.0	0.0	-2.0	
as found high NO ₂	4951	77.10	5028	500.0	276.0	780.0	505.0	501.0	507.0	0.988
adjusted high NO ₂	4951	77.10	5028	500.0	276.0	774.0	498.0	501.0	500.0	1.002
gpt mid	4951	77.10	5028	275.0	507.0	774.0	268.0	270.0	270.0	1.000
gpt low	4951	77.10	5028	100.0	684.0	774.0	91.0	93.0	93.0	1.000
Average NO ₂ C.F.=										1.001

Linear Regression/Calibration Results:

	NO	NOx	NO ₂	LIMITS
Correlation Coefficient =	1.000	1.000	1.000	> or = 0.995
Slope =	1.000	1.000	0.999	.95-1.05
b (Intercept as % of full scale)=	0.01%	0.01%	-0.10%	± 3% F.S.
% change in C.F. from last cal=	-0.11%	0.07%	2.16%	± 10%
NO ₂ converter efficiency			0.97	0.96 to 1.04




As found:	As left:
NOx SLOPE: 0.967	NOx SLOPE: 0.965
NOx OFFS: -1.6	NOx OFFS: 0.2
NO SLOPE: 0.964	NO SLOPE: 0.962
NO OFFS: -2.7	NO OFFS: -0.3
SAMP FLW: 483	SAMP FLW: 480
OZONE FL: 78	OZONE FL: 78
PMT: 1430	PMT: 18.3
NORM PMT: 1640	NORM PMT: 2.1
AZERO: 17	AZERO: 18
HVPS: 767	HVPS: 767
RCELL TEMP: 50	RCELL TEMP: 50
BOX TEMP: 29.8	BOX TEMP: 31.9
PMT TEMP: 6.7	PMT TEMP: 6.7
IZS TEMP: 40	IZS TEMP: 40.3
MOLY TEMP: 315.2	MOLY TEMP: 314
RCEL: 5.1	RCEL: 5.3
SAMP: 26.6	SAMP: 25.8
Expected Value NO: 420.0	Expected Value NO: 420.0
Expected Value NO ₂ : 7.1	Expected Value NO ₂ : 7.1
Expected Value NOx: 426.0	Expected Value NOx: 426.0


Comments:

Before the calibration, the GPT high point was checked for 40 minutes and NO₂ was found to be stable. NO₂ converter is deemed okay. Repeat calibration complete. The analyzer does not require replacing.

Date: March 7, 2017
Company/Airshed: LICA
Location/Station Name: St. Lina

Start/End Time 24 hr. (mst): 10:00 / 16:50
Calibration Purpose: repeat
Calibration Method: Gas Dilution & Gas Phase Titration





— NOX[ppb] — NO[ppb] — NO2[ppb]

OZONE

Maxxam Thermo 49i Ozone Analyzer Calibration

A Bureau Veritas Group Company

Date:	March 1, 2017	Barometric Pressure:	27.61 inHg
Company/Airshed:	LICA	Station Temperature °C:	21
Location/Station Name:	St. Lina	Weather Conditions:	Mainly sunny
Start/End Time 24 hr. (mst):	15:50 / 18:26	Calibration Purpose:	shut down
Ozone Calibration Method:	Direct G.P.T.	Performed By/Reviewer:	Limin Li / Trina Whitsitt
G.P.T. Date:	March 1, 2017	Cal Gas Expiry Date:	n/a

Analyzer:	
ID# or Serial Number:	1002240371
Last Calibration Date:	February 3, 2017
Previous Cal High Point C.F.:	1.000
Ozone Range ppb:	500
As Found C.F.:	1.024
New C.F.:	n/a

Calibrator:	
Flow Meter ID's:	n/a
Make & Model:	SABIO 2010
Serial #:	17200415
Cal Gas Cylinder I.D. #:	n/a

Point	AMD Required Range of Ozone Calibration Points
High	300-400 ppb
Mid	150-200 ppb
Low	50-75 ppb


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rate (cc/min)		Calculated Concentration:	Corrected Calculated Concentration:	Indicated Concentration:	Correction Factors:
	Total Flow @ Point Start	Total Flow @ Point Finish	(ppb)	(ppb)	(ppb)	
as found zero	5000	5000	0.0	n/a	-5.0	n/a
as found high	5000	5000	382.0	382.0	368.0	1.024
mid	5000	5000	176.0	176.0	172.0	0.994
low	5000	5000	55.0	55.0	50.0	1.000
Average C.F. =						1.006

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	1.024		0.90-1.10
b (Intercept as % of full scale) =	0.70%		± 3% F.S.
% change in C.F. from last cal =	-2.41%		± 10%


Thermo 49i Ozone Analyzer Calibration



As found:	O3 Bkg: 3.8	As left:	O3 Bkg: n/a
	O3 Coef: 0.937		O3 Coef: n/a
	Photo Lamp: 9.4		Photo Lamp: n/a
	O3 Lamp: 7.8		O3 Lamp: n/a
	Bench: 26.9		Bench: n/a
	Bench Lamp: 53.6		Bench Lamp: n/a
	O3 Lamp: 67.8		O3 Lamp: n/a
	Pressure: 680		Pressure: n/a
	Cell A lpm: 0.726		Cell A lpm: n/a
	Cell B lpm: 0.720		Cell B lpm: n/a
	O3 ppb: 28.9		O3 ppb: n/a
	Cell A ppb: 30.0		Cell A ppb: n/a
	Cell B ppb: 27		Cell B ppb: n/a
	Cell A int: 55429		Cell A int: n/a
	Expected Value: 377.0		Expected Value: n/a

Comments:

Cell B Int:70378; Shut-down calibration completed to rebuild a sample pump and clean a reaction cell. Leak check pass. Change charcoal container to address leaks.



— O3[ppb]

Maxxam Thermo 49i Ozone Analyzer Calibration

A Bureau Veritas Group Company

Date:	March 2, 2017	Barometric Pressure:	27.78 inHg
Company/Airshed:	LICA	Station Temperature °C:	20
Location/Station Name:	St. Lina	Weather Conditions:	Mainly sunny
Start/End Time 24 hr. (mst):	15:43 / 19:35	Calibration Purpose:	post repair
Ozone Calibration Method:	Varying UV Lamp Power	Performed By/Reviewer:	Alex Yakupov Trina Whitsitt
G.P.T. Date:	n/a-done by Varying UV Lamp Power	Cal Gas Expiry Date:	n/a

Analyzer:	
ID# or Serial Number:	1002240371
Last Calibration Date:	n/a
Previous Cal High Point C.F.:	n/a
Ozone Range ppb:	500
As Found C.F.:	n/a
New C.F.:	1.000

Calibrator:	
Flow Meter ID's:	n/a
Make & Model:	SABIO 2010 D
Serial #:	11900613
Cal Gas Cylinder I.D. #:	n/a


Point	AMD Required Range of Ozone Calibration Points
High	300-400 ppb
Mid	150-200 ppb
Low	50-75 ppb

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rate (cc/min)		Calculated Concentration:	Corrected Calculated Concentration:	Indicated Concentration:	Correction Factors:
	Total Flow @ Point Start	Total Flow @ Point Finish	(ppb)	(ppb)	(ppb)	
adjusted zero	5000	5000	0.0	n/a	0.0	n/a
adjusted high	5000	5000	380.0	380.0	380.0	1.000
mid	5000	5000	180.0	180.0	181.0	0.994
low	5000	5000	60.0	60.0	61.0	0.984
calibrator zero	5000	5000	0.0	0.0	1.0	n/a
Average C.F. =						0.993

Linear Regression/Calibration Results:


Correlation Coefficient =	<u>1.000</u>	LIMITS	> or = 0.995
Slope =	<u>1.000</u>		.95-1.05
b (Intercept as % of full scale) =	<u>0.00%</u>		± 3% F.S.
% change in C.F. from last cal =	<u>n/a</u>		± 10%



As found:		As left:	
O3 Bkg:	<u>n/a</u>	O3 Bkg:	<u>-1.5</u>
O3 Coef:	<u>n/a</u>	O3 Coef:	<u>0.978</u>
Photo Lamp:	<u>n/a</u>	Photo Lamp:	<u>10.7</u>
O3 Lamp:	<u>n/a</u>	O3 Lamp:	<u>8.2</u>
Bench:	<u>n/a</u>	Bench:	<u>27.0</u>
Bench Lamp:	<u>n/a</u>	Bench Lamp:	<u>53.6</u>
O3 Lamp:	<u>n/a</u>	O3 Lamp:	<u>67.7</u>
Pressure:	<u>n/a</u>	Pressure:	<u>680.0</u>
Cell A lpm:	<u>n/a</u>	Cell A lpm:	<u>0.730</u>
Cell B lpm:	<u>n/a</u>	Cell B lpm:	<u>0.764</u>
O3 ppb:	<u>n/a</u>	O3 ppb:	<u>1.5</u>
Cell A ppb:	<u>n/a</u>	Cell A ppb:	<u>1.5</u>
Cell B ppb:	<u>n/a</u>	Cell B ppb:	<u>0.2</u>
Cell A int:	<u>n/a</u>	Cell A int:	<u>86877</u>
Expected Value:	<u>n/a</u>	Expected Value:	<u>268.0</u>

Comments:
 The analyzer sample inlet filter was changed. The analyzer cooling fan filter(s) were cleaned.

A sample pump was rebuilt, and absorption cells A and B cleaned. Zero Air scrubber container was changed due to leak.



— O3[ppb]

PARTICULATE MATTER



R & P 1405F TEOM PM 2.5 Analyzer Calibration

Date: March 1, 2017
 Company: LICA
 Station Name/Location: St. Lina
 Previous Audit Date: February 3, 2017
 Parameter: PM 2.5

Performed By/Reviewer: Alex Yakupov | Trina Whatsitt
 Start Time (mst): 11:05
 End Time (mst): 13:49
 Calibration Purpose: Bi-monthly #1
 Weather Conditions: Mainly sunny

1400A Information and Status:

ID# or Serial Number: 1405A208301003 As Found Filter Loading %: 35.94
 Ko Factor: 13125 As Left Filter Loading %: 17.77
 Ambient Temperature °C: 0.919 As Found Noise: 0.004
 Ambient Pressure atm: -15.88 As Left Noise: 0.000
 Main Flow Reading lpm: 3.00 Pump Vacuum: 0.32
 Aux Flow Reading lpm: 13.67 Warnings: None

Reference Standards:

	Flow:	Pressure:	Temperature:
Make:	<u>Dwyer</u>	<u>Fisher</u>	<u>FLUKE</u>
Model:	<u>475 Mark III</u>	<u>FB1291</u>	<u>1551A Ex STIK</u>
Serial Number:	<u>#3</u>	<u>#05544</u>	<u>4295</u>
Calibration Date:	<u>January 1, 2017</u>	<u>December 5, 2016</u>	<u>November 15, 2016</u>

As found leak check:

		Base	Zero	Reference	Zero
PM 2.5 Flow	actual	0.00	0.03	0.00	0.03
	limit	0.15	0.15	0.15	0.15
Bypass Flow	actual	0.01	-0.68	0.01	-0.68
	limit	0.60	0.60	0.60	0.60

As left leak check (same as above if as found passes):

		Base	Zero	Reference	Zero
PM 2.5 Flow	actual	0.00	0.03	0.00	0.03
	limit	0.15	0.15	0.15	0.15
Bypass Flow	actual	0.01	-0.68	0.01	-0.68
	limit	0.60	0.60	0.60	0.60

As found temperature and pressure:

tolerance +/- 2.0°C	tolerance +/- 0.01 atm
1405F temperature °C: <u>-15.9</u>	1405F pressure atm: <u>0.919</u>
reference temperature °C: <u>-16.0</u>	reference pressure: <u>0.915</u>
difference °C: <u>-0.1</u>	difference: <u>0.004</u>

As left temperature and pressure (same as above if as found adequate):

tolerance +/- 2.0°C	tolerance +/- 0.01 atm
1405F temperature °C: <u>-16.0</u>	1405F pressure atm: <u>0.915</u>
reference temperature °C: <u>-16.0</u>	reference pressure: <u>0.915</u>
difference °C: <u>0.0</u>	difference: <u>0.000</u>

As found flows:

main flow tolerance 3.00 lpm +/- 0.20 lpm	total/aux flow tolerance 16.67/13.67 lpm +/- 1.00 lpm +/- 7%
1405F main flow lpm: <u>3.00</u>	1400A total/aux flow lpm: <u>16.67</u>
reference main flow lpm: <u>3.02</u>	reference total/aux flow lpm: <u>16.23</u>
difference lpm: <u>0.02</u>	difference lpm: <u>-0.44</u>

As left flows (same as above if as found adequate):

main flow tolerance 3.00 lpm +/- 0.20 lpm	total/aux flow tolerance 16.67/13.67 lpm +/- 1.00 lpm +/- 7%
1405F main flow lpm: <u>3.00</u>	1400A total/aux flow lpm: <u>16.67</u>
reference main flow lpm: <u>3.02</u>	reference total/aux flow lpm: <u>16.23</u>
difference lpm: <u>0.02</u>	difference lpm: <u>-0.44</u>

K_o Audit:

Last K_o audit date: February 3, 2017
 1405F K_o factor: 13125
 Measured K_o factor: 13126.5000
 % difference: 0.01

Comments:

The TEOM sample filter was changed. The TEOM intake head and associated sharp cut components were cleaned.
 The 47 mm FDMS filter was changed.



R & P 1405F TEOM PM 2.5 Analyzer Calibration

Date: March 28, 2017
 Company: LICA
 Station Name/Location: St. Lina
 Previous Audit Date: March 1, 2017
 Parameter: PM 2.5

Performed By/Reviewer: Alex Yakupov | Trina Whitsitt
 Start Time (mst): 9:58
 End Time (mst): 11:04
 Calibration Purpose: Bi-monthly #2
 Weather Conditions: Fog

1400A Information and Status:

ID# or Serial Number: 1405A208301003 As Found Filter Loading %: 22.67
 Ko Factor: 13125 As Left Filter Loading %: 21.44
 Ambient Temperature °C: 1.11 As Found Noise: 0.003
 Ambient Pressure atm: 0.912 As Left Noise: 0.000
 Main Flow Reading lpm: 3.00 Pump Vacuum: 0.29
 Aux Flow Reading lpm: 13.67 Warnings: None

Reference Standards:

	Flow:	Pressure:	Temperature:
Make:	<u>Dwyer</u>	<u>Fisher</u>	<u>FLUKE</u>
Model:	<u>475 Mark III</u>	<u>FB1291</u>	<u>1551A Ex STIK</u>
Serial Number:	<u>#3</u>	<u>#05544</u>	<u>4295</u>
Calibration Date:	<u>January 1, 2017</u>	<u>December 5, 2016</u>	<u>November 15, 2016</u>

As found leak check:

		Base	Zero	Reference	Zero
PM 2.5 Flow	actual	0.00	0.03	0.00	0.03
	limit	0.15	0.15	0.15	0.15
Bypass Flow	actual	0.00	-0.72	0.00	-0.72
	limit	0.60	0.60	0.60	0.60

As left leak check (same as above if as found passes):

		Base	Zero	Reference	Zero
PM 2.5 Flow	actual	0.00	0.03	0.00	0.03
	limit	0.15	0.15	0.15	0.15
Bypass Flow	actual	0.00	-0.72	0.00	-0.72
	limit	0.60	0.60	0.60	0.60

As found temperature and pressure:

tolerance +/- 2.0°C	tolerance +/- 0.01 atm
1405F temperature °C: <u>1.1</u>	1405F pressure atm: <u>0.912</u>
reference temperature °C: <u>1.3</u>	reference pressure: <u>0.916</u>
difference °C: <u>0.2</u>	difference: <u>-0.004</u>

As left temperature and pressure (same as above if as found adequate):

tolerance +/- 2.0°C	tolerance +/- 0.01 atm
1405F temperature °C: <u>1.3</u>	1405F pressure atm: <u>0.916</u>
reference temperature °C: <u>1.3</u>	reference pressure: <u>0.916</u>
difference °C: <u>0.0</u>	difference: <u>0.000</u>

As found flows:

main flow tolerance 3.00 lpm +/- 0.20 lpm	total/aux flow tolerance 16.67/13.67 lpm +/- 1.00 lpm +/- 7%
1405F main flow lpm: <u>3.00</u>	1400A total/aux flow lpm: <u>16.67</u>
reference main flow lpm: <u>2.94</u>	reference total/aux flow lpm: <u>16.05</u>
difference lpm: <u>-0.06</u>	difference lpm: <u>-0.62</u>

As left flows (same as above if as found adequate):

main flow tolerance 3.00 lpm +/- 0.20 lpm	total/aux flow tolerance 16.67/13.67 lpm +/- 1.00 lpm +/- 7%
1405F main flow lpm: <u>3.00</u>	1400A total/aux flow lpm: <u>16.67</u>
reference main flow lpm: <u>2.94</u>	reference total/aux flow lpm: <u>16.05</u>
difference lpm: <u>-0.06</u>	difference lpm: <u>-0.62</u>

K_o Audit:

Last K_o audit date: February 3, 2017
 1405F K_o factor: 13125
 Measured K_o factor: 13126.5000
 % difference: 0.01

Comments:

The TEOM intake head and associated sharp cut components were cleaned.

The 47 mm FDMS filter was changed.

WIND SYSTEM

CALIBRATORS

Company Maxxam Operator: Christopher Wesson

Calibrator:		Flow Measurement Device:	
Make/Model	<u>Sabio 2010</u>	Make/Model	<u>N/A</u>
Serial Number	<u>17200415</u>	Serial Number	<u>N/A</u>
Last Verification Date	<u>May 2015</u>	Temperature (°C)	<u>N/A</u>
NO Cylinder S/N	<u>LL42475</u>	Barometric Pressure	<u>N/A</u>
NO/NOx Concentration	<u>48.5/48.5</u>		

Dilution Flow (sccm)			
Pt. #1	<u>5000</u>	Pt. #2	<u>5000</u>
Pt. #3	<u>5000</u>		
Gas Flow (sccm)			
Pt. #1	<u>80</u>	Pt. #2	<u>40</u>
Pt. #3	<u>20</u>		

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
5029	0.0	0.000	0.000	0.000	0.000	0.000	Limit ± 10%	
5030	80.6	0.777	0.777	0.805	-0.005	0.800	4%	3%
5025	39.4	0.380	0.380	0.394	-0.002	0.392	4%	3%
5028	19.8	0.191	0.191	0.198	-0.001	0.197	4%	3%
Absolute Average Percent Difference							3.65%	3.09%

LINEAR REGRESSION ANALYSIS				$y=mx+b$ (where x=calculated concentration, y=indicated concentration)			
<u>NO</u>		<u>LIMITS</u>		<u>NOx</u>			
Correlation=	1.0000	≥ 0.990		Correlation=	1.0000		
m (Slope)=	1.0360	0.90-1.10		m (Slope)=	1.0295		
b (Intercept % of FS)=	0.0110	± 3% F.S.		b (Intercept % of FS)=	0.0293		

Flow	O ₂ Conc	NO Decrease	NO	NO ₂	NOx	% Diff. Vs Audit gas		
5030	Lamp C.	0.000	0.804	-0.004	0.800	NO ₂	% Diff. Limit	
5030	1.388	0.495	0.309	0.491	0.800	0%	± 10%	
5030	0.745	0.241	0.563	0.239	0.802	1%	± 10%	
5030	0.367	0.091	0.713	0.089	0.801	2%	± 10%	
Absolute Average Percent Difference							1%	± 10%

LINEAR REGRESSION ANALYSIS				$y=mx+b$ (where x=calculated concentration, y=indicated concentration)			
<u>NO₂</u>		<u>LIMITS</u>					
Correlation=	1.0000	≥ 0.995					
m (Slope)=	0.9988	0.90-1.10					
b (Intercept % of FS)=	-0.2760	± 3% F.S.					

AENV Standards Audit Calibrator		NO _x Analyzer	
Make/Model	<u>Teco 146i</u>	Make/Model	<u>Teco 42i</u>
Serial/AMU Number	<u>AMU 1809</u>	Serial/AMU Number	<u>AMU 1868</u>
		Last Calibration Date	<u>May 18, 2016</u>
		Full Scale (ppm)	<u>1.0</u>

COMMENTS: Contains 50.3 ppm SO₂. Flows not measured as per Chapter 7, Section 5 of AMD.

Auditor: AI Clark
Operator Signature: *AI Clark*

Date: May 18, 2016
Location: McIntyre Center Edmonton

Company Maxxam/SIA **Operator:** Chris

Calibrator:		Flow Measurement Device:	
Make/Model	<u>API 700</u>	Make/Model	<u>Definer 530</u>
Serial Number	<u>627</u>	Serial Number	<u>H-148944, L-152019</u>
Last Verification Date	<u>February 3, 2016</u>	Temperature (°C)	<u>23.5</u>
NO Cylinder S/N	<u>EY0000597</u>	Barometric Pressure	<u>707.1 mmHg</u>
NO [PPM]	<u>49.0</u>	NOx [PPM]	<u>49.0</u>
Expiry Date	<u>December 8, 2019</u>		

Dilution Flow (sccm)		
Pt. #1	<u>4892</u>	Pt. #3 <u>4951</u>
Pt. #2	<u>4975</u>	
Gas Flow (sccm)		
Pt. #1	<u>79.7</u>	Pt. #3 <u>19.4</u>
Pt. #2	<u>38.8</u>	

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
	0.0	0.0000	0.0000	0.0000	-0.0004	-0.0004	Limit ± 10%	
4972	79.7	0.7855	0.7855	0.7883	0.0004	0.7887	0.4%	0.5%
4936	38.8	0.3822	0.3822	0.3816	0.0005	0.3822	-0.2%	0.1%
4970	19.4	0.1913	0.1913	0.1902	0.0006	0.1913	-0.6%	0.2%
Absolute Average Percent Difference							0.1%	0.3%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO	LIMITS	NOx
Correlation= 1.0000	≥ 0.990	Correlation= 1.0000
m (Slope)= 1.0041	0.90-1.10	m (Slope)= 1.0046
b (Intercept % of FS)= -0.1118	± 3% F.S.	b (Intercept % of FS)= -0.0871

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4972	0	0.0000	0.7867	0.0014	0.7881	NO ₂	% Diff, Limit
4972	500	0.5127	0.2740	0.5104	0.7849	-0.7%	± 10%
4972	275	0.2863	0.5004	0.2860	0.7865	-0.6%	± 10%
4972	90	0.0940	0.6927	0.0954	0.7880	0.0%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO₂	LIMITS
Correlation= 1.0000	≥ 0.995
m (Slope)= 0.9924	0.90-1.10
b (Intercept % of FS)= 0.1755	± 3% F.S.

AENV Standards	NO_x Analyzer
Audit Calibrator	Make/Model <u>Thermo 42i</u>
Make/Model <u>Thermo 146i</u>	Serial/AMU Number <u>AMU 1868</u>
Serial/AMU Number <u>AMU1809</u>	Last Calibration Date <u>January 25, 2017</u>
SRM Gas Cylinder No. <u>CAL018140</u>	Full Scale (ppm) <u>1.0</u>
Cylinder Conc. (ppm) <u>48.79</u>	Cylinder Gas Expiry Date <u>March 25, 2019</u>

COMMENTS: _____

Auditor: Shea Beaton Date: January 27, 2017

Operator Signature: [Signature] Location: McIntyre Center Edmonton

Company Maxxam Operator: Mike

Calibrator:		Flow Measurement Device:	
Make/Model	<u>EnviroNics 2000</u>	Make/Model	<u>Bios Defender 530</u>
Serial Number	<u>1991</u>	Serial Number	<u>Hi148944 Lo 152019</u>
Last Verification Date	<u>March 31, 2016</u>	Temperature (°C)	<u>24.5</u>
NO Cylinder S/N	<u>EY0000597</u>	Barometric Pressure	<u>699</u>
NO [PPM]	<u>49.0</u>	NOx [PPM]	<u>49.0</u>
Expiry Date	<u>December 8, 2019</u>		

Dilution Flow (sccm)		
Pt. #1 <u>4902</u>	Pt. #2 <u>4935</u>	Pt. #3 <u>4957</u>
Gas Flow (sccm)		
Pt. #1 <u>79.3</u>	Pt. #2 <u>38.7</u>	Pt. #3 <u>19.4</u>

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
4976	0.0	0.0000	0.0000	0.0001	0.0000	0.0001	Limit ± 10%	
4981	79.3	0.7801	0.7801	0.7898	0.0000	0.7898	1%	1%
4972	38.7	0.3814	0.3814	0.3841	0.0002	0.3843	1%	1%
4976	19.4	0.1910	0.1910	0.1913	0.0003	0.1916	0%	0%
Absolute Average Percent Difference							1%	1%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO	LIMITS	NOx
Correlation= 1.0000	≥ 0.990	Correlation= 1.0000
m (Slope)= 1.0130	0.90-1.10	m (Slope)= 1.0129
b (Intercept % of FS)= -0.1190	± 3% F.S.	b (Intercept % of FS)= -0.1029

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4981	0.000	0.0000	0.7925	-0.0001	0.7924	NO ₂	% Diff. Limit
4981	0.400	0.5347	0.2578	0.5279	0.7857	-1%	± 10%
4981	0.200	0.2490	0.5435	0.2478	0.7913	0%	± 10%
4981	0.090	0.1090	0.6835	0.1095	0.7927	1%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO₂	LIMITS
Correlation= 1.0000	≥ 0.995
m (Slope)= 0.9864	0.90-1.10
b (Intercept % of FS)= 0.1136	± 3% F.S.

AENV Standards		NO_x Analyzer	
Audit Calibrator		Make/Model	<u>Thermo 42i</u>
Make/Model	<u>Thermo 146i</u>	Serial/AMU Number	<u>1868</u>
Serial/AMU Number	<u>1809</u>	Last Calibration Date	<u>March 15, 2017</u>
SRM Gas Cylinder No.	<u>CAL018140</u>	Full Scale (ppm)	<u>1.0</u>
Cylinder Conc. (ppm)	<u>48.79</u>	Cylinder Gas Expiry Date	<u>March 28, 2019</u>

COMMENTS: Gas has ~50ppm SO2

Auditor: Shea Beaton
Operator Signature: [Signature]

Date: March 16, 2017
Location: McIntyre Center Edmonton

Company <u>Maxxam</u>		Operator: <u>Mike</u>	
Calibrator:		Flow Measurement Device:	
Make/Model	<u>Sabio 2010D</u>	Make/Model	<u>Bios Defender 530</u>
Serial Number	<u>11900613</u>	Serial Number	<u>HI148944 Lo 152019</u>
Last Verification Date	<u>March 31, 2016</u>	Temperature (°C)	<u>23.9</u>
NO Cylinder S/N	<u>EY0000769</u>	Barometric Pressure	<u>698mmHg</u>
NO [PPM]	<u>51.1</u>	NOx [PPM]	<u>51.2</u>
Expiry Date	<u>December 8, 2019</u>		

Dilution Flow (sccm)		
Pt. #1 <u>4879</u>	Pt. #2 <u>4932</u>	Pt. #3 <u>4950</u>
Gas Flow (sccm)		
Pt. #1 <u>74.5</u>	Pt. #2 <u>36.4</u>	Pt. #3 <u>18.2</u>

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
4965	0.0	0.0000	0.0000	0.0001	0.0000	0.0001	Limit ± 10%	
4954	74.5	0.7685	0.7700	0.7915	0.0008	0.7923	3%	3%
4968	36.4	0.3744	0.3751	0.3832	0.0006	0.3838	2%	2%
4968	18.2	0.1872	0.1876	0.1916	0.0002	0.1918	2%	2%
Absolute Average Percent Difference							3%	2%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO	LIMITS	NOx
Correlation= 1.0000	≥ 0.990	Correlation= 1.0000
m (Slope)= 1.0301	0.90-1.10	m (Slope)= 1.0291
b (Intercept % of FS)= -0.0919	± 3% F.S.	b (Intercept % of FS)= -0.0881


Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4954	0.000	0.0000	0.7949	0.0005	0.7954	NO ₂	% Diff. Limit
4954	0.510	0.5104	0.2845	0.5072	0.7917	-1%	± 10%
4954	0.250	0.2516	0.5433	0.2514	0.7944	0%	± 10%
4954	0.100	0.1085	0.6864	0.1087	0.7951	0%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO₂	LIMITS
Correlation= 1.0000	≥ 0.995
m (Slope)= 0.9926	0.90-1.10
b (Intercept % of FS)= 0.0925	± 3% F.S.

AENV Standards	NO_x Analyzer
Audit Calibrator	Make/Model <u>Thermo 42i</u>
Make/Model <u>Thermo 146i</u>	Serial/AMU Number <u>1868</u>
Serial/AMU Number <u>1809</u>	Last Calibration Date <u>March 15, 2017</u>
SRM Gas Cylinder No. <u>CAL018140</u>	Full Scale (ppm) <u>1.0</u>
Cylinder Conc. (ppm) <u>48.79</u>	Cylinder Gas Expiry Date <u>March 28, 2019</u>

COMMENTS: Gas has ~50ppm SO2

Auditor: Shea Beaton
Operator Signature: 

Date: March 16, 2017
Location: McIntyre Center Edmonton

Company Maxxam Operator: Christopher Wesson

Calibrator:		Flow Measurement Device:	
Make/Model	<u>Sabio 2010</u>	Make/Model	<u>N/A</u>
Serial Number	<u>042531101(0911)</u>	Serial Number	<u>N/A</u>
Last Verification Date	<u>December 2014</u>	Temperature (°C)	<u>N/A</u>
NO Cylinder S/N	<u>L142475</u>	Barometric Pressure	<u>N/A</u>
NO/NOX Concentration	<u></u>		

Dilution Flow (sccm)			
Pt. #1	<u>5000</u>	Pt. #2	<u>5000</u>
Pt. #3	<u>5000</u>		
Gas Flow (sccm)			
Pt. #1	<u>80</u>	Pt. #2	<u>40</u>
Pt. #3	<u>20</u>		

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
5001	0.0	0.000	0.000	0.000	0.000	0.000	Limit ± 10%	
4999	80.4	0.780	0.780	0.802	-0.003	0.799	3%	2%
5000	39.2	0.380	0.380	0.391	-0.001	0.390	3%	3%
4999	19.6	0.190	0.190	0.195	-0.001	0.194	3%	2%
Absolute Average Percent Difference							3%	2%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO	LIMITS	NOx
Correlation= 1.0000	≥ 0.990	Correlation= 1.0000
m (Slope)= 1.0284	0.90-1.10	m (Slope)= 1.0247
b (Intercept % of FS)= -0.0085	± 3% F.S.	b (Intercept % of FS)= -0.0097


Flow	O ₂ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4999	0.000	0.000	0.802	-0.001	0.801	NO ₂	% Diff. Limit
4999		#VALUE!	Not completed O3 gen system issue.			#VALUE!	± 10%
4999		0.802				-100%	± 10%
4999		0.802				-100%	± 10%
Absolute Average Percent Difference						#VALUE!	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO ₂	LIMITS
Correlation= #VALUE!	≥ 0.995
m (Slope)= #VALUE!	0.90-1.10
b (Intercept % of FS)= #VALUE!	± 3% F.S.

AENV Standards Audit Calibrator		NO _x Analyzer	
Make/Model	<u>Teco 146i</u>	Make/Model	<u>Teco 42i</u>
Serial/AMU Number	<u>AMU 1809</u>	Serial/AMU Number	<u>AMU 1868</u>
		Last Calibration Date	<u>January 19, 2016</u>
		Full Scale (ppm)	<u>1.0</u>

COMMENTS: 50.3 ppm of SO2 in cylinder.

Auditor: Al Clark Date: January 19, 2016
 Operator Signature:  Location: McIntyre Center Edmonton

CALIBRATION GASES



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2015-342CGA

Company: Maxxam **Operator's Name:** Limin Li
Cylinder #: BLM002756T **Concentration PPM:** 49.9 **Tolerance(%)** 2 **Certified By:** Air Liquide

Reference Calibrator and Gas:

Make/Model: R&R MFC 201
 Serial Number: AMU 1690
 Last Verification Date: March 31, 2015
 Gas Type: SO2 Conc. 98.57
 Cylinder Number: CAL016720

Flow Measurement Device:

Make/Model: Bios DC2
 Serial Number: AMU 1659
 Temp. °C: 22.5 C
 B.P. 690 mmhg

Reference Analyzer:

Make/Model: Teco 43C Serial/AMU Number: 1623
 Instrument Settings: Zero: 7.9 Span: 1.028 Range: 1.0
 Last Calibration: Date: Mar 31/15 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.000	0.0000	0.000	0.000
4976	82.6	0.821	0.01660	60.242	49.5
4993	41.0	0.410	0.00821	121.780	49.9
4977	20.2	0.202	0.00406	246.386	49.8
Average Cylinder Concentration:					49.7

Previous Stated Concentration PPM: 49.9

Percent variance from Stated: 0.4

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration _____
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark
 Operator Signature: *Al Clark*

Date: March 31, 2015
 Location: McIntyre Center Edmonton



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2016-335CGA

Company: Maxxam **Operator's Name:** Russell Kirchner
Cylinder #: LL104222 **Concentration PPM:** 50.6 **Tolerance(%)** 1 **Certified By:** Praxair
Expiry Date: July 2019

Reference Calibrator and Gas:	Flow Measurement Device:
Make/Model: <u>R&R MFC 201</u>	Make/Model: <u>Bios DC2</u>
Serial Number: <u>AMU 1690</u>	Serial Number: <u>AMY 1659</u>
Last Verification Date: <u>October 19, 2016</u>	Temp. °C: <u>24.5 C</u>
Gas Type: <u>SO2</u> Conc. <u>98.07</u>	B.P. <u>706 mmhg</u>
Cylinder Number: <u>CA:016625</u>	
Expiry Date: <u>January 2019</u>	

Reference Analyzer:
Make/Model: Teco 43C **Serial/AMU Number:** 1623
Instrument Settings: **Zero:** 9.2 **Span:** 1.024 **Range:** 1.0
Last Calibration: **Date:** Oct 19/16 **C.F.** 1.000 **Done By:** Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.000	0.0000	0.0000	0.000
4935	82.0	0.830	0.01662	60.183	50.0
4968	40.8	0.412	0.00821	121.765	50.2
4955	20.2	0.203	0.00408	245.297	49.8
Average Cylinder Concentration:					50.0

Previous Stated Concentration PPM: 50.6

Percent variance from Stated: 1

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 < =5% Outside Manufacturer Tolerance. Use manufacturers concentration _____
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark
Operator Signature: *Al Clark*

Date: October 19, 2016
Location: McIntyre Center Edmonton



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2015-338CGA

Company: Maxxam **Operator's Name:** Limin Li
Cylinder #: BLM002508 **Concentration PPM:** 10.2 **Tolerance(%)** 2 **Certified By:** Air Liquide

Reference Calibrator and Gas:

Make/Model: R&R MFC 201
 Serial Number: AMU1690
 Last Verification Date: March 31, 2015
 Gas Type: H2S Conc. 20.43
 Cylinder Number: CAL015106

Flow Measurement Device:

Make/Model: Bios DC2
 Serial Number: AMU 1659
 Temp. °C: 23.0 C
 B.P. 689 mmhg

Reference Analyzer:

Make/Model: Teco 450i Serial/AMU Number: 1980
 Instrument Settings: Zero: 14.5 Span: 1.035 Range: 0.1
 Last Calibration: Date: Mar 31/15 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.0000	0.0000	132.984	9.6
5080	38.2	0.0725	0.00752	132.984	9.6
5078	17.9	0.0340	0.00353	283.687	9.6
5066	9.1	0.0170	0.00180	556.703	9.5
Average Cylinder Concentration:					9.6

Previous Stated Concentration PPM: 10.2

Percent variance from Stated: 6.0

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark
 Operator Signature: *Al Clark*

Date: March 31, 2015
 Location: McIntyre Center Edmonton



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2016-334CGA

Company: Maxxam **Operator's Name:** Russell Kirchner
Cylinder #: EY0000654 **Concentration PPM:** 10.2 **Tolerance(%)** 2 **Certified By:** Praxair
Expiry Date: June 2019

Reference Calibrator and Gas:	Flow Measurement Device:
Make/Model: <u>R&R MFC 201</u>	Make/Model: <u>Bios DC2</u>
Serial Number: <u>AMU 1690</u>	Serial Number: <u>AMU 1659</u>
Last Verification Date: <u>October 19, 2016</u>	Temp. °C: <u>24.0 C</u>
Gas Type: <u>H2S</u> Conc. <u>20.43</u>	B.P. <u>706 mmhg</u>
Cylinder Number: <u>CAL015584</u>	
Expiry Date: <u>January 2019</u>	

Reference Analyzer:
 Make/Model: Teco 450i Serial/AMU Number: 1980
 Instrument Settings: Zero: 16.6 Span: 1.231 Range: 0.1
 Last Calibration: Date: Oct 19/16 C.F. 1.000 Done By: Al Clark

Calibrator Flows (scm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.0000	0.00752	132.895	10.2
5050	38.0	0.0764	0.00752	132.895	10.2
5050	17.8	0.0355	0.00352	283.708	10.1
5023	9.1	0.0182	0.00181	551.978	10.0
Average Cylinder Concentration:					10.1

Previous Stated Concentration PPM: 10.2
 Percent variance from Stated: 1

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration _____
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark Date: October 19, 2016
 Operator Signature: *Al Clark* Location: McIntyre Center Edmonton



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2014-250CGA

Company: Maxxam Operator's Name: Limin Li
Cylinder #: LL74267 Concentration PPM: 9.88 Tolerance(%): 2 Certified By: Air Liquide

Reference Calibrator and Gas:

Make/Model: R&R MFC 201
Serial Number: AMU 1690
Last Verification Date: December 15, 2014
Gas Type: H2S Conc. 20.43
Cylinder Number: CAL015106

Flow Measurement Device:

Make/Model: Bios DC2
Serial Number: AMU 1659
Temp. °C: 23.0 C
B.P.: 702 mmhg

Reference Analyzer:

Make/Model: Teco 45C Serial/AMU Number: 1624
Instrument Settings: Zero: 6.4 Span: 1.160 Range: 0.1
Last Calibration: Date: Dec15/14 C.F.: 1.000 Done By: AI Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.0000	0.00755	282.889	9.67
5099	38.5	0.0731	0.00755	132.442	9.68
5092	18.0	0.0342	0.00353	282.889	9.67
5066	9.2	0.0173	0.00182	550.652	9.53
Average Cylinder Concentration:					9.63

Previous Stated Concentration PPM: 9.88

Percent variance from Stated: 2.6

Meets Manufacturer Tolerance. Use manufacturers stated concentration COMMENTS: _____
 < =5% Outside Manufacturer Tolerance. Use manufacturers concentration
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: AI Clark
Operator Signature: *AI Clark*

Date: December 16, 2014
Location: McIntyre Center Edmonton



Praxair
 5700 South Alameda Street
 Los Angeles, CA 90058
 Tel: (323) 585-2154 Fax: (714) 542-6689
 PGVPID: F22014

DocNumber: 000068924

CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

Customer & Order Information:

MAXXAM ANALYTICS INC *NA*
 9372 49TH ST
 EDMONTON AB T6B 2L

Praxair Order Number: 21137117
 Customer P. O. Number: 35-55963
 Customer Reference Number:

Fill Date: 7/1/2014
 Part Number: NI ME600P2E-AQ
 Lot Number: 109418203
 Cylinder Style & Outlet: AQ CGA 350
 Cylinder Pressure & Volume: 2200 psig 78 cu. ft.

Certified Concentration:

Expiration Date:	7/7/2022	NIST Traceable
Cylinder Number:	LL83638	Analytical Uncertainty:
582 ppm	METHANE	± 1.5 %
203 ppm	PROPANE	± 0.9 %
Balance	NITROGEN	

Certification Information: Certification Date: 7/7/2014 Term: 96 Months Expiration Date: 7/7/2022

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1. Do Not Use this Standard if Pressure is less than 100 PSIG.

Analytical Data:

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: METHANE

Requested Concentration: 600 ppm
 Certified Concentration: 582 ppm
 Instrument Used: MKS Multigas 2031 FTIR
 Analytical Method: Fourier Transform Infrared
 Last Multipoint Calibration: 6/24/2014

Reference Standard Type: GMIS
 Ref. Std. Cylinder #: CC139480
 Ref. Std. Conc: 246 ppm
 Ref. Std. Traceable to SRM #: 2751
 SRM Sample #: 212-09-AL
 SRM Cylinder #: SX-20000

First Analysis Data:		Date: 7/7/2014	
Z: 0	R: 249.5	C: 589.4	Conc: 581.21
R: 249.5	Z: 0	C: 589	Conc: 580.82
Z: 0	C: 592	R: 249.4	Conc: 583.77
UOM: ppm	Mean Test Assay:	581.93 ppm	

Second Analysis Data:		Date:	
Z: 0	R: 0	C: 0	Conc: 0
R: 0	Z: 0	C: 0	Conc: 0
Z: 0	C: 0	R: 0	Conc: 0
UOM: ppm	Mean Test Assay:	0 ppm	

2. Component: PROPANE

Requested Concentration: 200 ppm
 Certified Concentration: 203 ppm
 Instrument Used: MKS Multigas 2031 FTIR
 Analytical Method: Fourier Transform Infrared
 Last Multipoint Calibration: 6/24/2014

Reference Standard Type: GMIS
 Ref. Std. Cylinder #: CC 163442
 Ref. Std. Conc: 265.8 ppm
 Ref. Std. Traceable to SRM #: vs 2644a
 SRM Sample #: 101-C-45
 SRM Cylinder #: XF003829B

First Analysis Data:		Date: 7/7/2014	
Z: 0	R: 273.6	C: 208.4	Conc: 202.43
R: 273.7	Z: 0	C: 208.6	Conc: 202.63
Z: 0	C: 208.5	R: 273.6	Conc: 202.53
UOM: ppm	Mean Test Assay:	202.53 ppm	

Second Analysis Data:		Date:	
Z: 0	R: 0	C: 0	Conc: 0
R: 0	Z: 0	C: 0	Conc: 0
Z: 0	C: 0	R: 0	Conc: 0
UOM: ppm	Mean Test Assay:	0 ppm	

Analyzed by:

Jack Fu

Certified by:

Ying Yu

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.



Calibration Gas Audit

NO Cylinder Gas

File No. 2015-343CGA

Company: Maxxam **Operators name:** Limin Li
Cylinder #: BLM002756T **Conc (PPM)** 50.7/50.7 **Tolerance (%)** 2 **Certified By:** Air Liquide

Reference Calibrator and Gas:

Make/Model Teco 146i
Serial Number AMU 1809
Last Verification Date March 31, 2015
Gas Type NO **Conc.** 48.79
Cylinder Number CAL018024

Flow Measurement Device:

Make/Model Bios DC2
Serial Number AMU 1659
Temp. °C 22.5 C
B.P. 690 mmhg

Reference Analyzer:

Make/Model Teco 42i **Serial/AMU Number:** 1868
Instrument Settings **Zero:** 4.2 **Span:** 1.008 **Range:** 1.0
Last Calibration: **Date:** Mar 31/15 **C.F.** 1.000 **Done By:** Al Clark

Calibrator Flows (scm)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	NO	NOX			NO	NOX
5000	0.0	0.000	0.000	0.01660	60.242	50.7	49.5
4976	82.6	0.842	0.822	0.01660	60.242	50.7	49.5
4993	41.0	0.420	0.410	0.00821	121.780	51.1	49.9
4977	20.2	0.208	0.205	0.00406	246.386	51.2	50.5
Average Cylinder Concentration:						51.0	50.0

<u>NO</u>	<u>NOx</u>
Previous Stated Concentration PPM: <u>50.7</u>	<u>50.7</u>
Percent variance from Stated: <u>0.7</u>	<u>1.4</u>

Cylinder gas tolerances based on NO only

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 < =5% Outside Manufacturer Tolerance. Use manufacturers concentration Contains 49.9 ppm SO2 in cylinder
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark **Date:** March 31, 2015
Operator Signature: *Al Clark* **Location:** McIntyre Center Edmonton



Calibration Gas Audit

NO Cylinder Gas

File No. 2016-336CGA

Company: Maxxam **Operators name:** Russell Kirchner

Cylinder #: LL104222 Conc (PPM) 50.7/50.9 Tolerance (%) 1 Certified By: Praxair

Expiry Date: July 2019

Reference Calibrator and Gas:	Flow Measurement Device:
Make/Model: <u>Teco 146i</u>	Make/Model: <u>Bios DC2</u>
Serial Number: <u>AMU 1809</u>	Serial Number: <u>AMU 1659</u>
Last Verification Date: <u>October 19, 2019</u>	Temp. °C: <u>24.5 C</u>
Gas Type: <u>NO</u> Conc. <u>48.79</u>	B.P.: <u>706 mmhg</u>
Cylinder Number: <u>CAL018188</u>	
Expiry Date: <u>March 2019</u>	

Reference Analyzer:

Make/Model: Teco 42i Serial/AMU Number: 1868

Instrument Settings: Zero: 4.4 Span: 1.080 Range: 1.0

Last Calibration: Date: Oct 18/16 C.F.: 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	NO	NOX			NO	NOX
5000	0.0	0.000	0.000				
4935	82.0	0.838	0.837	0.017	60.183	50.4	50.4
4968	40.8	0.417	0.417	0.008	121.765	50.8	50.8
4955	20.2	0.207	0.207	0.004	245.297	50.8	50.8
Average Cylinder Concentration:						50.7	50.6

<u>NO</u>	<u>NOx</u>
Previous Stated Concentration PPM: <u>50.7</u>	<u>50.9</u>
Percent variance from Stated: <u>0</u>	<u>1</u>

Cylinder gas tolerances based on NO only

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:**

< =5% Outside Manufacturer Tolerance. Use manufacturers concentration Contains 50.6 ppm SO2.

> 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark Date: October 19, 2016

Operator Signature: *Al Clark* Location: McIntyre Center Edmonton



Calibration Gas Audit

NO Cylinder Gas

File No. 2015-343CGA

Company: Maxxam **Operators name:** Limin Li
Cylinder #: BLM002756T **Conc (PPM)** 50.7/50.7 **Tolerance (%)** 2 **Certified By:** Air Liquide

Reference Calibrator and Gas:				Flow Measurement Device:	
Make/Model	<u>Teco 146i</u>			Make/Model	<u>Bios DC2</u>
Serial Number	<u>AMU 1809</u>			Serial Number	<u>AMU 1659</u>
Last Verification Date	<u>March 31, 2015</u>			Temp. °C	<u>22.5 C</u>
Gas Type	<u>NO</u>	Conc.	<u>48.79</u>	B.P.	<u>690 mmhg</u>
Cylinder Number	<u>CAL018024</u>				

Reference Analyzer:
Make/Model Teco 42i **Serial/AMU Number:** 1868
Instrument Settings **Zero:** 4.2 **Span:** 1.008 **Range:** 1.0
Last Calibration: **Date:** Mar 31/15 **C.F.** 1.000 **Done By:** Al Clark

Calibrator Flows (scm)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	NO	NOX			NO	NOX
5000	0.0	0.000	0.000	0.01660	60.242	50.7	49.5
4976	82.6	0.842	0.822	0.01660	60.242	50.7	49.5
4993	41.0	0.420	0.410	0.00821	121.780	51.1	49.9
4977	20.2	0.208	0.205	0.00406	246.386	51.2	50.5
Average Cylinder Concentration:						51.0	50.0

<u>NO</u>	<u>NOx</u>
Previous Stated Concentration PPM: <u>50.7</u>	<u>50.7</u>
Percent variance from Stated: <u>0.7</u>	<u>1.4</u>

Cylinder gas tolerances based on NO only

- Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
- < =5% Outside Manufacturer Tolerance. Use manufacturers concentration Contains 49.9 ppm SO2 in cylinder
- > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark **Date:** March 31, 2015
Operator Signature: *Al Clark* **Location:** McIntyre Center Edmonton

***APPENDIX III
REPORT CERTIFICATION FORM***

Report Certification Form

Alberta Airshed (if applicable)	EPA Approval or Code of Practice Registration # (if applicable)
YES	NA
Company Name (if applicable)	Industrial Operation Name (if applicable)
Lakeland Industry & Community Association	St. Lina Continuous Monitoring Station
Name of the Representative of the Person Responsible (Last, First, Middle)	Position / Title of the Representative of the Person Responsible
Wunmi Adekanmbi	Project Manager, Customer Service, Air Services
Is an External Party Certifying the Report? (If 'Yes', fill in the fields below for the external person.)	
<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	
Name of External Person Certifying the Report (Last, First, Middle)	Position / Title of External Person Certifying the Report
NA	NA
Company Name for the External Person Certifying the Report	Identification of Qualifications / Professional Designations of the External Person Certifying the Report
NA	NA

I certify that I have reviewed and verified the submitted report. I also certify that the report presented with this certification form is complete, accurate and representative of the monitoring results and timeframe.

Maram Ghaleb

Signature of the Representative of the Person
Responsible / External Person Certifying the Report

May 12, 2017

Report Issued Date (dd-mm-yyyy)

APPENDIX IV
DATA VALIDATION CERTIFICATION FORM



Validation Certificate Form

Client: <u>Lakeland Industry & Community Association</u>	Project #: <u>2833-2017-03-31-C</u>
Site: <u>St. Lina Continuous Monitoring Station</u>	Contact: <u>Mike Bisaga</u>

Level 0 Preliminary Verification	<u>Maram Ghalab</u>	Date <u>April 30, 2017</u>
Level 1 Primary Validation	<u>Maram Ghalab</u>	Date <u>May 2, 2017</u>
Level 2 Final Validation	<u>Maram Ghalab</u>	Date <u>May 11, 2017</u>
Level 3 Independent Data Review	<u>Chadwick</u>	Date <u>May 12, 2017</u>
Post-Final Validation	<u>NA</u>	Date <u>NA</u>

Notes
The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. This validation is performed on an annual basis.



Alberta Environment and Parks (AEP)
Air.Reporting@gov.ab.ca

February 22, 2018

Subject: Monthly Report Submission for the LICA Portable (Bonnyville) station

Lakeland Industry & Community Association (LICA) is pleased to submit the ambient air monitoring monthly report for the LICA Portable (Bonnyville) AQM Station in the month of March 2017.

The air monitoring program consists of continuous air monitoring, intermittent sampling, including both VOC and PAH sampling program, and VOC canister sampling program. All the air monitoring activities were conducted by contractors.

Sampling Program	Monitoring Activities Conducted By	Sample Analysis Conducted By	Data/Report Review and Prepared By	Electronic Submission Conducted By
Continuous ambient air	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics
Intermittent	Maxxam Analytics	InnoTech Alberta Inc	InnoTech Alberta Inc	Not Applicable
VOC Canister	Maxxam Analytics	InnoTech Alberta Inc	InnoTech Alberta Inc	Not Applicable

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement systems.

All data collected in March 2017 was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016), with the exception of H₂S.

Non-Conformance: One instance of 1-hour exceedance for H₂S was recorded at 10.5 ppb on March 24 at hour 07. AE Reference #: 324242.

As the LICA Environmental Program Manager and Data & Reporting Specialist, we certify that we have reviewed and verified this report and that the information is complete, accurate and representative of the monitoring results, reporting timeframe and the specified analysis, summarization and reporting requirements. We also certify all air data that are required by the AMD to be electronically submitted to AEP and Alberta's Ambient Air Quality Data Warehouse have been submitted by the time of this report submission, with the exception of electronic submission for the results of intermittent samples and VOC canister samples. The results for both intermittent samples and VOC canister samples is scheduled to be submitted by the end of March 2018.

Should you have any questions, please don't hesitate to contact me.



Lakeland Industry & Community Association
5107 50 St
Bonnyville, AB T9N 2J7

Respectfully,

A handwritten signature in blue ink that reads 'Michael Bisaga'.

Michael Bisaga
Technical Program Manager
Lakeland Industry & Community Association
780-266-7068
mbisaga@otonabee.ca

A handwritten signature in blue ink that reads 'Lily Lin'.

Lily Lin
Data & Reporting Specialist
587-225-2248
rebbacaa@gmail.com



MAXXAM ANALYTICS
#1 2080 39 Ave. NE, Calgary, AB
T2E 6P7

maxxam.ca
Toll Free 800-386-7247
Fax 403-219-3673

AMBIENT AIR MONITORING MONTHLY DATA REPORT
LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
BONNYVILLE CONTINUOUS MONITORING STATION

JOB #: 2833-2017-03-35-C

March 2017

Prepared for:

LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
BOX 8237, 5107W - 50 STREET
BONNYVILLE, ALBERTA
T9N 2J5

Attention: MIKE BISAGA

DATE: **May 15, 2017**

Prepared by: *Maram Ghaleb*

Maram Ghaleb, B.Sc.
Project Manager, Customer Service, Air Services

Reviewed by: *Wunmi Adekanmbi*

Wunmi Adekanmbi, M.Sc., EPT.
Project Manager, Customer Service, Air Services

SUMMARY

In March 2017, Maxxam Analytics was contracted to manage the ambient air quality monitoring and maintenance activities at the Bonnyville Continuous Monitoring Station, near Bonnyville, Alberta. The monitoring station provides continuous meteorological measurements and air quality data for non-compliance parameters, as requested by the Lakeland Industry & Community Association.

All data collected this month, with the exception of H₂S on March 24 at hours 06:00-07:00, was compliant with the requirements outlined in the AMD, 2016.

Non-Conformance: One instance of 1-hour exceedance was recorded for H₂S this month at 10.5 ppb on March 24 at hour 07:00. Alberta Environment and Sustainable Resource Development (ESRD) Reference #: 324242.

Annual Maintenance: A scheduled annual station maintenance was completed between March 3 and March 4 on the following analyzers: SO₂, H₂S and O₃.

Twenty-one, twenty-four and one hours of downtime were recorded for SO₂, H₂S and O₃ respectively, due this maintenance event.

All Parameters: On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

PM_{2.5}: Thirty-two hours of data were recorded at concentrations lower than $-3 \mu\text{g}/\text{m}^3$ this month, rendering the data invalid.




The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

The summary of results is presented on the following pages.

Any deviations or modifications made to the sampling or analytical methods are outlined in Section 1.0, Discussion. On this basis, Maxxam Analytics is issuing this completed report to Lakeland Industry & Community Association, Bonnyville Continuous Monitoring Station.

Should you have any questions concerning the results or if we can be of further assistance, please contact us at 403-219-3678 or toll-free at 1-800-386-7247.

Pollutants		Monthly Records		1-Hour Records					24-Hour Records			
Name	Unit	Avg. Conc.	Uptime	Maximum			AAAQO Objective	Exceed. Hours	Maximum		AAAQO Objective	Exceed. Days
				Conc.	Date	Hour			Conc.	Date		
SO ₂	ppb	0.2	96.9%	1.9	March 7	14	172	0	1.0	March 14	48	0
H ₂ S	ppb	0.3	96.5%	10.5	March 24	7	10	1	2.2	March 24	3	0
THC	ppm	2.15	99.7%	3.28	March 31	5	-	-	2.41	March 3	-	-
CH ₄	ppm	2.15	99.7%	3.17	March 31	5	-	-	2.39	March 3	-	-
NMHC	ppm	0.01	99.7%	0.38	March 11	16	-	-	0.03	March 14	-	-
NO _x	ppb	9.9	99.7%	144.5	March 23	7	-	-	21.9	March 23	-	-
NO	ppb	2.9	99.7%	101.6	March 23	7	-	-	10.2	March 23	-	-
NO ₂	ppb	7.0	99.7%	42.9	March 23	7	159	0	12.8	March 3	-	-
O ₃	ppb	29.0	99.6%	45.6	March 16	13	82	0	38.6	March 19	-	-
PM _{2.5}	µg/m ³	6.3	95.4%	24.1	March 25	11	80	0	13.2	March 29	30	0
WS	kph	2.5	99.7%	22.7	March 31	22	-	-	14.0	March 19	-	-
WD	degree	80 (E)	99.7%	-	-	-	-	-	-	-	-	-



Monthly Update

All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.

All data collected this month, with the exception of H₂S on March 24 at hours 06:00-07:00, was compliant with the requirements outlined in the AMD, 2016.

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above 90%.

Operational Issues

Non-Conformance: One instance of 1-hour exceedance was recorded for H₂S this month at 10.5 ppb on March 24 at hour 07:00. Alberta Environment and Sustainable Resource Development (ESRD) Reference #: 324242.

Annual Maintenance: A scheduled annual station maintenance was completed between March 3 and March 4 on the following analyzers: SO₂, H₂S and O₃. Twenty-one, twenty-four and one hours of downtime were recorded for SO₂, H₂S and O₃ respectively, due this maintenance event.

All Parameters: On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

PM_{2.5}: Thirty-two hours of data were recorded at concentrations lower than -3 µg/m³ this month, rendering the data invalid.

Monthly Continuous Data Summary

Lakeland Industry & Community Association Bonnyville Continuous Monitoring Station						MAXIMUM VALUES							OPERATIONAL TIME (%)
PARAMETER	OBJECTIVES		EXCEEDANCES		MONTHLY AVERAGE	READING	DAY	1-HOUR			24-HOUR		
	1-hr	24-hr	1-hr	24-hr				HOUR	WIND SPEED (kph)	WIND DIRECTION (sector)	READING	DAY	
SO ₂ (ppb)	172	48	0	0	0.2	1.9	7	14	15.2	NW	1.0	14	96.9
H ₂ S (ppb)	10	3	1	0	0.3	10.5	24	7	1.2	ENE	2.2	24	96.5
THC (ppm)	-	-	-	-	2.15	3.28	31	5	1.2	N	2.41	3	99.7
CH ₄ (ppm)	-	-	-	-	2.15	3.17	31	5	1.2	N	2.39	3	99.7
NMHC (ppm)	-	-	-	-	0.01	0.38	11	16	9.5	ENE	0.03	14	99.7
NO ₂ (ppb)	159	-	0	-	7.0	42.9	23	7	1.2	N	12.8	3	99.7
NO (ppb)	-	-	-	-	2.9	101.6	23	7	1.2	N	10.2	23	99.7
NO _x (ppb)	-	-	-	-	9.9	144.5	23	7	1.2	N	21.9	23	99.7
O ₃ (ppb)	82	-	0	-	29.0	45.6	16	13	5.6	NE	38.6	19	99.6
PM _{2.5} (µg/m ³)	80	30	0	0	6.3	24.1	25	11	4.6	WSW	13.2	29	95.4
VECTOR WS (kph)	-	-	-	-	2.5	22.7	31	22	-	SSW	14.0	19	99.7
VECTOR WD (sec)	-	-	-	-	80 (E)	-	-	-	-	-	-	-	99.7

Exceedance Summary Report

SO₂ 1-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 1-hour AAAQO of 172 ppb.

SO₂ 24-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 24-hour AAAQO of 48.0 ppb.

H₂S 1-Hour Exceedances

DATE	TIME (MST)	READING (ppb)	WS (kph)	WD (deg)	ESRD Reference #
March 24	7	10.5	1.2	ENE	324242

H₂S 24-Hour Exceedances

Measured concentrations of hydrogen sulphide were below the 24-hour AAAQO of 3 ppb.

NO₂ 1-Hour Exceedances

Measured concentrations of nitrogen dioxide were below the 1-hour AAAQO of 159 ppb.

PM_{2.5} 1-Hour Exceedances

Measured concentrations of fine particulate matter were below the 1-hour AAAQO of 80 µg/m³.

PM_{2.5} 24-Hour Exceedances

Measured concentrations of fine particulate matter were below the 24-hour AAAQO of 30 µg/m³.

O₃ 1-Hour Exceedances

Measured concentrations of ozone were below the 1-hour AAAQO of 82 ppb.

In accordance with EPEA and the Substance Release Regulation.

In accordance with A Guide to Release Reporting and the Alberta Ambient Air Quality Objectives and Guidelines Summary.

Volatile Organics (VOCs) Data Summary

Sample Collection Date	Maximum Reading (ppb)	Volatile Organic Compound
March 2, 2017	2.37	n-Butane
March 8, 2017	1.4	Acetone
March 14, 2017	4.63	n-Butane
March 20, 2017	1.4	Acetone
March 26, 2017	3.2	Acetone

Note: NA

Polycyclic Aromatic Hydrocarbons (PAHs) Data Summary

Sample Collection Date	Maximum Reading ($\mu\text{g}/\text{puf}$)	Semi-Volatile Organic
March 2, 2017	0.56	2-Methylnaphthalene
March 8, 2017	3.89	Naphthalene
March 14, 2017	0.31	Phenanthrene
March 20, 2017	0.45	Naphthalene
March 26, 2017	0.16	Phenanthrene

Note: NA

Volatile Organics (VOCs) Data Summary - NMHC Canister System

Sample Collection Date	Maximum Reading (ppb)	Volatile Organic Compound
March 3, 2017	6.24	n-Butane
March 8, 2017	2.6	Ethanol
March 11, 2017	43.8	n-Butane
March 19, 2017	4.44	n-Butane
March 28, 2017	6.3	Acetone

Note: NA

TABLE OF CONTENTS

<u>Title</u>	<u>Page</u>
SUMMARY	1
MONTHLY DASHBOARD	2
MONTHLY CONTINUOUS DATA SUMMARY REPORT	3
EXCEEDANCE SUMMARY REPORT	4
VOC SAMPLE SUMMARY REPORT	5
PAH SAMPLE SUMMARY REPORT	6
NMHC CANISTER SAMPLE SUMMARY REPORT	7
TABLE OF CONTENTS	8
1.0 Discussion	10
.....	
2.0 Project Personnel	13
.....	
3.0 Plant Monthly Required AMD Summary	13
.....	
4.0 Calculations and Results	13
.....	
5.0 Methods and Procedures	14
.....	
Appendix I	17
.....	
Continuous Monitoring Data Results	17
Sulphur Dioxide	18
Hydrogen Sulphide	26
Total Hydrocarbon	34
Methane	42
Non-Methane Hydrocarbon	50
Oxides of Nitrogen	58
Nitric Oxides	66
Nitrogen Dioxide	73
Ozone	81
Particulate Matter 2.5	89
Wind Speed	94
Wind Direction	101
Standard Deviation Wind Direction	104

Appendix II	Non-Continuous Monitoring Data Results	107
	VOC Results	108
	PAH Results	124
	NMHC Canister Results	135
Appendix III	Equipment Calibration Results	151
	Sulphur Dioxide	152
	Hydrogen Sulphide	157
	Total Hydrocarbon	162
	Nitrogen Dioxide	166
	Ozone	170
	Particulate Matter	174
	Wind System	177
	Calibrators	180
	Calibration Gases	185
Appendix IV	Analytical Results	191
	VOCs Samples	192
	PAHs Samples	218
	NMHC Canister Samples	229
Appendix V	Report Certification Form	255
Appendix VI	Data Validation Certification Form	257

1.0 Discussion

This monthly report consists of continuous monitoring results for the following parameters: Sulphur Dioxide (SO₂), Hydrogen Sulphide (H₂S), Total Hydrocarbon (THC), Methane (CH₄), Non-Methane Hydrocarbon (NMHC), Oxides of Nitrogen (NO_x), Nitric Oxides (NO), Nitrogen Dioxide (NO₂), Ozone (O₃), Particulate Matter 2.5 (PM_{2.5}), Wind Speed (WS), Wind Direction (WD) and Standard Deviation Wind Direction (STDWD). The non-continuous monitoring data results for VOCs, PAHs and NMHC canister are also included in this report.

Sample filters for all continuous air monitors are changed before the calibration begins. The sample manifold is cleaned during the site visit each month.

Control checks, consisting of a zero and span, are conducted daily on all continuous air monitors. In place of the air sample, zero air (from scrubbed air or gas cylinders) is used for zero checks, and a known concentration of the pollutant being analyzed is used for span checks. These checks are controlled by automatic timers and valves. The total zero span cycle is completed within an hour, the commencement of the zero span cycle is at the beginning of the hour.

Multipoint calibrations are done a minimum of once a month for each continuous air monitor. An additional calibration is required under the following conditions: 1) within three days after the initial start-up and stabilization of a newly installed instrument, 2) prior to shut-down or moving of an instrument which has been working to specification, and 3) when major repair has been done on the instrument.

Time during the first multi-point calibration is not considered downtime (Data is flagged as C). If more than one calibration is performed during the month, the time during the additional calibration is considered as downtime (Data is flagged as C1).

Only one zero/span check is run per day. Time during the zero/span check is not considered as downtime (Data is flagged as S). If an extra zero/span check is performed, the time during the additional check is considered as downtime (Data is flagged as S1).

The AMD requires each instrument and accompanying data recording system to be operational 90% of the time, at a minimum, for each monthly monitoring period.

All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.

Data contained in this monthly report has undergone the verification and validation based on the requirements of the AMD Chapter 6: [Ambient Data Quality \(December, 2016\)](#). The descriptions of the data verification and validation process can be found in Section 5 of this report. Instantaneous data, where applicable, is provided for reference purposes and has not undergone zero correction. The minimum and maximum statistics are highlighted in the data table and are for reference only. The highlighted cells are based on the software's interpretation of the exact position of the minimum or maximum value. The visual presentation of these statistics may not be the obvious choice in a data range due to rounding, truncating or analyzer specifications.

Hourly/minute data have been reviewed based on daily zero/span results and multi-point calibration results. Data may be considered invalid if a zero-corrected span check in excess of +/- 10% of the span concentration (established by the previous multi-point calibration) is encountered and/or significant differences in the calibration factor occurs (greater than 10%).

SULPHUR DIOXIDE (SO₂)

- Operational time, for the monitoring period, was 96.9%, equivalent to twenty-three hours of downtime.
- A shut-down calibration was performed on March 3, prior to completing a scheduled annual maintenance on the analyzer. A new UV lamp was installed and the sample pump was rebuilt. A successful post-repair calibration was completed on March 4. Twenty-one hours of downtime were recorded due to this maintenance event.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

HYDROGEN SULPHIDE (H₂S)

- Operational time, for the monitoring period, was 96.5%, equivalent to twenty-six hours of downtime.
- A shut-down calibration was performed on March 3, prior to completing a scheduled annual maintenance on the analyzer. The reaction cell and sample valve were cleaned. The sinter filter was changed. A successful post-repair calibration was completed on March 4. Twenty-four hours of downtime were recorded due to this maintenance event.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

TOTAL HYDROCARBONS (THC), METHANE (CH₄) and NON-METHANE HYDROCARBONS (NMHC)

- Operational time, for the monitoring period, was 99.7%, equivalent to two hours of downtime.
- The routine monthly calibration was performed on March 3. The channel was placed in "maintenance" mode for few minutes on March 4, while the sample manifold was being cleaned. The maximum instantaneous data point at hour 10:00 was invalidated as a result.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

OXIDES OF NITROGEN (NO_x), NITRIC OXIDE (NO) and NITROGEN DIOXIDE (NO₂)

- Operational time, for the monitoring period, was 99.7%, equivalent to two hours of downtime.
- The routine monthly calibration was performed on March 3. The channel was placed in "maintenance" mode for few minutes on March 4, while the sample manifold was being cleaned. The maximum instantaneous data point at hour 10:00 was invalidated as a result.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

OZONE (O₃)

- Operational time, for the monitoring period, was 99.6%, equivalent to three hours of downtime.
- A shut-down calibration was performed on March 4, prior to completing a scheduled annual maintenance on the analyzer. The UV lamp was changed, a factory calibration was completed, a ZERO Air scrubber was re-positioned inside the analyzer and the sample pump was rebuilt. A successful post-repair calibration was subsequently completed. One hour of downtime was recorded due to this maintenance event.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.

PARTICULATE MATTER < 2.5 MICRONS (PM_{2.5})

- Operational time, for the monitoring period, was 95.4%, equivalent to thirty-four hours of downtime.
- Two routine TEOM audits were performed this month. The first was completed on March 3, and the second on March 28.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- Data was corrected in accordance with AMD (2016), Chapter 6, Table 2, Zero Adjustment Criteria. Data recorded between 0 and $-3 \mu\text{g}/\text{m}^3$ was corrected to $0 \mu\text{g}/\text{m}^3$. Data recorded below $-3 \mu\text{g}/\text{m}^3$ was invalidated. Thirty-two hours of data were invalidated as the data was below $-3 \mu\text{g}/\text{m}^3$ this month.

WIND SPEED (WS), WIND DIRECTION (WD) and STANDARD DEVIATION WIND DIRECTION (STDWD)

- Operational time, for the monitoring period, was 99.7%, equivalent to two hours of downtime.
- The wind system was calibrated on March 3.
- On March 11, the datalogger time was automatically synchronized to daylight savings time. This was immediately manually reverted to Mountain Standard Time (MST) as per AMD's requirement. However, two hours (7:00 and 8:00) of data were lost in the process.
- Wind data is reported as vector wind speed and vector wind direction. Wind direction is defined as the direction from which the wind is blowing from and is measured in degrees from true north.

VOC SAMPLES

- The sampler was programmed to run for 24 hours every 6th day per the NAPS (North American Pollution Surveillance Schedule).
- Samples were collected, as scheduled, on March 2, 8, 14, 20 and 26. Analysis was provided by InnoTech Alberta, results are included in this report.

PAH SAMPLES

- The sampler was programmed to run for 24 hours every 6th day per the NAPS (North American Pollution Surveillance Schedule).
- Samples were collected, as scheduled, on March 2, 8, 14, 20 and 26. Analysis was provided by InnoTech Alberta, results are included in this report.

NMHC CANISTER SAMPLES

The canister sampler is programmed to draw in a whole air sample when the 5-minute average concentration of NMHC is above 0.30 ppm. A representative sample of ambient air is collected over a one-hour period when the canister event is triggered.

Five canister events were recorded this month. The date, time and initial 5-min average concentration measurements are as follows:

- March 3 at 19:20 - 0.44 ppm
- March 8 at 19:05 - 0.48 ppm
- March 11 at 16:35 - 1.49 ppm
- March 19 at 06:50 - 0.36 ppm
- March 28 at 15:55 - 0.34 ppm

Other five-minute averages recorded at concentrations above 0.30 ppm are not considered sample-collection events as they occurred between events, before the canisters were replaced. Analysis was provided by InnoTech Alberta, results are included in this report.

2.0 Project Personnel

Mike Bisaga was the contact for Lakeland Industry & Community Association and the Maxxam field technician(s) was/were Alexander Yakupov and Limin Li.

3.0 Plant Monthly Required AMD Summary

All data collected this month, with the exception of H₂S on March 24 at hours 06:00-07:00, was compliant with the requirements outlined in the AMD, 2016.

Non-Conformance: One instance of 1-hour exceedance was recorded for H₂S this month at 10.5 ppb on March 24 at hour 07:00. Alberta Environment and Sustainable Resource Development (ESRD) Reference #: 324242.

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

4.0 Calculations and Results

All calculations and reporting of results follow the methods described in the AMD, 2016.

5.0 Methods and Procedures

The following methods and procedures were used to complete the monitoring program:

- Maxxam AIR SOP-00001: Methane, Non-Methane Hydrocarbon Analyzer Monitoring
- Maxxam AIR SOP-00208: RM Young Wind Monitor Calibration
- Maxxam AIR SOP-00209: Ambient Sulphur Monitoring
- Maxxam AIR SOP-00212: Ambient O₃ Monitoring
- Maxxam AIR SOP-00213: Ambient NO/NO₂/NO_x Monitoring
- Maxxam AIR SOP-00215: TEOM Operation
- Maxxam AIR SOP-00225: The Collection of VOCs in Ambient Air Using Canister and Xontech
- Maxxam AIR SOP-00007: TISCH PUF Sampler Operating, Calibration and Maintenance Procedures

There were no deviations from the prescribed methods.

The following instruments were used to perform the test program:

- Sulphur Dioxide - API 100E UV Fluorescent Analyzer
- Hydrogen Sulphide - API 101E UV Fluorescent Analyzer
- Methane, Non-Methane Hydrocarbon - Thermo 55i FID Analyzer
- Oxides of Nitrogen - API 200E Chemiluminescent Analyzer
- Ozone - Thermo 49i Photometric Analyzer
- Particulate Matter (PM_{2.5}) - R&P 1405F TEOM Unit
- Wind System - RM Young Unit
- Datalogger - ESC 8832
- VOC - XONTECH 910A Gaseous Air Sampler
- PAH - TISCH PUF Plus

The following steps were used to complete the data verification and validation process:

Level 0 Preliminary Verification

Level 0 data are raw data obtained directly from the data acquisition system (DAS). Under the step of Level 0, these data undergo a certain amount of manual or automated screening and flagging. It included a) identification of periods of missing data; b) verification of time stamps against reference time; c) verification that instrument diagnostics/datalogger flags indicate normal operation; d) comparison of data to upper and lower limits; e) rate of change flagging indicating that data changed too rapidly or not at all; and f) verification that zero, span and multipoint performance checks are within specifications. This level of verification is performed on a daily basis.

Level 1 Primary Validation

Validation actions under the step of Level 1 include a) review of all screening flags assigned during preliminary verification; b) review of all supporting site information and documentation; c) review of operational acceptance limits for each parameter/analyzer; d) review of daily zero/span and monthly calibration results for all gaseous parameters; and e) application of any necessary adjustments to data (e.g. baseline adjustments, below zero adjustments). This level of validation is performed on a monthly basis.

Level 2 Final Validation


The purpose of Level 2 validation is to verify that there are no inconsistencies among related data, or among regional data measured at nearby sites.

Level 3 Independent Data Review

Level 3 validation is the last step of data review, and it is completed by an individual that is independent of both field operations and primary data validation. A final independent QA review and endorsement is performed during this step before data is submitted to Alberta Environment.

Post-Final Validation

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. Any data issues or patterns which were not clear on a monthly basis are highlighted during this step. This validation is performed on an annual basis.



Source: Air Monitoring Directive (December 2016), Chapter 6, Ambient Data Quality; Figure 1 Data Collection and Management Process Flow Chart

APPENDIX I
CONTINUOUS MONITORING DATA RESULTS

SULPHUR DIOXIDE

SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.					
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.						
DAY																																	
1	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24		
2	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	1.0	0.0	24		
3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	0.0	0.0	0.0	12		
4	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	0.4	0.2	0.5	0.8	0.7	0.4	0.5	0.2	S	0.0	0.1	0.0	0.1	0.0	0.0	0.8	0.4	15		
5	0.0	0.0	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.0	0.1	0.0	0.1	0.1	0.0	0.1	0.1	0.2	0.3	0.0	0.1	0.2	0.0	0.3	0.1	24	
6	0.3	0.2	0.2	0.2	0.4	0.4	0.2	0.1	0.2	0.4	0.5	0.4	0.3	0.3	0.3	0.3	0.3	0.4	0.3	S	0.2	0.6	0.6	0.6	0.6	0.1	0.6	0.3	0.1	0.6	0.3	24	
7	0.5	0.5	0.4	0.4	0.4	0.3	0.4	0.8	1.0	1.0	1.1	1.4	1.7	1.7	1.9	1.8	1.8	1.9	S	0.8	0.8	0.5	0.5	0.5	0.3	1.9	1.0	24	0.3	1.9	1.0	24	
8	0.4	0.5	0.8	0.4	0.1	0.3	0.2	0.2	0.6	0.4	0.1	0.1	0.2	0.2	0.3	0.3	0.2	S	0.3	0.4	0.3	0.3	0.3	0.3	0.1	0.8	0.3	24	0.1	0.8	0.3	24	
9	0.1	0.0	0.1	0.3	0.0	0.0	0.0	0.1	0.4	0.8	0.8	0.4	1.0	1.2	0.9	0.6	S	0.7	0.7	0.7	0.4	0.7	0.6	0.2	0.0	1.2	0.5	24	0.0	1.2	0.5	24	
10	0.1	0.3	0.3	0.1	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.1	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.3	0.0	24	
11	0.0	0.0	0.0	0.0	0.0	0.0	0.0	X	X	0.1	0.1	0.0	0.1	0.0	0.1	S	0.0	0.0	0.0	0.0	0.1	0.1	0.1	0.2	0.0	0.2	0.0	0.0	0.0	0.2	0.0	22	
12	0.2	0.1	0.1	0.2	0.2	0.1	0.1	0.2	0.2	0.1	0.1	0.1	0.2	S	0.0	0.1	0.3	0.2	0.2	0.1	0.1	0.2	0.4	0.2	0.0	0.4	0.2	0.0	0.0	0.4	0.2	24	
13	0.2	0.1	0.3	0.4	0.3	0.2	0.3	0.4	0.6	1.1	1.2	1.2	S	1.0	1.1	0.9	1.0	1.1	1.2	1.0	0.9	0.7	0.8	0.9	0.1	1.2	0.7	24	0.1	1.2	0.7	24	
14	0.8	0.7	0.7	0.8	0.6	0.6	0.7	0.8	1.2	1.3	1.1	S	1.2	0.9	1.3	1.9	1.0	0.9	1.1	0.9	0.9	1.0	1.0	0.9	0.1	0.8	0.9	24	0.6	1.9	1.0	24	
15	0.7	0.7	0.7	0.7	0.7	0.8	0.5	0.8	0.7	0.7	S	0.3	0.5	0.5	0.5	0.3	0.2	0.3	0.2	0.1	0.0	0.0	0.1	0.0	0.0	0.8	0.4	24	0.0	0.8	0.4	24	
16	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.1	0.2	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	24	
17	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.1	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	24	
18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.1	0.2	0.5	0.7	0.9	1.0	1.0	1.0	1.1	1.2	1.2	1.3	1.4	1.4	1.1	0.0	1.4	0.6	24	0.0	1.4	0.6	24	
19	0.9	0.7	0.5	0.3	0.1	0.1	S	0.0	0.1	0.0	0.0	0.0	0.0	0.1	0.0	0.1	0.2	0.1	0.2	0.1	0.1	0.2	0.3	0.3	0.0	0.9	0.2	24	0.0	0.9	0.2	24	
20	0.1	0.1	0.1	0.1	0.1	S	0.0	0.2	0.2	0.2	0.1	0.0	0.1	0.1	0.0	0.0	0.1	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.1	24	0.0	0.2	0.1	24
21	0.0	0.0	0.1	0.1	S	0.0	0.0	0.0	0.3	0.3	0.1	0.1	0.0	0.0	0.1	0.0	0.1	0.1	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.3	0.1	24	0.0	0.3	0.1	24	
22	0.1	0.0	0.1	S	0.3	0.4	0.4	0.4	0.5	0.5	0.4	0.4	0.4	0.5	0.6	0.4	0.4	0.3	0.3	0.2	0.1	0.1	0.0	0.0	0.0	0.6	0.3	24	0.0	0.6	0.3	24	
23	0.0	0.0	S	0.0	0.0	0.3	0.0	1.6	0.6	0.5	0.9	0.7	0.7	0.9	1.1	1.1	0.8	0.4	0.4	0.5	0.7	0.2	0.2	0.3	0.0	1.6	0.5	24	0.0	1.6	0.5	24	
24	0.4	S	0.0	0.1	0.4	0.4	0.9	0.6	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.9	0.1	24	0.0	0.9	0.1	24	
25	S	0.0	0.0	0.0	0.1	0.1	0.0	0.0	0.0	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.2	0.0	24	0.0	0.2	0.0	24	
26	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.2	0.0	24	0.0	0.2	0.0	24	
27	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.1	0.1	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	S	0.0	0.0	0.0	0.2	0.0	24	0.0	0.2	0.0	24	
28	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	0.0	0.0	0.0	24
29	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.0	0.0	0.0	0.1	0.1	0.1	0.1	0.0	S	0.0	0.2	0.3	0.4	0.0	0.4	0.1	24	0.0	0.4	0.1	24	
30	0.2	0.2	0.3	0.3	0.4	0.3	0.3	0.4	0.4	0.4	0.3	0.2	0.1	0.1	0.1	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.4	0.2	24	0.0	0.4	0.2	24	
31	0.0	0.0	0.0	0.0	0.0	0.0	0.3	0.0	0.1	0.0	0.0	0.0	0.4	0.0	0.0	0.0	0.0	S	0.2	0.4	0.3	0.4	0.6	0.5	0.0	0.6	0.1	24	0.0	0.6	0.1	24	
HOURLY MAX	0.9	0.7	0.8	0.8	0.7	0.8	0.9	1.6	1.2	1.3	1.2	1.4	1.7	1.7	1.9	1.9	1.8	1.9	1.2	1.2	1.3	1.4	1.4	1.1									
HOURLY AVG	0.2	0.1	0.2	0.2	0.1	0.1	0.1	0.2	0.3	0.3	0.3	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.3	0.2									

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT:	1-HR	172	ppb	24-HR	48	ppb
-----------------------------	------	-----	-----	-------	----	-----


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0		
NUMBER OF 24-HR EXCEEDANCES:	0		
NUMBER OF NON-ZERO READINGS:	362		
MINIMUM 1-HR AVERAGE:	0.0 ppb @ HOUR(S) 0 ON DAY(S) 1		
MAXIMUM 1-HR AVERAGE:	1.9 ppb @ HOUR(S) 14 ON DAY(S) 7		
MAXIMUM 24-HR AVERAGE:	1.0 ppb ON DAY(S) 14		
	VAR-VARIOUS		
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	721 hrs
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	96.9 %
STANDARD DEVIATION:	0.4	MONTHLY AVERAGE:	0.2 ppb

24 HR AVERAGES March 2017



SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
2	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.0	1.6	0.4	0.4	0.5	0.6	0.6	0.6	S	0.0	1.6	0.3	24	
3	0.8	0.8	0.9	0.6	0.6	0.6	0.8	0.9	1.5	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	0.6	1.5	0.8	12
4	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	2.0	1.8	2.0	2.2	2.1	1.8	1.9	1.3	S	1.3	1.4	1.3	2.2	1.8	15	
5	1.2	1.2	1.3	1.3	1.2	1.2	1.2	1.2	1.2	1.2	1.3	1.3	1.4	1.6	1.5	1.5	1.5	1.5	1.5	1.4	S	1.7	1.7	1.7	1.2	1.7	1.4	24	
6	1.7	1.5	1.5	1.5	1.8	1.7	1.4	1.3	1.5	1.7	1.7	1.5	1.5	1.6	1.5	1.4	1.5	1.4	S	1.2	1.6	1.6	1.6	1.6	1.2	1.8	1.5	24	
7	1.5	1.3	1.2	1.2	1.1	1.0	1.2	1.3	1.6	1.5	1.7	2.0	2.2	2.1	2.2	2.1	2.2	2.3	S	1.0	1.1	0.8	0.8	0.7	0.7	2.3	1.5	24	
8	0.4	0.6	0.8	0.7	0.3	0.4	0.2	0.2	0.6	0.3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.3	0.0	0.1	0.0	0.0	0.0	0.8	0.2	24	
9	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
10	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
11	0.0	0.0	0.0	0.0	0.0	0.0	0.0	X	X	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	22	
12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.3	0.0	0.0	0.3	0.0	24	
13	0.2	0.1	0.2	0.3	0.2	0.1	0.5	0.6	1.0	1.3	1.5	1.5	S	1.5	1.5	1.3	1.4	1.6	1.5	1.5	1.2	1.2	1.3	1.3	0.1	1.6	1.0	24	
14	1.3	1.2	1.2	1.2	1.3	1.3	1.4	1.4	1.7	1.8	1.7	S	1.8	1.8	2.0	3.3	1.9	1.8	1.9	1.8	1.8	1.9	2.1	2.0	1.2	3.3	1.7	24	
15	1.9	1.9	1.9	2.1	2.1	2.1	1.8	2.5	2.6	2.3	S	2.0	2.1	2.1	2.1	2.0	1.7	1.8	1.8	1.7	1.7	1.5	1.6	1.6	1.5	2.6	2.0	24	
16	1.5	1.5	1.5	1.8	1.7	1.5	1.3	1.6	1.8	S	1.4	1.3	1.3	1.2	1.2	1.2	1.0	1.0	1.5	0.7	0.6	0.7	0.7	0.4	0.4	1.8	1.2	24	
17	0.4	0.4	0.5	0.6	0.6	0.6	0.6	0.7	S	0.6	0.7	0.7	0.9	1.0	1.4	1.0	1.2	1.3	1.2	1.2	1.2	1.1	1.1	1.2	0.4	1.4	0.9	24	
18	1.2	1.2	1.6	1.6	1.7	1.5	1.7	S	1.8	1.9	2.0	2.3	2.3	2.7	2.9	2.7	2.9	3.0	3.1	2.9	3.1	3.0	2.9	1.2	3.1	2.3	24		
19	2.5	2.3	2.2	2.1	1.7	1.5	S	1.4	1.6	1.3	1.3	1.3	1.2	1.2	1.0	1.0	1.1	1.0	1.0	0.9	0.9	0.7	0.8	0.8	0.7	2.5	1.3	24	
20	0.6	0.4	0.4	0.5	0.3	S	0.1	0.5	0.4	0.5	0.0	0.0	0.1	0.2	0.0	0.0	0.1	0.4	0.0	0.0	0.0	0.0	0.0	0.0	0.6	0.2	24		
21	0.0	0.0	0.5	0.1	S	0.0	0.0	0.2	0.4	0.5	0.4	0.3	0.4	0.5	0.6	0.7	0.6	0.7	0.7	0.8	0.9	0.9	0.9	1.0	0.0	1.0	0.5	24	
22	1.1	1.1	1.3	S	1.5	1.6	1.7	1.5	1.7	1.8	1.7	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.9	1.6	1.6	1.5	1.6	1.5	1.1	1.9	1.6	24	
23	1.6	1.4	S	1.2	1.3	1.8	1.2	3.9	2.9	1.9	2.1	1.8	1.8	2.1	2.1	2.0	1.7	1.3	1.2	1.2	2.3	0.9	0.9	0.7	0.7	3.9	1.7	24	
24	1.2	S	1.3	1.2	1.5	2.2	2.6	2.3	1.0	1.0	1.0	0.9	1.0	1.2	1.2	1.2	1.4	1.3	1.4	1.5	1.5	1.5	1.6	0.9	2.6	1.4	24		
25	S	1.5	1.7	1.7	1.9	1.6	1.5	1.6	1.7	1.8	1.6	1.5	1.2	1.5	1.3	1.4	1.2	1.3	1.3	1.2	1.2	1.2	1.3	S	1.2	1.9	1.5	24	
26	1.4	1.2	1.4	1.3	1.2	1.3	1.2	1.3	1.5	1.6	1.4	1.2	1.2	1.1	1.3	1.3	1.2	1.1	1.1	1.1	1.1	1.1	1.1	S	1.0	1.0	1.6	1.2	24
27	1.1	1.0	0.9	1.2	1.1	1.0	0.9	1.2	1.2	1.1	1.2	1.0	1.0	1.0	1.0	1.0	0.9	0.9	1.0	1.0	0.7	S	0.6	0.7	0.6	1.2	1.0	24	
28	0.8	0.8	1.0	0.8	0.8	1.0	0.9	0.8	0.8	1.0	1.0	0.9	0.9	0.8	1.1	1.0	1.3	1.2	1.0	1.1	S	0.9	1.0	1.2	0.8	1.3	1.0	24	
29	1.1	1.0	1.0	1.1	1.2	1.4	1.0	1.4	1.5	1.2	1.3	1.2	1.2	1.3	1.3	1.2	1.4	1.5	1.4	S	1.4	1.4	1.5	1.5	1.0	1.5	1.3	24	
30	1.5	1.4	1.4	1.5	1.3	1.5	1.3	1.5	1.3	1.6	1.3	1.1	1.0	1.0	0.9	0.8	0.6	0.7	S	0.5	0.4	0.6	0.6	0.4	0.4	1.6	1.1	24	
31	0.6	0.7	0.5	0.6	0.4	0.9	1.5	0.9	1.1	0.8	0.8	0.9	1.3	1.2	0.8	1.0	0.9	S	1.2	1.2	1.2	1.5	1.7	1.5	0.4	1.7	1.0	24	
HOURLY MAX	2.5	2.3	2.2	2.1	2.1	2.2	2.6	3.9	2.9	2.3	2.1	2.3	2.3	2.7	2.9	3.3	2.7	2.9	3.0	3.1	2.9	3.1	3.0	2.9					
HOURLY AVG	0.9	0.9	0.9	0.9	0.9	1.0	0.9	1.1	1.2	1.0	1.0	0.9	1.0	1.1	1.1	1.2	1.1	1.1	1.0	1.0	1.0	1.0	1.0	1.0					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	532
MAXIMUM INSTANTANEOUS VALUE:	3.9 ppb @ HOUR(S) 7 ON DAY(S) 23
VAR-VARIOUS	
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	7 hrs
OPERATIONAL TIME:	721 hrs
STANDARD DEVIATION:	0.7

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)



Wind: LICA Bonnyville
 Poll.: LICA Bonnyville-SO2[ppb]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 14.08%

Calm Avg: 0.23 [ppb]

Direction	0.0-0.4	0.4-0.8	0.8-1.2	1.2-1.6	1.6-2.0	>2.0	Total
N	5.4	2.8	0.6	0.2	0.2	0.0	9.1
NE	19.9	1.8	0.7	0.2	0.0	0.0	22.6
E	13.9	1.5	0.3	0.3	0.0	0.0	16.0
SE	13.8	1.8	2.5	0.2	0.2	0.0	18.3
S	2.5	0.2	1.3	0.3	0.0	0.0	4.3
SW	3.2	0.4	0.0	0.0	0.0	0.0	3.7
W	4.6	0.2	0.2	0.0	0.0	0.0	4.9
NW	4.1	1.2	0.9	0.3	0.7	0.0	7.2
Summary	67.5	9.7	6.5	1.3	1.0	0.0	86.0

% Icon Classes (ppb)

67 0.0-0.4

10 0.4-0.8


6 0.8-1.2

1 1.2-1.6

1 1.6-2.0

0 >2.0

LICA Bonnyville Poll.: LICA Bonnyville-SO2[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 14.08% Calm Poll Avg: 0.23[ppb]



SO2[ppb] Calibration: LICA Bonnyville Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

HYDROGEN SULPHIDE



HYDROGEN SULPHIDE Hourly Averages (H₂S ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	0.1	0.0	0.5	0.1	0.0	0.0	0.2	0.0	0.5	0.0	24	
2	S	0.1	0.1	0.2	0.1	0.3	0.2	0.4	0.3	0.1	0.2	0.1	0.1	0.2	0.3	0.3	0.1	0.5	0.3	0.3	0.1	0.1	0.2	S	0.1	0.5	0.2	24	
3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.5	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	0.0	0.5	0.1	12
4	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	0.1	0.3	0.2	0.1	0.0	S	0.0	0.0	0.0	0.3	0.1	12	
5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.2	0.3	0.3	S	0.6	0.2	0.0	0.0	0.6	0.1	24	
6	0.0	0.0	0.1	0.1	0.1	0.0	0.1	0.1	0.0	0.2	0.0	0.0	0.0	0.1	0.0	0.0	0.1	0.1	0.0	S	0.1	0.0	0.0	0.0	0.0	0.2	0.0	24	
7	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	S	0.3	0.3	0.0	0.0	0.3	0.0	0.3	0.0	24	
8	0.2	0.2	0.1	0.1	0.3	0.4	0.6	0.6	0.2	0.0	0.2	0.2	0.1	0.1	0.1	0.2	0.2	S	0.4	0.4	0.5	0.7	0.5	0.5	0.0	0.7	0.3	24	
9	0.5	0.5	0.4	0.7	0.4	0.4	0.5	0.5	0.6	0.8	0.7	0.0	0.1	0.3	0.2	0.3	S	0.0	0.4	0.9	1.0	1.2	1.0	1.3	0.0	1.3	0.6	24	
10	1.7	0.9	0.4	0.4	0.3	0.0	0.2	0.3	0.2	0.0	0.1	0.0	0.0	0.0	0.0	S	0.2	0.1	0.1	0.0	0.1	0.1	0.0	0.2	0.0	1.7	0.2	24	
11	0.0	0.1	0.1	0.2	0.2	0.0	0.0	X	X	0.3	0.1	0.0	0.0	0.1	S	0.1	0.1	0.2	0.2	0.2	0.1	0.2	0.2	0.1	0.0	0.3	0.1	22	
12	0.2	0.0	0.1	0.1	0.2	0.2	0.1	0.1	0.1	0.0	0.0	0.0	0.0	S	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.2	0.1	24	
13	0.1	0.1	0.0	0.0	0.0	0.2	0.0	0.2	0.1	0.1	0.1	0.1	S	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.0	0.2	24	
14	0.2	0.2	0.3	0.4	0.9	0.6	0.8	1.0	0.1	0.1	0.2	S	0.3	0.4	0.3	0.2	0.3	0.2	0.1	0.1	0.0	0.1	0.2	0.2	0.0	1.0	0.3	24	
15	0.2	0.2	0.1	0.0	0.1	0.1	0.5	1.2	0.5	0.6	S	0.4	0.2	0.1	0.1	0.1	0.2	0.3	0.1	0.1	0.0	0.1	0.2	0.0	0.0	1.2	0.2	24	
16	0.0	0.1	0.1	0.1	0.2	0.3	0.1	0.3	0.5	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.2	0.0	0.1	0.0	0.0	0.1	0.0	0.5	0.1	24	
17	0.2	0.1	0.1	0.0	0.0	0.0	0.0	0.0	S	0.1	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	24	
18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.1	0.1	0.2	0.1	0.2	0.3	0.4	0.4	0.3	0.4	0.6	0.6	0.4	0.5	0.0	0.6	0.2	24	
19	0.4	0.3	0.4	0.2	0.2	0.2	S	0.2	0.1	0.1	0.1	0.2	0.1	0.0	0.1	0.1	0.1	0.1	0.0	0.1	0.0	0.1	0.0	0.0	0.0	0.4	0.1	24	
20	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.2	0.2	0.0	0.1	0.1	0.0	0.2	0.0	24	
21	0.0	0.2	0.4	0.2	S	0.4	0.7	0.3	0.0	0.0	0.1	0.0	0.0	0.2	0.2	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.7	0.1	24	
22	0.1	0.0	0.0	S	0.2	0.1	0.2	0.2	0.2	0.1	0.2	0.1	0.3	0.2	0.1	0.2	0.3	0.2	0.2	0.1	0.2	0.5	0.6	0.0	0.6	0.2	24		
23	0.8	0.4	S	0.5	0.6	1.0	1.2	2.5	1.0	0.6	0.6	0.5	0.5	0.5	0.4	0.4	0.3	0.3	0.4	1.0	2.9	0.9	1.9	2.4	0.3	2.9	0.9	24	
24	1.2	S	3.8	2.8	4.7	8.1	10.4	10.5	3.6	0.5	0.4	0.3	0.3	0.4	0.3	0.3	0.3	0.3	0.3	0.2	0.3	0.2	0.2	0.2	0.2	10.5	2.2	24	
25	S	0.2	0.2	0.2	0.3	0.3	0.4	0.4	0.3	0.3	0.3	0.3	0.2	0.1	0.2	0.1	0.0	0.2	0.4	0.4	0.9	0.4	S	0.0	0.9	0.3	24		
26	1.0	0.3	0.4	0.9	1.0	0.9	0.6	0.4	0.5	0.5	0.4	0.3	0.2	0.1	0.0	0.0	0.0	0.1	0.2	0.2	0.4	0.3	S	0.1	1.0	0.4	24		
27	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	S	0.0	0.2	0.0	0.2	0.1	24	
28	0.3	0.1	0.0	0.0	0.0	0.1	0.1	0.2	0.1	0.1	0.0	0.0	0.1	0.0	0.2	0.1	0.1	0.0	0.2	0.2	S	0.2	0.3	0.1	0.0	0.3	0.1	24	
29	0.1	0.1	0.5	0.3	0.3	0.6	3.3	3.7	1.1	0.7	0.2	0.1	0.0	0.1	0.1	0.2	0.3	0.1	0.4	S	0.6	0.6	0.3	0.4	0.0	3.7	0.6	24	
30	0.2	0.2	0.2	0.4	0.3	0.3	0.3	0.4	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	S	0.2	0.3	0.3	0.2	0.2	0.2	0.4	0.3	24	
31	0.2	0.3	0.2	0.7	0.6	0.6	1.0	0.4	0.5	0.4	0.3	0.4	0.3	0.2	0.3	0.2	0.3	S	0.4	0.3	0.3	0.5	0.4	0.5	0.2	1.0	0.4	24	
HOURLY MAX	1.7	0.9	3.8	2.8	4.7	8.1	10.4	10.5	3.6	0.8	0.7	0.5	0.5	0.5	0.4	0.4	0.4	0.5	0.4	1.0	2.9	1.2	1.9	2.4					
HOURLY AVG	0.3	0.2	0.3	0.3	0.4	0.5	0.7	0.9	0.4	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.3	0.3	0.3	0.3					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT:	1-HR	10	ppb	24-HR	3	ppb
----------------------	------	----	-----	-------	---	-----


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	1		
NUMBER OF 24-HR EXCEEDANCES:	0		
NUMBER OF NON-ZERO READINGS:	470		
MINIMUM 1-HR AVERAGE:	0.0 ppb @ HOUR(S) 0 ON DAY(S) 1		
MAXIMUM 1-HR AVERAGE:	10.5 ppb @ HOUR(S) 7 ON DAY(S) 24		
MAXIMUM 24-HR AVERAGE:	2.2 ppb ON DAY(S) 24		
	VAR-VARIOUS		
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	718 hrs
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	96.5 %
STANDARD DEVIATION:	0.8	MONTHLY AVERAGE:	0.3 ppb

24 HR AVERAGES March 2017



HYDROGEN SULPHIDE Hourly Averages (H₂S ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

HYDROGEN SULPHIDE Instantaneous Maximum (H₂S ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	0.8	S	0.8	0.8	0.6	0.6	0.5	0.9	0.6	0.5	0.6	0.6	0.4	0.6	0.5	0.6	0.4	0.5	0.4	1.2	0.6	0.3	0.2	0.4	0.2	1.2	0.6	24	
2	S	0.3	0.3	0.4	0.3	0.7	0.6	1.6	0.7	0.5	0.5	0.5	0.6	0.6	0.7	0.9	0.8	1.1	0.9	0.9	0.7	0.6	0.7	S	0.3	1.6	0.7	24	
3	0.6	0.8	0.7	0.8	0.8	0.7	0.7	0.9	1.7	C	C	C	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	0.6	1.7	0.9	12
4	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y	C	C	C	C	0.3	0.6	0.4	0.4	0.4	0.4	S	0.3	2.6	0.3	2.6	0.7	12
5	0.3	0.1	0.3	0.4	0.4	0.4	0.4	0.3	0.4	0.4	0.3	0.3	0.3	0.3	0.4	0.4	0.4	0.6	0.6	0.7	S	1.0	0.9	0.4	0.1	1.0	0.4	24	
6	0.4	0.3	0.5	0.4	0.6	0.4	0.4	0.4	0.4	0.6	0.3	0.3	0.3	0.5	0.3	0.2	0.5	0.3	0.3	S	0.3	0.2	0.1	0.0	0.0	0.6	0.3	24	
7	0.2	0.2	0.1	0.2	0.2	0.2	0.2	0.0	0.1	0.2	0.2	0.0	0.1	0.0	0.0	0.0	0.0	0.1	S	0.5	0.4	0.0	0.0	0.2	0.0	0.5	0.1	24	
8	0.2	0.1	0.2	0.1	0.1	0.3	0.8	0.5	0.0	0.0	0.0	0.0	0.0	0.0	1.1	0.0	0.0	S	0.1	0.1	0.6	0.8	0.1	0.1	0.0	1.1	0.2	24	
9	0.0	0.0	0.1	0.6	0.0	0.0	0.0	0.2	0.5	0.6	1.6	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.7	0.6	1.2	1.1	0.8	0.0	1.6	0.3	24	
10	1.9	1.4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.9	0.1	24	
11	0.0	0.0	0.0	0.0	0.0	0.0	0.0	X	X	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	22	
12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	24	
14	0.3	0.4	0.2	0.4	1.0	1.0	0.9	1.2	0.2	0.0	0.1	S	0.3	0.2	0.2	0.1	0.2	0.2	0.0	0.0	0.0	0.0	0.0	0.2	0.0	1.2	0.3	24	
15	0.2	0.1	0.0	0.0	0.1	0.2	0.5	2.2	1.1	0.8	S	0.3	0.1	0.3	0.3	0.2	0.2	0.4	0.1	0.2	0.0	0.0	0.3	0.0	0.0	2.2	0.3	24	
16	0.0	0.1	0.1	0.0	0.4	1.0	0.1	0.5	0.6	S	0.2	0.0	0.4	0.0	0.0	0.0	0.0	0.4	1.0	0.0	0.0	0.0	0.0	0.5	0.0	1.0	0.2	24	
17	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	S	0.1	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	24	
18	0.0	0.0	0.0	0.0	0.2	0.1	0.2	S	0.2	0.3	0.3	0.2	0.2	0.2	0.2	0.4	0.5	0.4	0.4	0.8	0.6	0.4	0.4	0.0	0.8	0.3	0.3	24	
19	0.3	0.2	0.3	0.3	0.1	0.2	S	0.0	0.1	0.0	0.0	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.3	0.1	24	
20	0.0	0.0	0.0	0.0	0.0	S	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.4	0.0	0.4	0.0	0.0	0.0	0.0	0.0	0.0	0.4	0.0	24	
21	0.0	0.0	0.0	0.0	S	0.0	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.2	0.0	24		
22	0.0	0.0	0.1	S	0.2	0.1	0.1	0.3	0.1	0.0	0.2	0.1	0.3	0.1	0.1	0.3	0.2	0.2	0.2	0.2	0.3	0.3	0.4	0.7	0.0	0.7	0.2	24	
23	1.2	0.4	S	0.5	0.5	1.6	1.8	3.5	2.0	0.6	0.5	0.2	0.4	0.2	0.0	0.2	0.0	0.1	0.0	1.5	4.6	2.1	2.6	4.4	0.0	4.6	1.3	24	
24	2.2	S	6.4	4.3	7.1	9.4	26.6	23.8	6.3	0.2	0.2	0.0	0.1	0.2	0.0	0.1	0.0	0.2	0.1	0.1	0.1	0.0	0.1	0.0	0.1	0.0	26.6	3.8	24
25	S	0.0	0.1	0.1	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.1	0.0	0.0	0.0	0.0	0.3	0.3	0.5	1.4	0.8	1.0	S	0.0	1.4	0.3	24	
26	4.1	0.2	0.3	1.7	1.0	1.3	0.9	0.3	0.5	0.5	0.2	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.2	0.1	0.5	0.3	S	0.0	0.0	4.1	0.5	24	
27	0.0	0.0	0.0	0.0	0.1	0.1	0.1	0.0	0.0	0.0	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	S	0.0	0.5	0.0	0.5	0.0	24	
28	0.5	0.3	0.0	0.0	0.0	0.1	0.0	0.2	0.1	0.2	0.0	0.1	0.2	0.0	0.3	0.1	0.1	0.0	0.1	0.2	S	0.2	0.3	0.2	0.0	0.5	0.1	24	
29	0.2	0.1	0.6	0.4	0.2	0.8	12.5	12.5	2.0	1.2	0.3	0.1	0.0	0.0	0.1	0.1	0.5	0.1	0.4	S	1.1	1.0	0.3	0.3	0.0	12.5	1.5	24	
30	0.2	0.1	0.2	0.1	0.1	0.1	0.2	0.3	0.1	0.1	0.2	0.1	0.1	0.0	0.0	0.0	0.0	S	0.0	0.2	0.0	0.0	0.0	0.0	0.0	0.3	0.1	24	
31	0.0	0.0	0.0	0.5	0.3	0.3	1.1	0.2	0.2	0.1	0.1	0.0	0.0	0.0	0.1	0.1	S	0.2	0.1	0.2	0.3	0.2	0.3	0.0	1.1	0.2	24		
HOURLY MAX	4.1	1.4	6.4	4.3	7.1	9.4	26.6	23.8	6.3	1.2	1.6	0.6	0.6	1.1	0.9	0.8	1.1	1.0	1.5	4.6	2.1	2.6	4.4						
HOURLY AVG	0.5	0.2	0.4	0.4	0.5	0.7	1.7	1.8	0.6	0.3	0.2	0.1	0.1	0.1	0.2	0.1	0.1	0.2	0.2	0.3	0.5	0.3	0.3	0.4					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	399
MAXIMUM INSTANTANEOUS VALUE:	26.6 ppb @ HOUR(S) 6 ON DAY(S) 24
	VAR-VARIOUS
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	7 hrs
OPERATIONAL TIME:	718 hrs
STANDARD DEVIATION:	1.7

HYDROGEN SULPHIDE Instantaneous Maximum (H₂S ppb)




Wind: LICA Bonnyville
 Poll.: LICA Bonnyville-H2S[ppb]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 14.14% Calm Avg: 0.98 [ppb]


Direction	0.0-3.5	3.5-7.1	7.1-10.6	>10.6	Total
N	8.7	0.0	0.0	0.0	8.7
NE	22.7	0.0	0.0	0.0	22.7
E	16.1	0.0	0.0	0.0	16.1
SE	18.3	0.2	0.0	0.0	18.4
S	4.3	0.0	0.0	0.0	4.3
SW	3.7	0.0	0.0	0.0	3.7
W	4.9	0.0	0.0	0.0	4.9
NW	7.2	0.0	0.0	0.0	7.2
Summary	85.7	0.2	0.0	0.0	85.9

% Icon Classes (ppb) 86 0.0-3.5 0 3.5-7.1 0 7.1-10.6 0 >10.6

LICA Bonnyville Poll.: LICA Bonnyville-H2S[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 14.14% Calm Poll Avg: 0.98[ppb]



H2S[ppb] Calibration: LICA Bonnyville Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

TOTAL HYDROCARBON


TOTAL HYDROCARBONS Hourly Averages (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2.08	S	2.06	2.06	2.04	2.05	2.07	2.05	2.04	2.03	2.06	2.09	2.03	2.03	2.03	2.04	2.03	2.05	2.08	2.11	2.07	2.07	2.15	2.26	2.03	2.26	2.07	24	
2	S	2.27	2.22	2.24	2.23	2.31	2.27	2.38	2.43	2.46	2.39	2.34	2.28	2.22	2.23	2.48	2.48	2.43	2.40	2.43	2.39	2.33	2.25	S	2.22	2.48	2.34	24	
3	2.23	2.31	2.34	2.34	2.38	2.41	2.41	2.37	2.49	2.48	2.33	C	C	C	C	C	2.37	2.38	2.49	2.53	2.67	2.70	S	2.23	2.23	2.70	2.41	24	
4	3.02	2.66	2.04	2.07	2.10	2.07	2.01	2.00	2.00	2.00	2.00	2.01	2.05	2.07	2.04	2.04	2.03	2.04	2.05	2.01	2.03	S	2.00	1.99	1.99	3.02	2.10	24	
5	1.99	1.99	1.99	1.99	1.99	1.99	1.99	2.00	2.01	2.00	2.00	1.99	1.99	2.00	1.99	1.99	2.00	2.08	2.42	2.15	S	2.16	2.31	2.26	1.99	2.42	2.06	24	
6	2.27	2.12	2.00	2.02	2.04	2.01	2.01	2.02	2.06	2.10	2.05	2.05	2.05	2.03	2.03	2.03	2.05	2.03	2.04	S	2.04	2.03	2.02	2.03	2.00	2.27	2.05	24	
7	2.06	2.08	2.05	2.02	2.04	2.04	2.01	2.04	2.06	2.04	2.04	2.03	2.02	2.02	2.03	2.03	2.03	2.04	S	2.08	2.25	2.18	2.19	2.25	2.01	2.25	2.07	24	
8	2.20	2.17	2.07	2.08	2.13	2.20	2.28	2.29	2.12	2.01	2.01	2.01	2.01	2.00	1.99	2.00	1.99	S	2.00	2.17	2.20	2.20	2.44	2.30	1.99	2.44	2.12	24	
9	2.24	2.18	2.27	2.43	2.31	2.41	2.39	2.42	2.34	2.50	2.26	2.01	2.01	2.02	2.04	2.04	S	2.06	2.15	2.29	2.14	2.58	2.64	2.55	2.01	2.64	2.27	24	
10	2.90	2.77	2.50	2.43	2.33	2.22	2.14	2.11	2.09	2.08	2.06	2.03	2.01	2.01	2.01	S	2.03	2.03	2.03	2.07	2.09	2.10	2.16	2.18	2.01	2.90	2.19	24	
11	2.10	2.13	2.14	2.22	2.20	2.20	2.18	X	X	2.13	2.13	2.08	2.03	S	2.01	2.40	2.06	2.00	2.01	2.02	2.03	2.03	2.03	2.00	2.40	2.11	22		
12	2.02	2.02	2.01	2.02	2.01	2.01	2.02	2.03	2.03	2.02	2.00	2.01	2.00	S	2.01	2.04	2.01	2.02	2.05	2.06	2.05	2.08	2.07	2.06	2.00	2.08	2.03	24	
13	2.09	2.11	2.11	2.11	2.11	2.13	2.10	2.09	2.07	2.08	2.09	2.08	S	2.06	2.06	2.07	2.08	2.06	2.04	2.06	2.08	2.08	2.11	2.16	2.04	2.16	2.09	24	
14	2.32	2.28	2.31	2.79	2.45	2.37	2.93	2.44	2.38	2.37	2.38	S	2.38	2.42	2.45	2.32	2.40	2.46	2.34	2.37	2.18	2.16	2.08	2.15	2.08	2.93	2.38	24	
15	2.19	2.09	2.13	2.17	2.30	2.34	2.31	2.49	2.44	2.32	S	2.27	2.33	2.36	2.28	2.26	2.22	2.10	2.21	2.24	2.17	2.20	2.30	2.15	2.09	2.49	2.26	24	
16	2.05	2.15	2.09	2.05	2.22	2.37	2.05	2.49	2.11	S	2.10	2.02	1.97	1.97	1.98	1.99	1.99	2.00	2.13	2.11	2.17	2.09	2.21	2.25	1.97	2.49	2.11	24	
17	2.31	2.17	2.16	2.09	2.21	2.26	2.52	2.61	S	2.30	2.30	2.32	2.20	2.13	2.08	2.04	2.04	2.07	2.08	2.10	2.14	2.12	2.13	2.12	2.04	2.61	2.20	24	
18	2.08	2.06	2.08	2.11	2.10	2.09	2.05	S	2.00	2.02	2.03	2.02	2.01	2.01	2.01	2.02	2.05	2.08	2.09	2.05	2.02	2.07	2.06	1.99	1.99	2.11	2.05	24	
19	1.99	1.97	2.00	1.99	1.97	1.96	S	1.98	1.97	1.97	1.97	1.96	1.96	1.95	1.95	1.97	1.95	1.95	1.95	1.95	1.96	1.96	1.97	1.95	1.95	2.00	1.97	24	
20	1.98	2.01	2.01	2.00	2.01	S	2.00	2.02	2.01	1.99	1.99	1.98	1.98	1.98	1.98	1.99	2.00	2.02	2.02	2.06	2.03	2.08	2.05	1.98	2.08	2.01	24		
21	2.08	2.19	2.31	2.49	S	2.44	2.44	2.43	2.40	2.29	2.12	2.09	2.06	2.02	2.06	2.05	2.04	2.02	2.00	2.02	2.04	2.05	2.04	2.06	2.00	2.49	2.16	24	
22	2.06	2.04	2.05	S	2.03	2.04	2.05	2.05	2.04	2.02	2.02	2.02	2.01	1.99	1.99	2.00	2.04	2.11	2.08	2.02	2.04	2.06	2.06	2.06	1.99	2.11	2.04	24	
23	2.05	2.02	S	2.03	2.03	2.30	2.22	2.47	2.49	2.28	2.16	2.11	2.11	2.05	2.05	2.05	2.03	2.01	2.06	2.09	2.18	2.21	2.21	2.27	2.01	2.49	2.15	24	
24	2.84	S	2.29	2.34	2.46	2.67	2.86	2.88	2.48	2.57	2.34	2.40	2.53	2.44	2.26	2.13	2.09	2.03	2.02	2.01	2.01	2.04	2.02	2.00	2.00	2.88	2.34	24	
25	S	1.98	1.99	2.00	2.02	2.08	2.18	2.27	2.23	2.11	2.13	2.13	2.07	1.99	2.00	2.02	2.03	2.05	2.09	2.16	2.55	2.26	2.37	S	1.98	2.55	2.12	24	
26	2.89	2.32	2.43	2.50	2.47	2.51	2.63	2.70	2.52	2.40	2.41	2.37	2.31	2.27	2.18	2.19	2.23	2.23	2.18	2.09	2.25	2.30	S	2.16	2.09	2.89	2.37	24	
27	2.18	2.23	2.26	2.22	2.28	2.23	2.17	2.06	2.02	2.04	2.13	2.10	2.09	2.02	2.03	2.03	2.03	2.01	2.00	2.01	2.00	S	2.04	2.11	2.00	2.28	2.10	24	
28	2.17	2.15	2.13	2.06	2.08	2.09	2.12	2.10	2.09	2.11	2.12	2.10	2.11	2.11	2.12	2.12	2.09	2.11	2.07	2.09	S	2.11	2.08	2.08	2.06	2.17	2.10	24	
29	2.07	2.08	2.22	2.34	2.35	2.51	2.41	2.65	2.50	2.25	2.05	2.03	2.02	2.02	2.04	2.05	2.04	2.06	2.13	S	2.27	2.32	2.55	2.31	2.02	2.65	2.23	24	
30	2.14	2.00	2.00	1.98	1.98	1.98	1.98	1.97	1.97	1.96	1.96	1.97	1.97	1.98	1.97	1.98	1.97	1.98	S	2.30	2.05	2.66	2.18	2.26	1.96	2.66	2.05	24	
31	2.20	2.64	2.33	2.29	2.43	3.28	3.07	2.34	2.28	2.23	2.21	2.26	2.15	1.99	1.98	2.03	2.04	S	2.08	2.04	2.10	2.14	2.09	2.07	1.98	3.28	2.27	24	
HOURLY MAX	3.02	2.77	2.50	2.79	2.47	3.28	3.07	2.88	2.52	2.57	2.41	2.40	2.53	2.44	2.45	2.48	2.48	2.46	2.49	2.53	2.67	2.70	2.64	2.55					
HOURLY AVG	2.23	2.18	2.15	2.18	2.18	2.25	2.26	2.27	2.20	2.17	2.13	2.10	2.10	2.08	2.06	2.07	2.09	2.09	2.11	2.13	2.14	2.19	2.17	2.15					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	704			
MINIMUM 1-HR AVERAGE:	1.95 ppm	@ HOUR(S)	13	ON DAY(S) 19
MAXIMUM 1-HR AVERAGE:	3.28 ppm	@ HOUR(S)	5	ON DAY(S) 31
MAXIMUM 24-HR AVERAGE:	2.41 ppm			ON DAY(S) 3
				VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	742 hrs	
MONTHLY CALIBRATION TIME:	5 hrs	AMD OPERATION UPTIME:	99.7 %	
STANDARD DEVIATION:	0.19	MONTHLY AVERAGE:	2.15 ppm	

TOTAL HYDROCARBONS Hourly Averages (THC ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY 1	2.10	S	2.07	2.08	2.06	2.06	2.09	2.07	2.08	2.06	2.50	2.63	2.06	2.06	2.06	2.10	2.05	2.13	2.12	2.25	2.10	2.11	2.20	2.51	2.05	2.63	2.15	24
2	S	2.31	2.28	2.38	2.38	2.48	2.36	2.54	2.64	2.70	2.45	5.72	4.81	2.26	2.28	3.04	3.30	2.57	2.55	2.67	2.64	2.68	2.53	S	2.26	5.72	2.80	24
3	2.36	2.50	2.79	2.59	2.67	2.95	3.01	2.64	2.85	2.82	2.51	C	C	C	C	C	2.73	3.08	3.36	3.54	3.66	3.12	S	2.34	2.34	3.66	2.86	24
4	6.29	3.08	2.13	2.18	2.17	2.16	2.04	2.03	2.07	2.06	Y	2.09	2.18	2.19	2.17	2.14	2.14	2.15	2.20	2.05	2.15	S	2.09	2.04	2.03	6.29	2.35	23
5	2.01	2.03	2.03	2.02	2.02	2.03	2.02	2.07	2.18	2.04	2.20	2.12	2.05	2.06	2.04	2.03	2.18	2.44	3.16	3.53	S	2.34	2.74	2.44	2.01	3.53	2.25	24
6	2.42	2.34	2.06	2.10	2.23	2.20	2.10	2.59	2.23	2.29	2.15	2.17	2.14	2.11	2.10	2.13	2.54	2.09	2.13	S	2.16	2.11	2.09	2.19	2.06	2.59	2.20	24
7	2.16	2.25	2.10	2.10	2.38	2.10	2.06	2.09	2.09	2.16	2.12	2.10	2.07	2.07	2.09	2.12	2.11	2.16	S	2.28	2.68	2.54	2.43	2.54	2.06	2.68	2.21	24
8	2.66	2.32	2.22	2.18	2.50	2.59	2.49	2.43	2.25	2.04	2.04	2.07	2.10	2.04	2.09	2.08	2.10	S	2.06	6.06	2.82	3.27	4.16	2.67	2.04	6.06	2.58	24
9	2.43	2.23	2.42	3.32	2.42	2.52	2.52	2.59	2.41	2.75	2.60	2.06	2.07	2.03	2.11	2.20	S	2.22	3.36	2.96	2.73	3.66	3.50	3.58	2.03	3.66	2.64	24
10	3.05	3.52	2.86	2.61	2.62	2.32	2.19	2.12	2.12	2.29	2.12	2.05	2.13	2.03	2.03	S	2.06	2.09	2.07	2.10	2.16	2.20	2.19	2.38	2.03	3.52	2.32	24
11	2.15	2.16	2.20	2.55	2.27	2.37	2.27	X	X	2.23	2.27	2.21	2.13	2.10	S	2.03	4.76	2.96	2.02	2.07	2.08	2.06	2.08	2.05	2.02	4.76	2.33	22
12	2.04	2.05	2.06	2.04	2.03	2.03	2.05	2.05	2.05	2.15	2.17	2.06	2.04	S	2.04	2.44	2.02	2.04	2.07	2.07	2.08	2.29	2.08	2.08	2.02	2.44	2.09	24
13	2.10	2.13	2.13	2.12	2.36	2.16	2.23	2.28	2.18	2.08	2.16	2.09	S	2.08	2.07	2.09	2.17	2.09	2.05	2.07	2.09	2.11	2.16	2.37	2.05	2.37	2.15	24
14	3.51	2.58	2.61	4.46	4.46	3.61	4.32	2.85	2.66	2.51	2.60	S	2.66	2.53	2.72	2.49	2.67	2.91	2.58	2.67	2.32	2.43	2.17	2.47	2.17	4.46	2.90	24
15	2.63	2.35	2.23	2.36	2.74	2.90	3.19	3.29	2.74	2.45	S	2.36	2.42	2.56	2.35	2.33	2.45	2.14	2.51	2.33	2.37	2.35	4.49	2.44	2.14	4.49	2.61	24
16	2.21	2.34	2.31	2.31	4.98	4.55	2.38	6.20	2.59	S	2.21	2.21	2.03	2.22	2.12	2.10	2.04	2.19	2.68	2.30	2.87	2.19	2.90	2.87	2.03	6.20	2.73	24
17	3.96	2.63	2.70	2.22	2.40	2.50	3.35	3.17	S	2.51	2.42	2.44	2.29	2.15	2.16	2.15	2.06	2.14	2.10	2.13	2.15	2.22	2.14	2.13	2.06	3.96	2.44	24
18	2.10	2.11	2.18	2.43	2.23	2.34	2.13	S	2.06	2.23	2.17	2.13	2.05	2.10	2.14	2.12	2.31	2.18	2.23	2.10	2.12	2.40	2.12	2.06	2.05	2.43	2.18	24
19	2.01	1.99	2.02	2.00	1.98	1.98	S	1.99	1.98	1.98	1.98	1.98	1.97	1.97	1.96	2.03	2.04	1.96	1.96	1.96	1.96	1.97	1.98	1.96	1.96	2.04	1.98	24
20	2.05	2.32	2.06	2.06	2.06	S	2.02	2.07	2.05	2.10	2.05	2.01	2.01	1.99	1.99	2.00	2.23	2.10	2.14	2.15	2.66	2.25	2.09	2.12	1.99	2.66	2.11	24
21	2.13	2.46	2.51	2.70	S	2.57	2.65	2.60	2.45	2.56	2.15	2.13	2.18	2.10	2.16	2.17	2.13	2.18	2.07	2.13	2.14	2.29	2.15	2.20	2.07	2.70	2.30	24
22	2.17	2.13	2.14	S	2.15	2.14	2.15	2.16	2.13	2.10	2.17	2.15	2.03	2.01	2.03	2.11	2.17	2.23	2.48	2.05	2.08	2.08	2.07	2.09	2.01	2.48	2.13	24
23	2.13	2.04	S	2.08	2.14	3.39	2.91	4.73	3.69	2.82	2.30	2.16	2.20	2.10	2.12	2.13	2.14	2.03	2.35	2.16	3.47	2.64	3.38	3.39	2.03	4.73	2.63	24
24	5.88	S	2.83	2.81	2.66	3.37	3.72	3.43	2.62	3.35	2.42	2.53	2.65	2.55	2.37	2.24	2.14	2.07	2.12	2.11	2.04	2.26	2.07	2.05	2.04	5.88	2.71	24
25	S	2.00	2.01	2.03	2.15	2.28	2.38	2.51	2.47	2.25	2.14	2.19	2.13	2.00	2.01	2.03	2.05	2.21	2.33	2.51	3.00	2.89	2.88	S	2.00	3.00	2.29	24
26	3.82	2.48	2.91	2.70	3.24	3.04	3.19	3.06	2.82	2.48	2.64	2.60	2.34	2.30	2.41	2.21	2.26	2.68	2.63	2.23	3.33	2.42	S	2.30	2.21	3.82	2.70	24
27	2.27	2.49	2.40	2.29	2.38	2.38	2.29	2.10	2.08	2.14	2.18	2.17	2.14	2.17	2.24	2.07	2.05	2.02	2.02	2.03	2.04	S	2.09	2.19	2.02	2.49	2.18	24
28	2.37	2.43	2.30	2.13	2.41	2.32	2.13	2.16	2.10	2.22	2.24	2.13	2.16	2.14	2.24	2.86	2.17	2.26	2.25	2.16	S	2.97	2.25	2.28	2.10	2.97	2.29	24
29	2.12	2.13	2.41	2.74	2.73	2.74	2.86	3.99	2.95	2.41	2.14	2.27	2.03	2.03	2.08	2.20	2.10	2.15	2.45	S	2.62	2.43	2.92	2.55	2.03	3.99	2.48	24
30	2.36	2.02	2.03	2.01	2.00	2.01	2.00	2.00	2.01	2.01	1.98	2.02	1.99	2.00	2.04	2.05	2.02	2.13	S	3.23	2.10	4.43	3.07	4.30	1.98	4.43	2.34	24
31	2.95	3.43	3.64	2.62	2.96	4.36	4.10	3.27	2.32	2.27	2.23	2.36	2.32	2.02	2.03	2.61	2.06	S	2.61	2.15	2.26	2.15	2.12	2.12	2.02	4.36	2.65	24
HOURLY MAX	6.29	3.52	3.64	4.46	4.98	4.55	4.32	6.20	3.69	3.35	2.64	5.72	4.81	2.56	2.72	3.04	4.76	3.08	3.36	6.06	3.66	4.43	4.49	4.30				
HOURLY AVG	2.70	2.37	2.35	2.41	2.53	2.62	2.57	2.73	2.37	2.34	2.25	2.32	2.25	2.14	2.15	2.22	2.31	2.26	2.37	2.48	2.44	2.51	2.52	2.44				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	703
MAXIMUM INSTANTANEOUS VALUE:	6.29 ppm @ HOUR(S) 0 ON DAY(S) 4
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	5 hrs
OPERATIONAL TIME:	741 hrs
STANDARD DEVIATION:	0.57

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)




Wind: LICA Bonnyville
 Poll.: LICA Bonnyville-THC55[ppm]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 14.37% Calm Avg: 2.32 [ppm]


Direction	0.0-1.1	1.1-2.2	2.2-3.3	>3.3	Total
N	0.0	6.8	3.4	0.0	10.2
NE	0.0	18.4	4.4	0.0	22.8
E	0.0	8.8	6.7	0.0	15.5
SE	0.0	13.9	3.8	0.0	17.8
S	0.0	4.0	0.1	0.0	4.1
SW	0.0	1.6	2.0	0.0	3.6
W	0.0	4.7	0.0	0.0	4.7
NW	0.0	6.3	0.7	0.0	7.0
Summary	0.0	64.4	21.2	0.0	85.6

% Icon	Classes (ppm)	0	0.0-1.1	64	1.1-2.2	21	2.2-3.3	0	>3.3
--------	---------------	---	---------	----	---------	----	---------	---	------

LICA Bonnyville Poll.: LICA Bonnyville-THC55[ppm] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 14.37% Calm Poll Avg: 2.32[ppm]



THC55[ppm] Calibration: LICA Bonnyville Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

METHANE




METHANE Hourly Averages (CH₄ ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2.08	S	2.06	2.06	2.04	2.05	2.07	2.05	2.04	2.03	2.03	2.03	2.03	2.03	2.04	2.03	2.05	2.08	2.11	2.07	2.07	2.15	2.26	2.03	2.26	2.06	24		
2	S	2.27	2.21	2.24	2.23	2.31	2.27	2.38	2.43	2.46	2.39	2.29	2.24	2.22	2.23	2.38	2.37	2.43	2.40	2.43	2.38	2.32	2.25	S	2.21	2.46	2.32	24	
3	2.23	2.31	2.34	2.34	2.38	2.41	2.40	2.37	2.48	2.47	2.33	C	C	C	C	C	2.32	2.35	2.46	2.45	2.58	2.62	S	2.22	2.22	2.62	2.39	24	
4	2.88	2.56	2.04	2.07	2.10	2.07	2.01	2.00	2.00	2.00	2.00	2.00	2.03	2.07	2.04	2.04	2.03	2.04	2.05	2.01	2.03	S	2.00	1.99	1.99	2.88	2.09	24	
5	1.99	1.99	1.99	1.99	1.99	1.99	1.99	2.00	2.01	2.00	2.00	1.99	1.99	2.00	1.99	1.99	2.00	2.08	2.38	2.12	S	2.15	2.30	2.26	1.99	2.38	2.05	24	
6	2.27	2.12	2.00	2.02	2.04	2.01	2.01	2.02	2.06	2.10	2.05	2.05	2.05	2.03	2.03	2.03	2.04	2.03	2.04	S	2.04	2.03	2.02	2.03	2.00	2.27	2.05	24	
7	2.06	2.08	2.05	2.02	2.04	2.04	2.01	2.04	2.06	2.04	2.04	2.03	2.02	2.02	2.03	2.03	2.03	2.04	S	2.08	2.25	2.18	2.19	2.24	2.01	2.25	2.07	24	
8	2.20	2.17	2.07	2.08	2.13	2.19	2.28	2.29	2.12	2.01	2.01	2.01	2.01	2.00	1.99	2.00	1.99	S	2.00	2.04	2.19	2.16	2.41	2.29	1.99	2.41	2.11	24	
9	2.24	2.18	2.27	2.40	2.31	2.41	2.39	2.42	2.34	2.42	2.24	2.01	2.01	2.02	2.03	2.04	S	2.06	2.11	2.27	2.14	2.51	2.59	2.52	2.01	2.59	2.26	24	
10	2.89	2.76	2.50	2.43	2.33	2.22	2.14	2.11	2.09	2.07	2.06	2.03	2.01	2.01	2.01	S	2.03	2.03	2.03	2.07	2.09	2.10	2.16	2.18	2.01	2.89	2.19	24	
11	2.10	2.13	2.14	2.21	2.20	2.20	2.18	X	X	2.13	2.13	2.12	2.08	2.03	S	2.01	2.01	2.01	2.00	2.01	2.02	2.03	2.03	2.00	2.21	2.09	22		
12	2.02	2.02	2.01	2.02	2.01	2.01	2.02	2.03	2.03	2.02	2.00	2.01	2.00	S	2.01	2.01	2.01	2.02	2.05	2.06	2.05	2.07	2.07	2.06	2.00	2.07	2.03	24	
13	2.09	2.11	2.11	2.11	2.11	2.13	2.10	2.09	2.07	2.08	2.09	2.08	S	2.06	2.06	2.07	2.08	2.06	2.04	2.06	2.08	2.08	2.11	2.16	2.04	2.16	2.09	24	
14	2.29	2.27	2.29	2.68	2.41	2.34	2.79	2.39	2.38	2.36	2.34	S	2.36	2.41	2.43	2.32	2.38	2.42	2.30	2.31	2.17	2.15	2.08	2.15	2.08	2.79	2.35	24	
15	2.19	2.09	2.13	2.17	2.29	2.32	2.30	2.46	2.40	2.32	S	2.27	2.32	2.36	2.28	2.26	2.22	2.10	2.20	2.24	2.17	2.20	2.27	2.14	2.09	2.46	2.25	24	
16	2.05	2.15	2.09	2.05	2.20	2.33	2.05	2.44	2.09	S	2.10	2.02	1.97	1.97	1.97	1.99	1.99	2.00	2.13	2.10	2.17	2.09	2.21	2.25	1.97	2.44	2.10	24	
17	2.28	2.17	2.15	2.09	2.21	2.26	2.50	2.61	S	2.30	2.30	2.32	2.20	2.13	2.08	2.04	2.04	2.07	2.08	2.10	2.14	2.12	2.13	2.11	2.04	2.61	2.19	24	
18	2.08	2.06	2.08	2.11	2.10	2.08	2.05	S	2.00	2.02	2.03	2.02	2.01	2.01	2.01	2.02	2.04	2.08	2.09	2.05	2.02	2.07	2.06	1.99	1.99	2.11	2.05	24	
19	1.99	1.97	2.00	1.99	1.97	1.96	S	1.98	1.97	1.97	1.97	1.96	1.96	1.95	1.95	1.97	1.95	1.95	1.95	1.95	1.95	1.96	1.96	1.97	1.95	2.00	1.97	24	
20	1.98	2.01	2.01	2.00	2.01	S	2.00	2.02	2.01	1.99	1.99	1.98	1.98	1.98	1.98	1.99	1.99	2.01	2.02	2.06	2.01	2.08	2.05	1.98	2.08	2.01	24		
21	2.08	2.19	2.31	2.49	S	2.44	2.44	2.43	2.40	2.29	2.12	2.09	2.06	2.02	2.06	2.05	2.04	2.02	2.00	2.02	2.04	2.05	2.04	2.06	2.00	2.49	2.16	24	
22	2.06	2.04	2.05	S	2.03	2.04	2.04	2.05	2.04	2.02	2.02	2.02	2.01	1.99	1.99	1.99	2.04	2.11	2.08	2.02	2.04	2.06	2.06	2.06	1.99	2.11	2.04	24	
23	2.05	2.02	S	2.03	2.03	2.26	2.21	2.31	2.38	2.27	2.16	2.11	2.11	2.05	2.05	2.05	2.03	2.01	2.05	2.08	2.16	2.20	2.20	2.25	2.01	2.38	2.13	24	
24	2.73	S	2.28	2.32	2.45	2.61	2.77	2.84	2.48	2.56	2.34	2.40	2.53	2.44	2.26	2.13	2.09	2.02	2.02	2.01	2.01	2.03	2.02	2.00	2.00	2.84	2.32	24	
25	S	1.98	1.99	2.00	2.02	2.08	2.18	2.26	2.23	2.11	2.11	2.13	2.07	1.99	1.99	2.02	2.03	2.05	2.09	2.16	2.53	2.25	2.37	S	1.98	2.53	2.12	24	
26	2.85	2.32	2.42	2.50	2.46	2.50	2.61	2.69	2.52	2.39	2.40	2.37	2.31	2.27	2.18	2.19	2.23	2.23	2.18	2.08	2.24	2.29	S	2.16	2.08	2.85	2.36	24	
27	2.18	2.23	2.26	2.22	2.28	2.23	2.17	2.06	2.02	2.04	2.13	2.10	2.09	2.02	2.02	2.03	2.03	2.01	2.00	2.01	2.00	S	2.04	2.11	2.00	2.28	2.10	24	
28	2.17	2.15	2.13	2.06	2.08	2.09	2.12	2.10	2.09	2.11	2.12	2.10	2.11	2.11	2.11	2.08	2.09	2.10	2.07	2.09	S	2.09	2.08	2.08	2.06	2.17	2.10	24	
29	2.07	2.08	2.22	2.33	2.34	2.48	2.40	2.57	2.45	2.24	2.05	2.02	2.02	2.02	2.04	2.04	2.04	2.05	2.12	S	2.25	2.31	2.54	2.31	2.02	2.57	2.22	24	
30	2.14	2.00	2.00	1.98	1.98	1.98	1.98	1.97	1.97	1.96	1.96	1.97	1.97	1.98	1.97	1.98	1.97	1.98	S	2.27	2.05	2.56	2.16	2.24	1.96	2.56	2.04	24	
31	2.19	2.56	2.31	2.27	2.41	3.17	2.96	2.34	2.28	2.22	2.21	2.26	2.15	1.99	1.98	2.01	2.04	S	2.07	2.04	2.08	2.14	2.09	2.07	1.98	3.17	2.25	24	
HOURLY MAX	2.89	2.76	2.50	2.68	2.46	3.17	2.96	2.84	2.52	2.56	2.40	2.40	2.53	2.44	2.43	2.38	2.38	2.43	2.46	2.45	2.58	2.62	2.59	2.52					
HOURLY AVG	2.22	2.17	2.15	2.18	2.17	2.24	2.25	2.25	2.19	2.17	2.12	2.10	2.09	2.08	2.06	2.06	2.07	2.08	2.11	2.11	2.14	2.17	2.16	2.15					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	704			
MINIMUM 1-HR AVERAGE:	1.95 ppm	@ HOUR(S)	13	ON DAY(S) 19
MAXIMUM 1-HR AVERAGE:	3.17 ppm	@ HOUR(S)	5	ON DAY(S) 31
MAXIMUM 24-HR AVERAGE:	2.39 ppm			ON DAY(S) 3
				VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	742 hrs	
MONTHLY CALIBRATION TIME:	5 hrs	AMD OPERATION UPTIME:	99.7 %	
STANDARD DEVIATION:	0.18	MONTHLY AVERAGE:	2.15 ppm	

METHANE Hourly Averages (CH₄ ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

METHANE MAX Instantaneous Maximum (CH₄ ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	2.11	S	2.07	2.08	2.06	2.06	2.10	2.07	2.09	2.06	2.05	2.05	2.06	2.06	2.06	2.11	2.06	2.13	2.12	2.26	2.11	2.12	2.20	2.51	2.05	2.51	2.11	24	
2	S	2.31	2.28	2.39	2.38	2.47	2.37	2.54	2.64	2.71	2.45	2.50	2.49	2.27	2.28	2.76	2.71	2.53	2.50	2.58	2.55	2.58	2.52	S	2.27	2.27	2.76	2.49	24
3	2.36	2.49	2.70	2.59	2.61	2.87	2.85	2.60	2.78	2.73	2.51	C	C	C	C	C	2.44	2.55	3.12	3.07	3.39	2.93	S	2.27	2.27	3.39	2.71	24	
4	5.71	2.94	2.14	2.18	2.17	2.16	2.04	2.03	2.07	2.06	Y	2.04	2.15	2.20	2.17	2.15	2.15	2.16	2.16	2.05	2.16	S	2.10	2.05	2.03	5.71	2.32	23	
5	2.01	2.04	2.03	2.02	2.02	2.03	2.02	2.07	2.19	2.04	2.20	2.13	2.05	2.06	2.04	2.03	2.18	2.40	3.02	3.34	S	2.34	2.66	2.44	2.01	3.34	2.23	24	
6	2.42	2.34	2.06	2.11	2.22	2.20	2.10	2.54	2.23	2.29	2.16	2.17	2.15	2.11	2.10	2.14	2.14	2.10	2.14	S	2.17	2.12	2.10	2.19	2.06	2.54	2.19	24	
7	2.17	2.25	2.11	2.11	2.35	2.10	2.06	2.10	2.10	2.16	2.12	2.11	2.08	2.08	2.10	2.12	2.12	2.17	S	2.28	2.63	2.54	2.43	2.48	2.06	2.63	2.21	24	
8	2.60	2.32	2.19	2.19	2.44	2.52	2.44	2.43	2.26	2.04	2.04	2.08	2.10	2.04	2.10	2.08	2.01	S	2.02	2.29	2.74	2.43	3.89	2.60	2.01	3.89	2.34	24	
9	2.44	2.24	2.42	3.17	2.43	2.53	2.52	2.59	2.37	2.56	2.47	2.06	2.07	2.03	2.09	2.21	S	2.22	3.21	2.85	2.64	3.37	3.33	3.38	2.03	3.38	2.57	24	
10	3.05	3.41	2.87	2.58	2.62	2.32	2.19	2.13	2.12	2.10	2.13	2.06	2.13	2.03	2.03	S	2.06	2.10	2.08	2.10	2.16	2.21	2.19	2.39	2.03	3.41	2.31	24	
11	2.15	2.16	2.21	2.49	2.27	2.37	2.27	X	X	2.24	2.27	2.22	2.14	2.11	S	2.03	2.05	2.03	2.02	2.07	2.08	2.06	2.08	2.05	2.02	2.49	2.16	22	
12	2.04	2.05	2.06	2.04	2.03	2.03	2.06	2.05	2.06	2.16	2.18	2.06	2.04	S	2.04	2.04	2.02	2.04	2.07	2.08	2.08	2.13	2.09	2.08	2.02	2.18	2.07	24	
13	2.11	2.13	2.13	2.13	2.36	2.16	2.23	2.28	2.18	2.09	2.17	2.09	S	2.08	2.07	2.10	2.12	2.09	2.05	2.07	2.10	2.12	2.17	2.32	2.05	2.36	2.15	24	
14	3.32	2.52	2.52	4.10	4.10	3.37	4.00	2.73	2.57	2.45	2.36	S	2.60	2.48	2.61	2.49	2.57	2.81	2.51	2.58	2.32	2.37	2.18	2.40	2.18	4.10	2.78	24	
15	2.56	2.30	2.25	2.36	2.66	2.81	3.04	3.17	2.64	2.45	S	2.32	2.42	2.55	2.32	2.28	2.45	2.14	2.44	2.31	2.35	2.36	4.09	2.39	2.14	4.09	2.55	24	
16	2.22	2.35	2.31	2.31	4.61	4.22	2.33	5.68	2.42	S	2.22	2.22	2.03	2.20	2.10	2.10	2.04	2.20	2.62	2.24	2.77	2.19	2.79	2.79	2.03	5.68	2.65	24	
17	3.67	2.56	2.61	2.22	2.41	2.50	3.21	3.06	S	2.51	2.43	2.44	2.30	2.16	2.09	2.07	2.06	2.09	2.10	2.14	2.16	2.15	2.15	2.14	2.06	3.67	2.40	24	
18	2.10	2.11	2.18	2.38	2.24	2.32	2.14	S	2.06	2.23	2.18	2.14	2.05	2.11	2.15	2.13	2.23	2.18	2.24	2.11	2.13	2.35	2.13	2.06	2.05	2.38	2.17	24	
19	2.01	1.99	2.02	2.01	1.98	1.97	S	1.98	1.98	1.98	1.98	1.98	1.97	1.96	1.96	2.04	2.04	1.96	1.95	1.95	1.96	1.96	1.98	1.98	1.95	2.04	1.98	24	
20	2.05	2.27	2.06	2.06	2.06	S	2.02	2.07	2.05	2.10	2.06	2.01	2.01	1.99	1.99	2.00	2.02	2.11	2.15	2.15	2.04	2.17	2.09	2.13	1.99	2.27	2.07	24	
21	2.14	2.46	2.51	2.71	S	2.57	2.63	2.60	2.45	2.52	2.16	2.13	2.18	2.10	2.16	2.18	2.14	2.18	2.08	2.13	2.15	2.29	2.16	2.20	2.08	2.71	2.30	24	
22	2.18	2.13	2.15	S	2.15	2.15	2.16	2.17	2.14	2.10	2.18	2.16	2.03	2.01	2.03	2.06	2.17	2.24	2.42	2.05	2.08	2.08	2.07	2.10	2.01	2.42	2.13	24	
23	2.10	2.04	S	2.08	2.15	3.15	2.83	2.88	3.12	2.39	2.30	2.17	2.21	2.11	2.13	2.14	2.10	2.03	2.28	2.16	3.28	2.57	3.24	3.21	2.03	3.28	2.46	24	
24	5.35	S	2.72	2.70	2.61	3.17	3.44	3.30	2.62	2.77	2.42	2.53	2.65	2.55	2.37	2.25	2.15	2.08	2.13	2.12	2.04	2.06	2.07	2.06	2.04	5.35	2.62	24	
25	S	2.01	2.02	2.03	2.15	2.28	2.33	2.46	2.45	2.26	2.15	2.20	2.14	2.00	2.01	2.03	2.06	2.21	2.29	2.52	2.90	2.81	2.79	S	2.00	2.90	2.28	24	
26	3.60	2.48	2.83	2.66	3.10	2.95	2.84	3.06	2.82	2.48	2.54	2.51	2.34	2.30	2.41	2.21	2.26	2.69	2.64	2.23	3.10	2.34	S	2.30	2.21	3.60	2.64	24	
27	2.27	2.47	2.41	2.29	2.38	2.38	2.29	2.11	2.08	2.14	2.18	2.18	2.14	2.17	2.04	2.07	2.05	2.03	2.02	2.03	2.04	S	2.10	2.20	2.02	2.47	2.18	24	
28	2.37	2.39	2.30	2.14	2.37	2.33	2.13	2.16	2.11	2.16	2.19	2.14	2.16	2.15	2.19	2.12	2.12	2.19	2.10	2.17	S	2.18	2.25	2.28	2.10	2.39	2.20	24	
29	2.13	2.14	2.34	2.63	2.64	2.65	2.79	3.67	2.82	2.35	2.15	2.05	2.03	2.03	2.08	2.09	2.10	2.15	2.36	S	2.55	2.39	2.83	2.54	2.03	3.67	2.41	24	
30	2.36	2.02	2.03	2.01	2.00	2.01	2.00	1.99	2.01	2.02	1.98	2.01	1.99	2.00	2.04	2.05	2.02	2.14	S	3.08	2.10	4.06	2.89	4.03	1.98	4.06	2.30	24	
31	2.84	3.24	3.46	2.54	2.87	4.08	3.86	3.15	2.32	2.27	2.24	2.36	2.32	2.02	2.04	2.06	2.06	S	2.53	2.07	2.13	2.15	2.12	2.12	2.02	4.08	2.56	24	
HOURLY MAX	5.71	3.41	3.46	4.10	4.61	4.22	4.00	5.68	3.12	2.77	2.54	2.53	2.65	2.55	2.61	2.76	2.71	2.81	3.21	3.34	3.39	4.06	4.09	4.03					
HOURLY AVG	2.64	2.35	2.33	2.38	2.48	2.56	2.51	2.61	2.34	2.28	2.22	2.18	2.17	2.14	2.13	2.14	2.16	2.21	2.32	2.31	2.38	2.39	2.47	2.40					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	703
MAXIMUM INSTANTANEOUS VALUE:	5.71 ppm @ HOUR(S) 0 ON DAY(S) 4
VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	5 hrs
OPERATIONAL TIME:	741 hrs
STANDARD DEVIATION:	0.44

METHANE MAX Instantaneous Maximum (CH₄ ppm)




Wind: LICA Bonnyville
 Poll.: LICA Bonnyville-CH4[ppm]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 14.37% Calm Avg: 2.29 [ppm]


Direction	0.0-1.1	1.1-2.1	2.1-3.2	>3.2	Total
N	0.0	5.8	4.4	0.0	10.2
NE	0.0	17.2	5.6	0.0	22.8
E	0.0	7.1	8.4	0.0	15.5
SE	0.0	11.7	6.1	0.0	17.8
S	0.0	3.7	0.4	0.0	4.1
SW	0.0	1.3	2.3	0.0	3.6
W	0.0	4.6	0.1	0.0	4.7
NW	0.0	6.3	0.7	0.0	7.0
Summary	0.0	57.6	28.0	0.0	85.6

% Icon Classes (ppm) 0 0.0-1.1 58 1.1-2.1 28 2.1-3.2 0 >3.2

LICA Bonnyville Poll.: LICA Bonnyville-CH4[ppm] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 14.37% Calm Poll Avg: 2.29[ppm]



CH4[ppm] Calibration: LICA Bonnyville Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

NON-METHANE HYDROCARBON




NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.03	0.06	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.06	0.00	24
2	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.06	0.04	0.00	0.00	0.10	0.11	0.00	0.00	0.00	0.02	0.01	0.00	S	0.00	0.11	0.02	24
3	0.00	0.00	0.00	0.00	0.00	0.01	0.01	0.00	0.01	0.01	0.00	C	C	C	C	C	0.05	0.04	0.03	0.08	0.09	0.08	S	0.01	0.00	0.09	0.02	24
4	0.14	0.10	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.14	0.01	24
5	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.04	0.03	S	0.01	0.01	0.00	0.00	0.04	0.00	24
6	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.01	0.00	24
7	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
8	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.13	0.01	0.04	0.04	0.01	0.00	0.13	0.01	24
9	0.00	0.00	0.00	0.02	0.00	0.00	0.00	0.00	0.00	0.07	0.02	0.00	0.00	0.00	0.00	0.00	S	0.00	0.04	0.02	0.00	0.07	0.05	0.03	0.00	0.07	0.01	24
10	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	24
11	0.00	0.00	0.00	0.00	0.00	0.00	0.00	X	X	0.00	0.00	0.00	0.00	0.00	S	0.00	0.38	0.06	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.38	0.02	22
12	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.03	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.03	0.00	24
13	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
14	0.02	0.02	0.02	0.11	0.04	0.03	0.13	0.05	0.01	0.01	0.04	S	0.02	0.01	0.02	0.01	0.02	0.04	0.03	0.06	0.01	0.00	0.00	0.00	0.00	0.13	0.03	24
15	0.01	0.00	0.00	0.00	0.01	0.02	0.01	0.02	0.03	0.00	S	0.01	0.00	0.01	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.03	0.01	0.00	0.03	0.01	24
16	0.00	0.00	0.00	0.00	0.02	0.04	0.00	0.06	0.02	S	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.01	0.00	0.01	0.00	0.01	0.00	0.00	0.06	0.01	24
17	0.03	0.00	0.01	0.00	0.00	0.00	0.02	0.00	S	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.03	0.00	24
18	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
19	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
20	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00	0.00	0.00	0.00	0.02	0.00	24
21	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
22	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
23	0.00	0.00	S	0.00	0.00	0.04	0.01	0.16	0.11	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.01	0.01	0.02	0.00	0.16	0.02	24
24	0.11	S	0.01	0.02	0.01	0.06	0.09	0.04	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00	0.00	0.00	0.00	0.11	0.02	24
25	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.01	0.01	S	0.00	0.02	0.00	24
26	0.04	0.00	0.01	0.00	0.01	0.01	0.02	0.01	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	S	0.00	0.00	0.04	0.01	24
27	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.01	0.00	24
28	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.04	0.00	0.00	0.00	0.00	S	0.02	0.00	0.00	0.00	0.00	0.04	0.00	24
29	0.00	0.00	0.00	0.01	0.01	0.03	0.01	0.07	0.05	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.01	S	0.02	0.01	0.01	0.00	0.00	0.00	0.07	0.01	24
30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.03	0.00	0.10	0.02	0.02	0.00	0.10	0.01	24
31	0.00	0.08	0.02	0.02	0.02	0.11	0.11	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00	S	0.00	0.00	0.02	0.00	0.00	0.00	0.00	0.11	0.02	24
HOURLY MAX	0.14	0.10	0.02	0.11	0.04	0.11	0.13	0.16	0.11	0.07	0.04	0.06	0.04	0.01	0.02	0.10	0.38	0.06	0.04	0.13	0.09	0.10	0.05	0.03				
HOURLY AVG	0.01	0.01	0.00	0.01	0.00	0.01	0.01	0.01	0.01	0.00	0.00	0.01	0.00	0.00	0.00	0.01	0.02	0.00	0.01	0.01	0.01	0.01	0.01	0.00				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

24 HR AVERAGES March 2017




MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	146				
MINIMUM 1-HR AVERAGE:	0.00 ppm	@ HOUR(S)	0	ON DAY(S)	1
MAXIMUM 1-HR AVERAGE:	0.38 ppm	@ HOUR(S)	16	ON DAY(S)	11
MAXIMUM 24-HR AVERAGE:	0.03 ppm			ON DAY(S)	14
				VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	742 hrs		
MONTHLY CALIBRATION TIME:	5 hrs	AMD OPERATION UPTIME:	99.7 %		
STANDARD DEVIATION:	0.02	MONTHLY AVERAGE:	0.01 ppm		



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.47	0.61	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.61	0.05	24
2	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	3.23	2.32	0.00	0.00	0.67	0.95	0.10	0.05	0.09	0.12	0.11	0.00	S	0.00	0.00	3.23	0.35	24
3	0.00	0.04	0.10	0.00	0.07	0.09	0.17	0.05	0.12	0.10	0.00	C	C	C	C	C	0.29	0.53	0.24	0.84	0.29	0.45	S	0.08	0.00	0.84	0.19	24	
4	0.60	0.22	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	Y	0.10	0.17	0.00	0.00	0.00	0.00	0.00	0.08	0.00	0.00	S	0.00	0.00	0.00	0.60	0.05	23	
5	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.05	0.14	0.32	S	0.10	0.12	0.00	0.00	0.32	0.03	24	
6	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.47	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.47	0.02	24	
7	0.00	0.00	0.00	0.00	0.04	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.06	0.00	0.00	0.11	0.00	0.11	0.01	24	
8	0.06	0.00	0.03	0.00	0.06	0.08	0.11	0.02	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.09	S	0.05	3.80	0.15	1.08	0.26	0.09	0.00	3.80	0.26	24	
9	0.00	0.00	0.00	0.15	0.00	0.00	0.00	0.00	0.07	0.22	0.14	0.00	0.00	0.00	0.07	0.00	S	0.00	0.50	0.12	0.09	0.28	0.19	0.20	0.00	0.50	0.09	24	
10	0.11	0.13	0.00	0.12	0.00	0.00	0.00	0.00	0.00	0.20	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.20	0.02	24	
11	0.00	0.00	0.00	0.05	0.00	0.00	0.00	X	X	0.00	0.00	0.00	0.00	0.00	S	0.00	2.75	0.94	0.00	0.00	0.00	0.00	0.00	0.00	0.00	2.75	0.18	22	
12	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.45	0.00	0.00	0.00	0.00	0.16	0.00	0.00	0.00	0.45	0.03	24	
13	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.06	0.00	0.00	0.00	0.00	0.00	0.00	0.05	0.00	0.06	0.00	24	
14	0.18	0.11	0.09	0.35	0.35	0.24	0.36	0.13	0.10	0.11	0.25	S	0.11	0.12	0.11	0.13	0.12	0.21	0.10	0.16	0.08	0.08	0.05	0.08	0.05	0.36	0.16	24	
15	0.07	0.05	0.00	0.00	0.10	0.11	0.15	0.13	0.23	0.04	S	0.11	0.05	0.13	0.06	0.10	0.15	0.00	0.08	0.05	0.06	0.06	0.40	0.07	0.00	0.40	0.10	24	
16	0.00	0.00	0.00	0.00	0.37	0.34	0.05	0.53	0.18	S	0.00	0.00	0.00	0.04	0.13	0.06	0.00	0.00	0.10	0.07	0.12	0.00	0.12	0.09	0.00	0.53	0.10	24	
17	0.29	0.07	0.09	0.00	0.00	0.04	0.14	0.11	S	0.00	0.00	0.17	0.00	0.00	0.11	0.11	0.00	0.07	0.00	0.00	0.00	0.10	0.00	0.00	0.00	0.29	0.06	24	
18	0.00	0.00	0.00	0.06	0.00	0.05	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.09	0.00	0.00	0.00	0.00	0.05	0.00	0.00	0.00	0.09	0.01	24	
19	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
20	0.00	0.05	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.22	0.07	0.00	0.00	0.67	0.22	0.00	0.00	0.00	0.67	0.05	24	
21	0.00	0.00	0.00	0.00	S	0.10	0.04	0.00	0.00	0.04	0.00	0.00	0.00	0.06	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.10	0.01	24	
22	0.00	0.00	0.00	S	0.00	0.00	0.08	0.00	0.00	0.00	0.02	0.00	0.00	0.00	0.00	0.11	0.01	0.00	0.06	0.00	0.00	0.00	0.00	0.00	0.00	0.11	0.01	24	
23	0.07	0.00	S	0.00	0.00	0.25	0.11	2.11	0.89	0.57	0.00	0.00	0.00	0.00	0.00	0.10	0.00	0.00	0.08	0.10	0.20	0.09	0.15	0.18	0.00	2.11	0.21	24	
24	0.54	S	0.12	0.13	0.09	0.21	0.27	0.16	0.04	0.60	0.03	0.00	0.00	0.06	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.23	0.00	0.00	0.00	0.60	0.11	24	
25	S	0.00	0.00	0.00	0.00	0.05	0.08	0.09	0.07	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.06	0.00	0.34	0.33	0.10	S	0.00	0.34	0.05	24	
26	0.22	0.05	0.09	0.07	0.14	0.11	0.38	0.09	0.07	0.00	0.10	0.15	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.06	0.24	0.14	S	0.00	0.00	0.38	0.08	24	
27	0.00	0.07	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.21	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.21	0.01	24	
28	0.05	0.05	0.08	0.00	0.04	0.19	0.00	0.00	0.00	0.11	0.10	0.00	0.00	0.00	0.13	0.79	0.10	0.14	0.16	0.06	S	0.88	0.00	0.00	0.00	0.88	0.13	24	
29	0.05	0.00	0.12	0.12	0.10	0.12	0.09	0.32	0.20	0.09	0.00	0.22	0.00	0.00	0.00	0.16	0.00	0.00	0.10	S	0.15	0.10	0.10	0.00	0.00	0.32	0.09	24	
30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.04	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.16	0.00	0.37	0.18	0.27	0.00	0.37	0.04	24	
31	0.11	0.21	0.17	0.11	0.17	0.29	0.24	0.13	0.00	0.00	0.00	0.15	0.00	0.00	0.00	0.61	0.00	S	0.08	0.09	0.16	0.00	0.00	0.00	0.00	0.61	0.11	24	
HOURLY MAX	0.60	0.22	0.17	0.35	0.37	0.34	0.38	2.11	0.89	0.60	0.47	3.23	2.32	0.13	0.21	0.79	2.75	0.94	0.50	3.80	0.67	1.08	0.40	0.27					
HOURLY AVG	0.08	0.04	0.03	0.04	0.05	0.08	0.08	0.14	0.07	0.07	0.04	0.16	0.09	0.01	0.03	0.11	0.18	0.07	0.06	0.20	0.09	0.17	0.06	0.04					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	257
MAXIMUM INSTANTANEOUS VALUE:	3.80 ppm @ HOUR(S) 19 ON DAY(S) 8
	VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	5 hrs
OPERATIONAL TIME:	741 hrs
STANDARD DEVIATION:	0.28

NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)



Wind: LICA Bonnyville
 Poll.: LICA Bonnyville-NMHC[ppm]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 14.37%

Calm Avg: 0.02 [ppm]

Direction	0.0-0.4	0.4-0.8	0.8-1.1	1.1-1.5	1.5-1.9	>1.9	Total
N	10.2	0.0	0.0	0.0	0.0	0.0	10.2
NE	22.6	0.1	0.0	0.0	0.0	0.0	22.8
E	15.5	0.0	0.0	0.0	0.0	0.0	15.5
SE	17.8	0.0	0.0	0.0	0.0	0.0	17.8
S	4.1	0.0	0.0	0.0	0.0	0.0	4.1
SW	3.6	0.0	0.0	0.0	0.0	0.0	3.6
W	4.7	0.0	0.0	0.0	0.0	0.0	4.7
NW	7.0	0.0	0.0	0.0	0.0	0.0	7.0
Summary	85.5	0.1	0.0	0.0	0.0	0.0	85.6

% Icon Classes (ppm)

85 0.0-0.4

0 0.4-0.8


0 0.8-1.1

0 1.1-1.5

0 1.5-1.9

0 >1.9

LICA Bonnyville Poll.: LICA Bonnyville-NMHC[ppm] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 14.37% Calm Poll Avg: 0.02[ppm]



NMHC[ppm] Calibration: LICA Bonnyville Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

OXIDES OF NITROGEN




OXIDES OF NITROGEN Hourly Averages (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	4.5	S	5.3	4.4	4.4	4.8	6.0	7.0	6.7	5.3	4.9	4.6	4.7	5.8	7.1	6.3	4.0	7.2	13.2	15.3	11.8	7.7	7.4	8.1	4.0	15.3	6.8	24	
2	S	11.2	12.3	11.4	11.1	16.0	19.5	26.5	24.4	12.7	15.8	8.7	6.6	6.6	7.7	14.0	10.7	28.3	16.1	15.9	11.6	7.1	3.9	S	3.9	28.3	13.6	24	
3	5.6	6.0	6.1	5.2	7.6	8.7	10.6	9.3	42.1	C	C	C	C	C	C	C	10.9	21.2	20.3	28.6	37.8	32.6	S	12.2	5.2	42.1	16.6	24	
4	23.4	28.8	3.7	4.0	5.7	7.1	4.4	3.6	3.6	3.9	4.7	5.5	6.8	8.2	7.8	8.5	9.4	8.5	9.0	9.1	7.6	S	4.5	3.7	3.6	28.8	7.9	24	
5	2.8	3.6	3.0	3.1	3.1	3.5	3.5	4.6	3.4	6.0	4.4	3.4	2.8	2.4	3.6	2.1	3.0	7.8	16.0	21.2	S	41.9	21.8	11.2	2.1	41.9	7.7	24	
6	12.3	7.4	2.9	4.5	7.4	8.0	9.6	8.8	11.3	10.8	10.8	8.5	8.0	8.9	6.2	6.2	8.7	7.8	6.8	S	8.5	7.0	6.6	6.4	2.9	12.3	8.0	24	
7	5.0	4.1	4.2	3.9	4.4	6.2	7.3	6.2	6.6	8.8	7.4	7.1	7.2	6.3	6.3	8.7	9.5	13.7	S	25.7	31.9	13.2	14.9	25.1	3.9	31.9	10.2	24	
8	15.2	14.4	10.0	11.1	19.0	22.2	31.1	32.1	13.8	5.3	4.5	5.6	5.5	3.0	2.9	6.8	3.9	S	10.9	14.9	21.2	21.6	19.8	23.8	2.9	32.1	13.9	24	
9	16.4	14.7	15.5	31.2	15.1	13.1	16.4	18.8	34.9	52.7	37.0	3.8	6.5	7.1	6.4	11.1	S	10.7	12.2	32.8	20.5	42.2	37.8	19.0	3.8	52.7	20.7	24	
10	25.5	29.5	21.6	23.2	20.3	13.3	11.9	15.9	13.5	10.0	8.0	5.8	5.9	5.3	5.6	S	9.5	8.9	6.1	9.0	15.1	16.9	13.1	11.5	5.3	29.5	13.3	24	
11	8.2	10.0	10.6	11.5	10.9	8.3	5.4	X	X	3.9	4.0	4.8	4.4	4.3	S	4.6	5.0	6.8	6.5	5.4	6.6	8.2	8.2	6.3	3.9	11.5	6.9	22	
12	5.8	4.8	5.1	4.7	6.3	5.8	6.2	6.2	7.1	3.9	3.7	3.9	2.5	S	4.4	4.6	12.3	1.9	2.7	1.5	1.9	5.1	2.0	1.5	1.5	12.3	4.5	24	
13	1.9	2.0	2.1	2.1	3.2	4.5	4.7	7.3	3.1	3.6	5.2	4.5	S	5.8	5.5	5.3	4.7	2.8	3.1	3.1	2.8	4.1	4.6	5.2	1.9	7.3	4.0	24	
14	9.9	11.1	14.2	36.0	18.3	13.3	49.5	50.9	16.8	9.1	8.1	S	9.9	11.9	18.9	13.2	10.3	10.7	8.9	9.1	3.7	4.8	3.8	5.1	3.7	50.9	15.1	24	
15	5.8	3.7	4.0	5.1	8.0	13.9	18.7	48.0	37.2	35.7	S	13.3	12.9	14.8	14.9	10.3	10.0	16.5	10.5	5.8	7.1	10.5	15.4	8.0	3.7	48.0	14.4	24	
16	5.8	6.2	6.1	7.6	13.8	23.2	7.0	25.3	27.4	S	8.2	5.8	5.9	3.2	4.8	6.0	7.5	11.7	29.3	12.8	12.2	7.8	9.6	6.0	3.2	29.3	11.0	24	
17	8.2	3.4	5.1	2.1	6.1	11.4	13.6	12.6	S	10.2	6.8	6.0	5.4	3.7	6.7	4.1	3.9	3.1	3.3	3.8	3.6	2.6	2.3	2.9	2.1	13.6	5.7	24	
18	2.8	2.3	2.7	3.1	3.1	3.4	2.1	S	3.5	3.5	2.7	3.0	4.2	2.9	3.4	3.3	3.3	7.7	6.0	6.1	4.7	9.6	4.6	2.9	2.1	9.6	4.0	24	
19	2.8	2.2	2.5	2.5	2.2	2.0	S	3.7	2.7	2.6	2.5	1.9	1.5	1.3	1.5	2.6	2.6	1.8	1.2	1.2	1.5	1.4	2.9	2.2	1.2	3.7	2.1	24	
20	2.2	2.4	2.2	1.8	3.2	S	6.7	5.0	3.7	4.8	3.2	3.0	2.4	1.7	2.6	2.9	2.9	8.0	5.6	4.9	10.5	3.5	2.8	3.2	1.7	10.5	3.9	24	
21	3.5	8.0	18.0	10.7	S	13.4	19.0	12.4	5.2	5.6	3.1	5.3	3.2	5.5	6.5	4.4	4.3	3.5	3.0	2.6	3.2	2.8	2.3	2.7	2.3	19.0	6.4	24	
22	3.6	3.8	3.9	S	5.3	5.5	6.8	10.4	6.4	7.7	9.3	9.8	8.4	6.5	7.1	9.6	9.5	11.2	9.3	14.2	13.2	5.2	4.5	6.9	3.6	14.2	7.7	24	
23	5.3	6.4	S	5.9	5.5	48.9	16.4	144.5	53.6	26.4	16.9	9.4	9.8	8.7	8.2	9.4	7.2	9.0	10.8	18.5	55.6	10.9	9.2	7.7	5.3	144.5	21.9	24	
24	16.8	S	13.9	20.2	39.1	49.7	94.3	75.0	29.8	13.9	8.3	7.6	8.2	5.8	8.6	7.6	6.0	5.0	4.9	4.1	3.9	4.6	4.3	3.4	3.4	94.3	18.9	24	
25	S	3.4	3.1	3.8	19.1	13.7	16.9	16.4	9.9	8.9	6.6	6.5	4.5	1.7	2.8	3.1	2.3	6.0	12.2	19.7	18.0	12.1	14.7	S	1.7	19.7	9.3	24	
26	20.1	8.7	15.1	13.5	9.8	13.0	15.6	19.4	16.2	11.1	12.0	10.2	6.5	5.1	4.7	4.2	4.3	4.7	6.7	6.9	15.5	11.8	S	4.4	4.2	20.1	10.4	24	
27	7.5	9.6	11.5	7.3	8.4	10.2	9.6	11.1	9.6	7.0	5.1	5.0	4.6	3.1	4.2	3.5	5.9	3.5	3.9	3.4	2.4	S	4.1	5.2	2.4	11.5	6.3	24	
28	6.4	6.8	5.6	3.6	4.3	4.6	4.3	5.7	3.7	4.5	6.6	6.3	5.6	4.4	12.2	5.9	5.6	5.8	7.2	5.4	S	7.7	4.7	6.2	3.6	12.2	5.8	24	
29	6.0	6.2	7.5	8.8	10.8	28.8	17.1	24.8	32.7	21.8	4.5	2.9	2.9	2.9	4.4	5.8	8.7	8.1	10.2	S	25.6	13.2	11.9	8.1	2.9	32.7	11.9	24	
30	5.2	3.1	3.0	2.8	3.5	4.6	5.8	8.0	6.1	4.8	6.6	6.7	6.1	5.0	6.7	5.1	5.7	5.0	S	16.7	10.2	21.6	11.1	7.3	2.8	21.6	7.0	24	
31	6.8	16.1	7.5	17.3	21.9	44.2	73.2	16.2	11.9	10.4	10.1	12.0	7.9	3.8	6.0	4.4	3.3	S	7.4	7.1	4.5	3.9	3.5	4.8	3.3	73.2	13.2	24	
HOURLY MAX	25.5	29.5	21.6	36.0	39.1	49.7	94.3	144.5	53.6	52.7	37.0	13.3	12.9	14.8	18.9	14.0	12.3	28.3	29.3	32.8	55.6	42.2	37.8	25.1					
HOURLY AVG	8.5	8.3	7.6	9.1	10.0	14.2	17.1	21.9	15.4	10.9	8.0	6.2	5.9	5.4	6.5	6.3	6.5	8.5	9.1	11.2	12.9	11.8	8.8	7.6					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	702		
MINIMUM 1-HR AVERAGE:	1.2 ppb	@ HOUR(S)	18 ON DAY(S)
MAXIMUM 1-HR AVERAGE:	144.5 ppb	@ HOUR(S)	7 ON DAY(S)
MAXIMUM 24-HR AVERAGE:	21.9 ppb		23 ON DAY(S)
			VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	742 hrs
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	99.7 %
STANDARD DEVIATION:	10.9	MONTHLY AVERAGE:	9.9 ppb

OXIDES OF NITROGEN Hourly Averages (NO_x ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	8.0	S	9.6	7.6	8.3	8.8	11.5	11.6	10.9	9.8	8.9	9.6	10.7	24.5	36.2	36.9	7.6	15.2	20.1	27.4	38.2	11.8	11.6	13.5	7.6	38.2	15.6	24
2	S	15.7	17.2	16.7	16.5	23.8	29.2	38.7	38.6	21.6	24.1	14.9	11.9	10.6	13.0	40.7	38.1	40.1	23.4	29.4	25.3	20.4	8.2	S	8.2	40.7	23.6	24
3	11.6	9.7	9.9	8.8	11.9	12.7	47.5	32.7	82.3	C	C	C	C	C	C	C	17.9	28.7	30.6	45.1	115.3	83.6	S	19.6	8.8	115.3	35.5	24
4	45.6	48.1	11.3	8.2	7.9	9.3	6.4	5.9	8.4	7.7	Y	8.3	36.4	16.1	17.0	13.0	13.6	11.7	12.0	13.7	13.2	S	7.2	8.8	5.9	48.1	15.0	23
5	4.5	6.4	5.0	4.7	4.6	6.4	5.9	7.5	6.3	11.4	9.2	7.4	4.6	5.0	29.7	5.1	19.4	15.5	26.5	28.9	S	53.0	49.7	14.8	4.5	53.0	14.4	24
6	16.5	13.2	6.0	7.6	14.3	13.4	14.6	12.8	15.3	14.9	15.1	12.1	9.9	18.1	8.0	8.6	11.1	9.9	9.2	S	12.1	9.2	8.6	9.7	6.0	18.1	11.7	24
7	8.2	7.2	5.8	5.7	6.9	11.8	10.9	10.5	9.2	11.5	9.6	8.9	9.6	9.6	9.0	12.1	14.3	20.5	S	34.6	44.8	18.0	20.4	30.9	5.7	44.8	14.3	24
8	29.2	18.6	13.7	15.3	32.6	41.8	52.9	45.8	21.8	9.8	9.2	9.8	8.8	5.7	17.8	54.0	32.9	S	40.8	47.4	87.1	102.9	41.0	47.0	5.7	102.9	34.2	24
9	29.2	19.6	30.2	62.9	19.2	21.0	68.8	72.3	105.2	89.4	74.6	42.0	21.6	24.9	19.7	36.0	S	22.5	46.3	123.7	70.6	90.6	107.7	43.2	19.2	123.7	54.0	24
10	116.1	51.4	29.6	34.2	31.9	19.1	29.2	24.7	16.8	12.7	13.4	14.7	9.6	8.5	8.8	S	17.7	12.3	9.1	11.2	21.8	20.3	18.1	14.3	8.5	116.1	23.7	24
11	11.0	13.1	13.0	15.5	14.0	10.8	27.8	X	X	23.0	6.6	7.4	6.8	9.2	S	7.4	6.6	9.7	8.3	9.4	10.7	11.2	10.8	8.9	6.6	27.8	11.5	22
12	7.4	6.3	6.6	6.8	10.4	7.6	8.5	9.8	9.8	7.7	5.9	5.6	4.5	S	7.1	13.9	44.6	32.3	27.0	3.8	28.4	36.3	16.0	2.5	2.5	44.6	13.4	24
13	2.7	3.0	3.3	3.9	6.0	5.8	23.1	15.6	18.7	23.0	17.1	15.2	S	17.7	26.3	14.9	8.3	15.5	43.8	15.1	8.4	5.8	26.1	8.2	2.7	43.8	14.2	24
14	26.1	19.4	24.5	46.1	42.2	36.1	79.9	87.9	40.3	12.3	27.6	S	14.9	15.7	23.2	38.7	35.8	29.5	39.8	36.7	5.6	7.1	6.1	8.2	5.6	87.9	30.6	24
15	8.2	8.2	5.6	6.9	11.8	35.8	36.8	117.5	94.0	84.7	S	53.1	36.7	61.2	58.9	39.6	42.4	60.9	20.5	8.9	14.7	16.7	58.3	18.0	5.6	117.5	39.1	24
16	10.5	11.7	8.5	9.7	29.0	53.2	24.0	69.4	51.5	S	13.4	8.9	19.3	4.4	21.9	9.6	10.1	33.6	187.1	56.5	18.8	10.4	12.5	9.4	4.4	187.1	29.7	24
17	13.6	6.5	6.6	5.4	52.4	41.9	45.5	34.4	S	51.5	18.6	7.7	17.3	7.5	20.8	11.3	23.5	11.4	21.2	6.0	14.7	4.4	2.8	4.5	2.8	52.4	18.7	24
18	15.9	3.1	3.6	3.7	3.9	9.4	2.5	S	4.5	16.0	3.1	5.0	26.7	14.0	5.6	5.2	5.7	10.1	7.0	7.5	11.1	11.8	5.8	4.3	2.5	26.7	8.1	24
19	4.0	2.7	2.9	3.4	2.9	3.1	S	4.7	3.1	2.9	3.4	2.1	2.1	1.9	2.8	4.0	3.9	3.0	1.7	1.7	1.8	1.9	5.5	3.1	1.7	5.5	3.0	24
20	3.7	3.9	4.9	3.5	6.7	S	10.1	7.8	7.0	21.6	4.6	5.3	4.2	4.8	3.9	25.0	4.1	137.0	9.6	39.0	71.2	23.5	4.6	5.5	3.5	137.0	17.9	24
21	5.0	23.1	21.4	14.1	S	27.6	22.2	42.7	14.8	14.1	15.6	23.4	11.7	29.3	33.8	8.1	6.1	18.0	4.8	7.0	4.2	3.9	2.8	3.4	2.8	42.7	15.5	24
22	4.4	4.7	4.5	S	6.3	7.8	8.1	37.0	8.6	25.9	13.3	12.8	10.8	8.8	13.5	18.2	13.0	13.7	14.3	18.4	19.0	6.8	6.0	10.1	4.4	37.0	12.4	24
23	13.4	11.7	S	11.5	26.0	95.8	33.6	236.2	140.4	33.6	25.6	14.8	14.2	10.7	10.2	30.4	10.3	71.0	20.1	87.7	289.0	16.2	23.4	26.1	10.2	289.0	54.4	24
24	67.2	S	118.8	67.9	103.6	137.9	165.6	150.9	52.7	46.9	43.4	22.3	13.5	26.7	31.3	31.9	8.2	8.7	18.3	5.6	6.2	8.5	5.6	5.3	5.3	165.6	49.9	24
25	S	4.9	3.9	5.8	34.0	21.7	28.1	35.4	17.8	12.8	8.4	22.8	12.9	3.1	16.9	22.8	4.2	34.9	30.7	29.9	70.1	40.1	27.5	S	3.1	70.1	22.2	24
26	26.5	15.0	31.4	27.0	12.7	19.2	34.7	31.9	36.5	14.6	16.2	36.3	31.3	17.1	8.2	20.0	6.6	19.2	13.2	31.5	35.3	15.0	S	7.5	6.6	36.5	22.0	24
27	10.9	13.2	15.2	10.0	12.6	16.7	12.4	15.3	13.0	12.2	27.3	33.2	23.5	16.9	20.5	21.4	20.2	24.8	30.6	6.8	5.2	S	18.1	6.7	5.2	33.2	16.8	24
28	9.2	8.6	6.7	5.4	6.5	45.6	14.2	32.2	27.4	6.8	26.9	18.7	23.1	15.5	60.2	22.4	17.7	9.7	34.2	9.1	S	14.9	7.8	9.3	5.4	60.2	18.8	24
29	8.2	9.2	9.8	13.0	19.3	40.2	51.1	54.9	58.8	46.1	32.7	4.4	4.8	11.3	6.1	8.4	26.4	11.4	20.4	S	36.6	16.9	16.5	11.7	4.4	58.8	22.5	24
30	8.9	4.7	4.1	3.9	6.6	6.4	9.2	18.0	9.6	6.3	11.8	10.2	9.7	8.8	16.7	7.2	37.1	7.4	S	23.9	26.7	90.7	21.4	11.7	3.9	90.7	15.7	24
31	11.2	26.1	20.8	26.5	40.0	60.6	164.1	52.9	24.9	29.0	27.3	35.8	20.5	15.7	31.1	23.0	20.1	S	35.1	30.2	11.5	9.4	8.6	11.3	8.6	164.1	32.0	24
HOURLY MAX	116.1	51.4	118.8	67.9	103.6	137.9	165.6	236.2	140.4	89.4	74.6	53.1	36.7	61.2	60.2	54.0	44.6	137.0	187.1	123.7	289.0	102.9	107.7	47.0				
HOURLY AVG	18.5	13.4	15.1	15.4	20.0	28.4	35.9	45.4	32.7	23.4	18.3	16.3	14.9	14.6	19.9	20.3	17.6	25.5	27.8	27.5	38.5	26.3	19.3	13.0				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	701
MAXIMUM INSTANTANEOUS VALUE:	289.0 ppb @ HOUR(S) 20 ON DAY(S) 23
VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	7 hrs
OPERATIONAL TIME:	741 hrs
STANDARD DEVIATION:	27.2

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)




Wind: LICA Bonnyville
 Poll.: LICA Bonnyville-NOX[ppb]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 14.25% Calm Avg: 20.54 [ppb]

Direction	0.0-48.2	48.2-96.4	96.4-144.6	>144.6	Total
N	10.0	0.3	0.0	0.0	10.3
NE	22.8	0.0	0.0	0.0	22.8
E	15.5	0.0	0.0	0.0	15.5
SE	17.8	0.0	0.0	0.0	17.8
S	4.1	0.0	0.0	0.0	4.1
SW	3.6	0.0	0.0	0.0	3.6
W	4.7	0.0	0.0	0.0	4.7
NW	6.8	0.1	0.0	0.0	7.0
Summary	85.3	0.4	0.0	0.0	85.8

% Icon	Classes (ppb)	85	0	0	0
	0.0-48.2		48.2-96.4		96.4-144.6
	>144.6				

LICA Bonnyville Poll.: LICA Bonnyville-NOX[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 14.25% Calm Poll Avg: 20.54[ppb]



NOX[ppb] Calibration: LICA Bonnyville Monthly: 17/03 Type: Span



■ Span Meas
 — Span Ref
 — Span Low
 — Span High

NITRIC OXIDES


NITRIC OXIDE Hourly Averages (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	1.0	S	1.1	0.9	1.0	1.4	1.3	2.4	2.7	2.6	2.6	2.8	2.8	3.2	3.8	2.8	1.4	1.5	1.9	3.0	2.9	1.5	1.0	1.2	0.9	3.8	2.0	24
2	S	1.4	1.7	1.8	2.2	2.5	3.4	6.7	10.0	6.4	8.5	5.0	3.8	3.5	3.6	6.4	3.5	7.2	2.5	2.6	2.7	1.1	0.3	S	0.3	10.0	3.9	24
3	0.2	0.5	0.5	0.3	0.4	0.5	1.5	1.8	19.6	C	C	C	C	C	C	C	2.0	4.2	2.9	4.8	12.5	7.9	S	0.8	0.2	19.6	3.8	24
4	2.3	3.0	0.0	0.4	0.6	1.1	1.1	1.0	1.3	1.5	2.2	2.3	3.3	3.3	3.3	3.2	3.5	2.3	1.3	1.6	1.4	S	0.8	0.9	0.0	3.5	1.8	24
5	0.5	0.7	0.6	0.4	0.6	0.6	0.6	0.8	1.0	2.2	2.0	1.8	1.4	1.3	1.6	1.0	0.9	2.2	2.3	2.9	S	6.9	3.9	1.7	0.4	6.9	1.6	24
6	2.1	0.7	0.4	1.1	1.7	1.2	1.7	2.2	3.4	4.2	4.6	3.9	3.9	4.9	2.9	2.8	3.4	2.7	1.8	S	2.1	1.8	1.8	2.1	0.4	4.9	2.5	24
7	1.4	1.2	1.4	1.3	1.3	2.0	2.0	2.6	4.1	3.8	3.8	4.0	3.2	3.0	4.1	4.0	4.3	S	5.0	5.7	2.5	2.3	4.5	1.2	5.7	3.0	24	
8	2.8	3.0	2.1	1.6	2.2	4.3	9.3	10.8	5.2	2.3	2.2	2.9	2.9	1.5	1.8	3.7	1.7	S	2.4	3.1	5.9	8.3	3.5	3.1	1.5	10.8	3.8	24
9	0.8	0.0	0.9	8.8	0.2	0.3	3.2	6.7	20.9	34.0	23.8	2.7	4.4	4.1	3.8	6.4	S	3.3	1.8	10.6	4.5	21.7	16.9	4.1	0.0	34.0	8.0	24
10	8.1	7.8	4.7	5.2	4.3	3.2	3.7	6.7	7.1	6.1	5.4	3.9	4.2	3.5	2.3	S	2.7	2.7	1.9	2.2	3.1	3.4	2.6	2.5	1.9	8.1	4.2	24
11	1.8	2.4	2.4	2.0	2.2	2.1	2.0	X	X	1.9	1.9	2.5	2.3	2.2	S	1.4	1.4	1.6	1.3	1.2	1.3	1.4	1.5	1.3	1.2	2.5	1.8	22
12	1.2	1.0	1.0	0.7	1.2	1.2	1.2	1.7	2.5	1.7	1.9	1.9	1.1	S	1.5	1.8	5.6	0.7	0.8	0.1	0.3	1.9	0.0	0.0	0.0	5.6	1.3	24
13	0.0	0.0	0.0	0.0	0.4	0.7	0.9	2.7	1.2	1.8	2.8	2.8	S	2.5	2.8	2.6	1.7	0.9	0.6	0.5	0.0	0.2	0.3	0.3	0.0	2.8	1.1	24
14	0.7	1.1	0.3	4.8	2.3	2.4	15.0	20.7	6.0	3.7	3.4	S	3.5	3.9	5.8	3.9	2.6	2.0	1.1	1.5	0.0	0.4	0.1	0.3	0.0	20.7	3.7	24
15	0.0	0.0	0.2	0.2	0.7	1.7	2.4	21.7	17.6	19.7	S	6.0	5.4	5.9	6.1	3.6	2.5	3.1	1.2	0.8	1.2	0.4	1.8	0.4	0.0	21.7	4.5	24
16	0.4	0.4	0.8	0.5	1.0	1.2	0.4	5.4	7.8	S	2.7	1.6	1.8	0.9	1.4	1.4	1.3	1.9	6.0	1.2	0.7	0.6	1.0	0.3	0.3	7.8	1.8	24
17	0.7	0.1	0.2	0.1	0.4	1.5	1.7	2.7	S	4.1	2.7	2.5	2.1	1.3	2.7	1.4	1.2	0.4	0.7	0.7	0.6	0.3	0.2	0.3	0.1	4.1	1.2	24
18	0.3	0.1	0.2	0.1	0.3	0.3	0.0	S	0.7	0.9	0.6	0.9	1.4	1.0	1.1	1.1	0.8	1.2	1.0	0.9	0.6	1.5	0.4	0.2	0.0	1.5	0.7	24
19	0.3	0.2	0.5	0.4	0.3	0.3	S	0.9	0.7	0.9	0.8	0.6	0.4	0.3	0.4	0.7	0.5	0.2	0.1	0.3	0.3	0.1	0.3	0.1	0.1	0.9	0.4	24
20	0.3	0.5	0.8	0.7	1.0	S	1.5	1.5	1.6	2.2	1.4	1.3	0.8	0.5	0.9	0.9	0.7	2.8	0.5	1.0	3.2	0.4	0.0	0.0	0.0	3.2	1.1	24
21	0.0	0.5	1.6	0.9	S	2.1	3.2	3.7	1.4	2.4	1.3	2.1	1.1	2.4	3.0	1.3	1.3	0.9	0.5	0.3	0.7	0.3	0.0	0.1	0.0	3.7	1.4	24
22	0.4	0.4	0.4	S	0.7	0.7	1.1	3.0	1.5	2.2	2.6	3.0	2.4	2.1	2.9	3.7	2.6	2.5	1.5	1.1	1.0	0.2	0.0	0.2	0.0	3.7	1.6	24
23	0.2	0.4	S	0.2	0.3	17.4	2.6	101.6	31.3	14.5	9.0	4.5	4.6	3.8	3.1	3.6	2.3	3.2	1.1	4.0	25.8	0.6	0.7	0.7	0.2	101.6	10.2	24
24	3.3	S	5.5	6.6	19.0	28.0	62.2	49.7	14.7	5.7	3.3	3.3	3.7	2.2	3.6	2.4	1.7	1.4	0.9	0.6	0.6	0.5	0.5	0.4	0.4	62.2	9.6	24
25	S	0.5	0.3	0.5	3.8	2.0	3.9	4.6	3.2	3.1	2.6	2.6	1.6	0.3	0.7	0.7	0.2	1.2	1.0	2.2	2.2	1.7	0.6	S	0.2	4.6	1.8	24
26	1.8	0.6	1.4	0.3	0.3	1.4	2.5	5.0	6.0	4.4	4.9	4.0	2.2	1.3	1.0	1.3	0.8	0.7	1.1	1.3	2.3	1.1	S	0.2	0.2	6.0	2.0	24
27	0.6	0.7	0.8	0.9	0.8	1.3	1.1	2.8	3.0	2.6	2.1	2.5	1.7	1.0	1.8	1.1	2.1	1.0	1.2	1.2	0.2	S	0.8	0.4	0.2	3.0	1.4	24
28	0.2	0.5	0.2	0.2	0.3	0.9	0.4	1.8	1.3	1.8	3.1	3.1	2.3	1.8	6.3	2.6	1.8	1.6	2.4	0.9	S	1.5	0.3	0.3	0.2	6.3	1.5	24
29	0.3	0.4	0.7	0.7	0.9	7.2	3.7	10.6	15.5	9.4	1.0	0.6	0.6	0.7	1.0	1.5	2.8	1.4	1.0	S	3.5	1.1	0.8	0.5	0.3	15.5	2.9	24
30	0.6	0.3	0.4	0.3	0.6	0.6	1.2	2.2	1.5	1.3	2.1	2.2	1.8	0.9	1.8	1.3	1.7	0.9	S	2.0	1.9	3.6	0.2	0.4	0.2	3.6	1.3	24
31	0.3	2.1	0.4	3.2	6.0	22.3	49.1	7.3	5.7	5.1	4.6	5.7	2.9	0.9	1.7	1.3	0.5	S	1.9	1.8	0.9	0.4	0.0	0.0	0.0	49.1	5.4	24
HOURLY MAX	8.1	7.8	5.5	8.8	19.0	28.0	62.2	101.6	31.3	34.0	23.8	6.0	5.4	5.9	6.3	6.4	5.6	7.2	6.0	10.6	25.8	21.7	16.9	4.5				
HOURLY AVG	1.1	1.1	1.1	1.5	1.9	3.7	6.1	10.0	6.8	5.3	3.8	2.9	2.6	2.2	2.6	2.4	2.0	2.1	1.5	2.0	3.0	2.5	1.5	0.9				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	682		
MINIMUM 1-HR AVERAGE:	0.0 ppb	@ HOUR(S)	2 ON DAY(S) 4
MAXIMUM 1-HR AVERAGE:	101.6 ppb	@ HOUR(S)	7 ON DAY(S) 23
MAXIMUM 24-HR AVERAGE:	10.2 ppb		ON DAY(S) 23
			VAR-VARIOUS
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	742 hrs
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	99.7 %
STANDARD DEVIATION:	6.3	MONTHLY AVERAGE:	2.9 ppb

NITRIC OXIDE Hourly Averages (NO ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

NITRIC OXIDE Instantaneous Maximum (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2.6	S	2.8	2.4	2.9	3.1	3.0	4.8	4.8	4.8	4.8	6.0	6.3	13.3	31.3	31.3	3.1	4.3	5.0	8.1	11.3	3.3	2.8	3.0	2.4	31.3	7.2	24	
2	S	3.4	3.7	4.1	4.3	5.9	7.0	12.9	18.4	10.7	13.1	7.8	6.6	5.6	6.0	21.8	11.1	12.3	5.4	12.8	8.8	5.0	1.8	S	1.8	21.8	8.6	24	
3	3.0	2.1	2.3	2.3	2.7	2.7	26.8	21.4	50.7	C	C	C	C	C	C	C	6.5	7.9	7.5	13.6	78.4	43.4	S	3.6	2.1	78.4	17.2	24	
4	8.4	9.8	1.1	2.2	2.1	2.6	2.7	2.4	3.7	3.5	Y	4.2	22.3	8.7	7.3	5.9	5.6	5.3	3.7	3.6	3.7	S	2.0	3.4	1.1	22.3	5.2	23	
5	1.6	2.0	1.9	1.3	1.7	1.6	1.6	2.2	2.9	4.8	4.4	4.5	2.8	2.6	4.4	7.8	12.5	5.6	6.4	5.7	S	13.8	11.5	4.3	1.3	13.8	4.7	24	
6	4.1	2.4	1.7	3.1	3.3	2.7	2.7	4.0	5.6	6.2	7.3	5.9	5.4	11.9	4.2	4.6	4.9	4.0	3.0	S	3.4	3.2	3.1	4.1	1.7	11.9	4.4	24	
7	3.3	3.3	2.6	2.8	2.9	5.1	4.0	4.7	4.8	6.2	5.4	5.2	6.0	5.3	4.9	6.8	11.7	8.0	S	8.4	10.8	5.2	4.0	8.6	2.6	11.7	5.7	24	
8	6.8	5.3	4.1	4.1	5.2	12.2	21.5	17.1	8.6	5.3	4.9	6.0	5.0	3.3	20.8	37.7	15.7	S	19.0	25.6	36.5	56.2	10.3	10.5	3.3	56.2	14.9	24	
9	4.2	0.9	5.6	25.8	1.3	2.6	40.8	45.2	74.5	60.4	52.0	25.0	18.3	13.3	9.4	24.0	S	11.1	19.3	70.6	33.6	54.1	62.5	14.6	0.9	74.5	29.1	24	
10	69.6	30.7	8.5	8.8	7.9	5.2	15.5	11.5	9.6	8.1	9.0	5.6	6.6	6.0	3.7	S	14.1	4.2	3.4	3.9	5.6	5.6	4.3	3.8	3.4	69.6	10.9	24	
11	3.9	3.9	3.9	3.5	4.1	3.7	14.9	X	X	12.3	3.6	4.6	4.2	5.3	S	3.2	2.6	2.9	2.8	3.1	3.7	3.2	2.9	2.8	2.6	14.9	4.5	22	
12	2.4	2.3	2.3	2.0	3.1	2.8	2.7	3.7	3.9	4.2	4.3	3.5	2.5	S	3.4	6.7	25.5	25.3	15.0	1.5	11.5	25.1	12.6	0.9	0.9	25.5	7.3	24	
13	0.8	0.8	0.8	1.2	4.8	2.8	13.7	7.5	12.3	14.8	9.9	8.9	S	12.5	15.9	10.5	3.9	14.4	13.1	7.3	6.0	1.5	8.8	1.4	0.8	15.9	7.5	24	
14	10.7	3.1	1.9	9.8	11.4	20.0	39.3	54.7	15.2	5.6	16.1	S	6.0	5.9	7.8	22.0	17.8	16.1	14.2	19.4	0.9	1.7	1.3	1.3	0.9	54.7	13.1	24	
15	1.1	1.2	1.1	1.1	2.9	23.2	9.9	77.2	58.8	55.4	S	45.4	21.7	35.1	28.5	23.3	24.7	26.3	2.2	2.1	4.2	1.5	17.8	1.4	1.1	77.2	20.3	24	
16	1.7	1.2	1.7	1.5	5.0	23.3	4.6	35.2	26.1	S	5.8	3.0	10.1	1.5	10.2	3.1	2.0	6.3	127.3	17.7	2.2	1.9	2.2	1.2	1.2	127.3	12.8	24	
17	3.9	0.5	0.9	0.8	8.6	23.0	22.1	13.1	S	34.9	11.8	3.9	9.4	3.6	12.6	4.5	13.1	5.9	15.4	2.1	9.3	1.6	1.0	1.0	0.5	34.9	8.8	24	
18	4.7	1.1	1.0	0.9	1.0	6.8	1.0	S	1.7	8.6	1.2	1.8	10.5	6.3	2.9	2.8	6.1	2.2	2.6	2.0	3.2	3.7	1.2	0.9	0.9	10.5	3.2	24	
19	0.7	1.1	1.3	1.2	0.9	1.0	S	1.7	1.3	1.5	1.6	1.3	1.4	0.9	1.2	1.8	1.6	0.9	0.9	1.3	1.1	0.9	1.5	1.1	0.7	1.8	1.2	24	
20	1.4	1.9	2.4	2.3	3.5	S	3.2	3.1	3.3	10.6	2.8	3.3	2.3	2.6	2.7	11.1	1.7	60.7	1.8	26.2	30.2	12.7	0.8	0.6	0.6	60.7	8.3	24	
21	0.8	3.2	3.6	2.5	S	6.8	5.4	24.6	6.6	6.9	6.8	12.8	9.8	17.9	18.8	3.2	3.7	9.3	2.3	4.0	2.3	1.6	1.1	1.3	0.8	24.6	6.8	24	
22	1.9	1.7	1.6	S	2.3	2.4	2.7	16.0	3.3	17.8	5.2	5.1	4.1	3.7	7.4	9.0	4.6	4.3	4.0	3.3	2.3	1.6	1.2	1.5	1.2	17.8	4.7	24	
23	2.1	1.8	S	1.5	5.1	49.3	11.1	176.5	99.6	20.2	15.0	7.6	6.8	5.5	4.8	18.9	4.8	60.7	3.3	57.6	161.6	2.7	15.5	11.5	1.5	176.5	32.3	24	
24	28.9	S	86.1	36.1	62.1	94.2	121.5	114.6	48.1	30.4	27.7	12.5	6.8	16.1	24.5	16.7	3.6	3.9	11.9	2.4	2.2	1.7	1.6	1.3	1.3	121.5	32.8	24	
25	S	1.4	1.2	1.7	8.8	4.3	8.4	10.7	6.7	5.2	4.5	8.9	5.3	1.9	9.6	8.1	1.5	15.8	5.3	6.5	28.3	21.2	3.1	S	1.2	28.3	7.7	24	
26	3.7	2.1	4.6	2.8	3.0	3.0	10.0	11.5	22.7	7.2	7.4	22.0	26.5	8.1	3.0	10.2	2.2	10.1	5.4	11.2	11.7	2.9	S	1.3	1.3	26.5	8.4	24	
27	2.4	2.2	1.9	2.3	2.3	3.1	2.5	6.4	4.7	4.8	19.8	32.5	7.7	11.1	15.7	11.1	11.4	7.4	13.1	3.7	2.0	S	7.5	2.2	1.9	32.5	7.7	24	
28	1.5	1.7	1.2	1.7	1.9	36.1	10.3	14.8	21.0	3.8	19.6	9.5	15.6	11.9	43.6	22.0	12.5	4.6	17.0	3.3	S	5.1	2.3	1.1	1.1	43.6	11.4	24	
29	1.4	1.5	1.8	2.9	2.7	15.7	26.6	30.1	33.2	29.4	15.8	1.7	1.8	6.2	2.1	3.2	12.5	2.8	4.7	S	13.9	3.1	3.3	2.0	1.4	33.2	9.5	24	
30	2.0	1.4	1.5	1.5	1.8	1.6	2.6	13.3	3.1	2.4	4.4	4.1	3.9	2.7	7.2	2.9	18.6	2.3	S	5.4	20.1	63.3	1.5	1.5	1.4	63.3	7.4	24	
31	2.6	7.1	3.8	5.6	19.1	37.1	123.7	39.8	13.9	25.5	13.5	26.6	11.0	8.2	11.5	13.1	10.1	S	19.6	24.1	6.9	3.9	0.8	0.9	0.8	123.7	18.6	24	
HOURLY MAX	69.6	30.7	86.1	36.1	62.1	94.2	123.7	176.5	99.6	60.4	52.0	45.4	26.5	35.1	43.6	37.7	25.5	60.7	127.3	70.6	161.6	63.3	62.5	14.6					
HOURLY AVG	6.3	3.5	5.3	4.7	6.3	13.5	18.7	26.9	19.6	14.2	10.6	10.0	8.5	8.2	11.2	12.0	9.0	11.9	12.2	12.3	17.8	12.1	6.6	3.2					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	701
MAXIMUM INSTANTANEOUS VALUE:	176.5 ppb @ HOUR(S) 7 ON DAY(S) 23
VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	7 hrs
STANDARD DEVIATION:	18.0
OPERATIONAL TIME:	741 hrs

NITRIC OXIDE Instantaneous Maximum (NO ppb)



NITROGEN DIOXIDE

NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	3.5	S	4.1	3.5	3.4	3.5	4.7	4.6	3.9	2.7	2.3	1.8	2.0	2.6	3.2	3.4	2.6	5.7	11.3	12.3	8.9	6.3	6.5	6.9	1.8	12.3	4.8	24
2	S	9.8	10.6	9.7	9.0	13.6	16.2	19.9	14.4	6.3	7.3	3.7	2.8	3.1	4.1	7.6	7.1	21.1	13.6	13.3	8.9	5.9	3.6	S	2.8	21.1	9.6	24
3	5.4	5.5	5.6	4.9	7.1	8.2	9.1	7.5	22.5	C	C	C	C	C	C	C	8.9	17.0	17.4	23.8	25.3	24.7	S	11.4	4.9	25.3	12.8	24
4	21.1	25.8	3.7	3.6	5.1	6.0	3.3	2.7	2.3	2.4	2.6	3.2	3.5	4.8	4.5	5.3	5.9	6.2	7.7	7.5	6.2	S	3.8	2.8	2.3	25.8	6.1	24
5	2.3	2.9	2.3	2.7	2.6	2.9	2.9	3.8	2.4	3.8	2.4	1.6	1.4	1.2	2.1	1.2	2.1	5.6	13.7	18.2	S	35.0	17.9	9.4	1.2	35.0	6.1	24
6	10.2	6.7	2.5	3.4	5.7	6.8	7.9	6.6	8.0	6.6	6.1	4.7	4.1	4.0	3.3	3.4	5.3	5.1	5.0	S	6.4	5.2	4.8	4.3	2.5	10.2	5.5	24
7	3.6	2.9	2.8	2.5	3.1	4.1	5.3	4.2	4.0	4.7	3.6	3.3	3.2	3.1	3.3	4.5	5.5	9.4	S	20.8	26.2	10.6	12.6	20.6	2.5	26.2	7.1	24
8	12.4	11.4	7.9	9.5	16.8	17.9	21.9	21.3	8.5	2.9	2.3	2.7	2.6	1.5	1.2	3.1	2.1	S	8.5	11.8	15.2	13.2	16.3	20.7	1.2	21.9	10.1	24
9	15.6	14.7	14.6	22.5	14.9	12.8	13.2	12.2	14.0	18.6	13.2	1.1	2.2	3.0	2.6	4.8	S	7.4	10.4	22.2	16.0	20.5	20.9	14.9	1.1	22.5	12.7	24
10	17.4	21.7	16.8	18.0	16.0	10.1	8.2	9.2	6.4	3.8	2.6	1.9	1.8	1.8	3.3	S	6.7	6.2	4.2	6.7	12.0	13.5	10.4	9.0	1.8	21.7	9.0	24
11	6.3	7.5	8.2	9.5	8.8	6.1	3.4	X	X	2.0	2.1	2.3	2.1	2.2	S	3.1	3.6	5.3	5.2	4.2	5.3	6.9	6.7	5.0	2.0	9.5	5.0	22
12	4.6	3.8	4.1	3.9	5.1	4.6	5.0	4.5	4.6	2.2	1.8	1.9	1.5	S	2.9	2.9	6.7	1.3	1.9	1.4	1.6	3.2	2.0	1.5	1.3	6.7	3.2	24
13	1.9	2.0	2.1	2.1	2.8	3.8	3.7	4.6	1.9	1.9	2.4	1.8	S	3.3	2.7	2.7	3.0	1.9	2.4	2.5	2.8	3.9	4.3	5.0	1.8	5.0	2.8	24
14	9.2	10.1	13.9	31.2	16.0	11.0	34.5	30.2	10.7	5.4	4.6	S	6.4	8.0	13.1	9.3	7.7	8.7	7.7	7.6	3.7	4.4	3.7	4.8	3.7	34.5	11.4	24
15	5.8	3.7	3.8	4.9	7.3	12.2	16.3	26.3	19.6	16.0	S	7.3	7.5	8.8	8.8	6.7	7.5	13.4	9.3	4.9	5.9	10.1	13.5	7.6	3.7	26.3	9.9	24
16	5.5	5.8	5.3	7.0	12.8	22.0	6.6	19.8	19.6	S	5.6	4.2	4.1	2.3	3.3	4.6	6.3	9.8	23.3	11.6	11.4	7.1	8.6	5.6	2.3	23.3	9.2	24
17	7.6	3.3	4.9	2.0	5.7	10.0	11.9	9.9	S	6.1	4.1	3.5	3.3	2.4	4.0	2.8	2.7	2.7	2.6	3.1	3.0	2.3	2.1	2.5	2.0	11.9	4.5	24
18	2.5	2.2	2.6	3.0	2.9	3.0	2.1	S	2.8	2.6	2.1	2.2	2.7	1.9	2.2	2.2	2.6	6.5	5.1	5.2	4.1	8.0	4.2	2.6	1.9	8.0	3.3	24
19	2.6	2.0	2.1	2.1	1.9	1.8	S	2.8	1.9	1.7	1.6	1.3	1.1	1.1	1.1	1.9	2.1	1.7	1.1	1.0	1.3	1.3	2.6	2.2	1.0	2.8	1.8	24
20	1.9	1.8	1.4	1.0	2.2	S	5.3	3.5	2.1	2.6	1.7	1.6	1.6	1.2	1.7	2.0	2.2	5.2	5.2	3.9	7.3	3.0	2.8	3.2	1.0	7.3	2.8	24
21	3.5	7.6	16.4	9.8	S	11.3	15.9	8.7	3.8	3.2	1.8	3.2	2.1	3.1	3.5	3.2	3.0	2.6	2.5	2.2	2.5	2.5	2.3	2.6	1.8	16.4	5.1	24
22	3.2	3.4	3.5	S	4.6	4.8	5.7	7.4	4.9	5.5	6.7	6.8	6.0	4.4	4.2	5.8	6.8	8.8	7.7	13.1	12.2	5.0	4.5	6.6	3.2	13.1	6.2	24
23	5.1	6.1	S	5.7	5.2	31.5	13.8	42.9	22.3	11.9	7.9	4.9	5.2	4.9	5.1	5.8	4.9	5.8	9.7	14.5	29.8	10.3	8.6	7.1	4.9	42.9	11.7	24
24	13.5	S	8.4	13.7	20.1	21.7	32.1	25.3	15.1	8.2	5.1	4.2	4.5	3.5	5.0	5.2	4.3	3.6	4.0	3.5	3.3	4.0	3.8	3.0	3.0	32.1	9.4	24
25	S	2.9	2.9	3.3	15.3	11.7	12.9	11.8	6.7	5.8	4.0	3.9	2.9	1.4	2.1	2.4	2.1	4.8	11.2	17.5	15.8	10.4	14.1	S	1.4	17.5	7.5	24
26	18.3	8.1	13.7	13.2	9.5	11.6	13.1	14.4	10.1	6.7	7.1	6.2	4.3	3.8	3.7	2.9	3.5	3.9	5.6	5.6	13.2	10.7	S	4.2	2.9	18.3	8.4	24
27	6.9	8.9	10.7	6.5	7.5	8.9	8.5	8.3	6.6	4.3	3.1	2.5	2.9	2.2	2.4	2.4	3.7	2.6	2.7	2.3	2.2	S	3.4	4.8	2.2	10.7	5.0	24
28	6.2	6.3	5.4	3.5	4.0	3.7	3.9	3.9	2.4	2.8	3.6	3.2	3.3	2.6	5.9	3.3	3.8	4.2	4.8	4.5	S	6.2	4.4	5.9	2.4	6.3	4.3	24
29	5.7	5.8	6.9	8.1	9.9	21.6	13.4	14.2	17.2	12.3	3.5	2.3	2.3	2.2	3.4	4.2	5.9	6.7	9.2	S	22.1	12.1	11.0	7.6	2.2	22.1	9.0	24
30	4.6	2.8	2.6	2.5	2.9	4.0	4.6	5.8	4.6	3.5	4.5	4.5	4.3	4.0	4.9	3.8	4.1	4.1	S	14.7	8.3	18.1	11.0	7.0	2.5	18.1	5.7	24
31	6.4	14.0	7.2	14.1	16.0	21.8	24.1	8.9	6.2	5.3	5.5	6.3	5.0	2.9	4.3	3.0	2.8	S	5.5	5.3	3.7	3.5	3.5	4.8	2.8	24.1	7.8	24
HOURLY MAX	21.1	25.8	16.8	31.2	20.1	31.5	34.5	42.9	22.5	18.6	13.2	7.3	7.5	8.8	13.1	9.3	8.9	21.1	23.3	23.8	29.8	35.0	20.9	20.7				
HOURLY AVG	7.3	7.2	6.6	7.6	8.1	10.4	11.0	11.9	8.6	5.6	4.2	3.4	3.3	3.1	3.9	3.9	4.5	6.5	7.5	9.1	9.8	9.2	7.4	6.7				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 159 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0				
NUMBER OF NON-ZERO READINGS:	702				
MINIMUM 1-HR AVERAGE:	1.0 ppb	@ HOUR(S)	19	ON DAY(S)	19
MAXIMUM 1-HR AVERAGE:	42.9 ppb	@ HOUR(S)	7	ON DAY(S)	23
MAXIMUM 24-HR AVERAGE:	12.8 ppb			ON DAY(S)	3
				VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	742 hrs		
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	99.7 %		
STANDARD DEVIATION:	6.0	MONTHLY AVERAGE:	7.0 ppb		

24 HR AVERAGES March 2017



NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	6.3	S	7.0	5.9	5.8	6.4	9.3	7.8	6.7	5.5	5.0	4.4	4.9	11.4	9.1	6.9	5.4	11.9	16.9	22.7	26.9	9.3	9.4	11.0	4.4	26.9	9.4	24
2	S	13.4	14.1	14.3	13.1	18.6	22.5	26.3	20.2	11.5	11.2	7.4	5.6	5.7	7.5	20.8	27.9	28.0	18.6	20.8	16.8	16.5	7.0	S	5.6	28.0	15.8	24
3	8.8	8.0	8.4	7.3	9.9	10.7	21.6	19.2	34.7	C	C	C	C	C	C	C	13.9	20.9	25.0	32.1	36.9	43.6	S	17.7	7.3	43.6	19.9	24
4	37.4	38.3	10.7	6.8	7.0	7.8	4.6	4.4	5.1	4.8	Y	5.2	17.2	8.0	9.8	7.6	9.3	8.9	9.4	10.3	10.0	S	5.7	5.9	4.4	38.3	10.6	23
5	3.3	5.2	3.7	4.1	3.6	5.3	5.0	6.1	3.9	7.0	5.4	3.3	2.5	2.7	25.5	2.4	9.7	10.4	20.2	24.3	S	39.7	38.2	12.8	2.4	39.7	10.6	24
6	12.8	11.2	4.6	5.0	11.7	12.3	9.6	10.2	9.0	8.0	6.6	5.2	6.7	4.4	4.5	6.6	6.3	6.8	S	8.9	6.8	6.0	6.3	4.4	12.8	7.9	24	
7	5.5	4.5	3.9	3.5	4.7	7.8	7.4	6.8	4.9	6.2	5.0	4.2	4.4	4.8	4.5	5.7	7.8	13.2	S	27.9	34.3	13.4	16.9	24.5	3.5	34.3	9.6	24
8	22.4	14.1	9.8	11.7	27.6	30.5	32.0	28.8	13.3	5.0	4.7	4.5	4.0	2.7	3.5	16.6	18.4	S	24.1	29.6	51.0	51.0	32.6	37.8	2.7	51.0	20.7	24
9	25.9	19.0	25.1	39.6	18.6	19.2	34.8	33.2	32.5	33.0	31.3	19.2	10.2	18.9	13.9	19.6	S	11.3	28.4	62.6	38.3	39.9	50.0	29.0	10.2	62.6	28.4	24
10	49.9	30.2	22.1	25.7	24.2	14.5	14.1	13.6	8.5	5.2	4.9	9.7	3.6	3.1	5.4	S	8.8	8.8	6.4	8.5	17.2	15.2	14.2	11.1	3.1	49.9	14.1	24
11	8.0	9.8	9.9	12.6	11.0	8.2	14.7	X	X	11.3	3.6	3.4	3.1	4.6	S	4.6	4.4	7.9	6.4	6.9	7.4	9.2	8.2	6.6	3.1	14.7	7.7	22
12	5.8	4.7	5.0	5.3	7.8	5.4	6.6	6.6	6.6	3.9	2.8	2.9	2.5	S	4.4	7.6	21.4	10.0	13.3	3.2	18.6	20.7	6.1	2.2	2.2	21.4	7.5	24
13	2.5	2.8	3.1	3.0	4.9	5.0	10.7	8.4	8.8	9.6	8.2	10.0	S	9.9	11.7	10.4	5.0	3.5	35.0	10.1	5.6	5.5	20.2	7.7	2.5	35.0	8.8	24
14	16.2	16.8	22.8	36.8	36.0	19.9	42.5	47.0	25.5	7.3	11.7	S	9.6	10.1	15.9	25.2	19.8	16.0	26.0	21.6	5.1	6.1	5.5	7.2	5.1	47.0	19.6	24
15	7.4	7.3	5.2	6.3	9.5	18.7	27.1	46.7	35.8	29.9	S	19.3	18.0	31.5	30.8	27.8	23.7	39.0	18.8	7.6	10.9	16.1	41.7	16.6	5.2	46.7	21.6	24
16	9.8	11.4	7.3	8.8	25.5	34.8	19.4	38.2	31.8	S	8.1	6.1	13.2	3.3	12.7	7.0	8.5	27.5	64.1	42.3	16.8	9.8	10.8	8.7	3.3	64.1	18.5	24
17	12.1	6.8	6.8	4.8	43.6	20.7	25.6	21.6	S	29.8	9.1	4.9	9.6	4.4	11.0	8.0	11.1	9.9	9.7	4.7	5.6	3.6	2.5	4.1	2.5	43.6	11.7	24
18	11.7	2.6	3.2	3.3	3.5	6.7	2.3	S	3.3	7.8	2.5	3.5	19.5	8.4	3.0	3.3	4.0	8.5	5.7	6.6	7.9	10.1	4.9	3.8	2.3	19.5	5.9	24
19	3.7	2.1	2.2	2.7	2.5	2.5	S	3.5	2.1	2.1	2.3	1.3	1.3	1.4	1.9	2.9	3.0	2.8	1.3	1.2	1.5	1.7	4.7	2.7	1.2	4.7	2.3	24
20	2.9	2.7	2.9	1.7	3.7	S	7.6	5.2	4.5	11.9	2.7	3.0	2.5	2.7	2.3	14.0	3.1	76.7	8.7	16.3	40.7	12.7	4.6	5.4	1.7	76.7	10.4	24
21	4.6	21.5	18.4	12.3	S	20.8	18.5	22.0	8.6	11.8	11.1	15.3	5.4	15.6	15.8	5.3	3.6	8.9	2.8	3.5	2.9	3.1	2.5	2.7	2.5	22.0	10.3	24
22	3.7	3.9	3.7	S	4.8	6.5	6.6	22.5	6.1	11.9	8.6	8.2	7.1	5.8	7.0	9.6	8.9	10.1	10.5	17.1	17.8	6.0	5.2	8.9	3.7	22.5	8.7	24
23	11.7	10.6	S	10.6	20.8	54.8	22.6	60.1	41.4	14.4	10.9	7.2	7.6	5.5	5.6	13.3	6.0	21.0	18.0	39.5	143.5	15.2	16.1	14.6	5.5	143.5	24.8	24
24	38.8	S	38.4	33.5	41.6	44.5	48.8	39.9	20.7	18.7	17.1	14.0	7.1	10.4	11.6	18.6	5.4	5.6	6.3	4.4	4.4	7.4	5.1	4.4	4.4	48.8	19.4	24
25	S	4.0	3.3	4.3	25.7	17.8	19.9	24.8	11.3	8.1	5.2	14.8	7.7	1.7	10.8	14.7	3.3	21.7	25.6	23.7	44.7	20.2	24.5	S	1.7	44.7	15.4	24
26	23.5	13.3	27.0	24.5	12.0	17.6	24.9	23.3	15.9	8.2	9.0	15.9	9.6	13.7	5.4	10.3	4.9	10.0	10.3	23.7	25.0	12.9	S	6.4	4.9	27.0	15.1	24
27	8.9	12.2	14.3	8.6	10.7	14.4	10.5	10.3	8.7	7.7	16.3	8.8	15.8	9.7	14.7	15.6	12.2	18.7	18.4	4.1	3.9	S	10.7	6.1	3.9	18.7	11.4	24
28	8.1	7.2	6.2	4.5	5.2	10.8	9.6	17.5	13.7	3.6	11.1	9.8	12.4	6.3	22.6	9.2	8.3	5.8	18.7	6.7	S	10.3	6.0	8.6	3.6	22.6	9.7	24
29	7.2	8.2	8.5	10.5	16.9	24.9	28.0	25.0	26.1	19.8	17.0	3.3	3.6	8.2	4.6	5.9	17.4	9.0	16.0	S	26.9	14.8	13.8	10.9	3.3	28.0	14.2	24
30	7.8	3.9	3.3	3.3	5.2	5.5	6.8	7.7	7.2	4.7	7.6	6.6	6.8	6.4	9.8	5.0	19.7	6.0	S	19.8	15.3	31.5	20.4	10.9	3.3	31.5	9.6	24
31	10.3	19.2	17.1	21.2	21.3	26.5	50.2	20.8	14.6	15.5	16.0	16.2	10.6	12.6	22.2	13.6	18.1	S	15.7	11.7	7.6	6.1	8.3	10.9	6.1	50.2	16.8	24
HOURLY MAX	49.9	38.3	38.4	39.6	43.6	54.8	50.2	60.1	41.4	33.0	31.3	19.3	19.5	31.5	30.8	27.8	27.9	76.7	64.1	62.6	143.5	51.0	50.0	37.8				
HOURLY AVG	13.0	10.9	10.6	11.4	14.6	16.6	18.9	20.9	14.9	11.2	9.2	8.2	8.0	8.1	10.6	10.9	10.7	15.1	16.8	17.7	22.5	15.8	13.9	10.5				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	701
MAXIMUM INSTANTANEOUS VALUE:	143.5 ppb @ HOUR(S) 20 ON DAY(S) 23
VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	7 hrs
OPERATIONAL TIME:	741 hrs
STANDARD DEVIATION:	12.1

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)




Wind: LICA Bonnyville
Poll.: LICA Bonnyville-NO2[ppb]
Monthly: 17/03
Type: PollutionRose
Direction: Blowing From (Wind Frequency)
Based On 1 Hr.

Calm: 14.25% Calm Avg: 12.91 [ppb]


Direction	0.0-14.3	14.3-28.7	28.7-43.0	>43.0	Total
N	8.0	2.1	0.1	0.0	10.3
NE	20.8	1.9	0.1	0.0	22.8
E	14.1	1.4	0.0	0.0	15.5
SE	17.7	0.1	0.0	0.0	17.8
S	4.1	0.0	0.0	0.0	4.1
SW	3.4	0.1	0.0	0.0	3.6
W	4.7	0.0	0.0	0.0	4.7
NW	6.7	0.1	0.1	0.0	7.0
Summary	79.5	5.8	0.4	0.0	85.7

% Icon Classes (ppb) 79 0.0-14.3 6 14.3-28.7 0 28.7-43.0 0 >43.0

LICA Bonnyville Poll.: LICA Bonnyville-NO2[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 14.25% Calm Poll Avg: 12.91[ppb]



NO2[ppb] Calibration: LICA Bonnyville Monthly: 17/03 Type: Span



■ Span Meas
 — Span Ref
 — Span Low
 — Span High

OZONE

OZONE Hourly Averages (O₃ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	29.4	S	31.4	30.5	29.5	28.7	26.8	26.1	26.4	28.1	29.0	29.8	30.6	31.1	31.5	32.7	33.2	29.5	24.6	21.6	26.6	27.9	27.0	26.2	21.6	33.2	28.6	24
2	S	24.5	23.7	24.6	24.3	20.0	17.7	15.0	19.3	25.0	24.6	27.3	28.7	29.0	28.9	28.3	27.6	15.5	21.6	21.6	25.4	27.5	29.5	S	15.0	29.5	24.1	24
3	29.2	28.0	27.1	27.0	24.5	23.0	21.8	22.0	11.3	15.4	23.5	25.8	20.7	30.5	30.8	29.6	22.8	18.0	16.7	9.8	5.9	6.8	S	23.1	5.9	30.8	21.4	24
4	14.6	10.7	34.0	30.2	23.8	22.4	27.6	29.9	32.2	C	C	C	Y	C	C	C	C	33.7	32.4	32.8	34.0	S	37.2	37.9	10.7	37.9	28.9	23
5	38.5	38.4	38.8	38.1	37.9	37.4	37.0	36.4	37.4	36.6	37.6	38.2	38.3	38.4	38.1	38.4	37.5	34.3	26.8	22.8	S	7.4	21.6	26.8	7.4	38.8	34.0	24
6	25.8	29.3	34.2	34.1	31.5	30.7	29.4	30.0	28.7	30.3	31.0	32.3	33.0	33.1	33.9	33.5	31.7	32.4	32.4	S	31.1	30.6	30.9	31.6	25.8	34.2	31.4	24
7	32.3	32.4	32.2	32.5	31.1	29.6	27.7	29.6	31.1	31.5	33.1	33.9	34.3	35.2	35.3	34.4	33.2	29.7	S	21.4	17.7	30.2	26.0	17.3	17.3	35.3	30.1	24
8	22.5	21.4	23.2	21.7	14.2	12.9	9.0	13.3	24.9	32.1	34.5	35.0	35.5	36.6	37.2	36.1	36.6	S	30.8	28.2	22.0	22.8	18.7	15.2	9.0	37.2	25.4	24
9	16.2	16.2	15.0	10.3	14.4	15.7	15.3	16.8	17.6	18.7	24.2	28.3	27.3	28.3	29.1	29.3	S	24.4	18.8	9.0	13.4	10.7	10.4	11.5	9.0	29.3	18.3	24
10	8.1	9.8	15.6	14.4	14.7	18.6	17.5	16.8	18.9	19.8	20.6	21.5	21.9	23.9	26.0	S	23.6	22.3	24.4	22.9	18.7	17.7	20.3	22.3	8.1	26.0	19.1	24
11	26.0	24.3	23.5	23.8	24.7	27.5	30.4	X	X	31.7	31.5	32.2	32.6	33.2	S	33.8	33.8	33.3	35.6	34.8	32.0	29.3	29.8	31.2	23.5	35.6	30.2	22
12	31.7	32.0	31.9	32.4	31.0	31.0	30.4	30.9	31.2	33.4	34.2	35.1	36.9	S	38.7	41.0	40.9	41.8	40.7	41.8	41.9	39.5	38.5	39.2	30.4	41.9	35.9	24
13	38.8	38.7	38.5	38.0	36.9	35.2	35.0	33.3	35.5	36.1	35.8	36.2	S	38.2	38.9	39.3	39.4	40.0	40.0	39.5	38.8	37.9	36.9	35.5	33.3	40.0	37.5	24
14	31.1	30.1	25.8	11.6	23.0	26.6	6.5	12.3	29.7	33.3	35.2	S	37.4	38.0	35.5	39.8	42.0	42.1	42.7	41.1	42.6	40.8	40.1	38.3	6.5	42.7	32.4	24
15	37.0	37.0	35.0	32.4	29.1	23.9	19.4	11.7	17.6	21.2	S	30.9	33.4	34.3	37.0	39.6	39.4	36.2	38.0	40.5	39.0	33.5	28.8	39.5	11.7	40.5	31.9	24
16	43.2	42.2	43.1	41.8	33.9	25.5	37.3	27.9	27.9	S	42.7	44.5	44.0	45.6	43.6	42.5	40.9	34.3	24.0	29.3	29.1	33.4	31.6	32.8	24.0	45.6	36.6	24
17	30.8	35.0	34.2	35.4	31.5	26.0	23.9	23.9	S	33.9	35.4	36.1	37.7	38.5	40.1	43.0	41.7	42.0	40.9	39.3	38.2	37.9	37.6	35.7	23.9	43.0	35.6	24
18	34.2	33.5	32.2	30.8	30.6	29.6	30.1	S	29.9	30.1	30.3	30.6	31.2	31.8	32.3	31.8	30.8	25.9	27.3	26.2	27.2	22.8	26.3	28.4	22.8	34.2	29.7	24
19	28.7	31.0	29.7	33.4	37.0	37.3	S	34.9	37.2	38.9	38.9	41.6	43.6	44.7	45.0	43.2	42.7	42.2	41.1	40.8	40.1	38.9	38.4	37.9	28.7	45.0	38.6	24
20	36.6	37.0	38.6	40.7	39.5	S	36.9	38.0	40.1	39.7	39.8	40.5	41.7	43.9	43.5	41.5	40.4	39.1	36.9	33.8	31.7	35.0	34.1	34.1	31.7	43.9	38.4	24
21	33.0	30.0	21.7	26.9	S	26.8	22.4	30.6	35.3	36.1	38.4	37.2	37.9	38.6	37.6	38.2	37.3	37.0	36.2	35.7	35.0	34.7	34.8	34.1	21.7	38.6	33.7	24
22	33.3	32.4	31.5	S	30.6	29.6	28.2	26.5	27.6	26.5	25.3	25.5	27.5	29.8	29.9	28.5	27.1	24.8	25.6	19.7	19.0	24.3	23.7	21.4	19.0	33.3	26.9	24
23	22.7	22.6	S	23.2	22.3	2.9	14.5	3.9	11.4	17.5	24.1	29.3	30.6	31.7	32.2	31.6	32.4	31.4	26.8	23.8	15.5	23.8	24.0	23.6	2.9	32.4	22.7	24
24	17.8	S	20.7	14.5	9.3	7.8	3.1	7.8	17.6	26.5	29.8	31.4	31.6	33.5	32.4	31.6	34.1	38.0	36.2	35.1	34.1	32.6	31.5	32.3	3.1	38.0	25.6	24
25	S	32.8	31.8	31.2	19.7	23.3	21.4	22.0	26.2	26.4	28.5	30.7	34.1	38.7	39.7	41.4	42.9	40.4	33.6	27.7	27.6	32.1	25.8	S	19.7	42.9	30.8	24
26	18.9	26.9	19.6	17.6	20.3	19.3	17.8	17.1	19.6	23.6	26.3	31.3	36.4	40.0	43.0	41.7	40.4	38.9	35.7	33.5	26.5	27.6	S	32.2	17.1	43.0	28.4	24
27	28.5	25.5	22.8	26.0	25.2	23.1	21.5	25.7	29.2	31.3	32.5	33.7	36.7	40.3	37.4	36.2	35.6	35.2	32.8	31.4	31.5	S	29.5	27.8	21.5	40.3	30.4	24
28	26.6	26.6	27.0	26.6	24.1	22.5	21.8	22.1	22.5	21.4	22.2	23.0	23.6	23.2	22.2	24.6	24.9	24.1	22.5	22.6	S	22.3	23.1	20.7	20.7	27.0	23.5	24
29	19.8	19.7	18.0	16.2	14.6	2.7	8.2	7.1	7.2	13.2	28.0	31.0	32.0	33.6	33.6	34.1	33.3	31.7	27.9	S	13.9	18.5	16.8	17.2	2.7	34.1	20.8	24
30	20.0	23.7	22.6	23.7	24.9	23.5	23.1	23.4	24.6	25.3	26.2	27.0	28.3	31.1	32.0	32.2	32.7	32.6	S	20.1	19.1	10.2	12.2	15.6	10.2	32.7	24.1	24
31	15.3	7.1	13.0	8.4	5.6	0.9	1.8	14.3	16.8	20.0	21.9	26.6	33.0	41.7	42.0	39.5	34.7	S	33.4	34.9	35.3	35.1	34.1	30.8	0.9	42.0	23.7	24
HOURLY MAX	43.2	42.2	43.1	41.8	39.5	37.4	37.3	38.0	40.1	39.7	42.7	44.5	44.0	45.6	45.0	43.2	42.9	42.2	42.7	41.8	42.6	40.8	40.1	39.5				
HOURLY AVG	27.3	27.5	27.9	26.6	25.3	22.8	22.1	22.4	25.3	27.7	30.5	31.9	33.1	35.1	35.4	35.8	35.0	32.8	31.3	29.0	28.0	27.3	28.2	28.3				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 82 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0				
NUMBER OF NON-ZERO READINGS:	701				
MINIMUM 1-HR AVERAGE:	0.9 ppb	@ HOUR(S)	5	ON DAY(S)	31
MAXIMUM 1-HR AVERAGE:	45.6 ppb	@ HOUR(S)	13	ON DAY(S)	16
MAXIMUM 24-HR AVERAGE:	38.6 ppb			ON DAY(S)	19
				VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs	OPERATIONAL TIME:	741 hrs		
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	99.6 %		
STANDARD DEVIATION:	8.7	MONTHLY AVERAGE:	29.0 ppb		

24 HR AVERAGES March 2017



OZONE Hourly Averages (O₃ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

OZONE Instantaneous Maximum (O₃ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	31.5	S	32.2	31.6	30.3	30.3	29.0	28.0	28.7	29.1	30.2	30.9	32.0	33.4	33.5	34.4	34.3	32.9	28.7	31.5	29.0	29.7	29.3	28.2	28.0	34.4	30.8	24	
2	S	26.3	25.6	26.4	26.4	22.5	22.1	19.4	22.9	26.3	26.6	28.7	29.5	30.1	30.2	31.9	31.7	22.5	26.4	26.1	28.0	29.5	30.4	S	19.4	31.9	26.8	24	
3	30.3	28.7	28.0	28.2	26.6	24.8	24.0	23.4	15.4	18.6	25.4	28.2	45.6	32.5	32.8	31.7	30.9	22.3	22.1	16.5	13.2	15.4	S	26.4	13.2	45.6	25.7	24	
4	24.8	30.0	35.3	33.7	25.3	26.7	29.8	33.1	33.8	C	C	C	Y	C	C	C	C	36.2	34.9	35.4	37.2	S	38.4	39.3	24.8	39.3	32.9	23	
5	39.4	40.0	39.7	39.3	38.7	38.6	38.3	38.0	38.7	38.6	39.1	39.1	39.0	39.3	39.3	39.3	39.0	37.4	32.9	29.2	S	17.7	28.0	29.0	17.7	40.0	36.4	24	
6	28.1	34.0	35.6	36.2	34.9	33.5	32.4	33.1	33.2	32.1	32.6	33.8	34.4	35.0	35.0	34.7	33.8	34.0	34.1	S	32.8	31.7	32.1	32.6	28.1	36.2	33.5	24	
7	33.1	33.1	32.9	33.4	32.9	32.3	29.1	32.6	32.9	33.4	34.3	34.9	35.7	36.5	36.6	35.7	35.3	33.1	S	26.6	29.7	33.5	28.5	21.4	21.4	36.6	32.5	24	
8	26.0	23.1	24.5	24.1	22.8	21.3	18.5	21.1	30.6	33.8	36.1	36.2	36.9	37.2	38.0	37.6	37.4	S	35.1	38.6	33.8	30.6	29.0	26.1	18.5	38.6	30.4	24	
9	20.1	22.3	22.3	15.7	17.5	17.3	17.1	20.5	21.3	21.3	30.1	30.0	29.7	29.1	30.3	31.8	S	28.4	25.6	20.7	19.2	22.2	20.8	18.5	15.7	31.8	23.1	24	
10	15.2	16.2	19.8	20.5	17.0	21.0	19.1	19.7	20.8	20.5	21.9	22.1	23.2	25.2	27.5	S	25.8	24.6	25.7	24.6	23.7	21.0	22.5	25.7	15.2	27.5	21.9	24	
11	27.7	26.7	24.9	26.1	27.6	29.1	32.0	X	X	33.1	32.7	32.9	33.5	34.6	S	35.0	34.8	35.7	37.1	36.8	33.7	31.7	31.5	32.3	24.9	37.1	31.9	22	
12	32.8	33.4	33.4	34.0	32.8	32.2	31.7	32.5	32.5	34.9	35.0	36.6	38.1	S	40.0	40.0	44.4	44.2	43.3	41.9	42.6	42.9	42.2	39.3	39.9	31.7	44.4	37.4	24
13	39.3	39.3	39.2	38.8	38.6	36.6	36.5	36.1	36.5	37.1	37.0	37.2	S	39.5	40.3	40.5	40.5	40.7	40.8	40.5	39.8	39.6	38.7	39.0	36.1	40.8	38.8	24	
14	35.1	35.3	29.1	19.4	30.6	30.6	28.5	24.3	33.5	35.1	36.6	S	39.3	39.6	39.0	43.2	43.5	44.1	44.7	43.8	43.5	42.6	41.2	40.2	19.4	44.7	36.6	24	
15	38.7	38.4	36.5	34.6	31.6	27.9	26.1	23.7	23.4	26.0	S	34.0	35.9	37.2	40.2	42.6	43.9	44.8	41.9	42.3	42.0	37.4	35.4	45.5	23.4	45.5	36.1	24	
16	45.6	45.2	45.0	43.8	41.6	39.9	42.7	40.6	40.2	S	45.0	46.2	45.3	47.0	45.3	44.7	44.0	43.2	33.8	34.4	35.7	35.9	34.5	36.3	33.8	47.0	41.6	24	
17	36.8	37.3	36.9	36.9	33.4	32.8	27.0	26.7	S	37.4	38.6	39.2	39.1	40.8	42.3	44.3	43.4	42.6	42.2	40.2	39.0	38.8	38.3	37.4	26.7	44.3	37.9	24	
18	35.4	34.3	33.5	32.3	31.7	30.9	30.6	S	30.7	30.7	30.9	31.5	32.5	32.9	33.2	33.1	32.5	29.2	28.8	27.7	28.8	26.4	28.1	29.5	26.4	35.4	31.1	24	
19	30.9	31.8	32.0	35.3	39.0	38.4	S	36.8	38.0	39.9	41.6	42.9	44.7	45.8	46.1	45.1	43.9	42.8	42.2	41.7	40.8	40.3	39.8	39.4	30.9	46.1	40.0	24	
20	37.5	39.0	39.5	42.1	41.2	S	38.8	40.2	41.1	41.2	41.0	42.0	43.2	45.3	45.5	43.0	42.0	41.6	41.4	36.3	36.2	36.6	36.2	37.1	36.2	45.5	40.3	24	
21	35.6	35.4	25.6	29.8	S	31.2	27.6	35.4	36.6	38.3	39.6	38.8	39.1	39.7	39.9	39.6	38.6	38.6	37.2	36.8	35.7	35.9	35.1	25.6	39.9	35.9	24		
22	34.4	33.8	32.9	S	31.7	31.2	29.9	28.7	29.0	28.8	27.5	27.5	30.4	31.0	31.5	31.3	29.1	26.9	27.9	24.0	25.7	25.8	24.5	23.8	23.8	34.4	29.0	24	
23	26.7	26.7	S	25.4	24.6	13.8	21.9	8.7	16.5	20.4	27.0	31.7	32.0	32.8	33.4	33.4	33.8	33.5	32.5	30.7	26.7	28.5	27.5	28.1	8.7	33.8	26.8	24	
24	25.3	S	26.7	26.4	20.5	16.0	10.6	24.0	28.1	30.4	31.8	32.6	33.3	34.4	35.3	33.8	37.5	39.3	38.3	36.8	35.1	34.4	32.6	33.5	10.6	39.3	30.3	24	
25	S	33.8	32.8	32.6	30.1	27.0	25.2	24.5	28.1	30.5	30.1	34.6	37.8	39.7	41.2	42.7	44.1	43.8	41.2	33.7	35.9	36.8	31.7	S	24.5	44.1	34.5	24	
26	24.5	28.8	26.9	21.3	22.5	22.2	20.7	19.9	22.0	29.0	29.1	35.4	39.3	42.0	44.6	43.6	42.0	41.1	39.4	36.2	33.8	29.6	S	33.7	19.9	44.6	31.6	24	
27	31.4	29.1	25.6	27.5	27.5	26.9	23.7	28.8	31.2	33.7	33.8	35.7	40.0	41.6	40.2	37.8	37.6	37.6	35.7	33.1	33.2	S	30.6	30.0	23.7	41.6	32.7	24	
28	28.1	27.8	28.0	28.5	25.7	24.1	23.3	23.5	23.7	22.9	23.7	24.4	24.8	23.8	23.8	26.4	26.4	26.3	24.9	25.4	S	24.3	25.5	23.1	22.9	28.5	25.1	24	
29	21.5	21.5	20.1	18.2	18.8	7.5	14.3	14.8	12.3	23.3	30.6	32.2	33.7	34.7	35.1	35.6	35.4	34.5	32.9	S	22.1	22.5	18.9	19.7	7.5	35.6	24.4	24	
30	24.1	24.4	23.8	25.8	26.3	25.4	25.8	25.7	26.3	26.7	28.2	29.2	31.2	33.5	35.1	34.4	34.4	34.4	S	25.8	25.1	20.8	18.1	18.9	18.1	35.1	27.1	24	
31	18.5	12.7	18.6	11.5	11.4	2.7	4.8	17.1	19.2	21.7	25.3	29.3	38.8	48.0	47.1	42.3	37.2	S	37.4	37.8	36.9	36.5	36.3	34.3	2.7	48.0	27.2	24	
HOURLY MAX	45.6	45.2	45.0	43.8	41.6	39.9	42.7	40.6	41.1	41.2	45.0	46.2	45.6	48.0	47.1	45.1	44.2	44.8	44.7	43.8	43.5	42.6	41.2	45.5					
HOURLY AVG	30.3	30.6	30.2	29.3	28.7	26.5	26.0	26.9	28.5	30.2	32.5	33.7	35.8	36.6	37.3	37.7	37.1	35.7	34.7	33.0	32.4	31.0	31.2	31.2					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	701
MAXIMUM INSTANTANEOUS VALUE:	48.0 ppb @ HOUR(S) 13 ON DAY(S) 31
VAR-VARIOUS	
IZS CALIBRATION TIME:	33 hrs
MONTHLY CALIBRATION TIME:	7 hrs
STANDARD DEVIATION:	7.6
OPERATIONAL TIME:	741 hrs

OZONE Instantaneous Maximum (O₃ ppb)



Wind: LICA Bonnyville
Poll.: LICA Bonnyville-O3[ppb]
Monthly: 17/03
Type: PollutionRose
Direction: Blowing From (Wind Frequency)
Based On 1 Hr.


Calm: 14.47%

Calm Avg: 20.59 [ppb]


Direction	0.0-15.2	15.2-30.5	30.5-45.7	>45.7	Total
N	1.0	4.3	4.0	0.0	9.3
NE	1.2	11.6	10.0	0.0	22.8
E	0.3	10.7	4.6	0.0	15.6
SE	0.0	5.2	12.8	0.0	17.9
S	0.0	0.7	3.7	0.0	4.4
SW	0.0	0.9	2.9	0.0	3.7
W	0.0	1.0	3.7	0.0	4.7
NW	0.1	1.6	5.3	0.0	7.0
Summary	2.6	36.0	47.0	0.0	85.5

% Icon	Classes (ppb)	3	0.0-15.2	36	15.2-30.5	47	30.5-45.7	0	>45.7
--------	---------------	---	----------	----	-----------	----	-----------	---	-------

LICA Bonnyville Poll.: LICA Bonnyville-O3[ppb] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 14.47% Calm Poll Avg: 20.59[ppb]



O3[ppb] Calibration: LICA Bonnyville Monthly: 17/03 Type: Span



Span Meas Span Ref Span Low Span High

PARTICULATE MATTER 2.5

PARTICULATE MATTER < 2.5 MICRONS Hourly Averages (PM_{2.5} µg/m³)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	4.7	0.0	0.0	1.1	2.7	0.0	5.8	1.0	6.5	5.5	5.6	2.6	0.6	5.9	7.7	1.8	0.5	4.5	X	1.2	3.9	6.7	0.3	9.5	0.0	9.5	3.4	23	
2	4.2	4.4	10.1	4.7	0.5	3.4	4.2	6.4	6.8	6.1	2.4	3.4	7.0	3.1	6.7	11.4	14.9	12.1	13.4	10.4	11.8	14.9	11.4	9.4	0.5	14.9	7.6	24	
3	7.7	10.3	3.4	2.0	4.1	5.8	9.1	9.2	6.9	12.7	7.7	8.4	8.7	C	C	C	C	20.3	8.6	0.0	7.1	X	8.4	1.3	0.0	20.3	7.5	23	
4	17.4	X	X	X	10.0	X	2.8	3.5	0.0	0.0	0.0	0.0	0.0	0.0	5.2	9.0	6.9	X	X	X	9.8	7.9	1.6	0.0	0.0	17.4	4.4	17	
5	X	0.3	4.1	2.6	1.1	0.0	0.1	0.0	4.0	2.8	2.0	0.0	1.4	0.0	X	0.8	3.4	2.1	X	9.5	10.9	9.6	0.7	6.7	0.0	10.9	3.0	21	
6	7.7	5.7	0.5	0.0	11.6	4.0	3.5	15.6	9.3	6.5	0.0	1.3	0.0	0.2	5.5	3.6	0.1	0.1	3.2	3.1	0.0	6.0	0.0	9.0	0.0	15.6	4.0	24	
7	5.1	11.1	0.0	X	0.2	12.8	9.6	12.1	11.1	X	6.2	4.6	6.6	3.6	5.7	2.5	4.1	4.7	15.7	13.1	9.1	8.6	4.2	X	0.0	15.7	7.2	21	
8	0.0	4.7	4.1	11.1	10.1	5.6	7.2	2.6	8.6	6.6	0.1	5.6	6.6	3.1	1.6	4.7	0.0	0.0	3.6	13.1	0.1	7.7	4.1	0.0	0.0	13.1	4.6	24	
9	9.1	1.1	5.6	11.7	0.0	6.6	11.2	13.1	6.2	13.1	23.6	1.1	0.0	2.6	10.6	8.7	3.7	3.1	X	15.6	4.1	15.2	0.0	1.6	0.0	23.6	7.3	23	
10	14.6	14.1	2.6	8.7	X	12.1	6.2	X	10.1	0.7	7.7	9.6	6.1	11.1	14.5	3.8	0.0	5.8	9.1	8.9	6.6	0.0	3.9	15.1	0.0	15.1	7.8	22	
11	6.5	2.3	15.9	2.3	0.0	X	7.2	X	X	8.1	7.2	7.7	5.1	2.6	1.1	0.0	0.0	4.1	X	1.1	4.7	10.6	0.7	0.0	0.0	15.9	4.4	20	
12	0.0	7.7	4.7	0.0	0.0	6.1	8.6	12.1	11.1	3.6	1.6	5.6	0.1	3.6	4.6	3.1	0.7	1.6	2.6	6.2	X	X	1.1	9.1	0.0	12.1	4.3	22	
13	8.6	11.6	12.6	11.1	X	0.0	X	0.0	9.1	0.0	3.1	2.6	4.1	3.6	4.6	8.1	7.2	3.6	5.6	3.1	4.7	14.6	10.6	1.6	0.0	14.6	5.9	22	
14	9.1	10.6	6.6	12.6	12.6	6.6	5.1	15.6	7.2	7.7	8.6	9.6	8.6	12.1	16.1	13.1	16.1	11.7	12.7	12.6	8.6	8.1	7.2	10.1	5.1	16.1	10.4	24	
15	6.1	5.6	1.1	7.7	2.1	12.6	10.1	15.6	18.6	17.1	14.1	10.1	11.6	11.1	8.6	2.2	8.1	5.6	2.6	5.1	2.6	0.7	1.6	2.2	0.7	18.6	7.6	24	
16	0.0	X	1.1	2.2	4.7	5.6	3.1	4.1	0.7	0.7	2.6	5.6	0.0	6.2	7.2	1.6	4.7	7.7	7.7	5.1	3.1	3.1	4.1	3.1	0.0	7.7	3.7	23	
17	5.1	5.6	6.6	1.1	2.2	4.7	3.6	10.1	7.2	7.7	5.6	2.2	11.1	5.1	11.1	8.6	10.1	10.1	6.6	6.2	5.6	6.2	7.2	6.6	1.1	11.1	6.5	24	
18	3.1	3.1	3.1	5.1	4.1	2.6	0.1	6.2	0.0	2.2	1.1	5.1	3.6	7.2	X	6.6	3.1	2.6	6.2	3.1	1.6	0.1	2.2	6.2	0.0	7.2	3.4	23	
19	0.7	2.2	1.1	3.1	2.6	0.0	5.6	0.0	1.6	2.2	4.7	1.6	1.6	0.0	3.1	1.6	2.2	0.7	3.1	3.6	0.1	3.1	0.0	2.1	0.0	5.6	1.9	24	
20	0.7	1.1	0.0	0.7	6.2	1.1	2.6	13.7	4.1	0.6	1.6	3.6	2.1	3.6	2.6	3.6	4.7	3.1	6.1	6.2	2.2	7.2	X	4.1	0.0	13.7	3.5	23	
21	9.6	5.1	X	9.1	X	13.1	10.1	3.1	3.6	4.6	5.1	7.2	3.1	8.1	6.6	6.2	4.7	3.1	4.6	7.2	6.6	4.6	0.0	3.6	0.0	13.1	5.9	22	
22	2.6	4.7	1.1	4.1	2.6	3.6	10.6	7.7	6.6	8.1	11.6	7.2	5.6	0.7	5.6	6.6	3.6	4.6	8.6	8.6	3.6	8.1	1.6	12.1	0.7	12.1	5.8	24	
23	8.1	0.7	0.0	1.1	2.2	3.6	8.1	10.1	4.7	5.6	4.1	4.1	3.1	5.6	4.1	8.2	7.2	5.1	9.1	9.1	14.6	12.6	4.7	3.6	0.0	14.6	5.8	24	
24	6.6	7.7	5.1	8.1	7.2	23.1	12.7	23.1	15.2	10.1	1.6	5.1	6.2	7.2	10.1	8.2	7.7	6.2	6.6	7.7	7.2	4.1	3.1	4.1	1.6	23.1	8.5	24	
25	5.6	4.1	3.1	2.6	0.7	3.6	5.1	5.1	3.6	7.2	5.1	24.1	8.2	6.6	9.6	10.6	11.7	9.6	9.6	7.7	13.7	9.1	13.7	6.2	0.7	24.1	7.8	24	
26	12.7	8.7	6.6	11.6	14.1	10.6	12.1	12.6	15.2	13.7	13.7	12.1	13.7	9.1	7.7	7.7	6.2	10.6	7.2	6.6	3.1	10.6	4.7	3.7	3.1	15.2	9.8	24	
27	5.6	3.6	3.1	2.6	5.1	5.1	6.2	5.6	4.6	8.1	9.1	8.2	7.2	6.6	11.1	1.1	4.1	4.7	7.2	7.7	6.6	6.6	7.7	4.7	1.1	11.1	5.9	24	
28	3.7	8.2	5.6	6.2	10.1	11.1	15.2	10.6	9.6	13.7	12.7	8.7	11.7	C	C	3.6	7.2	6.2	5.6	6.2	7.2	10.6	12.1	11.2	3.6	15.2	9.0	24	
29	8.7	5.6	11.6	11.2	9.6	12.1	8.6	14.1	11.7	14.1	10.6	10.6	X	15.6	X	X	X	15.6	16.6	22.6	22.1	20.2	12.1	13.7	10.1	5.6	22.6	13.2	21
30	7.7	6.6	0.7	4.7	4.1	0.0	9.1	6.6	3.1	0.0	0.1	1.1	0.0	6.6	3.7	9.6	8.7	1.6	14.1	4.7	8.7	7.7	10.1	2.6	0.0	14.1	5.1	24	
31	5.1	6.6	7.2	8.6	12.6	11.1	13.1	7.7	12.6	14.6	17.1	12.1	12.1	14.1	7.2	4.1	4.1	6.6	13.1	13.1	12.7	6.2	11.1	8.7	4.1	17.1	10.1	24	
HOURLY MAX	17.4	14.1	15.9	12.6	14.1	23.1	15.2	23.1	18.6	17.1	23.6	24.1	13.7	15.6	16.1	13.1	16.1	20.3	22.6	22.1	20.2	15.2	13.7	15.1					
HOURLY AVG	6.2	5.6	4.4	5.4	5.1	6.4	7.2	8.5	7.3	6.8	6.3	6.2	5.2	5.7	7.0	5.5	5.7	5.9	8.3	7.6	6.7	7.7	5.1	5.6					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT:	1-HR	80 µg/m ³	24-HR	30 µg/m ³
----------------------	------	----------------------	-------	----------------------


24 HR AVERAGES March 2017



MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0			
NUMBER OF 24-HR EXCEEDANCES:	0			
NUMBER OF NON-ZERO READINGS:	651			
MINIMUM 1-HR AVERAGE:	0.0 µg/m ³	@ HOUR(S)	1	ON DAY(S)
MAXIMUM 1-HR AVERAGE:	24.1 µg/m ³	@ HOUR(S)	11	ON DAY(S)
MAXIMUM 24-HR AVERAGE:	13.2 µg/m ³			ON DAY(S)
				VAR-VARIOUS
MONTHLY CALIBRATION TIME:	6	hrs	OPERATIONAL TIME:	710
			AMD OPERATION UPTIME:	95.4
STANDARD DEVIATION:	4.6		MONTHLY AVERAGE:	6.3 µg/m ³

PARTICULATE MATTER < 2.5 MICRONS Hourly Averages (PM_{2.5} µg/m³)



Wind: LICA Bonnyville
 Poll.: LICA Bonnyville-PM25[ug/m3(L)]
 Monthly: 17/03
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 14.41%

Calm Avg: 8.46 [ug/m3(L)]

Direction	0.0-4.8	4.8-9.7	9.7-14.5	14.5-19.4	19.4-24.2	>24.2	Total
N	5.6	2.6	1.1	0.1	0.0	0.0	9.4
NE	12.8	6.7	2.3	0.7	0.4	0.0	23.0
E	5.3	6.0	3.7	0.3	0.1	0.0	15.4
SE	5.7	8.7	2.7	0.9	0.0	0.0	18.0
S	2.0	1.4	1.1	0.0	0.0	0.0	4.6
SW	0.4	1.7	1.0	0.4	0.0	0.0	3.6
W	4.0	0.4	0.3	0.0	0.1	0.0	4.9
NW	3.7	2.3	0.7	0.1	0.0	0.0	6.8
Summary	39.5	29.8	13.0	2.6	0.7	0.0	85.6

% Icon	Classes (ug/m3(L))	40	0.0-4.8	30	4.8-9.7	13	9.7-14.5	3	14.5-19.4	1	19.4-24.2	0	>24.2
--------	--------------------	----	---------	----	---------	----	----------	---	-----------	---	-----------	---	-------

LICA Bonnyville Poll.: LICA Bonnyville-PM25[ug/m3(L)] 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 14.41% Calm Poll Avg: 8.46[ug/m3(L)]



WIND SPEED




WIND SPEED Hourly Averages (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY 1	12.8	13.7	12.2	10.4	11.5	11.8	10.4	13.5	12.9	12.9	14.0	14.9	12.5	10.9	9.2	8.2	8.5	3.9	2.9	3.3	6.3	7.0	5.5	5.6	2.9	14.9	9.7	24
2	4.3	4.1	4.7	6.9	6.2	4.4	4.9	3.3	3.4	6.5	6.1	8.7	11.5	9.8	7.3	4.4	6.7	6.2	7.7	6.0	7.9	10.8	13.2	13.5	3.3	13.5	5.9	24
3	9.8	10.2	10.1	11.2	8.0	6.2	8.1	10.1	3.9	1.7	C	C	C	3.7	14.3	9.7	3.5	4.9	5.5	0.6	1.3	3.7	4.9	0.5	0.5	14.3	3.7	24
4	1.7	2.2	10.4	11.6	16.9	14.0	16.1	14.5	15.3	16.2	14.6	14.4	10.6	11.6	12.8	9.0	9.0	8.1	6.3	6.9	6.3	8.3	7.8	8.4	1.7	16.9	10.3	24
5	8.3	10.1	8.5	7.0	7.6	8.5	7.0	5.8	6.2	6.6	5.0	6.2	6.0	3.9	5.1	2.9	1.3	3.7	2.6	3.1	0.1	1.4	4.8	5.2	0.1	10.1	4.9	24
6	7.2	6.3	7.6	7.5	7.2	6.1	5.8	4.9	4.2	10.9	11.9	11.7	12.6	12.1	12.8	14.1	11.5	13.4	12.2	13.7	13.9	14.1	13.8	13.5	4.2	14.1	9.8	24
7	14.3	18.9	16.2	16.6	14.6	14.5	13.3	17.7	16.2	14.1	14.7	15.5	14.8	14.9	15.2	14.4	12.3	10.3	7.9	6.5	7.1	8.2	7.3	5.3	5.3	18.9	12.6	24
8	5.6	6.5	5.6	2.2	1.6	1.5	3.9	5.4	6.7	6.6	6.9	6.6	3.6	3.8	3.2	1.5	4.9	4.0	0.7	0.2	1.3	0.3	3.4	2.8	0.2	6.9	2.8	24
9	1.4	1.5	0.9	1.8	0.8	0.1	0.7	0.3	0.2	0.4	1.5	1.2	1.9	5.5	4.9	4.0	3.4	2.0	0.1	0.7	1.2	1.1	1.0	1.6	0.1	5.5	0.5	24
10	2.7	6.1	5.9	5.1	6.6	8.7	9.0	10.2	10.4	10.0	11.6	12.9	13.8	13.4	11.5	11.0	10.3	7.7	7.2	6.3	5.2	6.0	4.8	4.8	2.7	13.8	8.3	24
11	6.0	5.9	7.5	5.3	6.6	7.2	6.5	X	X	6.6	7.2	8.1	9.3	10.1	11.1	10.0	9.5	11.0	10.7	9.5	10.1	8.3	9.8	9.9	5.3	11.1	7.9	22
12	9.9	9.9	8.6	7.8	8.3	9.3	6.9	8.2	9.2	9.1	9.3	8.4	8.9	6.6	8.7	5.8	16.3	14.3	12.9	15.0	14.3	16.8	10.8	14.9	5.8	16.8	6.8	24
13	13.6	9.3	10.0	12.1	8.6	10.4	9.9	9.4	13.3	14.9	13.9	15.8	15.8	17.4	18.0	17.2	18.4	19.6	16.8	14.0	10.2	5.4	8.8	6.7	5.4	19.6	12.7	24
14	3.1	2.8	1.2	2.0	2.3	4.3	3.4	1.2	3.0	3.4	0.2	2.2	0.3	5.0	4.8	7.8	10.3	8.4	10.6	9.4	15.5	12.1	10.5	7.1	0.2	15.5	4.6	24
15	6.6	13.1	11.3	9.7	7.5	6.7	2.7	3.2	4.8	1.9	1.2	1.6	2.3	2.2	6.9	8.5	6.2	2.4	7.4	9.2	8.8	3.8	0.6	5.6	0.6	13.1	0.7	24
16	8.7	6.9	8.4	6.7	1.0	0.7	1.2	0.4	1.5	2.0	5.6	8.1	6.5	5.6	6.9	3.5	4.1	0.4	2.6	2.7	1.5	3.9	3.3	2.8	0.4	8.7	3.4	24
17	1.5	1.5	4.5	4.6	4.8	1.5	1.4	1.2	8.9	6.6	8.0	10.7	10.5	14.0	11.8	17.4	15.0	18.2	18.7	16.6	19.4	18.3	15.0	13.8	1.2	19.4	9.5	24
18	18.4	12.8	9.2	11.6	9.1	9.7	14.2	11.1	12.9	12.0	11.5	10.8	12.7	11.1	12.0	10.6	9.3	8.4	10.6	9.9	7.4	9.3	16.3	13.5	7.4	18.4	7.7	24
19	13.7	18.0	20.6	21.5	17.2	16.9	14.7	13.1	14.1	15.0	14.0	14.9	16.5	15.2	12.7	17.0	15.6	12.5	11.1	11.3	11.5	9.0	9.7	13.4	9.0	21.5	14.0	24
20	15.0	17.7	16.9	20.8	15.5	13.5	15.1	16.7	18.3	12.7	10.5	7.1	7.1	9.6	9.7	9.5	8.4	6.5	3.3	0.4	0.8	0.0	1.2	1.2	0.0	20.8	9.3	24
21	3.0	2.4	3.3	8.3	6.8	7.2	5.8	9.4	10.2	15.1	20.3	15.3	17.6	17.1	15.9	14.9	16.8	18.0	19.0	16.7	14.2	13.9	14.9	12.8	2.4	20.3	12.1	24
22	12.2	10.1	10.8	11.7	11.2	9.7	9.1	10.4	9.9	7.9	6.7	8.3	8.2	10.2	10.8	8.4	5.2	4.9	5.5	2.7	3.6	5.5	4.0	2.7	2.7	12.2	5.2	24
23	6.0	4.1	3.2	1.3	1.0	2.4	0.6	1.2	3.1	4.5	4.8	2.2	4.9	5.8	3.8	3.3	0.8	0.7	0.7	1.2	1.1	4.0	0.5	0.8	0.5	6.0	1.3	24
24	1.5	0.3	0.6	0.9	1.1	0.3	1.3	1.2	3.1	5.8	8.8	12.6	8.1	6.4	6.1	7.0	11.2	13.7	11.4	12.3	9.2	7.8	9.4	8.8	0.3	13.7	5.7	24
25	9.2	9.5	8.4	8.1	6.0	4.7	4.9	4.8	4.9	3.2	5.0	4.6	8.0	7.9	7.6	5.2	1.7	1.0	2.4	4.8	4.5	5.0	1.9	3.7	1.0	9.5	1.2	24
26	3.1	8.2	2.6	1.3	0.6	7.2	3.0	3.9	1.6	0.4	2.0	1.4	6.4	5.9	6.3	10.1	6.0	4.2	3.3	5.6	3.6	6.4	7.3	8.1	0.4	10.1	1.7	24
27	7.4	5.1	6.1	7.1	5.6	7.5	5.9	9.4	6.8	6.4	8.6	9.3	7.2	10.5	11.1	8.6	7.9	8.3	8.0	13.0	12.1	9.2	6.1	5.0	5.0	13.0	6.4	24
28	2.2	3.3	5.4	7.9	5.3	7.8	10.5	8.6	10.3	7.7	7.4	9.0	9.5	10.6	8.2	8.0	6.7	6.0	9.1	6.6	8.8	5.9	3.8	3.4	2.2	10.6	6.8	24
29	4.0	4.8	2.5	1.6	1.1	0.2	0.9	1.2	0.6	0.6	1.4	2.5	5.2	1.4	3.3	1.8	0.9	3.9	2.1	2.6	2.9	5.4	4.8	3.6	0.2	5.4	1.4	24
30	5.5	7.9	7.9	9.4	9.3	8.8	9.8	10.0	10.8	9.8	10.1	9.5	8.4	7.6	6.0	4.8	3.9	2.6	1.9	2.7	0.6	1.8	1.6	0.6	0.6	10.8	5.9	24
31	0.6	1.8	0.7	5.1	1.4	1.2	0.5	1.3	8.2	10.0	4.9	9.1	16.9	19.5	18.5	19.0	19.4	14.4	4.9	9.4	15.6	19.8	22.7	16.7	0.5	22.7	8.8	24
HOURLY MAX	18.4	18.9	20.6	21.5	17.2	16.9	16.1	17.7	18.3	16.2	20.3	15.8	17.6	19.5	18.5	19.0	19.4	19.6	19.0	16.7	19.4	19.8	22.7	16.7				
HOURLY AVG	2.8	3.6	3.9	3.7	3.5	3.5	3.0	3.4	2.5	2.9	3.1	2.6	1.8	1.3	1.8	1.9	2.6	2.6	2.8	3.3	3.3	3.2	2.2	2.1				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


LAST CALIBRATION:	March 3, 2017
DECLINATION :	MAGNETIC DECLINATION 19 DEGREE EAST



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	738
MINIMUM 1-HR AVERAGE:	0.0 kph @ HOUR(S) 21 ON DAY(S) 20
MAXIMUM 1-HR AVERAGE:	22.7 kph @ HOUR(S) 22 ON DAY(S) 31
MAXIMUM 24-HR AVERAGE:	14.0 kph ON DAY(S) 19
	VAR-VARIOUS
MONTHLY CALIBRATION TIME:	3 hrs
OPERATIONAL TIME:	742 hrs
AMD OPERATION UPTIME:	99.7 %
STANDARD DEVIATION:	4.9
MONTHLY AVERAGE:	2.5 kph

WIND SPEED Hourly Averages (WS kph)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

WIND SPEED Instantaneous Maximum (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	26.8	28.0	27.0	22.2	21.1	24.9	24.8	28.4	31.0	27.5	30.3	27.4	27.2	23.6	20.7	18.9	18.1	11.7	7.4	14.0	13.7	15.9	12.3	12.0	7.4	31.0	21.5	24
2	10.0	10.0	10.5	13.6	12.8	10.2	10.7	8.2	11.7	13.2	14.0	17.7	19.9	19.2	17.5	15.3	16.5	13.7	16.4	16.1	18.4	21.1	27.3	27.4	8.2	27.4	15.5	24
3	21.2	19.9	17.5	19.6	15.6	11.9	16.3	19.0	8.7	7.7	C	C	C	C	26.3	22.8	10.7	10.7	14.0	5.3	5.3	8.3	10.7	6.4	5.3	26.3	13.9	24
4	5.9	9.8	27.6	31.3	35.8	31.5	40.2	32.9	36.8	40.2	32.4	34.5	24.8	23.7	32.0	21.4	22.9	17.4	16.8	20.0	17.7	19.9	17.5	20.1	5.9	40.2	25.5	24
5	18.8	25.0	20.7	15.3	16.5	20.6	16.8	11.3	12.5	13.4	11.4	16.2	13.5	11.7	12.3	8.2	7.9	6.8	7.0	7.6	6.0	6.6	10.3	13.5	6.0	25.0	12.9	24
6	15.9	16.5	16.5	16.3	18.7	16.1	19.1	12.2	12.1	22.6	24.1	26.1	26.3	25.4	28.1	36.0	23.5	28.4	31.1	33.2	33.1	33.7	32.6	28.8	12.1	36.0	24.0	24
7	29.5	37.8	33.4	34.3	42.7	30.2	26.9	40.0	32.9	32.1	30.8	33.1	32.1	29.7	32.3	30.4	25.9	19.2	19.0	12.1	15.3	17.8	14.7	10.8	10.8	42.7	27.6	24
8	14.6	13.9	13.8	9.0	6.1	5.4	10.9	12.9	14.2	13.5	16.0	15.1	11.1	11.5	9.8	8.5	8.8	9.3	4.7	2.1	4.8	5.4	9.2	7.0	2.1	16.0	9.9	24
9	4.6	4.9	3.6	6.5	6.5	1.4	4.5	3.9	2.4	2.5	5.7	6.3	8.3	9.4	10.2	11.0	9.3	6.8	1.6	6.0	7.7	6.8	5.6	5.9	1.4	11.0	5.9	24
10	6.7	14.3	13.9	13.6	14.1	18.3	18.8	18.0	17.9	21.4	23.5	27.5	29.8	29.1	23.2	23.7	21.8	21.8	18.8	14.0	11.7	11.9	11.8	11.7	6.7	29.8	18.2	24
11	12.7	11.6	14.1	12.3	12.4	13.2	14.2	X	X	12.4	13.5	18.5	17.0	20.5	21.2	18.5	22.6	22.1	23.2	23.1	20.1	18.0	22.4	19.4	11.6	23.2	17.4	22
12	21.2	20.9	23.1	16.3	18.5	18.4	16.0	17.4	18.4	21.3	19.1	17.5	17.3	13.8	18.7	22.0	23.0	23.5	22.4	25.8	29.3	30.9	19.1	28.2	13.8	30.9	20.9	24
13	25.4	17.9	21.8	23.4	17.1	18.8	20.8	18.5	31.7	28.5	24.7	26.4	26.5	29.6	27.5	27.3	31.5	32.4	29.5	27.8	19.5	15.5	19.9	21.7	15.5	32.4	24.3	24
14	18.5	20.6	8.3	7.2	7.9	12.1	8.8	7.4	9.0	9.8	5.1	8.1	8.4	11.2	14.2	17.8	20.9	17.7	22.3	19.5	30.9	29.0	22.3	16.6	5.1	30.9	14.7	24
15	18.5	27.3	21.2	17.5	14.1	11.7	9.3	8.8	24.6	6.5	10.2	6.5	9.1	9.5	12.0	12.8	12.6	11.3	15.9	17.0	19.4	14.3	6.9	20.5	6.5	27.3	14.1	24
16	18.6	16.2	16.0	17.7	6.4	7.3	8.6	4.2	6.6	5.9	13.2	18.0	16.1	14.2	22.0	9.1	9.8	3.6	8.8	10.4	7.0	9.8	8.0	11.8	3.6	22.0	11.2	24
17	7.7	6.6	10.4	9.8	10.1	6.5	7.1	6.7	20.9	19.7	16.3	19.2	19.6	23.6	23.7	27.6	30.7	33.9	33.4	34.3	34.8	31.6	31.6	6.5	6.5	34.8	20.5	24
18	33.4	29.1	22.7	24.7	21.9	23.5	28.6	24.4	26.2	22.4	22.5	22.7	25.0	21.6	21.8	22.3	22.8	21.0	23.5	25.4	19.7	23.5	43.0	38.9	19.7	43.0	25.4	24
19	36.4	42.6	54.4	49.9	43.3	41.4	40.7	30.9	33.0	36.7	37.1	36.2	40.8	37.2	42.0	46.1	33.6	31.8	23.3	24.3	26.0	21.1	28.5	29.9	21.1	54.4	36.1	24
20	37.7	45.5	45.7	45.3	36.7	32.8	34.0	36.2	42.5	28.3	22.7	19.8	21.5	25.6	26.9	22.2	18.4	19.8	10.4	5.1	4.3	6.1	3.9	6.1	3.9	45.7	24.9	24
21	7.6	9.1	11.4	16.6	13.2	14.1	16.7	18.4	19.5	37.3	39.5	33.3	34.2	34.4	33.6	30.4	34.7	36.3	39.1	36.4	33.0	30.7	32.4	30.9	7.6	39.5	26.8	24
22	26.1	23.3	25.2	22.0	22.3	22.5	21.4	21.4	20.6	20.0	16.4	20.6	19.0	22.3	20.9	18.7	11.5	9.6	10.6	7.9	11.8	11.3	9.3	13.8	7.9	26.1	17.9	24
23	12.7	11.4	8.2	5.7	6.0	6.6	4.6	4.5	6.7	13.6	11.9	10.0	10.7	16.8	11.3	9.3	12.1	4.8	5.4	6.3	7.2	9.1	5.7	6.4	4.5	16.8	8.6	24
24	9.0	6.1	3.0	3.0	4.3	4.7	5.8	9.8	10.5	14.9	18.7	26.5	22.3	13.5	19.0	19.6	27.1	31.4	25.4	27.1	21.4	19.4	20.6	19.0	3.0	31.4	15.9	24
25	22.6	21.2	19.4	19.1	12.6	10.2	9.6	10.0	10.6	8.2	10.1	16.9	15.1	15.2	13.1	11.5	7.8	7.6	7.5	10.2	13.4	11.2	8.7	11.0	7.5	22.6	12.6	24
26	12.1	14.7	12.1	6.7	7.3	15.6	10.9	12.5	6.6	6.1	7.5	10.7	13.9	13.3	15.2	17.2	10.7	11.1	11.4	11.4	8.2	11.7	18.2	19.5	6.1	19.5	11.9	24
27	13.7	11.3	13.1	16.2	13.7	15.6	16.1	18.9	15.4	14.8	17.5	16.4	16.7	18.7	17.8	20.3	16.6	18.3	15.7	31.1	29.6	18.8	14.1	10.8	10.8	31.1	17.1	24
28	7.5	7.7	13.7	15.3	12.3	17.4	21.1	16.3	19.3	16.8	14.3	17.5	18.4	18.3	16.1	16.8	13.3	13.0	19.5	15.8	20.4	14.3	15.3	8.3	7.5	21.1	15.4	24
29	8.5	9.7	7.2	6.4	6.7	2.4	2.7	3.9	5.6	5.1	9.0	9.7	12.9	11.0	13.0	9.3	5.9	7.7	6.8	8.5	8.6	11.3	10.7	9.9	2.4	13.0	8.0	24
30	14.8	16.5	21.6	23.5	21.1	19.2	22.1	25.2	23.9	20.8	23.1	22.6	18.2	17.1	16.8	12.1	11.9	9.9	6.3	8.3	6.4	5.4	5.6	4.7	4.7	25.2	15.7	24
31	3.8	6.0	5.7	12.2	7.9	5.4	6.2	7.4	15.3	15.1	10.2	19.4	26.6	32.1	28.0	30.1	31.5	24.4	14.9	18.1	23.9	36.2	34.9	30.4	3.8	36.2	18.6	24
HOURLY MAX	37.7	45.5	54.4	49.9	43.3	41.4	40.7	40.0	42.5	40.2	39.5	36.2	40.8	37.2	42.0	46.1	34.7	36.3	39.1	36.4	34.3	36.2	43.0	38.9				
HOURLY AVG	16.9	17.9	18.2	17.8	16.3	15.8	16.6	16.3	18.2	18.0	18.4	20.0	20.1	20.1	20.9	19.9	18.4	17.2	16.5	16.9	17.0	17.1	17.2	17.2				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

MAXIMUM INSTANTANEOUS VALUE:	54.4	kph	@ HOUR(S)	2	ON DAY(S)	19
					VAR-VARIOUS	
OPERATIONAL TIME:					742	hrs

WIND SPEED Instantaneous Maximum (WS kph)



Wind: LICA Bonnyville
 Monitor: WSP [kph]
 Monthly: 17/03
 Type: WindRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 14.07%

Calm Avg: 0.00 [kph]

Direction	1.8-4.6	4.6-9.1	9.1-13.7	13.7-18.2	18.2-22.8	>22.8	Total
N	3.1	4.2	1.9	1.1	0.0	0.0	10.3
NE	4.2	11.2	6.2	1.1	0.0	0.0	22.7
E	3.3	8.3	3.5	0.1	0.0	0.0	15.2
SE	1.6	5.4	7.0	3.1	0.8	0.0	18.0
S	0.3	1.0	0.5	2.0	1.0	0.0	4.7
SW	0.4	1.9	0.7	0.3	0.3	0.0	3.5
W	0.8	1.1	1.4	1.5	0.0	0.0	4.7
NW	0.8	0.8	1.4	3.1	0.7	0.0	6.8
Summary	14.5	33.8	22.6	12.3	2.7	0.0	85.9

% Icon	Classes (kph)	14		1.8-4.6	34		4.6-9.1	23		9.1-13.7	12		13.7-18.2	3		18.2-22.8	0		>22.8
--------	---------------	----	---	---------	----	---	---------	----	---	----------	----	---	-----------	---	---	-----------	---	---	-------

LICA Bonnyville 2017/03/01 00:00 - 2017/03/31 23:00 Calm: 14.07% Calm Wind Avg Speed: 0.99(kph)



WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

WIND DIRECTION Hourly Averages (WD)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	24-HOUR AVG	24-HR	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	QUADRANT	RDGS.	
DAY 1	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	NE	ENE	ENE	NE	NE	ENE	ENE	NE	NE	ENE	NE	E	ENE	24	
2	E	ENE	ENE	ENE	E	ENE	ENE	ENE	ENE	ENE	NE	NE	NE	NE	NE	ESE	ESE	NE	ENE	E	E	ESE	ESE	SE	ENE	24	
3	SE	ESE	ESE	ESE	ESE	E	ESE	ESE	NNE	N	C	C	C	SW	S	SSE	NNE	NNE	N	N	NNW	NNW	N	SE	ESE	24	
4	NNW	N	NE	NNE	N	NNE	NNE	NE	NNE	NNE	NNE	NNE	N	N	N	N	N	NE	NE	NNE	NE	NE	NE	NE	NNE	24	
5	NE	NE	NE	NE	NE	NE	NE	NE	NE	NE	NE	NE	NE	NE	NE	NE	N	N	NE	SE	NNW	NNW	NNW	NNW	NE	24	
6	NNE	NNE	NE	NNE	NNE	NE	NNE	NNE	N	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	N	24
7	NW	NW	NW	NNW	NW	NW	NW	NW	NW	NNW	NNW	NNW	NW	NW	NW	NNW	NNW	NNW	NNW	NNW	N	N	N	N	NNW	24	
8	N	NNE	NNE	NNE	NNW	NNW	N	NNE	NNE	NE	NE	NE	NNE	NNE	N	SW	SSE	SE	N	NW	N	NW	NNE	N	NNE	24	
9	N	NNE	NNW	N	N	NNW	WNW	NW	N	NNW	SW	SW	S	S	S	SE	SSE	E	N	NNE	SSE	NE	NNE	NNE	SSE	24	
10	NE	NE	ENE	ENE	NE	NE	ENE	ENE	ENE	ENE	ENE	ENE	NE	NE	NE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	24
11	ENE	ENE	E	E	E	E	ESE	X	X	ESE	ESE	NE	NE	NE	NE	NE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	22	
12	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	E	ENE	ENE	NE	NE	NE	NE	SSE	S	S	SSE	SE	SSE	SSE	SSE	ESE	24	
13	SSE	SE	SE	SE	SE	SE	SE	SE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SE	SE	SE	ESE	SSE	24	
14	ESE	NE	ENE	ENE	E	ESE	N	E	E	ESE	NNW	S	E	NNE	ENE	SE	ESE	ESE	ESE	ESE	SE	ESE	SE	E	ESE	24	
15	ESE	SE	SE	SE	ESE	E	ENE	NW	WSW	W	S	N	N	SW	SW	SSW	WNW	NNW	NW	NNW	N	NE	N	ESE	24		
16	N	N	N	N	N	ESE	SW	NE	N	NE	ENE	ENE	NE	NE	NNE	NNE	NE	NNE	NNE	NE	NE	ENE	NE	ESE	NNE	24	
17	ESE	SE	ESE	SE	SE	E	E	ESE	SSE	SSE	SE	SSW	SSW	SSE	SSE	S	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SE	SSE	24	
18	SE	SE	ESE	E	E	ESE	ESE	ESE	ESE	ESE	SE	SE	SE	SE	SE	SE	ESE	ENE	ENE	ENE	NNW	NW	NW	ESE	24		
19	W	WNW	WNW	WNW	WNW	WNW	WNW	W	W	W	W	W	WSW	W	W	NW	WNW	W	WSW	WSW	WSW	WSW	WNW	WNW	W	24	
20	WNW	NNW	NW	NW	NW	WNW	WNW	NW	NW	NNW	NNW	WNW	WNW	W	W	W	W	NW	NW	WNW	NW	E	NNW	SE	NW	24	
21	SE	SE	E	ESE	ESE	ESE	E	ESE	SE	SE	SSE	SE	SE	SE	ESE	ESE	ESE	ESE	SE	ESE	ESE	ESE	ESE	ESE	SE	24	
22	ESE	E	E	E	E	E	E	E	E	E	ENE	ENE	NE	NE	NE	NE	N	NNW	NW	WNW	W	W	W	W	ENE	24	
23	WSW	WNW	W	W	WNW	NNW	NNW	N	NNW	NE	NE	NNE	NNW	NW	NNW	N	NW	N	NE	E	S	ESE	SSW	SSE	NNW	24	
24	E	NW	NNE	NNE	NNE	N	NNE	ENE	ESE	E	ESE	SE	ESE	ESE	ESE	E	E	ESE	ESE	E	E	ENE	ENE	ENE	E	24	
25	NE	NE	NE	NE	NNE	NNE	N	NNW	NNW	NW	SW	WSW	SW	SW	SSW	SSW	SW	E	ENE	E	E	ESE	ENE	ENE	NE	24	
26	ENE	ESE	NE	SSE	NE	E	E	NNE	W	W	NNE	NNW	WSW	SW	SW	SW	SW	SW	SE	SE	E	E	SE	SE	SSE	24	
27	E	ENE	ENE	E	ENE	ENE	NE	ENE	ENE	E	SE	SE	SE	SSE	SSE	SE	SE	SSE	SSE	SSE	SE	SE	SSE	ESE	ESE	24	
28	E	E	ESE	ESE	ESE	SE	SE	SE	SSE	SE	SE	SE	SE	SE	SE	SE	SE	ESE	SE	ESE	ESE	E	E	NE	SE	24	
29	NE	NE	ENE	ENE	NE	NNE	NNE	N	NE	N	W	W	WSW	SW	N	NNE	ESE	NE	ENE	NE	NE	E	E	E	NE	24	
30	E	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	NE	NE	NE	NE	NE	NE	NNE	NNW	NNW	SE	NNW	SE	ENE	ENE	24	
31	NNE	NNW	NE	E	NE	N	ENE	ESE	SW	SW	SSW	SSW	SSW	SSW	SSW	S	S	SSE	SSE	SSE	SSE	S	SSW	SSW	S	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION:	March 3, 2017
DECLINATION :	MAGNETIC DECLINATION 19 DEGREE EAST

MONTHLY CALIBRATION TIME:	3 hrs	OPERATIONAL TIME:	742 hrs
STANDARD DEVIATION:	106	AMD OPERATION UPTIME:	99.7 %
		MONTHLY AVERAGE:	80 (E)

WIND DIRECTION Hourly Averages (WD)



STANDARD DEVIATION WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Bonnyville Continuous Monitoring Station - March 2017

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59		
DAY																										
1	16	16	17	18	16	17	18	16	17	18	18	16	17	19	20	20	18	28	28	17	18	19	16	19	24	
2	27	22	21	15	19	24	20	24	28	18	17	16	13	17	19	34	21	15	17	22	23	12	13	14	24	
3	15	14	13	10	14	15	13	17	14	16	C	C	C	39	13	18	30	15	16	11	6	6	12	57	24	
4	12	12	18	19	17	19	20	19	18	20	20	18	18	15	17	16	17	17	17	17	15	17	16	17	24	
5	15	18	17	14	17	16	14	16	15	16	22	19	21	24	17	17	18	11	9	14	40	12	10	11	24	
6	15	16	17	16	15	17	15	14	13	14	15	14	14	14	15	15	14	15	14	15	16	15	15	15	24	
7	15	13	14	15	15	14	13	12	14	16	16	16	16	16	15	15	14	13	12	8	11	14	12	10	24	
8	12	14	16	14	14	22	11	15	16	15	19	21	24	25	19	30	21	36	39	13	9	43	12	7	24	
9	11	6	8	11	15	8	10	19	15	17	45	51	69	14	31	54	32	29	9	11	59	40	11	10	24	
10	13	15	18	15	13	13	16	14	14	15	16	17	17	16	16	16	16	19	20	18	17	15	18	17	24	
11	17	16	13	16	15	14	19	X	X	16	22	20	15	15	14	15	17	16	16	20	17	16	16	15	22	
12	15	18	18	17	16	16	17	18	16	20	18	18	16	18	14	24	7	6	6	10	13	10	11	10	24	
13	11	12	13	13	15	13	14	15	15	14	13	9	10	10	10	10	11	10	11	10	13	21	17	32	24	
14	34	23	29	23	29	18	16	26	28	34	59	55	70	15	22	16	14	15	13	14	15	13	17	20	24	
15	20	15	14	13	13	14	24	15	33	39	68	60	26	35	42	10	21	17	13	14	15	17	35	15	24	
16	13	13	14	12	19	37	35	29	13	16	19	18	17	17	16	15	12	7	10	12	14	16	15	28	24	
17	22	29	21	21	21	24	20	39	21	22	18	12	16	9	12	8	11	10	10	11	10	12	13	15	24	
18	11	14	19	14	18	18	13	17	13	15	14	15	14	14	13	15	18	16	16	18	23	13	15	18	24	
19	18	19	18	16	18	18	18	19	19	19	21	20	19	18	22	17	16	17	15	15	15	14	17	16	24	
20	16	17	15	15	15	15	14	15	16	18	19	30	28	25	26	25	23	22	16	16	28	71	39	61	24	
21	46	42	24	13	15	17	22	17	16	13	12	15	14	15	16	18	15	14	13	14	15	15	15	16	24	
22	15	18	16	15	15	18	15	15	17	18	20	19	16	16	14	16	13	12	13	23	25	19	21	25	24	
23	17	25	25	29	18	29	27	11	14	14	20	38	20	20	18	19	58	31	18	33	45	17	54	35	24	
24	24	20	14	8	24	25	8	17	31	24	18	14	22	21	26	23	17	15	16	16	20	19	16	16	24	
25	17	17	17	16	14	13	10	12	17	22	19	24	24	16	13	25	55	51	19	18	18	21	18	21	24	
26	17	12	18	35	19	17	27	22	38	47	25	63	20	35	36	9	19	37	33	14	20	12	13	20	24	
27	14	14	13	16	19	15	15	16	19	21	16	16	23	12	9	16	15	13	12	11	14	14	19	17	24	
28	22	20	16	13	19	16	13	16	13	16	18	16	15	13	16	15	18	20	14	17	16	19	17	15	24	
29	12	12	16	15	17	3	15	10	36	37	48	44	27	73	24	47	49	17	17	11	14	16	19	18	24	
30	22	18	18	17	17	18	18	17	17	18	17	19	17	17	21	23	21	18	13	12	68	11	35	23	24	
31	37	17	46	18	17	7	14	48	45	7	25	11	8	9	7	11	8	7	22	11	6	6	7	11	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION: March 3, 2017

CALIBRATION TIME: 0 hrs OPERATIONAL TIME: 742 hrs

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)



APPENDIX II
NON-CONTINUOUS MONITORING DATA RESULTS

VOC RESULTS

Maxxam Analytics

VOC Sample Collection Data Sheet Alberta Air FCD AIR FCD-01320 / 2

Client: LICA Sampler S/N: 6200
 Location: Bonnyville - AER Canister ID: 14995
 Station ID: LICA 37 Installation Date/Time (mst): Feb 27, 2017 @ 12:48
 Sample ID: LICA/VOC/Bonnyville/Mar 02, 2017 Removal Date/Time (mst): Mar 07, 2017 @ 14:49

Date and Time Information

Sample Date:	Start Time (mst)	End Time (mst)	Elapsed Time (hours)
<u>Mar 02, 2017</u>	<u>00:00</u>	<u>00:00</u> <u>Mar 03, 2017</u>	<u>24.0</u>

Canister Pressure/Vacuum	
Initial Vacuum (in. Hg)	Final Pressure (psi)
<u>-27.6</u>	<u>+18.1</u>

Flow Settings		
Flow Reading (sccm)	Pot Set Point	Pump Set (psi)
<u>10.0</u>	<u>4.94</u>	<u>27.5</u>

Deployment/Collection and Maintenance Checklist

Initial leak check deployment vacuum (in. Hg) = — @ — mst
 Final leak check deployment vacuum (in. Hg) = — @ — mst
 Total leak rate = — psi over — minutes
 Timer reset to zero prior to sampling? YES (yes/no)
 Date of last flow calibration: Jan 27, 2017 (due every 3 months)
 Last date of sample line & fitting replacement: Sep 28, 2016 (due every 6 months)

Leak rate must be 0.0 psi over a minimum of 5 minutes or repair is required

Comments: The canister is not equipped with a pressure gauge.

Deployment Technician Signature: Alex Yakupov

Collection Technician Signature: Alex Yakupov Date: Mar 07, 2017

Sample ID: 17030077-001

Customer ID: LICA
 Cust Samp ID: LICAVOC/Bonnyville/Mar 2, 2017



Volatile Organics Data Results

Date: March 2, 2017
Canister ID: 14995

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.02
1,1,2,2-Tetrachloroethane	< 0.02
1,1,2-Trichloroethane	< 0.02
1,1-Dichloroethane	< 0.02
1,1-Dichloroethylene	< 0.04
1,2,3-Trimethylbenzene	< 0.05
1,2,4-Trichlorobenzene	< 0.8
1,2,4-Trimethylbenzene	0.07
1,2-Dibromoethane	< 0.02
1,2-Dichlorobenzene	< 0.03
1,2-Dichloroethane	0.02
1,2-Dichloropropane	0.01
1,3,5-Trimethylbenzene	0.04
1,3-Butadiene	0.02
1,3-Dichlorobenzene	< 0.3
1,4-Dichlorobenzene	< 0.4
1,4-Dioxane	< 0.4
1-Butene	0.11
1-Hexene	0.02
1-Pentene	0.02
2,2,4-Trimethylpentane	0.15
2,2-Dimethylbutane	0.04
2,3,4-Trimethylpentane	0.05
2,3-Dimethylbutane	0.11
2,3-Dimethylpentane	0.15
2,4-Dimethylpentane	0.07
2-Methylheptane	0.06
2-Methylhexane	0.09
2-Methylpentane	0.32
3-Methylheptane	0.04
3-Methylhexane	0.12
3-Methylpentane	0.17
Acetone	1.7
Acrolein	< 0.3
Benzene	0.27
Benzyl chloride	< 0.4
Bromodichloromethane	< 0.02
Bromoform	< 0.02
Bromomethane	0.01
Carbon disulfide	0.01
Carbon tetrachloride	0.13
Chlorobenzene	< 0.02
Chloroethane	< 0.02
Chloroform	0.03
Chloromethane	0.64
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.04
cis-2-Butene	0.03
cis-2-Pentene	< 0.02
Cyclohexane	0.23
Cyclopentane	0.11
Dibromochloromethane	< 0.01
Ethanol	2.6
Ethyl acetate	< 0.4
Ethylbenzene	0.07
Freon-11	0.29
Freon-113	0.12

Volatile Organics Data Results

Date: March 2, 2017
Canister ID: 14995

PARAMETERS	CONCENTRATION (PPB)
Freon-114	0.02
Freon-12	0.61
Hexachloro-1,3-butadiene	< 0.50
Isobutane	1.01
Isopentane	0.99
Isoprene	< 0.01
Isopropyl alcohol	< 0.4
Isopropylbenzene	< 0.01
m,p-Xylene	0.25
m-Diethylbenzene	< 0.04
m-Ethyltoluene	< 0.08
Methyl butyl ketone	< 0.50
Methyl ethyl ketone	< 0.3
Methyl isobutyl ketone	< 0.4
Methyl methacrylate	< 0.07
Methyl tert butyl ether	< 0.03
Methylcyclohexane	0.29
Methylcyclopentane	0.27
Methylene chloride	< 0.3
n-Butane	2.37
n-Decane	< 0.06
n-Dodecane	< 0.4
n-Heptane	0.14
n-Hexane	0.23
n-Nonane	0.06
n-Octane	0.6
n-Pentane	< 0.05
n-Propylbenzene	< 0.5
n-Undecane	< 0.5
Naphthalene	0.03
o-Ethyltoluene	0.02
o-Xylene	0.08
p-Diethylbenzene	< 0.04
p-Ethyltoluene	< 0.07
Styrene	< 0.04
Tetrachloroethylene	0.05
Tetrahydrofuran	< 0.4
Toluene	0.33
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.04
trans-2-Butene	0.03
trans-2-Pentene	0.03
Trichloroethylene	< 0.04
Vinyl acetate	< 0.04
Vinyl chloride	< 0.02

Customer ID: LICA
 Cust Samp ID: LICA/VOC/Bonnyville/Mar 08, 2017

Maxxam Analytics

VOC Sample Collection Data Sheet Alberta Air FCD AIR FCD-01320 / 2

Client: LICA Sampler S/N: 6200
 Location: Bonnyville - AER Canister ID: 1831
 Station ID: LICA 37 Installation Date/Time (mst): Mar 07, 2017 @ 14:49
 Sample ID: LICA/VOC/Bonnyville/Mar 08, 2017 Removal Date/Time (mst): Mar 09, 2017 @ 14:29

Date and Time Information

Sample Date:	Start Time (mst)	End Time (mst)	Elapsed Time (hours)
<u>Mar 09, 2017</u>	<u>00:00</u>	<u>00:00</u> <u>Mar 09, 2017</u>	<u>24.0</u>

Canister Pressure/Vacuum	
Initial Vacuum (in. Hg)	Final Pressure (psi)
<u>-27.0</u>	<u>+18.5</u>

Flow Settings		
Flow Reading (sccm)	Pot Set Point	Pump Set (psi)
<u>10.0</u>	<u>4.94</u>	<u>27.5</u>

Deployment/Collection and Maintenance Checklist

Initial leak check deployment vacuum (in. Hg) = - @ - mst
 Final leak check deployment vacuum (in. Hg) = - @ - mst
 Total leak rate = - psi over - minutes
 Timer reset to zero prior to sampling? YES (yes/no)
 Date of last flow calibration: Jan 27, 2017 (due every 3 months)
 Last date of sample line & fitting replacement: Sep 28, 2016 (due every 6 months)

****Leak rate must be 0.0 psi over a minimum of 5 minutes or repair is required****

Comments: The canister is not equipped with a pressure gauge

Deployment Technician Signature: Alex Yakupov

Collection Technician Signature: Alex Yakupov Date: Mar 09, 2017



Volatile Organics Data Results

Date: March 8, 2017
Canister ID: 1831

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.02
1,1,2,2-Tetrachloroethane	< 0.02
1,1,2-Trichloroethane	< 0.02
1,1-Dichloroethane	< 0.02
1,1-Dichloroethylene	< 0.04
1,2,3-Trimethylbenzene	< 0.05
1,2,4-Trichlorobenzene	< 0.8
1,2,4-Trimethylbenzene	0.1
1,2-Dibromoethane	< 0.02
1,2-Dichlorobenzene	< 0.03
1,2-Dichloroethane	0.02
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	0.05
1,3-Butadiene	0.03
1,3-Dichlorobenzene	< 0.3
1,4-Dichlorobenzene	< 0.4
1,4-Dioxane	< 0.4
1-Butene	0.14
1-Hexene	0.03
1-Pentene	0.04
2,2,4-Trimethylpentane	0.14
2,2-Dimethylbutane	0.05
2,3,4-Trimethylpentane	0.06
2,3-Dimethylbutane	0.08
2,3-Dimethylpentane	0.12
2,4-Dimethylpentane	0.06
2-Methylheptane	0.06
2-Methylhexane	0.08
2-Methylpentane	0.21
3-Methylheptane	0.05
3-Methylhexane	0.1
3-Methylpentane	0.16
Acetone	1.4
Acrolein	< 0.3
Benzene	0.3
Benzyl chloride	< 0.4
Bromodichloromethane	< 0.02
Bromoform	< 0.02
Bromomethane	0.01
Carbon disulfide	0.08
Carbon tetrachloride	0.13
Chlorobenzene	< 0.02
Chloroethane	< 0.02
Chloroform	0.03
Chloromethane	0.66
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.04
cis-2-Butene	0.05
cis-2-Pentene	0.03
Cyclohexane	0.14
Cyclopentane	0.08
Dibromochloromethane	< 0.01
Ethanol	1.3
Ethyl acetate	< 0.4
Ethylbenzene	0.08
Freon-11	0.3
Freon-113	0.13

Volatile Organics Data Results

Date: March 8, 2017
Canister ID: 1831

PARAMETERS	CONCENTRATION (PPB)
Freon-114	0.02
Freon-12	0.63
Hexachloro-1,3-butadiene	< 0.50
Isobutane	0.59
Isopentane	0.61
Isoprene	0.03
Isopropyl alcohol	< 0.4
Isopropylbenzene	0.02
m,p-Xylene	0.27
m-Diethylbenzene	< 0.04
m-Ethyltoluene	< 0.08
Methyl butyl ketone	< 0.50
Methyl ethyl ketone	< 0.3
Methyl isobutyl ketone	< 0.4
Methyl methacrylate	< 0.07
Methyl tert butyl ether	< 0.03
Methylcyclohexane	0.15
Methylcyclopentane	0.28
Methylene chloride	< 0.3
n-Butane	1.19
n-Decane	< 0.06
n-Dodecane	< 0.4
n-Heptane	0.11
n-Hexane	0.4
n-Nonane	0.05
n-Octane	0.07
n-Pentane	0.3
n-Propylbenzene	< 0.05
n-Undecane	< 0.5
Naphthalene	< 0.5
o-Ethyltoluene	0.04
o-Xylene	0.1
p-Diethylbenzene	< 0.04
p-Ethyltoluene	< 0.07
Styrene	0.05
Tetrachloroethylene	0.05
Tetrahydrofuran	< 0.4
Toluene	0.31
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.04
trans-2-Butene	0.06
trans-2-Pentene	0.05
Trichloroethylene	< 0.04
Vinyl acetate	< 0.4
Vinyl chloride	< 0.02

Sample ID: 17030212-001

Customer ID: LICA

Cust Samp ID: LICA/VOC/Bonnyville/Mar 14, 2017

Maxxam Analytics

VOC Sample Collection Data Sheet Alberta Air FCD AIR FCD-01320 / 2

Client: <u>LICA</u>	Sampler S/N: <u>6200</u>
Location: <u>Bonnyville - AER</u>	Canister ID: <u>2665</u>
Station ID: <u>LICA 37</u>	Installation Date/Time (mst): <u>Mar 09, 2017 @ 14:29</u>
Sample ID: <u>LICA/VOC/Bonnyville/Mar 14, 2017</u>	Removal Date/Time (mst): <u>MAR 16, 2017 @ 11:35</u>

Date and Time Information

Sample Date:	Start Time (mst)	End Time (mst)	Elapsed Time (hours)
<u>Mar 14, 2017</u>	<u>00:00</u>	<u>00:00 Mar 15, 2017</u>	<u>24.0</u>

Canister Pressure/Vacuum	
Initial Vacuum (in. Hg)	Final Pressure (psi)
<u>-27.7</u>	<u>20</u>

Flow Settings		
Flow Reading (sccm)	Pot Set Point	Pump Set (psi)
<u>10.0</u>	<u>4.94</u>	<u>27.5</u>

Deployment/Collection and Maintenance Checklist

Initial leak check deployment vacuum (in. Hg) = - @ - mst

Final leak check deployment vacuum (in. Hg) = - @ - mst

Total leak rate = - psi over - minutes

Timer reset to zero prior to sampling? YES (yes/no)

Date of last flow calibration: Jan 27, 2017 (due every 3 months)

Last date of sample line & fitting replacement: Sep 28, 2016 (due every 6 months)

Leak rate must be 0.0 psi over a minimum of 5 minutes or repair is required

Comments: _____

Deployment Technician Signature: Alex Yakupov

Collection Technician Signature: [Signature] Date: Mar 16, 2017.



Volatile Organics Data Results

Date: March 14, 2017
Canister ID: 2665

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.02
1,1,2,2-Tetrachloroethane	< 0.02
1,1,2-Trichloroethane	< 0.02
1,1-Dichloroethane	< 0.02
1,1-Dichloroethylene	< 0.04
1,2,3-Trimethylbenzene	< 0.05
1,2,4-Trichlorobenzene	< 0.8
1,2,4-Trimethylbenzene	< 0.05
1,2-Dibromoethane	< 0.02
1,2-Dichlorobenzene	< 0.03
1,2-Dichloroethane	0.02
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	0.02
1,3-Butadiene	< 0.02
1,3-Dichlorobenzene	< 0.3
1,4-Dichlorobenzene	< 0.4
1,4-Dioxane	< 0.4
1-Butene	0.14
1-Hexene	< 0.02
1-Pentene	0.02
2,2,4-Trimethylpentane	0.05
2,2-Dimethylbutane	0.03
2,3,4-Trimethylpentane	0.02
2,3-Dimethylbutane	0.07
2,3-Dimethylpentane	0.07
2,4-Dimethylpentane	0.03
2-Methylheptane	0.03
2-Methylhexane	0.07
2-Methylpentane	0.25
3-Methylheptane	0.02
3-Methylhexane	0.09
3-Methylpentane	0.16
Acetone	2.7
Acrolein	< 0.3
Benzene	0.2
Benzyl chloride	< 0.4
Bromodichloromethane	< 0.02
Bromoform	< 0.02
Bromomethane	< 0.01
Carbon disulfide	< 0.01
Carbon tetrachloride	0.1
Chlorobenzene	< 0.02
Chloroethane	< 0.02
Chloroform	0.02
Chloromethane	0.63
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.04
cis-2-Butene	< 0.02
cis-2-Pentene	< 0.02
Cyclohexane	0.11
Cyclopentane	0.07
Dibromochloromethane	< 0.01
Ethanol	1.8
Ethyl acetate	< 0.4
Ethylbenzene	0.03
Freon-11	0.29
Freon-113	0.09

Volatile Organics Data Results

Date: March 14, 2017
Canister ID: 2665

PARAMETERS	CONCENTRATION (PPB)
Freon-114	0.02
Freon-12	0.63
Hexachloro-1,3-butadiene	< 0.50
Isobutane	2.37
Isopentane	1.26
Isoprene	< 0.01
Isopropyl alcohol	< 0.4
Isopropylbenzene	< 0.01
m,p-Xylene	0.09
m-Diethylbenzene	< 0.04
m-Ethyltoluene	< 0.08
Methyl butyl ketone	< 0.50
Methyl ethyl ketone	0.6
Methyl isobutyl ketone	< 0.4
Methyl methacrylate	< 0.07
Methyl tert butyl ether	< 0.03
Methylcyclohexane	0.16
Methylcyclopentane	0.19
Methylene chloride	< 0.3
n-Butane	4.63
n-Decane	< 0.06
n-Dodecane	< 0.4
n-Heptane	0.12
n-Hexane	0.34
n-Nonane	0.02
n-Octane	0.04
n-Pentane	1.1
n-Propylbenzene	< 0.05
n-Undecane	< 0.5
Naphthalene	< 0.5
o-Ethyltoluene	0.01
o-Xylene	0.04
p-Diethylbenzene	< 0.04
p-Ethyltoluene	< 0.07
Styrene	< 0.04
Tetrachloroethylene	0.04
Tetrahydrofuran	< 0.4
Toluene	0.14
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.04
trans-2-Butene	0.03
trans-2-Pentene	< 0.02
Trichloroethylene	< 0.04
Vinyl acetate	< 0.4
Vinyl chloride	< 0.02

Maxxam Analytics

VOC Sample Collection Data Sheet Alberta Air FCD AIR FCD-01320 / 2

Client: LICA
 Location: Bonnyville - AER
 Station ID: LICA 37
 Sample ID: LICA/VOC/Bonnyville/Mar 20, 2017

Sampler S/N: 6200
 Canister ID: 2471
 Installation Date/Time (mst): MAR 16, 2017 @ 11:40
 Removal Date/Time (mst): MAR 24, 2017 @ 08:15

Date and Time Information

Sample Date:	Start Time (mst)	End Time (mst)	Elapsed Time (hours)
MAR 20, 2017	00:00	24:00	24:00

Canister Pressure/Vacuum	
Initial Vacuum (in. Hg)	Final Pressure (psi)
-29	18

Flow Settings		
Flow Reading (sccm)	Pot Set Point	Pump Set (psi)
10.0	4.94	27.5

Deployment/Collection and Maintenance Checklist

Initial leak check deployment vacuum (in. Hg) = - @ - mst
 Final leak check deployment vacuum (in. Hg) = - @ - mst
 Total leak rate = - psi over - minutes
 Timer reset to zero prior to sampling? YES (yes/no)
 Date of last flow calibration: JAN 27, 2017 (due every 3 months)
 Last date of sample line & fitting replacement: SEP 28, 2016 (due every 6 months)

Leak rate must be 0.0 psi over a minimum of 5 minutes or repair is required

Comments: _____

Deployment Technician Signature: _____

Collection Technician Signature: _____

Sample ID: 17030303-001

Customer ID: LICA
 Cust Samp ID: LICA/VOC/Bonnyville/Mar 20, 2017



Volatile Organics Data Results

Date: March 20, 2017
Canister ID: 2417

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.02
1,1,2,2-Tetrachloroethane	< 0.02
1,1,2-Trichloroethane	< 0.02
1,1-Dichloroethane	< 0.02
1,1-Dichloroethylene	< 0.04
1,2,3-Trimethylbenzene	< 0.05
1,2,4-Trichlorobenzene	< 0.8
1,2,4-Trimethylbenzene	< 0.05
1,2-Dibromoethane	< 0.02
1,2-Dichlorobenzene	< 0.03
1,2-Dichloroethane	0.02
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	< 0.02
1,3-Butadiene	< 0.02
1,3-Dichlorobenzene	< 0.3
1,4-Dichlorobenzene	< 0.4
1,4-Dioxane	< 0.4
1-Butene	0.02
1-Hexene	< 0.02
1-Pentene	< 0.01
2,2,4-Trimethylpentane	0.02
2,2-Dimethylbutane	0.01
2,3,4-Trimethylpentane	< 0.01
2,3-Dimethylbutane	< 0.02
2,3-Dimethylpentane	0.02
2,4-Dimethylpentane	0.01
2-Methylheptane	< 0.01
2-Methylhexane	0.02
2-Methylpentane	0.05
3-Methylheptane	< 0.02
3-Methylhexane	0.02
3-Methylpentane	0.01
Acetone	1.4
Acrolein	< 0.3
Benzene	0.14
Benzyl chloride	< 0.4
Bromodichloromethane	< 0.02
Bromoform	< 0.02
Bromomethane	0.01
Carbon disulfide	0.12
Carbon tetrachloride	0.08
Chlorobenzene	< 0.02
Chloroethane	< 0.02
Chloroform	< 0.02
Chloromethane	0.63
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.04
cis-2-Butene	< 0.02
cis-2-Pentene	< 0.02
Cyclohexane	0.03
Cyclopentane	0.01
Dibromochloromethane	< 0.01
Ethanol	0.4
Ethyl acetate	< 0.4
Ethylbenzene	< 0.01
Freon-11	0.23
Freon-113	0.09

Volatile Organics Data Results

Date: March 20, 2017
Canister ID: 2417

PARAMETERS	CONCENTRATION (PPB)
Freon-114	< 0.02
Freon-12	0.38
Hexachloro-1,3-butadiene	< 0.50
Isobutane	0.29
Isopentane	0.32
Isoprene	< 0.01
Isopropyl alcohol	< 0.4
Isopropylbenzene	< 0.01
m,p-Xylene	< 0.03
m-Diethylbenzene	< 0.04
m-Ethyltoluene	< 0.08
Methyl butyl ketone	< 0.50
Methyl ethyl ketone	< 0.3
Methyl isobutyl ketone	< 0.4
Methyl methacrylate	< 0.07
Methyl tert butyl ether	< 0.03
Methylcyclohexane	0.04
Methylcyclopentane	0.02
Methylene chloride	< 0.3
n-Butane	0.97
n-Decane	< 0.06
n-Dodecane	< 0.4
n-Heptane	0.03
n-Hexane	< 0.01
n-Nonane	< 0.01
n-Octane	< 0.02
n-Pentane	0.2
n-Propylbenzene	< 0.05
n-Undecane	< 0.5
Naphthalene	< 0.5
o-Ethyltoluene	< 0.01
o-Xylene	0.02
p-Diethylbenzene	< 0.04
p-Ethyltoluene	< 0.07
Styrene	< 0.04
Tetrachloroethylene	< 0.04
Tetrahydrofuran	< 0.4
Toluene	0.07
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.04
trans-2-Butene	0.02
trans-2-Pentene	< 0.02
Trichloroethylene	< 0.04
Vinyl acetate	< 0.4
Vinyl chloride	< 0.02

Sample ID: 17030303-003

Customer ID: LICA

Cust Samp ID: LICA/VOC/Bonnyville/Mar 26, 2017

Maxxam Analytics

Collection Data Sheet Alberta Air FCD AIR FCD-01320 / 2

Client: LICA Sampler S/N: ~~55600ME~~ 6200
 Location: Bonnyville - AER Canister ID: S5600
 Station ID: LICA 37 Installation Date/Time (mst): MAR 24, 2017 @ 08:32
 Sample ID: LICA/VOC/Bonnyville/Mar 26, 2017 Removal Date/Time (mst): MAR 27, 2017 @ 11:39

Date and Time Information

Sample Date:	Start Time (mst)	End Time (mst)	Elapsed Time (hours)
<u>MARCH 26, 2017</u>	<u>00:00</u>	<u>24:00</u>	<u>24</u>

Canister Pressure/Vacuum	
Initial Vacuum (in. Hg)	Final Pressure (psi)
<u>-28</u>	<u>+19.2</u>

Flow Settings		
Flow Reading (sccm)	Pot Set Point	Pump Set (psi)
<u>10</u>	<u>4.94</u>	<u>27.5</u>

Deployment/Collection and Maintenance Checklist

Initial leak check deployment vacuum (in. Hg) = — @ — mst
 Final leak check deployment vacuum (in. Hg) = — @ — mst
 Total leak rate = — psi over — minutes
 Timer reset to zero prior to sampling? YES (yes/no)
 Date of last flow calibration: JANUARY 27, 2017 (due every 3 months)
 Last date of sample line & fitting replacement: SEPTEMBER 28, 2016 (due every 6 months)

****Leak rate must be 0.0 psi over a minimum of 5 minutes or repair is required****

Comments: n/a

Deployment Technician Signature: [Signature]

Collection Technician Signature: Alex Yakupov

Date: Mar 27, 2017



Volatile Organics Data Results

Date: March 26, 2017
Canister ID: 2471

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.02
1,1,2,2-Tetrachloroethane	< 0.02
1,1,2-Trichloroethane	< 0.02
1,1-Dichloroethane	< 0.02
1,1-Dichloroethylene	< 0.04
1,2,3-Trimethylbenzene	< 0.05
1,2,4-Trichlorobenzene	< 0.8
1,2,4-Trimethylbenzene	0.05
1,2-Dibromoethane	< 0.02
1,2-Dichlorobenzene	< 0.03
1,2-Dichloroethane	0.02
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	< 0.02
1,3-Butadiene	< 0.02
1,3-Dichlorobenzene	< 0.3
1,4-Dichlorobenzene	< 0.4
1,4-Dioxane	< 0.4
1-Butene	0.06
1-Hexene	< 0.02
1-Pentene	< 0.01
2,2,4-Trimethylpentane	< 0.01
2,2-Dimethylbutane	0.02
2,3,4-Trimethylpentane	0.01
2,3-Dimethylbutane	0.06
2,3-Dimethylpentane	0.11
2,4-Dimethylpentane	0.04
2-Methylheptane	0.11
2-Methylhexane	0.26
2-Methylpentane	0.23
3-Methylheptane	0.06
3-Methylhexane	0.27
3-Methylpentane	0.14
Acetone	3.2
Acrolein	< 0.3
Benzene	0.3
Benzyl chloride	< 0.4
Bromodichloromethane	< 0.02
Bromoform	< 0.02
Bromomethane	< 0.01
Carbon disulfide	0.09
Carbon tetrachloride	0.07
Chlorobenzene	< 0.02
Chloroethane	< 0.02
Chloroform	< 0.02
Chloromethane	0.65
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.04
cis-2-Butene	0.02
cis-2-Pentene	< 0.02
Cyclohexane	0.26
Cyclopentane	0.1
Dibromochloromethane	< 0.01
Ethanol	1.3
Ethyl acetate	< 0.4
Ethylbenzene	0.1
Freon-11	0.22
Freon-113	0.08

Volatile Organics Data Results

Date: March 26, 2017
Canister ID: 2471

PARAMETERS	CONCENTRATION (PPB)
Freon-114	< 0.02
Freon-12	0.36
Hexachloro-1,3-butadiene	< 0.50
Isobutane	0.99
Isopentane	0.92
Isoprene	< 0.01
Isopropyl alcohol	< 0.4
Isopropylbenzene	< 0.01
m,p-Xylene	0.39
m-Diethylbenzene	< 0.04
m-Ethyltoluene	< 0.08
Methyl butyl ketone	< 0.50
Methyl ethyl ketone	0.3
Methyl isobutyl ketone	< 0.4
Methyl methacrylate	< 0.07
Methyl tert butyl ether	< 0.03
Methylcyclohexane	0.54
Methylcyclopentane	0.31
Methylene chloride	0.3
n-Butane	2.53
n-Decane	< 0.06
n-Dodecane	< 0.4
n-Heptane	0.43
n-Hexane	0.36
n-Nonane	0.08
n-Octane	0.16
n-Pentane	0.7
n-Propylbenzene	< 0.05
n-Undecane	< 0.5
Naphthalene	< 0.5
o-Ethyltoluene	0.01
o-Xylene	0.19
p-Diethylbenzene	< 0.04
p-Ethyltoluene	< 0.07
Styrene	< 0.04
Tetrachloroethylene	< 0.04
Tetrahydrofuran	< 0.4
Toluene	0.8
trans-1,2-Dichloroethylene	0.08
trans-1,3-Dichloropropylene	< 0.04
trans-2-Butene	0.02
trans-2-Pentene	< 0.02
Trichloroethylene	< 0.04
Vinyl acetate	< 0.4
Vinyl chloride	< 0.02

PAH RESULTS

Sample ID: 17030077-002

Customer ID: LICA

Cust Samp ID: LICA/PUF/Bonnyville/Mar
2, 2017

TISCH PUF PLUS Sample Collection Data Sheet

Client:	<u>LICA</u>	Puf+ S/N:	<u>TE-02</u>
Location:	<u>Bonnyville - AER</u>	Motor S/N:	<u>1139/100-1015</u>
Station ID:	<u>LICA 37</u>	Installation Date/Time:	<u>Feb 27, 2017/12:56</u>
Field Sample ID:	<u>LICA/PUF/Bonnyville/Mar 02, 2017</u>	Removal Date/Time:	<u>Mar 07, 2017/14:59</u>


Sample Data Collection Information

Sample Date:	<u>Mar 02, 2017</u>	Average Pressure (mmHg)	<u>708</u>
Start Time (mst):	<u>00:00</u>	Average Flow (Q _{std})	<u>229</u>
End Time (mst):	<u>00:00 Mar 03, 2017</u>	Average Temperature (°C)	<u>-11.7°</u>
Elapsed Time (Hours):	<u>24.0</u>	Volume (V _{std} m ³)	<u>330.15</u>

Sample Recovery Checklist

(circle one)

Flow Rate 230 slpm +/- 0.2 slpm ?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Average temperature appears correct?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Average pressure appears correct?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Any error messages? (if yes list below)	<input type="radio"/> YES	<input checked="" type="radio"/> NO
Sample duration 24 hours?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Date of last calibration/audit:	<u>Jan 27, 2017</u>	
Other observations?	<u>n/a</u>	



Deployed By:	<u>Alex Yakupov</u>	
Collected By:	<u>Alex Yakupov</u>	<u>Date: Mar 07, 2017</u>

Polycyclic Aromatic Hydrocarbons (PAHs) Data Results

Date: March 2 , 2017
PUF S/N: TE02

PARAMETERS	CONCENTRATION (µg/puf)
1-Methylnaphthalene	0.44
2-Methylnaphthalene	0.56
3-Methylcholanthrene	< 0.01
7,12-Dimethylbenz(a)anthracene	< 0.01
Acenaphthene	0.15
Acenaphthylene	0.04
Acridine	< 0.01
Anthracene	0.01
Benzo(a)anthracene	< 0.01
Benzo(a)pyrene	0.01
Benzo(b,j,k)fluoranthene	0.04
Benzo(c)phenanthrene	< 0.01
Benzo(e)pyrene	0.01
Benzo(ghi)perylene	< 0.01
Chrysene	0.02
Dibenzo(a,h)pyrene	< 0.01
Dibenzo(a,i)pyrene	< 0.01
Dibenzo(a,l)pyrene	< 0.01
Dibenzo(ah)anthracene	< 0.01
Fluoranthene	0.09
Fluorene	0.13
Indeno(1,2,3-cd)pyrene	0.01
Naphthalene	0.33
Perylene	< 0.01
Phenanthrene	0.31
Pyrene	0.06
Retene	0.05

Sample ID: 17030116-002

Customer ID: LICA

Cust Samp ID: LICA/PUF/Bonnyville/Mar 08, 2017



TISCH PUF PLUS Sample Collection Data Sheet

Client:	<u>LICA</u>	Puf+ S/N:	<u>9801</u>
Location:	<u>Bonnyville - AER</u>	Motor S/N:	<u>1139/100-1015</u>
Station ID:	<u>LICA 37</u>	Installation Date/Time:	<u>Mar 07, 2017/14:59</u>
Field Sample ID:	<u>LICA/PUF/Bonnyville/Mar 08, 2017</u>	Removal Date/Time:	<u>Mar 09, 2017/14:33</u>

Sample Data Collection Information

Sample Date:	<u>Mar 08, 2017</u>	Average Pressure (mmHg)	<u>714</u>
Start Time (mst):	<u>00:00</u>	Average Flow (Q _{std})	<u>229</u>
End Time (mst):	<u>Mar 09, 2017/00:00</u>	Average Temperature (°C)	<u>-21.5°</u>
Elapsed Time (Hours):	<u>24.0</u>	Volume (V _{std} m ³)	<u>330.17</u>

Sample Recovery Checklist

(circle one)

Flow Rate 230 slpm +/- 0.2 slpm ?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Average temperature appears correct?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Average pressure appears correct?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Any error messages? (if yes list below)	<input type="radio"/> YES	<input checked="" type="radio"/> NO
Sample duration 24 hours?	<input checked="" type="radio"/> YES	<input type="radio"/> NO
Date of last calibration/audit:	<u>Jan 27, 2017</u>	
Other observations?	<u>n/a</u>	

Deployed By: Alex Yakupov

Collected By: Alex Yakupov Date: Mar 09, 2017

Polycyclic Aromatic Hydrocarbons (PAHs) Data Results

Date: March 8 , 2017
PUF S/N: 9801

PARAMETERS	CONCENTRATION (µg/puf)
1-Methylnaphthalene	1.21
2-Methylnaphthalene	2.07
3-Methylcholanthrene	< 0.01
7,12-Dimethylbenz(a)anthracene	< 0.01
Acenaphthene	0.34
Acenaphthylene	0.12
Acridine	< 0.01
Anthracene	0.01
Benzo(a)anthracene	0.02
Benzo(a)pyrene	0.01
Benzo(b,j,k)fluoranthene	0.06
Benzo(c)phenanthrene	< 0.01
Benzo(e)pyrene	0.02
Benzo(ghi)perylene	0.03
Chrysene	0.04
Dibenzo(a,h)pyrene	< 0.01
Dibenzo(a,i)pyrene	0.06
Dibenzo(a,l)pyrene	< 0.01
Dibenzo(ah)anthracene	< 0.01
Fluoranthene	0.09
Fluorene	0.12
Indeno(1,2,3-cd)pyrene	0.03
Naphthalene	3.89
Perylene	< 0.01
Phenanthrene	0.36
Pyrene	0.06
Retene	0.08

Sample ID: 17030212-002

Customer ID: LICA
Cust Samp ID: LICA/PUF/Bonnyville/Mar
14, 2017

TISCH PUF PLUS Sample Collection Data Sheet

Client:	<u>LICA</u>	Puf+ S/N:	<u>P13-01</u>
Location:	<u>Bonnyville - AER</u>	Motor S/N:	<u>1139/100-1015</u>
Station ID:	<u>LICA 37</u>	Installation Date/Time:	<u>Mar 09, 2017 / 14:33</u>
Field Sample ID:	<u>LICA/PUF/Bonnyville / Mar 14, 2017</u>	Removal Date/Time:	<u>MAR 16, 2017 @ 11:50</u>

Sample Data Collection Information

Sample Date:	<u>Mar 14, 2017</u>	Average Pressure (mmHg)	<u>701</u>
Start Time (mst):	<u>00:00</u>	Average Flow (Q _{std})	<u>229</u>
End Time (mst):	<u>00:00 Mar 15, 2017</u>	Average Temperature (°C)	<u>-4.7</u>
Elapsed Time (Hours):	<u>24.0</u>	Volume (Vstd m ³)	<u>330.18</u>

Sample Recovery Checklist

(circle one)

Flow Rate 230 slpm +/- 0.2 slpm ?	<u>YES</u>	NO
Average temperature appears correct?	<u>YES</u>	NO
Average pressure appears correct?	<u>YES</u>	NO
Any error messages? (if yes list below)	<u>YES</u>	NO
Sample duration 24 hours?	<u>YES</u>	NO
Date of last calibration/audit:	<u>Jan 27, 2017</u>	
Other observations?		

Deployed By: Alex Yakupov

Collected By: [Signature] Date: Mar 16, 2017



Polycyclic Aromatic Hydrocarbons (PAHs) Data Results

Date: March 14, 2017
PUF S/N: P13-01

PARAMETERS	CONCENTRATION (µg/puf)
1-Methylnaphthalene	0.11
2-Methylnaphthalene	0.12
3-Methylcholanthrene	< 0.01
7,12-Dimethylbenz(a)anthracene	< 0.01
Acenaphthene	0.07
Acenaphthylene	0.06
Acridine	< 0.01
Anthracene	0.02
Benzo(a)anthracene	0.01
Benzo(a)pyrene	< 0.01
Benzo(b,j,k)fluoranthene	0.03
Benzo(c)phenanthrene	< 0.01
Benzo(e)pyrene	0.01
Benzo(ghi)perylene	0.01
Chrysene	< 0.01
Dibenzo(a,h)pyrene	< 0.01
Dibenzo(a,i)pyrene	< 0.01
Dibenzo(a,l)pyrene	< 0.01
Dibenzo(ah)anthracene	< 0.01
Fluoranthene	0.1
Fluorene	0.1
Indeno(1,2,3-cd)pyrene	0.02
Naphthalene	0.11
Perylene	< 0.01
Phenanthrene	0.31
Pyrene	0.08
Retene	0.04

Sample ID: 17030303-002

Customer ID: LICA

Cust Samp ID: LICA/PUF/Bonnyville/Mar
20, 2017

ISCH PUF PLUS Sample Collection Data Sheet

Client:	<u>LICA</u>	Puf+ S/N:	<u>TE09</u>
Location:	<u>Bonnyville - AER</u>	Motor S/N:	<u>1139/100-1015</u>
Station ID:	<u>LICA 37</u>	Installation Date/Time:	<u>MAR 16, 2017 @ 11:52</u>
Field Sample ID:	<u>LICA/PUF/Bonnyville/MAR 20, 2017</u>	Removal Date/Time:	<u>MAR 24, 2017 @</u>


Sample Data Collection Information

Sample Date:	<u>MAR 20, 2017</u>	Average Pressure (mmHg)	<u>706</u>
Start Time (mst):	<u>00:00</u>	Average Flow (Q _{std})	<u>229</u>
End Time (mst):	<u>24:00</u>	Average Temperature (°C)	<u>-7.8</u>
Elapsed Time (Hours):	<u>24:00</u>	Volume (Vstd m ³)	<u>330.18</u>

Sample Recovery Checklist

(circle one)

Flow Rate 230 slpm +/- 0.2 slpm ?	YES	NO
Average temperature appears correct?	YES	NO
Average pressure appears correct?	YES	NO
Any error messages? (if yes list below)	YES	NO
Sample duration 24 hours?	YES	NO
Date of last calibration/audit:	<u>JAN 27, 2017</u>	
Other observations?		



Deployed By: [Signature]

Collected By: [Signature]

Polycyclic Aromatic Hydrocarbons (PAHs) Data Results

Date: March 20, 2017
PUF S/N: TE-09

PARAMETERS	CONCENTRATION (µg/puf)
1-Methylnaphthalene	0.21
2-Methylnaphthalene	0.4
3-Methylcholanthrene	< 0.01
7,12-Dimethylbenz(a)anthracene	< 0.01
Acenaphthene	0.03
Acenaphthylene	0.01
Acridine	< 0.01
Anthracene	< 0.01
Benzo(a)anthracene	< 0.01
Benzo(a)pyrene	< 0.01
Benzo(b,j,k)fluoranthene	0.01
Benzo(c)phenanthrene	< 0.01
Benzo(e)pyrene	< 0.01
Benzo(ghi)perylene	< 0.01
Chrysene	0.01
Dibenzo(a,h)pyrene	< 0.01
Dibenzo(a,i)pyrene	< 0.01
Dibenzo(a,l)pyrene	< 0.01
Dibenzo(ah)anthracene	< 0.01
Fluoranthene	0.02
Fluorene	0.03
Indeno(1,2,3-cd)pyrene	< 0.01
Naphthalene	0.45
Perylene	< 0.01
Phenanthrene	0.07
Pyrene	0.02
Retene	0.01

Sample ID: 17030303-004

Customer ID: LICA

Cust Samp ID: LICA/PUF/Bonnyville/Mar 26, 2017

TISCH PUF PLUS Sample Collection Data Sheet

Client: LICA Puf+ S/N: TE 09 01
Location: Bonnyville - AER Motor S/N: 1139 / 100-1015
Station ID: LICA 37 Installation Date/Time: MARCH 24, 2017
Field Sample ID: LICA / PUF / Bonnyville / Mar 26, 2017 Removal Date/Time: Mar 27, 2017 / 11:39

Sample Data Collection Information

Sample Date: MARCH 26, 2017 Average Pressure (mmHg) 700
Start Time (mst): 00:00 Average Flow (Q_{std}) 229 ~~330.20~~ A.V.
End Time (mst): 24:00 Average Temperature (°C) + 1.6°
Elapsed Time (Hours): 24:00 Volume (Vstd m³) 330.20

Sample Recovery Checklist

(circle one)

Flow Rate 230 slpm +/- 0.2 slpm ? YES NO
Average temperature appears correct? YES NO
Average pressure appears correct? YES NO
Any error messages? (if yes list below) YES NO
Sample duration 24 hours? YES NO
Date of last calibration/audit: JANUARY 27, 2017
Other observations? n/a



Deployed By: [Signature]
Collected By: Alex Yakupov Date: Mar 27, 2017

Polycyclic Aromatic Hydrocarbons (PAHs) Data Results

Date: March 26, 2017
PUF S/N: TE01

PARAMETERS	CONCENTRATION (µg/puf)
1-Methylnaphthalene	0.08
2-Methylnaphthalene	0.14
3-Methylcholanthrene	< 0.01
7,12-Dimethylbenz(a)anthracene	< 0.01
Acenaphthene	0.02
Acenaphthylene	0.01
Acridine	< 0.01
Anthracene	< 0.01
Benzo(a)anthracene	< 0.01
Benzo(a)pyrene	< 0.01
Benzo(b,j,k)fluoranthene	0.01
Benzo(c)phenanthrene	< 0.01
Benzo(e)pyrene	< 0.01
Benzo(ghi)perylene	< 0.01
Chrysene	0.01
Dibenzo(a,h)pyrene	< 0.01
Dibenzo(a,i)pyrene	< 0.01
Dibenzo(a,l)pyrene	< 0.01
Dibenzo(ah)anthracene	< 0.01
Fluoranthene	0.05
Fluorene	0.07
Indeno(1,2,3-cd)pyrene	< 0.01
Naphthalene	0.12
Perylene	< 0.01
Phenanthrene	0.16
Pyrene	0.04
Retene	0.06

NMHC CANISTER RESULTS

Sample ID: 17030077-004

AIR FCD-01320/2

Customer ID: LICA

Cust Samp ID: LICA/NMHC
VOC/Bonnyville/Mar 3, 2017

Maxxam

VOC Sample Collection Data Sheet



Client: LICA

Sampler S/N: n/a

Location: Bonnyville - AER

Canister ID: H 3286

Station ID: LICA 37

Canister Installation Date/Time: Mar 03, 2017 / 16:05

Field Sample ID: LICA/NMHC/VOC/Bonnyville
Mar 03, 2017

Canister Removal Date/Time: Mar 04, 2017 / 15:19

Date and Time Information			
Sample Date	Start Time (MST)	End Time (MST)	Elapsed Time (Hours)
Mar 03, 2017	19:25	n/a	n/a

Flow Settings		
Meter Reading (sccm)	Pot Set Pt.	Pump Pressure Setting (psig)
n/a	n/a	n/a

Canister Information	
Initial Canister Vacuum (inHg)	Final Canister Vacuum (inHg)
-27.0	-2.0

Canister valve open prior to sampling?: YES / NO

Canister valve closed prior to disconnection?: YES / NO

Comments: NMHC canister

Technician Signature: Alex Yakupov

Date: Mar 04, 2017

Volatile Organics Data Results (NMHC Canister System)

Date: March 3, 2017
Canister ID: H3286

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.03
1,1,2,2-Tetrachloroethane	< 0.03
1,1,2-Trichloroethane	< 0.03
1,1-Dichloroethane	< 0.03
1,1-Dichloroethylene	< 0.05
1,2,3-Trimethylbenzene	0.11
1,2,4-Trichlorobenzene	< 1.1
1,2,4-Trimethylbenzene	0.37
1,2-Dibromoethane	< 0.03
1,2-Dichlorobenzene	< 0.04
1,2-Dichloroethane	0.03
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	0.14
1,3-Butadiene	0.17
1,3-Dichlorobenzene	< 0.4
1,4-Dichlorobenzene	< 0.5
1,4-Dioxane	< 0.5
1-Butene	0.52
1-Hexene	0.09
1-Pentene	0.11
2,2,4-Trimethylpentane	0.57
2,2-Dimethylbutane	0.22
2,3,4-Trimethylpentane	0.15
2,3-Dimethylbutane	0.36
2,3-Dimethylpentane	0.43
2,4-Dimethylpentane	0.2
2-Methylheptane	0.19
2-Methylhexane	0.42
2-Methylpentane	1.16
3-Methylheptane	0.16
3-Methylhexane	0.48
3-Methylpentane	0.65
Acetone	4.0
Acrolein	< 0.4
Benzene	0.69
Benzyl chloride	< 0.5
Bromodichloromethane	< 0.03
Bromoform	< 0.03
Bromomethane	< 0.01
Carbon disulfide	0.3
Carbon tetrachloride	0.14
Chlorobenzene	< 0.03
Chloroethane	< 0.03
Chloroform	0.04
Chloromethane	0.72
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.05
cis-2-Butene	0.18
cis-2-Pentene	0.1
Cyclohexane	0.48
Cyclopentane	0.29
Dibromochloromethane	< 0.01
Ethanol	10.1
Ethyl acetate	< 0.5
Ethylbenzene	0.26
Freon-11	0.31
Freon-113	0.15

Volatile Organics Data Results (NMHC Canister System)

Date: March 3, 2017
Canister ID: H3286

PARAMETERS	CONCENTRATION (PPB)
Freon-114	< 0.03
Freon-12	0.59
Hexachloro-1,3-butadiene	< 0.68
Isobutane	3.02
Isopentane	3.59
Isoprene	0.1
Isopropyl alcohol	< 0.5
Isopropylbenzene	0.03
m,p-Xylene	0.97
m-Diethylbenzene	< 0.05
m-Ethyltoluene	0.24
Methyl butyl ketone	< 0.68
Methyl ethyl ketone	< 0.4
Methyl isobutyl ketone	< 0.5
Methyl methacrylate	< 0.09
Methyl tert butyl ether	< 0.04
Methylcyclohexane	0.58
Methylcyclopentane	0.79
Methylene chloride	< 0.4
n-Butane	6.24
n-Decane	< 0.08
n-Dodecane	< 0.5
n-Heptane	0.44
n-Hexane	0.71
n-Nonane	0.17
n-Octane	1.7
n-Pentane	0.07
n-Propylbenzene	< 0.7
n-Undecane	1.7
Naphthalene	0.1
o-Ethyltoluene	0.1
o-Xylene	0.34
p-Diethylbenzene	0.1
p-Ethyltoluene	0.15
Styrene	0.08
Tetrachloroethylene	0.3
Tetrahydrofuran	< 0.5
Toluene	1.06
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.05
trans-2-Butene	0.21
trans-2-Pentene	0.21
Trichloroethylene	< 0.05
Vinyl acetate	< 0.5
Vinyl chloride	< 0.03

Sample ID: 17030116-003

AIR FCD-01320/2

Customer ID: LICA
Cust Samp ID: LICA/NMHC/Bonnyville/Mar 08, 2017

Maxxam

VOC Sample Collection Data Sheet

Client: LICA Sampler S/N: n/a
 Location: Bonnyville - AER Canister ID: 1061
 Station ID: LICA 37 Canister Installation Date/Time: Mar 04, 2017 / 15:19
 Field Sample ID: LICA/NMHC VOC/Bonnyville/Mar 08, 2017 Canister Removal Date/Time: Mar 09, 2017 / 13:41 A.V.

Date and Time Information			
Sample Date	Start Time (MST)	End Time (MST)	Elapsed Time (Hours)
<u>Mar 08, 2017</u>	<u>19:05</u>	<u>n/a</u>	<u>n/a</u>

Flow Settings		
Meter Reading (sccm)	Pot Set Pt.	Pump Pressure Setting (psig)
<u>n/a</u>	<u>n/a</u>	<u>n/a</u>

Canister Information	
Initial Canister Vacuum (inHg)	Final Canister Vacuum (inHg)
<u>-27.7</u>	<u>n/a</u>

- no pressure gauge (A.V.)

Canister valve open prior to sampling?: YES / NO

Canister valve closed prior to disconnection?: YES / NO

Comments: NMHC canister
The canister is not equipped with a pressure gauge.

Technician Signature: Alex Yampov Date: Mar 09, 2017



Volatile Organics Data Results (NMHC Canister System)

Date: March 8, 2017
Canister ID: 1061

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.02
1,1,2,2-Tetrachloroethane	< 0.02
1,1,2-Trichloroethane	< 0.02
1,1-Dichloroethane	< 0.02
1,1-Dichloroethylene	< 0.04
1,2,3-Trimethylbenzene	0.11
1,2,4-Trichlorobenzene	< 0.8
1,2,4-Trimethylbenzene	0.36
1,2-Dibromoethane	< 0.02
1,2-Dichlorobenzene	< 0.03
1,2-Dichloroethane	0.03
1,2-Dichloropropane	0.01
1,3,5-Trimethylbenzene	0.13
1,3-Butadiene	0.1
1,3-Dichlorobenzene	< 0.3
1,4-Dichlorobenzene	< 0.4
1,4-Dioxane	< 0.4
1-Butene	0.47
1-Hexene	0.09
1-Pentene	0.09
2,2,4-Trimethylpentane	0.75
2,2-Dimethylbutane	0.06
2,3,4-Trimethylpentane	0.31
2,3-Dimethylbutane	0.18
2,3-Dimethylpentane	0.58
2,4-Dimethylpentane	0.25
2-Methylheptane	0.21
2-Methylhexane	0.44
2-Methylpentane	0.56
3-Methylheptane	0.2
3-Methylhexane	0.52
3-Methylpentane	0.34
Acetone	1.9
Acrolein	< 0.3
Benzene	0.57
Benzyl chloride	< 0.4
Bromodichloromethane	< 0.02
Bromoform	< 0.02
Bromomethane	0.01
Carbon disulfide	< 0.01
Carbon tetrachloride	0.12
Chlorobenzene	< 0.02
Chloroethane	< 0.02
Chloroform	0.04
Chloromethane	0.72
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.04
cis-2-Butene	0.14
cis-2-Pentene	0.09
Cyclohexane	0.17
Cyclopentane	0.1
Dibromochloromethane	< 0.01
Ethanol	2.6
Ethyl acetate	< 0.4
Ethylbenzene	0.31
Freon-11	0.29
Freon-113	0.13

Volatile Organics Data Results (NMHC Canister System)

Date: March 8, 2017
Canister ID: 1061

PARAMETERS	CONCENTRATION (PPB)
Freon-114	0.02
Freon-12	0.52
Hexachloro-1,3-butadiene	< 0.50
Isobutane	0.83
Isopentane	1.1
Isoprene	0.09
Isopropyl alcohol	< 0.4
Isopropylbenzene	0.04
m,p-Xylene	1.11
m-Diethylbenzene	< 0.04
m-Ethyltoluene	0.25
Methyl butyl ketone	< 0.50
Methyl ethyl ketone	< 0.3
Methyl isobutyl ketone	< 0.4
Methyl methacrylate	< 0.07
Methyl tert butyl ether	< 0.03
Methylcyclohexane	0.3
Methylcyclopentane	0.41
Methylene chloride	< 0.3
n-Butane	2.18
n-Decane	0.09
n-Dodecane	< 0.4
n-Heptane	0.47
n-Hexane	0.37
n-Nonane	0.1
n-Octane	0.22
n-Pentane	0.5
n-Propylbenzene	0.08
n-Undecane	< 0.5
Naphthalene	1.5
o-Ethyltoluene	0.1
o-Xylene	0.4
p-Diethylbenzene	0.08
p-Ethyltoluene	0.12
Styrene	0.06
Tetrachloroethylene	< 0.04
Tetrahydrofuran	< 0.4
Toluene	1.29
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.04
trans-2-Butene	0.13
trans-2-Pentene	0.15
Trichloroethylene	< 0.04
Vinyl acetate	< 0.4
Vinyl chloride	< 0.02

Sample ID: 17030212-003

Customer ID: LICA

Cust Samp ID: LICA/NMHC
VOC/Bonnyville/Mar 14,
2017



Maxxam

VOC Sample Collection Data Sheet

Client: LICA Sampler S/N: n/a
 Location: Bonnyville - AER Canister ID: 17121
 Station ID: LICA 37 Canister Installation Date/Time: Mar 09, 2017 / 13:41
 Field Sample ID: LICA/NMHC VOC/Bonnyville/MAR 11, 2017 Canister Removal Date/Time: Mar 16, 2017 @ 12:21

Date and Time Information			
Sample Date	Start Time (MST)	End Time (MST)	Elapsed Time (Hours)
<u>Mar 11, 2017</u>	<u>16:35</u>	<u>n/a</u>	<u>n/a</u>

Flow Settings		
Meter Reading (sccm)	Pot Set Pt.	Pump Pressure Setting (psig)
<u>n/a</u>	<u>n/a</u>	<u>n/a</u>

Canister Information	
Initial Canister Vacuum (inHg)	Final Canister Vacuum (inHg)
<u>-27.7</u>	<u>0</u>

Canister valve open prior to sampling?: YES / NO

Canister valve closed prior to disconnection?: YES / NO

Comments:

NMHC canister

5mm NMHC core = 1.49ppm.

Technician Signature:

Alex Vakepov (Installation)

Date: Mar 16, 2017

(of removal)

Volatile Organics Data Results (NMHC Canister System)

Date: March 11, 2017
Canister ID: 17121

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.02
1,1,2,2-Tetrachloroethane	< 0.02
1,1,2-Trichloroethane	< 0.02
1,1-Dichloroethane	< 0.02
1,1-Dichloroethylene	< 0.05
1,2,3-Trimethylbenzene	< 0.06
1,2,4-Trichlorobenzene	< 1.0
1,2,4-Trimethylbenzene	< 0.06
1,2-Dibromoethane	< 0.02
1,2-Dichlorobenzene	< 0.04
1,2-Dichloroethane	0.02
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	0.04
1,3-Butadiene	< 0.02
1,3-Dichlorobenzene	< 0.4
1,4-Dichlorobenzene	< 0.5
1,4-Dioxane	< 0.5
1-Butene	2.38
1-Hexene	0.1
1-Pentene	0.29
2,2,4-Trimethylpentane	0.28
2,2-Dimethylbutane	0.03
2,3,4-Trimethylpentane	0.06
2,3-Dimethylbutane	0.33
2,3-Dimethylpentane	0.37
2,4-Dimethylpentane	0.2
2-Methylheptane	0.02
2-Methylhexane	0.12
2-Methylpentane	0.73
3-Methylheptane	< 0.02
3-Methylhexane	0.13
3-Methylpentane	0.43
Acetone	2.9
Acrolein	< 0.4
Benzene	1.45
Benzyl chloride	< 0.5
Bromodichloromethane	< 0.02
Bromoform	< 0.02
Bromomethane	< 0.01
Carbon disulfide	< 0.01
Carbon tetrachloride	0.1
Chlorobenzene	< 0.02
Chloroethane	< 0.02
Chloroform	< 0.02
Chloromethane	0.65
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.05
cis-2-Butene	1.64
cis-2-Pentene	0.3
Cyclohexane	0.05
Cyclopentane	0.09
Dibromochloromethane	< 0.01
Ethanol	5.7
Ethyl acetate	< 0.5
Ethylbenzene	0.04
Freon-11	0.29
Freon-113	0.09

Volatile Organics Data Results (NMHC Canister System)

Date: March 11, 2017
Canister ID: 17121

PARAMETERS	CONCENTRATION (PPB)
Freon-114	< 0.02
Freon-12	0.63
Hexachloro-1,3-butadiene	< 0.62
Isobutane	18.9
Isopentane	9.79
Isoprene	0.01
Isopropyl alcohol	< 0.5
Isopropylbenzene	< 0.01
m,p-Xylene	0.12
m-Diethylbenzene	< 0.05
m-Ethyltoluene	< 0.10
Methyl butyl ketone	< 0.62
Methyl ethyl ketone	0.4
Methyl isobutyl ketone	< 0.5
Methyl methacrylate	< 0.09
Methyl tert butyl ether	< 0.04
Methylcyclohexane	0.06
Methylcyclopentane	0.32
Methylene chloride	< 0.4
n-Butane	43.8
n-Decane	< 0.08
n-Dodecane	< 0.5
n-Heptane	0.09
n-Hexane	0.22
n-Nonane	0.02
n-Octane	0.03
n-Pentane	1.5
n-Propylbenzene	< 0.06
n-Undecane	< 0.6
Naphthalene	< 0.6
o-Ethyltoluene	0.02
o-Xylene	0.05
p-Diethylbenzene	< 0.05
p-Ethyltoluene	< 0.09
Styrene	< 0.05
Tetrachloroethylene	< 0.05
Tetrahydrofuran	< 0.5
Toluene	0.38
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.05
trans-2-Butene	1.96
trans-2-Pentene	0.54
Trichloroethylene	< 0.05
Vinyl acetate	0.8
Vinyl chloride	< 0.02

Sample ID: 17030303-005

AIR FCD-01320/2

Customer ID: LICA
Cust Samp ID: LICA/NMHC
VOC/Bonnyville/Mar 19,
2017

Maxxam

VOC Sample Collection Data Sheet



Client: LICA Sampler S/N: n/a
Location: Bonnyville - AER Canister ID: 1684
Station ID: LICA 37 Canister Installation Date/Time: 16-Mar-2017 @ 11:28 MST
Field Sample ID: LICA/NMHC VOC/Bonnyville/March 19, 2017 Canister Removal Date/Time: 24 MARCH 2017 @ 07:58 MST

Date and Time Information			
Sample Date	Start Time (MST)	End Time (MST)	Elapsed Time (Hours)
<u>MARCH 19, 2017</u>	<u>06:55</u>	<u>n/a</u>	<u>n/a</u>

Flow Settings		
Meter Reading (sccm)	Pot Set Pt.	Pump Pressure Setting (psig)
<u>n/a</u>	<u>n/a</u>	<u>n/a</u>

Canister Information	
Initial Canister Vacuum (inHg)	Final Canister Vacuum (inHg)
<u>-28</u>	<u>0</u>

Canister valve open prior to sampling?: YES / NO

Canister valve closed prior to disconnection?: YES / NO

Comments: NMHC canister

Technician Signature: Installed: [Signature]

Date: Mar 16, 2017

Removed: [Signature]

Date: March 24, 2017

Volatile Organics Data Results (NMHC Canister System)

Date: March 19, 2017
Canister ID: 1684

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.03
1,1,2,2-Tetrachloroethane	< 0.03
1,1,2-Trichloroethane	< 0.03
1,1-Dichloroethane	< 0.03
1,1-Dichloroethylene	< 0.06
1,2,3-Trimethylbenzene	< 0.07
1,2,4-Trichlorobenzene	< 1.1
1,2,4-Trimethylbenzene	< 0.07
1,2-Dibromoethane	< 0.03
1,2-Dichlorobenzene	< 0.04
1,2-Dichloroethane	< 0.01
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	< 0.03
1,3-Butadiene	< 0.03
1,3-Dichlorobenzene	< 0.4
1,4-Dichlorobenzene	< 0.6
1,4-Dioxane	< 0.6
1-Butene	0.1
1-Hexene	< 0.03
1-Pentene	< 0.01
2,2,4-Trimethylpentane	0.02
2,2-Dimethylbutane	< 0.01
2,3,4-Trimethylpentane	< 0.01
2,3-Dimethylbutane	< 0.03
2,3-Dimethylpentane	< 0.03
2,4-Dimethylpentane	< 0.01
2-Methylheptane	< 0.01
2-Methylhexane	0.03
2-Methylpentane	0.08
3-Methylheptane	< 0.03
3-Methylhexane	0.03
3-Methylpentane	0.03
Acetone	1.6
Acrolein	< 0.4
Benzene	0.13
Benzyl chloride	< 0.6
Bromodichloromethane	< 0.03
Bromoform	< 0.03
Bromomethane	< 0.01
Carbon disulfide	0.04
Carbon tetrachloride	0.07
Chlorobenzene	< 0.03
Chloroethane	< 0.03
Chloroform	< 0.03
Chloromethane	0.6
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.06
cis-2-Butene	< 0.03
cis-2-Pentene	< 0.03
Cyclohexane	< 0.03
Cyclopentane	0.01
Dibromochloromethane	< 0.01
Ethanol	0.9
Ethyl acetate	< 0.6
Ethylbenzene	< 0.01
Freon-11	0.22
Freon-113	0.09

Volatile Organics Data Results (NMHC Canister System)

Date: March 19, 2017
Canister ID: 1684

PARAMETERS	CONCENTRATION (PPB)
Freon-114	< 0.03
Freon-12	0.37
Hexachloro-1,3-butadiene	< 0.70
Isobutane	0.56
Isopentane	0.63
Isoprene	< 0.01
Isopropyl alcohol	< 0.6
Isopropylbenzene	< 0.01
m,p-Xylene	< 0.04
m-Diethylbenzene	< 0.06
m-Ethyltoluene	< 0.11
Methyl butyl ketone	< 0.70
Methyl ethyl ketone	< 0.4
Methyl isobutyl ketone	< 0.6
Methyl methacrylate	< 0.10
Methyl tert butyl ether	< 0.04
Methylcyclohexane	0.03
Methylcyclopentane	< 0.03
Methylene chloride	< 0.4
n-Butane	4.44
n-Decane	< 0.08
n-Dodecane	< 0.6
n-Heptane	0.03
n-Hexane	< 0.01
n-Nonane	< 0.01
n-Octane	< 0.03
n-Pentane	0.2
n-Propylbenzene	< 0.07
n-Undecane	< 0.7
Naphthalene	< 0.7
o-Ethyltoluene	< 0.01
o-Xylene	0.03
p-Diethylbenzene	< 0.06
p-Ethyltoluene	< 0.10
Styrene	< 0.06
Tetrachloroethylene	< 0.06
Tetrahydrofuran	< 0.6
Toluene	0.05
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.06
trans-2-Butene	0.04
trans-2-Pentene	< 0.03
Trichloroethylene	< 0.06
Vinyl acetate	< 0.6
Vinyl chloride	< 0.03

Sample ID: 17040067-003

Customer ID: LICA

Cust Samp ID: LICA/NMHC/Bonnyville/April 04, 2017

AIR FCD-01320/2



Maxxam

VOC Sample Collection Data Sheet

Client: LICA
 Location: Bonnyville AER
 Station ID: LICA 37
 Field Sample ID: LICA/NMHC/Bonnyville/Apr 04, 2017

Sampler S/N: n/a
 Canister ID: K517
 Canister Installation Date/Time: 24 MARCH 2017 @ 08:05 MST
 Canister Removal Date/Time: Apr 05, 2017 / 12:44

Date and Time Information			
Sample Date	Start Time (MST)	End Time (MST)	Elapsed Time (Hours)
<u>Apr 04, 2017</u>	<u>18:10</u>	<u>n/a</u>	<u>n/a</u>

Flow Settings		
Meter Reading (scm)	Pot Set Pt.	Pump Pressure Setting (psig)
<u>n/a</u>	<u>n/a</u>	<u>n/a</u>

Canister Information	
Initial Canister Vacuum (inHg)	Final Canister Vacuum (inHg)
<u>-28</u>	<u>0.0</u>

Canister valve open prior to sampling?: YES / NO

Canister valve closed prior to disconnection?: YES / NO

Comments:

NMHC canister

Technician Signiture: Installed: [Signature]
removed Alex Vakupov

Date: 24 MARCH 2017

Date: Apr 05, 2017

Volatile Organics Data Results (NMHC Canister System)

Date: March 28, 2017
Canister ID: 1517

PARAMETERS	CONCENTRATION (PPB)
1,1,1-Trichloroethane	< 0.03
1,1,2,2-Tetrachloroethane	< 0.03
1,1,2-Trichloroethane	< 0.03
1,1-Dichloroethane	< 0.03
1,1-Dichloroethylene	< 0.05
1,2,3-Trimethylbenzene	0.09
1,2,4-Trichlorobenzene	< 1.1
1,2,4-Trimethylbenzene	0.24
1,2-Dibromoethane	< 0.03
1,2-Dichlorobenzene	< 0.04
1,2-Dichloroethane	0.03
1,2-Dichloropropane	< 0.01
1,3,5-Trimethylbenzene	0.11
1,3-Butadiene	0.04
1,3-Dichlorobenzene	< 0.4
1,4-Dichlorobenzene	< 0.5
1,4-Dioxane	< 0.5
1-Butene	0.48
1-Hexene	0.03
1-Pentene	0.03
2,2,4-Trimethylpentane	0.07
2,2-Dimethylbutane	0.04
2,3,4-Trimethylpentane	0.05
2,3-Dimethylbutane	0.13
2,3-Dimethylpentane	0.15
2,4-Dimethylpentane	0.08
2-Methylheptane	0.23
2-Methylhexane	0.26
2-Methylpentane	0.71
3-Methylheptane	0.18
3-Methylhexane	0.32
3-Methylpentane	0.40
Acetone	6.3
Acrolein	< 0.4
Benzene	0.53
Benzyl chloride	< 0.5
Bromodichloromethane	< 0.03
Bromoform	< 0.03
Bromomethane	0.02
Carbon disulfide	0.26
Carbon tetrachloride	0.14
Chlorobenzene	< 0.03
Chloroethane	0.04
Chloroform	0.03
Chloromethane	0.68
cis-1,2-Dichloroethene	< 0.01
cis-1,3-Dichloropropene	< 0.05
cis-2-Butene	0.05
cis-2-Pentene	< 0.03
Cyclohexane	0.70
Cyclopentane	0.11
Dibromochloromethane	< 0.01
Ethanol	2.0
Ethyl acetate	< 0.5
Ethylbenzene	0.12
Freon-11	0.31
Freon-113	0.14

Volatile Organics Data Results (NMHC Canister System)

Date: March 28, 2017
Canister ID: 1517

PARAMETERS	CONCENTRATION (PPB)
Freon-114	< 0.03
Freon-12	0.51
Hexachloro-1,3-butadiene	< 0.68
Isobutane	1.06
Isopentane	1.34
Isoprene	0.02
Isopropyl alcohol	< 0.5
Isopropylbenzene	0.03
m,p-Xylene	0.70
m-Diethylbenzene	< 0.05
m-Ethyltoluene	< 0.11
Methyl butyl ketone	< 0.68
Methyl ethyl ketone	0.4
Methyl isobutyl ketone	< 0.5
Methyl methacrylate	< 0.09
Methyl tert butyl ether	< 0.04
Methylcyclohexane	1.17
Methylcyclopentane	0.46
Methylene chloride	< 0.4
n-Butane	2.73
n-Decane	0.19
n-Dodecane	< 0.5
n-Heptane	0.40
n-Hexane	0.58
n-Nonane	0.30
n-Octane	0.33
n-Pentane	1.2
n-Propylbenzene	< 0.07
n-Undecane	< 0.7
Naphthalene	< 0.7
o-Ethyltoluene	0.05
o-Xylene	0.24
p-Diethylbenzene	0.06
p-Ethyltoluene	< 0.09
Styrene	< 0.05
Tetrachloroethylene	0.06
Tetrahydrofuran	< 0.5
Toluene	0.98
trans-1,2-Dichloroethylene	< 0.01
trans-1,3-Dichloropropylene	< 0.05
trans-2-Butene	0.04
trans-2-Pentene	0.03
Trichloroethylene	0.07
Vinyl acetate	< 0.5
Vinyl chloride	< 0.03

APPENDIX III
EQUIPMENT CALIBRATION RESULTS

SULPHUR DIOXIDE



API 100E Sulphur Dioxide Analyzer Calibration

Date: <u>March 3, 2017</u>	Barometric Pressure: <u>27.73 inHg</u>
Company/Airshed: <u>LICA</u>	Station Temperature °C: <u>21</u>
Location/Station Name: <u>Bonnyville - AER</u>	Weather Conditions: <u>Mainly sunny</u>
Parameter: <u>Sulphur Dioxide</u>	Calibration Purpose: <u>shut down</u>
Start Time 24 hr. (mst): <u>9:10</u>	Performed By/Reviewer: <u>Limin Li</u> <u>Trina Whitsitt</u>
End Time 24 hr. (mst): <u>11:40</u>	Cal Gas Expiry Date: <u>December 25, 2018</u>
Calibration Method: <u>Gas Dilution</u>	Converter Model & s/n (if applicable): <u>n/a</u>

Analyzer:	
ID# or Serial Number: <u>467</u>	Range ppb: <u>1000</u>
Last Calibration Date: <u>February 17, 2017</u>	As Found C.F.: <u>1.012</u>
Previous C.F.: <u>1.000</u>	New C.F.: <u>n/a</u>

Calibrator: Flow Meter ID's: <u>n/a</u> Make & Model: <u>Sabio 2010</u> Serial #: <u>17200415</u> Cal Gas Cylinder I.D. #: <u>BLM002756T</u> Cal Gas Conc. (ppm): <u>49.9</u>	Standard Calibration Points for Ranges <table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>780</td></tr> <tr><td>Mid</td><td>380</td></tr> <tr><td>Low</td><td>190</td></tr> </table>	Point	ppb	High	780	Mid	380	Low	190
Point	ppb								
High	780								
Mid	380								
Low	190								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated Concentration:		Indicated Concentration:		Correction Factors (C.F.):
	Diluent	Cal Gas	Total	(ppb)	(ppb)			
as found zero	5028	0.00	5028	0.0	-0.8			n/a
as found high	4951	77.10	5028	765.2	755.0			1.012
mid	4991	37.70	5029	374.1	369.0			1.012
low	5009	19.00	5028	188.6	185.0			1.015
Average C.F.=								1.013

Linear Regression/Calibration Results:

Correlation Coefficient = <u>1.000</u>	LIMITS > or = 0.995
Slope = <u>1.012</u>	0.90-1.10
b (Intercept as % of full scale) = <u>0.09%</u>	± 3% F.S.
% change in C.F. from last cal = <u>-1.24%</u>	± 10%


API 100E Sulphur Dioxide Analyzer Calibration



As found: SLOPE: <u>0.952</u> OFFSET: <u>130.6</u> HVPS: <u>524</u> RCELL TEMP: <u>50.0</u> BOX TEMP: <u>32.2</u> PMT TEMP: <u>8.1</u> IZS TEMP: <u>50.0</u> PRES: <u>25.1</u> SAMP FL: <u>536</u> NORM PMT: <u>125</u> UV LAMP: <u>2699</u> LAMP RATIO: <u>97.1</u> STR. LGT: <u>62.2</u> DRK PMT: <u>15.8</u> DRK LMP: <u>2.7</u> Expected Value: <u>450.0</u>	As left: SLOPE: <u>n/a</u> OFFSET: <u>n/a</u> HVPS: <u>n/a</u> RCELL TEMP: <u>n/a</u> BOX TEMP: <u>n/a</u> PMT TEMP: <u>n/a</u> IZS TEMP: <u>n/a</u> PRES: <u>n/a</u> SAMP FL: <u>n/a</u> NORM PMT: <u>n/a</u> UV LAMP: <u>n/a</u> LAMP RATIO: <u>n/a</u> STR. LGT: <u>n/a</u> DRK PMT: <u>n/a</u> DRK LMP: <u>n/a</u> Expected Value: <u>n/a</u>
---	--

Comments:

After the shutdown calibration, the UV lamp was changed. A factory calibration was completed. A ZERO Air scrubber was re-positioned inside. Sample pump rebuilt.



— SO2[ppb]



API 100E Sulphur Dioxide Analyzer Calibration

Date: March 4, 2017	Barometric Pressure: 0.921 atm
Company/Airshed: LICA	Station Temperature °C: 22
Location/Station Name: Bonnyville - AER	Weather Conditions: Mainly cloudy with snow
Parameter: Sulphur Dioxide	Calibration Purpose: post repair
Start Time 24 hr. (mst): 9:17	Performed By/Reviewer: Alex Yakupov Trina Whitsitt
End Time 24 hr. (mst): 12:48	Cal Gas Expiry Date: July 18, 2019
Calibration Method: Gas Dilution	Converter Model & s/n (if applicable): n/a

Analyzer:	
ID# or Serial Number: 467	Range ppb: 1000
Last Calibration Date: n/a	As Found C.F.: n/a
Previous C.F.: n/a	New C.F.: 1.000


Calibrator:	Standard Calibration Points for Ranges								
Flow Meter ID's: n/a	<table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>780</td></tr> <tr><td>Mid</td><td>380</td></tr> <tr><td>Low</td><td>190</td></tr> </table>	Point	ppb	High	780	Mid	380	Low	190
Point	ppb								
High	780								
Mid	380								
Low	190								
Make & Model: API 700									
Serial #: 627									
Cal Gas Cylinder I.D. #: LL104222									
Cal Gas Conc. (ppm): 50.6									

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	(ppb)	(ppb)	
adjusted zero	5000	0.00	5000	0.0	0.0	n/a
adjusted high	4924	77.10	5001	780.1	780.0	1.000
mid	4964	37.50	5002	379.4	377.0	1.006
low	4979	18.80	4998	190.3	187.0	1.018
calibrator zero	5000	0.00	5000	0.0	0.0	n/a
Average C.F. =						1.008

Linear Regression/Calibration Results:


Correlation Coefficient = 1.000	LIMITS
Slope = 0.999	> or = 0.995
b (Intercept as % of full scale) = 0.18%	.95-1.05
% change in C.F. from last cal = n/a	± 3% F.S.
	± 10%



As found:	As left:
SLOPE: n/a	SLOPE: 1.047
OFFSET: n/a	OFFSET: 117.3
HVPS: n/a	HVPS: 488
RCELL TEMP: n/a	RCELL TEMP: 50.0
BOX TEMP: n/a	BOX TEMP: 71.7
PMT TEMP: n/a	PMT TEMP: 8.1
IZS TEMP: n/a	IZS TEMP: 50.0
PRES: n/a	PRES: 25.2
SAMP FL: n/a	SAMP FL: 560
NORM PMT: n/a	NORM PMT: 117.0
UV LAMP: n/a	UV LAMP: 4316.7
LAMP RATIO: n/a	LAMP RATIO: 99.6
STR. LGT n/a	STR. LGT 61.4
DRK PMT: n/a	DRK PMT: 15.5
DRK LMP: n/a	DRK LMP: 3.3
Expected Value: n/a	Expected Value: 414.5

Comments:
The analyzer sample inlet filter was changed.

A new UV lamp was installed. A factory calibration was completed. A ZERO Air scrubber was re-positioned inside. Sample pump rebuilt.



— SO2[ppb]

HYDROGEN SULPHIDE



API 101E Hydrogen Sulphide Analyzer Calibration

Date: <u>March 3, 2017</u>	Barometric Pressure: <u>27.73 inHg</u>
Company/Airshed: <u>LICA</u>	Station Temperature °C: <u>21</u>
Location/Station Name: <u>Bonnyville - AER</u>	Weather Conditions: <u>Mainly sunny</u>
Parameter: <u>Hydrogen Sulphide</u>	Calibration Purpose: <u>shut down</u>
Start Time 24 hr. (mst): <u>9:10</u>	Performed By/Reviewer: <u>Limin Li</u> <u>Trina Whitsitt</u>
End Time 24 hr. (mst): <u>11:20</u>	Cal Gas Expiry Date: <u>January 6, 2018</u>
Calibration Method: <u>Gas Dilution</u>	Converter Model & s/n (if applicable): <u>n/a</u>

Analyzer:	
ID# or Serial Number: <u>510</u>	Range ppb: <u>100</u>
Last Calibration Date: <u>February 8, 2017</u>	As Found C.F.: <u>1.019</u>
Previous C.F.: <u>0.998</u>	New C.F.: <u>n/a</u>

Calibrator: Flow Meter ID's: <u>n/a</u> Make & Model: <u>Enviroonica 2000</u> Serial #: <u>1991</u> Cal Gas Cylinder I.D. #: <u>BLM002508</u> Cal Gas Conc. (ppm): <u>10.2</u>	Standard Calibration Points for Ranges <table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19
Point	ppb								
High	78								
Mid	38								
Low	19								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated Concentration:		Indicated Concentration:		Correction Factors (C.F.):
	Diluent	Cal Gas	Total	(ppb)	(ppb)			
as found zero	7507	0.00	7507	0.0	0.2			n/a
as found high	7443	57.34	7500	78.0	76.7			1.019
mid	7473	27.92	7501	38.0	37.2			1.026
low	7502	13.92	7516	18.9	18.5			1.033
Average C.F.=								1.026

Linear Regression/Calibration Results:

Correlation Coefficient = <u>1.000</u>	LIMITS > or = 0.995
Slope = <u>1.019</u>	0.90-1.10
b (Intercept as % of full scale) = <u>-0.05%</u>	± 3% F.S.
% change in C.F. from last cal = <u>-2.14%</u>	± 10%


API 101E Hydrogen Sulphide Analyzer Calibration



As found: SLOPE: <u>0.951</u> OFFSET: <u>30.8</u> HVPS: <u>530</u> RCELL TEMP: <u>50.0</u> BOX TEMP: <u>35.1</u> PMT TEMP: <u>8.4</u> IZS TEMP: <u>45.0</u> Converter Temp: <u>314.4</u> PRES: <u>20.5</u> SAMP FL: <u>535</u> UV LAMP: <u>3426</u> LAMP RATIO: <u>90.3</u> STR. LGT: <u>14.6</u> DRK PMT: <u>35.0</u> DRK LMP: <u>-1.9</u> Expected Value: <u>50.6</u>	As left: SLOPE: <u>n/a</u> OFFSET: <u>n/a</u> HVPS: <u>n/a</u> RCELL TEMP: <u>n/a</u> BOX TEMP: <u>n/a</u> PMT TEMP: <u>n/a</u> IZS TEMP: <u>n/a</u> Converter Temp: <u>n/a</u> PRES: <u>n/a</u> SAMP FL: <u>n/a</u> UV LAMP: <u>n/a</u> LAMP RATIO: <u>n/a</u> STR. LGT: <u>n/a</u> DRK PMT: <u>n/a</u> DRK LMP: <u>n/a</u> Expected Value: <u>n/a</u>
--	--

Comments:

Pump:-19inHg. After the shutdown calibration, the reaction cell and sample valve were cleaned. The sinter filter was changed.



— H2S[ppb]



API 101E Hydrogen Sulphide Analyzer Calibration

Date: March 4, 2017	Barometric Pressure: 0.921 atm
Company/Airshed: LICA	Station Temperature °C: 22
Location/Station Name: Bonnyville - AER	Weather Conditions: Mainly cloudy with snow
Parameter: Hydrogen Sulphide	Calibration Purpose: post repair
Start Time 24 hr. (mst): 12:07	Performed By/Reviewer: Alex Yakupov Trina Whitsitt
End Time 24 hr. (mst): 15:45	Cal Gas Expiry Date: June 14, 2019
Calibration Method: Gas Dilution	Converter Model & s/n (if applicable): n/a

Analyzer:	
ID# or Serial Number: 510	Range ppb: 100
Last Calibration Date: n/a	As Found C.F.: n/a
Previous C.F.: n/a	New C.F.: 0.999

Calibrator:	Standard Calibration Points for Ranges								
Flow Meter ID's: n/a	<table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19
Point	ppb								
High	78								
Mid	38								
Low	19								
Make & Model: API 700									
Serial #: 627									
Cal Gas Cylinder I.D. #: EY0000654									
Cal Gas Conc. (ppm): 10.2									


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total	(ppb)	(ppb)	
adjusted zero	7500	0.00	7500	0.0	0.0	n/a
adjusted high	7443	57.40	7500	78.1	78.1	0.999
mid	7471	27.90	7499	37.9	37.3	1.017
low	7490	14.00	7504	19.0	18.3	1.040
calibrator zero	7500	0.00	7500	0.0	0.0	n/a
Average C.F. =						1.019

Linear Regression/Calibration Results:

Correlation Coefficient = 1.000	LIMITS
Slope = 0.997	> or = 0.995
b (Intercept as % of full scale) = 0.44%	.95-1.05
% change in C.F. from last cal = n/a	± 3% F.S.
	± 10%


API 101E Hydrogen Sulphide Analyzer Calibration



As found:	As left:
SLOPE: n/a	SLOPE: 1.039
OFFSET: n/a	OFFSET: 30.7
HVPS: n/a	HVPS: 530
RCELL TEMP: n/a	RCELL TEMP: 50.0
BOX TEMP: n/a	BOX TEMP: 35.0
PMT TEMP: n/a	PMT TEMP: 8.4
IZS TEMP: n/a	IZS TEMP: 45.0
Converter Temp: n/a	Converter Temp: 314.5
PRES: n/a	PRES: 20.7
SAMP FL: n/a	SAMP FL: 535
UV LAMP: n/a	UV LAMP: 3354.4
LAMP RATIO: n/a	LAMP RATIO: 100.0
STR. LGT: n/a	STR. LGT: 16.0
DRK PMT: n/a	DRK PMT: 35.7
DRK LMP: n/a	DRK LMP: -1.8
Expected Value: n/a	Expected Value: 49.3

Comments:
The analyzer sample inlet filter was changed.

Post-repair calibration completed after annual maintenance.



— H2S[ppb]

TOTAL HYDROCARBON



Thermo 55i Methane/Non-Methane Analyzer Calibration

Date:	March 3, 2017	Barometric Pressure:	27.74 inHg
Company/Airshed:	LICA	Station Temperature °C:	21
Location/Station Name:	Bonnyville - AER	Weather Conditions:	Mainly sunny
Parameter:	CH ₄ / NMHC / THC	Calibration Purpose:	routine monthly
Start/End Time 24 hr. (mst):	11:30 / 15:45	Performed By/Reviewer:	Limin Li / Trina Whitsitt
Calibration Method:	Gas Dilution	Cal Gas Expiry Date:	December 25, 2023

Analyzer:		Correction Factors:			
ID# or Serial Number:	1236656107	Previous C.F.:	As Found C.F.:	New C.F.:	
Measured Flow:	1.15 lpm	CH ₄ =	1.001	0.988	1.000
Last Calibration Date:	February 9, 2017	NMHC =	1.000	0.989	1.000
Range ppm:	20 CH ₄ /20 NMHC/40 THC	THC =	1.000	0.989	1.000

Calibrator:		Standard Calibration Points for Analyzer Range of 20/20/40 ppm			
Flow Meter ID's:	n/a	Point	CH ₄	NMHC	THC
Make & Model:	Enviroics 2000	High	13.00	13.00	26.00
Serial #:	1991	Mid	7.00	7.00	14.00
Cal Gas Cylinder I.D. #:	LL165372	Low	3.00	3.00	6.00
CH ₄ Cylinder Conc. =	606.0	212.0	=C ₃ H ₈ Cylinder Conc.		
CH ₄ as C ₃ H ₈ =	583.0	1189.0	=total CH ₄ equivalent		

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)										Correction Factors:		
Point	Diluent	Cal Gas	Total Flow	Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	CH ₄	NMHC	THC
as found zero	2503	0.00	2503	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
as found high	2503	62.00	2565	14.65	14.09	28.74	14.82	14.25	29.07	0.988	0.989	0.989
adjusted zero	2503	0.00	2503	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
adjusted high	2503	62.00	2565	14.65	14.09	28.74	14.65	14.09	28.74	1.000	1.000	1.000
mid	2503	30.50	2534	7.30	7.02	14.31	7.31	7.02	14.33	0.998	1.000	0.999
low	2503	13.50	2517	3.25	3.13	6.38	3.27	3.16	6.43	0.994	0.990	0.992
calibrator zero	2503	0.00	2503	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
										Average C.F. =		
										0.997	0.997	0.997




Linear Regression/Calibration Results:


	CH ₄	NMHC	THC	LIMITS
Correlation Coefficient =	1.000	1.000	1.000	> or = 0.995
Slope =	1.000	0.999	0.999	.95-1.05
b (Intercept as % of full scale) =	0.05%	0.07%	0.06%	± 3% F.S.
% change in C.F. from last cal =	1.26%	1.11%	1.14%	± 10%

As found:		As left:		
Interface Board Voltages:	Bias Supply: -292.8	Calibration History cnt'd:	NM Peak Area: 86981	
Temperatures:	Detector Oven: 175.0	Crucial Settings:	Methane Start: 8	
	Filter: 175.0		Methane End: 16	
	Column Oven: 75.2		Backflush: 18	
Cylinder Pressures/reg.:	Internal: 31.9	Run History>1:	NMHV Start: 24.5	
	Carrier: 1400		50	NMHC End: 56
	Fuel: 1450		55	Date: 03MAR17
	Span Gas: 1050		22	Time: 09:28
Internal Pressures:	Zero Air Generator: 55	CH ₄ PK HT: 3450		
	Carrier: 31.1	CH ₄ RT: 13.2		
	Fuel: 40.3	CH ₄ Baseline: 2451		
FID Status:	Air: 32.1	CH ₄ LOD: 52		
	Status: LIT	CH ₄ SD: 17		
	Counts: 28049	CH ₄ CONC: 2.57		
Flame and Power Stats:	Flame: 370.1	NM PK HT: 20		
	Det Base: 175	NM Peak Area: 338		
	Last Power On: August 3, 2016	NM CONC: 0.05		
	Flameouts: 3	NM Base Start: 2372		
Calibration History:	Det Oven at Start: 169.0	NM Base End: 2388		
	Col Oven at Start: 74.5	NM LOD: 12		
	Time: 09FEB17 12:09	NM Start IDX: 30		
	Type: SPAN	NM End IDX: 64		
	Status: GOOD	NM Max Slope: 1.6e+00		
	Check/Adjust: ADJUST	NM Min Slope: -1.4e+00		
	CH ₄ Span Conc: 13.62	NM PT Count: 28		
	CH ₄ SP Ratio: 0.000744	Expected Values:		
	CH ₄ RT: 13.2	Previous CH ₄ : 9.01		
	CH ₄ PK IDX: 26	Previous NMHC: 13.93		
CH ₄ PK HT: 18312	Previous THC: 22.97			
NM Span Conc: 13.11	New CH ₄ : 9.34			
NM SP Ratio: 0.000151	New NMHC: 13.77			
	New THC: 23.13			

Date: March 3, 2017
Company/Airshed: LICA
Location/Station Name: Bonnyville - AER

Start/End Time 24 hr. (mst): 11:30 / 15:45
Calibration Purpose: routine monthly
Calibration Method: Gas Dilution





— CH4[ppm] — NMHC[ppm]

NITROGEN DIOXIDE



API 200E NO-NO2-NOx Analyzer Calibration

Date: March 3, 2017	Barometric Pressure: 27.74 inHg
Company/Airshed: LICA	Station Temperature °C: 21
Location/Station Name: Bonnyville - AER	Weather Conditions: Mainly sunny
Start/End Time 24 hr. (mst): 09:10 / 15:30	Calibration Purpose: routine monthly
G.P.T. to be used for Ozone? Yes with 1000 ppb NOx full scale	Performed By/Reviewer: Limin Li Trina Whitsitt
Calibration Method: Gas Dilution & Gas Phase Titration	Cal Gas Expiry Date: December 25, 2018

Analyzer: ID# or Serial Number: 593 Last Calibration Date: February 24, 2017 Range ppb: 1000	Correction Factors: <table border="1" style="width: 100%; text-align: center;"> <thead> <tr> <th></th> <th>Previous C.F.:</th> <th>As Found C.F.:</th> <th>New C.F.:</th> </tr> </thead> <tbody> <tr> <td>NO =</td> <td>1.001</td> <td>1.032</td> <td>1.000</td> </tr> <tr> <td>NO₂ =</td> <td>1.000</td> <td>0.994</td> <td>1.002</td> </tr> <tr> <td>NOx =</td> <td>1.001</td> <td>1.031</td> <td>1.000</td> </tr> </tbody> </table>		Previous C.F.:	As Found C.F.:	New C.F.:	NO =	1.001	1.032	1.000	NO ₂ =	1.000	0.994	1.002	NOx =	1.001	1.031	1.000
	Previous C.F.:	As Found C.F.:	New C.F.:														
NO =	1.001	1.032	1.000														
NO ₂ =	1.000	0.994	1.002														
NOx =	1.001	1.031	1.000														

Calibrator: Flow Meter ID's: n/a Make & Model: Sabio 2010 Serial #: 17200415 Cal Gas Cylinder I.D. #: BLM002756T NO/NOx Gas Conc. (ppm): 50.7 50.7	Standard Calibration Points for a Range of: 1000 ppb <table border="1" style="width: 100%; text-align: center;"> <thead> <tr> <th>Point</th> <th>Target NO (ppb)</th> <th>Target NO₂ (ppb)</th> <th>Cc Ozone ?</th> </tr> </thead> <tbody> <tr> <td>High</td> <td>610</td> <td>375</td> <td><-high ozone</td> </tr> <tr> <td>Mid</td> <td>380</td> <td>190</td> <td><-mid ozone</td> </tr> <tr> <td>Low</td> <td>190</td> <td>70</td> <td><-low ozone</td> </tr> <tr> <td>Extra Point #1</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> <tr> <td>Extra Point #2</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> </tbody> </table>	Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?	High	610	375	<-high ozone	Mid	380	190	<-mid ozone	Low	190	70	<-low ozone	Extra Point #1	n/a	n/a	n/a	Extra Point #2	n/a	n/a	n/a
Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?																						
High	610	375	<-high ozone																						
Mid	380	190	<-mid ozone																						
Low	190	70	<-low ozone																						
Extra Point #1	n/a	n/a	n/a																						
Extra Point #2	n/a	n/a	n/a																						

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated NO	Calculated NOx	Indicated NO	Indicated NOx	NO C.F.	NOx C.F.
Point	Diluent	Cal Gas	Total Flow	(ppb)	(ppb)	(ppb)	(ppb)		
as found zero	5028	0.0	5028	0	0	0.4	0.6	n/a	n/a
as found high	4951	77.1	5028	777.4	777.4	754.0	755.0	1.032	1.031
adjusted zero	5028	0.00	5028	0.0	0.0	0.1	-0.2	n/a	n/a
adjusted high	4951	77.10	5028	777.4	777.4	777.4	777.4	1.000	1.000
mid	4991	37.70	5029	380.1	380.1	380.0	380.0	1.001	1.000
low	5009	19.00	5028	191.6	191.6	191.0	191.0	1.004	1.002
calibrator zero	5028	0.00	5028	0	0	0.2	-0.6	n/a	n/a
								Average C.F.=	1.001

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calibrator Setting	Indicated NO	Indicated NOx	Indicated NO ₂	NO drop	NO ₂ gain	NO ₂ C.F.
Point	Diluent	Cal Gas	Total Flow	volts or ppb	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)
NOx reference	4953	77.10	5030	0.0	779.0	779.0	0.0	0.1	0.0	
as found high NO ₂	4953	77.10	5030	510.0	266.0	782.0	516.0	513.0	516.0	0.994
adjusted high NO ₂	4953	77.10	5030	510.0	266.0	778.0	512.0	513.0	512.0	1.002
gpt mid	4953	77.10	5030	275.0	508.0	779.0	271.0	271.0	271.0	1.000
gpt low	4953	77.10	5030	100.0	684.0	778.0	94.0	95.0	94.0	1.011
									Average NO ₂ C.F.=	1.004

Linear Regression/Calibration Results:




	NO	NOx	NO ₂	LIMITS
Correlation Coefficient =	1.000	1.000	1.000	> or = 0.995
Slope =	1.000	1.000	1.001	.95-1.05
b (Intercept as % of full scale)=	-0.02%	-0.04%	-0.03%	± 3% F.S.
% change in C.F. from last cal=	-3.06%	-2.95%	0.58%	± 10%
NO ₂ converter efficiency			1.00	0.96 to 1.04


As found:	As left:
NOx SLOPE: 0.853	NOx SLOPE: 0.881
NOx OFFS: -0.9	NOx OFFS: 0.5
NO SLOPE: 0.856	NO SLOPE: 0.882
NO OFFS: -2.1	NO OFFS: -1.6
SAMP FLW: 472	SAMP FLW: 467
OZONE FL: 76	OZONE FL: 76
PMT: 91.1	PMT: 17.6
NORM PMT: 53.6	NORM PMT: 6.3
AZERO: 9.5	AZERO: 10.6
HVPS: 670	HVPS: 670
RCELL TEMP: 50	RCELL TEMP: 50.0
BOX TEMP: 31.2	BOX TEMP: 32.1
PMT TEMP: 6.7	PMT TEMP: 6.7
IZS TEMP: 45.2	IZS TEMP: 45.1
MOLY TEMP: 315.8	MOLY TEMP: 316
RCEL: 5.3	RCEL: 5.3
SAMP: 26.6	SAMP: 26.6
Expected Value NO: 8.6	Expected Value NO: 7.9
Expected Value NO ₂ : 415.0	Expected Value NO ₂ : 421.0
Expected Value NOx: 423.0	Expected Value NOx: 428.7

Comments:
 The analyzer sample inlet filter was changed.

Date: March 3, 2017
Company/Airshed: LICA
Location/Station Name: Bonnyville - AER

Start/End Time 24 hr. (mst): 09:10 / 15:30
Calibration Purpose: routine monthly
Calibration Method: Gas Dilution & Gas Phase Titration





— NOX[ppb] — NO[ppb] — NO2[ppb]

OZONE

Maxxam Thermo 49i Ozone Analyzer Calibration

A Bureau Veritas Group Company

Date: March 4, 2017 Barometric Pressure: 0.921 atm
 Company/Airshed: LICA Station Temperature °C: 22
 Location/Station Name: Bonnyville Weather Conditions: Mainly cloudy with snow
 Start/End Time 24 hr. (mst): 9:17 / 11:12 Calibration Purpose: shut down
 Ozone Calibration Method: Varying UV Lamp Power Performed By/Reviewer: Alex Yakupov Trina Whitsitt
 G.P.T. Date: n/a-done by Varying UV Lamp Power Cal Gas Expiry Date: n/a

Analyzer:
 ID# or Serial Number: 1002240372 Ozone Range ppb: 500
 Last Calibration Date: February 9, 2017 As Found C.F.: 0.999
 Previous Cal High Point C.F.: 1.000 New C.F.: n/a

Calibrator:
 Flow Meter ID's: n/a
 Make & Model: SABIO 2010 D
 Serial #: 11900613
 Cal Gas Cylinder I.D. #: n/a

Point	AMD Required Range of Ozone Calibration Points
High	300-400 ppb
Mid	150-200 ppb
Low	50-75 ppb


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rate (cc/min)		Calculated Concentration:	Corrected Calculated Concentration:	Indicated Concentration:	Correction Factors:
	Total Flow @ Point Start	Total Flow @ Point Finish	(ppb)	(ppb)	(ppb)	
as found zero	5000	5000	0.0	n/a	-0.4	n/a
as found high	5000	5000	380.0	380.0	380.0	0.999
mid	5000	5000	180.0	180.0	180.0	0.998
low	5000	5000	60.0	60.0	60.0	0.993
Average C.F. =						0.997

Linear Regression/Calibration Results:

Correlation Coefficient = 1.000 > or = 0.995
 Slope = 0.999 0.90-1.10
 b (Intercept as % of full scale) = 0.04% ± 3% F.S.
 % change in C.F. from last cal = 0.11% ± 10%

Thermo 49i Ozone Analyzer Calibration



As found:	As left:
O3 Bkg: <u>-0.1</u>	O3 Bkg: <u>n/a</u>
O3 Coef: <u>0.984</u>	O3 Coef: <u>n/a</u>
Photo Lamp: <u>14.2</u>	Photo Lamp: <u>n/a</u>
O3 Lamp: <u>5.8</u>	O3 Lamp: <u>n/a</u>
Bench: <u>28.9</u>	Bench: <u>n/a</u>
Bench Lamp: <u>54.0</u>	Bench Lamp: <u>n/a</u>
O3 Lamp: <u>68.1</u>	O3 Lamp: <u>n/a</u>
Pressure: <u>695.9</u>	Pressure: <u>n/a</u>
Cell A lpm: <u>0.734</u>	Cell A lpm: <u>n/a</u>
Cell B lpm: <u>0.746</u>	Cell B lpm: <u>n/a</u>
O3 ppb: <u>-0.2</u>	O3 ppb: <u>n/a</u>
Cell A ppb: <u>1.8</u>	Cell A ppb: <u>n/a</u>
Cell B ppb: <u>-2.1</u>	Cell B ppb: <u>n/a</u>
Cell A int: <u>79343</u>	Cell A int: <u>n/a</u>
Expected Value: <u>303.0</u>	Expected Value: <u>n/a</u>

Comments:

Shutdown calibration completed to perform annual maintenance. No ZERO adjustment made. No High Point adjustment made.

Maxxam Thermo 49i Ozone Analyzer Calibration

A Bureau Veritas Group Company

Date:	March 4, 2017	Barometric Pressure:	0.921 atm
Company/Airshed:	LICA	Station Temperature °C:	22
Location/Station Name:	Bonnyville	Weather Conditions:	Mainly cloudy with snow
Start/End Time 24 hr. (mst):	12:45 / 16:43	Calibration Purpose:	post repair
Ozone Calibration Method:	Varying UV Lamp Power	Performed By/Reviewer:	Alex Yakupov Trina Whitsitt
G.P.T. Date:	n/a-done by Varying UV Lamp Power	Cal Gas Expiry Date:	n/a

Analyzer:	
ID# or Serial Number:	1002240372
Last Calibration Date:	n/a
Previous Cal High Point C.F.:	n/a
Ozone Range ppb:	500
As Found C.F.:	n/a
New C.F.:	1.000

Calibrator:	
Flow Meter ID's:	n/a
Make & Model:	SABIO 2010 D
Serial #:	11900613
Cal Gas Cylinder I.D. #:	n/a

Point	AMD Required Range of Ozone Calibration Points
High	300-400 ppb
Mid	150-200 ppb
Low	50-75 ppb


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rate (cc/min)		Calculated Concentration:	Corrected Calculated Concentration:	Indicated Concentration:	Correction Factors:
	Total Flow @ Point Start	Total Flow @ Point Finish	(ppb)	(ppb)	(ppb)	
adjusted zero	5000	5000	0.0	n/a	0.0	n/a
adjusted high	5000	5000	380.0	380.0	380.0	1.000
mid	5000	5000	180.0	180.0	181.0	0.994
low	5000	5000	60.0	60.0	61.0	0.984
calibrator zero	5000	5000	0.0	0.0	0.0	n/a
Average C.F. =						0.993

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	1.000		.95-1.05
b (Intercept as % of full scale) =	0.00%		± 3% F.S.
% change in C.F. from last cal =	n/a		± 10%

Thermo 49i Ozone Analyzer Calibration




As found:		As left:	
O3 Bkg:	n/a	O3 Bkg:	-0.2
O3 Coef:	n/a	O3 Coef:	1.000
Photo Lamp:	n/a	Photo Lamp:	14.2
O3 Lamp:	n/a	O3 Lamp:	5.8
Bench:	n/a	Bench:	27.7
Bench Lamp:	n/a	Bench Lamp:	54.0
O3 Lamp:	n/a	O3 Lamp:	68.1
Pressure:	n/a	Pressure:	700.0
Cell A lpm:	n/a	Cell A lpm:	0.754
Cell B lpm:	n/a	Cell B lpm:	0.766
O3 ppb:	n/a	O3 ppb:	0.0
Cell A ppb:	n/a	Cell A ppb:	1.9
Cell B ppb:	n/a	Cell B ppb:	-1.9
Cell A int:	n/a	Cell A int:	82069
Expected Value:	n/a	Expected Value:	282.0

Comments:
 The analyzer sample inlet filter was changed. The analyzer cooling fan filter(s) were cleaned.

Post-repair calibration completed after annual maintenance. The Optical Cell A and B was cleaned and the sample pump was rebuilt.

O3[ppb] Station: LICA Bonnyville Daily: 2017/03/04 Type: AVG 1 Min. [1 Min.]



— O3[ppb]

PARTICULATE MATTER



R & P 1405F TEOM PM 2.5 Analyzer Calibration

Date: March 3, 2017
 Company: LICA
 Station Name/Location: Bonnyville - AER
 Previous Audit Date: February 21, 2017
 Parameter: PM 2.5

Performed By/Reviewer: Alex Yakupov | Trina Whitsitt
 Start Time (mst): 13:55
 End Time (mst): 16:21
 Calibration Purpose: Bi-monthly #1
 Weather Conditions: Mainly sunny

1400A Information and Status:

ID# or Serial Number: 1405A207691003 As Found Filter Loading %: 27.36
 Ko Factor: 15635 As Left Filter Loading %: 16.95
 Ambient Temperature °C: 3.74 As Found Noise: 0.004
 Ambient Pressure atm: 0.919 As Left Noise: 0.000
 Main Flow Reading lpm: 3.00 Pump Vacuum: 0.30
 Aux Flow Reading lpm: 13.67 Warnings: None

Reference Standards:

	Flow:	Pressure:	Temperature:
Make:	<u>Dwyer</u>	<u>Fisher</u>	<u>FLUKE</u>
Model:	<u>475 Mark III</u>	<u>FB1291</u>	<u>1551A Ex STIK</u>
Serial Number:	<u>#3</u>	<u>#05544</u>	<u>2329070</u>
Calibration Date:	<u>January 1, 2017</u>	<u>December 5, 2016</u>	<u>November 15, 2016</u>

As found leak check:

		Base	Zero	Reference	Zero
PM 2.5 Flow	actual	0.00	0.07	0.00	0.07
	limit	0.15	0.15	0.15	0.15
Bypass Flow	actual	0.00	-0.31	0.00	-0.31
	limit	0.60	0.60	0.60	0.60

As left leak check (same as above if as found passes):

		Base	Zero	Reference	Zero
PM 2.5 Flow	actual	0.00	0.07	0.00	0.07
	limit	0.15	0.15	0.15	0.15
Bypass Flow	actual	0.00	-0.31	0.00	-0.31
	limit	0.60	0.60	0.60	0.60

As found temperature and pressure:

tolerance +/- 2.0°C	tolerance +/- 0.01 atm
1405F temperature °C: <u>3.7</u>	1405F pressure atm: <u>0.919</u>
reference temperature °C: <u>3.5</u>	reference pressure: <u>0.919</u>
difference °C: <u>-0.3</u>	difference: <u>0.000</u>

As left temperature and pressure (same as above if as found adequate):

tolerance +/- 2.0°C	tolerance +/- 0.01 atm
1405F temperature °C: <u>3.5</u>	1405F pressure atm: <u>0.919</u>
reference temperature °C: <u>3.5</u>	reference pressure: <u>0.919</u>
difference °C: <u>0.0</u>	difference: <u>0.000</u>

As found flows:

main flow tolerance 3.00 lpm +/- 0.20 lpm	total/aux flow tolerance 16.67/13.67 lpm +/- 1.00 lpm +/- 7%
1405F main flow lpm: <u>3.00</u>	1400A total/aux flow lpm: <u>16.67</u>
reference main flow lpm: <u>2.99</u>	reference total/aux flow lpm: <u>16.23</u>
difference lpm: <u>-0.01</u>	difference lpm: <u>-0.44</u>

As left flows (same as above if as found adequate):

main flow tolerance 3.00 lpm +/- 0.20 lpm	total/aux flow tolerance 16.67/13.67 lpm +/- 1.00 lpm +/- 7%
1405F main flow lpm: <u>3.00</u>	1400A total/aux flow lpm: <u>16.67</u>
reference main flow lpm: <u>2.99</u>	reference total/aux flow lpm: <u>16.23</u>
difference lpm: <u>-0.01</u>	difference lpm: <u>-0.44</u>

K_o Audit:

Last K_o audit date: February 6, 2017
 1405F K_o factor: 15635
 Measured K_o factor: 15808.3000
 % difference: 1.11

Comments:

The TEOM sample filter was changed. The TEOM intake head and associated sharp cut components were cleaned.

The 47 mm FDMS filter was changed.

The total and by-pass flows were measured and the main flow was calculated.



R & P 1405F TEOM PM 2.5 Analyzer Calibration

Date: March 28, 2017
Company: LICA
Station Name/Location: Bonnyville - AER
Previous Audit Date: March 3, 2017
Parameter: PM 2.5

Performed By/Reviewer: Alex Yakupov | Trina Whitsitt
Start Time (mst): 13:39
End Time (mst): 14:29
Calibration Purpose: Bi-monthly #2
Weather Conditions: Fog

1400A Information and Status:

ID# or Serial Number: <u>1405A207691003</u>	As Found Filter Loading %: <u>25.93</u>
Ko Factor: <u>15635</u>	As Left Filter Loading %: <u>26.11</u>
Ambient Temperature °C: <u>5.08</u>	As Found Noise: <u>0.004</u>
Ambient Pressure atm: <u>0.930</u>	As Left Noise: <u>0.000</u>
Main Flow Reading lpm: <u>3.00</u>	Pump Vacuum: <u>0.31</u>
Aux Flow Reading lpm: <u>13.67</u>	Warnings: <u>None</u>

Reference Standards:

Make: <u>Dwyer</u>	Flow: <u>Fisher</u>	Pressure: <u>FLUKE</u>	Temperature: <u>FLUKE</u>
Model: <u>475 Mark III</u>	Pressure: <u>FB1291</u>	Temperature: <u>1551A Ex STIK</u>	
Serial Number: <u>#3</u>	Pressure: <u>#05544</u>	Temperature: <u>4295</u>	
Calibration Date: <u>January 1, 2017</u>	Pressure: <u>December 5, 2016</u>	Temperature: <u>November 15, 2016</u>	

As found leak check:

		Base	Zero	Reference	Zero
PM 2.5 Flow	actual	0.00	0.07	0.00	0.07
	limit	0.15	0.15	0.15	0.15
Bypass Flow	actual	0.00	-0.31	0.00	-0.31
	limit	0.60	0.60	0.60	0.60

As left leak check (same as above if as found passes):

		Base	Zero	Reference	Zero
PM 2.5 Flow	actual	0.00	0.07	0.00	0.07
	limit	0.15	0.15	0.15	0.15
Bypass Flow	actual	0.00	-0.31	0.00	-0.31
	limit	0.60	0.60	0.60	0.60

As found temperature and pressure:

tolerance +/- 2.0°C 1405F temperature °C: <u>5.1</u> reference temperature °C: <u>6.7</u> difference °C: <u>1.7</u>	tolerance +/- 0.01 atm 1405F pressure atm: <u>0.930</u> reference pressure: <u>0.932</u> difference : <u>-0.002</u>
---	---

As left temperature and pressure (same as above if as found adequate):

tolerance +/- 2.0°C 1405F temperature °C: <u>6.7</u> reference temperature °C: <u>6.7</u> difference °C: <u>0.0</u>	tolerance +/- 0.01 atm 1405F pressure atm: <u>0.932</u> reference pressure: <u>0.932</u> difference : <u>0.000</u>
---	--

As found flows:

main flow tolerance 3.00 lpm +/- 0.20 lpm 1405F main flow lpm: <u>3.00</u> reference main flow lpm: <u>2.96</u> difference lpm: <u>-0.04</u>	total/aux flow tolerance 16.67/13.67 lpm +/- 1.00 lpm/+/- 7% 1400A total/aux flow lpm: <u>16.67</u> reference total/aux flow lpm: <u>16.09</u> difference lpm: <u>-0.58</u>
--	---

As left flows (same as above if as found adequate):

main flow tolerance 3.00 lpm +/- 0.20 lpm 1405F main flow lpm: <u>3.00</u> reference main flow lpm: <u>2.96</u> difference lpm: <u>-0.04</u>	total/aux flow tolerance 16.67/13.67 lpm +/- 1.00 lpm/+/- 7% 1400A total/aux flow lpm: <u>16.67</u> reference total/aux flow lpm: <u>16.09</u> difference lpm: <u>-0.58</u>
--	---

K_o Audit:

Last K_o audit date: February 6, 2017
 1405F K_o factor: 15635
 Measured K_o factor: 15808.3000
 % difference: 1.11

Comments:

The TEOM intake head and associated sharp cut components were cleaned.

The 47 mm FDMS filter was changed.

The total and by-pass flows were measured and the main flow was calculated.

WIND SYSTEM



Meteorological Sensor Audit

Station Information

Company:	<u>LICA</u>	Performed By:	<u>Limin Li</u>
Location:	<u>Bonnyville (in Calgary shop)</u>	Reason:	<u>Annual maintenance</u>
Audit Date:	<u>26-Jan-16</u>	Start Time (mst):	<u>11:00</u>
Previous Audit Date:	<u>NA</u>	End Time (mst):	<u>15:00</u>

Wind Speed

Sensor make:	<u>R. M. Young</u>	Sensor height:	<u>n/a</u>
Sensor model:	<u>5103VK</u>	Serial Number:	<u>56589</u>
Calibrator:	<u>Young 18802</u>	Variable speed motor:	<u>CA 03309</u>
Voltage range:	<u>0-1</u>	Output signal range:	<u>200KPH</u>

Wind Speed Audit Data

RPM	Wind Speed Actual	Indicated WS - CW	Indicated WS-CCW	Correction Factor
0	0.0	0.032	0.032	-
1000	17.6	17.66	17.64	1.00
2000	35.28	35.3	35.29	1.00
3000	52.92	52.99	52.99	1.00
4000	70.56	70.66	70.65	1.00
5000	88.2	88.35	88.33	1.00
6000	105.84	106	106	1.00
7000	123.48	123.7	123.7	1.00
8000	141.12	141.4	141.3	1.00
9000	158.76	159.1	159.1	1.00
10000	176.4	176.7	176.7	1.00
Average Correction Factor:				1.00

Wind Direction

Sensor make:	<u>R. M. Young</u>	Sensor height:	<u>n/a</u>
Sensor model:	<u>5103VK</u>	Serial Number:	<u>56589</u>
Calibrator:	<u>Young 18802</u>	Variable speed motor:	<u>CA 03309</u>
Voltage range:	<u>0-1</u>	Output signal range:	<u>0-360DEG</u>

Wind Direction Audit Data

Wind Direction	Indicated	Correction Factor
0	0.5	NA
45	44.9	1.00
90	92.0	0.98
135	136.5	0.99
180	180.6	1.00
225	224.4	1.00
270	270.3	1.00
315	312.2	1.01
359	355.0	1.01
Average Correction Factor:		1.00

Remarks: Annual maintenance. Changed 05163PG, 05124VG bearings. 05131D, 05133B & 05135D

Audit Performed by: Limin Li



Meteorological Sensor Audit/Calibration

Location Information

Company:	LICA	Performed By:	Alex Yakupov
Audit Location:	Bonnyville - AER	Reviewed By:	Trina Whitsitt
Audit Date:	March 3, 2017	Start /EndTime (mst):	10:11 / 13:14

Wind Sensor Information

Sensor ID Data:		Sensor Outputs:	
Sensor Make:	R.M. Young	Velocity Voltage Output Range:	0-1
Sensor Model:	5103 VK	Velocity Unit Output Range:	0-200
Serial #:	56589	Direction Voltage Output Range:	0-1
Previous Cal/Audit Date:	January 26, 2016	Direction Unit Output Range:	0-360

Wind Calibrator Information

Calibrator Make/ Model:	RM Young / Model 18802	Serial #:	CA 03309
Maxxam Unit ID #:	13-3357	Certification Date:	October 6, 2016

Wind Speed Audit Data ****+/- 2% of the average correction factor is the limit****

RPM	Wind Speed Generated kph	Clockwise Wind Speed kph	Counter Clockwise Wind Speed kph	Correction Factor
0	0	0.0	0.0	-
1000	17.6	17.7	17.6	0.998
2000	35.3	35.3	35.3	0.999
3000	52.9	53.0	53.0	0.999
4000	70.6	70.6	70.7	0.999
5000	88.2	88.3	88.4	0.998
6000	105.8	106.0	106.0	0.998
7000	123.5	123.6	123.7	0.999
8000	141.1	141.3	141.4	0.998
9000	158.8	159.0	159.1	0.998
10000	176.4	176.7	176.7	0.998
The audit meets AMD requirements.			Average Correction Factor=	0.998

Wind Direction Audit Data ****+/- 5° of the absolute average degrees difference for all points is the limit****

Generated Wind Direction 0-360 (Up)	Generated Wind Direction 360-0 (Down)	Indicated Wind Direction 0-360 (Up)	Indicated Wind Direction 360-0 (Down)	Degrees Difference 0-360 (Up)	Degrees Difference 360-0 (Down)	Average Absolute Degrees Difference
0	355	0	354	0.2	1.3	0.8
30	330	30	329	-0.4	0.9	0.6
60	300	60	300	-0.4	0.3	0.4
90	270	91	271	-0.8	-0.6	0.7
120	240	120	240	0.3	0.4	0.4
150	210	149	210	0.7	-0.3	0.5
180	180	180	181	0.5	-0.9	0.7
210	150	209	150	1.1	-0.2	0.6
240	120	239	121	1.4	-0.5	1.0
270	90	269	90	1.5	-0.2	0.9
300	60	296	60	3.6	0.0	1.8
330	30	325	30	4.7	0.0	2.4
355	0	352	0	3.2	0.2	1.7
The audit meets AMD requirements.				Average Absolute Degrees Difference=		1.0

Comments:

CALIBRATORS

Company Maxxam Operator: Christopher Wesson

Calibrator:		Flow Measurement Device:	
Make/Model	<u>Sabio 2010</u>	Make/Model	<u>N/A</u>
Serial Number	<u>17200415</u>	Serial Number	<u>N/A</u>
Last Verification Date	<u>May 2015</u>	Temperature (°C)	<u>N/A</u>
NO Cylinder S/N	<u>LL42475</u>	Barometric Pressure	<u>N/A</u>
NO/NOx Concentration	<u>48.5/48.5</u>		

Dilution Flow (sccm)			
Pt. #1	<u>5000</u>	Pt. #2	<u>5000</u>
Pt. #3	<u>5000</u>		
Gas Flow (sccm)			
Pt. #1	<u>80</u>	Pt. #2	<u>40</u>
Pt. #3	<u>20</u>		

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
5029	0.0	0.000	0.000	0.000	0.000	0.000	Limit ± 10%	
5030	80.6	0.777	0.777	0.805	-0.005	0.800	4%	3%
5025	39.4	0.380	0.380	0.394	-0.002	0.392	4%	3%
5028	19.8	0.191	0.191	0.198	-0.001	0.197	4%	3%
Absolute Average Percent Difference							3.65%	3.09%

LINEAR REGRESSION ANALYSIS				$y=mx+b$ (where x=calculated concentration, y=indicated concentration)			
NO		LIMITS		NOx			
Correlation=	1.0000	≥ 0.990		Correlation=	1.0000		
m (Slope)=	1.0360	0.90-1.10		m (Slope)=	1.0295		
b (Intercept % of FS)=	0.0110	± 3% F.S.		b (Intercept % of FS)=	0.0293		

Flow	O ₂ Conc	NO Decrease	NO	NO ₂	NOx	% Diff. Vs Audit gas	
5030	Lamp C.	0.000	0.804	-0.004	0.800	NO ₂	% Diff. Limit
5030	1.388	0.495	0.309	0.491	0.800	0%	± 10%
5030	0.745	0.241	0.563	0.239	0.802	1%	± 10%
5030	0.367	0.091	0.713	0.089	0.801	2%	± 10%
Absolute Average Percent Difference						1%	± 10%

LINEAR REGRESSION ANALYSIS				$y=mx+b$ (where x=calculated concentration, y=indicated concentration)			
NO ₂		LIMITS					
Correlation=	1.0000	≥ 0.995					
m (Slope)=	0.9988	0.90-1.10					
b (Intercept % of FS)=	-0.2760	± 3% F.S.					

AENV Standards Audit Calibrator		NO _x Analyzer	
Make/Model	<u>Teco 146i</u>	Make/Model	<u>Teco 42i</u>
Serial/AMU Number	<u>AMU 1809</u>	Serial/AMU Number	<u>AMU 1868</u>
		Last Calibration Date	<u>May 18, 2016</u>
		Full Scale (ppm)	<u>1.0</u>

 COMMENTS: Contains 50.3 ppm SO₂. Flows not measured as per Chapter 7, Section 5 of AMD.

 Auditor: AI Clark
 Operator Signature: *AI Clark*

 Date: May 18, 2016
 Location: McIntyre Center Edmonton

Company Maxxam/SIA **Operator:** Chris

Calibrator:		Flow Measurement Device:	
Make/Model	<u>API 700</u>	Make/Model	<u>Definer 530</u>
Serial Number	<u>627</u>	Serial Number	<u>H-148944, L-152019</u>
Last Verification Date	<u>February 3, 2016</u>	Temperature (°C)	<u>23.5</u>
NO Cylinder S/N	<u>EY0000597</u>	Barometric Pressure	<u>707.1 mmHg</u>
NO [PPM]	<u>49.0</u>	NOx [PPM]	<u>49.0</u>
Expiry Date	<u>December 8, 2019</u>		

Dilution Flow (sccm)		
Pt. #1	<u>4892</u>	Pt. #3 <u>4951</u>
Pt. #2	<u>4975</u>	
Gas Flow (sccm)		
Pt. #1	<u>79.7</u>	Pt. #3 <u>19.4</u>
Pt. #2	<u>38.8</u>	

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
	0.0	0.0000	0.0000	0.0000	-0.0004	-0.0004	Limit ± 10%	
4972	79.7	0.7855	0.7855	0.7883	0.0004	0.7887	0.4%	0.5%
4936	38.8	0.3822	0.3822	0.3816	0.0005	0.3822	-0.2%	0.1%
4970	19.4	0.1913	0.1913	0.1902	0.0006	0.1913	-0.6%	0.2%
Absolute Average Percent Difference							0.1%	0.3%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO	LIMITS	NOx
Correlation= 1.0000	≥ 0.990	Correlation= 1.0000
m (Slope)= 1.0041	0.90-1.10	m (Slope)= 1.0046
b (Intercept % of FS)= -0.1118	± 3% F.S.	b (Intercept % of FS)= -0.0871

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4972	0	0.0000	0.7867	0.0014	0.7881	NO ₂	% Diff, Limit
4972	500	0.5127	0.2740	0.5104	0.7849	-0.7%	± 10%
4972	275	0.2863	0.5004	0.2860	0.7865	-0.6%	± 10%
4972	90	0.0940	0.6927	0.0954	0.7880	0.0%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO₂	LIMITS
Correlation= 1.0000	≥ 0.995
m (Slope)= 0.9924	0.90-1.10
b (Intercept % of FS)= 0.1755	± 3% F.S.

AENV Standards	NO_x Analyzer
Audit Calibrator	Make/Model <u>Thermo 42i</u>
Make/Model <u>Thermo 146i</u>	Serial/AMU Number <u>AMU 1868</u>
Serial/AMU Number <u>AMU1809</u>	Last Calibration Date <u>January 25, 2017</u>
SRM Gas Cylinder No. <u>CAL018140</u>	Full Scale (ppm) <u>1.0</u>
Cylinder Conc. (ppm) <u>48.79</u>	Cylinder Gas Expiry Date <u>March 25, 2019</u>

COMMENTS: _____

Auditor: Shea Beaton Date: January 27, 2017
Operator Signature: _____ Location: McIntyre Center Edmonton

Company Maxxam Operator: Mike

Calibrator:			Flow Measurement Device:		
Make/Model	<u>EnviroNics 2000</u>		Make/Model	<u>Bios Defender 530</u>	
Serial Number	<u>1991</u>		Serial Number	<u>HI148944 Lo 152019</u>	
Last Verification Date	<u>March 31, 2016</u>		Temperature (°C)	<u>24.5</u>	
NO Cylinder S/N	<u>EY0000597</u>		Barometric Pressure	<u>699</u>	
NO [PPM]	<u>49.0</u>	NOx [PPM]	<u>49.0</u>		
Expiry Date	<u>December 8, 2019</u>				

Dilution Flow (sccm)					
Pt. #1	<u>4902</u>	Pt. #2	<u>4935</u>	Pt. #3	<u>4957</u>
Gas Flow (sccm)					
Pt. #1	<u>79.3</u>	Pt. #2	<u>38.7</u>	Pt. #3	<u>19.4</u>

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
4976	0.0	0.0000	0.0000	0.0001	0.0000	0.0001	Limit ± 10%	
4981	79.3	0.7801	0.7801	0.7898	0.0000	0.7898	1%	1%
4972	38.7	0.3814	0.3814	0.3841	0.0002	0.3843	1%	1%
4976	19.4	0.1910	0.1910	0.1913	0.0003	0.1916	0%	0%
Absolute Average Percent Difference							1%	1%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO	LIMITS	NOx
Correlation= 1.0000	≥ 0.990	Correlation= 1.0000
m (Slope)= 1.0130	0.90-1.10	m (Slope)= 1.0129
b (Intercept % of FS)= -0.1190	± 3% F.S.	b (Intercept % of FS)= -0.1029

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4981	0.000	0.0000	0.7925	-0.0001	0.7924	NO ₂	% Diff. Limit
4981	0.400	0.5347	0.2578	0.5279	0.7857	-1%	± 10%
4981	0.200	0.2490	0.5435	0.2478	0.7913	0%	± 10%
4981	0.090	0.1090	0.6835	0.1095	0.7927	1%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO₂	LIMITS
Correlation= 1.0000	≥ 0.995
m (Slope)= 0.9864	0.90-1.10
b (Intercept % of FS)= 0.1136	± 3% F.S.

AENV Standards	NO_x Analyzer
Audit Calibrator	Make/Model <u>Thermo 42i</u>
Make/Model <u>Thermo 146i</u>	Serial/AMU Number <u>1868</u>
Serial/AMU Number <u>1809</u>	Last Calibration Date <u>March 15, 2017</u>
SRM Gas Cylinder No. <u>CAL018140</u>	Full Scale (ppm) <u>1.0</u>
Cylinder Conc. (ppm) <u>48.79</u>	Cylinder Gas Expiry Date <u>March 28, 2019</u>

COMMENTS: Gas has ~50ppm SO2

Auditor: Shea Beaton Date: March 16, 2017

Operator Signature: [Signature] Location: McIntyre Center Edmonton

Company <u>Maxxam</u>		Operator: <u>Mike</u>	
Calibrator:		Flow Measurement Device:	
Make/Model	<u>Sabio 2010D</u>	Make/Model	<u>Bios Defender 530</u>
Serial Number	<u>11900613</u>	Serial Number	<u>HI148944 Lo 152019</u>
Last Verification Date	<u>March 31, 2016</u>	Temperature (°C)	<u>23.9</u>
NO Cylinder S/N	<u>EY0000769</u>	Barometric Pressure	<u>698mmHg</u>
NO [PPM]	<u>51.1</u> NOx [PPM]		<u>51.2</u>
Expiry Date	<u>December 8, 2019</u>		

Dilution Flow (sccm)		
Pt. #1 <u>4879</u>	Pt. #2 <u>4932</u>	Pt. #3 <u>4950</u>
Gas Flow (sccm)		
Pt. #1 <u>74.5</u>	Pt. #2 <u>36.4</u>	Pt. #3 <u>18.2</u>

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
4965	0.0	0.0000	0.0000	0.0001	0.0000	0.0001	Limit ± 10%	
4954	74.5	0.7685	0.7700	0.7915	0.0008	0.7923	3%	3%
4968	36.4	0.3744	0.3751	0.3832	0.0006	0.3838	2%	2%
4968	18.2	0.1872	0.1876	0.1916	0.0002	0.1918	2%	2%
Absolute Average Percent Difference							3%	2%


LINEAR REGRESSION ANALYSIS			<i>y=mx+b (where x=calculated concentration, y=indicated concentration)</i>		
NO		LIMITS		NOx	
Correlation=	1.0000	≥ 0.990		Correlation=	1.0000
m (Slope)=	1.0301	0.90-1.10		m (Slope)=	1.0291
b (Intercept % of FS)=	-0.0919	± 3% F.S.		b (Intercept % of FS)=	-0.0881

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4954	0.000	0.0000	0.7949	0.0005	0.7954	NO ₂	% Diff. Limit
4954	0.510	0.5104	0.2845	0.5072	0.7917	-1%	± 10%
4954	0.250	0.2516	0.5433	0.2514	0.7944	0%	± 10%
4954	0.100	0.1085	0.6864	0.1087	0.7951	0%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS			<i>y=mx+b (where x=calculated concentration, y=indicated concentration)</i>		
NO₂		LIMITS			
Correlation=	1.0000	≥ 0.995			
m (Slope)=	0.9926	0.90-1.10			
b (Intercept % of FS)=	0.0925	± 3% F.S.			

AENV Standards		NO_x Analyzer	
Audit Calibrator			
Make/Model	<u>Thermo 146i</u>	Make/Model	<u>Thermo 42i</u>
Serial/AMU Number	<u>1809</u>	Serial/AMU Number	<u>1868</u>
SRM Gas Cylinder No.	<u>CAL018140</u>	Last Calibration Date	<u>March 15, 2017</u>
Cylinder Conc. (ppm)	<u>48.79</u>	Full Scale (ppm)	<u>1.0</u>
		Cylinder Gas Expiry Date	<u>March 28, 2019</u>

COMMENTS: Gas has ~50ppm SO2

Auditor: Shea Beaton
Operator Signature: 

Date: March 16, 2017
Location: McIntyre Center Edmonton

CALIBRATION GASES



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2016-335CGA

Company: Maxxam **Operator's Name:** Russell Kirchner

Cylinder #: LL104222 Concentration PPM: 50.6 Tolerance(%) 1 Certified By: Praxair

Expiry Date: July 2019

Reference Calibrator and Gas:	Flow Measurement Device:
Make/Model: <u>R&R MFC 201</u>	Make/Model: <u>Bios DC2</u>
Serial Number: <u>AMU 1690</u>	Serial Number: <u>AMY 1659</u>
Last Verification Date: <u>October 19, 2016</u>	Temp. °C: <u>24.5 C</u>
Gas Type: <u>SO2</u> Conc. <u>98.07</u>	B.P. <u>706 mmhg</u>
Cylinder Number: <u>CA:016625</u>	
Expiry Date: <u>January 2019</u>	

Reference Analyzer:

Make/Model: Teco 43C Serial/AMU Number: 1623

Instrument Settings: Zero: 9.2 Span: 1.024 Range: 1.0

Last Calibration: Date: Oct 19/16 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.000	0.0000	0.0000	0.000
4935	82.0	0.830	0.01662	60.183	50.0
4968	40.8	0.412	0.00821	121.765	50.2
4955	20.2	0.203	0.00408	245.297	49.8
Average Cylinder Concentration:					50.0

Previous Stated Concentration PPM: 50.6

Percent variance from Stated: 1

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____

< =5% Outside Manufacturer Tolerance. Use manufacturers concentration _____

> 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark Date: October 19, 2016

Operator Signature: *Al Clark* Location: McIntyre Center Edmonton



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2015-338CGA

Company: Maxxam **Operator's Name:** Limin Li
Cylinder #: BLM002508 **Concentration PPM:** 10.2 **Tolerance(%)** 2 **Certified By:** Air Liquide

Reference Calibrator and Gas:

Make/Model: R&R MFC 201
 Serial Number: AMU1690
 Last Verification Date: March 31, 2015
 Gas Type: H2S Conc. 20.43
 Cylinder Number: CAL015106

Flow Measurement Device:

Make/Model: Bios DC2
 Serial Number: AMU 1659
 Temp. °C: 23.0 C
 B.P. 689 mmhg

Reference Analyzer:

Make/Model: Teco 450i Serial/AMU Number: 1980
 Instrument Settings: Zero: 14.5 Span: 1.035 Range: 0.1
 Last Calibration: Date: Mar 31/15 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.0000	0.0000	132.984	9.6
5080	38.2	0.0725	0.00752	132.984	9.6
5078	17.9	0.0340	0.00353	283.687	9.6
5066	9.1	0.0170	0.00180	556.703	9.5
Average Cylinder Concentration:					9.6

Previous Stated Concentration PPM: 10.2

Percent variance from Stated: 6.0

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark
 Operator Signature: *Al Clark*

Date: March 31, 2015
 Location: McIntyre Center Edmonton



Calibration Gas Audit

NO Cylinder Gas

File No. 2015-343CGA

Company: Maxxam **Operators name:** Limin Li
Cylinder #: BLM002756T **Conc (PPM)** 50.7/50.7 **Tolerance (%)** 2 **Certified By:** Air Liquide

Reference Calibrator and Gas:

Make/Model Teco 146i
Serial Number AMU 1809
Last Verification Date March 31, 2015
Gas Type NO **Conc.** 48.79
Cylinder Number CAL018024

Flow Measurement Device:

Make/Model Bios DC2
Serial Number AMU 1659
Temp. °C 22.5 C
B.P. 690 mmhg

Reference Analyzer:

Make/Model Teco 42i **Serial/AMU Number:** 1868
Instrument Settings **Zero:** 4.2 **Span:** 1.008 **Range:** 1.0
Last Calibration: **Date:** Mar 31/15 **C.F.** 1.000 **Done By:** Al Clark

Calibrator Flows (scm)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	NO	NOX			NO	NOX
5000	0.0	0.000	0.000	0.01660	60.242	50.7	49.5
4976	82.6	0.842	0.822	0.01660	60.242	50.7	49.5
4993	41.0	0.420	0.410	0.00821	121.780	51.1	49.9
4977	20.2	0.208	0.205	0.00406	246.386	51.2	50.5
Average Cylinder Concentration:						51.0	50.0

<u>NO</u>	<u>NOx</u>
Previous Stated Concentration PPM: <u>50.7</u>	<u>50.7</u>
Percent variance from Stated: <u>0.7</u>	<u>1.4</u>

Cylinder gas tolerances based on NO only

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 < =5% Outside Manufacturer Tolerance. Use manufacturers concentration Contains 49.9 ppm SO2 in cylinder
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark **Date:** March 31, 2015
Operator Signature: *Al Clark* **Location:** McIntyre Center Edmonton



Calibration Gas Audit

CH4 / C3H8 Cylinder Gas

File No. 2015-092CGA

Company: Maxxam Operators name: Chris Wesson
Cylinder #: LL165372 Conc CH4 (PPM) 606/212 Tolerance (%) 0.5 Certified By: Praxair

Reference Calibrator and Gas:

Make/Model R&R MFC 201
Serial Number AMU 1698
Last Verification Date January 18, 2016
Gas Type CH4 Conc. 999.2
Cylinder Number D751932
Gas Type C3H8 Conc. 246.5
Cylinder Number XF0037998

Flow Measurement Device:

Make/Model Bios DC-2
Serial Number Blos D
Temp. °C 24.5
B.P. 688mmHg

Reference Analyzer:

Make/Model Thermo 55C Serial/AMU Number: 1643
Instrument Settings Zero: NA Span: NA Range: 20.0
Last Calibration: Date: 18-Jan-16 C.F. 1.000 Done By: SB

Calibrator Flows (scem)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	CH4	C3H8			CH4	C3H8
2568	0.00	0.00	0.00	0.02140	46.722	607	214
2630	56.29	12.99	12.62	0.02140	46.722	607	214
2588	19.73	4.62	4.50	0.00762	131.171	606	215
2580	9.69	2.29	2.24	0.00376	266.254	610	217
Average Cylinder Concentration:						608	215

<u>CH4</u>	<u>C3H8</u>
Previous Stated Concentration PPM: <u>606</u>	<u>212</u>
Percent variance from Stated: <u>0.3</u>	<u>1.6</u>

Cylinder gas tolerances based on CH4 only

Meets Manufacturer Tolerance. Use manufacturers stated concentration COMMENTS: _____
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration C3H8 manufacturers tolerance 1.1%
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Shea Beaton Date: January 19, 2016
Operator Signature: _____ Location: McIntyre Center Edmonton

APPENDIX IV
ANALYTICAL RESULTS

VOCS SAMPLES

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
LICA/VOC/Bonnyville/Mar 2, 2017	14995	Ambient Air	02-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030077	REPORT CREATED:	13-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030077-001	1,1,1-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	1,1,2,2-Tetrachloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	1,1,2-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	1,1-Dichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	1,1-Dichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030077-001	1,2,3-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030077-001	1,2,4-Trichlorobenzene	K, T, U	< 0.8	ppbv	0.8	AC-058	15-Mar-17
17030077-001	1,2,4-Trimethylbenzene		0.07	ppbv	0.05	AC-058	15-Mar-17
17030077-001	1,2-Dibromoethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	1,2-Dichlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030077-001	1,2-Dichloroethane	I	0.02	ppbv	0.01	AC-058	15-Mar-17
17030077-001	1,2-Dichloropropane	I	0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-001	1,3,5-Trimethylbenzene		0.04	ppbv	0.02	AC-058	15-Mar-17
17030077-001	1,3-Butadiene	I	0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	1,3-Dichlorobenzene	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030077-001	1,4-Dichlorobenzene	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030077-001	1,4-Dioxane	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030077-001	1-Butene		0.11	ppbv	0.02	AC-058	15-Mar-17
17030077-001	1-Hexene	I	0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	1-Pentene		0.02	ppbv	0.01	AC-058	15-Mar-17
17030077-001	2,2,4-Trimethylpentane		0.15	ppbv	0.01	AC-058	15-Mar-17
17030077-001	2,2-Dimethylbutane		0.04	ppbv	0.01	AC-058	15-Mar-17
17030077-001	2,3,4-Trimethylpentane		0.05	ppbv	0.01	AC-058	15-Mar-17
17030077-001	2,3-Dimethylbutane		0.11	ppbv	0.02	AC-058	15-Mar-17
17030077-001	2,3-Dimethylpentane		0.15	ppbv	0.02	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
LICA/VOC/Bonnyville/Mar 2, 2017	14995	Ambient Air	02-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030077	REPORT CREATED:	13-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030077-001	2,4-Dimethylpentane		0.07	ppbv	0.01	AC-058	15-Mar-17
17030077-001	2-Methylheptane		0.06	ppbv	0.01	AC-058	15-Mar-17
17030077-001	2-Methylhexane		0.09	ppbv	0.01	AC-058	15-Mar-17
17030077-001	2-Methylpentane		0.32	ppbv	0.01	AC-058	15-Mar-17
17030077-001	3-Methylheptane		0.04	ppbv	0.02	AC-058	15-Mar-17
17030077-001	3-Methylhexane		0.12	ppbv	0.02	AC-058	15-Mar-17
17030077-001	3-Methylpentane		0.17	ppbv	0.01	AC-058	15-Mar-17
17030077-001	Acetone		1.7	ppbv	0.4	AC-058	15-Mar-17
17030077-001	Acrolein	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030077-001	Benzene		0.27	ppbv	0.01	AC-058	15-Mar-17
17030077-001	Benzyl chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030077-001	Bromodichloromethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	Bromoform	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	Bromomethane	I	0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-001	Carbon disulfide	I	0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-001	Carbon tetrachloride	I	0.13	ppbv	0.01	AC-058	15-Mar-17
17030077-001	Chlorobenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	Chloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	Chloroform	I	0.03	ppbv	0.02	AC-058	15-Mar-17
17030077-001	Chloromethane		0.64	ppbv	0.02	AC-058	15-Mar-17
17030077-001	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-001	cis-1,3-Dichloropropene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030077-001	cis-2-Butene		0.03	ppbv	0.02	AC-058	15-Mar-17
17030077-001	cis-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	Cyclohexane		0.23	ppbv	0.02	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
LICA/VOC/Bonnyville/Mar 2, 2017	14995	Ambient Air	02-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030077	REPORT CREATED:	13-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030077-001	Cyclopentane		0.11	ppbv	0.01	AC-058	15-Mar-17
17030077-001	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-001	Ethanol		2.6	ppbv	0.3	AC-058	15-Mar-17
17030077-001	Ethyl acetate	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030077-001	Ethylbenzene		0.07	ppbv	0.01	AC-058	15-Mar-17
17030077-001	Freon-11	I	0.29	ppbv	0.02	AC-058	15-Mar-17
17030077-001	Freon-113	I	0.12	ppbv	0.01	AC-058	15-Mar-17
17030077-001	Freon-114	I	0.02	ppbv	0.02	AC-058	15-Mar-17
17030077-001	Freon-12		0.61	ppbv	0.02	AC-058	15-Mar-17
17030077-001	Hexachloro-1,3-butadiene	K, T, U	< 0.50	ppbv	0.50	AC-058	15-Mar-17
17030077-001	Isobutane		1.01	ppbv	0.02	AC-058	15-Mar-17
17030077-001	Isopentane		0.99	ppbv	0.03	AC-058	15-Mar-17
17030077-001	Isoprene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-001	Isopropyl alcohol	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030077-001	Isopropylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-001	m,p-Xylene		0.25	ppbv	0.03	AC-058	15-Mar-17
17030077-001	m-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030077-001	m-Ethyltoluene	K, T, U	< 0.08	ppbv	0.08	AC-058	15-Mar-17
17030077-001	Methyl butyl ketone	K, T, U	< 0.50	ppbv	0.50	AC-058	15-Mar-17
17030077-001	Methyl ethyl ketone	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030077-001	Methyl isobutyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030077-001	Methyl methacrylate	K, T, U	< 0.07	ppbv	0.07	AC-058	15-Mar-17
17030077-001	Methyl tert butyl ether	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030077-001	Methylcyclohexane		0.29	ppbv	0.01	AC-058	15-Mar-17
17030077-001	Methylcyclopentane		0.27	ppbv	0.02	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
LICA/VOC/Bonnyville/Mar 2, 2017	14995	Ambient Air	02-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030077	REPORT CREATED:	13-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030077-001	Methylene chloride	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030077-001	n-Butane		2.37	ppbv	0.03	AC-058	15-Mar-17
17030077-001	n-Decane	K, T, U	< 0.06	ppbv	0.06	AC-058	15-Mar-17
17030077-001	n-Dodecane	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030077-001	n-Heptane		0.14	ppbv	0.01	AC-058	15-Mar-17
17030077-001	n-Hexane		0.23	ppbv	0.01	AC-058	15-Mar-17
17030077-001	n-Octane		0.06	ppbv	0.02	AC-058	15-Mar-17
17030077-001	n-Pentane		0.6	ppbv	0.1	AC-058	15-Mar-17
17030077-001	n-Propylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030077-001	n-Undecane	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030077-001	Naphthalene	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030077-001	n-Nonane		0.03	ppbv	0.01	AC-058	15-Mar-17
17030077-001	o-Ethyltoluene	I	0.02	ppbv	0.01	AC-058	15-Mar-17
17030077-001	o-Xylene		0.08	ppbv	0.01	AC-058	15-Mar-17
17030077-001	p-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030077-001	p-Ethyltoluene	K, T, U	< 0.07	ppbv	0.07	AC-058	15-Mar-17
17030077-001	Styrene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030077-001	Tetrachloroethylene	I	0.05	ppbv	0.04	AC-058	15-Mar-17
17030077-001	Tetrahydrofuran	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030077-001	Toluene		0.33	ppbv	0.01	AC-058	15-Mar-17
17030077-001	trans-1,2-Dichloroethylene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-001	trans-1,3-Dichloropropylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030077-001	trans-2-Butene		0.03	ppbv	0.01	AC-058	15-Mar-17
17030077-001	trans-2-Pentene		0.03	ppbv	0.02	AC-058	15-Mar-17
17030077-001	Trichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
LICA/VOC/Bonnyville/Mar 2, 2017	14995	Ambient Air	02-Mar-17	0:00
DESCRIPTION:	Bonnyville - AER			
REPORT NUMBER:	17030077	REPORT CREATED:	13-Apr-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030077-001	Vinyl acetate	K, T, U	< 0.4 ppbv	0.4	AC-058	15-Mar-17
17030077-001	Vinyl chloride	K, T, U	< 0.02 ppbv	0.02	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 08, 2017	1831	Ambient Air	08-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030116	REPORT CREATED:	11-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030116-001	1,1,1-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-001	1,1,2,2-Tetrachloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-001	1,1,2-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-001	1,1-Dichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-001	1,1-Dichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030116-001	1,2,3-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030116-001	1,2,4-Trichlorobenzene	K, T, U	< 0.8	ppbv	0.8	AC-058	15-Mar-17
17030116-001	1,2,4-Trimethylbenzene		0.10	ppbv	0.05	AC-058	15-Mar-17
17030116-001	1,2-Dibromoethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-001	1,2-Dichlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030116-001	1,2-Dichloroethane	I	0.02	ppbv	0.01	AC-058	15-Mar-17
17030116-001	1,2-Dichloropropane	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030116-001	1,3,5-Trimethylbenzene		0.05	ppbv	0.02	AC-058	15-Mar-17
17030116-001	1,3-Butadiene	I	0.03	ppbv	0.02	AC-058	15-Mar-17
17030116-001	1,3-Dichlorobenzene	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030116-001	1,4-Dichlorobenzene	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-001	1,4-Dioxane	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-001	1-Butene		0.14	ppbv	0.02	AC-058	15-Mar-17
17030116-001	1-Hexene	I	0.03	ppbv	0.02	AC-058	15-Mar-17
17030116-001	1-Pentene		0.04	ppbv	0.01	AC-058	15-Mar-17
17030116-001	2,2,4-Trimethylpentane		0.14	ppbv	0.01	AC-058	15-Mar-17
17030116-001	2,2-Dimethylbutane		0.05	ppbv	0.01	AC-058	15-Mar-17
17030116-001	2,3,4-Trimethylpentane		0.06	ppbv	0.01	AC-058	15-Mar-17
17030116-001	2,3-Dimethylbutane		0.08	ppbv	0.02	AC-058	15-Mar-17
17030116-001	2,3-Dimethylpentane		0.12	ppbv	0.02	AC-058	15-Mar-17

Report certified by:	Krista Gegolick, Account Coordinator	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services
Date:	May-11-17	Inquiries:	(780) 632 8455
		E-mail:	EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
.ICA/VOC/Bonnyville/Mar 08, 2017	1831	Ambient Air	08-Mar-17	0:00
DESCRIPTION:	Bonnyville - AER			
REPORT NUMBER:	17030116	REPORT CREATED:	11-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030116-001	2,4-Dimethylpentane		0.06	ppbv	0.01	AC-058	15-Mar-17
17030116-001	2-Methylheptane		0.06	ppbv	0.01	AC-058	15-Mar-17
17030116-001	2-Methylhexane		0.08	ppbv	0.01	AC-058	15-Mar-17
17030116-001	2-Methylpentane		0.21	ppbv	0.01	AC-058	15-Mar-17
17030116-001	3-Methylheptane		0.05	ppbv	0.02	AC-058	15-Mar-17
17030116-001	3-Methylhexane		0.10	ppbv	0.02	AC-058	15-Mar-17
17030116-001	3-Methylpentane		0.16	ppbv	0.01	AC-058	15-Mar-17
17030116-001	Acetone		1.4	ppbv	0.4	AC-058	15-Mar-17
17030116-001	Acrolein	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030116-001	Benzene		0.30	ppbv	0.01	AC-058	15-Mar-17
17030116-001	Benzyl chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-001	Bromodichloromethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-001	Bromoform	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-001	Bromomethane	I	0.01	ppbv	0.01	AC-058	15-Mar-17
17030116-001	Carbon disulfide	I	0.08	ppbv	0.01	AC-058	15-Mar-17
17030116-001	Carbon tetrachloride	I	0.13	ppbv	0.01	AC-058	15-Mar-17
17030116-001	Chlorobenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-001	Chloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-001	Chloroform	I	0.03	ppbv	0.02	AC-058	15-Mar-17
17030116-001	Chloromethane		0.66	ppbv	0.02	AC-058	15-Mar-17
17030116-001	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030116-001	cis-1,3-Dichloropropene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030116-001	cis-2-Butene		0.05	ppbv	0.02	AC-058	15-Mar-17
17030116-001	cis-2-Pentene		0.03	ppbv	0.02	AC-058	15-Mar-17
17030116-001	Cyclohexane		0.14	ppbv	0.02	AC-058	15-Mar-17

Report certified by: Krista Gegolick, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-11-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 08, 2017	1831	Ambient Air	08-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030116	REPORT CREATED:	11-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030116-001	Cyclopentane		0.08	ppbv	0.01	AC-058	15-Mar-17
17030116-001	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030116-001	Ethanol		1.3	ppbv	0.3	AC-058	15-Mar-17
17030116-001	Ethyl acetate	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-001	Ethylbenzene		0.08	ppbv	0.01	AC-058	15-Mar-17
17030116-001	Freon-11		0.30	ppbv	0.02	AC-058	15-Mar-17
17030116-001	Freon-113	I	0.13	ppbv	0.01	AC-058	15-Mar-17
17030116-001	Freon-114	I	0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-001	Freon-12		0.63	ppbv	0.02	AC-058	15-Mar-17
17030116-001	Hexachloro-1,3-butadiene	K, T, U	< 0.50	ppbv	0.50	AC-058	15-Mar-17
17030116-001	Isobutane		0.59	ppbv	0.02	AC-058	15-Mar-17
17030116-001	Isopentane		0.61	ppbv	0.03	AC-058	15-Mar-17
17030116-001	Isoprene		0.03	ppbv	0.01	AC-058	15-Mar-17
17030116-001	Isopropyl alcohol	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-001	Isopropylbenzene		0.02	ppbv	0.01	AC-058	15-Mar-17
17030116-001	m,p-Xylene		0.27	ppbv	0.03	AC-058	15-Mar-17
17030116-001	m-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030116-001	m-Ethyltoluene	K, T, U	< 0.08	ppbv	0.08	AC-058	15-Mar-17
17030116-001	Methyl butyl ketone	K, T, U	< 0.50	ppbv	0.50	AC-058	15-Mar-17
17030116-001	Methyl ethyl ketone	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030116-001	Methyl isobutyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-001	Methyl methacrylate	K, T, U	< 0.07	ppbv	0.07	AC-058	15-Mar-17
17030116-001	Methyl tert butyl ether	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030116-001	Methylcyclohexane		0.15	ppbv	0.01	AC-058	15-Mar-17
17030116-001	Methylcyclopentane		0.28	ppbv	0.02	AC-058	15-Mar-17

Report certified by:	Krista Gegolick, Account Coordinator	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services
Date:	May-11-17	Inquiries:	(780) 632 8455
		E-mail:	EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 08, 2017	1831	Ambient Air	08-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030116	REPORT CREATED:	11-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030116-001	Methylene chloride	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030116-001	n-Butane		1.19	ppbv	0.03	AC-058	15-Mar-17
17030116-001	n-Decane	K, T, U	< 0.06	ppbv	0.06	AC-058	15-Mar-17
17030116-001	n-Dodecane	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-001	n-Heptane		0.11	ppbv	0.01	AC-058	15-Mar-17
17030116-001	n-Hexane		0.40	ppbv	0.01	AC-058	15-Mar-17
17030116-001	n-Octane		0.07	ppbv	0.02	AC-058	15-Mar-17
17030116-001	n-Pentane		0.3	ppbv	0.1	AC-058	15-Mar-17
17030116-001	n-Propylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030116-001	n-Undecane	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030116-001	Naphthalene	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030116-001	n-Nonane		0.05	ppbv	0.01	AC-058	15-Mar-17
17030116-001	o-Ethyltoluene	I	0.04	ppbv	0.01	AC-058	15-Mar-17
17030116-001	o-Xylene		0.10	ppbv	0.01	AC-058	15-Mar-17
17030116-001	p-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030116-001	p-Ethyltoluene	K, T, U	< 0.07	ppbv	0.07	AC-058	15-Mar-17
17030116-001	Styrene	I	0.05	ppbv	0.04	AC-058	15-Mar-17
17030116-001	Tetrachloroethylene	I	0.05	ppbv	0.04	AC-058	15-Mar-17
17030116-001	Tetrahydrofuran	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-001	Toluene		0.31	ppbv	0.01	AC-058	15-Mar-17
17030116-001	trans-1,2-Dichloroethylene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030116-001	trans-1,3-Dichloropropylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030116-001	trans-2-Butene		0.06	ppbv	0.01	AC-058	15-Mar-17
17030116-001	trans-2-Pentene		0.05	ppbv	0.02	AC-058	15-Mar-17
17030116-001	Trichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17

Report certified by: Krista Gegolick, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-11-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 08, 2017	1831	Ambient Air	08-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030116	REPORT CREATED:	11-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030116-001	Vinyl acetate	K, T, U	< 0.4 ppbv	0.4	AC-058	15-Mar-17
17030116-001	Vinyl chloride	K, T, U	< 0.02 ppbv	0.02	AC-058	15-Mar-17

Report certified by: Krista Gegolick, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-11-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 14, 2017	2665	Ambient Air	14-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030212	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030212-001	1,1,1-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	1,1,2,2-Tetrachloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	1,1,2-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	1,1-Dichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	1,1-Dichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	22-Mar-17
17030212-001	1,2,3-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	22-Mar-17
17030212-001	1,2,4-Trichlorobenzene	K, T, U	< 0.8	ppbv	0.8	AC-058	22-Mar-17
17030212-001	1,2,4-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	22-Mar-17
17030212-001	1,2-Dibromoethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	1,2-Dichlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	22-Mar-17
17030212-001	1,2-Dichloroethane	I	0.02	ppbv	0.01	AC-058	22-Mar-17
17030212-001	1,2-Dichloropropane	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030212-001	1,3,5-Trimethylbenzene		0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	1,3-Butadiene	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	1,3-Dichlorobenzene	K, T, U	< 0.3	ppbv	0.3	AC-058	22-Mar-17
17030212-001	1,4-Dichlorobenzene	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030212-001	1,4-Dioxane	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030212-001	1-Butene		0.14	ppbv	0.02	AC-058	22-Mar-17
17030212-001	1-Hexene	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	1-Pentene		0.02	ppbv	0.01	AC-058	22-Mar-17
17030212-001	2,2,4-Trimethylpentane		0.05	ppbv	0.01	AC-058	22-Mar-17
17030212-001	2,2-Dimethylbutane		0.03	ppbv	0.01	AC-058	22-Mar-17
17030212-001	2,3,4-Trimethylpentane		0.02	ppbv	0.01	AC-058	22-Mar-17
17030212-001	2,3-Dimethylbutane		0.07	ppbv	0.02	AC-058	22-Mar-17
17030212-001	2,3-Dimethylpentane		0.07	ppbv	0.02	AC-058	22-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 14, 2017	2665	Ambient Air	14-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030212	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030212-001	2,4-Dimethylpentane		0.03	ppbv	0.01	AC-058	22-Mar-17
17030212-001	2-Methylheptane		0.03	ppbv	0.01	AC-058	22-Mar-17
17030212-001	2-Methylhexane		0.07	ppbv	0.01	AC-058	22-Mar-17
17030212-001	2-Methylpentane		0.25	ppbv	0.01	AC-058	22-Mar-17
17030212-001	3-Methylheptane		0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	3-Methylhexane		0.09	ppbv	0.02	AC-058	22-Mar-17
17030212-001	3-Methylpentane		0.16	ppbv	0.01	AC-058	22-Mar-17
17030212-001	Acetone		2.7	ppbv	0.4	AC-058	22-Mar-17
17030212-001	Acrolein	K, T, U	< 0.3	ppbv	0.3	AC-058	22-Mar-17
17030212-001	Benzene		0.20	ppbv	0.01	AC-058	22-Mar-17
17030212-001	Benzyl chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030212-001	Bromodichloromethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	Bromoform	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	Bromomethane	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030212-001	Carbon disulfide	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030212-001	Carbon tetrachloride	I	0.10	ppbv	0.01	AC-058	22-Mar-17
17030212-001	Chlorobenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	Chloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	Chloroform	I	0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	Chloromethane		0.63	ppbv	0.02	AC-058	22-Mar-17
17030212-001	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030212-001	cis-1,3-Dichloropropene	K, T, U	< 0.04	ppbv	0.04	AC-058	22-Mar-17
17030212-001	cis-2-Butene	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	cis-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	Cyclohexane		0.11	ppbv	0.02	AC-058	22-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 14, 2017	2665	Ambient Air	14-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030212	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030212-001	Cyclopentane		0.07	ppbv	0.01	AC-058	22-Mar-17
17030212-001	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030212-001	Ethanol		1.8	ppbv	0.3	AC-058	22-Mar-17
17030212-001	Ethyl acetate	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030212-001	Ethylbenzene		0.03	ppbv	0.01	AC-058	22-Mar-17
17030212-001	Freon-11	I	0.29	ppbv	0.02	AC-058	22-Mar-17
17030212-001	Freon-113	I	0.09	ppbv	0.01	AC-058	22-Mar-17
17030212-001	Freon-114	I	0.02	ppbv	0.02	AC-058	22-Mar-17
17030212-001	Freon-12		0.63	ppbv	0.02	AC-058	22-Mar-17
17030212-001	Hexachloro-1,3-butadiene	K, T, U	< 0.50	ppbv	0.50	AC-058	22-Mar-17
17030212-001	Isobutane		2.37	ppbv	0.02	AC-058	22-Mar-17
17030212-001	Isopentane		1.26	ppbv	0.03	AC-058	22-Mar-17
17030212-001	Isoprene	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030212-001	Isopropyl alcohol	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030212-001	Isopropylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	22-Mar-17
17030212-001	m,p-Xylene		0.09	ppbv	0.03	AC-058	22-Mar-17
17030212-001	m-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	22-Mar-17
17030212-001	m-Ethyltoluene	K, T, U	< 0.08	ppbv	0.08	AC-058	22-Mar-17
17030212-001	Methyl butyl ketone	K, T, U	< 0.50	ppbv	0.50	AC-058	22-Mar-17
17030212-001	Methyl ethyl ketone		0.6	ppbv	0.3	AC-058	22-Mar-17
17030212-001	Methyl isobutyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	22-Mar-17
17030212-001	Methyl methacrylate	K, T, U	< 0.07	ppbv	0.07	AC-058	22-Mar-17
17030212-001	Methyl tert butyl ether	K, T, U	< 0.03	ppbv	0.03	AC-058	22-Mar-17
17030212-001	Methylcyclohexane		0.16	ppbv	0.01	AC-058	22-Mar-17
17030212-001	Methylcyclopentane		0.19	ppbv	0.02	AC-058	22-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 14, 2017	2665	Ambient Air	14-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030212	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030212-001	Methylene chloride	K, T, U	< 0.3 ppbv	0.3	AC-058	22-Mar-17
17030212-001	n-Butane		4.63 ppbv	0.03	AC-058	22-Mar-17
17030212-001	n-Decane	K, T, U	< 0.06 ppbv	0.06	AC-058	22-Mar-17
17030212-001	n-Dodecane	K, T, U	< 0.4 ppbv	0.4	AC-058	22-Mar-17
17030212-001	n-Heptane		0.12 ppbv	0.01	AC-058	22-Mar-17
17030212-001	n-Hexane		0.34 ppbv	0.01	AC-058	22-Mar-17
17030212-001	n-Octane		0.04 ppbv	0.02	AC-058	22-Mar-17
17030212-001	n-Pentane		1.1 ppbv	0.1	AC-058	22-Mar-17
17030212-001	n-Propylbenzene	K, T, U	< 0.05 ppbv	0.05	AC-058	22-Mar-17
17030212-001	n-Undecane	K, T, U	< 0.5 ppbv	0.5	AC-058	22-Mar-17
17030212-001	Naphthalene	K, T, U	< 0.5 ppbv	0.5	AC-058	22-Mar-17
17030212-001	n-Nonane		0.02 ppbv	0.01	AC-058	22-Mar-17
17030212-001	o-Ethyltoluene	I	0.01 ppbv	0.01	AC-058	22-Mar-17
17030212-001	o-Xylene		0.04 ppbv	0.01	AC-058	22-Mar-17
17030212-001	p-Diethylbenzene	K, T, U	< 0.04 ppbv	0.04	AC-058	22-Mar-17
17030212-001	p-Ethyltoluene	K, T, U	< 0.07 ppbv	0.07	AC-058	22-Mar-17
17030212-001	Styrene	K, T, U	< 0.04 ppbv	0.04	AC-058	22-Mar-17
17030212-001	Tetrachloroethylene	I	0.04 ppbv	0.04	AC-058	22-Mar-17
17030212-001	Tetrahydrofuran	K, T, U	< 0.4 ppbv	0.4	AC-058	22-Mar-17
17030212-001	Toluene		0.14 ppbv	0.01	AC-058	22-Mar-17
17030212-001	trans-1,2-Dichloroethylene	K, T, U	< 0.01 ppbv	0.01	AC-058	22-Mar-17
17030212-001	trans-1,3-Dichloropropylene	K, T, U	< 0.04 ppbv	0.04	AC-058	22-Mar-17
17030212-001	trans-2-Butene		0.03 ppbv	0.01	AC-058	22-Mar-17
17030212-001	trans-2-Pentene	K, T, U	< 0.02 ppbv	0.02	AC-058	22-Mar-17
17030212-001	Trichloroethylene	K, T, U	< 0.04 ppbv	0.04	AC-058	22-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
.ICA/VOC/Bonnyville/Mar 14, 2017	2665	Ambient Air	14-Mar-17	0:00
DESCRIPTION:	Bonnyville - AER			
REPORT NUMBER:	17030212	REPORT CREATED:	02-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030212-001	Vinyl acetate	K, T, U	< 0.4 ppbv	0.4	AC-058	22-Mar-17
17030212-001	Vinyl chloride	K, T, U	< 0.02 ppbv	0.02	AC-058	22-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 20, 2017	2471	Ambient Air	20-Mar-17 0:00
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030303-001	1,1,1-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	1,1,2,2-Tetrachloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	1,1,2-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	1,1-Dichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	1,1-Dichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030303-001	1,2,3-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	04-Apr-17
17030303-001	1,2,4-Trichlorobenzene	K, T, U	< 0.8	ppbv	0.8	AC-058	04-Apr-17
17030303-001	1,2,4-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	04-Apr-17
17030303-001	1,2-Dibromoethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	1,2-Dichlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-001	1,2-Dichloroethane	I	0.02	ppbv	0.01	AC-058	04-Apr-17
17030303-001	1,2-Dichloropropane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	1,3,5-Trimethylbenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	1,3-Butadiene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	1,3-Dichlorobenzene	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030303-001	1,4-Dichlorobenzene	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-001	1,4-Dioxane	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-001	1-Butene		0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	1-Hexene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	1-Pentene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	2,2,4-Trimethylpentane		0.02	ppbv	0.01	AC-058	04-Apr-17
17030303-001	2,2-Dimethylbutane		0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	2,3,4-Trimethylpentane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	2,3-Dimethylbutane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	2,3-Dimethylpentane		0.02	ppbv	0.02	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 20, 2017	2471	Ambient Air	20-Mar-17 0:00
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030303-001	2,4-Dimethylpentane		0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	2-Methylheptane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	2-Methylhexane		0.02	ppbv	0.01	AC-058	04-Apr-17
17030303-001	2-Methylpentane		0.05	ppbv	0.01	AC-058	04-Apr-17
17030303-001	3-Methylheptane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	3-Methylhexane		0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	3-Methylpentane		0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	Acetone		1.4	ppbv	0.4	AC-058	04-Apr-17
17030303-001	Acrolein	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030303-001	Benzene		0.14	ppbv	0.01	AC-058	04-Apr-17
17030303-001	Benzyl chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-001	Bromodichloromethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	Bromoform	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	Bromomethane	I	0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	Carbon disulfide	I	0.12	ppbv	0.01	AC-058	04-Apr-17
17030303-001	Carbon tetrachloride	I	0.08	ppbv	0.01	AC-058	04-Apr-17
17030303-001	Chlorobenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	Chloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	Chloroform	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	Chloromethane		0.63	ppbv	0.02	AC-058	04-Apr-17
17030303-001	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	cis-1,3-Dichloropropene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030303-001	cis-2-Butene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	cis-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	Cyclohexane		0.03	ppbv	0.02	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 20, 2017	2471	Ambient Air	20-Mar-17 0:00
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030303-001	Cyclopentane		0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	Ethanol		0.4	ppbv	0.3	AC-058	04-Apr-17
17030303-001	Ethyl acetate	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-001	Ethylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	Freon-11	I	0.23	ppbv	0.02	AC-058	04-Apr-17
17030303-001	Freon-113	I	0.09	ppbv	0.01	AC-058	04-Apr-17
17030303-001	Freon-114	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-001	Freon-12		0.38	ppbv	0.02	AC-058	04-Apr-17
17030303-001	Hexachloro-1,3-butadiene	K, T, U	< 0.50	ppbv	0.50	AC-058	04-Apr-17
17030303-001	Isobutane		0.29	ppbv	0.02	AC-058	04-Apr-17
17030303-001	Isopentane		0.32	ppbv	0.03	AC-058	04-Apr-17
17030303-001	Isoprene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	Isopropyl alcohol	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-001	Isopropylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-001	m,p-Xylene	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-001	m-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030303-001	m-Ethyltoluene	K, T, U	< 0.08	ppbv	0.08	AC-058	04-Apr-17
17030303-001	Methyl butyl ketone	K, T, U	< 0.50	ppbv	0.50	AC-058	04-Apr-17
17030303-001	Methyl ethyl ketone	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030303-001	Methyl isobutyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-001	Methyl methacrylate	K, T, U	< 0.07	ppbv	0.07	AC-058	04-Apr-17
17030303-001	Methyl tert butyl ether	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-001	Methylcyclohexane		0.04	ppbv	0.01	AC-058	04-Apr-17
17030303-001	Methylcyclopentane		0.02	ppbv	0.02	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 20, 2017	2471	Ambient Air	20-Mar-17 0:00
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030303-001	Methylene chloride	K, T, U	< 0.3 ppbv	0.3	AC-058	04-Apr-17
17030303-001	n-Butane		0.97 ppbv	0.03	AC-058	04-Apr-17
17030303-001	n-Decane	K, T, U	< 0.06 ppbv	0.06	AC-058	04-Apr-17
17030303-001	n-Dodecane	K, T, U	< 0.4 ppbv	0.4	AC-058	04-Apr-17
17030303-001	n-Heptane		0.03 ppbv	0.01	AC-058	04-Apr-17
17030303-001	n-Hexane	K, T, U	< 0.01 ppbv	0.01	AC-058	04-Apr-17
17030303-001	n-Octane	K, T, U	< 0.02 ppbv	0.02	AC-058	04-Apr-17
17030303-001	n-Pentane		0.2 ppbv	0.1	AC-058	04-Apr-17
17030303-001	n-Propylbenzene	K, T, U	< 0.05 ppbv	0.05	AC-058	04-Apr-17
17030303-001	n-Undecane	K, T, U	< 0.5 ppbv	0.5	AC-058	04-Apr-17
17030303-001	Naphthalene	K, T, U	< 0.5 ppbv	0.5	AC-058	04-Apr-17
17030303-001	n-Nonane	K, T, U	< 0.01 ppbv	0.01	AC-058	04-Apr-17
17030303-001	o-Ethyltoluene	K, T, U	< 0.01 ppbv	0.01	AC-058	04-Apr-17
17030303-001	o-Xylene		0.02 ppbv	0.01	AC-058	04-Apr-17
17030303-001	p-Diethylbenzene	K, T, U	< 0.04 ppbv	0.04	AC-058	04-Apr-17
17030303-001	p-Ethyltoluene	K, T, U	< 0.07 ppbv	0.07	AC-058	04-Apr-17
17030303-001	Styrene	K, T, U	< 0.04 ppbv	0.04	AC-058	04-Apr-17
17030303-001	Tetrachloroethylene	K, T, U	< 0.04 ppbv	0.04	AC-058	04-Apr-17
17030303-001	Tetrahydrofuran	K, T, U	< 0.4 ppbv	0.4	AC-058	04-Apr-17
17030303-001	Toluene		0.07 ppbv	0.01	AC-058	04-Apr-17
17030303-001	trans-1,2-Dichloroethylene	K, T, U	< 0.01 ppbv	0.01	AC-058	04-Apr-17
17030303-001	trans-1,3-Dichloropropylene	K, T, U	< 0.04 ppbv	0.04	AC-058	04-Apr-17
17030303-001	trans-2-Butene		0.02 ppbv	0.01	AC-058	04-Apr-17
17030303-001	trans-2-Pentene	K, T, U	< 0.02 ppbv	0.02	AC-058	04-Apr-17
17030303-001	Trichloroethylene	K, T, U	< 0.04 ppbv	0.04	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 20, 2017	2471	Ambient Air	20-Mar-17 0:00
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030303-001	Vinyl acetate	K, T, U	< 0.4 ppbv	0.4	AC-058	04-Apr-17
17030303-001	Vinyl chloride	K, T, U	< 0.02 ppbv	0.02	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 26, 2017	S5600	Ambient Air	26-Mar-17 0:00
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030303-003	1,1,1-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	1,1,2,2-Tetrachloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	1,1,2-Trichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	1,1-Dichloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	1,1-Dichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030303-003	1,2,3-Trimethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	04-Apr-17
17030303-003	1,2,4-Trichlorobenzene	K, T, U	< 0.8	ppbv	0.8	AC-058	04-Apr-17
17030303-003	1,2,4-Trimethylbenzene		0.05	ppbv	0.05	AC-058	04-Apr-17
17030303-003	1,2-Dibromoethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	1,2-Dichlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-003	1,2-Dichloroethane	I	0.02	ppbv	0.01	AC-058	04-Apr-17
17030303-003	1,2-Dichloropropane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-003	1,3,5-Trimethylbenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	1,3-Butadiene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	1,3-Dichlorobenzene	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030303-003	1,4-Dichlorobenzene	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-003	1,4-Dioxane	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-003	1-Butene		0.06	ppbv	0.02	AC-058	04-Apr-17
17030303-003	1-Hexene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	1-Pentene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-003	2,2,4-Trimethylpentane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-003	2,2-Dimethylbutane		0.02	ppbv	0.01	AC-058	04-Apr-17
17030303-003	2,3,4-Trimethylpentane		0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-003	2,3-Dimethylbutane		0.06	ppbv	0.02	AC-058	04-Apr-17
17030303-003	2,3-Dimethylpentane		0.11	ppbv	0.02	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 26, 2017	S5600	Ambient Air	26-Mar-17 0:00
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030303-003	2,4-Dimethylpentane		0.04	ppbv	0.01	AC-058	04-Apr-17
17030303-003	2-Methylheptane		0.11	ppbv	0.01	AC-058	04-Apr-17
17030303-003	2-Methylhexane		0.26	ppbv	0.01	AC-058	04-Apr-17
17030303-003	2-Methylpentane		0.23	ppbv	0.01	AC-058	04-Apr-17
17030303-003	3-Methylheptane		0.06	ppbv	0.02	AC-058	04-Apr-17
17030303-003	3-Methylhexane		0.27	ppbv	0.02	AC-058	04-Apr-17
17030303-003	3-Methylpentane		0.14	ppbv	0.01	AC-058	04-Apr-17
17030303-003	Acetone		3.2	ppbv	0.4	AC-058	04-Apr-17
17030303-003	Acrolein	K, T, U	< 0.3	ppbv	0.3	AC-058	04-Apr-17
17030303-003	Benzene		0.30	ppbv	0.01	AC-058	04-Apr-17
17030303-003	Benzyl chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-003	Bromodichloromethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	Bromoform	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	Bromomethane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-003	Carbon disulfide	I	0.09	ppbv	0.01	AC-058	04-Apr-17
17030303-003	Carbon tetrachloride	I	0.07	ppbv	0.01	AC-058	04-Apr-17
17030303-003	Chlorobenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	Chloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	Chloroform	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	Chloromethane		0.65	ppbv	0.02	AC-058	04-Apr-17
17030303-003	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-003	cis-1,3-Dichloropropene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030303-003	cis-2-Butene		0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	cis-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	Cyclohexane		0.26	ppbv	0.02	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 26, 2017	S5600	Ambient Air	26-Mar-17 0:00
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030303-003	Cyclopentane		0.10	ppbv	0.01	AC-058	04-Apr-17
17030303-003	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-003	Ethanol		1.3	ppbv	0.3	AC-058	04-Apr-17
17030303-003	Ethyl acetate	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-003	Ethylbenzene		0.10	ppbv	0.01	AC-058	04-Apr-17
17030303-003	Freon-11	I	0.22	ppbv	0.02	AC-058	04-Apr-17
17030303-003	Freon-113	I	0.08	ppbv	0.01	AC-058	04-Apr-17
17030303-003	Freon-114	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	Freon-12		0.36	ppbv	0.02	AC-058	04-Apr-17
17030303-003	Hexachloro-1,3-butadiene	K, T, U	< 0.50	ppbv	0.50	AC-058	04-Apr-17
17030303-003	Isobutane		0.99	ppbv	0.02	AC-058	04-Apr-17
17030303-003	Isopentane		0.92	ppbv	0.03	AC-058	04-Apr-17
17030303-003	Isoprene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-003	Isopropyl alcohol	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-003	Isopropylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-003	m,p-Xylene		0.39	ppbv	0.03	AC-058	04-Apr-17
17030303-003	m-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030303-003	m-Ethyltoluene	K, T, U	< 0.08	ppbv	0.08	AC-058	04-Apr-17
17030303-003	Methyl butyl ketone	K, T, U	< 0.50	ppbv	0.50	AC-058	04-Apr-17
17030303-003	Methyl ethyl ketone		0.3	ppbv	0.3	AC-058	04-Apr-17
17030303-003	Methyl isobutyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-003	Methyl methacrylate	K, T, U	< 0.07	ppbv	0.07	AC-058	04-Apr-17
17030303-003	Methyl tert butyl ether	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-003	Methylcyclohexane		0.54	ppbv	0.01	AC-058	04-Apr-17
17030303-003	Methylcyclopentane		0.31	ppbv	0.02	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/VOC/Bonnyville/Mar 26, 2017	S5600	Ambient Air	26-Mar-17 0:00
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030303-003	Methylene chloride		0.3	ppbv	0.3	AC-058	04-Apr-17
17030303-003	n-Butane		2.53	ppbv	0.03	AC-058	04-Apr-17
17030303-003	n-Decane	K, T, U	< 0.06	ppbv	0.06	AC-058	04-Apr-17
17030303-003	n-Dodecane	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-003	n-Heptane		0.43	ppbv	0.01	AC-058	04-Apr-17
17030303-003	n-Hexane		0.36	ppbv	0.01	AC-058	04-Apr-17
17030303-003	n-Octane		0.16	ppbv	0.02	AC-058	04-Apr-17
17030303-003	n-Pentane		0.7	ppbv	0.1	AC-058	04-Apr-17
17030303-003	n-Propylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	04-Apr-17
17030303-003	n-Undecane	K, T, U	< 0.5	ppbv	0.5	AC-058	04-Apr-17
17030303-003	Naphthalene	K, T, U	< 0.5	ppbv	0.5	AC-058	04-Apr-17
17030303-003	n-Nonane		0.08	ppbv	0.01	AC-058	04-Apr-17
17030303-003	o-Ethyltoluene	I	0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-003	o-Xylene		0.19	ppbv	0.01	AC-058	04-Apr-17
17030303-003	p-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030303-003	p-Ethyltoluene	K, T, U	< 0.07	ppbv	0.07	AC-058	04-Apr-17
17030303-003	Styrene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030303-003	Tetrachloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030303-003	Tetrahydrofuran	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-003	Toluene		0.80	ppbv	0.01	AC-058	04-Apr-17
17030303-003	trans-1,2-Dichloroethylene	I	0.08	ppbv	0.01	AC-058	04-Apr-17
17030303-003	trans-1,3-Dichloropropylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030303-003	trans-2-Butene		0.02	ppbv	0.01	AC-058	04-Apr-17
17030303-003	trans-2-Pentene	K, T, U	< 0.02	ppbv	0.02	AC-058	04-Apr-17
17030303-003	Trichloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
.ICA/VOC/Bonnyville/Mar 26, 2017	S5600	Ambient Air	26-Mar-17	0:00
DESCRIPTION:	Bonnyville- AER			
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030303-003	Vinyl acetate	K, T, U	< 0.4 ppbv	0.4	AC-058	04-Apr-17
17030303-003	Vinyl chloride	K, T, U	< 0.02 ppbv	0.02	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

PAHS SAMPLES

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
LICA/PUF/Bonnyville/Mar 2, 2017	TE02	Air Filter	02-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030077	REPORT CREATED:	13-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030077-002	1-Methylnaphthalene		0.44	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	2-Methylnaphthalene		0.56	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	3-Methylcholanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	7,12-Dimethylbenz(a)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Acenaphthene		0.15	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Acenaphthylene		0.04	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Acridine	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Anthracene		0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Benzo(a)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Benzo(a)pyrene		0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Benzo(b,j,k)fluoranthene		0.04	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Benzo(c)phenanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Benzo(e)pyrene		0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Benzo(ghi)perylene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Chrysene		0.02	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Dibenzo(a,h)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Dibenzo(a,i)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Dibenzo(a,l)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Dibenzo(ah)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Fluoranthene		0.09	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Fluorene		0.13	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Indeno(1,2,3-cd)pyrene		0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Naphthalene		0.33	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Perylene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Phenanthrene		0.31	ug/Filter	0.01	NA-017	16-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
LICA/PUF/Bonnyville/Mar 2, 2017	TE02	Air Filter	02-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030077	REPORT CREATED:	13-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030077-002	Pyrene		0.06 ug/Filter	0.01	NA-017	16-Mar-17
17030077-002	Retene		0.05 ug/Filter	0.01	NA-017	16-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/PUF/Bonnyville/Mar 08, 2017	9801	Air Filter	08-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030116	REPORT CREATED:	11-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030116-002	1-Methylnaphthalene		1.21	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	2-Methylnaphthalene		2.07	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	3-Methylcholanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	7,12-Dimethylbenz(a)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Acenaphthene		0.34	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Acenaphthylene		0.12	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Acridine	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Anthracene		0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Benzo(a)anthracene		0.02	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Benzo(a)pyrene		0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Benzo(b,j,k)fluoranthene		0.06	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Benzo(c)phenanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Benzo(e)pyrene		0.02	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Benzo(ghi)perylene		0.03	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Chrysene		0.04	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Dibenzo(a,h)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Dibenzo(a,i)pyrene		0.06	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Dibenzo(a,l)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Dibenzo(ah)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Fluoranthene		0.09	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Fluorene		0.12	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Indeno(1,2,3-cd)pyrene		0.03	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Naphthalene		3.89	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Perylene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Phenanthrene		0.36	ug/Filter	0.01	NA-017	22-Mar-17

Report certified by: Alberto dos Santos Pereira, Supervisor

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-11-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/PUF/Bonnyville/Mar 08, 2017	9801	Air Filter	08-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030116	REPORT CREATED:	11-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030116-002	Pyrene		0.06 ug/Filter	0.01	NA-017	22-Mar-17
17030116-002	Retene		0.08 ug/Filter	0.01	NA-017	22-Mar-17

Report certified by: Alberto dos Santos Pereira, Supervisor

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-11-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/PUF/Bonnyville/Mar 14, 2017	P13-01	Air Filter	14-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030212	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030212-002	1-Methylnaphthalene		0.11	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	2-Methylnaphthalene		0.12	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	3-Methylcholanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	7,12-Dimethylbenz(a)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Acenaphthene		0.07	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Acenaphthylene		0.06	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Acridine	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Anthracene		0.02	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Benzo(a)anthracene		0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Benzo(a)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Benzo(b,j,k)fluoranthene		0.03	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Benzo(c)phenanthrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Benzo(e)pyrene		0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Benzo(ghi)perylene		0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Chrysene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Dibenzo(a,h)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Dibenzo(a,i)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Dibenzo(a,l)pyrene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Dibenzo(ah)anthracene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Fluoranthene		0.10	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Fluorene		0.10	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Indeno(1,2,3-cd)pyrene		0.02	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Naphthalene		0.11	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Perylene	K, T, U	< 0.01	ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Phenanthrene		0.31	ug/Filter	0.01	NA-017	22-Mar-17

Report certified by: Alberto dos Santos Pereira, Supervisor

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/PUF/Bonnyville/Mar 14, 2017	P13-01	Air Filter	14-Mar-17 0:00
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030212	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030212-002	Pyrene		0.08 ug/Filter	0.01	NA-017	22-Mar-17
17030212-002	Retene		0.04 ug/Filter	0.01	NA-017	22-Mar-17

Report certified by: Alberto dos Santos Pereira, Supervisor

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/PUF/Bonnyville/Mar 20, 2017	TE09	Air Filter	20-Mar-17 0:00
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030303-002	1-Methylnaphthalene		0.21 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	2-Methylnaphthalene		0.40 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	3-Methylcholanthrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	7,12-Dimethylbenz(a)anthracene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Acenaphthene		0.03 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Acenaphthylene		0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Acridine	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Anthracene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Benzo(a)anthracene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Benzo(a)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Benzo(b,j,k)fluoranthene		0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Benzo(c)phenanthrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Benzo(e)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Benzo(ghi)perylene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Chrysene		0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Dibenzo(a,h)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Dibenzo(a,i)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Dibenzo(a,l)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Dibenzo(ah)anthracene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Fluoranthene		0.02 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Fluorene		0.03 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Indeno(1,2,3-cd)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Naphthalene		0.45 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Perylene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Phenanthrene		0.07 ug/Filter	0.01	NA-017	13-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/PUF/Bonnyville/Mar 20, 2017	TE09	Air Filter	20-Mar-17 0:00
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030303-002	Pyrene		0.02 ug/Filter	0.01	NA-017	13-Apr-17
17030303-002	Retene		0.01 ug/Filter	0.01	NA-017	13-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
.ICA/PUF/Bonnyville/Mar 26, 2017	TE01	Air Filter	26-Mar-17 0:00
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030303-004	1-Methylnaphthalene		0.08 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	2-Methylnaphthalene		0.14 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	3-Methylcholanthrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	7,12-Dimethylbenz(a)anthracene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Acenaphthene		0.02 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Acenaphthylene		0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Acridine	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Anthracene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Benzo(a)anthracene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Benzo(a)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Benzo(b,j,k)fluoranthene		0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Benzo(c)phenanthrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Benzo(e)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Benzo(ghi)perylene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Chrysene		0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Dibenzo(a,h)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Dibenzo(a,i)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Dibenzo(a,l)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Dibenzo(ah)anthracene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Fluoranthene		0.05 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Fluorene		0.07 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Indeno(1,2,3-cd)pyrene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Naphthalene		0.12 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Perylene	K, T, U	< 0.01 ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Phenanthrene		0.16 ug/Filter	0.01	NA-017	13-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
.ICA/PUF/Bonnyville/Mar 26, 2017	TE01	Air Filter	26-Mar-17	0:00
DESCRIPTION:	Bonnyville- AER			
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030303-004	Pyrene		0.04	ug/Filter	0.01	NA-017	13-Apr-17
17030303-004	Retene		0.06	ug/Filter	0.01	NA-017	13-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

NMHC CANISTER SAMPLES

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
\\NMHC VOC/Bonnyville/Mar 3, 2	H3286	Ambient Air	03-Mar-17 19:25
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030077	REPORT CREATED:	13-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030077-004	1,1,1-Trichloroethane	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030077-004	1,1,2,2-Tetrachloroethane	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030077-004	1,1,2-Trichloroethane	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030077-004	1,1-Dichloroethane	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030077-004	1,1-Dichloroethylene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030077-004	1,2,3-Trimethylbenzene	I	0.11	ppbv	0.07	AC-058	15-Mar-17
17030077-004	1,2,4-Trichlorobenzene	K, T, U	< 1.1	ppbv	1.1	AC-058	15-Mar-17
17030077-004	1,2,4-Trimethylbenzene		0.37	ppbv	0.07	AC-058	15-Mar-17
17030077-004	1,2-Dibromoethane	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030077-004	1,2-Dichlorobenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030077-004	1,2-Dichloroethane	I	0.03	ppbv	0.01	AC-058	15-Mar-17
17030077-004	1,2-Dichloropropane	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-004	1,3,5-Trimethylbenzene		0.14	ppbv	0.03	AC-058	15-Mar-17
17030077-004	1,3-Butadiene	I	0.17	ppbv	0.03	AC-058	15-Mar-17
17030077-004	1,3-Dichlorobenzene	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030077-004	1,4-Dichlorobenzene	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030077-004	1,4-Dioxane	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030077-004	1-Butene		0.52	ppbv	0.03	AC-058	15-Mar-17
17030077-004	1-Hexene	I	0.09	ppbv	0.03	AC-058	15-Mar-17
17030077-004	1-Pentene		0.11	ppbv	0.01	AC-058	15-Mar-17
17030077-004	2,2,4-Trimethylpentane		0.57	ppbv	0.01	AC-058	15-Mar-17
17030077-004	2,2-Dimethylbutane		0.22	ppbv	0.01	AC-058	15-Mar-17
17030077-004	2,3,4-Trimethylpentane		0.15	ppbv	0.01	AC-058	15-Mar-17
17030077-004	2,3-Dimethylbutane		0.36	ppbv	0.03	AC-058	15-Mar-17
17030077-004	2,3-Dimethylpentane		0.43	ppbv	0.03	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
\\NMHC VOC/Bonnyville/Mar 3, 2	H3286	Ambient Air	03-Mar-17 19:25
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030077	REPORT CREATED:	13-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030077-004	2,4-Dimethylpentane		0.20	ppbv	0.01	AC-058	15-Mar-17
17030077-004	2-Methylheptane		0.19	ppbv	0.01	AC-058	15-Mar-17
17030077-004	2-Methylhexane		0.42	ppbv	0.01	AC-058	15-Mar-17
17030077-004	2-Methylpentane		1.16	ppbv	0.01	AC-058	15-Mar-17
17030077-004	3-Methylheptane		0.16	ppbv	0.03	AC-058	15-Mar-17
17030077-004	3-Methylhexane		0.48	ppbv	0.03	AC-058	15-Mar-17
17030077-004	3-Methylpentane		0.65	ppbv	0.01	AC-058	15-Mar-17
17030077-004	Acetone		4.0	ppbv	0.5	AC-058	15-Mar-17
17030077-004	Acrolein	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030077-004	Benzene		0.69	ppbv	0.01	AC-058	15-Mar-17
17030077-004	Benzyl chloride	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030077-004	Bromodichloromethane	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030077-004	Bromoform	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030077-004	Bromomethane	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-004	Carbon disulfide	I	0.30	ppbv	0.01	AC-058	15-Mar-17
17030077-004	Carbon tetrachloride	I	0.14	ppbv	0.01	AC-058	15-Mar-17
17030077-004	Chlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030077-004	Chloroethane	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030077-004	Chloroform	I	0.04	ppbv	0.03	AC-058	15-Mar-17
17030077-004	Chloromethane		0.72	ppbv	0.03	AC-058	15-Mar-17
17030077-004	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-004	cis-1,3-Dichloropropene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030077-004	cis-2-Butene		0.18	ppbv	0.03	AC-058	15-Mar-17
17030077-004	cis-2-Pentene		0.10	ppbv	0.03	AC-058	15-Mar-17
17030077-004	Cyclohexane		0.48	ppbv	0.03	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
\\NMHC VOC/Bonnyville/Mar 3, 2	H3286	Ambient Air	03-Mar-17 19:25
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030077	REPORT CREATED:	13-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030077-004	Cyclopentane		0.29	ppbv	0.01	AC-058	15-Mar-17
17030077-004	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-004	Ethanol		10.1	ppbv	0.4	AC-058	15-Mar-17
17030077-004	Ethyl acetate	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030077-004	Ethylbenzene		0.26	ppbv	0.01	AC-058	15-Mar-17
17030077-004	Freon-11	I	0.31	ppbv	0.03	AC-058	15-Mar-17
17030077-004	Freon-113	I	0.15	ppbv	0.01	AC-058	15-Mar-17
17030077-004	Freon-114	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030077-004	Freon-12		0.59	ppbv	0.03	AC-058	15-Mar-17
17030077-004	Hexachloro-1,3-butadiene	K, T, U	< 0.68	ppbv	0.68	AC-058	15-Mar-17
17030077-004	Isobutane		3.02	ppbv	0.03	AC-058	15-Mar-17
17030077-004	Isopentane		3.59	ppbv	0.04	AC-058	15-Mar-17
17030077-004	Isoprene		0.10	ppbv	0.01	AC-058	15-Mar-17
17030077-004	Isopropyl alcohol	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030077-004	Isopropylbenzene		0.03	ppbv	0.01	AC-058	15-Mar-17
17030077-004	m,p-Xylene		0.97	ppbv	0.04	AC-058	15-Mar-17
17030077-004	m-Diethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030077-004	m-Ethyltoluene	I	0.24	ppbv	0.11	AC-058	15-Mar-17
17030077-004	Methyl butyl ketone	K, T, U	< 0.68	ppbv	0.68	AC-058	15-Mar-17
17030077-004	Methyl ethyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030077-004	Methyl isobutyl ketone	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030077-004	Methyl methacrylate	K, T, U	< 0.09	ppbv	0.09	AC-058	15-Mar-17
17030077-004	Methyl tert butyl ether	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030077-004	Methylcyclohexane		0.58	ppbv	0.01	AC-058	15-Mar-17
17030077-004	Methylcyclopentane		0.79	ppbv	0.03	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
\\NMHC VOC/Bonnyville/Mar 3, 2	H3286	Ambient Air	03-Mar-17 19:25
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030077	REPORT CREATED:	13-Apr-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030077-004	Methylene chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030077-004	n-Butane		6.24	ppbv	0.04	AC-058	15-Mar-17
17030077-004	n-Decane	K, T, U	< 0.08	ppbv	0.08	AC-058	15-Mar-17
17030077-004	n-Dodecane	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030077-004	n-Heptane		0.44	ppbv	0.01	AC-058	15-Mar-17
17030077-004	n-Hexane		0.71	ppbv	0.01	AC-058	15-Mar-17
17030077-004	n-Octane		0.17	ppbv	0.03	AC-058	15-Mar-17
17030077-004	n-Pentane		1.7	ppbv	0.1	AC-058	15-Mar-17
17030077-004	n-Propylbenzene		0.07	ppbv	0.07	AC-058	15-Mar-17
17030077-004	n-Undecane	K, T, U	< 0.7	ppbv	0.7	AC-058	15-Mar-17
17030077-004	Naphthalene		1.7	ppbv	0.7	AC-058	15-Mar-17
17030077-004	n-Nonane		0.10	ppbv	0.01	AC-058	15-Mar-17
17030077-004	o-Ethyltoluene	I	0.10	ppbv	0.01	AC-058	15-Mar-17
17030077-004	o-Xylene		0.34	ppbv	0.01	AC-058	15-Mar-17
17030077-004	p-Diethylbenzene	I	0.10	ppbv	0.05	AC-058	15-Mar-17
17030077-004	p-Ethyltoluene	I	0.15	ppbv	0.09	AC-058	15-Mar-17
17030077-004	Styrene	I	0.08	ppbv	0.05	AC-058	15-Mar-17
17030077-004	Tetrachloroethylene	I	0.30	ppbv	0.05	AC-058	15-Mar-17
17030077-004	Tetrahydrofuran	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030077-004	Toluene		1.06	ppbv	0.01	AC-058	15-Mar-17
17030077-004	trans-1,2-Dichloroethylene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030077-004	trans-1,3-Dichloropropylene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17
17030077-004	trans-2-Butene		0.21	ppbv	0.01	AC-058	15-Mar-17
17030077-004	trans-2-Pentene		0.21	ppbv	0.03	AC-058	15-Mar-17
17030077-004	Trichloroethylene	K, T, U	< 0.05	ppbv	0.05	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca



PO Bag 4000
 Vegreville, Alberta
 Canada T9C 1T4
 (780) 632-8211

ENVIRONMENTAL ANALYTICAL SERVICES

TEST REPORT

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
\\NMHC VOC/Bonnyville/Mar 3, 2017	H3286	Ambient Air	03-Mar-17	19:25
DESCRIPTION:	Bonnyville - AER			
REPORT NUMBER:	17030077	REPORT CREATED:	13-Apr-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030077-004	Vinyl acetate	K, T, U	< 0.5 ppbv	0.5	AC-058	15-Mar-17
17030077-004	Vinyl chloride	K, T, U	< 0.03 ppbv	0.03	AC-058	15-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: April-13-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

RESULTS: Adewunmi Adekanmbi Lakeland Industry and Community Assn 4000, 19 St NE Calgary AB T2E 6P8	CLIENT SAMPLE ID CA/NMHC/Bonnyville/Mar 08, 20	CANISTER ID	Matrix Ambient Air	Priority Normal
	DESCRIPTION: Bonnyville - AER	DATE SAMPLED: 08-Mar-17 19:05	DATE RECEIVED: 13-Mar-17	REPORT NUMBER: 17030116
INVOICE: Arianna Cook 780 812 2182 PO Box 8237 5107W-50 St Bonnyville AB T9N 2J5	REPORT CREATED: 11-May-17	VERSION: Version 01		

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030116-003	1,1,1-Trichloroethane	K, T, U	< 0.02 ppbv	0.02	AC-058	15-Mar-17
17030116-003	1,1,2,2-Tetrachloroethane	K, T, U	< 0.02 ppbv	0.02	AC-058	15-Mar-17
17030116-003	1,1,2-Trichloroethane	K, T, U	< 0.02 ppbv	0.02	AC-058	15-Mar-17
17030116-003	1,1-Dichloroethane	K, T, U	< 0.02 ppbv	0.02	AC-058	15-Mar-17
17030116-003	1,1-Dichloroethylene	K, T, U	< 0.04 ppbv	0.04	AC-058	15-Mar-17
17030116-003	1,2,3-Trimethylbenzene	I	0.11 ppbv	0.05	AC-058	15-Mar-17
17030116-003	1,2,4-Trichlorobenzene	K, T, U	< 0.8 ppbv	0.8	AC-058	15-Mar-17
17030116-003	1,2,4-Trimethylbenzene		0.36 ppbv	0.05	AC-058	15-Mar-17
17030116-003	1,2-Dibromoethane	K, T, U	< 0.02 ppbv	0.02	AC-058	15-Mar-17
17030116-003	1,2-Dichlorobenzene	K, T, U	< 0.03 ppbv	0.03	AC-058	15-Mar-17
17030116-003	1,2-Dichloroethane	I	0.03 ppbv	0.01	AC-058	15-Mar-17
17030116-003	1,2-Dichloropropane	I	0.01 ppbv	0.01	AC-058	15-Mar-17
17030116-003	1,3,5-Trimethylbenzene		0.13 ppbv	0.02	AC-058	15-Mar-17
17030116-003	1,3-Butadiene	I	0.10 ppbv	0.02	AC-058	15-Mar-17
17030116-003	1,3-Dichlorobenzene	K, T, U	< 0.3 ppbv	0.3	AC-058	15-Mar-17
17030116-003	1,4-Dichlorobenzene	K, T, U	< 0.4 ppbv	0.4	AC-058	15-Mar-17
17030116-003	1,4-Dioxane	K, T, U	< 0.4 ppbv	0.4	AC-058	15-Mar-17
17030116-003	1-Butene		0.47 ppbv	0.02	AC-058	15-Mar-17

Report certified by: Krista Gegolick, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-11-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
CA/NMHC/Bonnyville/Mar 08, 201		Ambient Air	08-Mar-17 19:05
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030116	REPORT CREATED:	11-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030116-003	1-Hexene	I	0.09	ppbv	0.02	AC-058	15-Mar-17
17030116-003	1-Pentene		0.09	ppbv	0.01	AC-058	15-Mar-17
17030116-003	2,2,4-Trimethylpentane		0.75	ppbv	0.01	AC-058	15-Mar-17
17030116-003	2,2-Dimethylbutane		0.06	ppbv	0.01	AC-058	15-Mar-17
17030116-003	2,3,4-Trimethylpentane		0.31	ppbv	0.01	AC-058	15-Mar-17
17030116-003	2,3-Dimethylbutane		0.18	ppbv	0.02	AC-058	15-Mar-17
17030116-003	2,3-Dimethylpentane		0.58	ppbv	0.02	AC-058	15-Mar-17
17030116-003	2,4-Dimethylpentane		0.25	ppbv	0.01	AC-058	15-Mar-17
17030116-003	2-Methylheptane		0.21	ppbv	0.01	AC-058	15-Mar-17
17030116-003	2-Methylhexane		0.44	ppbv	0.01	AC-058	15-Mar-17
17030116-003	2-Methylpentane		0.56	ppbv	0.01	AC-058	15-Mar-17
17030116-003	3-Methylheptane		0.20	ppbv	0.02	AC-058	15-Mar-17
17030116-003	3-Methylhexane		0.52	ppbv	0.02	AC-058	15-Mar-17
17030116-003	3-Methylpentane		0.34	ppbv	0.01	AC-058	15-Mar-17
17030116-003	Acetone		1.9	ppbv	0.4	AC-058	15-Mar-17
17030116-003	Acrolein	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030116-003	Benzene		0.57	ppbv	0.01	AC-058	15-Mar-17
17030116-003	Benzyl chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-003	Bromodichloromethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-003	Bromoform	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-003	Bromomethane	I	0.01	ppbv	0.01	AC-058	15-Mar-17
17030116-003	Carbon disulfide	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030116-003	Carbon tetrachloride	I	0.12	ppbv	0.01	AC-058	15-Mar-17
17030116-003	Chlorobenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-003	Chloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	15-Mar-17

Report certified by:	Krista Gegolick, Account Coordinator	On behalf of:	PJ Pretorius, Manager, Analysis and Testing Services
Date:	May-11-17	Inquiries:	(780) 632 8455
		E-mail:	EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
CA/NMHC/Bonnyville/Mar 08, 201		Ambient Air	08-Mar-17 19:05
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030116	REPORT CREATED:	11-May-17
		VERSION:	Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030116-003	Chloroform	I	0.04	ppbv	0.02	AC-058	15-Mar-17
17030116-003	Chloromethane		0.72	ppbv	0.02	AC-058	15-Mar-17
17030116-003	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030116-003	cis-1,3-Dichloropropene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030116-003	cis-2-Butene		0.14	ppbv	0.02	AC-058	15-Mar-17
17030116-003	cis-2-Pentene		0.09	ppbv	0.02	AC-058	15-Mar-17
17030116-003	Cyclohexane		0.17	ppbv	0.02	AC-058	15-Mar-17
17030116-003	Cyclopentane		0.10	ppbv	0.01	AC-058	15-Mar-17
17030116-003	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	15-Mar-17
17030116-003	Ethanol		2.6	ppbv	0.3	AC-058	15-Mar-17
17030116-003	Ethyl acetate	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-003	Ethylbenzene		0.31	ppbv	0.01	AC-058	15-Mar-17
17030116-003	Freon-11	I	0.29	ppbv	0.02	AC-058	15-Mar-17
17030116-003	Freon-113	I	0.13	ppbv	0.01	AC-058	15-Mar-17
17030116-003	Freon-114	I	0.02	ppbv	0.02	AC-058	15-Mar-17
17030116-003	Freon-12		0.52	ppbv	0.02	AC-058	15-Mar-17
17030116-003	Hexachloro-1,3-butadiene	K, T, U	< 0.50	ppbv	0.50	AC-058	15-Mar-17
17030116-003	Isobutane		0.83	ppbv	0.02	AC-058	15-Mar-17
17030116-003	Isopentane		1.10	ppbv	0.03	AC-058	15-Mar-17
17030116-003	Isoprene		0.09	ppbv	0.01	AC-058	15-Mar-17
17030116-003	Isopropyl alcohol	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-003	Isopropylbenzene		0.04	ppbv	0.01	AC-058	15-Mar-17
17030116-003	m,p-Xylene		1.11	ppbv	0.03	AC-058	15-Mar-17
17030116-003	m-Diethylbenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17
17030116-003	m-Ethyltoluene	I	0.25	ppbv	0.08	AC-058	15-Mar-17

Report certified by: Krista Gegolick, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-11-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
CA/NMHC/Bonnyville/Mar 08, 201		Ambient Air	08-Mar-17 19:05
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030116	REPORT CREATED:	11-May-17
		VERSION:	Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030116-003	Methyl butyl ketone	K, T, U	< 0.50	ppbv	0.50	AC-058	15-Mar-17
17030116-003	Methyl ethyl ketone	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030116-003	Methyl isobutyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-003	Methyl methacrylate	K, T, U	< 0.07	ppbv	0.07	AC-058	15-Mar-17
17030116-003	Methyl tert butyl ether	K, T, U	< 0.03	ppbv	0.03	AC-058	15-Mar-17
17030116-003	Methylcyclohexane		0.30	ppbv	0.01	AC-058	15-Mar-17
17030116-003	Methylcyclopentane		0.41	ppbv	0.02	AC-058	15-Mar-17
17030116-003	Methylene chloride	K, T, U	< 0.3	ppbv	0.3	AC-058	15-Mar-17
17030116-003	n-Butane		2.18	ppbv	0.03	AC-058	15-Mar-17
17030116-003	n-Decane		0.09	ppbv	0.06	AC-058	15-Mar-17
17030116-003	n-Dodecane	K, T, U	< 0.4	ppbv	0.4	AC-058	15-Mar-17
17030116-003	n-Heptane		0.47	ppbv	0.01	AC-058	15-Mar-17
17030116-003	n-Hexane		0.37	ppbv	0.01	AC-058	15-Mar-17
17030116-003	n-Octane		0.22	ppbv	0.02	AC-058	15-Mar-17
17030116-003	n-Pentane		0.5	ppbv	0.1	AC-058	15-Mar-17
17030116-003	n-Propylbenzene		0.08	ppbv	0.05	AC-058	15-Mar-17
17030116-003	n-Undecane	K, T, U	< 0.5	ppbv	0.5	AC-058	15-Mar-17
17030116-003	Naphthalene		1.5	ppbv	0.5	AC-058	15-Mar-17
17030116-003	n-Nonane		0.10	ppbv	0.01	AC-058	15-Mar-17
17030116-003	o-Ethyltoluene	I	0.10	ppbv	0.01	AC-058	15-Mar-17
17030116-003	o-Xylene		0.40	ppbv	0.01	AC-058	15-Mar-17
17030116-003	p-Diethylbenzene	I	0.08	ppbv	0.04	AC-058	15-Mar-17
17030116-003	p-Ethyltoluene	I	0.12	ppbv	0.07	AC-058	15-Mar-17
17030116-003	Styrene	I	0.06	ppbv	0.04	AC-058	15-Mar-17
17030116-003	Tetrachloroethylene	K, T, U	< 0.04	ppbv	0.04	AC-058	15-Mar-17

Report certified by: Krista Gegolick, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-11-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
CA/NMHC/Bonnyville/Mar 08, 201		Ambient Air	08-Mar-17 19:05
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030116	REPORT CREATED:	11-May-17
		VERSION:	Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030116-003	Tetrahydrofuran	K, T, U	< 0.4 ppbv	0.4	AC-058	15-Mar-17
17030116-003	Toluene		1.29 ppbv	0.01	AC-058	15-Mar-17
17030116-003	trans-1,2-Dichloroethylene	K, T, U	< 0.01 ppbv	0.01	AC-058	15-Mar-17
17030116-003	trans-1,3-Dichloropropylene	K, T, U	< 0.04 ppbv	0.04	AC-058	15-Mar-17
17030116-003	trans-2-Butene		0.13 ppbv	0.01	AC-058	15-Mar-17
17030116-003	trans-2-Pentene		0.15 ppbv	0.02	AC-058	15-Mar-17
17030116-003	Trichloroethylene	K, T, U	< 0.04 ppbv	0.04	AC-058	15-Mar-17
17030116-003	Vinyl acetate	K, T, U	< 0.4 ppbv	0.4	AC-058	15-Mar-17
17030116-003	Vinyl chloride	K, T, U	< 0.02 ppbv	0.02	AC-058	15-Mar-17

Report certified by: Krista Gegolick, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-11-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

RESULTS: Adewunmi Adekanmbi Lakeland Industry and Community Assn 4000, 19 St NE Calgary AB T2E 6P8	CLIENT SAMPLE ID /NMHC VOC/Bonnyville/Mar 14, 2017	CANISTER ID 17121	Matrix Ambient Air	Priority Normal
	DESCRIPTION: Bonnyville - AER			
INVOICE: Arianna Cook PO Box 8237 5107W-50 St Bonnyville AB T9N 2J5	DATE SAMPLED: 11-Mar-17 16:35	DATE RECEIVED: 20-Mar-17	REPORT NUMBER: 17030212	
	REPORT CREATED: 02-May-17	VERSION: Version 01		
	780 812 2182			

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030212-003	1,1,1-Trichloroethane	K, T, U	< 0.02 ppbv	0.02	AC-058	23-Mar-17
17030212-003	1,1,2,2-Tetrachloroethane	K, T, U	< 0.02 ppbv	0.02	AC-058	23-Mar-17
17030212-003	1,1,2-Trichloroethane	K, T, U	< 0.02 ppbv	0.02	AC-058	23-Mar-17
17030212-003	1,1-Dichloroethane	K, T, U	< 0.02 ppbv	0.02	AC-058	23-Mar-17
17030212-003	1,1-Dichloroethylene	K, T, U	< 0.05 ppbv	0.05	AC-058	23-Mar-17
17030212-003	1,2,3-Trimethylbenzene	K, T, U	< 0.06 ppbv	0.06	AC-058	23-Mar-17
17030212-003	1,2,4-Trichlorobenzene	K, T, U	< 1.0 ppbv	1.0	AC-058	23-Mar-17
17030212-003	1,2,4-Trimethylbenzene	K, T, U	< 0.06 ppbv	0.06	AC-058	23-Mar-17
17030212-003	1,2-Dibromoethane	K, T, U	< 0.02 ppbv	0.02	AC-058	23-Mar-17
17030212-003	1,2-Dichlorobenzene	K, T, U	< 0.04 ppbv	0.04	AC-058	23-Mar-17
17030212-003	1,2-Dichloroethane	I	0.02 ppbv	0.01	AC-058	23-Mar-17
17030212-003	1,2-Dichloropropane	K, T, U	< 0.01 ppbv	0.01	AC-058	23-Mar-17
17030212-003	1,3,5-Trimethylbenzene		0.04 ppbv	0.02	AC-058	23-Mar-17
17030212-003	1,3-Butadiene	K, T, U	< 0.02 ppbv	0.02	AC-058	23-Mar-17
17030212-003	1,3-Dichlorobenzene	K, T, U	< 0.4 ppbv	0.4	AC-058	23-Mar-17
17030212-003	1,4-Dichlorobenzene	K, T, U	< 0.5 ppbv	0.5	AC-058	23-Mar-17
17030212-003	1,4-Dioxane	K, T, U	< 0.5 ppbv	0.5	AC-058	23-Mar-17
17030212-003	1-Butene		2.38 ppbv	0.02	AC-058	23-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
/NMHC VOC/Bonnyville/Mar 14, 2	17121	Ambient Air	11-Mar-17 16:35
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030212	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030212-003	1-Hexene	I	0.10	ppbv	0.02	AC-058	23-Mar-17
17030212-003	1-Pentene		0.29	ppbv	0.01	AC-058	23-Mar-17
17030212-003	2,2,4-Trimethylpentane		0.28	ppbv	0.01	AC-058	23-Mar-17
17030212-003	2,2-Dimethylbutane		0.03	ppbv	0.01	AC-058	23-Mar-17
17030212-003	2,3,4-Trimethylpentane		0.06	ppbv	0.01	AC-058	23-Mar-17
17030212-003	2,3-Dimethylbutane		0.33	ppbv	0.02	AC-058	23-Mar-17
17030212-003	2,3-Dimethylpentane		0.37	ppbv	0.02	AC-058	23-Mar-17
17030212-003	2,4-Dimethylpentane		0.20	ppbv	0.01	AC-058	23-Mar-17
17030212-003	2-Methylheptane		0.02	ppbv	0.01	AC-058	23-Mar-17
17030212-003	2-Methylhexane		0.12	ppbv	0.01	AC-058	23-Mar-17
17030212-003	2-Methylpentane		0.73	ppbv	0.01	AC-058	23-Mar-17
17030212-003	3-Methylheptane	K, T, U	< 0.02	ppbv	0.02	AC-058	23-Mar-17
17030212-003	3-Methylhexane		0.13	ppbv	0.02	AC-058	23-Mar-17
17030212-003	3-Methylpentane		0.43	ppbv	0.01	AC-058	23-Mar-17
17030212-003	Acetone		2.9	ppbv	0.5	AC-058	23-Mar-17
17030212-003	Acrolein	K, T, U	< 0.4	ppbv	0.4	AC-058	23-Mar-17
17030212-003	Benzene		1.45	ppbv	0.01	AC-058	23-Mar-17
17030212-003	Benzyl chloride	K, T, U	< 0.5	ppbv	0.5	AC-058	23-Mar-17
17030212-003	Bromodichloromethane	K, T, U	< 0.02	ppbv	0.02	AC-058	23-Mar-17
17030212-003	Bromoform	K, T, U	< 0.02	ppbv	0.02	AC-058	23-Mar-17
17030212-003	Bromomethane	K, T, U	< 0.01	ppbv	0.01	AC-058	23-Mar-17
17030212-003	Carbon disulfide	K, T, U	< 0.01	ppbv	0.01	AC-058	23-Mar-17
17030212-003	Carbon tetrachloride	I	0.10	ppbv	0.01	AC-058	23-Mar-17
17030212-003	Chlorobenzene	K, T, U	< 0.02	ppbv	0.02	AC-058	23-Mar-17
17030212-003	Chloroethane	K, T, U	< 0.02	ppbv	0.02	AC-058	23-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
/NMHC VOC/Bonnyville/Mar 14, 2	17121	Ambient Air	11-Mar-17 16:35
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030212	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030212-003	Chloroform	K, T, U	< 0.02	ppbv	0.02	AC-058	23-Mar-17
17030212-003	Chloromethane		0.65	ppbv	0.02	AC-058	23-Mar-17
17030212-003	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	23-Mar-17
17030212-003	cis-1,3-Dichloropropene	K, T, U	< 0.05	ppbv	0.05	AC-058	23-Mar-17
17030212-003	cis-2-Butene		1.64	ppbv	0.02	AC-058	23-Mar-17
17030212-003	cis-2-Pentene		0.30	ppbv	0.02	AC-058	23-Mar-17
17030212-003	Cyclohexane		0.05	ppbv	0.02	AC-058	23-Mar-17
17030212-003	Cyclopentane		0.09	ppbv	0.01	AC-058	23-Mar-17
17030212-003	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	23-Mar-17
17030212-003	Ethanol		5.7	ppbv	0.4	AC-058	23-Mar-17
17030212-003	Ethyl acetate	K, T, U	< 0.5	ppbv	0.5	AC-058	23-Mar-17
17030212-003	Ethylbenzene		0.04	ppbv	0.01	AC-058	23-Mar-17
17030212-003	Freon-11	I	0.29	ppbv	0.02	AC-058	23-Mar-17
17030212-003	Freon-113	I	0.09	ppbv	0.01	AC-058	23-Mar-17
17030212-003	Freon-114	K, T, U	< 0.02	ppbv	0.02	AC-058	23-Mar-17
17030212-003	Freon-12		0.63	ppbv	0.02	AC-058	23-Mar-17
17030212-003	Hexachloro-1,3-butadiene	K, T, U	< 0.62	ppbv	0.62	AC-058	23-Mar-17
17030212-003	Isobutane		18.9	ppbv	0.02	AC-058	23-Mar-17
17030212-003	Isopentane		9.79	ppbv	0.04	AC-058	23-Mar-17
17030212-003	Isoprene		0.01	ppbv	0.01	AC-058	23-Mar-17
17030212-003	Isopropyl alcohol	K, T, U	< 0.5	ppbv	0.5	AC-058	23-Mar-17
17030212-003	Isopropylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	23-Mar-17
17030212-003	m,p-Xylene		0.12	ppbv	0.04	AC-058	23-Mar-17
17030212-003	m-Diethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	23-Mar-17
17030212-003	m-Ethyltoluene	K, T, U	< 0.10	ppbv	0.10	AC-058	23-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
/NMHC VOC/Bonnyville/Mar 14, 2	17121	Ambient Air	11-Mar-17 16:35
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030212	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030212-003	Methyl butyl ketone	K, T, U	< 0.62	ppbv	0.62	AC-058	23-Mar-17
17030212-003	Methyl ethyl ketone		0.4	ppbv	0.4	AC-058	23-Mar-17
17030212-003	Methyl isobutyl ketone	K, T, U	< 0.5	ppbv	0.5	AC-058	23-Mar-17
17030212-003	Methyl methacrylate	K, T, U	< 0.09	ppbv	0.09	AC-058	23-Mar-17
17030212-003	Methyl tert butyl ether	K, T, U	< 0.04	ppbv	0.04	AC-058	23-Mar-17
17030212-003	Methylcyclohexane		0.06	ppbv	0.01	AC-058	23-Mar-17
17030212-003	Methylcyclopentane		0.32	ppbv	0.02	AC-058	23-Mar-17
17030212-003	Methylene chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	23-Mar-17
17030212-003	n-Butane		43.8	ppbv	0.11	AC-058	23-Mar-17
17030212-003	n-Decane	K, T, U	< 0.08	ppbv	0.08	AC-058	23-Mar-17
17030212-003	n-Dodecane	K, T, U	< 0.5	ppbv	0.5	AC-058	23-Mar-17
17030212-003	n-Heptane		0.09	ppbv	0.01	AC-058	23-Mar-17
17030212-003	n-Hexane		0.22	ppbv	0.01	AC-058	23-Mar-17
17030212-003	n-Octane		0.03	ppbv	0.02	AC-058	23-Mar-17
17030212-003	n-Pentane		1.5	ppbv	0.1	AC-058	23-Mar-17
17030212-003	n-Propylbenzene	K, T, U	< 0.06	ppbv	0.06	AC-058	23-Mar-17
17030212-003	n-Undecane	K, T, U	< 0.6	ppbv	0.6	AC-058	23-Mar-17
17030212-003	Naphthalene	K, T, U	< 0.6	ppbv	0.6	AC-058	23-Mar-17
17030212-003	n-Nonane		0.02	ppbv	0.01	AC-058	23-Mar-17
17030212-003	o-Ethyltoluene	I	0.02	ppbv	0.01	AC-058	23-Mar-17
17030212-003	o-Xylene		0.05	ppbv	0.01	AC-058	23-Mar-17
17030212-003	p-Diethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	23-Mar-17
17030212-003	p-Ethyltoluene	K, T, U	< 0.09	ppbv	0.09	AC-058	23-Mar-17
17030212-003	Styrene	K, T, U	< 0.05	ppbv	0.05	AC-058	23-Mar-17
17030212-003	Tetrachloroethylene	K, T, U	< 0.05	ppbv	0.05	AC-058	23-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
/NMHC VOC/Bonnyville/Mar 14, 2	17121	Ambient Air	11-Mar-17 16:35
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17030212	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030212-003	Tetrahydrofuran	K, T, U	< 0.5	ppbv	0.5	AC-058	23-Mar-17
17030212-003	Toluene		0.38	ppbv	0.01	AC-058	23-Mar-17
17030212-003	trans-1,2-Dichloroethylene	K, T, U	< 0.01	ppbv	0.01	AC-058	23-Mar-17
17030212-003	trans-1,3-Dichloropropylene	K, T, U	< 0.05	ppbv	0.05	AC-058	23-Mar-17
17030212-003	trans-2-Butene		1.96	ppbv	0.01	AC-058	23-Mar-17
17030212-003	trans-2-Pentene		0.54	ppbv	0.02	AC-058	23-Mar-17
17030212-003	Trichloroethylene	K, T, U	< 0.05	ppbv	0.05	AC-058	23-Mar-17
17030212-003	Vinyl acetate		0.8	ppbv	0.5	AC-058	23-Mar-17
17030212-003	Vinyl chloride	K, T, U	< 0.02	ppbv	0.02	AC-058	23-Mar-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

RESULTS: Adewunmi Adekanmbi Lakeland Industry and Community Assn 4000, 19 St NE Calgary AB T2E 6P8	CLIENT SAMPLE ID /NMHC VOC/Bonnyville/Mar 19, :	CANISTER ID 1684	Matrix Ambient Air	Priority Normal
	DESCRIPTION: Bonnyville- AER			
INVOICE: Arianna Cook PO Box 8237 5107W-50 St Bonnyville AB T9N 2J5	DATE SAMPLED: 19-Mar-17	6:55	DATE RECEIVED: 30-Mar-17	
	REPORT CREATED: 02-May-17		REPORT NUMBER: 17030303	
			VERSION: Version 01	

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030303-005	1,1,1-Trichloroethane	K, T, U	< 0.03 ppbv	0.03	AC-058	04-Apr-17
17030303-005	1,1,2,2-Tetrachloroethane	K, T, U	< 0.03 ppbv	0.03	AC-058	04-Apr-17
17030303-005	1,1,2-Trichloroethane	K, T, U	< 0.03 ppbv	0.03	AC-058	04-Apr-17
17030303-005	1,1-Dichloroethane	K, T, U	< 0.03 ppbv	0.03	AC-058	04-Apr-17
17030303-005	1,1-Dichloroethylene	K, T, U	< 0.06 ppbv	0.06	AC-058	04-Apr-17
17030303-005	1,2,3-Trimethylbenzene	K, T, U	< 0.07 ppbv	0.07	AC-058	04-Apr-17
17030303-005	1,2,4-Trichlorobenzene	K, T, U	< 1.1 ppbv	1.1	AC-058	04-Apr-17
17030303-005	1,2,4-Trimethylbenzene	K, T, U	< 0.07 ppbv	0.07	AC-058	04-Apr-17
17030303-005	1,2-Dibromoethane	K, T, U	< 0.03 ppbv	0.03	AC-058	04-Apr-17
17030303-005	1,2-Dichlorobenzene	K, T, U	< 0.04 ppbv	0.04	AC-058	04-Apr-17
17030303-005	1,2-Dichloroethane	K, T, U	< 0.01 ppbv	0.01	AC-058	04-Apr-17
17030303-005	1,2-Dichloropropane	K, T, U	< 0.01 ppbv	0.01	AC-058	04-Apr-17
17030303-005	1,3,5-Trimethylbenzene	K, T, U	< 0.03 ppbv	0.03	AC-058	04-Apr-17
17030303-005	1,3-Butadiene	K, T, U	< 0.03 ppbv	0.03	AC-058	04-Apr-17
17030303-005	1,3-Dichlorobenzene	K, T, U	< 0.4 ppbv	0.4	AC-058	04-Apr-17
17030303-005	1,4-Dichlorobenzene	K, T, U	< 0.6 ppbv	0.6	AC-058	04-Apr-17
17030303-005	1,4-Dioxane	K, T, U	< 0.6 ppbv	0.6	AC-058	04-Apr-17
17030303-005	1-Butene		0.10 ppbv	0.03	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
/NMHC VOC/Bonnyville/Mar 19, 2	1684	Ambient Air	19-Mar-17 6:55
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030303-005	1-Hexene	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	1-Pentene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	2,2,4-Trimethylpentane		0.02	ppbv	0.01	AC-058	04-Apr-17
17030303-005	2,2-Dimethylbutane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	2,3,4-Trimethylpentane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	2,3-Dimethylbutane	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	2,3-Dimethylpentane	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	2,4-Dimethylpentane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	2-Methylheptane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	2-Methylhexane		0.03	ppbv	0.01	AC-058	04-Apr-17
17030303-005	2-Methylpentane		0.08	ppbv	0.01	AC-058	04-Apr-17
17030303-005	3-Methylheptane	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	3-Methylhexane		0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	3-Methylpentane		0.03	ppbv	0.01	AC-058	04-Apr-17
17030303-005	Acetone		1.6	ppbv	0.6	AC-058	04-Apr-17
17030303-005	Acrolein	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-005	Benzene		0.13	ppbv	0.01	AC-058	04-Apr-17
17030303-005	Benzyl chloride	K, T, U	< 0.6	ppbv	0.6	AC-058	04-Apr-17
17030303-005	Bromodichloromethane	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	Bromoform	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	Bromomethane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	Carbon disulfide	I	0.04	ppbv	0.01	AC-058	04-Apr-17
17030303-005	Carbon tetrachloride	I	0.07	ppbv	0.01	AC-058	04-Apr-17
17030303-005	Chlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	Chloroethane	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
/NMHC VOC/Bonnyville/Mar 19, 2	1684	Ambient Air	19-Mar-17 6:55
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030303-005	Chloroform	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	Chloromethane		0.60	ppbv	0.03	AC-058	04-Apr-17
17030303-005	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	cis-1,3-Dichloropropene	K, T, U	< 0.06	ppbv	0.06	AC-058	04-Apr-17
17030303-005	cis-2-Butene	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	cis-2-Pentene	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	Cyclohexane	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	Cyclopentane		0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	Ethanol		0.9	ppbv	0.4	AC-058	04-Apr-17
17030303-005	Ethyl acetate	K, T, U	< 0.6	ppbv	0.6	AC-058	04-Apr-17
17030303-005	Ethylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	Freon-11	I	0.22	ppbv	0.03	AC-058	04-Apr-17
17030303-005	Freon-113	I	0.09	ppbv	0.01	AC-058	04-Apr-17
17030303-005	Freon-114	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	Freon-12	I	0.37	ppbv	0.03	AC-058	04-Apr-17
17030303-005	Hexachloro-1,3-butadiene	K, T, U	< 0.70	ppbv	0.70	AC-058	04-Apr-17
17030303-005	Isobutane		0.56	ppbv	0.03	AC-058	04-Apr-17
17030303-005	Isopentane		0.63	ppbv	0.04	AC-058	04-Apr-17
17030303-005	Isoprene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	Isopropyl alcohol	K, T, U	< 0.6	ppbv	0.6	AC-058	04-Apr-17
17030303-005	Isopropylbenzene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	m,p-Xylene	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030303-005	m-Diethylbenzene	K, T, U	< 0.06	ppbv	0.06	AC-058	04-Apr-17
17030303-005	m-Ethyltoluene	K, T, U	< 0.11	ppbv	0.11	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
/NMHC VOC/Bonnyville/Mar 19, 2	1684	Ambient Air	19-Mar-17 6:55
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17030303-005	Methyl butyl ketone	K, T, U	< 0.70	ppbv	0.70	AC-058	04-Apr-17
17030303-005	Methyl ethyl ketone	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-005	Methyl isobutyl ketone	K, T, U	< 0.6	ppbv	0.6	AC-058	04-Apr-17
17030303-005	Methyl methacrylate	K, T, U	< 0.10	ppbv	0.10	AC-058	04-Apr-17
17030303-005	Methyl tert butyl ether	K, T, U	< 0.04	ppbv	0.04	AC-058	04-Apr-17
17030303-005	Methylcyclohexane		0.03	ppbv	0.01	AC-058	04-Apr-17
17030303-005	Methylcyclopentane	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	Methylene chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	04-Apr-17
17030303-005	n-Butane		4.44	ppbv	0.04	AC-058	04-Apr-17
17030303-005	n-Decane	K, T, U	< 0.08	ppbv	0.08	AC-058	04-Apr-17
17030303-005	n-Dodecane	K, T, U	< 0.6	ppbv	0.6	AC-058	04-Apr-17
17030303-005	n-Heptane		0.03	ppbv	0.01	AC-058	04-Apr-17
17030303-005	n-Hexane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	n-Octane	K, T, U	< 0.03	ppbv	0.03	AC-058	04-Apr-17
17030303-005	n-Pentane		0.2	ppbv	0.1	AC-058	04-Apr-17
17030303-005	n-Propylbenzene	K, T, U	< 0.07	ppbv	0.07	AC-058	04-Apr-17
17030303-005	n-Undecane	K, T, U	< 0.7	ppbv	0.7	AC-058	04-Apr-17
17030303-005	Naphthalene	K, T, U	< 0.7	ppbv	0.7	AC-058	04-Apr-17
17030303-005	n-Nonane	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	o-Ethyltoluene	K, T, U	< 0.01	ppbv	0.01	AC-058	04-Apr-17
17030303-005	o-Xylene		0.03	ppbv	0.01	AC-058	04-Apr-17
17030303-005	p-Diethylbenzene	K, T, U	< 0.06	ppbv	0.06	AC-058	04-Apr-17
17030303-005	p-Ethyltoluene	K, T, U	< 0.10	ppbv	0.10	AC-058	04-Apr-17
17030303-005	Styrene	K, T, U	< 0.06	ppbv	0.06	AC-058	04-Apr-17
17030303-005	Tetrachloroethylene	K, T, U	< 0.06	ppbv	0.06	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
/NMHC VOC/Bonnyville/Mar 19, 2	1684	Ambient Air	19-Mar-17 6:55
DESCRIPTION:	Bonnyville- AER		
REPORT NUMBER:	17030303	REPORT CREATED:	02-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result Units	RDL	Method	Analysis Date
17030303-005	Tetrahydrofuran	K, T, U	< 0.6 ppbv	0.6	AC-058	04-Apr-17
17030303-005	Toluene		0.05 ppbv	0.01	AC-058	04-Apr-17
17030303-005	trans-1,2-Dichloroethylene	K, T, U	< 0.01 ppbv	0.01	AC-058	04-Apr-17
17030303-005	trans-1,3-Dichloropropylene	K, T, U	< 0.06 ppbv	0.06	AC-058	04-Apr-17
17030303-005	trans-2-Butene		0.04 ppbv	0.01	AC-058	04-Apr-17
17030303-005	trans-2-Pentene	K, T, U	< 0.03 ppbv	0.03	AC-058	04-Apr-17
17030303-005	Trichloroethylene	K, T, U	< 0.06 ppbv	0.06	AC-058	04-Apr-17
17030303-005	Vinyl acetate	K, T, U	< 0.6 ppbv	0.6	AC-058	04-Apr-17
17030303-005	Vinyl chloride	K, T, U	< 0.03 ppbv	0.03	AC-058	04-Apr-17

Report certified by: Colleen McGerrigle, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: May-02-17

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

<p>RESULTS: Adewunmi Adekanmbi Lakeland Industry and Community Assn 4000, 19 St NE</p> <p>Calgary AB T2E 6P8</p> <p>INVOICE: Arianna Cook 780 812 2182 PO Box 8237 5107W-50 St Bonnyville AB T9N 2J5</p>	<p>CLIENT SAMPLE ID A/NMHC/Bonnyville/March 28, 2017</p> <p>CANISTER ID 1517</p> <p>Matrix Ambient Air</p> <p>Priority Normal</p> <p>DESCRIPTION: Bonnyville - AER</p> <p>DATE SAMPLED: 28-Mar-17 18:10</p> <p>REPORT CREATED: 03-May-17</p> <p>DATE RECEIVED: 06-Apr-17</p> <p>REPORT NUMBER: 17040067</p> <p>VERSION: Version 01</p>
--	--

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17040067-003	1,1,1-Trichloroethane	K, T, U	< 0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	1,1,2,2-Tetrachloroethane	K, T, U	< 0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	1,1,2-Trichloroethane	K, T, U	< 0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	1,1-Dichloroethane	K, T, U	< 0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	1,1-Dichloroethylene	K, T, U	< 0.05	ppbv	0.05	AC-058	14-Apr-17
17040067-003	1,2,3-Trimethylbenzene	I	0.09	ppbv	0.07	AC-058	14-Apr-17
17040067-003	1,2,4-Trichlorobenzene	K, T, U	< 1.1	ppbv	1.1	AC-058	14-Apr-17
17040067-003	1,2,4-Trimethylbenzene		0.24	ppbv	0.07	AC-058	14-Apr-17
17040067-003	1,2-Dibromoethane	K, T, U	< 0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	1,2-Dichlorobenzene	K, T, U	< 0.04	ppbv	0.04	AC-058	14-Apr-17
17040067-003	1,2-Dichloroethane	I	0.03	ppbv	0.01	AC-058	14-Apr-17
17040067-003	1,2-Dichloropropane	K, T, U	< 0.01	ppbv	0.01	AC-058	14-Apr-17
17040067-003	1,3,5-Trimethylbenzene		0.11	ppbv	0.03	AC-058	14-Apr-17
17040067-003	1,3-Butadiene	I	0.04	ppbv	0.03	AC-058	14-Apr-17
17040067-003	1,3-Dichlorobenzene	K, T, U	< 0.4	ppbv	0.4	AC-058	14-Apr-17
17040067-003	1,4-Dichlorobenzene	K, T, U	< 0.5	ppbv	0.5	AC-058	14-Apr-17
17040067-003	1,4-Dioxane	K, T, U	< 0.5	ppbv	0.5	AC-058	14-Apr-17
17040067-003	1-Butene		0.48	ppbv	0.03	AC-058	14-Apr-17

Report certified by: Krista Gegoick, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Wednesday, May 03, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
A/NMHC/Bonnyville/March 28, 2017	1517	Ambient Air	28-Mar-17	18:10
DESCRIPTION:	Bonnyville - AER			
REPORT NUMBER:	17040067	REPORT CREATED:	03-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17040067-003	1-Hexene	I	0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	1-Pentene		0.03	ppbv	0.01	AC-058	14-Apr-17
17040067-003	2,2,4-Trimethylpentane		0.07	ppbv	0.01	AC-058	14-Apr-17
17040067-003	2,2-Dimethylbutane		0.04	ppbv	0.01	AC-058	14-Apr-17
17040067-003	2,3,4-Trimethylpentane		0.05	ppbv	0.01	AC-058	14-Apr-17
17040067-003	2,3-Dimethylbutane		0.13	ppbv	0.03	AC-058	14-Apr-17
17040067-003	2,3-Dimethylpentane		0.15	ppbv	0.03	AC-058	14-Apr-17
17040067-003	2,4-Dimethylpentane		0.08	ppbv	0.01	AC-058	14-Apr-17
17040067-003	2-Methylheptane		0.23	ppbv	0.01	AC-058	14-Apr-17
17040067-003	2-Methylhexane		0.26	ppbv	0.01	AC-058	14-Apr-17
17040067-003	2-Methylpentane		0.71	ppbv	0.01	AC-058	14-Apr-17
17040067-003	3-Methylheptane		0.18	ppbv	0.03	AC-058	14-Apr-17
17040067-003	3-Methylhexane		0.32	ppbv	0.03	AC-058	14-Apr-17
17040067-003	3-Methylpentane		0.40	ppbv	0.01	AC-058	14-Apr-17
17040067-003	Acetone		6.3	ppbv	0.5	AC-058	14-Apr-17
17040067-003	Acrolein	K, T, U	< 0.4	ppbv	0.4	AC-058	14-Apr-17
17040067-003	Benzene		0.53	ppbv	0.01	AC-058	14-Apr-17
17040067-003	Benzyl chloride	K, T, U	< 0.5	ppbv	0.5	AC-058	14-Apr-17
17040067-003	Bromodichloromethane	K, T, U	< 0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	Bromoform	K, T, U	< 0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	Bromomethane	I	0.02	ppbv	0.01	AC-058	14-Apr-17
17040067-003	Carbon disulfide	I	0.26	ppbv	0.01	AC-058	14-Apr-17
17040067-003	Carbon tetrachloride	I	0.14	ppbv	0.01	AC-058	14-Apr-17
17040067-003	Chlorobenzene	K, T, U	< 0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	Chloroethane	I	0.04	ppbv	0.03	AC-058	14-Apr-17

Report certified by: Krista Gegoick, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Wednesday, May 03, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
A/NMHC/Bonnyville/March 28, 2C	1517	Ambient Air	28-Mar-17 18:10
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17040067	REPORT CREATED:	03-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17040067-003	Chloroform	I	0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	Chloromethane		0.68	ppbv	0.03	AC-058	14-Apr-17
17040067-003	cis-1,2-Dichloroethene	K, T, U	< 0.01	ppbv	0.01	AC-058	14-Apr-17
17040067-003	cis-1,3-Dichloropropene	K, T, U	< 0.05	ppbv	0.05	AC-058	14-Apr-17
17040067-003	cis-2-Butene		0.05	ppbv	0.03	AC-058	14-Apr-17
17040067-003	cis-2-Pentene	K, T, U	< 0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	Cyclohexane		0.70	ppbv	0.03	AC-058	14-Apr-17
17040067-003	Cyclopentane		0.11	ppbv	0.01	AC-058	14-Apr-17
17040067-003	Dibromochloromethane	K, T, U	< 0.01	ppbv	0.01	AC-058	14-Apr-17
17040067-003	Ethanol		2.0	ppbv	0.4	AC-058	14-Apr-17
17040067-003	Ethyl acetate	K, T, U	< 0.5	ppbv	0.5	AC-058	14-Apr-17
17040067-003	Ethylbenzene		0.12	ppbv	0.01	AC-058	14-Apr-17
17040067-003	Freon-11	I	0.31	ppbv	0.03	AC-058	14-Apr-17
17040067-003	Freon-113	I	0.14	ppbv	0.01	AC-058	14-Apr-17
17040067-003	Freon-114	K, T, U	< 0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	Freon-12		0.51	ppbv	0.03	AC-058	14-Apr-17
17040067-003	Hexachloro-1,3-butadiene	K, T, U	< 0.68	ppbv	0.68	AC-058	14-Apr-17
17040067-003	Isobutane		1.06	ppbv	0.03	AC-058	14-Apr-17
17040067-003	Isopentane		1.34	ppbv	0.04	AC-058	14-Apr-17
17040067-003	Isoprene		0.02	ppbv	0.01	AC-058	14-Apr-17
17040067-003	Isopropyl alcohol	K, T, U	< 0.5	ppbv	0.5	AC-058	14-Apr-17
17040067-003	Isopropylbenzene		0.03	ppbv	0.01	AC-058	14-Apr-17
17040067-003	m,p-Xylene		0.70	ppbv	0.04	AC-058	14-Apr-17
17040067-003	m-Diethylbenzene	K, T, U	< 0.05	ppbv	0.05	AC-058	14-Apr-17
17040067-003	m-Ethyltoluene	K, T, U	< 0.11	ppbv	0.11	AC-058	14-Apr-17

Report certified by: Krista Gegoick, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Wednesday, May 03, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED
A/NMHC/Bonnyville/March 28, 2017	1517	Ambient Air	28-Mar-17 18:10
DESCRIPTION:	Bonnyville - AER		
REPORT NUMBER:	17040067	REPORT CREATED:	03-May-17
			VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17040067-003	Methyl butyl ketone	K, T, U	< 0.68	ppbv	0.68	AC-058	14-Apr-17
17040067-003	Methyl ethyl ketone		0.4	ppbv	0.4	AC-058	14-Apr-17
17040067-003	Methyl isobutyl ketone	K, T, U	< 0.5	ppbv	0.5	AC-058	14-Apr-17
17040067-003	Methyl methacrylate	K, T, U	< 0.09	ppbv	0.09	AC-058	14-Apr-17
17040067-003	Methyl tert butyl ether	K, T, U	< 0.04	ppbv	0.04	AC-058	14-Apr-17
17040067-003	Methylcyclohexane		1.17	ppbv	0.01	AC-058	14-Apr-17
17040067-003	Methylcyclopentane		0.46	ppbv	0.03	AC-058	14-Apr-17
17040067-003	Methylene chloride	K, T, U	< 0.4	ppbv	0.4	AC-058	14-Apr-17
17040067-003	n-Butane		2.73	ppbv	0.04	AC-058	14-Apr-17
17040067-003	n-Decane		0.19	ppbv	0.08	AC-058	14-Apr-17
17040067-003	n-Dodecane	K, T, U	< 0.5	ppbv	0.5	AC-058	14-Apr-17
17040067-003	n-Heptane		0.40	ppbv	0.01	AC-058	14-Apr-17
17040067-003	n-Hexane		0.58	ppbv	0.01	AC-058	14-Apr-17
17040067-003	n-Octane		0.33	ppbv	0.03	AC-058	14-Apr-17
17040067-003	n-Pentane		1.2	ppbv	0.1	AC-058	14-Apr-17
17040067-003	n-Propylbenzene	K, T, U	< 0.07	ppbv	0.07	AC-058	14-Apr-17
17040067-003	n-Undecane	K, T, U	< 0.7	ppbv	0.7	AC-058	14-Apr-17
17040067-003	Naphthalene	K, T, U	< 0.7	ppbv	0.7	AC-058	14-Apr-17
17040067-003	n-Nonane		0.30	ppbv	0.01	AC-058	14-Apr-17
17040067-003	o-Ethyltoluene	I	0.05	ppbv	0.01	AC-058	14-Apr-17
17040067-003	o-Xylene		0.24	ppbv	0.01	AC-058	14-Apr-17
17040067-003	p-Diethylbenzene	I	0.06	ppbv	0.05	AC-058	14-Apr-17
17040067-003	p-Ethyltoluene	K, T, U	< 0.09	ppbv	0.09	AC-058	14-Apr-17
17040067-003	Styrene	K, T, U	< 0.05	ppbv	0.05	AC-058	14-Apr-17
17040067-003	Tetrachloroethylene	I	0.06	ppbv	0.05	AC-058	14-Apr-17

Report certified by: Krista Gegolick, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Wednesday, May 03, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

CLIENT SAMPLE ID	CANISTER ID	Matrix	DATE SAMPLED	
A/NMHC/Bonnyville/March 28, 2C	1517	Ambient Air	28-Mar-17	18:10
DESCRIPTION:	Bonnyville - AER			
REPORT NUMBER:	17040067	REPORT CREATED:	03-May-17	VERSION: Version 01

Lab ID	Parameter	Qualifier	Result	Units	RDL	Method	Analysis Date
17040067-003	Tetrahydrofuran	K, T, U	< 0.5	ppbv	0.5	AC-058	14-Apr-17
17040067-003	Toluene		0.98	ppbv	0.01	AC-058	14-Apr-17
17040067-003	trans-1,2-Dichloroethylene	K, T, U	< 0.01	ppbv	0.01	AC-058	14-Apr-17
17040067-003	trans-1,3-Dichloropropylene	K, T, U	< 0.05	ppbv	0.05	AC-058	14-Apr-17
17040067-003	trans-2-Butene		0.04	ppbv	0.01	AC-058	14-Apr-17
17040067-003	trans-2-Pentene		0.03	ppbv	0.03	AC-058	14-Apr-17
17040067-003	Trichloroethylene	I	0.07	ppbv	0.05	AC-058	14-Apr-17
17040067-003	Vinyl acetate	K, T, U	< 0.5	ppbv	0.5	AC-058	14-Apr-17
17040067-003	Vinyl chloride	K, T, U	< 0.03	ppbv	0.03	AC-058	14-Apr-17

Report certified by: Krista Gegolick, Account Coordinator

On behalf of: PJ Pretorius, Manager, Analysis and Testing Services

Date: Wednesday, May 03, 2017

Inquiries: (780) 632 8455

E-mail: EAS.Results@innotechalberta.ca

APPENDIX V
REPORT CERTIFICATION FORM

Report Certification Form

Alberta Airshed (if applicable)	EPA Approval or Code of Practice Registration # (if applicable)
YES	NA
Company Name (if applicable)	Industrial Operation Name (if applicable)
Lakeland Industry & Community Association	Bonnyville Continuous Monitoring Station
Name of the Representative of the Person Responsible (Last, First, Middle)	Position / Title of the Representative of the Person Responsible
Maram Ghaleb	Project Manager, Customer Service, Air Services
Is an External Party Certifying the Report? (If 'Yes', fill in the fields below for the external person.)	
<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	
Name of External Person Certifying the Report (Last, First, Middle)	Position / Title of External Person Certifying the Report
NA	NA
Company Name for the External Person Certifying the Report	Identification of Qualifications / Professional Designations of the External Person Certifying the Report
NA	NA

I certify that I have reviewed and verified the submitted report. I also certify that the report presented with this certification form is complete, accurate and representative of the monitoring results and timeframe.

Maram Ghaleb

Signature of the Representative of the Person Responsible / External Person Certifying the Report

May 12, 2017

Report Issued Date (dd-mm-yyyy)

APPENDIX VI
DATA VALIDATION CERTIFICATION FORM



Validation Certificate Form

Client: <u>Lakeland Industry & Community Association</u> Site: <u>Bonnyville Continuous Monitoring Station</u>	Project #: <u>2833-2017-03-35-C</u> Contact: <u>Mike Bisaga</u>
---	--

Level 0 Preliminary Verification	 <hr style="width: 80%; margin: 0 auto;"/>	Date <u>April 30, 2017</u>
Level 1 Primary Validation	 <hr style="width: 80%; margin: 0 auto;"/>	Date <u>May 5, 2017</u>
Level 2 Final Validation	 <hr style="width: 80%; margin: 0 auto;"/>	Date <u>May 9, 2017</u>
Level 3 Independent Data Review	 <hr style="width: 80%; margin: 0 auto;"/>	Date <u>May 12, 2017</u>
Post-Final Validation	<u>NA</u>	Date <u>NA</u>

Notes

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. This validation is performed on an annual basis.
