

Alberta Environment and Parks (AEP)
Air.Reporting@gov.ab.ca

December 2, 2018

Subject: Monthly Report Submission for the LICA Cold Lake South station

Lakeland Industry & Community Association (LICA) is pleased to submit the ambient air monitoring monthly report for the LICA Cold Lake South AQM Station in the month of October 2018.

The air monitoring program consists of continuous air monitoring, passive sampling, intermittent sampling, including both VOC and PAH sampling program, and Partisol sampling program. All the air monitoring activities were conducted by contractors.

Sampling Program	Monitoring Activities Conducted By	Sample Analysis Conducted By	Data/Report Review and Prepared By	Electronic Submission Conducted By
Continuous ambient air	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics
Passive	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics
Intermittent	Maxxam Analytics	InnoTech Alberta Inc	InnoTech Alberta Inc	Not Applicable
Partisol	Maxxam Analytics	InnoTech Alberta Inc	InnoTech Alberta Inc	Not Applicable

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

All data collected in October 2018 was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

As the LICA Environmental Program Manager and Data & Reporting Specialist, we have reviewed and verified this report and that the information is complete, accurate and representative of the monitoring results, reporting timeframe and the specified analysis, summarization and reporting requirements. We also verify all air data that are required by the AMD to be electronically submitted to AEP and Alberta’s Ambient Air Quality Data Warehouse have been submitted by the time of this report submission, with the exception of electronic submission for the results of intermittent samples and Partisol samples. We are currently working with the airdata warehouse to set up codes for some VOC/PAH species that are missing in the parameter list. The results for these data will be submitted once all needed codes are available.

Should you have any questions, please don’t hesitate to contact us.

Respectfully,



Lakeland Industry & Community Association
5107 50 St
Bonnyville, AB T9N 2J7

A handwritten signature in blue ink that reads "Michael Bisaga".

Michael Bisaga
Technical Program Manager
Lakeland Industry & Community Association
780-266-7068
monitoring@lica.ca

A handwritten signature in blue ink that reads "Lily Lin".

Lily Lin
Data & Reporting Specialist
587-225-2248
monitoring@lica.ca



MAXXAM ANALYTICS
#1 2080 39 Ave. NE, Calgary, AB
T2E 6P7

maxxam.ca
Toll Free 800-386-7247
Fax 403-219-3673

AMBIENT AIR MONITORING MONTHLY DATA REPORT
LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
COLD LAKE SOUTH CONTINUOUS MONITORING STATION

JOB #: 2833-2018-10-1-C

October 2018

Prepared for:

LAKELAND INDUSTRY & COMMUNITY ASSOCIATION

5107 50 ST.

BONNYVILLE, ALBERTA

T9N 2J7

Attention: MIKE BISAGA

DATE: **November 23, 2018**

Prepared by: *Maram Ghaleb*

Maram Ghaleb, B.Sc.
Project Manager, Customer Service, Air Services

Reviewed by:

Wunmi Adekanmbi

Wunmi Adekanmbi, M.Sc., EPT, PMP
Project Team Lead, Customer Service, Air Services

SUMMARY

In October 2018, Maxxam Analytics was contracted to manage the ambient air quality monitoring and maintenance activities at the Cold Lake Continuous Monitoring Station, near Cold Lake, Alberta. The monitoring station provides continuous meteorological measurements and air quality data for non-compliance parameters, as requested by the Lakeland Industry & Community Association.

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

All data collected this month were within the Alberta Ambient Air Quality Objectives and Guidelines (November, 2018).

PM_{2.5}: One hour of downtime was incurred on October 1 at hour 11:00, due to datalogger activities surrounding the addition of another PM_{2.5} channel that records minimum concentrations.

The summary of results is presented on the following pages.

Any deviations or modifications made to the sampling or analytical methods are outlined in Section 1.0, Discussion. On this basis, Maxxam Analytics is issuing this completed report to Lakeland Industry & Community Association, Cold Lake South Continuous Monitoring Station.

Should you have any questions concerning the results or if we can be of further assistance, please contact us at 403-219-3677 or toll-free at 1-800-386-7247.

Monthly Continuous Data Summary

Lakeland Industry & Community Association						MAXIMUM VALUES							OPERATIONAL TIME (%)
Cold Lake South Continuous Monitoring Station						1-HOUR					24-HOUR		
PARAMETER	OBJECTIVES		EXCEEDANCES		MONTHLY AVERAGE	READING	DAY	HOUR	WIND SPEED (kph)	WIND DIRECTION (sector)	READING	DAY	
	1-hr	24-hr	1-hr	24-hr									
SO ₂ (ppb)	172	48	0	0	0	2	13	2	12.0	NW	1	13	100.0
TRS (ppb)	-	-	-	-	0	1	16	1	3.9	WNW	0	1	100.0
THC (ppm)	-	-	-	-	2.07	2.75	24	1	1.1	ESE	2.30	24	100.0
CH ₄ (ppm)	-	-	-	-	2.07	2.75	24	1	1.1	ESE	2.29	24	100.0
NMHC (ppm)	-	-	-	-	0.00	0.13	10	18	1.1	SSW	0.01	10	100.0
NO ₂ (ppb)	159	-	0	-	3	36	26	7	6.2	SW	9	24	100.0
NO (ppb)	-	-	-	-	1	79	29	10	10.2	NW	9	25	100.0
NO _x (ppb)	-	-	-	-	5	79	29	10	10.2	NW	18	25	100.0
O ₃ (ppb)	82	-	0	-	22.3	52.4	17	15	10.7	WSW	34.4	18	100.0
PM _{2.5} (µg/m ³)	80	30	0	0	3	27	9	7	0.9	NE	14	25	99.9
RELATIVE HUMIDITY (%)	-	-	-	-	66	98	12	7	1.2	NW	89	29	100.0
AMBIENT TEMPERATURE (°C)	-	-	-	-	2.0	21.9	17	15	10.7	WSW	9.9	17	100.0
VECTOR WS (kph)	-	-	-	-	1.5	17.4	21	13	-	WNW	9.9	13	100.0
VECTOR WD (sec)	-	-	-	-	279 (W)	-	-	-	-	-	-	-	100.0

* Precipitation: data represents the total (sum) for the indicated time frame.

Exceedance Summary Report

SO₂ 1-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 1-hour AAAQO of 172 ppb.

SO₂ 24-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 24-hour AAAQO of 48.0 ppb.

NO₂ 1-Hour Exceedances

Measured concentrations of nitrogen dioxide were below the 1-hour AAAQO of 159 ppb.

PM_{2.5} 1-Hour Exceedances

Measured concentrations of fine particulate matter were below the 1-hour AAAQO of 80 µg/m³.

PM_{2.5} 24-Hour Exceedances

Measured concentrations of fine particulate matter were below the 24-hour AAAQO of 30 µg/m³.

O₃ 1-Hour Exceedances

Measured concentrations of ozone were below the 1-hour AAAQO of 82 ppb.

In accordance with EPEA and the Substance Release Regulation.

In accordance with A Guide to Release Reporting and the Alberta Ambient Air Quality Objectives and Guidelines Summary.

TABLE OF CONTENTS

<u>Title</u>	<u>Page</u>
SUMMARY	2
MONTHLY CONTINUOUS DATA SUMMARY REPORT	3
EXCEEDANCE SUMMARY REPORT	4
TABLE OF CONTENTS	5
1.0 Discussion	7
2.0 Project Personnel	9
3.0 Plant Monthly Required AMD Summary	9
4.0 Calculations and Results	9
5.0 Methods and Procedures	10
Appendix I	13
Continuous Monitoring Data Results	
Sulphur Dioxide	14
Total Reduced Sulphur	20
Total Hydrocarbon	26
Methane	32
Non-Methane Hydrocarbon	38
Oxides of Nitrogen	44
Nitric Oxides	50
Nitrogen Dioxide	55
Ozone	61
Particulate Matter 2.5	67
Wind Speed	72
Wind Direction	77
Standard Deviation Wind Direction	81
Relative Humidity	84
Ambient Temperature	87

Appendix III	Equipment Calibration Results	90
	Sulphur Dioxide	91
	Total Reduced Sulphur	94
	Total Hydrocarbon	97
	Nitrogen Dioxide	101
	Ozone	105
	Particulate Matter	108
	Wind System	110
	Calibrators	112
	Calibration Gases	115
Appendix III	Maximum Instantaneous Data	120
Appendix IV	Report Certification Form	141
Appendix V	Data Validation Certification Form	143

1.0 Discussion

This monthly report consists of continuous monitoring results for the following parameters: Sulphur Dioxide (SO₂), Total Reduced Sulphur (TRS), Total Hydrocarbon (THC), Methane (CH₄), Non-Methane Hydrocarbon (NMHC), Oxides of Nitrogen (NO_x), Nitric Oxides (NO), Nitrogen Dioxide (NO₂), Ozone (O₃), Particulate Matter 2.5 (PM_{2.5}), Relative Humidity (RH), Ambient Temperature (AmbTPX), Wind Speed (WS), Wind Direction (WD) and Standard Deviation Wind Direction (STDWD).

The sample inlet filter for all continuous air analyzers are replaced before the calibration begins. The sample manifold is cleaned during the site visit each month.

Control checks, consisting of a zero and span, are conducted daily on all continuous air monitors. In place of the air sample, zero air (from scrubbed air or gas cylinders) is used for zero checks, and a known concentration of the pollutant being analyzed is used for span checks. These checks are controlled by automatic timers and valves. The total zero span cycle is completed within an hour, the commencement of the zero span cycle is at the beginning of the hour.

Multipoint calibrations are done a minimum of once a month for each continuous air monitor. An additional calibration is required under the following conditions: 1) within three days after the initial start-up and stabilization of a newly installed instrument, 2) prior to shut-down or moving of an instrument which has been working to specification, and 3) when major repair has been done on the instrument.

Time during the first multi-point calibration is not considered downtime (Data is flagged as C). If more than one calibration is performed during the month, the time during the additional calibration is considered as downtime (Data is flagged as C1).

Only one zero/span check is run per day. Time during the zero/span check is not considered as downtime (Data is flagged as S). If an extra zero/span check is performed, the time during the additional check is considered as downtime (Data is flagged as S1).

The AMD requires each instrument and accompanying data recording system to be operational 90% of the time, at a minimum, for each monthly monitoring period.

All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.

Data contained in this monthly report has undergone the verification and validation based on the requirements of the AMD Chapter 6: Ambient Data Quality (December, 2016). The descriptions of the data verification and validation process can be found in Section 5 of this report. Instantaneous data, where applicable, is provided for reference purposes and has not undergone zero correction. The minimum and maximum statistics are highlighted in the data table and are for reference only. The highlighted cells are based on the software's interpretation of the exact position of the minimum or maximum value. The visual presentation of these statistics may not be the obvious choice in a data range due to rounding, truncating or analyzer specifications.

Hourly/minute data have been reviewed based on daily zero/span results and multi-point calibration results. Data may be considered invalid if a zero-corrected span check in excess of +/- 10% of the span concentration (established by the previous multi-point calibration) is encountered and/or significant differences in the calibration factor occurs (greater than 10%).

SULPHUR DIOXIDE (SO₂)

- Operational time for the monitoring period was 100%.
- The routine monthly calibration was performed on October 17.

TOTAL REDUCED SULPHUR (TRS)

- Operational time for the monitoring period was 100%.
- The routine monthly calibration was performed on October 16.

TOTAL HYDROCARBONS (THC), METHANE (CH₄) and NON-METHANE HYDROCARBONS (NMHC)

- Operational time for the monitoring period was 100%.
- The routine monthly calibration was performed on October 17.

OXIDES OF NITROGEN (NO_x), NITRIC OXIDE (NO) and NITROGEN DIOXIDE (NO₂)

- Operational time for the monitoring period was 100%.
- The routine monthly calibration was performed on October 17.
- The City of Cold Lake, Municipal Road and Environmental Service, carried out a clean-up activity around the station towards the end of the month. As a result, elevated concentrations were recorded by the analyzer during this period. As maximum instantaneous data do not undergo the standard data correction process that hourly average data are subjected to, the NO/NO_x/NO₂ parameter relationship was not preserved on some maximum instantaneous data recorded during the said period.

OZONE (O₃)

- Operational time for the monitoring period was 100%.
- The routine monthly calibration was performed on October 16.

PARTICULATE MATTER < 2.5 MICRONS (PM_{2.5})

- Operational time for the monitoring period was 99.9%, equivalent to 1 hour of downtime. This was incurred on October 1 at hour 11:00, due to datalogger activities surrounding the addition of another PM_{2.5} channel that records minimum concentrations.
- The routine monthly check was performed on October 16.

WIND SPEED (WS), WIND DIRECTION (WD) and STANDARD DEVIATION WIND DIRECTION (STDWD)

- Operational time for the monitoring period was 100%.
- Wind data is reported as vector wind speed and vector wind direction. Wind direction is defined as the direction from which the wind is blowing from and is measured in degrees from true north.

RELATIVE HUMIDITY (RH)

- Operational time for the monitoring period was 100%.

AMBIENT TEMPERATURE (AmbTPX)

- Operational time for the monitoring period was 100%.

2.0 Project Personnel

Mike Bisaga and Lily Lin were the contacts for Lakeland Industry & Community Association and the Maxxam field technician was Alexander Yakupov.

3.0 Plant Monthly Required AMD Summary

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

All data collected this month were within the Alberta Ambient Air Quality Objectives and Guidelines (November, 2018).

4.0 Calculations and Results

All calculations and reporting of results follow the methods described in the AMD, 2016.

5.0 Methods and Procedures

The following methods and procedures were used to complete the monitoring program:

- Maxxam AIR SOP-00001: Methane, Non-Methane Hydrocarbon Analyzer Monitoring
- Maxxam AIR SOP-00014: Measurement of Particulate Concentration Using the THERMO SHARP
- Maxxam AIR SOP-00209: Ambient Sulphur Monitoring
- Maxxam AIR SOP-00212: Ambient O₃ Monitoring
- Maxxam AIR SOP-00213: Ambient NO/NO₂/NO_x Monitoring
- MET One Instruments: Operation Manual Document No. 50.5-9800

There were no deviations from the prescribed methods.

The following instruments were used to perform the test program:

- Sulphur Dioxide - Thermo 43i UV Fluorescent Analyzer
- Total Reduced Sulphur - Thermo 450i UV Fluorescent Analyzer
- Methane, Non-Methane Hydrocarbon - Thermo 55i FID Analyzer
- Oxides of Nitrogen - Thermo 42i Chemiluminescent Analyzer
- Ozone - Thermo 49i Photometric Analyzer
- Particulate Matter (PM_{2.5}) - Thermo SHARP 5030 Unit
- Wind System - Met One Unit
- Relative Humidity - Met One Unit
- Ambient Temperature - Met One Unit
- Datalogger - Envista Ultimate

The following steps were used to complete the data verification and validation process:

Level 0 Preliminary Verification

Level 0 data are raw data obtained directly from the data acquisition system (DAS). Under the step of Level 0, these data undergo a certain amount of manual or automated screening and flagging. It included a) identification of periods of missing data; b) verification of time stamps against reference time; c) verification that instrument diagnostics/datalogger flags indicate normal operation; d) comparison of data to upper and lower limits; e) rate of change flagging indicating that data changed too rapidly or not at all; and f) verification that zero, span and multipoint performance checks are within specifications. This level of verification is performed on a daily basis.

Level 1 Primary Validation

Validation actions under the step of Level 1 include a) review of all screening flags assigned during preliminary verification; b) review of all supporting site information and documentation; c) review of operational acceptance limits for each parameter/analyzer; d) review of daily zero/span and monthly calibration results for all gaseous parameters; and e) application of any necessary adjustments to data (e.g. baseline adjustments, below zero adjustments). This level of validation is performed on a monthly basis.

Level 2 Final Validation


The purpose of Level 2 validation is to verify that there are no inconsistencies among related data, or among regional data measured at nearby sites.

Level 3 Independent Data Review

Level 3 validation is the last step of data review, and it is completed by an individual that is independent of both field operations and primary data validation. A final independent QA review and endorsement is performed during this step before data is submitted to Alberta Environment.

Post-Final Validation

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. Any data issues or patterns which were not clear on a monthly basis are highlighted during this step. This validation is performed on an annual basis.



Source: Air Monitoring Directive (December 2016), Chapter 6, Ambient Data Quality; Figure 1 Data Collection and Management Process Flow Chart

APPENDIX I
CONTINUOUS MONITORING DATA RESULTS

SULPHUR DIOXIDE

SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
2	0	1	0	0	S	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
3	0	0	0	S	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
4	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	1	0	24	
5	0	S	0	0	0	0	0	0	0	0	1	1	1	1	0	0	0	0	0	1	0	0	0	0	0	0	0	0	1	0	24	
6	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	24	
7	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	24	
8	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	24	
9	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	24	
10	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	24	
11	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	1	0	24	
12	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	1	1	0	0	0	0	0	1	0	24	
13	1	1	2	2	1	1	2	1	1	1	1	1	1	0	0	0	S	0	0	0	1	0	0	0	0	0	0	0	2	1	24	
14	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
15	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	1	0	0	0	1	0	24	
16	0	1	1	1	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
17	0	0	0	0	0	0	0	0	0	C	C	C	C	C	C	C	C	1	2	1	0	0	0	0	0	0	0	0	2	0	24	
18	0	0	0	1	1	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
19	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
20	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
21	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
22	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
23	0	0	0	0	0	0	S	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
24	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
25	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
26	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
27	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
28	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
29	S	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	1	1	1	1	S	0	0	0	0	1	0	24	
30	1	1	1	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	1	0	24	
31	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	24	
HOURLY MAX	1	1	2	2	1	1	2	1	1	1	1	1	1	1	0	0	1	1	2	1	1	1	1	1	1	1	1					
HOURLY AVG	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT:	1-HR	172	ppb	24-HR	48	ppb
----------------------	------	-----	-----	-------	----	-----


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0		
NUMBER OF 24-HR EXCEEDANCES:	0		
NUMBER OF NON-ZERO READINGS:	51		
MINIMUM 1-HR AVERAGE:	0 ppb @ HOUR	0 ON DAY	1
MAXIMUM 1-HR AVERAGE:	2 ppb @ HOUR	2 ON DAY	13
MAXIMUM 24-HR AVERAGE:	1 ppb	ON DAY	13
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	744 hrs
MONTHLY CALIBRATION TIME:	8 hrs	AMD OPERATION UPTIME:	100.0 %
STANDARD DEVIATION:	0	MONTHLY AVERAGE:	0 ppb

24 HR AVERAGES October 2018



SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)









Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-SO2[ppb]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 21.59%

Calm Avg: 0.08 [ppb]

Direction	0.0-0.6	0.6-1.2	1.2-1.8	1.8-2.4	2.4-3.0	>3.0	Total
N	4.8	0.1	0.0	0.0	0.0	0.0	5.0
NE	6.1	0.0	0.0	0.0	0.0	0.0	6.1
E	7.0	0.0	0.0	0.0	0.0	0.0	7.0
SE	14.8	0.3	0.0	0.0	0.0	0.0	15.1
S	2.6	0.0	0.0	0.0	0.0	0.0	2.6
SW	12.6	0.4	0.1	0.0	0.0	0.0	13.2
W	16.1	0.6	0.0	0.0	0.0	0.0	16.6
NW	8.8	3.0	1.0	0.1	0.0	0.0	12.9
Summary	72.7	4.4	1.1	0.1	0.0	0.0	78.4


% Icon	Classes (ppb)	73		0.0-0.6	4		0.6-1.2	1		1.2-1.8	0		1.8-2.4	0		2.4-3.0	0		>3.0
--------	---------------	----	---	---------	---	---	---------	---	--	---------	---	---	---------	---	---	---------	---	---	------

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-SO2[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 21.59% Calm Poll Avg: 0.08[ppb]



SO2[ppb] Calibration: LICA COLD LAKE SOUTH Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



TOTAL REDUCED SULPHUR


TOTAL REDUCED SULPHUR Hourly Averages (TRS ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY																													
1	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
2	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
3	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
4	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
5	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
6	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	24	
7	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	24	
8	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	24	
9	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	24	
10	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	24	
11	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	24	
12	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	24	
13	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	24	
14	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	24	
15	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	24	
16	0	1	0	0	0	0	0	0	0	0	0	C	C	C	C	C	C	C	0	0	0	0	0	0	0	1	0	24	
17	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
18	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
19	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
20	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
21	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
22	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
23	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
24	0	0	0	0	0	S	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
25	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
26	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
27	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
28	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
29	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	24	
30	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	24	
31	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	24	
HOURLY MAX	0	1	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	
HOURLY AVG	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	2			
MINIMUM 1-HR AVERAGE:	0	ppb @ HOUR	0	ON DAY 1
MAXIMUM 1-HR AVERAGE:	1	ppb @ HOUR	1	ON DAY 16
MAXIMUM 24-HR AVERAGE:	0	ppb		ON DAY 1
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	744 hrs
MONTHLY CALIBRATION TIME:	7	hrs	AMD OPERATION UPTIME:	100.0 %
STANDARD DEVIATION:	0		MONTHLY AVERAGE:	0 ppb

TOTAL REDUCED SULPHUR Hourly Averages (TRS ppb)




Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-TRS[ppb]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 21.42% Calm Avg: 0.23 [ppb]

Direction	0.0-0.7	0.7-1.3	1.3-2.0	>2.0	Total
N	5.0	0.0	0.0	0.0	5.0
NE	6.1	0.0	0.0	0.0	6.1
E	7.0	0.0	0.0	0.0	7.0
SE	15.0	0.0	0.0	0.0	15.0
S	2.6	0.0	0.0	0.0	2.6
SW	13.6	0.0	0.0	0.0	13.6
W	16.5	0.0	0.0	0.0	16.5
NW	12.9	0.0	0.0	0.0	12.9
Summary	78.6	0.0	0.0	0.0	78.6

% Icon Classes (ppb) 79 0.0-0.7 0 0.7-1.3 0 1.3-2.0 0 >2.0

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-TRS[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 21.42% Calm Poll Avg: 0.23[ppb]




TRS[ppb] Calibration: LICA COLD LAKE SOUTH Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



TOTAL HYDROCARBON

TOTAL HYDROCARBONS Hourly Averages (THC ppm)




Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-THC[ppm]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 21.50% Calm Avg: 2.20 [ppm]

Direction	0.0-0.9	0.9-1.8	1.8-2.8	>2.8	Total
N	0.0	0.0	5.0	0.0	5.0
NE	0.0	0.0	6.1	0.0	6.1
E	0.0	0.0	6.9	0.0	6.9
SE	0.0	0.0	15.0	0.0	15.0
S	0.0	0.0	2.6	0.0	2.6
SW	0.0	0.0	13.6	0.0	13.6
W	0.0	0.0	16.6	0.0	16.6
NW	0.0	0.0	12.9	0.0	12.9
Summary	0.0	0.0	78.5	0.0	78.5


% Icon Classes (ppm) 0 0.0-0.9 0 0.9-1.8 79 1.8-2.8 0 >2.8

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-THC[ppm] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 21.50% Calm Poll Avg: 2.20[ppm]



THC[ppm] Calibration: LICA COLD LAKE SOUTH Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



METHANE


METHANE Hourly Averages (CH₄ ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2.08	2.09	2.08	2.08	2.09	S	2.12	2.21	2.34	2.24	2.21	2.12	2.08	2.07	2.07	2.06	2.06	2.10	2.13	2.10	2.07	2.08	2.09	2.05	2.05	2.34	2.11	24	
2	2.05	2.06	2.07	2.04	S	2.01	2.01	2.07	2.06	1.99	2.00	1.99	1.99	1.99	1.99	1.98	1.98	1.99	2.04	2.04	2.05	2.09	2.09	1.98	1.98	2.09	2.02	24	
3	2.08	2.12	2.14	S	2.08	2.07	2.06	2.06	2.03	1.99	1.98	1.97	1.97	1.97	1.97	1.97	1.98	2.00	2.04	2.02	2.04	2.04	2.03	1.97	2.14	2.03	24		
4	2.02	2.03	S	2.02	2.01	2.01	2.01	2.01	2.02	2.05	2.04	2.02	2.02	2.01	2.02	2.03	2.04	2.08	2.16	2.22	2.15	2.15	2.18	2.20	2.01	2.22	2.07	24	
5	2.19	S	2.24	2.25	2.28	2.26	2.28	2.34	2.20	2.15	2.19	2.08	2.01	2.00	1.99	1.99	1.99	2.00	2.00	2.00	2.00	2.02	2.05	2.04	1.99	2.34	2.11	24	
6	S	2.05	2.04	2.01	2.01	2.01	2.01	2.00	2.01	2.00	2.00	2.01	2.02	2.02	2.03	2.03	2.04	2.05	2.05	2.05	2.08	2.07	2.04	S	2.00	2.08	2.03	24	
7	2.04	2.02	1.99	1.99	1.99	1.99	1.99	1.98	1.98	1.98	1.98	1.97	1.98	1.97	1.98	1.98	1.98	1.97	1.97	1.97	1.97	1.97	1.97	S	1.97	1.97	2.04	1.98	24
8	1.97	1.97	1.97	1.97	1.97	1.97	1.98	1.98	1.97	1.97	1.97	1.97	1.97	1.96	1.96	1.96	1.96	1.96	1.96	1.96	1.97	S	2.03	2.07	1.96	2.07	1.97	24	
9	2.12	2.14	2.10	2.13	2.17	2.25	2.24	2.27	2.34	2.21	2.01	1.95	1.95	1.95	1.95	1.95	1.95	1.96	1.98	2.03	S	2.12	2.13	2.15	1.95	2.34	2.09	24	
10	2.14	2.20	2.20	2.16	2.15	2.18	2.22	2.17	2.07	2.00	2.00	1.98	1.98	1.97	1.97	1.97	1.97	1.98	2.01	S	2.03	2.05	2.16	2.24	1.97	2.24	2.08	24	
11	2.27	2.35	2.57	2.31	2.26	2.24	2.20	2.27	2.43	2.43	2.10	2.13	2.09	2.11	2.13	2.09	2.03	2.05	2.03	S	2.02	2.02	2.10	2.08	2.10	2.02	2.57	2.19	24
12	2.11	2.11	2.13	2.17	2.16	2.23	2.25	2.28	2.22	2.27	2.26	2.17	2.01	1.99	2.00	1.99	1.97	S	1.94	1.94	1.94	1.96	1.94	1.93	1.93	2.28	2.08	24	
13	1.95	1.95	1.96	1.97	1.96	1.96	1.96	1.96	1.95	1.95	1.96	1.95	1.95	1.95	1.95	S	1.95	1.95	1.96	1.97	1.99	2.01	2.00	2.10	1.95	2.10	1.97	24	
14	2.14	2.19	2.25	2.26	2.22	2.20	2.22	2.25	2.10	2.05	2.00	1.99	1.96	1.98	1.99	S	1.98	1.99	2.03	2.06	2.03	2.04	2.03	2.00	1.96	2.26	2.08	24	
15	2.02	2.08	2.03	1.97	1.97	1.98	1.99	2.00	1.98	1.95	1.91	1.91	1.91	1.91	S	1.92	1.92	1.91	1.93	1.91	1.90	1.93	1.91	1.91	1.90	2.08	1.95	24	
16	1.92	1.96	1.95	1.92	1.96	2.05	2.14	2.29	2.18	2.10	2.09	2.04	1.98	S	1.92	1.92	1.92	1.93	1.96	2.18	2.20	2.24	2.22	2.14	1.92	2.29	2.05	24	
17	2.17	2.19	2.25	2.22	2.25	2.29	2.29	2.30	2.05	C	C	C	C	C	1.97	1.96	1.96	1.99	2.08	2.09	2.11	2.05	2.05	2.08	1.96	2.30	2.12	24	
18	2.10	2.17	2.24	2.27	2.21	2.12	2.14	2.11	2.07	2.03	1.97	S	1.96	1.94	1.93	1.94	1.95	1.96	1.96	1.97	1.97	2.00	2.05	2.05	1.93	2.27	2.05	24	
19	2.06	2.12	2.10	2.08	2.15	2.10	2.09	2.14	2.22	2.13	S	1.96	1.96	1.96	1.96	1.96	1.96	1.98	1.98	2.00	2.01	1.97	1.97	1.96	2.22	2.04	24		
20	2.00	2.02	1.99	1.99	2.08	2.02	2.08	2.03	S	2.07	2.10	2.12	2.06	1.98	1.97	1.97	1.96	1.95	1.97	1.97	2.01	2.03	2.02	1.95	2.12	2.02	24		
21	2.00	1.99	1.96	1.98	1.98	2.00	2.02	2.02	S	1.99	1.96	1.95	1.94	1.95	1.95	1.94	1.95	1.96	1.98	2.00	2.01	2.04	2.10	2.14	1.94	2.14	1.99	24	
22	2.15	2.18	2.26	2.30	2.40	2.36	2.37	S	2.43	2.13	2.08	2.07	2.05	2.01	2.01	2.06	2.18	2.22	2.33	2.14	2.18	2.21	2.20	2.14	2.01	2.43	2.19	24	
23	2.14	2.16	2.11	2.10	2.02	2.03	S	2.14	2.05	2.05	2.04	2.03	2.03	2.00	1.99	2.01	2.05	2.09	2.20	2.18	2.23	2.39	2.51	1.99	2.51	2.11	24		
24	2.58	2.75	2.70	2.74	2.65	S	2.58	2.57	2.51	2.20	2.02	2.00	1.99	1.98	1.99	1.99	2.02	2.02	2.14	2.19	2.19	2.29	2.31	2.36	1.98	2.75	2.29	24	
25	2.31	2.31	2.42	2.49	S	2.42	2.53	2.60	2.55	2.33	2.02	1.97	1.98	1.98	1.98	1.97	1.98	1.99	2.02	2.13	2.17	2.24	2.19	2.15	1.97	2.60	2.21	24	
26	2.27	2.22	2.15	S	2.11	2.14	2.23	2.55	2.30	2.10	2.03	1.96	1.94	1.95	1.94	1.94	1.94	1.95	1.99	2.04	2.10	2.07	2.14	2.21	1.94	2.55	2.10	24	
27	2.18	2.27	S	2.44	2.50	2.26	2.16	2.07	2.08	2.01	1.97	1.98	1.93	1.92	1.92	1.93	1.93	1.94	1.96	1.97	1.97	1.95	1.95	1.95	1.92	2.50	2.05	24	
28	1.97	S	2.01	2.08	2.10	2.10	2.15	2.21	2.20	2.07	2.01	1.97	1.96	1.97	1.97	1.97	1.99	2.01	2.00	1.98	1.98	1.97	1.96	1.98	1.96	2.21	2.03	24	
29	S	1.95	1.95	1.95	1.96	1.94	1.93	1.94	1.93	1.93	1.95	1.93	1.92	1.92	1.94	1.94	1.94	1.94	1.95	1.97	1.94	1.95	1.97	1.97	S	1.92	1.97	1.94	24
30	1.99	2.01	2.01	2.02	2.05	2.05	1.98	1.99	2.01	1.97	1.99	1.99	1.97	1.98	1.98	2.00	2.01	2.03	2.08	2.10	2.11	2.11	S	2.16	1.97	2.16	2.03	24	
31	2.25	2.26	2.18	2.17	2.12	2.13	2.18	2.18	2.18	2.18	2.08	2.00	1.98	1.98	1.97	1.98	1.98	1.99	2.02	2.08	2.12	S	2.13	2.18	1.97	2.26	2.10	24	
HOURLY MAX	2.58	2.75	2.70	2.74	2.65	2.42	2.58	2.60	2.55	2.43	2.26	2.17	2.12	2.13	2.09	2.06	2.18	2.22	2.33	2.22	2.20	2.29	2.39	2.51					
HOURLY AVG	2.11	2.14	2.14	2.14	2.13	2.12	2.15	2.17	2.15	2.08	2.03	2.01	1.99	1.98	1.98	1.98	1.98	2.00	2.02	2.04	2.05	2.07	2.08	2.10					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	707						
MINIMUM 1-HR AVERAGE:	1.90	ppm	@ HOUR	20	ON DAY	15	
MAXIMUM 1-HR AVERAGE:	2.75	ppm	@ HOUR	1	ON DAY	24	
MAXIMUM 24-HR AVERAGE:	2.29	ppm			ON DAY	24	
IZS CALIBRATION TIME:		32	hrs	OPERATIONAL TIME:		744	hrs
MONTHLY CALIBRATION TIME:		5	hrs	AMD OPERATION UPTIME:		100.0	%
STANDARD DEVIATION:		0.13		MONTHLY AVERAGE:		2.07	ppm

METHANE Hourly Averages (CH₄ ppm)




Wind: LICA COLD LAKE SOUTH
Poll.: LICA COLD LAKE SOUTH-CH4[ppm]
Monthly: 18/10
Type: PollutionRose
Direction: Blowing From (Wind Frequency)
Based On 1 Hr.

Calm: 21.50% Calm Avg: 2.19 [ppm]

Direction	0.0-0.9	0.9-1.8	1.8-2.8	>2.8	Total
N	0.0	0.0	5.0	0.0	5.0
NE	0.0	0.0	6.1	0.0	6.1
E	0.0	0.0	6.9	0.0	6.9
SE	0.0	0.0	15.0	0.0	15.0
S	0.0	0.0	2.6	0.0	2.6
SW	0.0	0.0	13.6	0.0	13.6
W	0.0	0.0	16.6	0.0	16.6
NW	0.0	0.0	12.9	0.0	12.9
Summary	0.0	0.0	78.5	0.0	78.5

% Icon Classes (ppm) 0 0.0-0.9 0.9-1.8 79 1.8-2.8 0 >2.8

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-CH4[ppm] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 21.50% Calm Poll Avg: 2.19[ppm]



CH4[ppm] Calibration: LICA COLD LAKE SOUTH Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



NON-METHANE HYDROCARBON


NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.								
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.									
DAY																																				
1	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24							
2	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24							
3	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24							
4	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24							
5	0.00	S	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	24							
6	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	24							
7	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.03	0.00	0.00	24							
8	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24						
9	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.03	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.03	0.00	0.00	24						
10	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.13	S	0.00	0.00	0.00	0.00	0.00	0.13	0.01	0.00	24						
11	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24					
12	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24					
13	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24					
14	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
15	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
16	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	24				
17	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	C	C	C	C	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
18	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
19	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24				
20	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24			
21	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24			
22	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24			
23	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24			
24	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.11	0.02	0.00	0.00	0.00	0.00	0.00	0.11	0.01	0.00	0.00	0.00	0.00	24			
25	0.01	0.00	0.00	0.00	S	0.00	0.00	0.01	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	24			
26	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
27	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
28	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
29	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24		
30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
31	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
HOURLY MAX	0.01	0.00	0.00	0.00	0.00	0.00	0.01	0.03	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.03	0.00	0.07	0.13	0.02	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
HOURLY AVG	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	12					
MINIMUM 1-HR AVERAGE:	0.00	ppm	@ HOUR	0	ON DAY	1
MAXIMUM 1-HR AVERAGE:	0.13	ppm	@ HOUR	18	ON DAY	10
MAXIMUM 24-HR AVERAGE:	0.01	ppm			ON DAY	10
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	744	hrs	
MONTHLY CALIBRATION TIME:	5	hrs	AMD OPERATION UPTIME:	100.0	%	
STANDARD DEVIATION:	0.01		MONTHLY AVERAGE:	0.00	ppm	

NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)



Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-NMHC[ppm]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 21.50%

Calm Avg: 0.00 [ppm]

Direction	0.0-0.4	0.4-0.8	0.8-1.2	1.2-1.6	1.6-2.0	>2.0	Total
N	5.0	0.0	0.0	0.0	0.0	0.0	5.0
NE	6.1	0.0	0.0	0.0	0.0	0.0	6.1
E	6.9	0.0	0.0	0.0	0.0	0.0	6.9
SE	15.0	0.0	0.0	0.0	0.0	0.0	15.0
S	2.6	0.0	0.0	0.0	0.0	0.0	2.6
SW	13.6	0.0	0.0	0.0	0.0	0.0	13.6
W	16.6	0.0	0.0	0.0	0.0	0.0	16.6
NW	12.9	0.0	0.0	0.0	0.0	0.0	12.9
Summary	78.5	0.0	0.0	0.0	0.0	0.0	78.5


% Icon Classes (ppm)	79	0	0	0	0	0
	0.0-0.4	0.4-0.8	0.8-1.2	1.2-1.6	1.6-2.0	>2.0

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-NMHC[ppm] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 21.50% Calm Poll Avg: 0.00[ppm]



NMHC[ppm] Calibration: LICA COLD LAKE SOUTH Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



OXIDES OF NITROGEN



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018


OXIDES OF NITROGEN Hourly Averages (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY																												
1	1	0	0	1	1	S	4	3	3	2	3	3	1	1	3	1	1	2	3	2	2	1	1	1	0	4	2	24
2	1	2	2	1	S	1	1	3	2	2	3	0	1	0	0	0	0	0	1	2	3	3	3	4	0	4	2	24
3	3	4	5	S	5	4	6	6	3	2	1	0	0	0	0	0	0	1	8	9	14	6	2	0	14	3	24	
4	2	3	S	1	0	2	2	2	1	1	1	1	1	1	1	1	2	4	4	2	2	2	2	2	0	4	2	24
5	2	S	3	2	2	3	8	9	6	5	6	5	2	1	1	1	0	1	1	2	2	2	4	7	0	9	3	24
6	S	11	8	2	2	1	1	3	2	1	1	1	1	1	1	2	3	3	3	4	2	2	1	S	1	11	2	24
7	3	2	1	1	1	1	1	2	2	1	0	0	1	1	1	1	1	1	1	1	1	0	S	2	0	3	1	24
8	2	1	1	1	1	2	2	1	1	1	1	0	0	1	1	1	1	1	1	0	1	0	1	S	5	5	1	24
9	8	9	6	5	4	14	9	13	14	10	4	0	0	0	0	0	0	2	7	13	S	24	20	15	0	24	8	24
10	10	22	18	11	8	11	32	25	18	1	0	0	0	1	1	1	2	7	S	5	5	4	3	0	32	8	24	
11	10	10	5	2	3	3	4	5	6	4	3	2	2	2	2	2	2	2	S	2	2	3	2	2	2	10	3	24
12	2	2	3	4	3	4	7	8	8	11	10	8	3	2	2	1	1	S	1	1	2	4	1	1	1	11	4	24
13	2	2	3	3	2	3	3	3	2	2	2	2	1	1	1	0	S	1	1	3	4	5	3	4	0	5	2	24
14	5	6	5	4	3	4	4	4	2	3	2	1	0	1	1	S	1	2	2	3	3	3	4	4	0	6	3	24
15	4	4	4	2	2	3	5	7	5	2	0	0	0	0	S	1	0	0	1	1	1	2	1	3	0	7	2	24
16	2	4	6	7	5	10	13	28	22	5	5	3	4	S	2	2	2	4	6	4	5	4	3	2	2	28	6	24
17	3	2	1	1	3	4	5	17	7	C	C	C	C	C	C	C	C	C	C	4	4	4	3	4	1	17	4	24
18	3	4	4	5	4	4	6	6	7	4	1	S	2	1	0	1	1	1	0	1	1	3	3	3	0	7	3	24
19	3	4	4	4	5	4	7	12	9	7	S	2	2	2	1	1	2	6	8	9	9	6	1	0	12	5	24	
20	0	0	1	1	1	1	2	1	2	S	2	3	3	3	1	1	1	1	0	1	1	2	3	3	0	3	1	24
21	2	2	1	2	2	2	3	3	S	3	1	1	0	0	0	0	0	1	2	3	3	3	7	7	0	7	2	24
22	4	4	11	22	25	12	26	S	43	10	4	3	2	1	1	1	3	6	5	3	2	2	3	3	1	43	9	24
23	3	2	2	2	2	2	S	4	3	3	3	2	2	3	2	7	14	11	7	15	12	7	12	2	15	5	24	
24	8	14	12	12	5	S	9	16	31	32	4	3	3	4	4	5	8	13	18	17	14	18	17	16	3	32	12	24
25	9	8	18	18	S	16	20	33	63	46	7	2	3	4	10	3	4	12	14	23	23	36	19	13	2	63	18	24
26	16	11	4	S	3	6	18	46	17	9	5	2	2	2	1	1	3	8	4	6	6	8	7	1	46	8	24	
27	5	8	S	6	8	8	6	6	5	4	3	3	1	1	1	1	1	1	5	3	2	2	2	2	1	8	4	24
28	3	S	3	4	4	5	11	9	7	3	2	1	2	1	2	4	3	2	2	2	2	1	1	2	1	11	3	24
29	S	3	2	2	3	3	2	12	30	46	79	5	11	11	9	1	1	2	2	2	2	4	5	S	1	79	11	24
30	6	6	5	3	4	5	5	8	5	3	2	3	3	3	4	4	5	5	6	4	4	S	4	2	8	4	24	
31	5	4	3	3	2	3	6	10	9	7	7	5	4	4	3	4	4	8	11	12	10	S	7	6	2	12	6	24
HOURLY MAX	16	22	18	22	25	16	32	46	63	46	79	8	11	11	10	5	8	14	18	23	23	36	20	16				
HOURLY AVG	4	5	5	5	4	5	8	10	11	8	6	2	2	2	2	1	2	3	5	5	5	6	5	5				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	648			
MINIMUM 1-HR AVERAGE:	0	ppb @ HOUR	1	ON DAY
MAXIMUM 1-HR AVERAGE:	79	ppb @ HOUR	10	ON DAY
MAXIMUM 24-HR AVERAGE:	18	ppb		ON DAY
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	744
MONTHLY CALIBRATION TIME:	10	hrs	AMD OPERATION UPTIME:	100.0
STANDARD DEVIATION:	7		MONTHLY AVERAGE:	5
				ppb





OXIDES OF NITROGEN Hourly Averages (NO_x ppb)




Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-NOX[ppb]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 21.65% Calm Avg: 10.51 [ppb]

Direction	0.0-26.7	26.7-53.3	53.3-80.0	>80.0	Total
N	5.0	0.0	0.0	0.0	5.0
NE	6.1	0.0	0.0	0.0	6.1
E	6.8	0.1	0.0	0.0	7.0
SE	15.1	0.0	0.0	0.0	15.1
S	2.6	0.0	0.0	0.0	2.6
SW	12.8	0.1	0.0	0.0	13.0
W	16.7	0.0	0.0	0.0	16.7
NW	12.5	0.3	0.1	0.0	13.0
Summary	77.7	0.6	0.1	0.0	78.4


% Icon	Classes (ppb)	78	1	0	0
	0.0-26.7		26.7-53.3		53.3-80.0
	>80.0				

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-NOX[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 21.65% Calm Poll Avg: 10.51[ppb]



NOX[ppb] Calibration: LICA COLD LAKE SOUTH Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



NITRIC OXIDE


NITRIC OXIDE Hourly Averages (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	0	0	0	0	0	S	0	0	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
2	0	0	0	0	S	0	0	0	1	1	1	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
3	0	0	0	S	0	0	0	1	1	1	0	0	0	0	0	0	0	0	0	1	0	2	0	0	0	0	2	0	24
4	0	0	S	0	0	0	0	0	0	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
5	0	S	0	0	0	0	1	2	2	2	2	2	1	0	0	0	0	0	0	0	0	0	0	0	0	0	2	1	24
6	S	1	1	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	1	0	24
7	0	0	0	0	0	0	0	1	1	0	0	0	0	1	0	0	0	0	0	0	0	0	S	0	0	0	1	0	24
8	1	0	0	0	0	1	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	1	0	24
9	1	1	0	0	0	6	2	7	8	5	2	0	0	0	0	0	0	0	0	2	S	9	6	3	0	9	2	24	
10	2	10	7	2	1	3	20	17	12	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	20	3	24	
11	2	1	0	0	0	0	0	1	1	1	0	1	1	0	0	0	0	0	0	S	0	0	0	0	0	2	0	24	
12	0	0	0	0	1	2	2	3	2	4	3	3	1	1	0	0	0	0	S	0	0	0	0	0	0	4	1	24	
13	0	0	0	0	0	0	0	0	0	1	1	1	1	0	0	0	S	0	0	0	0	0	0	0	0	1	0	24	
14	0	0	0	0	0	0	0	2	1	1	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	2	0	24	
15	0	0	0	0	0	0	0	1	1	1	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	1	0	24	
16	0	0	0	0	0	1	4	17	14	1	1	1	1	S	0	0	0	1	1	0	0	0	0	0	0	17	2	24	
17	0	0	0	0	1	0	0	8	2	C	C	C	C	C	C	C	C	C	C	C	0	0	0	0	0	8	1	24	
18	0	0	0	0	0	0	0	0	1	1	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
19	0	0	0	0	0	0	1	2	2	2	S	0	0	1	0	0	0	1	0	0	1	0	0	0	0	2	1	24	
20	0	0	0	0	0	0	0	0	0	S	0	1	1	1	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
21	0	0	0	0	0	0	0	0	S	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
22	0	0	2	9	13	3	15	S	27	5	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	27	3	24	
23	0	0	0	0	0	0	S	0	0	1	1	1	1	0	1	0	1	1	1	1	1	2	1	0	2	2	1	24	
24	1	4	2	4	0	S	4	10	20	19	2	1	1	1	1	1	0	0	0	1	0	4	2	4	0	20	4	24	
25	1	2	8	9	S	9	12	23	42	47	2	1	1	2	8	0	0	0	1	7	8	19	6	1	0	47	9	24	
26	2	1	0	S	0	0	6	10	5	3	1	1	1	1	0	0	0	0	0	0	0	0	0	1	0	10	1	24	
27	0	1	S	0	1	0	0	1	1	1	1	0	0	0	0	0	0	0	1	0	0	0	0	0	0	1	0	24	
28	0	S	0	0	0	0	2	3	3	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	3	1	24	
29	S	0	0	0	0	0	0	8	21	39	79	1	8	8	7	0	0	0	0	0	0	0	0	S	0	79	8	24	
30	0	0	0	0	1	0	0	1	1	1	0	1	1	0	1	0	0	0	0	0	0	0	S	0	0	1	0	24	
31	0	0	0	0	0	0	3	5	5	2	3	2	1	1	0	0	0	0	1	1	0	S	0	0	0	5	1	24	
HOURLY MAX	2	10	8	9	13	9	20	23	42	47	79	3	8	8	8	1	1	1	1	7	8	19	6	4					
HOURLY AVG	0	1	1	1	1	1	2	4	6	5	4	1	1	1	1	0	0	0	0	0	0	1	0	0					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	215			
MINIMUM 1-HR AVERAGE:	0	ppb @ HOUR	0	ON DAY 1
MAXIMUM 1-HR AVERAGE:	79	ppb @ HOUR	10	ON DAY 29
MAXIMUM 24-HR AVERAGE:	9	ppb		ON DAY 25
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	744
MONTHLY CALIBRATION TIME:	10	hrs	AMD OPERATION UPTIME:	100.0 %
STANDARD DEVIATION:	5		MONTHLY AVERAGE:	1
				ppb

NITRIC OXIDE Hourly Averages (NO ppb)




Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-NO[ppb]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 21.65% Calm Avg: 3.51 [ppb]

Direction	0.0-26.7	26.7-53.3	53.3-80.0	>80.0	Total
N	5.0	0.0	0.0	0.0	5.0
NE	6.1	0.0	0.0	0.0	6.1
E	7.0	0.0	0.0	0.0	7.0
SE	15.1	0.0	0.0	0.0	15.1
S	2.6	0.0	0.0	0.0	2.6
SW	13.0	0.0	0.0	0.0	13.0
W	16.7	0.0	0.0	0.0	16.7
NW	12.7	0.1	0.1	0.0	13.0
Summary	78.1	0.1	0.1	0.0	78.4

% Icon Classes (ppb) 78 0.0-26.7 0 26.7-53.3 0 53.3-80.0 0 >80.0

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-NO[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 21.65% Calm Poll Avg: 3.51[ppb]



NITROGEN DIOXIDE



NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	1	0	0	1	1	S	4	3	3	2	2	2	0	0	2	1	1	2	3	2	2	1	1	1	0	4	2	24
2	1	2	2	1	S	1	1	3	2	1	2	0	0	0	0	0	0	0	1	2	3	3	3	4	0	4	1	24
3	3	3	4	S	5	4	6	5	2	1	0	0	0	0	0	0	0	0	1	7	8	12	5	1	0	12	3	24
4	1	2	S	1	0	2	2	2	1	1	1	0	0	1	1	1	2	4	4	2	2	2	2	2	0	4	1	24
5	2	S	3	2	2	3	7	7	4	3	4	3	2	1	1	0	1	1	2	4	2	2	4	7	0	7	3	24
6	S	10	7	2	1	1	1	2	1	1	0	0	1	1	1	1	2	3	3	4	2	2	1	S	0	10	2	24
7	3	2	1	1	0	0	1	1	1	1	0	0	0	1	0	1	1	1	1	0	0	0	S	1	0	3	1	24
8	1	1	0	0	1	1	2	1	0	1	1	0	0	0	0	0	0	1	1	0	1	S	5	5	0	5	1	24
9	7	8	6	5	3	8	7	5	6	4	2	0	0	0	0	0	0	2	6	11	S	15	14	12	0	15	5	24
10	9	12	11	9	7	9	12	8	7	0	0	0	0	1	1	1	2	7	S	5	5	4	3	0	0	12	5	24
11	8	9	5	2	3	3	4	5	5	3	2	2	2	2	2	2	2	2	S	2	2	3	2	2	2	9	3	24
12	2	2	3	4	2	3	4	6	6	7	6	5	2	2	2	1	1	S	1	1	2	4	1	1	1	7	3	24
13	2	2	3	3	2	3	3	3	2	2	1	1	1	1	0	0	S	1	1	3	4	5	3	4	0	5	2	24
14	5	6	4	4	3	3	4	2	2	2	1	1	0	1	1	S	1	2	2	3	3	3	4	4	0	6	3	24
15	4	4	3	2	2	3	5	6	4	1	0	0	0	0	S	1	0	0	1	1	1	2	1	3	0	6	2	24
16	2	4	6	7	5	9	9	10	8	4	4	3	3	S	2	2	2	2	4	5	4	4	3	2	2	10	5	24
17	3	2	1	1	2	3	5	9	5	C	C	C	C	C	C	C	C	C	C	4	4	4	3	4	1	9	4	24
18	3	4	4	5	4	4	6	6	6	3	1	S	2	1	0	0	1	1	0	1	1	3	3	3	0	6	3	24
19	3	4	4	4	5	4	6	9	7	5	S	1	1	1	1	1	2	5	8	9	8	6	1	0	0	9	4	24
20	0	0	1	1	1	1	1	1	2	S	2	2	2	2	1	1	1	1	0	1	1	2	3	3	0	3	1	24
21	2	2	1	2	1	2	3	3	S	3	1	1	0	0	0	0	0	1	2	3	3	2	7	7	0	7	2	24
22	4	4	8	13	12	9	11	S	16	6	2	2	1	1	0	1	3	6	5	3	2	2	3	3	0	16	5	24
23	3	2	1	2	2	2	S	3	2	2	2	2	1	2	3	2	6	13	10	6	13	11	7	10	1	13	5	24
24	7	11	10	8	5	S	6	6	10	14	2	3	3	3	3	3	7	12	18	16	13	14	15	13	2	18	9	24
25	8	7	9	8	S	7	7	10	22	0	5	2	2	2	1	3	4	12	14	17	15	17	13	11	0	22	8	24
26	14	10	4	S	3	5	11	36	12	6	3	2	2	1	1	1	3	8	4	6	6	8	6	1	1	36	7	24
27	5	7	S	6	7	8	6	5	4	3	2	2	1	1	0	1	1	1	5	3	2	1	2	2	0	8	3	24
28	3	S	3	4	4	4	9	7	5	2	1	1	1	1	2	3	3	2	2	1	1	1	2	1	1	9	3	24
29	S	3	2	2	2	2	2	4	9	8	0	4	4	3	2	1	1	2	2	2	4	5	S	0	0	9	3	24
30	6	6	5	3	3	4	5	7	5	2	2	2	2	2	2	3	3	4	5	5	4	4	S	4	2	7	4	24
31	4	4	3	2	2	2	4	5	4	4	4	3	3	3	3	3	4	8	10	11	10	S	7	6	2	11	5	24
HOURLY MAX	14	12	11	13	12	9	12	36	22	14	6	5	4	3	3	3	7	13	18	17	15	17	15	13				
HOURLY AVG	4	5	4	4	3	4	5	6	5	3	2	2	1	1	1	1	2	3	4	4	4	5	5	4				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 159 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0				
NUMBER OF NON-ZERO READINGS:	621				
MINIMUM 1-HR AVERAGE:	0	ppb	@ HOUR	1 ON DAY	1
MAXIMUM 1-HR AVERAGE:	36	ppb	@ HOUR	7 ON DAY	26
MAXIMUM 24-HR AVERAGE:	9	ppb		ON DAY	24
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	744	hrs
MONTHLY CALIBRATION TIME:	10	hrs	AMD OPERATION UPTIME:	100.0	%
STANDARD DEVIATION:	4		MONTHLY AVERAGE:	3	ppb

24 HR AVERAGES October 2018



NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)




Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-NO2[ppb]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 21.65% Calm Avg: 7.00 [ppb]

Direction	0.0-12.3	12.3-24.7	24.7-37.0	>37.0	Total
N	5.0	0.0	0.0	0.0	5.0
NE	5.8	0.3	0.0	0.0	6.1
E	6.8	0.1	0.0	0.0	7.0
SE	14.8	0.3	0.0	0.0	15.1
S	2.6	0.0	0.0	0.0	2.6
SW	12.8	0.0	0.1	0.0	13.0
W	16.7	0.0	0.0	0.0	16.7
NW	13.0	0.0	0.0	0.0	13.0
Summary	77.5	0.7	0.1	0.0	78.3

% Icon Classes (ppb) 77 0.0-12.3 1 12.3-24.7 0 24.7-37.0 0 >37.0

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-NO2[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 21.65% Calm Poll Avg: 7.00[ppb]



NO2[ppb] Calibration: LICA COLD LAKE SOUTH Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



OZONE

OZONE Hourly Averages (O₃ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.																							
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.																								
DAY																																																			
1	24.2	24.7	24.7	23.3	22.1	S	18.4	18.3	18.4	20.3	21.9	26.5	29.1	30.6	29.6	30.8	29.8	24.7	20.1	21.6	26.9	28.3	26.2	25.6	18.3	30.8	24.6	24																							
2	24.4	22.7	22.8	23.2	S	24.8	24.8	21.1	23.1	27.7	27.0	29.3	30.5	32.0	33.9	33.6	32.8	32.0	30.4	25.5	25.8	26.7	25.0	23.4	21.1	33.9	27.1	24																							
3	23.3	23.2	21.3	S	22.7	23.3	20.8	21.5	24.9	28.9	31.2	31.9	32.9	34.8	34.7	34.5	33.6	31.5	27.1	17.0	14.8	7.9	16.6	21.7	7.9	34.8	25.2	24																							
4	23.9	23.2	S	24.4	25.0	23.6	23.0	23.3	24.0	24.2	23.8	25.0	26.4	28.1	29.8	33.0	37.1	34.0	28.0	26.2	28.4	28.8	29.7	29.5	23.0	37.1	27.1	24																							
5	29.1	S	26.3	24.4	22.8	20.5	12.1	12.3	17.4	20.0	22.2	28.1	31.7	31.7	32.1	32.9	34.2	33.0	29.9	28.3	26.7	19.3	13.6	16.8	12.1	34.2	24.6	24																							
6	S	10.2	14.8	24.0	24.5	24.8	25.4	24.8	25.1	25.5	25.8	26.0	25.9	26.2	26.7	25.5	24.2	23.0	20.9	18.2	16.7	16.8	20.1	S	10.2	26.7	22.5	24																							
7	12.9	15.7	22.0	22.2	22.0	22.5	21.7	21.5	23.7	24.1	23.6	23.3	23.8	24.2	24.7	24.8	25.0	25.6	26.2	27.5	28.0	28.8	S	28.7	12.9	28.8	23.6	24																							
8	29.5	28.7	26.2	26.4	26.6	26.6	25.5	28.5	30.6	30.4	30.4	30.8	31.5	31.0	30.6	30.4	31.0	31.0	31.2	30.7	23.6	S	11.5	7.8	7.8	31.5	27.4	24																							
9	3.8	2.7	4.9	5.9	6.1	1.3	1.5	4.3	9.3	15.5	25.6	30.1	30.2	30.2	30.7	30.4	30.2	24.9	15.6	5.8	S	0.9	1.9	2.5	0.9	30.7	13.7	24																							
10	3.8	0.7	1.0	1.4	2.0	2.4	0.9	2.1	15.1	27.9	27.9	28.5	30.6	30.3	30.3	29.4	29.0	25.3	13.8	S	11.0	11.4	14.7	11.5	0.7	30.6	15.3	24																							
11	7.0	7.7	16.5	20.9	20.8	18.8	16.7	15.6	14.8	18.1	21.8	24.1	24.4	24.4	24.6	24.3	23.5	24.0	S	24.2	22.4	20.3	18.6	17.4	7.0	24.6	19.6	24																							
12	16.2	16.3	15.2	9.1	6.4	2.5	2.1	5.1	6.8	5.8	7.2	9.4	16.0	19.6	20.6	23.0	28.1	S	31.1	31.7	30.1	26.8	28.9	30.3	2.1	31.7	16.9	24																							
13	28.5	28.2	27.6	26.9	27.3	26.2	25.3	26.2	27.3	27.7	28.7	29.2	30.7	31.1	31.5	32.8	S	31.6	30.7	29.1	27.3	19.5	13.8	16.8	13.8	32.8	27.1	24																							
14	12.2	9.3	14.7	16.4	14.7	14.9	9.1	10.6	20.0	21.5	26.0	28.1	29.8	30.0	29.0	S	28.9	27.0	24.0	22.5	22.1	21.1	19.8	20.3	9.1	30.0	20.5	24																							
15	24.1	24.6	27.2	30.9	30.4	28.4	25.8	24.3	26.3	30.4	33.0	34.6	35.7	36.8	S	38.9	37.7	35.1	32.4	32.0	30.9	28.4	27.8	25.6	24.1	38.9	30.5	24																							
16	26.5	23.0	19.2	16.5	8.5	4.5	1.3	1.3	7.1	18.3	20.0	C	C	C	C	C	C	C	14.6	12.3	14.5	13.6	11.3	16.8	1.3	26.5	13.5	24																							
17	13.8	6.3	2.6	2.5	1.5	7.0	7.8	4.3	21.2	29.2	34.1	37.9	S	48.4	50.5	52.4	51.6	48.3	47.8	50.6	48.5	44.1	38.5	36.8	1.5	52.4	29.8	24																							
18	29.8	28.3	27.8	30.2	32.3	33.5	30.1	31.2	31.4	35.6	38.2	S	39.0	40.6	41.2	40.9	40.1	39.1	39.5	37.9	36.4	32.0	29.5	26.4	26.4	41.2	34.4	24																							
19	25.8	24.3	24.3	24.2	20.3	20.3	16.8	10.3	18.7	24.1	S	33.2	32.6	30.7	31.8	31.6	31.4	30.4	26.4	18.4	21.5	20.1	35.5	36.7	10.3	36.7	25.6	24																							
20	35.1	34.6	33.5	32.3	32.3	31.5	29.3	27.7	26.0	S	26.5	25.4	24.8	28.5	37.8	40.4	39.9	40.8	40.6	38.8	37.0	32.2	30.0	29.6	24.8	40.8	32.8	24																							
21	27.5	29.7	30.9	29.2	29.1	27.5	26.6	26.1	S	31.9	34.3	36.0	36.3	36.2	36.0	35.9	35.5	33.9	30.3	28.5	22.7	17.3	10.8	8.3	8.3	36.3	28.7	24																							
22	9.6	8.2	3.7	0.9	0.6	1.5	0.9	S	4.2	21.5	26.7	27.9	30.7	35.3	36.2	35.4	31.7	26.0	26.5	29.2	29.1	28.2	26.0	23.9	0.6	36.2	20.2	24																							
23	24.6	26.3	27.0	28.2	27.9	27.1	S	25.0	25.9	26.5	28.3	29.8	33.4	35.7	31.7	35.8	25.3	16.2	19.6	17.3	6.9	5.9	6.5	3.3	3.3	35.8	23.2	24																							
24	4.2	1.1	1.7	1.2	1.7	S	1.0	1.5	2.6	14.9	28.4	30.7	31.2	32.0	33.7	33.5	28.9	22.4	9.9	6.5	6.8	2.7	2.6	1.8	1.0	33.7	13.1	24																							
25	2.4	1.7	0.6	0.7	S	0.7	0.5	0.8	1.6	6.1	21.4	25.6	26.6	28.0	28.4	29.4	25.5	12.6	7.1	2.0	1.0	1.7	1.0	3.2	0.5	29.4	9.9	24																							
26	2.8	6.5	13.9	S	14.6	11.1	9.6	6.0	15.4	22.4	28.1	31.1	33.2	34.9	36.0	35.4	34.5	28.5	18.7	18.2	11.1	12.1	7.1	7.2	2.8	36.0	19.1	24																							
27	7.2	6.5	S	7.4	8.6	13.1	16.8	18.3	21.6	25.2	29.8	30.9	33.1	33.0	33.2	32.8	32.3	31.3	24.4	24.8	29.2	30.3	27.4	25.5	6.5	33.2	23.6	24																							
28	24.4	S	19.0	20.2	19.1	12.2	4.8	3.7	10.0	21.7	26.9	28.1	29.6	32.7	33.2	32.4	30.2	31.0	30.9	31.0	31.0	30.4	30.2	27.6	3.7	33.2	24.4	24																							
29	S	25.5	25.8	25.1	24.6	23.7	23.1	18.1	16.0	15.3	15.0	18.6	22.3	23.1	22.1	22.4	20.6	19.5	16.8	18.4	19.4	15.3	13.5	S	13.5	25.8	20.2	24																							
30	12.7	9.9	7.7	2.8	1.8	5.9	13.2	9.2	12.6	23.2	23.2	23.6	22.6	20.3	19.3	16.9	13.8	14.8	13.3	9.8	8.5	7.3	S	6.6	1.8	23.6	13.0	24																							
31	7.3	7.3	8.0	8.7	8.0	5.9	1.5	1.5	1.6	9.4	15.6	19.3	23.7	24.8	28.0	26.8	24.2	17.1	9.7	5.0	3.6	S	10.5	7.9	1.5	28.0	12.0	24																							
HOURLY MAX	35.1	34.6	33.5	32.3	32.3	33.5	30.1	31.2	31.4	35.6	38.2	37.9	39.0	48.4	50.5	52.4	51.6	48.3	47.8	50.6	48.5	44.1	38.5	36.8																											
HOURLY AVG	17.8	16.5	17.6	17.6	17.4	16.8	14.5	14.8	17.4	22.4	25.8	27.7	29.3	30.8	31.3	31.9	30.7	28.1	24.6	23.0	22.1	19.9	18.9	18.6																											

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 82 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0			
NUMBER OF NON-ZERO READINGS:	705			
MINIMUM 1-HR AVERAGE:	0.5	ppb	@ HOUR	6 ON DAY 25
MAXIMUM 1-HR AVERAGE:	52.4	ppb	@ HOUR	15 ON DAY 17
MAXIMUM 24-HR AVERAGE:	34.4	ppb		ON DAY 18
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	744 hrs
MONTHLY CALIBRATION TIME:	7	hrs	AMD OPERATION UPTIME:	100.0 %
STANDARD DEVIATION:	10.6		MONTHLY AVERAGE:	22.3 ppb

24 HR AVERAGES October 2018



OZONE Hourly Averages (O₃ ppb)




Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-O3[ppb]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 21.42% Calm Avg: 8.01 [ppb]

Direction	0.0-17.5	17.5-35.0	35.0-52.5	>52.5	Total
N	0.1	4.8	0.0	0.0	5.0
NE	0.3	5.8	0.0	0.0	6.1
E	0.4	6.5	0.0	0.0	7.0
SE	1.7	12.5	0.9	0.0	15.0
S	0.4	1.8	0.3	0.0	2.6
SW	4.3	7.7	1.7	0.0	13.6
W	1.7	12.1	2.7	0.0	16.5
NW	1.4	9.7	1.8	0.0	12.9
Summary	10.4	60.9	7.4	0.0	78.6

% Icon	Classes (ppb)	10	61	7	0
	0.0-17.5				
	17.5-35.0				
	35.0-52.5				
	>52.5				

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-O3[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 21.42% Calm Poll Avg: 8.01[ppb]



O3[ppb] Calibration: LICA COLD LAKE SOUTH Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



PARTICULATE MATTER 2.5



PARTICULATE MATTER < 2.5 MICRONS Hourly Averages (PM_{2.5} µg/m³)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	0	0	0	0	0	1	1	1	1	1	1	X	1	1	1	1	0	1	1	1	1	0	0	1	0	0	1	1	23
2	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0
3	0	0	0	0	0	0	1	1	1	0	0	0	0	0	0	0	0	0	0	1	2	2	1	0	0	0	2	0	24
4	0	0	0	0	0	0	0	0	0	0	1	1	1	1	1	2	3	4	4	3	3	4	4	0	4	0	4	1	24
5	4	4	6	7	7	8	9	12	17	12	7	3	1	1	1	1	1	1	1	1	2	2	3	4	6	1	17	5	24
6	4	4	3	2	2	2	2	3	2	2	2	2	2	3	3	4	4	4	7	6	4	4	4	4	4	2	7	3	24
7	8	4	3	2	2	1	0	0	2	4	0	0	0	1	1	2	1	1	0	0	0	0	0	0	0	0	8	1	24
8	1	0	0	0	0	0	0	0	0	5	3	0	0	0	1	1	0	8	0	0	0	0	1	8	3	0	8	1	24
9	2	3	3	2	1	2	2	27	6	1	0	0	0	0	0	0	0	0	0	1	2	2	3	2	2	0	27	3	24
10	4	6	4	5	4	3	7	4	1	0	0	0	0	0	0	0	0	0	2	1	1	1	1	1	0	0	7	2	24
11	1	2	3	3	3	3	5	6	6	7	6	5	5	5	6	6	6	6	6	7	8	8	8	8	8	1	8	5	24
12	8	8	8	7	6	5	6	7	10	9	7	4	2	2	2	1	1	0	0	0	0	0	0	0	0	0	10	4	24
13	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
14	0	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
15	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
16	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	1	0	24
17	0	0	0	1	1	1	1	2	2	1	1	1	1	1	1	1	1	1	2	2	2	3	3	6	6	0	6	1	24
18	7	10	11	8	8	8	9	10	10	8	5	2	3	3	4	4	2	0	0	0	0	0	0	0	0	0	11	5	24
19	0	1	1	1	1	1	1	2	1	1	0	0	1	1	1	0	0	1	6	2	2	1	0	0	0	0	6	1	24
20	0	0	0	0	0	0	0	1	1	1	1	1	2	2	2	1	1	0	0	0	0	1	1	1	0	0	2	1	24
21	1	1	1	1	1	1	1	1	1	1	1	1	1	0	0	0	0	0	0	1	0	0	1	1	0	0	1	1	24
22	0	1	1	2	2	2	3	5	6	1	1	1	1	0	0	0	2	1	1	1	2	2	3	4	0	0	6	2	24
23	4	4	4	4	4	4	4	4	4	3	4	3	3	4	3	6	11	9	9	12	9	7	10	3	3	3	12	6	24
24	8	9	9	8	6	8	10	10	12	15	9	7	6	7	7	7	8	10	14	14	12	12	13	12	6	15	10	24	
25	9	9	10	9	8	7	7	11	16	15	14	13	13	14	15	14	19	18	19	22	19	26	17	15	7	26	14	24	
26	18	19	19	19	18	17	16	12	6	3	2	1	1	0	0	0	0	1	1	1	1	1	2	2	0	0	19	7	24
27	1	2	2	2	2	2	2	1	1	1	1	1	1	1	2	2	2	2	2	2	2	1	1	1	1	1	2	1	24
28	1	1	1	1	1	1	1	1	1	2	1	1	1	1	1	2	2	1	2	1	2	2	2	2	2	1	2	1	24
29	2	2	3	4	4	5	3	1	8	5	6	3	4	2	2	1	2	2	2	3	4	3	2	2	2	1	8	3	24
30	2	2	3	2	2	3	3	5	5	4	4	4	5	6	6	7	7	6	5	4	3	3	3	4	2	7	4	24	
31	6	6	6	5	5	5	5	6	6	5	4	3	2	2	1	1	1	2	4	5	5	4	3	3	1	6	4	24	
HOURLY MAX	18	19	19	19	18	17	16	27	17	15	14	13	13	14	15	14	19	18	19	22	19	26	17	15					
HOURLY AVG	3	3	3	3	3	3	3	4	4	3	3	2	2	2	2	2	3	2	3	3	3	3	3	3					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT:	1-HR	80 µg/m ³	24-HR	30 µg/m ³
----------------------	------	----------------------	-------	----------------------


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0			
NUMBER OF 24-HR EXCEEDANCES:	0			
NUMBER OF NON-ZERO READINGS:	500			
MINIMUM 1-HR AVERAGE:	0 µg/m ³ @ HOUR	0	ON DAY	1
MAXIMUM 1-HR AVERAGE:	27 µg/m ³ @ HOUR	7	ON DAY	9
MAXIMUM 24-HR AVERAGE:	14 µg/m ³		ON DAY	25
MONTHLY CALIBRATION TIME:	1 hrs	OPERATIONAL TIME:	743 hrs	
STANDARD DEVIATION:	4	AMD OPERATION UPTIME:	99.9 %	
		MONTHLY AVERAGE:	3 µg/m ³	

24 HR AVERAGES October 2018



PARTICULATE MATTER < 2.5 MICRONS Hourly Averages (PM_{2.5} µg/m³)









Wind: LICA COLD LAKE SOUTH
 Poll.: LICA COLD LAKE SOUTH-PM2.5_2[ug/m3]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 21.43%

Calm Avg: 5.45 [ug/m3]

Direction	0.0-5.6	5.6-11.2	11.2-16.8	16.8-22.4	22.4-28.0	>28.0	Total
N	4.6	0.0	0.4	0.0	0.0	0.0	5.0
NE	5.0	0.7	0.3	0.1	0.0	0.0	6.1
E	5.7	1.1	0.3	0.0	0.0	0.0	7.0
SE	14.2	0.7	0.0	0.0	0.0	0.0	14.8
S	1.2	1.4	0.0	0.0	0.0	0.0	2.6
SW	10.9	2.0	0.3	0.4	0.0	0.0	13.6
W	15.4	1.4	0.1	0.0	0.0	0.0	16.8
NW	12.4	0.3	0.0	0.0	0.0	0.0	12.7
Summary	69.3	7.4	1.3	0.5	0.0	0.0	78.6

% Icon	Classes (ug/m3(L))	69	 0.0-5.6	7	 5.6-11.2	1	 11.2-16.8	1	 16.8-22.4	0	 22.4-28.0	0	 >28.0
--------	--------------------	----	---	---	--	---	---	---	---	---	---	---	---

LICA COLD LAKE SOUTH Poll.: LICA COLD LAKE SOUTH-PM2.5_2[ug/m3(L)] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 21.43% Calm Poll Avg: 5.45[ug/m3(L)]



WIND SPEED



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018

WIND SPEED Hourly Averages (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	4.3	6.3	5.9	2.9	3.9	2.9	2.6	2.5	2.3	1.9	3.0	3.0	1.6	4.0	3.3	3.2	1.8	0.2	1.4	4.0	3.8	5.0	3.7	5.1	0.2	6.3	0.7	24
2	4.8	5.0	5.0	7.3	7.8	7.7	4.2	4.5	6.9	6.1	6.3	7.7	7.8	8.2	11.9	12.1	12.5	8.6	5.4	3.6	5.1	6.1	6.5	6.4	3.6	12.5	5.8	24
3	6.2	7.2	5.9	5.5	6.7	7.4	6.5	5.9	6.7	11.9	12.0	13.1	11.1	14.0	11.3	10.9	7.8	4.8	2.5	0.9	1.5	1.9	2.7	4.3	0.9	14.0	5.2	24
4	4.4	4.4	7.0	5.8	7.7	6.8	8.5	10.1	10.3	8.4	6.9	8.0	8.8	9.0	8.6	9.0	6.1	3.9	3.5	5.0	5.7	3.1	3.6	3.8	3.1	10.3	6.5	24
5	3.1	2.8	2.5	2.7	2.7	2.3	1.4	3.7	6.3	7.0	7.9	7.6	10.9	10.6	8.1	6.5	6.9	6.4	7.8	6.4	3.5	0.7	1.7	2.9	0.7	10.9	2.7	24
6	1.9	1.8	3.6	4.4	6.3	7.0	8.7	7.5	8.8	10.7	10.9	10.0	8.6	7.8	5.8	5.4	4.8	3.2	1.9	1.0	1.1	2.3	3.3	2.3	1.0	10.9	4.6	24
7	0.4	1.9	6.4	7.2	8.6	11.0	7.0	5.7	7.5	7.4	6.6	6.8	8.2	8.7	6.5	8.8	9.3	8.1	10.1	11.6	12.1	12.0	9.8	8.4	0.4	12.1	7.7	24
8	9.6	7.2	9.2	11.8	9.3	7.5	5.4	8.5	10.2	8.0	7.3	6.8	5.6	5.3	5.6	5.4	6.4	5.2	5.1	3.5	0.2	0.9	0.0	0.1	0.0	11.8	5.7	24
9	0.1	0.4	0.5	0.9	0.6	0.3	0.5	0.9	1.0	1.4	3.3	4.2	2.9	6.2	4.5	5.0	5.3	1.1	0.2	0.4	0.6	0.8	1.2	0.7	0.1	6.2	1.1	24
10	0.2	0.7	0.3	0.7	0.6	0.6	0.4	0.4	3.6	6.9	5.3	3.0	1.8	2.6	3.0	3.8	2.2	1.5	1.1	0.4	1.3	2.3	1.4	0.7	0.2	6.9	0.7	24
11	0.9	1.7	3.1	1.8	2.0	2.4	2.9	2.9	3.4	4.7	4.7	5.7	7.0	7.4	5.7	5.2	5.1	3.6	4.0	3.7	3.2	1.7	3.0	0.9	7.4	3.3	24	
12	3.5	5.0	3.6	1.2	1.2	0.7	1.6	1.2	3.0	4.3	5.0	6.2	9.7	13.5	15.4	16.2	17.2	13.1	16.2	13.0	9.8	9.8	13.5	11.8	0.7	17.2	6.7	24
13	11.2	11.8	12.0	13.4	11.7	13.1	10.6	10.5	12.2	15.9	16.0	11.8	13.1	12.3	13.0	12.8	9.9	7.5	7.7	6.9	4.2	2.4	2.6	4.1	2.4	16.0	9.9	24
14	3.2	2.0	4.7	4.8	4.1	2.6	1.1	1.8	3.5	5.3	9.5	10.0	9.5	9.0	9.7	6.5	8.0	5.0	5.5	5.9	6.3	6.4	6.7	6.2	1.1	10.0	5.6	24
15	10.2	10.6	9.7	10.3	9.5	7.6	7.0	8.1	10.4	11.3	12.6	13.9	13.2	15.1	17.0	16.4	12.1	6.5	4.8	5.6	5.7	3.9	9.1	8.2	3.9	17.0	9.4	24
16	5.6	3.9	3.4	2.2	1.1	1.0	1.3	0.3	1.0	3.4	4.8	4.8	4.1	3.8	3.7	4.8	2.1	1.4	2.3	1.7	2.0	2.4	1.0	1.7	0.3	5.6	1.5	24
17	2.0	1.2	1.6	1.1	0.5	4.8	2.6	3.0	8.1	8.4	10.4	11.0	13.0	11.6	11.0	10.7	10.5	9.3	9.7	8.4	9.2	6.9	4.9	5.4	0.5	13.0	6.7	24
18	5.1	4.6	5.7	6.2	7.1	6.4	5.6	6.4	6.7	9.8	8.2	11.4	10.0	13.6	14.9	14.4	13.8	13.9	8.6	7.2	6.1	6.2	6.7	5.7	4.6	14.9	7.8	24
19	6.2	6.0	6.0	5.5	6.2	6.1	2.3	5.0	7.0	8.3	8.2	14.0	13.7	12.0	10.5	6.3	7.4	3.5	2.8	2.1	2.7	1.8	8.0	11.5	1.8	14.0	1.9	24
20	10.5	10.9	7.2	6.9	10.8	8.0	5.9	7.0	0.6	5.3	7.5	10.9	10.5	8.8	16.0	13.7	9.1	11.6	11.7	8.6	8.7	4.6	6.1	7.3	0.6	16.0	4.6	24
21	5.9	7.8	7.4	8.5	7.3	8.0	8.2	8.1	9.5	12.2	12.4	13.9	14.5	17.4	14.6	13.0	9.8	5.2	5.4	4.6	2.6	2.7	0.7	0.4	0.4	17.4	7.6	24
22	0.5	0.9	1.1	1.2	0.8	2.1	0.9	1.9	1.4	3.8	5.2	5.0	5.4	6.8	6.5	6.4	3.6	3.8	4.5	4.6	5.4	5.1	3.0	3.9	0.5	6.8	2.9	24
23	4.3	5.0	6.2	7.4	6.9	7.3	5.7	6.1	7.4	7.8	7.2	8.1	6.7	5.9	5.3	5.5	2.1	2.7	2.5	1.0	0.9	1.1	0.1	1.1	0.1	8.1	4.3	24
24	0.3	1.1	0.5	0.7	0.7	0.4	0.2	0.5	0.2	2.6	3.9	4.5	6.1	5.6	4.7	5.5	3.2	2.2	0.8	1.1	0.3	0.6	1.2	0.6	0.2	6.1	1.5	24
25	0.8	0.2	0.7	0.8	0.8	0.4	0.8	0.2	0.1	1.3	3.6	3.9	7.6	5.1	5.4	2.7	1.3	0.6	0.4	1.3	0.3	1.4	0.8	0.2	0.1	7.6	1.1	24
26	2.4	1.3	2.3	1.3	2.2	0.4	1.6	6.2	5.0	6.5	8.6	9.4	10.2	11.3	10.3	7.5	5.2	3.3	1.4	1.0	1.6	0.8	0.7	0.6	0.4	11.3	3.6	24
27	0.1	0.4	2.4	1.7	3.5	6.1	6.2	6.3	6.4	7.9	10.7	11.0	12.6	12.9	13.5	11.0	8.6	4.7	0.8	4.3	5.3	6.2	3.9	5.4	0.1	13.5	5.8	24
28	6.4	4.0	2.6	5.6	5.1	0.9	1.5	1.1	1.3	4.0	4.7	5.9	6.2	7.7	10.5	7.4	4.0	5.5	6.0	5.4	5.7	6.1	7.8	5.5	0.9	10.5	3.5	24
29	6.6	5.8	7.7	6.1	3.1	3.6	8.3	8.5	8.6	9.7	10.2	10.9	10.8	11.1	10.1	9.4	6.7	6.9	6.9	8.6	10.1	9.1	7.9	7.7	3.1	11.1	5.6	24
30	6.5	4.0	2.0	1.1	1.3	2.0	1.5	0.9	3.2	3.5	5.1	3.7	4.3	4.0	3.3	3.6	4.3	2.7	1.9	1.3	1.3	1.3	1.0	2.3	0.9	6.5	2.3	24
31	2.9	1.4	2.2	1.4	1.5	0.5	0.6	1.4	0.9	4.6	7.6	7.2	5.4	4.9	5.2	3.1	3.2	1.3	0.2	0.1	1.1	2.6	1.8	0.0	0.0	7.6	1.9	24
HOURLY MAX	11.2	11.8	12.0	13.4	11.7	13.1	10.6	10.5	12.2	15.9	16.0	14.0	14.5	17.4	17.0	16.4	17.2	13.9	16.2	13.0	12.1	12.0	13.5	11.8				

STATUS FLAG CODES


C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


LAST CALIBRATION:	November 9, 2017
DECLINATION :	MAGNETIC DECLINATION 19 DEGREE EAST

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	742				
MINIMUM 1-HR AVERAGE	0.0	kph	@ HOUR	22 ON DAY	8
MAXIMUM 1-HR AVERAGE:	17.4	kph	@ HOUR	13 ON DAY	21
MAXIMUM 24-HR AVERAGE:	9.9	kph		ON DAY	13
MONTHLY CALIBRATION TIME:	0	hrs	OPERATIONAL TIME:	744 hrs	
STANDARD DEVIATION:	3.9		AMD OPERATION UPTIME:	100.0 %	
			MONTHLY AVERAGE:	1.5 kph	

24 HR AVERAGES October 2018






Wind: LICA COLD LAKE SOUTH
 Monitor: WSP [kph]
 Monthly: 18/10
 Type: WindRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 21.37%

Direction	1.8-3.5	3.5-7.0	7.0-10.5	10.5-14.0	14.0-17.5	>17.5	Total
N	0.8	1.8	0.9	0.8	0.7	0.0	5.0
NE	1.1	2.3	2.0	0.7	0.0	0.0	6.1
E	1.5	3.8	1.8	0.0	0.0	0.0	7.0
SE	4.0	6.3	3.6	0.8	0.0	0.0	14.8
S	0.9	1.5	0.1	0.0	0.0	0.0	2.6
SW	2.7	6.7	3.4	0.9	0.0	0.0	13.7
W	2.0	7.4	4.7	2.4	0.3	0.0	16.8
NW	0.5	3.0	3.2	4.8	1.2	0.0	12.8
Summary	13.6	32.7	19.8	10.5	2.2	0.0	78.7

% Icon	Classes (kph)	14		1.8-3.5	33		3.5-7.0	20		7.0-10.5	10		10.5-14.0	2		14.0-17.5	0		>17.5
--------	---------------	----	---	---------	----	---	---------	----	---	----------	----	---	-----------	---	---	-----------	---	---	-------

LICA COLD LAKE SOUTH 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 21.37% Calm Wind Avg Speed: 0.90(kph)



WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018

WIND DIRECTION Hourly Averages (WD)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	24-HOUR AVG	24-HR	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	QUADRANT	RDGS.	
DAY 1	SE	SE	SE	SE	SE	SE	SE	SE	SE	S	WSW	NW	NE	NNE	E	N	NNW	E	NNW	NNW	N	NNW	NW	NNW	ENE	24	
2	NW	NNW	NNW	N	N	NNW	NW	WNW	NW	NNW	WNW	W	WNW	W	W	W	W	W	WSW	W	WSW	WSW	WSW	WSW	WNW	24	
3	W	W	WSW	WSW	WSW	WSW	WSW	W	WNW	NW	NW	NW	NW	NW	NW	NW	NW	N	N	ENE	SE	SE	SE	SE	WNW	24	
4	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	ESE	SE	SSE	SSE	SE	SE	SE	SE	SE	SE	SE	24	
5	SSE	SE	SE	SE	S	SW	SW	WSW	WSW	W	WNW	WNW	NW	WNW	WNW	WNW	NW	N	N	NNE	SSE	ESE	E	NW	24		
6	E	ENE	ENE	ESE	ESE	ESE	ESE	E	ESE	SE	ESE	ESE	ESE	ESE	E	E	E	NE	ENE	NE	NW	NW	N	NNE	ESE	24	
7	SW	NNE	NNE	NE	NE	NE	ENE	ENE	ENE	ENE	ENE	NE	NE	NE	ENE	NE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	24	
8	ENE	ENE	NE	NE	ENE	ENE	ENE	ENE	ENE	E	ENE	ENE	ENE	NE	NE	NE	ENE	ESE	SE	S	ESE	ESE	WSW	ENE	24		
9	ENE	SW	SSW	SW	SSW	SSE	SW	NE	WNW	NW	ESE	E	ENE	NNE	NNE	NE	ENE	SSE	SE	ESE	E	ESE	SE	NE	24		
10	S	ESE	SSE	SW	WSW	S	ENE	WSW	SE	SE	SE	SSE	SSE	SE	NNE	NNE	N	NNW	SSW	SE	WSW	WSW	WSW	S	SE	24	
11	SE	ESE	SE	SSW	SE	S	S	SSE	S	SSW	S	S	SSW	SSW	SSW	SSW	SSW	SSE	S	SSW	S	SE	SSW	SW	S	24	
12	SSW	SW	SW	SW	S	SW	SW	NW	W	W	W	WNW	NNW	N	N	NNW	N	NNW	NNW	NW	NNW	NNW	NNW	NNW	NNW	NNW	24
13	NW	NW	NW	NW	NW	NW	NW	NW	NW	NW	NW	NW	NW	NW	NNW	NNW	NNW	NW	NW	NW	WNW	WSW	SW	WSW	NW	24	
14	WSW	SW	WSW	SW	SW	SW	SSW	SSE	SW	SW	SW	SW	SW	SW	SSW	SW	SSW	SW	SW	SW	SW	SW	WSW	SW	SW	24	
15	W	W	W	W	W	W	W	W	W	W	WNW	WNW	WNW	WNW	NW	NW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	24
16	NW	WNW	W	WNW	SSE	SE	SSE	SSE	SSW	WSW	WSW	WSW	W	WSW	W	SW	SW	S	SE	SSE	SE	ESE	SE	SE	WSW	24	
17	SE	SSE	SW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	SW	SW	SW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	24
18	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	W	W	WNW	WNW	W	W	WNW	WNW	NW	NNW	W	WSW	WSW	WSW	WSW	WSW	W	24	
19	WSW	WSW	W	WSW	SW	SW	WSW	WSW	WSW	W	NW	NNE	N	N	NNW	N	NE	ENE	E	E	ESE	SE	SE	SE	NW	24	
20	SE	SE	SE	SE	SE	SE	SE	SE	SW	SW	SW	SW	SW	WSW	W	W	W	WNW	WNW	W	W	SW	SW	SW	SW	24	
21	SW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	W	W	W	WNW	WNW	WNW	WNW	WNW	W	W	WSW	WSW	WSW	ESE	SW	W	24	
22	SSW	SW	E	E	SSE	WSW	E	ENE	NE	SE	SE	SE	SSE	SSE	SSE	SW	S	SE	SE	SE	SE	SE	SE	SE	SE	24	
23	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	SE	NE	SE	NNE	SE	SE	ESE	ENE	ENE	W	E	SE	24	
24	WSW	ESE	S	SW	SW	WSW	S	SW	E	ENE	E	ENE	NNE	NNE	E	E	ENE	ENE	ESE	E	WSW	E	ESE	ESE	ENE	24	
25	WSW	NNE	ESE	NE	SW	NNE	E	NNW	NE	NE	ENE	NNE	ENE	NE	N	NNW	ENE	SSE	SW	E	SE	ENE	WNW	WSW	NE	24	
26	ENE	ESE	SW	SSE	SW	ESE	SW	SW	SW	W	W	W	WSW	WSW	WSW	WSW	WSW	WSW	SE	SSW	SW	WSW	SW	SSW	WSW	24	
27	SW	ENE	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	W	WNW	W	W	WNW	NW	NNW	NW	WNW	W	W	W	W	W	WSW	W	24
28	WSW	WSW	SSW	SW	WSW	S	E	SSW	SE	SE	SE	SE	ESE	ESE	SE	SE	ESE	E	ESE	ESE	E	ESE	SE	ESE	SE	24	
29	E	E	E	E	E	NNE	NNW	NNW	NW	NW	NW	NW	WNW	WNW	WNW	WNW	WNW	W	W	NW	WNW	NW	NW	NW	NW	24	
30	NW	WNW	WNW	SSW	WSW	W	WNW	W	WSW	W	WSW	W	SW	WSW	WSW	WSW	WSW	W	WSW	SW	SSW	S	SSE	SE	WSW	24	
31	SSE	SSE	SSE	S	S	SE	ESE	SSE	SW	WSW	WSW	WSW	SW	SW	WSW	W	NNW	ESE	N	S	SSW	W	WSW	ENE	SW	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION:	November 9, 2017
DECLINATION :	MAGNETIC DECLINATION 19 DEGREE EAST


MONTHLY CALIBRATION TIME:	0	hrs	OPERATIONAL TIME:	744	hrs
STANDARD DEVIATION:	91		AMD OPERATION UPTIME:	100.0	%
			MONTHLY AVERAGE:	279	(W)

WIND DIRECTION Hourly Averages (WD)



WDR[degwdr] Station: LICA COLD LAKE SOUTH Monthly: 18/10 Type: AVG 1 Hr. [1 Hr.]

— WDR[degwdr]



STANDARD DEVIATION WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59		
DAY																										
1	8	4	7	22	5	9	9	8	12	40	34	20	63	39	32	32	44	72	32	15	11	8	14	5	24	
2	6	11	8	6	10	8	28	15	16	27	27	24	32	22	14	18	11	12	16	8	10	5	5	6	24	
3	10	9	8	7	10	7	8	15	13	18	17	10	16	12	16	14	17	10	12	41	34	21	10	6	24	
4	8	10	4	5	3	4	4	3	7	9	17	14	13	18	18	13	24	14	4	3	5	10	15	13	24	
5	20	11	21	18	21	19	44	10	16	15	19	23	18	17	20	29	30	10	13	9	18	48	24	9	24	
6	28	28	17	11	12	16	15	9	11	11	14	14	16	13	19	15	18	19	49	50	39	14	19	21	24	
7	70	46	13	8	11	4	9	13	8	14	20	16	10	13	19	10	8	7	7	7	6	9	8	9	24	
8	5	10	5	4	6	11	7	8	7	14	13	12	29	30	19	29	17	18	19	17	70	45	80	74	24	
9	72	62	53	25	51	53	54	39	53	58	33	35	52	24	31	25	19	52	74	50	42	36	21	38	24	
10	72	37	66	46	51	56	58	65	8	11	19	39	53	44	26	14	33	44	33	59	30	9	32	45	24	
11	55	21	18	31	21	27	29	17	28	23	25	28	18	19	21	22	17	15	19	19	29	13	31	12	24	
12	12	10	15	40	37	52	25	43	12	10	10	18	16	8	8	13	9	10	10	10	7	7	9	8	24	
13	7	8	7	7	9	8	7	8	8	7	8	15	14	15	12	11	12	11	4	5	16	11	14	11	24	
14	10	14	7	7	10	38	41	23	13	17	10	12	12	11	12	13	8	12	8	7	8	7	5	7	24	
15	10	5	7	5	5	7	7	5	6	11	11	14	12	14	8	9	7	7	9	9	11	10	7	5	24	
16	11	18	14	34	31	32	28	65	47	20	22	23	29	34	13	41	27	18	38	32	27	61	27	24		
17	26	30	29	23	61	5	20	20	7	8	8	9	7	10	8	6	4	4	4	6	9	11	19	7	24	
18	13	16	9	9	8	9	8	9	9	7	10	12	10	10	11	18	7	9	20	8	9	3	3	4	24	
19	3	7	4	10	4	6	60	7	8	11	24	13	15	15	16	34	7	13	17	29	28	36	4	4	24	
20	3	5	6	5	4	5	13	6	74	32	9	8	10	21	9	8	7	11	7	9	7	22	10	5	24	
21	6	6	8	3	5	5	4	3	14	7	10	10	17	12	11	11	8	15	15	8	21	36	53	68	24	
22	58	46	38	34	54	14	39	14	32	12	16	19	28	26	25	16	33	7	6	5	5	5	7	7	24	
23	11	6	5	5	3	5	8	6	6	8	15	10	15	24	24	25	31	12	12	43	43	42	71	41	24	
24	66	19	58	36	43	53	70	48	69	17	19	25	18	17	17	10	29	15	36	25	66	39	41	65	24	
25	53	72	58	56	57	69	46	74	78	48	28	42	18	25	24	29	42	61	69	45	72	39	54	76	24	
26	17	53	45	63	32	74	46	7	11	11	12	13	15	11	11	10	7	13	33	37	42	62	55	62	24	
27	77	71	35	29	15	5	6	5	6	12	11	9	10	12	10	9	8	11	62	19	10	9	25	5	24	
28	5	13	39	12	6	54	27	47	38	11	24	17	19	15	6	13	12	9	13	11	11	13	10	11	24	
29	7	8	9	10	17	13	6	9	9	20	7	8	6	7	9	8	10	9	10	10	7	8	6	5	24	
30	6	9	28	29	37	33	42	57	17	33	16	23	15	17	18	19	10	13	19	33	33	33	53	23	24	
31	31	36	28	40	38	57	57	45	59	10	10	13	13	12	19	23	18	50	76	79	43	26	32	78	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION: November 9, 2017

CALIBRATION TIME: 0 hrs OPERATIONAL TIME: 744 hrs

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)



RELATIVE HUMIDITY



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018


RELATIVE HUMIDITY Hourly Averages (RH %)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	59	60	59	60	62	64	66	66	64	60	55	47	41	40	40	40	43	49	56	58	59	60	62	65	40	66	56	24				
2	67	68	74	79	81	82	79	79	72	56	51	45	41	38	36	35	39	43	49	58	57	58	60	64	35	82	59	24				
3	66	67	70	71	68	67	68	69	65	56	50	47	42	37	34	33	32	38	45	54	59	66	63	63	32	71	55	24				
4	61	61	63	64	64	65	66	65	63	61	59	54	51	47	43	40	36	42	52	57	58	60	59	61	36	66	56	24				
5	62	67	70	74	76	78	83	84	76	67	57	49	46	44	44	43	43	49	55	58	62	72	76	74	43	84	63	24				
6	78	81	80	78	78	77	76	74	73	71	67	62	58	55	52	54	57	60	63	65	71	75	80	85	52	85	70	24				
7	87	88	86	86	86	83	85	86	79	75	71	71	71	71	69	70	68	67	68	69	72	72	74	86	67	88	77	24				
8	90	91	87	83	84	84	86	79	72	70	65	65	62	64	65	65	68	67	66	66	75	82	85	87	62	91	75	24				
9	87	86	86	85	85	84	84	85	80	72	61	49	44	46	39	36	36	44	58	66	71	75	75	78	36	87	67	24				
10	80	82	83	85	86	86	84	84	68	56	51	45	39	40	40	43	44	51	66	67	73	72	72	78	39	86	66	24				
11	79	80	79	77	76	78	79	80	80	78	75	72	68	66	64	67	69	68	70	71	75	77	79	87	64	87	75	24				
12	93	96	96	96	97	97	97	98	98	98	98	97	91	78	76	72	66	60	56	53	55	61	60	59	53	98	81	24				
13	62	61	63	64	65	66	67	66	65	64	59	55	50	46	44	40	41	46	53	57	60	70	79	78	40	79	59	24				
14	82	84	85	85	87	87	88	88	80	74	66	59	51	48	48	48	48	52	58	62	64	67	69	68	48	88	69	24				
15	65	65	64	63	64	67	70	69	61	53	48	41	36	35	37	38	43	51	55	59	68	73	76	75	35	76	57	24				
16	75	76	79	80	87	90	90	89	85	79	76	76	72	69	68	71	73	78	86	89	89	92	94	94	68	94	82	24				
17	92	93	94	94	94	92	87	90	64	43	32	27	23	19	17	15	18	22	24	26	27	31	38	39	15	94	50	24				
18	50	52	53	47	44	42	45	41	38	35	36	35	35	33	32	33	36	38	38	43	44	50	56	63	32	63	42	24				
19	66	70	71	74	78	80	82	85	71	60	53	56	60	59	54	52	53	52	55	63	58	63	50	50	50	85	63	24				
20	54	55	57	61	60	61	66	68	65	56	54	53	49	42	34	32	32	30	31	35	38	45	50	53	30	68	49	24				
21	59	59	58	60	62	63	65	66	57	49	41	35	32	32	32	32	34	40	48	54	64	73	75	77	32	77	53	24				
22	81	81	82	82	83	87	87	86	76	60	53	45	39	34	32	32	37	47	52	53	55	58	63	66	32	87	61	24				
23	65	63	62	60	61	63	65	66	60	54	46	40	33	30	38	29	45	53	51	62	69	71	75	76	29	76	56	24				
24	79	79	81	83	87	87	86	86	75	60	52	46	48	45	40	39	43	53	63	69	73	75	78	80	39	87	67	24				
25	83	84	83	84	86	87	88	88	84	75	68	66	63	57	53	52	59	71	78	80	82	80	81	82	52	88	76	24				
26	82	84	86	85	85	86	87	88	87	78	66	57	50	46	44	45	48	58	68	75	80	81	81	83	44	88	72	24				
27	82	83	84	85	84	80	80	82	78	66	59	57	56	55	54	54	55	58	62	65	63	68	76	80	54	85	69	24				
28	81	80	86	86	88	91	93	93	86	78	71	66	61	56	56	58	59	59	61	61	62	63	72	56	93	72	24					
29	77	82	85	88	89	90	95	96	97	96	94	93	93	90	86	85	86	87	88	88	86	85	86	87	77	97	89	24				
30	88	90	92	94	96	97	94	92	93	90	88	84	81	81	81	82	85	85	85	86	89	90	91	93	81	97	89	24				
31	94	94	96	96	96	96	97	97	98	92	76	67	60	59	56	60	67	76	80	82	86	93	96	56	98	82	24					
HOURLY MAX	94	96	96	96	97	97	97	98	98	98	98	97	93	90	86	85	86	87	88	89	89	92	94	96								
HOURLY AVG	75	76	77	78	79	79	80	80	75	67	61	57	53	50	49	48	50	54	59	63	66	69	72	74								

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY


MINIMUM 1-HR AVERAGE:	15	%	@ HOUR	15	ON DAY	17
MAXIMUM 1-HR AVERAGE:	98	%	@ HOUR	7	ON DAY	12
MAXIMUM 24-HR AVERAGE:	89	%			ON DAY	29
OPERATIONAL TIME:						744 hrs
AMD OPERATION UPTIME:						100.0 %
STANDARD DEVIATION:	18					MONTHLY AVERAGE: 66 %

RELATIVE HUMIDITY Hourly Averages (RH %)



AMBIENT TEMPERATURE

AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)



APPENDIX II
EQUIPMENT CALIBRATION RESULTS

SULPHUR DIOXIDE



Thermo 43i Sulphur Dioxide Analyzer Calibration

Date:	October 17, 2018	Barometer/B.P./units:	F.S. 05544 expires January 15, 2019	950	millibars
Company/Airshed:	LICA	Thermometer/Station Temp:	F.S. 170286131 expires April 19, 2019	22	°C
Location/Station Name:	Cold Lake South	Weather Conditions:	Mix of sun and clouds		
Parameter:	Sulphur Dioxide	Calibration Purpose:	routine monthly		
Start Time 24 hr. (mst):	9:03	Performed By/Reviewer:	Alex Yakupov	Rob Fisher	
End Time 24 hr. (mst):	16:18	Cal Gas Expiry Date:	October 24, 2020		
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	n/a		
Analyzer:					
Serial Number/Owner:	806528242 LICA	Range ppb:	500		
Last Calibration Date:	September 5, 2018	As Found C.F.:	1.051		
Previous C.F.:	0.997	New C.F.:	0.999		

Calibration Standards: Low Flow Meter ID/Expiry Date: Defender Low 152019 expires December 13, 2018 High Flow Meter ID/Expiry Date: Defender High 148944 expires December 13, 2018 Calibrator ID/Expiry Date: API id# 690 expires March 15, 2019 Cal Gas Cylinder I.D. #: LL 104225 Cal Gas Conc. (ppm): 49.2	Standard Calibration Points for Ranges <table border="1"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>380</td></tr> <tr><td>Mid</td><td>180</td></tr> <tr><td>Low</td><td>90</td></tr> </table>	Point	ppb	High	380	Mid	180	Low	90
Point	ppb								
High	380								
Mid	180								
Low	90								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration (ppb):	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total			
as found zero	4779	0.00	4779	0.0	0	n/a
as found high	4913	36.81	4950	365.9	348	1.051
adjusted zero	4940	0.00	4940	0.0	0	n/a
adjusted high	4909	36.97	4946	367.8	368	0.999
mid	4927	17.38	4944	173.0	176	0.983
low	4941	8.82	4950	87.7	87.3	1.004
calibrator zero	4940	0.00	4940	0.0	0	n/a
Average C.F. =						0.995

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	0.998		0.95-1.05
b (Intercept as % of full scale) =	-0.09%		± 3% F.S.
% change in C.F. from last cal =	-5.45%		± 10%

Thermo 43i Sulphur Dioxide Analyzer Calibration




As found:		As left:	
Bkg:	8.5	Bkg:	9.2
Coef:	0.944	Coef:	0.981
Pmt:	-624.2	Pmt:	-624.2
Flash:	762	Flash:	766
Internal:	30.6	Internal:	29.7
Chamber:	45.2	Chamber:	45.2
Perm Oven Gas:	45.01	Perm Oven Gas:	45.00
Perm Oven Heater:	44.26	Perm Oven Heater:	44.26
Pressure:	680.4	Pressure:	678.0
Sample Flow:	0.477	Sample Flow:	0.474
Lamp Intensity:	96	Lamp Intensity:	96
Converter:	n/a	Converter:	n/a
Converter Set:	n/a	Converter Set:	n/a
Averaging Time:	120	Averaging Time:	120
Expected Value:	243.3	Expected Value:	260.0

Comments:

The analyzer sample inlet filter was changed. The analyzer cooling fan filter(s) were cleaned. The manifold blower was found to be working normally.

— SO2[ppb]



TOTAL REDUCED SULPHUR



Thermo 450i Total Reduced Sulphur Analyzer Calibration

Date:	October 16, 2018	Barometer/B.P./units:	F.S. 05544 expires January 15, 2019	957	millibars
Company/Airshed:	LICA	Thermometer/Station Temp:	F.S. 170286131 expires April 19, 2019	22	°C
Location/Station Name:	Cold Lake South	Weather Conditions:	Cloudy/Overcast		
Parameter:	Total Reduced Sulphur	Calibration Purpose:	routine monthly		
Start Time 24 hr. (mst):	11:29	Performed By/Reviewer:	Alex Yakupov	Rob Fisher	
End Time 24 hr. (mst):	17:19	Cal Gas Expiry Date:	October 20, 2020		
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	CDNOVA / Model CDN-101 / #501		
Analyzer:					
Serial Number/Owner:	812728560 LICA	Range ppb:	100		
Last Calibration Date:	September 5, 2018	As Found C.F.:	1.009		
Previous C.F.:	1.000	New C.F.:	1.000		

Calibration Standards: Low Flow Meter ID/Expiry Date: Defender Low 152019 expires December 13, 2018 High Flow Meter ID/Expiry Date: Defender High 148944 expires December 13, 2018 Calibrator ID/Expiry Date: API id# 690 expires March 15, 2019 Cal Gas Cylinder I.D. #: EY 0001003 Cal Gas Conc. (ppm): 9.55	Standard Calibration Points for Ranges <table border="1"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19	SO2 Scrubber Check (10 minutes): Start/End Time 24 hr.: 11:36 / 11:53 SO2 Analyzer Range: 500 Target Concentration (ppb): 380 As Found Zero: 0.0 Analyzer Response (ppb): 0.0 Zero Corrected Result (ppb): 0.0
Point	ppb									
High	78									
Mid	38									
Low	19									


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration (ppb):	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total			
as found zero	7448	0.00	7448	0.0	0	n/a
as found high	7382	60.92	7443	78.2	77.5	1.009
adjusted zero	7452	0.00	7452	0.0	0	n/a
adjusted high	7381	60.84	7442	78.1	78.1	1.000
mid	7420	29.59	7450	37.9	37.7	1.006
low	7420	14.81	7435	19.0	18.9	1.007
calibrator zero	7448	0.00	7448	0.0	0	n/a
Average C.F. =						1.004

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	0.999		0.95-1.05
b (Intercept as % of full scale) =	0.10%		± 3% F.S.
% change in C.F. from last cal =	-0.86%		± 10%

Thermo 450i Total Reduced Sulphur Analyzer Calibration




As found:		As left:	
Bkg:	15.3	Bkg:	15.4
Coef:	0.910	Coef:	0.928
Pmt:	-650.8	Pmt:	-650.8
Flash:	742	Flash:	742
Internal:	33.0	Internal:	32.6
Chamber:	45.0	Chamber:	45.2
Converter Temp:	825	Converter Temp:	825
Converter Set:	825	Converter Set:	825
Perm Oven Gas:	45.01	Perm Oven Gas:	45.00
Perm Oven Htr:	44.38	Perm Oven Htr:	44.37
Pressure:	638.2	Pressure:	638.2
Sample Flow:	0.493	Sample Flow:	0.495
Lamp Intensity:	92	Lamp Intensity:	92
Averaging Time:	120	Averaging Time:	120
Expected Value:	37.7	Expected Value:	38.8

Comments:

The analyzer sample inlet filter was changed. The analyzer cooling fan filter(s) were cleaned. The manifold blower was found to be working normally.

— TRS[ppb]



TOTAL HYDROCARBON



Thermo 55i Methane/Non-Methane Analyzer Calibration

Date: October 17, 2018	Barometer/B.P./units: F.S. 05544 expires January 15, 2019 950 millibars
Company/Airshed: LICA	Thermometer/Station Temp: F.S. 170286131 expires April 19, 2019 22 °C
Location/Station Name: Cold Lake South	Weather Conditions: Mix of sun and clouds
Parameter: CH4 / NMHC / THC	Calibration Purpose: routine monthly
Start/End Time 24 hr. (mst): 9:02 / 14:14	Performed By/Reviewer: Alex Yakupov Rob Fisher
Calibration Method: Gas Dilution	Cal Gas Expiry Date: October 18, 2025

Analyzer:	Correction Factors:
Serial Number/Owner: 1180320044 LICA	Previous C.F.: As Found C.F.: New C.F.:
Measured Flow: 0.946	CH ₄ = 0.999 1.014 1.000
Last Calibration Date: September 5, 2018	NMHC = 1.001 1.006 1.000
Range ppm: 20 CH4/20 NMHC/40 THC	THC = 1.000 1.011 1.000

Calibration Standards:																	
Low Flow Meter ID/Expiry Date: Defender Low 152019 expires December 13, 2018 High Flow Meter ID/Expiry Date: Defender High 148944 expires December 13, 2018 Calibrator ID/Expiry Date: Sabio id# 11900613 expires August 22, 2019 Cal Gas Cylinder I.D. #: LL 119471 CH4 Cylinder Conc.: 599.0 207.0 = C ₂ H ₆ Cylinder Conc. CH₄ expressed as C₂H₆: 569.3 1168.3 = total CH ₄ equivalent	Standard Calibration Points for Analyzer Range of 20/20/40 ppm <table border="1" style="margin: auto;"> <thead> <tr> <th>Point</th> <th>CH4</th> <th>NMHC</th> <th>THC</th> </tr> </thead> <tbody> <tr> <td>High</td> <td>13.00</td> <td>13.00</td> <td>26.00</td> </tr> <tr> <td>Mid</td> <td>7.00</td> <td>7.00</td> <td>14.00</td> </tr> <tr> <td>Low</td> <td>3.00</td> <td>3.00</td> <td>6.00</td> </tr> </tbody> </table>	Point	CH4	NMHC	THC	High	13.00	13.00	26.00	Mid	7.00	7.00	14.00	Low	3.00	3.00	6.00
Point	CH4	NMHC	THC														
High	13.00	13.00	26.00														
Mid	7.00	7.00	14.00														
Low	3.00	3.00	6.00														

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	Correction Factors:		
	Diluent	Cal Gas	Total Flow							CH ₄	NMHC	THC
as found zero	2505	0.00	2505	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
as found high	2448	55.49	2503	13.28	12.62	25.90	13.09	12.54	25.63	1.014	1.006	1.011
adjusted zero	2501	0.00	2501	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
adjusted high	2451	55.48	2506	13.26	12.60	25.86	13.26	12.60	25.86	1.000	1.000	1.000
mid	2470	30.09	2500	7.21	6.85	14.06	7.19	6.86	14.06	1.003	0.999	1.000
low	2486	12.99	2499	3.11	2.96	6.07	3.10	2.97	6.07	1.004	0.996	1.000
calibrator zero	2505	0.00	2505	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
Average C.F. =										1.002	0.998	1.000

Linear Regression/Calibration Results:				
Correlation Coefficient =	CH ₄	NMHC	THC	LIMITS
	1.000	1.000	1.000	> or = 0.995
Slope =	1.000	1.000	1.000	0.95-1.05
b (Intercept as % of full scale) =	-0.04%	0.03%	0.00%	± 3% F.S.
% change in C.F. from last cal =	-1.55%	-0.54%	-1.05%	± 10%




As Left Instrument Diagnostics:

Interface Board Voltages: Bias Supply: 297.4 Temperatures: Detector Oven: 175.0 Filter: 175.1 Column Oven: 75.6 Internal: 35.5 Cylinder Pressures/reg.: Carrier: 2400 50 Fuel: 1000 50 Span Gas: 1400 10 Zero Air Generator: 42 Internal Pressures: Carrier: 29.4 Fuel: 44.2 Air: 30.2 FID Status: Status: LIT Counts: 32538 Flame: 348.9 Det Base: 175.1 Flame and Power Stats: Last Power On: Sep 08, 2018 Flameouts: 1 Det Oven at Start: 170.5 Col Oven at Start: 74.6 Calibration History: Time: Jan 01, 1970/ERROR Type: n/a Status: n/a Check/Adjust: n/a CH ₄ Span Conc: n/a CH ₄ SP Ratio: n/a CH ₄ RT: n/a CH ₄ PK IDX: n/a CH ₄ PK HT: n/a NM Span Conc: n/a NM SP Ratio: n/a	Calibration History cnt'd: NM Peak Area: n/a Methane Start: n/a Methane End: n/a Backflush: n/a NMHV Start: n/a NMHC End: n/a Run History>1: Date: Oct 17, 2018 Time: 09:36 CH ₄ PK HT: 0 CH ₄ RT: 12.6 CH ₄ Baseline: 2886 CH ₄ LOD: 29 CH ₄ SD: 9 CH ₄ CONC: 0.00 NM PK HT: 0 NM Peak Area: 0 NM CONC: 0.00 NM Base Start: 2869 NM Base End: 2863 NM LOD: 4 NM Start IDX: 5 NM End IDX: 59 NM Max Slope: 9.1e-01 NM Min Slope: -1.2e+00 NM PT Count: 0 Expected Values: Previous CH ₄ : 10.47 Previous NMHC: 10.95 Previous THC: 21.42 New CH ₄ : 10.21 New NMHC: 10.86 New THC: 21.07
---	---


Comments:
 The analyzer sample inlet filter was changed.
 No zero adjustment was required/made. As found zero values were copied to adjusted zero values for linearity calculation purposes.
 The analyzer cooling fan filter(s) were cleaned.
 The manifold blower was found to be working normally.

Date: October 17, 2018
Company/Airshed: LICA
Location/Station Name: Cold Lake South

Start/End Time 24 hr. (mst): 9:02 / 14:14
Calibration Purpose: routine monthly
Calibration Method: Gas Dilution



CH4[ppm] NMHC[ppm]



NITROGEN DIOXIDE



Thermo 42i NO-NO2-NOx Analyzer Calibration

Date: <u>October 17, 2018</u>	Barometer/B.P./units: <u>F.S. 05544 expires January 15, 2019</u>	<u>950</u>	millibars
Company/Airshed: <u>LICA</u>	Thermometer/Station Temp: <u>F.S. 170286131 expires April 19, 2019</u>	<u>22</u>	°C
Location/Station Name: <u>Cold Lake South</u>	Weather Conditions: <u>Mix of sun and clouds</u>		
Start/End Time 24 hr. (mst): <u>9:02 / 18:29</u>	Calibration Purpose: <u>routine monthly</u>		
G.P.T. to be used for Ozone?: <u>No</u>	Performed By/Reviewer: <u>Alex Yakupov</u>	<u>Rob Fisher</u>	
Calibration Method: <u>Gas Dilution & Gas Phase Titration</u>	Cal Gas Expiry Date: <u>October 24, 2020</u>		

Analyzer: Serial Number/Owner: <u>1505664393</u> <u>LICA</u> Last Calibration Date: <u>September 5, 2018</u> Range ppb: <u>500</u>	Correction Factors: <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th>Previous C.F.:</th> <th>As Found C.F.:</th> <th>New C.F.:</th> </tr> </thead> <tbody> <tr> <td>NO =</td> <td>0.999</td> <td>1.008</td> <td>1.000</td> </tr> <tr> <td>NO₂ =</td> <td>1.001</td> <td>1.004</td> <td>1.000</td> </tr> <tr> <td>NOx =</td> <td>0.998</td> <td>1.008</td> <td>0.999</td> </tr> </tbody> </table>		Previous C.F.:	As Found C.F.:	New C.F.:	NO =	0.999	1.008	1.000	NO ₂ =	1.001	1.004	1.000	NOx =	0.998	1.008	0.999
	Previous C.F.:	As Found C.F.:	New C.F.:														
NO =	0.999	1.008	1.000														
NO ₂ =	1.001	1.004	1.000														
NOx =	0.998	1.008	0.999														

Calibration Standards: Low Flow Meter ID/Expiry Date: <u>Defender Low 152019 expires December 13, 2018</u> High Flow Meter ID/Expiry Date: <u>Defender High 148944 expires December 13, 2018</u> Calibrator ID/Expiry Date: <u>API id# 690 expires March 15, 2019</u> Cal Gas Cylinder I.D. #: <u>LL 104225</u> Cal Gas Conc. (ppm): <u>51.5</u> <u>51.6</u>	Standard Calibration Points for a Range of: <u>500 ppb</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Point</th> <th>Target NO (ppb)</th> <th>Target NO₂ (ppb)</th> <th>Cc Ozone ?</th> </tr> </thead> <tbody> <tr> <td>High</td> <td>380</td> <td>250</td> <td>n/a</td> </tr> <tr> <td>Mid</td> <td>180</td> <td>145</td> <td>n/a</td> </tr> <tr> <td>Low</td> <td>90</td> <td>50</td> <td>n/a</td> </tr> <tr> <td>Extra Point #1</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> <tr> <td>Extra Point #2</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> </tbody> </table>	Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?	High	380	250	n/a	Mid	180	145	n/a	Low	90	50	n/a	Extra Point #1	n/a	n/a	n/a	Extra Point #2	n/a	n/a	n/a
Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?																						
High	380	250	n/a																						
Mid	180	145	n/a																						
Low	90	50	n/a																						
Extra Point #1	n/a	n/a	n/a																						
Extra Point #2	n/a	n/a	n/a																						

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated NO	Calculated NOx	Indicated NO	Indicated NOx	NO C.F.	NOx C.F.
Point	Diluent	Cal Gas	Total Flow	(ppb)	(ppb)	(ppb)	(ppb)		
as found zero	4779	0.0	4779	0	0	0.0	0.2	n/a	n/a
as found high	4913	36.8	4950	383.0	383.7	380.0	381.0	1.008	1.008
adjusted zero	4940	0.00	4940	0.0	0.0	0.0	0.0	n/a	n/a
adjusted high	4909	36.97	4946	384.9	385.7	385.0	386.0	1.000	0.999
mid	4927	17.38	4944	181.0	181.4	183.0	183.0	0.989	0.991
low	4941	8.82	4950	91.8	91.9	92.0	92.0	0.997	0.999
calibrator zero	4940	0.00	4940	0	0	0.0	0.0	n/a	n/a
Average C.F.=								0.996	0.997

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calibrator Setting	Indicated NO	Indicated NOx	Indicated NO ₂	NO drop	NO ₂ gain	NO ₂ C.F.
Point	Diluent	Cal Gas	Total Flow	volts or ppb	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)
NOx reference	4909	36.97	4946	0.0	386.0	386.0	0.0	0.0	0.0	
as found high NO2	4909	36.97	4946	250.0	140.0	385.0	245.0	246.0	245.0	1.004
adjusted high NO2	4909	36.97	4946	250.0	140.0	386.0	246.0	246.0	246.0	1.000
gpt mid	4909	36.97	4946	145.0	245.0	386.0	141.0	141.0	141.0	1.000
gpt low	4909	36.97	4946	55.0	330.0	386.0	56.0	56.0	56.0	1.000
Average NO ₂ C.F.=										1.000

Linear Regression/Calibration Results:

	NO	NOx	NO ₂	LIMITS
Correlation Coefficient =	1.000	1.000	1.000	> or = 0.995
Slope =	1.000	0.999	1.000	0.95-1.05
b (Intercept as % of full scale)=	0.10%	0.06%	0.00%	± 3% F.S.
% change in C.F. from last cal=	-0.88%	-0.97%	-0.31%	± 10%
NO2 converter efficiency			1.00	0.96 to 1.04

As found:		As left:	
NO Bkg:	4.2	NO Bkg:	4.3
NOx Bkg:	4.3	NOx Bkg:	4.6
NO Coef:	1.064	NO Coef:	1.076
NO2 Coef:	1.000	NO2 Coef:	1.000
NOx Coef:	0.999	NOx Coef:	0.999
PMT:	-854.7	PMT:	-855.1
Internal:	29.1	Internal:	27.0
Chamber:	50.1	Chamber:	50.4
Cooler:	-2.7	Cooler:	-2.9
NO2 Converter:	323.4	NO2 Converter:	326.6
NO2 Converter Set:	325.0	NO2 Converter Set:	325.0
Perm Oven Gas:	35.02	Perm Oven Gas:	35.00
Perm Oven Heater:	34.28	Perm Oven Heater:	34.25
Pressure:	181.7	Pressure:	181.1
Flow:	0.732	Flow:	0.725
Ozonator Flow:	OK	Ozonator Flow:	OK
Expected Value NO:	5	Expected Value NO:	3
Expected Value NO2:	267	Expected Value NO2:	277
Expected Value NOx:	271	Expected Value NOx:	280

Comments:

The analyzer sample inlet filter was changed.

The manifold blower was found to be working normally.




The converter cooling fan filter was cleaned.

The analyzer cooling fan filter(s) were cleaned.

Date: October 17, 2018
Company/Airshed: LICA
Location/Station Name: Cold Lake South


Start/End Time 24 hr. (mst): 9:02 / 18:29
Calibration Purpose: routine monthly
Calibration Method: Gas Dilution & Gas Phase Titration

Thermo 42i NO-NO2-NOx Analyzer Calibration



Station: LICA COLD LAKE SOUTH Daily: 18/10/17 Type: AVG 1 Min. [1 Min.]

— NOX[ppb] — NO[ppb] — NO2[ppb]



OZONE



Thermo 49i Ozone Analyzer Calibration

Date: October 16, 2018 Company/Airshed: LICA Location/Station Name: Cold Lake South Start/End Time 24 hr. (mst): 11:29 / 17:39 Ozone Calibration Method: Varying UV Lamp Power G.P.T. Date: n/a-done by Varying UV Lamp Power Analyzer: Serial Number/Owner: 700419951 LICA Last Calibration Date: September 5, 2018 Previous Cal High Point C.F.: 1.001	Barometer/B.P./units: F.S. 05544 expires January 15, 2019 957 millibars Thermometer/Station Temp: F.S. 170286131 expires April 19, 2019 22 °C Weather Conditions: Cloudy/Overcast Calibration Purpose: routine monthly Performed By/Reviewer: Alex Yakupov Rob Fisher Cal Gas Expiry Date: n/a-done by Varying UV Lamp Power Ozone Range ppb: 500 As Found C.F.: 0.992 New C.F.: 1.000
---	---

Calibration Standards: Low Flow Meter ID/Expiry Date: Defender Low 152019 expires December 13, 2018 High Flow Meter ID/Expiry Date: Defender High 148944 expires December 13, 2018 Calibrator ID/Expiry Date: Sabio id# 11900613 expires August 22, 2019 Cal Gas Cylinder I.D. #: n/a	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th>Point</th> <th>AMD Required Range of Ozone Calibration Points</th> </tr> <tr> <td>High</td> <td>300-400 ppb</td> </tr> <tr> <td>Mid</td> <td>150-200 ppb</td> </tr> <tr> <td>Low</td> <td>50-75 ppb</td> </tr> </table>	Point	AMD Required Range of Ozone Calibration Points	High	300-400 ppb	Mid	150-200 ppb	Low	50-75 ppb
Point	AMD Required Range of Ozone Calibration Points								
High	300-400 ppb								
Mid	150-200 ppb								
Low	50-75 ppb								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rate (cc/min)		Calculated Concentration:	Corrected Calculated Concentration:	Indicated Concentration:	Correction Factors:
	Total Flow @ Point Start	Total Flow @ Point Finish	(ppb)	(ppb)	(ppb)	
as found zero	5000	5000	0.0	n/a	0.0	n/a
as found high	5000	5000	380.0	380.0	383.0	0.992
adjusted zero	5000	5000	0.0	0.0	0.0	n/a
adjusted high	5000	5000	380.0	380.0	380.0	1.000
mid	5000	5000	180.0	180.0	180.0	1.000
low	5000	5000	60.0	60.0	60.0	1.000
calibrator zero	5000	5000	0.0	n/a	0.0	n/a
Average C.F.=						1.000

Linear Regression/Calibration Results:

Correlation Coefficient = <u>1.000</u>	LIMITS
Slope = <u>1.000</u>	> or = 0.995
b (Intercept as % of full scale) = <u>0.00%</u>	0.95-1.05
% change in C.F. from last cal = <u>0.88%</u>	± 3% F.S.
	± 10%

Thermo 49i Ozone Analyzer Calibration



As found:

O3 Bkg:	0.0
O3 Coef:	1.059
Photo Lamp:	9.6
O3 Lamp:	8.1
Bench:	29.8
Bench Lamp:	53.5
O3 Lamp:	67.5
Pressure:	710.1
Cell A lpm:	0.713
Cell B lpm:	0.757
O3 ppb:	7.8
Cell A ppb:	9.8
Cell B ppb:	5.8
Cell A int (Hz):	78623
Cell B int (Hz):	80195
Expected Value:	304.4

As left:


O3 Bkg:	0.0
O3 Coef:	1.051
Photo Lamp:	9.6
O3 Lamp:	5.7
Bench:	29.4
Bench Lamp:	53.5
O3 Lamp:	67.5
Pressure:	712.3
Cell A lpm:	0.714
Cell B lpm:	0.758
O3 ppb:	0.0
Cell A ppb:	9.7
Cell B ppb:	-9.7
Cell A int (Hz):	78487
Cell B int (Hz):	80097
Expected Value:	306.0

Comments:

The analyzer sample inlet filter was changed.
 The analyzer cooling fan filter(s) were cleaned.
 The manifold blower was found to be working normally.

The daily ZS check started at 13:00. The As Found High Point was started at 13:04.

O3[ppb]



PARTICULATE MATTER



Thermo 5030 SHARP Monitor Audit

Date:	October 16, 2018	Performed By/Reviewer:	Alex Yakupov	Rob Fisher
Company:	LICA	Start Time (mst):	16:22	
Station Name/Location:	Cold Lake South	End Time (mst):	16:57	
Previous Audit Date:	September 4, 2018	Calibration Purpose:	routine monthly	
Parameter:	PM 2.5	Weather Conditions:	Cloudy/Overcast	

SHARP Information and Status:				
Serial Number/Owner:	CM - 2209	LICA	Status Code	0
Approx. % Tape Reaming	50%		Error Code	0

Reference Standards/I.D./Cert. Date:	
High Flow:	Airmetrics/Chinook High Maxxam ID #2 expires April 24, 2019
Digital Manometer:	Dwyer 475 Mark III id# 3 expires January 9, 2019
Temperature:	F.S. 170286131 expires April 19, 2019
Pressure:	F.S. 05544 expires January 15, 2019

As Found Temperatures, Pressure, Humidity:						
	T1 (°C)	T2 (°C)	T3 (°C)	T4 (°C)	P3 (hPa)	RH (%)
SHARP:	9	n/a	n/a	n/a	956	n/a
Reference:	8.8	n/a	n/a	n/a	957.0	n/a
Difference:	0.2	n/a	n/a	n/a	1.0	n/a
	Temp Limit: ± 4 °C Pressure Limit: ± 13.33 hPa RH Limit: ± 2%					

As Left Temperature and Pressure (same as above if as found adequate):						
	T1 (°C)	T2 (°C)	T3 (°C)	T4 (°C)	P3 (hPa)	RH (%)
SHARP:	9	n/a	n/a	n/a	956	n/a
Reference:	8.8	n/a	n/a	n/a	957.0	n/a
Difference:	0.2	n/a	n/a	n/a	1.0	n/a
	Temp Limit: ± 4 °C Pressure Limit: ± 13.33 hPa RH Limit: ± 2%					

Mass Foil Calibration:			
	Mass Foil:	ZERO:	Span Sensitivity
Mass Foil ID:	n/a	QLF:	n/a
Spanfoil Value (µg):	n/a	CONFID:	n/a
		OLD:	n/a
		NEW:	n/a

Nephelometer Zero:				
	As Found		As Left	
Analog	n/a		n/a	
NEPH	n/a		n/a	
C14	n/a		n/a	
Conc	n/a		n/a	

Flow rate:				
	As Found		As Left	
SHARP AirFlow l/hr	1000		1000	
Reference AirFlow (l/min)	16.65		16.65	
Reference AirFlow (l/hr)	999		999	
% Difference:	0.1%		0.1%	
	Tolerance +/- 5%			
	$%D = 100 \times \frac{Q_m - Q_i}{Q_i}$			

Inlet Assembly:		
	Yes/No?	If no, explain:
PM10 Inlet Cleaned	yes	
PM2.5 Cyclone Cleaned	yes	

Pump Assembly:		
	Yes/No?	If no, explain:
Pump Inspected / Cleaned	yes	
Pump Vanes Replaced	no	not required

Comments:	

WIND SYSTEM

CALIBRATORS

Company: <u>Maxxam</u>		Operator: <u>Chris W</u>	
Calibrator:		Flow Measurement Device:	
Make/Model	<u>API 700</u>	Make/Model	<u>Mesa Defender 530</u>
Serial Number	<u>690</u>	Serial Number	<u>L-153351 H-152571</u>
Last Verification Date	<u>March 2016</u>	Temperature (°C)	<u>23.5 C</u>
NO Cylinder S/N	<u>LL108015</u>	Barometric Pressure	<u>695 mmHg</u>
NO [PPM]	<u>52.2</u>	NOx [PPM]	<u>52.3</u>
Expiry Date	<u>Oct 2020</u>		

Dilution Flow (sccm)			
Pt. #1	<u>5000</u>	Pt. #2	<u>5000</u>
		Pt. #3	<u>5000</u>
Gas Flow (sccm)			
Pt. #1	<u>80</u>	Pt. #2	<u>40</u>
		Pt. #3	<u>20</u>

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
5000	0.0	0.000	0.000	0.000	0.000	0.000	Limit ± 10%	
4959	75.0	0.789	0.791	0.793	0.000	0.793	1%	0%
4971	36.5	0.383	0.384	0.384	0.000	0.384	0%	0%
4967	18.2	0.191	0.192	0.191	0.000	0.191	0%	-1%
Absolute Average Percent Difference							0%	0%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO		LIMITS		NOx	
Correlation=	1.0000	≥ 0.990		Correlation=	1.0000
m (Slope)=	1.0054	0.90-1.10		m (Slope)=	1.0031
b (Intercept % of FS)=	-0.0583	± 3% F.S.		b (Intercept % of FS)=	-0.0795

Flow	O ₂ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4959	0.000	0.000	0.790	-0.001	0.789	NO ₂	% Diff. Limit
4959	0.500	0.497	0.293	0.493	0.786	-1%	± 10%
4959	0.275	0.273	0.517	0.269	0.787	-1%	± 10%
4959	0.100	0.102	0.688	0.099	0.787	-2%	± 10%
Absolute Average Percent Difference						1%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO₂		LIMITS	
Correlation=	1.0000	≥ 0.995	
m (Slope)=	0.9946	0.90-1.10	
b (Intercept % of FS)=	-0.1817	± 3% F.S.	

<p align="center">AENV Standards Audit Calibrator</p> <p>Make/Model <u>Teco 146i</u></p> <p>Serial/AMU Number <u>AMU 1809</u></p> <p>SRM Gas Cylinder No. <u>APEX1170572</u></p> <p>Cylinder Conc. (ppm) <u>49.99</u></p>	<p align="center">NO_x Analyzer</p> <p>Make/Model <u>Teco 42i</u></p> <p>Serial/AMU Number <u>AMU 1868</u></p> <p>Last Calibration Date <u>March 14, 2018</u></p> <p>Full Scale (ppm) <u>1.0</u></p> <p>Cylinder Gas Expiry Date <u>November 2020</u></p>
---	--

COMMENTS: Cylinder contains 47.9 ppm SO₂.

Auditor: Al Clark

Operator Signature: *Chris W*

Date: March 15, 2018

Location: McIntyre Center Edmonton

Company Maxxam Operator: Mike

Calibrator:		Flow Measurement Device:	
Make/Model	<u>Sabio</u>	Make/Model	<u>Bios Definer 220</u>
Serial Number	<u>11900613</u>	Serial Number	<u>H=128686; L=129069</u>
Last Verification Date	<u>March 16, 2018</u>	Temperature (°C)	<u>22.9 C</u>
NO Cylinder S/N	<u>LL104183</u>	Barometric Pressure	<u>698 mmHg</u>
NO [PPM]	<u>50.8</u>	NOx [PPM]	<u>50.9</u>
Expiry Date	<u>October 24, 2020</u>		

Dilution Flow (sccm)			
Pt. #1	<u>5059</u>	Pt. #2	<u>5073</u>
		Pt. #3	<u>5073</u>
Gas Flow (sccm)			
Pt. #1	<u>77.5</u>	Pt. #2	<u>38.2</u>
		Pt. #3	<u>19.1</u>

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
5124	0.0	0.0000	0.0000	0.0000	-0.0001	0.0000	Limit ± 10%	
5059	77.5	0.7782	0.7797	0.7763	0.0005	0.7767	0%	0%
5073	38.2	0.3825	0.3833	0.3794	0.0000	0.3795	-1%	-1%
5073	19.1	0.1913	0.1916	0.1904	0.0000	0.1904	0%	-1%
Absolute Average Percent Difference							1%	1%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

<u>NO</u>		<u>LIMITS</u>		<u>NOx</u>	
Correlation=	1.0000	≥ 0.990		Correlation=	1.0000
m (Slope)=	0.9975	0.90-1.10		m (Slope)=	0.9960
b (Intercept % of FS)=	-0.0616	± 3% F.S.		b (Intercept % of FS)=	-0.0661

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
5059	0.0	0.0000	0.7741	0.0000	0.7741	NO ₂	% Diff. Limit
5059	500.0	0.4918	0.2823	0.4916	0.7739	0%	± 10%
5059	275.0	0.2774	0.4967	0.2780	0.7747	0%	± 10%
5059	100.0	0.1031	0.6710	0.1032	0.7743	0%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

<u>NO₂</u>		<u>LIMITS</u>	
Correlation=	1.0000	≥ 0.995	
m (Slope)=	0.9998	0.90-1.10	
b (Intercept % of FS)=	0.0173	± 3% F.S.	

<u>AENV Standards</u>		<u>NO_x Analyzer</u>	
Audit Calibrator		Make/Model <u>Thermo 42i</u>	
Make/Model	<u>Thermo 146i</u>	Serial/AMU Number	<u>1868</u>
Serial/AMU Number	<u>1809</u>	Last Calibration Date	<u>August 16, 2018</u>
SRM Gas Cylinder No.	<u>APEX1170572</u>	Full Scale (ppm)	<u>1.0</u>
Cylinder Conc. (ppm)	<u>49.99</u>	Cylinder Gas Expiry Date	<u>November 15, 2020</u>

COMMENTS: _____

Auditor: Shea Beaton
Operator Signature: _____

Date: August 22, 2018
Location: McIntyre Center Edmonton

CALIBRATION GASES



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2017-482CGA

Company: Maxxam Operator's Name: Mike
 Cylinder #: LL104225 Concentration PPM: 49.2 Tolerance(%) 2 Certified By: Praxair
 Expiry Date: October 2020

Reference Calibrator and Gas:	Flow Measurement Device:
Make/Model: <u>R&R MFC 201</u>	Make/Model: <u>Mesa Definer 220</u>
Serial Number: <u>AMU 1690</u>	Serial Number: <u>H-133034 / L-132702</u>
Last Verification Date: <u>December 13, 2017</u>	Temp. °C: <u>23.4 C</u>
Gas Type: <u>SO2</u> Conc. <u>98.07</u>	B.P. <u>707 mmHg</u>
Cylinder Number: <u>CAL016625</u>	
Expiry Date: <u>January 2019</u>	


Reference Analyzer:
 Make/Model: Teco 43C Serial/AMU Number: 1623
 Instrument Settings: Zero: 10.0 Span: 1.006 Range: 1.0
 Last Calibration: Date: Dec12/17 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.000	X	X	X
4989	79.5	0.764	0.01594	62.755	47.9
4995	39.6	0.380	0.00793	126.136	47.9
4992	19.6	0.188	0.00393	254.694	47.9
Average Cylinder Concentration:					47.9

Previous Stated Concentration PPM: 49.2
 Percent variance from Stated: 3

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 < =5% Outside Manufacturer Tolerance. Use manufacturers concentration _____
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark Date: December 13, 2017
 Operator Signature: *Al Clark* Location: McIntyre Center Edmonton



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2017-493CGA

Company: Maxxam Operator's Name: Mike
 Cylinder #: EY0001003 Concentration PPM: 9.55 Tolerance(%) 2 Certified By: Praxair
 Expiry Date: October 2020

Reference Calibrator and Gas:
 Make/Model: Sabio 2010
 Serial Number: AMU 2092
 Last Verification Date: January 17, 2018
 Gas Type: H2S Conc. 20.43
 Cylinder Number: CAL015272
 Expiry Date: January 2019

Flow Measurement Device:
 Make/Model: Mesa Defender 530
 Serial Number: H-153961 / L-153874
 Temp. °C: 23.0 C
 B.P.: 697 mmHg

Reference Analyzer:
 Make/Model: Teco 450i Serial/AMU Number: 1980
 Instrument Settings: Zero: 12.9 Span: 0.955 Range: 0.1
 Last Calibration: Date: Jan 17/18 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.0000	0.0000	0.0000	0.0000
5051	39.6	0.0753	0.00784	127.551	9.60
5028	20.2	0.0387	0.00402	248.911	9.63
5033	10.5	0.0198	0.00209	479.333	9.49
Average Cylinder Concentration:					9.58

Previous Stated Concentration PPM: 9.55

Percent variance from Stated: 0

Meets Manufacturer Tolerance. Use manufacturers stated concentration COMMENTS: Used AEP regulator
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark

Date: January 18, 2018

Operator Signature: [Signature]

Location: McIntyre Center Edmonton



Calibration Gas Audit

NO Cylinder Gas

File No. 2017-483CGA

Company: Maxxam **Operators name:** Mike

Cylinder #: LL104225 Conc (PPM) 51.5/51.6 Tolerance (%) 2 Certified By: Praxair

Expiry Date: October 2020

Reference Calibrator and Gas:				Flow Measurement Device:	
Make/Model	<u>Teco 146i</u>			Make/Model	<u>Mesa Definer 220</u>
Serial Number	<u>AMU 1809</u>			Serial Number	<u>H-133034 / L-132702</u>
Last Verification Date	<u>December 13, 2017</u>			Temp. °C	<u>23.4 C</u>
Gas Type	<u>NO</u>	Conc.	<u>50.03</u>	B.P.	<u>707 mmHg</u>
Cylinder Number	<u>APEX 1223938</u>				
Expiry Date	<u>June 2020</u>				

Reference Analyzer:

Make/Model Teco 42i Serial/AMU Number: 1868

Instrument Settings Zero: 4.7 Span: 1.004 Range: 1.0

Last Calibration: Date: Dec12/17 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	NO	NOX			NO	NOX
5000	0.0	0.000	0.000				
4989	79.5	0.813	0.812	0.016	62.755	51.0	51.0
4995	39.6	0.407	0.406	0.008	126.136	51.3	51.2
4992	19.6	0.202	0.201	0.004	254.694	51.4	51.2
Average Cylinder Concentration:						51.3	51.1

	<u>NO</u>		<u>NOx</u>
Previous Stated Concentration PPM:	<u>51.5</u>		<u>51.6</u>
Percent variance from Stated:	<u>0</u>		<u>1</u>

Cylinder gas tolerances based on NO only

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:**

<=5% Outside Manufacturer Tolerance. Use manufacturers concentration

> 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark Date: December 13, 2017

Operator Signature: *Al Clark* Location: McIntyre Center Edmonton

***APPENDIX III
MAXIMUM INSTANTANEOUS DATA***



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	1	0	1	1	1	S	1	1	1	1	1	0	1	0	1	0	1	1	1	1	1	1	1	1	0	1	1	24
2	1	1	1	1	S	1	1	0	1	1	2	1	1	1	0	0	0	0	0	0	0	0	0	0	0	2	1	24
3	0	1	0	S	0	1	1	0	1	1	1	1	1	0	0	1	1	1	1	0	1	1	1	1	0	1	1	24
4	1	1	S	1	1	1	1	1	1	1	1	0	0	1	1	1	2	2	1	1	1	1	1	1	0	2	1	24
5	1	S	1	1	1	1	1	1	1	1	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	24
6	S	1	1	0	0	0	1	0	0	1	1	1	1	1	1	1	1	1	1	0	0	0	1	S	0	1	1	24
7	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	S	0	0	1	0	24
8	1	0	0	0	0	0	1	1	0	0	0	0	0	0	1	0	1	1	1	0	1	S	1	0	0	1	0	24
9	1	1	1	1	0	0	1	1	0	1	1	1	0	0	0	0	1	0	1	0	S	1	1	1	0	1	1	24
10	1	0	1	0	0	0	1	1	1	0	0	0	0	0	1	1	0	0	1	S	1	1	1	1	0	1	1	24
11	1	0	1	0	0	1	1	0	1	1	1	1	1	1	1	1	1	0	S	1	1	1	1	0	1	1	24	
12	0	1	1	1	0	0	0	1	0	1	1	1	1	1	1	0	0	S	1	1	1	2	2	1	0	2	1	24
13	2	2	2	3	2	2	2	2	2	1	2	1	2	1	1	1	S	1	1	1	1	1	1	0	0	3	1	24
14	0	0	1	0	0	0	0	0	1	2	2	1	1	1	1	S	1	1	1	1	1	1	1	1	0	2	1	24
15	1	1	1	1	1	0	0	1	0	1	1	1	1	0	S	1	0	0	1	1	0	1	1	1	0	1	1	24
16	1	1	2	1	1	1	1	1	1	1	1	0	0	S	1	0	1	0	0	0	0	0	0	0	0	2	1	24
17	0	0	0	0	0	1	0	1	1	C	C	C	C	C	C	C	C	2	2	1	1	0	0	0	0	2	1	24
18	0	0	1	1	1	0	0	0	0	0	0	S	0	0	0	0	1	0	0	0	0	0	0	0	0	1	0	24
19	0	0	0	0	0	0	0	0	0	0	S	1	0	0	0	1	1	0	0	0	0	0	0	0	0	1	0	24
20	0	0	0	0	0	0	0	0	0	0	S	0	1	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
21	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
22	0	0	0	0	0	0	0	S	1	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
23	0	0	0	0	0	0	S	0	0	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
24	0	0	0	0	0	S	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	1	0	24
25	0	0	0	0	S	0	0	1	2	3	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	3	0	24
26	0	0	0	S	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
27	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
28	0	S	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	1	0	24
29	S	0	0	0	0	0	0	0	0	1	3	1	0	0	0	0	0	0	0	1	1	1	1	S	0	3	1	24
30	2	2	1	0	0	1	1	1	1	0	1	0	1	1	0	0	0	0	0	0	0	0	S	0	0	2	1	24
31	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	S	0	0	1	0	24
HOURLY MAX	2	2	2	3	2	2	2	2	2	3	3	2	2	1	1	1	2	2	2	1	1	2	2	1				
HOURLY AVG	1	0	1	0	0	0	1	1	1	1	1	1	0	0	1	0	0	0	1	0	0	0	0	0				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	290
MAXIMUM INSTANTANEOUS VALUE:	3 ppb @ HOUR 3 ON DAY 13
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	8 hrs
OPERATIONAL TIME:	744 hrs
STANDARD DEVIATION:	1

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018

TOTAL REDUCED SULPHUR Instantaneous Maximum (TRS ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	1	0	1	1	0	S	1	0	0	0	0	0	0	1	1	0	0	0	0	0	1	0	0	0	0	0	1	0	24
2	1	0	0	0	S	1	0	1	0	1	1	1	1	1	0	0	0	1	1	1	1	0	0	0	0	0	1	1	24
3	0	0	0	S	0	0	0	0	1	0	1	0	0	1	0	1	0	1	1	1	0	1	0	1	0	0	1	0	24
4	1	1	S	1	0	0	1	0	0	0	1	0	0	1	0	0	1	0	0	1	1	0	1	0	0	0	1	0	24
5	0	S	1	0	0	1	0	1	0	0	0	1	1	1	1	1	0	0	1	0	0	1	1	1	1	0	1	1	24
6	S	1	1	0	0	1	1	1	1	1	1	1	1	0	1	1	1	0	0	1	0	0	1	S	0	0	1	1	24
7	0	0	1	0	0	0	0	0	1	0	1	1	1	1	0	0	1	0	1	1	1	1	1	S	0	0	1	0	24
8	0	1	0	1	0	1	1	0	1	0	0	1	1	0	0	1	1	1	1	1	1	0	S	0	1	0	1	1	24
9	1	0	1	0	0	0	0	1	1	0	1	0	1	0	0	1	1	1	1	1	0	S	1	1	0	0	1	1	24
10	0	1	1	1	0	1	1	1	1	1	1	0	1	1	1	0	0	1	1	0	S	0	1	1	0	0	1	1	24
11	1	0	1	1	0	0	1	1	1	1	1	0	0	1	1	1	1	0	S	1	1	1	1	1	1	0	1	1	24
12	0	1	0	1	1	0	0	1	1	1	1	1	1	0	0	0	1	S	0	0	1	1	1	1	0	0	1	1	24
13	1	1	0	0	0	1	1	1	0	1	1	1	0	1	1	0	S	1	0	1	1	1	1	0	0	0	1	1	24
14	0	1	1	1	1	1	1	0	0	1	0	1	0	0	1	S	0	0	1	0	1	0	1	1	0	0	1	1	24
15	1	1	1	1	1	1	1	1	1	1	1	0	1	1	S	1	1	0	0	0	0	0	1	1	1	0	1	1	24
16	1	1	1	1	1	1	1	1	1	1	1	C	C	C	C	C	C	C	1	1	1	1	1	1	1	1	1	1	24
17	1	1	1	1	1	1	1	1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
18	1	1	1	1	1	1	1	1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
19	1	1	1	1	1	1	1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
20	1	1	1	1	1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
21	1	1	1	1	1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
22	1	1	1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
23	1	1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
24	1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
25	1	1	1	1	S	1	1	1	2	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	24
26	1	1	1	S	1	1	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	24
27	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
28	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24
29	S	1	1	1	1	1	1	1	1	1	2	1	1	1	1	1	1	1	1	1	1	1	1	1	S	1	2	1	24
30	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	S	1	1	1	1	24
31	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	S	1	1	1	1	1	24
HOURLY MAX	1	1	1	1	1	1	2	1	2	2	2	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
HOURLY AVG	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	550
MAXIMUM INSTANTANEOUS VALUE:	2 ppb @ HOUR 8 ON DAY 25
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	7 hrs
STANDARD DEVIATION:	0
OPERATIONAL TIME:	744 hrs

TOTAL REDUCED SULPHUR Instantaneous Maximum (TRS ppb)



TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018

METHANE MAX Instantaneous Maximum (CH₄ ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	2.10	2.10	2.10	2.10	2.11	S	2.14	2.36	2.42	2.27	2.25	2.12	2.11	2.08	2.08	2.07	2.12	2.21	2.21	2.20	2.12	2.15	2.15	2.07	2.07	2.42	2.16	24	
2	2.07	2.08	2.14	2.10	S	2.05	2.03	2.13	2.11	2.01	2.01	2.01	2.00	2.01	1.99	2.00	2.00	1.99	2.05	2.12	2.06	2.11	2.11	2.10	1.99	2.14	2.05	24	
3	2.10	2.14	2.16	S	2.10	2.11	2.07	2.08	2.08	2.02	1.99	1.98	1.99	1.98	2.00	1.98	1.99	1.99	2.04	2.11	2.09	2.10	2.10	2.06	1.98	2.16	2.05	24	
4	2.03	2.11	S	2.03	2.02	2.02	2.03	2.02	2.05	2.07	2.08	2.03	2.03	2.03	2.04	2.04	2.07	2.14	2.24	2.30	2.19	2.20	2.25	2.24	2.02	2.30	2.10	24	
5	2.25	S	2.29	2.28	2.29	2.28	2.33	2.40	2.33	2.19	2.26	2.13	2.04	2.01	2.00	2.02	2.00	2.02	2.01	2.02	2.03	2.09	2.11	2.12	2.00	2.40	2.15	24	
6	S	2.09	2.08	2.05	2.04	2.04	2.03	2.02	2.03	2.01	2.01	2.02	2.03	2.05	2.05	2.04	2.06	2.09	2.09	2.09	2.19	2.16	2.06	S	2.01	2.19	2.06	24	
7	2.10	2.05	2.02	2.01	2.01	2.01	2.00	2.00	2.03	2.09	1.99	1.99	1.98	1.99	1.99	2.01	1.98	1.98	1.99	1.98	1.99	1.98	S	1.98	1.98	2.10	2.01	24	
8	1.98	1.98	1.99	1.99	1.98	1.99	1.99	2.00	1.98	1.98	1.99	1.98	1.99	1.97	1.98	1.97	1.98	1.97	1.99	1.98	1.99	S	2.10	2.10	1.97	2.10	1.99	24	
9	2.18	2.30	2.16	2.16	2.23	2.29	2.30	2.42	2.49	2.26	2.11	1.96	1.98	1.96	1.97	1.96	1.98	2.01	2.03	2.12	S	2.22	2.17	2.25	1.96	2.49	2.15	24	
10	2.23	2.23	2.23	2.20	2.20	2.28	2.28	2.25	2.21	2.03	2.04	2.02	2.00	2.00	1.99	1.98	1.98	2.07	2.07	S	2.07	2.13	2.25	2.28	1.98	2.28	2.13	24	
11	2.34	2.42	2.90	2.55	2.31	2.31	2.23	2.38	2.49	2.51	2.31	2.10	2.14	2.15	2.12	2.05	2.06	2.06	S	2.04	2.09	2.16	2.16	2.12	2.04	2.90	2.26	24	
12	2.14	2.14	2.17	2.38	2.38	2.46	2.51	2.37	2.26	2.29	2.30	2.26	2.06	2.00	2.02	2.00	1.98	S	1.96	1.95	1.96	1.97	1.96	1.95	1.95	1.95	2.51	2.15	24
13	1.97	1.96	1.98	2.00	1.97	1.97	1.98	1.99	1.97	1.97	1.97	1.97	1.96	1.96	1.97	1.96	S	1.97	1.99	1.98	2.01	2.06	2.05	2.23	1.96	2.23	1.99	24	
14	2.18	2.22	2.28	2.28	2.29	2.27	2.26	2.32	2.26	2.08	2.03	2.00	1.98	2.00	2.00	S	1.99	2.02	2.05	2.08	2.07	2.08	2.08	2.01	1.98	2.32	2.12	24	
15	2.05	2.10	2.08	1.99	1.98	1.99	2.02	2.01	1.99	2.02	1.92	1.92	1.94	1.92	S	1.93	1.93	1.92	1.94	1.93	1.92	1.96	1.94	1.95	1.92	2.10	1.97	24	
16	1.95	2.01	1.97	1.96	2.10	2.18	2.33	2.47	2.30	2.12	2.12	2.08	2.06	S	1.94	1.94	1.94	1.95	2.03	2.54	2.48	2.32	2.35	2.17	1.94	2.54	2.14	24	
17	2.23	2.25	2.36	2.30	2.31	2.34	2.40	2.42	2.22	C	C	C	C	C	C	1.97	1.98	2.02	2.11	2.12	2.17	2.08	2.11	2.11	1.97	2.42	2.19	24	
18	2.16	2.23	2.31	2.31	2.29	2.15	2.17	2.15	2.09	2.08	2.00	S	1.98	1.96	1.94	1.95	1.97	1.97	1.98	1.98	1.98	2.05	2.06	2.07	1.94	2.31	2.08	24	
19	2.08	2.14	2.14	2.15	2.18	2.12	2.12	2.30	2.25	2.25	S	1.97	1.97	1.97	1.97	1.97	1.97	2.02	2.02	2.04	2.04	2.08	2.03	1.99	1.97	2.30	2.08	24	
20	2.09	2.08	2.04	2.03	2.18	2.12	2.17	2.17	2.06	S	2.09	2.14	2.16	2.12	2.01	1.98	1.98	1.98	1.96	1.99	1.99	2.04	2.08	2.04	1.96	2.18	2.06	24	
21	2.03	2.01	1.99	1.99	1.99	2.01	2.08	2.07	S	2.02	1.97	1.96	1.97	1.95	1.95	1.95	1.96	1.98	2.01	2.02	2.12	2.13	2.19	2.28	1.95	2.28	2.03	24	
22	2.31	2.31	2.39	2.36	2.49	2.45	2.42	S	2.54	2.38	2.10	2.09	2.07	2.03	2.02	2.14	2.23	2.36	2.37	2.33	2.24	2.25	2.25	2.20	2.02	2.54	2.27	24	
23	2.18	2.27	2.17	2.15	2.06	2.12	S	2.18	2.17	2.07	2.06	2.05	2.05	2.05	2.02	2.01	2.03	2.23	2.21	2.41	2.28	2.32	2.56	2.77	2.01	2.77	2.19	24	
24	2.77	2.80	2.82	2.94	2.81	S	2.61	2.68	2.65	2.49	2.03	2.02	2.00	1.99	2.00	2.01	2.08	2.11	2.23	2.31	2.26	2.34	2.38	2.46	1.99	2.94	2.38	24	
25	2.42	2.48	2.49	2.57	S	2.48	2.62	2.68	2.74	2.51	2.12	1.99	1.99	1.98	1.99	1.98	2.07	2.07	2.06	2.24	2.22	2.27	2.32	2.23	1.98	2.74	2.28	24	
26	2.32	2.33	2.26	S	2.14	2.18	2.39	2.64	2.48	2.20	2.06	1.99	1.96	2.02	2.03	1.95	1.97	1.99	2.17	2.17	2.28	2.14	2.20	2.32	1.95	2.64	2.18	24	
27	2.26	2.38	S	2.55	2.60	2.39	2.22	2.12	2.11	2.07	1.99	1.99	1.96	1.94	1.93	1.94	1.94	1.96	1.98	2.00	2.00	1.95	1.97	1.97	1.93	2.60	2.10	24	
28	2.00	S	2.05	2.15	2.16	2.16	2.20	2.34	2.33	2.14	2.15	1.98	1.99	2.00	2.03	2.06	2.06	2.05	2.00	2.03	2.03	1.98	2.04	1.98	2.34	2.08	24		
29	S	1.95	1.96	1.97	1.99	1.95	1.95	1.96	1.95	1.95	1.97	1.95	1.93	1.94	2.06	1.98	2.01	1.97	1.98	1.96	1.96	2.00	1.99	S	1.93	2.06	1.97	24	
30	2.01	2.02	2.05	2.07	2.08	2.10	2.04	2.04	2.07	1.99	2.03	2.30	2.00	2.10	2.01	2.01	2.03	2.05	2.15	2.16	2.17	2.19	S	2.24	1.99	2.30	2.08	24	
31	2.32	2.32	2.29	2.24	2.25	2.18	2.33	2.32	2.24	2.21	2.17	2.04	2.03	2.05	2.01	2.08	2.02	2.05	2.08	2.18	2.22	S	2.20	2.24	2.01	2.33	2.18	24	
HOURLY MAX	2.77	2.80	2.90	2.94	2.81	2.48	2.62	2.68	2.74	2.51	2.31	2.30	2.16	2.15	2.12	2.14	2.23	2.36	2.37	2.54	2.48	2.34	2.56	2.77					
HOURLY AVG	2.17	2.19	2.20	2.20	2.19	2.17	2.21	2.24	2.23	2.15	2.07	2.04	2.01	2.01	2.00	2.00	2.01	2.04	2.07	2.11	2.11	2.12	2.14	2.16					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	706
MAXIMUM INSTANTANEOUS VALUE:	2.94 ppm @ HOUR 3 ON DAY 24
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	6 hrs
OPERATIONAL TIME:	744 hrs
STANDARD DEVIATION:	0.17

METHANE MAX Instantaneous Maximum (CH₄ ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018

NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
2	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
3	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
4	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
5	0.00	S	0.00	0.00	0.00	0.00	0.49	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.49	24	
6	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	24	
7	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.37	0.48	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.48	0.04	24	
8	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	24	
9	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.44	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.44	0.02	24	
10	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.69	S	0.00	0.00	0.00	0.00	0.00	0.69	0.03	24	
11	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
12	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
13	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
14	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
15	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
16	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.35	0.00	0.00	0.00	0.00	S	0.26	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.35	0.03	24
17	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	C	C	C	C	C	C	0.26	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.26	0.01	24
18	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
19	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
20	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
21	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
22	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
23	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.80	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.80	0.03	24
24	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.51	0.37	0.00	0.00	0.00	0.00	0.00	0.51	0.04	24	
25	0.47	0.00	0.00	0.00	S	0.00	0.00	0.56	0.66	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.31	0.00	0.00	0.00	0.00	0.66	0.09	24	
26	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
27	0.00	0.00	S	0.00	0.27	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.27	0.01	24	
28	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
29	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	24
30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	24
31	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	24
HOURLY MAX	0.47	0.00	0.00	0.00	0.27	0.00	0.49	0.56	0.66	0.00	0.00	0.00	0.00	0.00	0.37	0.48	0.00	0.80	0.69	0.37	0.31	0.00	0.00	0.00					
HOURLY AVG	0.02	0.00	0.00	0.00	0.01	0.00	0.02	0.05	0.02	0.00	0.00	0.00	0.00	0.00	0.02	0.02	0.00	0.03	0.04	0.01	0.01	0.00	0.00	0.00					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	16
MAXIMUM INSTANTANEOUS VALUE:	0.80 ppm @ HOUR 17 ON DAY 23
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	6 hrs
OPERATIONAL TIME:	744 hrs
STANDARD DEVIATION:	0.07

NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	4	2	2	4	6	S	8	6	9	5	6	3	2	2	7	3	3	9	7	5	5	2	3	3	2	9	5	24	
2	3	5	3	3	S	3	5	6	5	7	18	3	13	4	2	2	1	2	3	4	7	7	5	7	1	18	5	24	
3	6	6	7	S	7	7	10	10	6	4	3	5	2	4	1	1	2	2	5	16	17	32	14	6	1	32	8	24	
4	6	9	S	5	2	8	4	5	4	4	4	5	3	3	3	6	3	4	9	8	4	5	4	4	2	9	5	24	
5	4	S	6	5	4	5	24	12	10	9	9	7	5	3	2	2	2	4	3	3	8	4	11	11	2	24	7	24	
6	S	22	16	5	4	3	4	8	5	5	2	5	3	3	3	5	7	6	6	7	5	3	3	S	2	22	6	24	
7	9	5	3	2	2	4	5	4	10	3	2	4	3	6	2	3	11	5	4	4	3	2	S	4	2	11	4	24	
8	6	5	4	2	3	6	4	4	2	4	4	5	3	2	2	7	3	4	3	5	5	S	9	8	2	9	4	24	
9	13	17	10	12	9	21	16	23	25	15	9	2	1	4	2	4	2	8	14	35	S	45	33	25	1	45	15	24	
10	16	31	25	18	14	23	44	44	45	4	3	3	2	2	3	6	4	11	11	S	10	8	7	9	2	45	15	24	
11	20	19	11	6	7	5	8	10	8	6	6	3	4	4	4	4	3	S	4	5	5	4	3	3	3	20	7	24	
12	3	3	4	4	10	18	15	11	47	11	15	12	15	13	8	5	3	1	S	2	2	3	6	3	2	1	47	9	24
13	3	4	4	4	4	5	5	4	4	3	3	4	3	3	2	2	S	3	3	5	7	10	6	9	2	10	4	24	
14	7	8	7	8	6	11	9	20	4	5	4	3	2	3	3	S	3	3	4	4	4	5	6	6	2	20	6	24	
15	5	6	6	5	4	6	10	10	7	5	2	1	1	1	S	2	1	2	3	3	2	4	3	5	1	10	4	24	
16	3	10	7	13	10	20	24	50	45	7	8	5	6	S	4	5	10	19	18	8	9	9	8	4	3	50	13	24	
17	6	6	3	2	10	7	9	88	10	C	C	C	C	C	C	C	C	C	C	5	5	5	4	5	2	88	12	24	
18	4	6	7	7	7	7	10	7	9	8	3	S	3	2	4	2	2	2	2	2	3	5	5	6	2	10	5	24	
19	7	6	7	8	7	5	35	23	12	11	S	4	5	4	3	4	8	13	15	13	19	16	4	1	1	35	10	24	
20	1	1	3	2	1	2	6	4	4	S	4	3	7	4	3	1	2	1	1	1	3	3	3	4	1	7	3	24	
21	2	4	3	4	4	4	5	6	S	5	2	2	2	1	1	1	1	2	6	4	7	8	10	15	1	15	4	24	
22	10	10	22	31	31	20	40	S	54	42	6	4	5	3	2	2	7	10	9	9	4	5	6	6	2	54	15	24	
23	8	5	3	3	3	S	13	4	3	7	10	4	5	13	6	17	24	18	27	36	15	12	19	3	36	11	24		
24	12	20	17	30	7	S	19	29	47	60	18	5	5	6	11	14	14	22	23	24	18	32	23	28	5	60	21	24	
25	21	15	23	28	S	29	30	280	385	170	20	4	6	14	65	8	10	19	19	34	31	54	32	17	4	385	57	24	
26	21	18	9	S	6	9	349	435	121	20	9	3	7	5	3	4	2	7	13	9	11	14	13	16	2	435	48	24	
27	10	14	S	11	20	9	10	9	6	5	4	3	3	1	1	1	1	2	29	4	4	4	4	6	1	29	7	24	
28	5	S	12	5	6	12	17	23	21	6	5	4	5	2	3	5	6	6	4	3	3	3	2	4	2	23	7	24	
29	S	5	5	3	5	7	3	54	109	163	498	6	59	70	48	2	2	3	4	3	3	6	5	S	2	498	48	24	
30	6	7	7	5	5	6	12	21	8	6	3	3	9	7	8	9	5	7	6	8	5	7	S	7	3	21	7	24	
31	8	5	5	5	3	14	19	18	14	8	11	8	5	9	7	8	11	21	21	21	14	S	8	9	3	21	11	24	
HOURLY MAX	21	31	25	31	31	29	349	435	385	170	498	15	59	70	65	14	17	24	29	35	36	54	33	28					
HOURLY AVG	8	9	8	8	7	10	25	42	33	21	24	5	6	6	7	4	5	8	9	9	9	11	9	9					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	702
MAXIMUM INSTANTANEOUS VALUE:	498 ppb @ HOUR 10 ON DAY 29
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	10 hrs
OPERATIONAL TIME:	744 hrs
STANDARD DEVIATION:	36

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018

NITRIC OXIDE Instantaneous Maximum (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	1	0	0	1	1	S	2	1	5	1	2	1	1	1	1	1	2	0	0	1	0	1	0	0	0	5	1	24	
2	0	1	0	1	S	1	3	2	1	3	7	1	6	1	0	0	0	0	0	0	2	3	2	1	0	7	2	24	
3	2	2	2	S	2	2	2	2	1	1	1	4	0	2	0	0	0	0	1	4	2	9	1	1	0	9	2	24	
4	1	2	S	1	0	2	1	1	1	2	2	5	2	1	1	3	2	0	1	0	1	0	0	0	0	5	1	24	
5	0	S	1	1	0	0	9	3	4	4	4	3	1	1	0	1	1	1	1	0	2	0	1	2	0	9	2	24	
6	S	5	3	1	1	1	1	2	1	1	1	2	1	1	1	2	3	1	1	1	0	0	0	S	0	5	1	24	
7	1	1	2	1	1	1	2	2	3	2	1	2	1	2	2	1	4	2	2	2	2	2	1	S	1	1	4	2	24
8	2	2	2	1	1	2	2	1	1	2	2	2	1	1	1	3	1	2	1	3	0	S	0	1	0	3	2	24	
9	4	5	1	2	2	11	5	14	16	8	6	1	1	2	0	2	1	3	4	21	S	25	15	7	0	25	7	24	
10	4	16	12	7	4	10	32	30	32	2	1	2	1	1	1	2	1	1	1	S	1	1	0	1	0	32	7	24	
11	6	4	1	1	2	1	1	2	1	1	4	1	2	1	1	1	1	0	S	0	5	0	0	0	0	6	2	24	
12	0	0	1	0	11	9	5	30	3	5	5	6	7	4	1	1	0	S	0	0	0	0	0	0	0	30	4	24	
13	0	0	0	0	0	0	0	0	1	1	1	1	1	1	1	0	S	0	0	0	0	1	2	0	2	0	2	1	24
14	0	0	2	1	2	3	16	1	1	1	1	0	1	1	0	S	0	0	0	0	1	1	1	1	0	16	2	24	
15	1	1	1	1	1	1	2	1	2	1	1	0	0	0	S	0	0	0	0	0	0	0	0	0	0	2	1	24	
16	0	2	1	3	1	7	10	35	32	2	2	2	S	1	1	2	11	7	2	2	1	2	1	0	0	35	5	24	
17	1	1	1	1	6	2	2	37	3	C	C	C	C	C	C	C	C	C	C	0	0	0	0	0	0	37	4	24	
18	0	1	2	1	1	1	0	1	2	1	0	S	1	0	1	0	0	0	0	1	0	1	1	2	0	2	1	24	
19	2	2	2	2	0	0	19	8	4	3	S	1	3	2	1	1	3	3	2	2	5	2	1	0	0	19	3	24	
20	0	0	1	0	0	0	2	1	1	S	1	1	3	2	1	0	0	0	0	0	1	0	0	1	0	3	1	24	
21	0	2	2	2	1	1	1	1	S	1	1	0	0	0	0	1	0	1	1	1	0	2	2	3	0	3	1	24	
22	3	3	9	15	17	9	28	S	46	25	4	1	4	1	5	1	2	1	1	2	1	1	1	1	1	46	8	24	
23	2	1	1	0	1	0	S	4	1	1	3	7	2	2	4	1	6	5	2	16	15	3	2	6	0	16	4	24	
24	2	9	6	15	1	S	13	21	30	37	17	1	1	3	4	19	3	1	1	8	1	15	5	15	1	37	10	24	
25	8	8	12	18	S	21	21	106	223	495	11	2	3	90	108	2	4	3	3	15	16	34	14	4	2	495	53	24	
26	6	4	2	S	2	1	179	73	29	11	3	1	3	1	1	1	0	0	2	0	1	3	2	4	0	179	14	24	
27	1	4	S	1	10	1	2	3	1	1	1	0	1	0	0	0	0	0	11	1	2	2	1	2	0	11	2	24	
28	1	S	3	1	2	2	5	9	11	2	2	3	2	1	2	1	1	2	1	0	1	1	0	1	0	11	2	24	
29	S	1	1	1	1	2	1	73	76	172	723	1	81	78	61	0	0	1	1	1	0	0	0	S	0	723	58	24	
30	0	1	1	2	1	1	4	12	1	4	1	1	3	3	5	5	3	3	1	2	0	2	S	3	0	12	3	24	
31	2	1	1	2	1	8	11	12	10	5	5	3	3	3	3	3	3	6	5	3	1	S	0	1	0	12	4	24	
HOURLY MAX	8	16	12	18	17	21	179	106	223	495	723	7	81	90	108	19	6	11	11	21	16	34	15	15					
HOURLY AVG	2	3	3	3	3	3	12	17	18	27	28	2	5	7	7	2	1	2	2	3	2	4	2	2					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	561
MAXIMUM INSTANTANEOUS VALUE:	723 ppb @ HOUR 10 ON DAY 29
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	10 hrs
STANDARD DEVIATION:	37
OPERATIONAL TIME:	744 hrs

NITRIC OXIDE Instantaneous Maximum (NO ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	3	2	2	3	4	S	7	6	4	3	4	2	2	2	6	3	3	7	7	5	5	2	3	3	2	7	4	24	
2	3	4	3	2	S	3	4	4	4	4	11	2	7	2	1	2	1	2	3	4	6	5	5	6	1	11	4	24	
3	6	5	6	S	6	6	9	8	5	3	2	2	1	2	1	1	2	2	5	15	15	23	12	6	1	23	6	24	
4	5	7	S	4	2	6	4	4	3	2	3	2	2	2	2	3	2	3	8	8	4	4	4	4	2	8	4	24	
5	3	S	5	5	4	5	16	10	6	5	5	5	3	3	2	2	2	3	3	3	6	4	10	10	2	16	5	24	
6	S	17	13	4	3	2	3	6	4	4	1	3	2	2	2	3	5	5	6	5	3	2	S	1	17	5	24		
7	8	4	2	1	1	2	3	3	7	2	2	2	2	4	2	3	7	3	2	2	2	1	S	3	1	8	3	24	
8	4	4	2	1	2	4	3	3	1	2	2	3	2	2	2	5	3	2	3	2	4	S	9	7	1	9	3	24	
9	10	12	9	10	6	10	11	10	10	7	5	1	1	3	1	3	1	8	12	16	S	24	18	17	1	24	9	24	
10	13	16	14	12	11	14	16	13	14	3	2	2	1	2	3	4	3	10	11	S	10	8	6	8	1	16	9	24	
11	14	15	10	4	6	5	8	8	6	5	3	3	3	3	3	3	3	S	4	4	5	4	3	3	3	15	5	24	
12	3	3	4	9	7	7	8	19	8	9	8	10	6	4	4	2	1	S	2	2	3	6	3	2	1	19	6	24	
13	3	3	4	4	4	5	5	4	3	3	2	3	2	2	2	1	S	3	3	5	6	8	6	7	1	8	4	24	
14	7	8	6	6	5	9	7	10	3	3	2	2	1	2	2	S	3	3	3	4	4	4	5	6	1	10	5	24	
15	5	6	5	4	4	5	9	9	6	4	2	1	1	1	1	S	2	1	1	3	3	2	3	3	5	1	9	4	24
16	3	8	7	10	9	13	14	15	14	5	6	4	5	S	4	4	8	8	12	6	8	8	7	3	3	15	8	24	
17	5	5	2	2	4	7	8	53	7	C	C	C	C	C	C	C	C	C	C	5	5	5	4	4	2	53	8	24	
18	4	5	6	6	6	6	10	7	8	6	3	S	3	2	2	2	2	2	2	2	3	5	5	5	2	10	4	24	
19	4	5	6	6	7	5	17	15	8	8	S	3	3	3	2	3	6	10	13	12	15	14	4	1	1	17	7	24	
20	1	1	2	2	1	2	5	3	3	S	3	2	5	3	2	1	2	1	1	1	3	3	3	4	1	5	2	24	
21	2	3	2	3	4	3	4	5	S	4	2	1	1	1	1	1	1	1	6	4	6	6	10	13	1	13	4	24	
22	8	8	14	16	14	12	15	S	21	17	3	3	3	2	1	2	4	10	9	7	4	4	5	6	1	21	8	24	
23	6	4	2	2	3	3	S	9	3	3	4	4	3	4	10	5	10	19	15	14	23	14	11	13	2	23	8	24	
24	11	12	12	15	6	S	11	8	18	23	7	4	3	5	7	9	12	21	23	21	17	18	19	16	3	23	13	24	
25	13	9	12	11	S	10	10	217	306	111	10	3	4	5	14	6	9	16	17	19	18	20	18	13	3	306	38	24	
26	16	15	8	S	5	9	227	401	98	11	6	3	5	4	3	3	2	7	12	9	10	11	12	12	2	401	39	24	
27	9	12	S	9	11	9	9	8	5	4	3	3	2	1	1	1	1	2	18	4	3	2	3	4	1	18	5	24	
28	4	S	10	5	6	9	14	14	11	4	4	2	3	2	2	3	5	5	3	2	2	2	2	3	2	14	5	24	
29	S	4	3	3	4	5	3	19	49	91	243	5	35	44	30	1	2	3	3	3	3	5	S	1	243	26	24		
30	6	7	6	4	4	6	9	10	7	3	3	3	6	5	3	7	4	6	6	6	5	5	S	6	3	10	5	24	
31	6	5	4	3	3	6	8	8	5	5	5	5	3	6	5	5	8	15	17	18	13	S	8	7	3	18	7	24	
HOURLY MAX	16	17	14	16	14	14	227	401	306	111	243	10	35	44	30	9	12	21	23	21	23	24	19	17					
HOURLY AVG	6	7	6	6	5	6	16	30	22	12	12	3	4	4	4	3	4	6	8	7	7	8	7	7					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	702
MAXIMUM INSTANTANEOUS VALUE:	401 ppb @ HOUR 7 ON DAY 26
VAR-VARIOUS	
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	10 hrs
OPERATIONAL TIME:	744 hrs
STANDARD DEVIATION:	25

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018

OZONE Instantaneous Maximum (O₃ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	25.2	25.1	25.2	24.3	23.0	S	19.9	19.3	19.8	21.2	24.1	26.7	30.2	31.4	31.2	31.3	31.5	28.9	23.5	26.6	27.9	28.7	27.9	26.7	19.3	31.5	26.1	24
2	25.3	23.7	23.3	23.7	S	25.5	26.4	22.0	26.4	28.7	28.9	30.0	31.3	33.6	34.9	34.2	33.2	33.0	31.7	28.8	26.5	27.8	25.9	24.3	22.0	34.9	28.2	24
3	24.0	23.5	22.5	S	23.6	24.1	23.2	23.0	27.5	30.5	32.8	32.7	34.5	35.3	35.3	35.1	34.7	32.5	30.6	24.0	20.0	13.5	22.2	24.8	13.5	35.3	27.4	24
4	25.0	25.2	S	25.0	25.2	24.9	23.8	23.9	24.3	24.6	24.6	25.5	27.3	29.1	31.3	36.1	38.2	35.7	32.3	28.3	30.1	30.0	30.6	30.2	23.8	38.2	28.3	24
5	30.3	S	27.5	25.6	23.4	22.4	17.5	15.9	19.4	21.2	24.8	31.1	33.2	32.7	32.8	33.8	34.6	34.1	31.3	28.9	28.4	21.5	16.7	20.1	15.9	34.6	26.4	24
6	S	14.2	22.0	25.0	25.1	25.8	25.9	25.8	25.6	25.8	26.3	26.3	26.1	26.9	27.0	26.4	25.3	24.3	22.7	20.6	19.2	19.0	21.1	S	14.2	27.0	23.9	24
7	17.3	21.3	22.6	22.6	22.6	22.9	22.7	22.6	24.5	25.7	23.9	23.8	24.7	24.6	25.3	25.3	25.5	26.2	27.1	28.0	28.6	29.3	S	29.6	17.3	29.6	24.6	24
8	30.2	30.6	26.9	26.7	27.1	27.5	26.1	30.6	31.2	31.2	31.1	32.0	32.2	31.7	31.4	31.3	32.3	32.3	32.4	31.6	30.0	S	13.4	9.2	9.2	32.4	28.6	24
9	8.2	4.7	7.8	8.7	8.8	3.2	3.2	6.7	12.5	22.1	29.3	30.9	31.1	30.8	31.2	30.9	30.6	29.7	20.9	8.1	S	2.7	4.5	8.0	2.7	31.2	16.3	24
10	9.1	3.4	4.8	3.0	4.3	3.7	2.6	2.7	26.3	29.0	28.7	30.4	31.4	31.1	31.2	30.4	29.8	28.2	17.6	S	15.3	16.6	20.0	16.2	2.6	31.4	18.1	24
11	12.5	14.3	21.0	23.3	22.3	20.1	19.2	19.2	16.1	20.4	23.7	24.7	24.8	24.9	25.2	24.7	23.8	24.2	S	24.6	23.2	21.1	19.8	17.9	12.5	25.2	21.3	24
12	16.9	17.2	16.4	13.6	9.7	4.7	5.8	6.4	7.3	6.5	7.7	12.1	20.5	20.6	22.8	24.0	31.1	S	31.9	32.3	31.1	28.0	31.2	31.2	4.7	32.3	18.6	24
13	30.0	28.6	28.2	27.3	27.9	27.4	25.8	26.7	27.5	28.3	29.4	30.2	31.5	31.6	32.5	33.1	S	32.9	31.3	30.4	28.1	25.6	19.1	20.9	19.1	33.1	28.4	24
14	16.1	11.7	18.1	18.1	18.2	19.0	15.5	18.0	21.8	24.1	27.7	29.5	30.5	30.6	29.6	S	29.6	28.3	25.4	23.4	22.6	21.9	20.2	21.3	11.7	30.6	22.7	24
15	25.6	25.1	29.5	31.4	30.9	29.7	27.3	25.3	28.1	33.0	33.4	35.8	36.6	37.3	S	39.5	38.2	36.7	33.7	32.5	32.4	29.9	28.6	26.5	25.1	39.5	31.6	24
16	27.1	25.8	20.7	19.3	13.1	9.4	2.9	2.6	17.9	19.3	22.6	C	C	C	C	C	C	C	18.5	16.6	20.0	16.7	17.5	17.8	2.6	27.1	16.9	24
17	16.3	9.6	3.9	3.5	2.3	12.6	13.4	12.9	26.3	31.9	36.7	41.0	S	51.5	52.3	53.6	52.9	50.8	50.7	51.6	50.5	46.1	42.6	39.2	2.3	53.6	33.3	24
18	35.1	32.5	30.7	31.8	34.6	34.9	31.7	31.9	33.0	36.8	39.7	S	40.5	41.2	41.6	42.0	40.9	40.5	41.3	38.4	37.9	34.4	30.6	29.7	29.7	42.0	36.2	24
19	27.4	26.7	25.2	25.9	22.3	21.8	20.8	16.3	21.3	25.8	S	34.0	34.0	31.8	33.0	32.6	33.8	33.6	31.5	21.5	32.5	30.3	36.7	37.2	16.3	37.2	28.5	24
20	36.1	35.3	34.4	33.1	32.6	32.3	30.6	28.5	27.0	S	27.0	26.1	25.6	33.9	40.1	41.1	40.4	41.3	41.0	39.9	37.7	35.9	31.3	30.4	25.6	41.3	34.0	24
21	30.0	30.3	31.9	30.4	29.5	28.0	27.3	26.6	S	33.0	35.3	36.3	36.5	36.5	36.1	36.1	36.0	35.1	32.7	30.1	28.6	25.4	14.2	12.3	12.3	36.5	30.4	24
22	14.0	12.8	8.3	2.6	0.7	3.0	2.3	S	5.8	28.0	27.6	29.0	34.0	35.8	36.5	36.4	33.6	30.0	28.6	30.3	29.6	29.1	27.5	25.9	0.7	36.5	22.2	24
23	26.3	27.4	28.5	28.5	28.4	27.5	S	25.7	26.7	27.4	29.2	31.2	35.5	36.6	36.3	38.1	36.1	24.4	26.2	26.3	15.2	8.4	12.3	7.1	7.1	38.1	26.5	24
24	8.9	2.0	5.8	2.0	2.4	S	2.3	2.5	5.3	26.4	30.3	32.6	32.1	33.3	34.7	35.1	33.3	25.5	13.7	9.1	10.2	7.5	4.3	3.9	2.0	35.1	15.8	24
25	3.8	3.6	1.1	2.3	S	1.1	0.8	1.7	2.7	20.6	25.8	26.7	27.6	28.7	30.4	30.3	29.4	16.8	9.9	5.7	2.2	2.7	1.9	5.2	0.8	30.4	12.2	24
26	4.5	13.1	16.2	S	15.4	15.2	12.0	10.5	17.9	26.4	30.3	33.1	34.8	36.4	36.6	36.0	35.3	33.9	24.7	21.8	14.0	17.3	10.7	13.3	4.5	36.6	22.2	24
27	9.5	10.3	S	9.5	12.6	13.9	18.2	21.3	22.4	28.6	31.0	31.7	33.8	33.4	33.5	33.2	32.5	32.1	30.9	28.4	30.5	30.7	29.4	26.8	9.5	33.8	25.4	24
28	25.3	S	23.5	21.4	20.5	17.7	10.1	6.5	17.9	24.9	28.6	29.0	31.7	33.6	33.9	34.2	31.3	32.1	32.0	31.9	31.8	31.1	30.8	29.5	6.5	34.2	26.5	24
29	S	26.0	26.2	25.7	25.2	25.1	24.7	20.8	19.0	18.0	16.9	21.6	23.4	25.2	23.4	23.7	21.4	20.6	17.6	19.4	20.1	17.4	14.1	S	14.1	26.2	21.6	24
30	13.5	11.1	9.2	3.9	2.4	16.7	21.4	22.2	23.4	24.2	24.0	24.7	24.4	21.3	20.3	19.3	14.7	16.9	15.6	11.1	11.9	11.9	S	9.1	2.4	24.7	16.2	24
31	9.1	8.8	9.9	9.8	9.7	9.2	3.4	6.0	3.0	13.6	17.9	22.3	25.5	26.0	30.9	29.3	26.3	25.7	17.0	7.9	4.7	S	12.2	9.1	3.0	30.9	14.7	24
HOURLY MAX	36.1	35.3	34.4	33.1	34.6	34.9	31.7	31.9	33.0	36.8	39.7	41.0	40.5	51.5	52.3	53.6	52.9	50.8	50.7	51.6	50.5	46.1	42.6	39.2				
HOURLY AVG	20.1	18.8	19.7	18.9	18.8	18.9	16.9	17.5	20.3	25.2	27.3	29.0	30.5	31.9	32.5	33.0	32.3	30.6	27.5	25.3	24.7	22.8	21.3	20.8				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	705
MAXIMUM INSTANTANEOUS VALUE:	53.6 ppb @ HOUR 15 ON DAY 17
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	7 hrs
STANDARD DEVIATION:	10.0
OPERATIONAL TIME:	744 hrs

OZONE Instantaneous Maximum (O₃ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Cold Lake South Continuous Monitoring Station - October 2018

WIND SPEED Instantaneous Maximum (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	11.4	13.5	14.4	10.6	8.8	11.5	12.0	6.9	7.5	9.6	10.1	6.9	15.9	15.4	11.3	10.1	10.1	2.9	8.6	12.0	9.3	12.3	9.6	11.8	2.9	15.9	10.5	24	
2	12.5	13.5	15.4	20.3	19.6	20.3	14.4	14.2	22.7	20.1	21.4	35.4	31.5	31.8	49.6	39.5	45.9	36.2	23.5	14.0	14.0	17.4	15.9	15.7	12.5	49.6	23.5	24	
3	17.4	21.3	19.3	17.9	18.2	23.3	19.6	19.6	20.1	34.1	34.9	30.5	33.0	35.4	42.7	45.9	23.5	16.6	12.0	11.5	12.0	17.4	10.8	12.0	10.8	45.9	22.9	24	
4	13.0	13.5	15.9	12.8	16.2	16.4	22.7	20.3	22.5	24.8	18.4	21.5	23.2	24.5	22.5	28.6	27.4	12.7	8.4	11.5	13.0	12.3	12.3	13.5	8.4	28.6	17.8	24	
5	13.2	10.1	8.8	10.5	12.0	12.7	12.7	12.3	15.9	18.6	22.0	23.7	33.7	29.6	22.1	22.4	25.4	15.9	20.8	17.4	14.1	5.9	13.0	12.7	5.9	33.7	16.9	24	
6	11.8	12.3	14.7	14.2	21.0	20.8	28.6	22.5	27.5	30.8	32.0	31.3	34.0	24.9	19.6	14.7	16.9	9.1	8.1	7.1	3.7	5.9	12.7	13.2	3.7	34.0	18.2	24	
7	5.5	13.5	18.1	22.7	22.7	26.2	20.3	17.1	22.7	20.8	20.8	25.9	21.0	28.2	23.3	24.2	24.2	22.3	28.6	39.1	30.8	34.7	27.9	27.1	5.5	39.1	23.7	24	
8	28.6	23.7	24.7	30.5	27.6	26.6	21.0	23.7	27.0	22.3	23.5	19.6	19.8	20.3	19.1	19.8	17.6	17.6	19.7	12.8	2.1	7.4	22.7	19.3	2.1	30.5	20.7	24	
9	8.9	9.6	17.4	21.3	15.2	7.4	18.6	10.6	10.6	13.0	14.0	16.7	19.6	20.8	17.4	15.2	21.5	8.5	6.2	2.3	2.5	3.3	7.1	9.3	2.3	21.5	12.4	24	
10	17.4	13.8	8.8	18.7	23.2	20.1	20.1	17.6	12.6	16.7	15.9	16.2	13.2	14.4	9.8	11.1	7.9	5.9	3.3	2.6	11.0	10.8	12.7	10.1	2.6	23.2	13.1	24	
11	11.0	12.0	9.1	12.7	10.3	10.8	10.3	11.3	11.3	14.6	17.4	21.5	22.0	22.5	30.5	23.2	22.5	20.1	16.6	17.4	13.7	10.1	9.8	11.0	9.1	30.5	15.5	24	
12	12.0	13.5	12.3	11.3	11.0	11.0	6.4	11.0	9.8	14.7	13.5	17.4	38.3	42.7	58.1	54.1	55.9	41.8	57.9	48.7	30.8	31.0	52.5	36.3	6.4	58.1	28.8	24	
13	31.3	32.3	29.1	32.5	29.6	31.5	31.8	28.3	32.5	40.1	40.1	36.7	35.7	34.0	38.8	37.7	31.8	23.0	19.6	18.1	13.7	12.5	10.4	15.7	10.4	40.1	28.6	24	
14	14.0	12.1	14.7	14.7	12.5	16.2	3.1	11.0	13.2	18.1	25.2	26.9	27.9	25.5	34.5	25.2	26.9	26.2	16.2	19.7	20.9	17.6	19.3	20.4	3.1	34.5	19.2	24	
15	39.9	34.7	33.0	31.4	31.8	22.0	17.6	22.2	34.2	34.9	35.5	38.3	41.8	48.6	45.9	44.4	34.9	19.1	16.9	18.1	14.9	11.5	26.2	20.5	11.5	48.6	29.9	24	
16	17.9	9.3	10.1	8.4	3.6	10.8	6.9	11.0	9.8	9.6	13.1	15.2	14.1	12.3	13.5	11.5	7.1	4.6	6.7	6.9	8.4	8.4	7.4	9.9	3.6	17.9	9.8	24	
17	8.6	6.4	11.1	3.8	11.0	10.5	11.5	16.9	22.7	25.2	34.9	32.7	33.5	34.7	32.9	33.0	26.5	24.9	25.9	25.0	24.7	17.9	18.8	13.5	3.8	34.9	21.1	24	
18	16.7	13.3	14.2	16.4	22.1	18.8	15.4	18.6	18.8	30.3	23.5	37.1	34.2	43.2	58.6	43.9	35.4	41.0	38.5	20.5	18.2	14.4	14.2	12.5	12.5	58.6	25.8	24	
19	12.3	14.8	13.0	12.6	14.2	14.1	15.2	15.9	19.8	20.3	23.5	50.6	38.1	44.7	31.6	22.5	16.9	15.1	7.6	12.5	11.0	12.5	21.0	27.9	7.6	50.6	20.3	24	
20	23.0	25.3	21.8	19.1	25.2	21.8	15.4	17.2	15.2	18.6	25.2	29.6	28.3	37.4	53.5	57.8	37.9	41.3	46.4	26.2	23.7	19.6	16.7	18.9	15.2	57.8	27.7	24	
21	16.6	23.7	20.5	20.8	21.0	20.4	18.8	19.8	34.5	37.1	37.4	49.6	53.7	56.6	54.0	37.9	29.3	15.3	15.9	12.1	7.5	12.0	4.2	10.8	4.2	56.6	26.2	24	
22	12.5	4.7	17.6	11.3	10.9	12.3	16.7	18.1	5.1	11.8	17.9	19.6	23.5	28.1	24.0	21.0	16.2	10.1	10.5	11.9	12.5	14.7	10.6	10.5	4.7	28.1	14.7	24	
23	13.0	12.5	17.9	19.1	16.9	18.6	15.7	16.9	18.1	17.5	20.8	20.5	19.9	20.3	19.9	19.8	8.6	10.1	8.1	4.2	3.0	10.3	2.1	10.8	2.1	20.8	14.4	24	
24	2.4	3.2	2.5	2.6	2.4	2.5	2.5	2.5	3.0	10.3	10.5	14.0	15.2	15.5	13.2	15.7	10.1	6.2	3.5	3.1	2.8	3.2	4.9	4.2	2.4	15.7	6.5	24	
25	4.8	4.9	5.5	6.8	4.8	6.2	4.7	28.4	23.3	20.3	14.1	14.2	18.1	15.2	14.0	9.3	7.1	5.8	5.4	6.0	5.9	7.6	8.8	7.6	4.7	28.4	10.4	24	
26	6.2	9.6	11.3	10.3	7.9	8.6	9.1	16.8	16.0	25.2	33.0	35.7	35.7	44.4	30.3	23.6	15.2	9.7	10.5	11.0	6.3	7.1	3.8	5.7	3.8	44.4	16.4	24	
27	2.8	5.8	13.2	7.2	14.4	15.9	17.1	16.7	16.7	42.9	46.9	32.3	58.6	55.9	63.0	28.1	22.5	15.2	7.0	15.4	14.9	18.4	12.0	11.8	2.8	63.0	23.1	24	
28	15.3	13.5	11.0	15.9	11.8	5.2	3.3	5.4	6.7	10.7	14.2	16.5	19.3	23.2	28.3	24.7	15.5	19.7	19.6	22.3	19.6	25.4	22.4	16.7	3.3	28.3	16.1	24	
29	16.9	16.4	22.7	20.3	13.5	9.8	23.7	20.2	21.0	24.2	25.9	27.6	28.6	28.8	29.3	27.4	20.8	19.1	18.8	25.2	25.9	22.0	19.1	18.9	9.8	29.3	21.9	24	
30	14.2	10.4	8.1	4.0	3.7	6.3	9.1	5.4	11.3	11.8	15.7	12.7	12.3	9.1	12.7	14.4	10.3	7.9	6.3	4.7	5.0	5.3	7.6	3.7	3.7	15.7	9.2	24	
31	10.1	6.7	8.2	7.4	6.7	5.2	3.5	6.4	5.7	14.7	18.4	15.7	15.9	13.3	15.7	11.3	9.3	5.2	4.2	4.1	3.6	9.7	7.5	2.3	2.3	18.4	8.8	24	
HOURLY MAX	39.9	34.7	33.0	32.5	31.8	31.5	31.8	28.4	34.5	42.9	46.9	50.6	58.6	56.6	63.0	57.8	55.9	41.8	57.9	48.7	30.8	34.7	52.5	36.3					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

MAXIMUM INSTANTANEOUS VALUE:	63.0	kph	@ HOUR	14	ON DAY	27	
OPERATIONAL TIME:						744	hrs

WIND SPEED Instantaneous Maximum (WS kph)



***APPENDIX IV
REPORT CERTIFICATION FORM***

Report Certification Form

Alberta Airshed (if applicable)	EPA Approval or Code of Practice Registration # (if applicable)
YES	NA
Company Name (if applicable)	Industrial Operation Name (if applicable)
LAKELAND INDUSTRY & COMMUNITY ASSOCIATION	COLD LAKE SOUTH CONTINUOUS MONITORING STATION
Name of the Representative of the Person Responsible	Position / Title of the Representative of the Person Responsible
Mike Bisaga	Environment Monitoring Program Manager
Is an External Party Certifying the Report?	
<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	
Name of External Person Certifying the Report	Position / Title of External Person Certifying the Report
Wunmi Adekanmbi	Project Team Lead, Customer Service - Air Services
Company Name for External Person Certifying the Report	Identification of Qualifications / Professional Designations of the External Person Certifying the Report
Maxxam Analytics, A Bureau Veritas Group Company	M.Sc., EPt., PMP

Maxxam Analytics is the designated contractor conducting monitoring and reporting activities. I certify that the submitted data has been (a) reviewed and validated as per the AMD Chapter 6: Ambient Data Quality. I certify that the submitted report (b) accurately reflects the monitoring results and reporting timeframe and (c) meets the specified analysis, summarization and reporting requirements as per the AMD Chapter 9: Reporting.



Signature of the External Person Certifying the Report

23-Nov-2018

Report Issued Date (dd-mon-yyyy)

APPENDIX V
DATA VALIDATION CERTIFICATION FORM



Validation Certificate Form

Client: <u>Lakeland Industry & Community Association</u>	Project #: <u>2833-2018-10-1-C</u>
Site: <u>Cold Lake South Continuous Monitoring Station</u>	Contact: <u>Mike Bisaga</u>

Level 0 Preliminary Verification	<u>Maram Ghaleb</u>	Date <u>21-Nov-2018</u>
Level 1 Primary Validation	<u>Maram Ghaleb</u>	Date <u>21-Nov-2018</u>
Level 2 Final Validation	<u>Maram Ghaleb</u>	Date <u>23-Nov-2018</u>
Level 3 Independent Data Review	<u>MSB</u>	Date <u>23-Nov-2018</u>
Post-Final Validation	<u>NA</u>	Date <u>NA</u>

Notes
The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. This validation is performed on an annual basis.

Alberta Environment and Parks (AEP)
Air.Reporting@gov.ab.ca

December 10, 2018

Subject: Monthly Report Submission for the LICA Maskwa station

Lakeland Industry & Community Association (LICA) is pleased to submit the ambient air monitoring monthly report for the LICA Maskwa AQM Station in the month of October 2018.

The air monitoring program consists of continuous air monitoring results for Sulphur Dioxide (SO₂), Hydrogen Sulphide (H₂S), Total Hydrocarbon (THC), Oxides of Nitrogen (NO_x), Nitric Oxides (NO), Nitrogen Dioxide (NO₂), Relative Humidity (RH), Barometric Pressure (BP), Precipitation, Ambient Temperature (AmbTPX), Wind Speed (WS), Wind Direction (WD) and Standard Deviation Wind Direction (STDWD).

Sampling Program	Monitoring Activities Conducted By	Sample Analysis Conducted By	Data/Report Review and Prepared By	Electronic Submission Conducted By
Continuous ambient air	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics

All data collected in October was compliant with the requirements outlined in the Air Monitoring Directive 2016 (AMD 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement. The operational uptime was 7.9% for the CH₄ and NMHC channels. The Thermo 55i Methane/Non-Methane Hydrocarbon analyzer was removed from the site, and a Thermo 51C THC analyzer was installed. No CH₄ and NMHC data was collected on the 51C analyzer. The AMD's monthly data completeness criteria (Chapter 6, Section 4.1.3, DQ 4-C) was not applicable to these channels in the October monitoring period.

THC/CH₄/NMHC: The Thermo 55i Methane/Non-Methane analyzer (s/n: 1236656107), was replaced with an alternate model, Thermo 51C Total Hydrocarbon analyzer (s/n: 46609738), On October 3. The Thermo 55i was removed from the site to be installed at LICA Bonnyville-East station. No Methane and Non-Methane Hydrocarbon was collected on the Thermo 51C analyzer.

Precipitation: configuration error occurred during the datalogger upgrade on May 31 and was corrected on November 1. Consequently, data collected during this time frame, at rates greater than or equal to 0.3 millimeter /minute, might be compromised and should therefore be used with caution.

Should you have any questions, please don't hesitate to contact us.



Lakeland Industry & Community Association
5107 50 St
Bonnyville, AB T9N 2J7

Respectfully,

A handwritten signature in blue ink that reads "Michael Bisaga".

Michael Bisaga
Technical Program Manager
Lakeland Industry & Community Association
780-266-7068
monitoring@lica.ca

A handwritten signature in blue ink that reads "Lily Lin".

Lily Lin
Data & Reporting Specialist
587-225-2248
monitoring@lica.ca



MAXXAM ANALYTICS
#1 2080 39 Ave. NE, Calgary, AB
T2E 6P7

maxxam.ca
Toll Free 800-386-7247
Fax 403-219-3673

AMBIENT AIR MONITORING MONTHLY DATA REPORT
LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
MASKWA CONTINUOUS MONITORING STATION

JOB #: 2833-2018-10-30-C

October 2018

Prepared for:

LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
5107 50 ST.
BONNYVILLE, ALBERTA
T9N 2J7

Attention: MIKE BISAGA

DATE: **November 29, 2018**

Prepared by: *Maram Ghaleb*

Maram Ghaleb, B.Sc.
Project Manager, Customer Service, Air Services

Reviewed by: *Wunmi Adekanmbi*

Wunmi Adekanmbi, M.Sc., EPt, PMP
Project Team Lead, Customer Service, Air Services

SUMMARY

In October 2018, Maxxam Analytics was contracted to manage the ambient air quality monitoring and maintenance activities at the Maskwa Continuous Monitoring Station, near Bonnyville, Alberta. The monitoring station provides continuous meteorological measurements and air quality data for non-compliance parameters, as requested by Lakeland Industry & Community Association.

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

With the exception of CH₄ and NMHC, the operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

The operational uptime was 7.9% for the CH₄ and NMHC channels. The Thermo 55i Methane/Non-Methane Hydrocarbon analyzer was removed from site, and a Thermo 51C THC analyzer was installed. No CH₄ and NMHC data was collected on the 51C analyzer. The AMD's monthly data completeness criteria (Chapter 6, Section 4.1.3, DQ 4-C) was not applicable to these channels in the October monitoring period.

All data collected this month were within the Alberta Ambient Air Quality Objectives and Guidelines (November, 2018).

THC/CH₄/NMHC:

- On October 3, the LICA-owned resident Thermo 55i Methane/Non-Methane analyzer (s/n: 1236656107), was replaced with an alternate model, Thermo 51c Total Hydrocarbon analyzer (s/n: 46609738), also LICA-owned, upon client request. Three hours of downtime were incurred due to this analyzer replacement event.
- On October 12, a repeat calibration was performed on the Thermo 51c analyzer, to correct a biased low zero drift. Six hours of downtime were incurred as a result.

Precipitation: A configuration error occurred during the datalogger upgrade on May 31, and was corrected on November 1. Consequently, data collected during this time frame, at rates greater than or equal to 0.3 mm/min, might be compromised and should therefore be used with caution.

The summary of results is presented on the following pages.

Any deviations or modifications made to the sampling or analytical methods are outlined in Section 1.0, Discussion. On this basis, Maxxam Analytics is issuing this completed report to Lakeland Industry & Community Association, Maskwa Continuous Monitoring Station.

Should you have any questions concerning the results or if we can be of further assistance, please contact us at 403-219-3677 or toll-free at 1-800-386-7247.

Monthly Continuous Data Summary

Lakeland Industry & Community Association Maskwa Continuous Monitoring Station						MAXIMUM VALUES							OPERATIONAL TIME (%)
PARAMETER	OBJECTIVES		EXCEEDANCES		MONTHLY AVERAGE	READING	DAY	1-HOUR			24-HOUR		
	1-hr	24-hr	1-hr	24-hr				HOUR	WIND SPEED (kph)	WIND DIRECTION (sector)	READING	DAY	
SO ₂ (ppb)	172	48	0	0	1	21	18	22	5.9	WNW	3	15	100.0
H ₂ S (ppb)	10	3	0	0	0	3	24	17	1.4	SE	1	24	100.0
THC (ppm)	-	-	-	-	2.11	3.85	8	22	0.3	E	2.25	11	98.8
CH ₄ (ppm)	-	-	-	-	1.99	2.16	1	9	1.2	SW	2.02	1	7.9
NMHC (ppm)	-	-	-	-	0.00	0.00	1	0	4.0	SE	0.00	1	7.9
NO ₂ (ppb)	159	-	0	-	4	30	18	22	5.9	WNW	8	17	100.0
NO (ppb)	-	-	-	-	1	32	25	7	0.2	NNE	4	25	100.0
NO _x (ppb)	-	-	-	-	5	55	25	7	0.2	NNE	11	17	100.0
RELATIVE HUMIDITY (%)	-	-	-	-	72	100	12	0	7.2	SSW	98	29	100.0
BAROMETRIC PRESSURE (millibar)	-	-	-	-	938	954	9	10	4.5	ENE	952	9	100.0
AMBIENT TEMPERATURE (°C)	-	-	-	-	1.4	21.8	17	14	5.4	WSW	10.4	17	100.0
PRECIPITATION (mm)	-	-	-	-	6.6	0.0	29	4	5.8	NNE	0.0	29	100.0
VECTOR WS (kph)	-	-	-	-	0.9	11.7	12	14	-	N	8.1	7	100.0
VECTOR WD (sec)	-	-	-	-	284 (WNW)	-	-	-	-	-	-	-	100.0

* Precipitation: data represents the total (sum) for the indicated time frame.

Exceedance Summary Report

SO₂ 1-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 1-hour AAAQO of 172 ppb.

SO₂ 24-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 24-hour AAAQO of 48.0 ppb.

H₂S 1-Hour Exceedances

Measured concentrations of hydrogen sulphide were below the 1-hour AAAQO of 10 ppb.

H₂S 24-Hour Exceedances

Measured concentrations of hydrogen sulphide were below the 24-hour AAAQO of 3 ppb.

NO₂ 1-Hour Exceedances

Measured concentrations of nitrogen dioxide were below the 1-hour AAAQO of 159 ppb.

In accordance with EPEA and the Substance Release Regulation.

In accordance with A Guide to Release Reporting and the Alberta Ambient Air Quality Objectives and Guidelines Summary.

TABLE OF CONTENTS

<u>Title</u>	<u>Page</u>
SUMMARY	2
MONTHLY CONTINUOUS DATA SUMMARY REPORT	3
EXCEEDANCE SUMMARY REPORT	4
TABLE OF CONTENTS	5
1.0 Discussion	7
2.0 Project Personnel	10
3.0 Plant Monthly Required AMD Summary	10
4.0 Calculations and Results	10
5.0 Methods and Procedures	11
Appendix I	14
Continuous Monitoring Data Results	
Sulphur Dioxide	15
Hydrogen Sulphide	21
Total Hydrocarbon	27
Methane	33
Non-Methane Hydrocarbon	39
Oxides of Nitrogen	45
Nitric Oxides	51
Nitrogen Dioxide	56
Wind Speed	62
Wind Direction	67
Standard Deviation Wind Direction	71
Relative Humidity	74
Barometric Pressure	77
Ambient Temperature	80
Precipitation	83

Appendix II	Equipment Calibration Results	86
	Sulphur Dioxide	87
	Hydrogen Sulphide	90
	Total Hydrocarbon	93
	Nitrogen Dioxide	101
	Wind System	105
	Meteorological System Check	107
	Calibrators	109
	Calibration Gases	112
Appendix III	Maximum Instantaneous Data	117
Appendix IV	Report Certification Form	136
Appendix V	Data Validation Certification Form	138

1.0 Discussion

This monthly report consists of continuous monitoring results for the following parameters: Sulphur Dioxide (SO₂), Hydrogen Sulphide (H₂S), Total Hydrocarbon (THC), Methane (CH₄), Non-Methane Hydrocarbon (NMHC), Oxides of Nitrogen (NO_x), Nitric Oxides (NO), Nitrogen Dioxide (NO₂), Particulate Matter 2.5 (PM_{2.5}), Relative Humidity (RH), Barometric Pressure (BP), Precipitation, Ambient Temperature (AmbTPX), Wind Speed (WS), Wind Direction (WD) and Standard Deviation Wind Direction (STDWD).

The sample inlet filter for all continuous air analyzers are replaced before the calibration begins. The sample manifold is cleaned during the site visit each month.

Control checks, consisting of a zero and span, are conducted daily on all continuous air monitors. In place of the air sample, zero air (from scrubbed air or gas cylinders) is used for zero checks, and a known concentration of the pollutant being analyzed is used for span checks. These checks are controlled by automatic timers and valves. The total zero span cycle is completed within an hour, the commencement of the zero span cycle is at the beginning of the hour.

Multipoint calibrations are done a minimum of once a month for each continuous air monitor. An additional calibration is required under the following conditions: 1) within three days after the initial start-up and stabilization of a newly installed instrument, 2) prior to shut-down or moving of an instrument which has been working to specification, and 3) when major repair has been done on the instrument.

Time during the first multi-point calibration is not considered downtime (Data is flagged as C). If more than one calibration is performed during the month, the time during the additional calibration is considered as downtime (Data is flagged as C1).

Only one zero/span check is run per day. Time during the zero/span check is not considered as downtime (Data is flagged as S). If an extra zero/span check is performed, the time during the additional check is considered as downtime (Data is flagged as S1).

The AMD requires each instrument and accompanying data recording system to be operational 90% of the time, at a minimum, for each monthly monitoring period.

All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.

Data contained in this monthly report has undergone the verification and validation based on the requirements of the AMD Chapter 6: Ambient Data Quality (December, 2016). The descriptions of the data verification and validation process can be found in Section 5 of this report. Instantaneous data, where applicable, is provided for reference purposes and has not undergone zero correction. The minimum and maximum statistics are highlighted in the data table and are for reference only. The highlighted cells are based on the software's interpretation of the exact position of the minimum or maximum value. The visual presentation of these statistics may not be the obvious choice in a data range due to rounding, truncating or analyzer specifications.

Hourly/minute data have been reviewed based on daily zero/span results and multi-point calibration results. Data may be considered invalid if a zero-corrected span check in excess of +/- 10% of the span concentration (established by the previous multi-point calibration) is encountered and/or significant differences in the calibration factor occurs (greater than 10%).

SULPHUR DIOXIDE (SO₂)

- Operational time for the monitoring period was 100%.
- The routine monthly calibration was performed on October 12.

HYDROGEN SULPHIDE (H₂S)

- Operational time for the monitoring period was 100%.
- The routine monthly calibration was performed on October 12.

TOTAL HYDROCARBONS (THC), METHANE (CH₄) and NON-METHANE HYDROCARBONS (NMHC)

- Operational time for the monitoring period was 98.8% for THC, equivalent to 9 hours of downtime. The CH₄/NMHC channel operational uptime was 7.9%. The AMD's monthly data completeness criteria (Chapter 6, Section 4.1.3, DQ 4-C) was not applicable to these channels in the October monitoring period.
- Following a successful shutdown calibration on October 3, the LICA-owned resident analyzer, Thermo 55i Methane/ Non- Methane (s/n: 1236656107), was removed to be installed at another LICA station, upon client request. An alternate analyzer model, Thermo 51c Total Hydrocarbon (s/n: 46609738), which is also LICA-owned and monitors THC only, was installed. A successful installation calibration was subsequently completed. Three hours of downtime were incurred on October 3 due to this analyzer replacement event. No methane and Non-methane hydrocarbon data was collected on the Thermo 51 C analyzer.
- A repeat calibration was performed on October 12 to correct a biased low zero drift, incurring 6 hours of downtime.

OXIDES OF NITROGEN (NO_x), NITRIC OXIDE (NO) and NITROGEN DIOXIDE (NO₂)

- Operational time for the monitoring period was 100%.
- The routine monthly calibration was performed on October 12.

WIND SPEED (WS), WIND DIRECTION (WD) and STANDARD DEVIATION WIND DIRECTION (STDWD)

- Operational time for the monitoring period was 100%.
- Wind data is reported as vector wind speed and vector wind direction. Wind direction is defined as the direction from which the wind is blowing from and is measured in degrees from true north.

RELATIVE HUMIDITY (RH)

- Operational time for the monitoring period was 100%.

BAROMETRIC PRESSURE (BP)

- Operational time for the monitoring period was 100%.

PRECIPITATION (PRECIP)

- Operational time for the monitoring period was 100%.
- A precipitation sensor audit was performed on October 12. The result was satisfactory.
- A configuration error occurred during the datalogger upgrade on May 31, and was corrected on November 1. Consequently, data collected during this time frame, at rates greater than or equal to 0.3 mm/min, might be compromised and should therefore be used with caution.

AMBIENT TEMPERATURE (AmbTPX)

- Operational time for the monitoring period was 100%.

2.0 Project Personnel

Mike Bisaga and Lily Lin were the contacts for Lakeland Industry & Community Association and the Maxxam field technician was Alexander Yakupov.

3.0 Plant Monthly Required AMD Summary

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

With the exception of CH₄ and NMHC, the operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

The operational uptime was 7.9% for the CH₄ and NMHC channels. The Thermo 55i Methane/Non-Methane Hydrocarbon analyzer was removed from the site, and a Thermo 51C THC analyzer was installed. No CH₄ and NMHC data was collected on the 51C analyzer. The AMD's monthly data completeness criteria (Chapter 6, Section 4.1.3, DQ 4-C) was not applicable to these channels in the October monitoring period.

All data collected this month were within the Alberta Ambient Air Quality Objectives and Guidelines (November, 2018).

4.0 Calculations and Results

All calculations and reporting of results follow the methods described in the AMD, 2016.

5.0 Methods and Procedures

The following methods and procedures were used to complete the monitoring program:

- Maxxam AIR SOP-00001: Methane, Non-Methane Hydrocarbon Analyzer Monitoring
- Maxxam AIR SOP-00013: RM Young Wind Monitor Calibration
- Maxxam AIR SOP-00209: Ambient Sulphur Monitoring
- Maxxam AIR SOP-00213: Ambient NO/NO₂/NO_x Monitoring
- Maxxam AIR SOP-00214: Ambient Hydrocarbon (THC) Monitoring
- Maxxam AIR SOP-00242: Precipitation Collector Installation/Maintenance

There were no deviations from the prescribed methods.

The following instruments were used to perform the test program:

- Sulphur Dioxide - Thermo 43I - TLE UV Fluorescent Analyzer
- Hydrogen Sulphide - Thermo 450i UV Fluorescent Analyzer
- Total Hydrocarbons - Thermo 51C FID Analyzer
- Methane, Non-Methane Hydrocarbon - Thermo 55i FID Analyzer
- Oxides of Nitrogen - Thermo 42i Chemiluminescent Analyzer
- Wind System - RM Young Unit
- Relative Humidity - Rotronic Hygroclip Unit
- Barometric Pressure - Met One Unit
- Ambient Temperature - Rotronic Hygroclip Unit
- Precipitation - Met One Unit
- Datalogger - Envidas Ultimate

The following steps were used to complete the data verification and validation process:

Level 0 Preliminary Verification

Level 0 data are raw data obtained directly from the data acquisition system (DAS). Under the step of Level 0, these data undergo a certain amount of manual or automated screening and flagging. It included a) identification of periods of missing data; b) verification of time stamps against reference time; c) verification that instrument diagnostics/datalogger flags indicate normal operation; d) comparison of data to upper and lower limits; e) rate of change flagging indicating that data changed too rapidly or not at all; and f) verification that zero, span and multipoint performance checks are within specifications. This level of verification is performed on a daily basis.

Level 1 Primary Validation

Validation actions under the step of Level 1 include a) review of all screening flags assigned during preliminary verification; b) review of all supporting site information and documentation; c) review of operational acceptance limits for each parameter/analyzer; d) review of daily zero/span and monthly calibration results for all gaseous parameters; and e) application of any necessary adjustments to data (e.g. baseline adjustments, below zero adjustments). This level of validation is performed on a monthly basis.

Level 2 Final Validation


The purpose of Level 2 validation is to verify that there are no inconsistencies among related data, or among regional data measured at nearby sites.

Level 3 Independent Data Review

Level 3 validation is the last step of data review, and it is completed by an individual that is independent of both field operations and primary data validation. A final independent QA review and endorsement is performed during this step before data is submitted to Alberta Environment.

Post-Final Validation

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. Any data issues or patterns which were not clear on a monthly basis are highlighted during this step. This validation is performed on an annual basis.




Source: Air Monitoring Directive (December 2016), Chapter 6, Ambient Data Quality; Figure 1 Data Collection and Management Process Flow Chart

APPENDIX I
CONTINUOUS MONITORING DATA RESULTS

SULPHUR DIOXIDE

SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)



Wind: LICA MASKWA
 Poll.: LICA MASKWA-SO2[ppb]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 20.11%

Calm Avg: 0.38 [ppb]

Direction	0.0-4.4	4.4-8.8	8.8-13.2	13.2-17.6	17.6-22.0	>22.0	Total
N	7.5	0.0	0.0	0.0	0.0	0.0	7.5
NE	11.8	0.0	0.0	0.0	0.0	0.0	11.8
E	3.8	0.0	0.0	0.0	0.0	0.0	3.8
SE	7.7	0.0	0.0	0.0	0.0	0.0	7.7
S	9.1	0.0	0.0	0.0	0.0	0.0	9.1
SW	15.0	0.0	0.0	0.0	0.0	0.0	15.0
W	15.0	0.3	0.0	0.0	0.0	0.0	15.3
NW	5.8	3.1	0.7	0.0	0.1	0.0	9.8
Summary	75.6	3.4	0.7	0.0	0.1	0.0	79.9

% Icon	Classes (ppb)	76	 0.0-4.4	3	 4.4-8.8	1	 8.8-13.2	0	 13.2-17.6	0	 17.6-22.0	0	 >22.0
--------	---------------	----	---	---	---	---	--	---	---	---	---	---	---

LICA MASKWA Poll.: LICA MASKWA-SO₂[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 20.11% Calm Poll Avg: 0.38[ppb]



SO2[ppb] Calibration: LICA MASKWA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



HYDROGEN SULPHIDE



HYDROGEN SULPHIDE Hourly Averages (H₂S ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
2	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
3	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	24
4	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	1	0	0	0	0	0	0	S	0	0	0	0	0	1	0	24
5	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	24
6	0	0	0	0	0	1	0	0	1	0	0	0	0	0	0	1	0	0	0	0	0	S	0	0	0	0	0	0	0	1	0	24
7	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	24
8	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	24
9	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	1	0	0	0	1	0	24
10	1	1	1	2	1	0	0	0	0	1	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	2	0	24
11	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
12	0	0	0	0	0	0	0	0	0	0	0	0	0	0	C	C	C	C	C	C	C	C	0	0	0	0	0	0	0	0	0	24
13	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
14	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
15	0	0	0	0	0	0	0	0	0	1	1	S	0	1	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
16	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
17	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
18	0	0	0	0	0	0	0	0	0	S	1	0	0	0	0	0	0	1	0	0	0	0	0	1	1	0	0	0	0	1	0	24
19	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
20	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
21	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
22	0	0	0	0	S	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
23	0	0	1	S	1	0	0	0	0	0	0	0	0	1	1	0	0	1	0	0	0	1	1	1	0	0	0	0	0	1	0	24
24	1	0	S	2	1	1	1	1	0	0	0	0	0	0	0	0	0	0	3	3	2	1	1	1	0	0	0	0	3	3	1	24
25	0	S	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	1	1	1	0	0	0	0	0	1	0	24
26	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	24
27	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	S	0	0	0	0	0	1	0	24
28	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	1	0	S	1	1	0	0	0	0	1	0	24
29	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	S	0	0	0	0	0	0	1	0	24
30	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	24
31	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	24
HOURLY MAX	1	1	1	2	1	1	1	1	1	1	1	1	0	1	1	1	1	3	3	2	1	1	1	1	0	0	0	0				
HOURLY AVG	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT:	1-HR	10	ppb	24-HR	3	ppb
----------------------	------	----	-----	-------	---	-----


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0			
NUMBER OF 24-HR EXCEEDANCES:	0			
NUMBER OF NON-ZERO READINGS:	52			
MINIMUM 1-HR AVERAGE:	0 ppb @ HOUR	0	ON DAY	1
MAXIMUM 1-HR AVERAGE:	3 ppb @ HOUR	17	ON DAY	24
MAXIMUM 24-HR AVERAGE:	1 ppb		ON DAY	24
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	744 hrs	
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	100.0 %	
STANDARD DEVIATION:	0	MONTHLY AVERAGE:	0 ppb	

24 HR AVERAGES October 2018



HYDROGEN SULPHIDE Hourly Averages (H₂S ppb)



Wind: LICA MASKWA
Poll.: LICA MASKWA-H2S[ppb]
Monthly: 18/10
Type: PollutionRose
Direction: Blowing From (Wind Frequency)
Based On 1 Hr.

Calm: 20.14%

Calm Avg: 0.29 [ppb]

Direction	0.0-1.3	1.3-2.7	2.7-4.0	>4.0	Total
N	7.2	0.0	0.0	0.0	7.2
NE	11.8	0.0	0.0	0.0	11.8
E	3.8	0.0	0.0	0.0	3.8
SE	7.7	0.0	0.0	0.0	7.7
S	9.1	0.0	0.0	0.0	9.1
SW	15.0	0.0	0.0	0.0	15.0
W	15.3	0.0	0.0	0.0	15.3
NW	9.8	0.1	0.0	0.0	9.9
Summary	79.7	0.1	0.0	0.0	79.9

% Icon Classes (ppb)


80 0.0-1.3

0 1.3-2.7

0 2.7-4.0

0 >4.0

LICA MASKWA Poll.: LICA MASKWA-H2S[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 20.14% Calm Poll Avg: 0.29[ppb]



H2S[ppb] Calibration: LICA MASKWA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



TOTAL HYDROCARBON




TOTAL HYDROCARBONS Hourly Averages (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2.02	2.02	S	2.01	2.01	2.03	2.05	2.07	2.12	2.16	2.11	2.04	1.99	1.99	1.98	1.97	1.97	1.98	1.99	1.98	1.99	1.99	1.99	2.01	1.97	2.16	2.02	24	
2	2.01	S	2.00	1.98	1.95	1.95	1.96	1.95	1.95	1.94	1.94	1.94	1.93	1.94	1.94	1.93	1.94	1.94	1.95	1.95	1.96	1.97	1.98	1.93	2.01	1.95	24		
3	S	1.99	1.98	1.99	1.97	1.97	1.96	2.01	1.93	C	C	Y	Y	Y	C	C	C	C	2.13	2.10	2.10	2.08	2.10	S	1.93	2.13	2.02	21	
4	2.07	2.07	2.06	2.05	2.06	2.08	2.11	2.12	2.11	2.05	2.07	2.04	2.02	2.04	2.04	2.05	2.04	2.07	2.17	2.25	2.25	2.24	S	2.28	2.02	2.28	2.10	24	
5	2.32	2.37	2.41	2.37	2.33	2.30	2.33	2.30	2.29	2.23	2.16	2.08	2.07	2.05	2.02	2.03	2.10	2.23	2.09	2.09	2.10	S	2.10	2.13	2.02	2.41	2.20	24	
6	2.10	2.12	2.16	2.12	2.14	2.09	2.09	2.10	2.07	2.06	2.06	2.06	2.06	2.05	2.11	2.06	2.06	2.08	2.11	2.11	S	2.14	2.13	2.09	2.05	2.16	2.09	24	
7	2.07	2.09	2.10	2.10	2.09	2.09	2.09	2.10	2.09	2.08	2.07	2.07	2.07	2.06	2.05	2.06	2.06	2.05	2.06	S	2.05	2.07	2.06	2.07	2.05	2.10	2.07	24	
8	2.06	2.08	2.09	2.09	2.09	2.09	2.10	2.11	2.10	2.09	2.08	2.07	2.06	2.05	2.04	2.03	2.05	2.05	S	2.08	2.11	2.34	3.85	3.84	2.03	3.85	2.24	24	
9	2.46	2.24	2.21	2.22	2.21	2.23	2.20	2.22	2.23	2.12	2.12	2.09	2.09	2.08	2.10	2.10	2.08	S	2.11	2.13	2.12	2.13	2.15	2.26	2.08	2.46	2.17	24	
10	2.28	2.22	2.20	2.21	2.18	2.20	2.24	2.24	2.29	2.20	2.16	2.14	2.09	2.08	2.08	2.09	S	2.08	2.11	2.14	2.17	2.24	2.27	2.26	2.08	2.29	2.18	24	
11	2.28	2.29	2.32	2.33	2.34	2.35	2.25	2.21	2.20	2.28	2.39	2.46	2.35	2.31	2.27	S	2.14	2.13	2.13	2.12	2.18	2.16	2.16	2.16	2.12	2.46	2.25	24	
12	2.14	2.12	2.07	2.06	2.07	2.12	2.13	2.17	2.06	C1	C1	C1	C1	C1	C1	2.12	2.08	2.09	2.12	2.15	2.15	2.15	2.15	2.12	2.06	2.17	2.11	18	
13	2.15	2.13	2.14	2.14	2.16	2.13	2.11	2.11	2.12	2.12	2.09	2.06	2.04	S	2.04	2.06	2.06	2.07	2.09	2.12	2.14	2.13	2.22	2.17	2.04	2.22	2.11	24	
14	2.19	2.19	2.21	2.23	2.22	2.30	2.68	2.52	2.33	2.27	2.17	2.07	S	2.02	2.00	2.00	1.98	2.01	2.03	2.03	2.03	2.05	2.08	2.11	1.98	2.68	2.16	24	
15	2.00	1.96	1.95	1.95	1.97	1.99	1.97	1.98	2.01	2.09	2.06	S	1.95	1.99	1.96	1.96	2.01	2.00	2.09	2.12	2.07	2.04	2.05	2.04	1.95	2.12	2.01	24	
16	2.06	2.05	2.06	2.08	2.10	2.19	2.22	2.22	2.20	2.15	S	2.17	2.23	2.12	2.13	2.09	2.14	2.13	2.08	2.09	2.07	2.29	2.25	2.29	2.05	2.29	2.15	24	
17	2.32	2.35	2.33	2.30	2.35	2.32	2.29	2.28	2.30	S	2.41	2.32	2.05	1.97	2.00	2.01	1.96	1.91	1.98	2.01	2.09	2.12	2.09	2.15	1.91	2.41	2.17	24	
18	2.17	2.11	2.09	2.34	2.31	2.13	2.03	2.00	S	2.03	1.97	1.91	1.90	1.86	1.87	1.91	1.96	2.19	1.98	1.99	2.02	2.05	2.19	2.23	1.86	2.34	2.05	24	
19	2.11	2.01	2.02	2.01	2.02	2.03	2.05	S	2.08	2.05	2.06	2.07	2.07	2.08	2.07	2.07	2.09	2.12	2.12	2.14	2.14	2.20	2.16	2.13	2.01	2.20	2.08	24	
20	2.12	2.11	2.12	2.11	2.11	2.12	S	2.08	2.13	2.15	2.18	2.19	2.15	1.99	1.87	1.82	1.86	1.87	1.90	1.94	1.98	1.99	2.01	1.98	1.82	2.19	2.03	24	
21	1.98	2.00	2.03	2.04	2.04	S	2.06	2.06	2.04	2.05	2.02	1.97	1.95	1.92	1.93	1.96	1.99	2.07	2.06	2.01	2.03	2.07	2.05	2.06	1.92	2.07	2.02	24	
22	2.06	2.09	2.18	2.36	S	2.17	2.17	2.18	2.24	2.19	2.15	2.22	2.22	2.12	2.06	2.03	2.15	2.26	2.31	2.25	2.19	2.19	2.22	2.23	2.03	2.36	2.18	24	
23	2.23	2.22	2.25	S	2.27	2.19	2.12	2.16	2.18	2.17	2.20	2.13	2.07	1.99	1.99	1.97	2.02	2.06	2.08	2.08	2.09	2.16	2.14	2.16	1.97	2.27	2.13	24	
24	2.17	2.13	S	2.16	2.16	2.15	2.17	2.21	2.15	2.12	2.10	2.07	2.00	2.00	1.98	1.96	2.01	2.08	2.09	2.14	2.15	2.17	2.15	2.19	1.96	2.21	2.11	24	
25	2.15	S	2.19	2.20	2.20	2.19	2.17	2.24	2.19	2.22	2.11	2.15	2.08	2.02	2.05	2.07	2.10	2.11	2.14	2.24	2.26	2.28	2.20	2.20	2.02	2.28	2.16	24	
26	S	2.19	2.20	2.26	2.31	2.28	2.24	2.16	2.11	2.08	2.03	2.00	1.94	1.92	1.93	1.93	1.96	2.02	2.07	2.11	2.13	2.29	2.24	S	1.92	2.31	2.11	24	
27	2.27	2.40	2.31	2.23	2.19	2.16	2.27	2.29	2.11	2.01	2.01	2.01	2.00	1.99	2.05	2.05	2.04	2.02	2.03	2.01	2.00	2.01	S	2.01	1.99	2.40	2.11	24	
28	2.02	2.04	2.06	2.09	2.13	2.24	2.29	2.26	2.29	2.19	2.09	2.13	2.14	2.09	2.09	2.06	2.10	2.09	2.11	2.07	2.04	S	2.02	2.02	2.02	2.29	2.12	24	
29	1.99	1.96	1.94	1.95	1.94	1.94	1.94	1.96	1.97	1.99	1.99	2.00	2.02	2.09	2.09	2.09	2.08	2.12	2.10	2.06	S	2.05	2.08	2.10	1.94	2.12	2.02	24	
30	2.09	2.10	2.09	2.07	2.08	2.09	2.09	2.06	2.04	2.05	2.05	2.24	2.09	2.11	2.12	2.12	2.14	2.13	2.07	S	2.09	2.10	2.22	2.31	2.04	2.31	2.11	24	
31	2.31	2.32	2.35	2.30	2.26	2.23	2.22	2.23	2.24	2.29	2.28	2.16	2.20	2.14	2.16	2.14	2.16	2.15	S	2.14	2.19	2.23	2.19	2.12	2.12	2.35	2.22	24	
HOURLY MAX	2.46	2.40	2.41	2.37	2.35	2.35	2.68	2.52	2.33	2.29	2.41	2.46	2.35	2.31	2.27	2.14	2.16	2.26	2.31	2.25	2.26	2.34	3.85	3.84					
HOURLY AVG	2.14	2.14	2.14	2.15	2.14	2.15	2.15	2.15	2.14	2.12	2.11	2.10	2.07	2.04	2.04	2.03	2.05	2.07	2.08	2.09	2.10	2.14	2.19	2.20					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	697			
MINIMUM 1-HR AVERAGE:	1.82 ppm	@ HOUR	15	ON DAY 20
MAXIMUM 1-HR AVERAGE:	3.85 ppm	@ HOUR	22	ON DAY 8
MAXIMUM 24-HR AVERAGE:	2.25 ppm			ON DAY 11
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	735 hrs	
MONTHLY CALIBRATION TIME:	6 hrs	AMD OPERATION UPTIME:	98.8 %	
STANDARD DEVIATION:	0.14	MONTHLY AVERAGE:	2.11 ppm	

TOTAL HYDROCARBONS Hourly Averages (THC ppm)




Wind: LICA MASKWA
 Poll.: LICA MASKWA-THC[ppm]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 20.37% Calm Avg: 2.18 [ppm]

Direction	0.0-1.3	1.3-2.6	2.6-3.9	>3.9	Total
N	0.0	6.9	0.0	0.0	6.9
NE	0.0	11.9	0.0	0.0	11.9
E	0.0	3.9	0.0	0.0	3.9
SE	0.0	7.8	0.0	0.0	7.8
S	0.0	9.2	0.0	0.0	9.2
SW	0.0	15.1	0.1	0.0	15.2
W	0.0	15.5	0.0	0.0	15.5
NW	0.0	9.3	0.0	0.0	9.3
Summary	0.0	79.5	0.1	0.0	79.6


% Icon Classes (ppm) 0 0.0-1.3 79 1.3-2.6 0 2.6-3.9 0 >3.9

LICA MASKWA Poll.: LICA MASKWA-THC[ppm] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 20.37% Calm Poll Avg: 2.18[ppm]



THC[ppm] Calibration: LICA MASKWA Monthly: 18/10 Type: Span

■ Span Meas
 — Span Ref
 — Span Low
 — Span High



METHANE




METHANE Hourly Averages (CH₄ ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	2.02	2.02	S	2.01	2.01	2.03	2.05	2.07	2.12	2.16	2.11	2.04	1.99	1.99	1.98	1.97	1.97	1.98	1.99	1.98	1.99	1.99	1.99	2.01	1.97	2.16	2.02	24				
2	2.01	S	2.00	1.98	1.95	1.95	1.96	1.95	1.95	1.94	1.94	1.94	1.94	1.93	1.94	1.94	1.93	1.94	1.94	1.95	1.95	1.96	1.97	1.98	1.93	2.01	1.95	24				
3	S	1.99	1.98	1.99	1.97	1.97	1.96	2.01	1.93	C	C	-	-	-	-	-	-	-	-	-	-	-	-	-	1.93	2.01	1.98	11				
4	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
5	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
6	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
7	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
8	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
9	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
10	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
11	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
12	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
13	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
14	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
15	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
16	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
17	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
18	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
19	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
20	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
21	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
22	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
23	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
24	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
25	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
26	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
27	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
28	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
29	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
30	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
31	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0				
HOURLY MAX	2.02	2.02	2.00	2.01	2.01	2.03	2.05	2.07	2.12	2.16	2.11	2.04	1.99	1.99	1.98	1.97	1.97	1.98	1.99	1.98	1.99	1.99	1.99	2.01								
HOURLY AVG	2.02	2.01	1.99	1.99	1.98	1.98	1.99	2.01	2.00	2.05	2.03	1.99	1.97	1.96	1.96	1.95	1.96	1.97	1.97	1.97	1.97	1.98	1.98	2.00								

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	54			
MINIMUM 1-HR AVERAGE:	1.93 ppm	@ HOUR	13	ON DAY 2
MAXIMUM 1-HR AVERAGE:	2.16 ppm	@ HOUR	9	ON DAY 1
MAXIMUM 24-HR AVERAGE:	2.02 ppm			ON DAY 1
IZS CALIBRATION TIME:	3 hrs	OPERATIONAL TIME:	59	hrs
MONTHLY CALIBRATION TIME:	2 hrs	AMD OPERATION UPTIME:	7.9	%
STANDARD DEVIATION:	0.05	MONTHLY AVERAGE:	1.99	ppm

METHANE Hourly Averages (CH₄ ppm)




Wind: LICA MASKWA
 Poll.: LICA MASKWA-CH4[ppm]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 25.93% Calm Avg: 2.03 [ppm]

Direction	0.0-0.7	0.7-1.4	1.4-2.2	>2.2	Total
N	0.0	0.0	24.1	0.0	24.1
NE	0.0	0.0	0.0	0.0	0.0
E	0.0	0.0	0.0	0.0	0.0
SE	0.0	0.0	3.7	0.0	3.7
S	0.0	0.0	0.0	0.0	0.0
SW	0.0	0.0	3.7	0.0	3.7
W	0.0	0.0	31.5	0.0	31.5
NW	0.0	0.0	11.1	0.0	11.1
Summary	0.0	0.0	74.1	0.0	74.1


% Icon Classes (ppm) 0 0.0-0.7 0 0.7-1.4 74 1.4-2.2 0 >2.2

LICA MASKWA Poll.: LICA MASKWA-CH4[ppm] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 25.93% Calm Poll Avg: 2.03[ppm]



CH4[ppm] Calibration: LICA MASKWA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



NON-METHANE HYDROCARBON




NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.					
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.						
DAY																																	
1	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
2	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24	
3	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	C	C	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0.00	0.00	0.00	0.00	11		
4	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
5	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
6	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
7	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
8	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
9	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
10	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
11	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
12	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
13	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
14	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
15	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
16	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
17	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
18	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
19	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
20	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
21	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
22	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
23	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
24	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
25	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
26	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
27	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
28	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
29	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
30	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
31	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0		
HOURLY MAX	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	
HOURLY AVG	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	0
MINIMUM 1-HR AVERAGE:	0.00 ppm @ HOUR 0 ON DAY 1
MAXIMUM 1-HR AVERAGE:	0.00 ppm @ HOUR 0 ON DAY 1
MAXIMUM 24-HR AVERAGE:	0.00 ppm ON DAY 1
IZS CALIBRATION TIME:	3 hrs OPERATIONAL TIME: 59 hrs
MONTHLY CALIBRATION TIME:	2 hrs AMD OPERATION UPTIME: 7.9 %
STANDARD DEVIATION:	0.00 MONTHLY AVERAGE: 0.00 ppm

NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)




Wind: LICA MASKWA
 Poll.: LICA MASKWA-NMHC[ppm]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 25.93% Calm Avg: 0.00 [ppm]

Direction	0.0-0.7	0.7-1.3	1.3-2.0	>2.0	Total
N	24.1	0.0	0.0	0.0	24.1
NE	0.0	0.0	0.0	0.0	0.0
E	0.0	0.0	0.0	0.0	0.0
SE	3.7	0.0	0.0	0.0	3.7
S	0.0	0.0	0.0	0.0	0.0
SW	3.7	0.0	0.0	0.0	3.7
W	31.5	0.0	0.0	0.0	31.5
NW	11.1	0.0	0.0	0.0	11.1
Summary	74.1	0.0	0.0	0.0	74.1

% Icon Classes (ppm) 74 0.0-0.7 0 0.7-1.3 0 1.3-2.0 0 >2.0

LICA MASKWA Poll.: LICA MASKWA-NMHC[ppm] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 25.93% Calm Poll Avg: 0.00[ppm]



NMHC[ppm] Calibration: LICA MASKWA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



OXIDES OF NITROGEN




OXIDES OF NITROGEN Hourly Averages (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	2	0	S	1	1	2	4	5	5	3	4	7	5	4	2	1	1	1	1	1	1	3	1	3	0	7	3	24				
2	2	S	5	3	2	2	4	3	3	2	6	4	1	0	0	3	0	4	0	1	2	1	1	1	0	6	2	24				
3	S	3	2	3	14	29	14	9	2	1	0	3	3	2	2	0	0	1	1	0	1	0	0	S	0	29	4	24				
4	2	1	1	1	1	11	15	7	4	3	4	4	2	2	1	2	1	2	8	5	2	S	5	1	15	4	24					
5	5	5	4	4	4	6	11	13	16	12	9	9	3	4	3	3	1	2	2	3	1	S	3	2	1	16	5	24				
6	1	6	6	1	1	1	3	11	10	9	5	5	7	1	5	3	3	1	1	2	S	3	3	1	1	11	4	24				
7	1	1	1	0	0	3	3	5	3	0	0	2	0	0	0	0	0	0	0	S	1	0	0	1	0	5	1	24				
8	1	1	1	1	1	1	1	1	3	3	2	1	1	1	0	0	1	2	S	4	1	2	2	1	0	4	1	24				
9	0	0	0	0	0	0	1	1	12	0	3	2	3	1	2	2	0	S	0	0	0	0	2	1	0	12	1	24				
10	0	0	0	0	0	0	0	7	18	8	5	1	0	2	1	3	S	3	1	1	4	4	5	5	0	18	3	24				
11	3	3	2	4	3	4	5	10	3	3	4	5	4	4	3	S	4	3	3	2	3	4	4	4	2	10	4	24				
12	4	5	6	6	6	9	16	14	6	C	C	C	C	C	C	C	C	C	2	2	4	2	1	1	1	16	6	24				
13	1	1	1	1	19	2	1	6	9	2	4	12	7	S	3	1	1	5	0	1	11	7	4	2	0	19	4	24				
14	1	2	4	3	4	7	20	11	8	7	6	3	S	2	1	1	1	2	3	3	4	7	9	7	1	20	5	24				
15	2	1	1	0	2	9	1	1	3	20	16	S	9	20	14	12	16	7	18	9	11	13	11	1	0	20	9	24				
16	5	9	3	5	5	8	7	22	27	6	S	8	8	4	4	3	4	4	3	2	2	9	4	4	2	27	7	24				
17	4	13	5	6	9	17	23	35	29	S	15	12	6	1	1	2	5	4	7	12	14	9	9	8	1	35	11	24				
18	6	5	4	11	11	7	5	7	S	17	6	3	2	1	2	4	9	22	6	1	8	15	43	26	1	43	10	24				
19	14	5	2	2	2	3	14	S	19	7	3	3	2	2	9	6	5	1	0	0	0	7	2	0	0	19	5	24				
20	0	0	0	1	1	1	S	6	3	5	6	5	5	3	2	1	1	1	0	0	0	0	1	0	0	6	2	24				
21	0	2	1	1	2	S	12	28	4	2	2	1	1	0	2	15	11	13	13	10	7	9	9	7	0	28	7	24				
22	3	3	3	2	S	4	5	3	4	5	3	4	3	2	2	2	3	6	7	3	4	5	5	2	7	7	4	24				
23	4	4	2	S	5	3	3	4	5	5	4	3	3	3	2	2	2	2	1	4	5	2	2	1	1	5	3	24				
24	1	1	S	2	2	1	5	6	12	13	14	11	9	9	5	13	2	2	3	2	4	2	3	5	1	14	6	24				
25	1	S	23	7	14	1	5	55	22	8	7	23	6	6	4	6	5	4	9	3	9	7	3	17	1	55	11	24				
26	S	4	5	12	9	14	23	25	10	2	1	2	2	1	1	1	3	1	2	4	3	3	S	1	25	6	24					
27	8	8	8	6	5	5	10	16	10	5	4	1	1	1	17	9	9	1	4	3	3	3	S	4	1	17	6	24				
28	2	5	7	5	1	2	4	4	15	5	2	2	4	10	6	1	5	5	9	1	1	S	2	2	1	15	4	24				
29	2	2	2	2	7	4	6	7	8	8	4	12	17	12	18	8	9	10	22	17	S	5	6	6	2	22	8	24				
30	5	5	5	7	6	8	8	11	3	1	4	6	6	5	5	4	4	3	1	S	5	3	5	7	1	11	5	24				
31	6	6	6	5	4	4	7	12	9	12	10	7	7	5	9	6	7	5	S	13	10	7	6	4	4	13	7	24				
HOURLY MAX	14	13	23	12	19	29	23	55	29	20	16	23	17	20	18	15	16	22	22	17	14	15	43	26								
HOURLY AVG	3	3	4	3	5	6	8	12	10	6	5	6	4	4	4	4	4	4	4	4	4	4	5	5								

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	641				
MINIMUM 1-HR AVERAGE:	0 ppb	@ HOUR	1	ON DAY	1
MAXIMUM 1-HR AVERAGE:	55 ppb	@ HOUR	7	ON DAY	25
MAXIMUM 24-HR AVERAGE:	11 ppb			ON DAY	17
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	744 hrs		
MONTHLY CALIBRATION TIME:	9 hrs	AMD OPERATION UPTIME:	100.0 %		
STANDARD DEVIATION:	6	MONTHLY AVERAGE:	5 ppb		

OXIDES OF NITROGEN Hourly Averages (NO_x ppb)



Wind: LICA MASKWA
 Poll.: LICA MASKWA-NOX[ppb]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 20.20% Calm Avg: 4.87 [ppb]

Direction	0.0-18.7	18.7-37.3	37.3-56.0	>56.0	Total
N	7.1	0.0	0.0	0.0	7.1
NE	11.7	0.1	0.0	0.0	11.8
E	3.8	0.0	0.0	0.0	3.8
SE	7.7	0.0	0.0	0.0	7.7
S	9.1	0.0	0.0	0.0	9.1
SW	14.4	0.7	0.0	0.0	15.1
W	14.9	0.4	0.0	0.0	15.4
NW	8.7	1.0	0.1	0.0	9.8
Summary	77.4	2.3	0.1	0.0	79.8

% Icon Classes (ppb)


77 0.0-18.7

2 18.7-37.3

0 37.3-56.0

0 >56.0

LICA MASKWA Poll.: LICA MASKWA-NOX[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 20.20% Calm Poll Avg: 4.87[ppb]



NOX[ppb] Calibration: LICA MASKWA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



NITRIC OXIDE


NITRIC OXIDE Hourly Averages (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.																						
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.																							
DAY																																																		
1	0	0	S	0	0	0	0	0	1	1	1	3	1	1	0	0	0	0	0	0	0	0	0	0	0	0	3	0	24																					
2	0	S	0	0	0	0	0	0	1	1	3	1	0	0	0	1	0	0	0	0	0	0	0	0	0	0	3	0	24																					
3	S	0	0	0	1	4	1	2	0	0	0	1	1	1	1	0	0	0	0	0	0	0	0	S	0	4	1	24																						
4	0	0	0	0	0	2	6	2	1	1	1	1	1	0	0	0	0	0	0	0	0	0	S	0	0	6	1	24																						
5	0	0	0	0	0	1	1	2	6	5	4	4	1	1	1	1	0	0	0	0	0	0	S	0	0	6	1	24																						
6	0	1	1	0	0	0	0	2	2	2	2	1	2	0	1	0	0	0	0	0	0	S	0	0	0	2	1	24																						
7	0	0	0	0	0	1	1	2	1	0	0	1	0	0	0	0	0	0	0	0	S	0	0	0	0	2	0	24																						
8	0	0	0	0	0	0	0	0	1	1	1	0	0	0	0	0	0	0	S	0	0	0	0	0	0	1	0	24																						
9	0	0	0	0	0	0	0	0	6	0	1	1	1	1	1	0	0	S	0	0	0	0	0	0	0	6	0	24																						
10	0	0	0	0	0	0	0	4	12	3	1	0	0	0	0	S	0	0	0	0	0	0	0	0	0	12	1	24																						
11	0	0	0	0	0	0	0	2	0	0	1	1	1	1	0	S	0	0	0	0	0	0	0	0	0	2	0	24																						
12	0	0	0	0	0	2	6	4	1	C	C	C	C	C	C	C	C	C	0	0	0	0	0	0	0	6	1	24																						
13	0	0	0	0	6	0	0	2	3	1	2	6	3	S	1	0	0	1	0	0	1	0	0	0	0	6	1	24																						
14	0	0	0	0	0	0	2	2	2	3	2	1	S	0	0	0	0	0	0	0	0	0	0	0	0	3	1	24																						
15	0	0	0	0	0	0	0	0	1	8	6	S	3	7	4	3	3	1	3	1	0	0	2	0	0	8	2	24																						
16	1	1	0	0	0	0	1	12	18	1	S	2	2	1	1	0	0	0	0	0	0	5	0	0	0	18	2	24																						
17	0	7	0	0	0	3	4	13	12	S	5	4	1	0	0	0	0	0	0	0	0	0	0	0	0	13	2	24																						
18	0	0	0	0	0	1	0	0	S	4	1	0	0	0	1	1	2	5	0	0	0	0	13	5	0	13	1	24																						
19	2	0	0	0	0	1	2	S	7	2	1	1	0	0	3	2	1	0	0	0	0	0	0	0	0	7	1	24																						
20	0	0	0	0	0	0	S	0	0	2	2	2	2	1	1	0	0	0	0	0	0	0	0	0	0	2	0	24																						
21	0	0	0	0	0	S	1	8	2	0	0	0	0	0	0	5	2	2	0	0	0	0	0	0	0	8	1	24																						
22	0	0	0	0	S	0	1	0	1	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0	1	0	24																						
23	0	0	0	S	0	0	0	0	1	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0	1	0	24																						
24	0	0	S	0	0	0	3	1	6	7	8	5	3	4	1	4	0	0	1	0	1	0	0	1	0	8	2	24																						
25	0	S	13	1	3	0	2	32	11	3	2	11	2	2	0	1	0	0	0	0	1	1	0	4	0	32	4	24																						
26	S	0	0	3	2	3	6	8	3	0	0	1	1	0	0	0	0	0	0	0	0	0	0	S	0	8	1	24																						
27	0	0	0	0	0	0	0	3	3	1	1	0	0	0	6	2	1	0	0	0	0	0	S	0	0	6	1	24																						
28	0	0	0	0	0	0	0	1	8	2	0	1	1	2	1	0	0	0	0	0	0	S	0	0	0	8	1	24																						
29	0	0	0	0	2	1	3	3	2	2	1	5	7	4	8	2	1	3	8	6	S	0	1	1	0	8	3	24																						
30	1	1	0	1	0	1	0	1	0	0	1	1	2	1	1	0	0	0	0	S	0	0	0	0	0	2	1	24																						
31	0	0	0	0	0	0	2	6	2	4	4	3	3	1	2	1	0	0	S	3	0	0	0	0	0	6	1	24																						
HOURLY MAX	2	7	13	3	6	4	6	32	18	8	8	11	7	7	8	5	3	5	8	6	1	5	13	5																										
HOURLY AVG	0	0	0	0	0	1	1	4	4	2	2	2	1	1	1	0	0	0	0	0	0	0	1	0																										

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	256			
MINIMUM 1-HR AVERAGE:	0 ppb	@ HOUR	0	ON DAY 1
MAXIMUM 1-HR AVERAGE:	32 ppb	@ HOUR	7	ON DAY 25
MAXIMUM 24-HR AVERAGE:	4 ppb			ON DAY 25
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	744 hrs	
MONTHLY CALIBRATION TIME:	9 hrs	AMD OPERATION UPTIME:	100.0 %	
STANDARD DEVIATION:	2	MONTHLY AVERAGE:	1 ppb	

NITRIC OXIDE Hourly Averages (NO ppb)



% Icon Classes (ppb)


79 0.0-11.0

1 11.0-22.0

0 22.0-33.0

0 >33.0

LICA MASKWA Poll.: LICA MASKWA-NO[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 20.20% Calm Poll Avg: 1.01[ppb]



NITROGEN DIOXIDE

NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	2	0	S	1	1	2	4	5	4	3	3	4	3	3	2	1	1	1	1	1	1	3	1	3	0	5	2	24				
2	2	S	5	3	2	2	3	2	2	1	4	2	0	0	0	2	0	3	0	1	1	1	1	1	0	5	2	24				
3	S	3	2	3	13	25	13	7	1	1	0	2	2	1	1	0	0	1	1	0	1	0	0	S	0	25	4	24				
4	2	1	1	1	1	9	9	5	3	2	3	2	1	1	1	2	1	2	8	5	2	S	5	1	9	3	24					
5	5	4	4	3	4	6	10	11	10	6	5	5	2	3	2	2	1	2	2	3	1	S	3	2	1	11	4					
6	1	5	6	1	1	1	3	9	8	7	3	3	5	1	3	3	2	1	1	2	S	3	3	1	1	9	3					
7	1	1	1	0	0	2	2	3	2	0	0	1	0	0	0	0	0	0	0	S	1	0	0	1	0	3	1					
8	1	0	0	0	1	1	1	0	2	2	1	1	1	0	0	0	0	0	1	S	4	1	2	1	0	4	1					
9	0	0	0	0	0	0	1	1	6	0	2	1	2	1	1	1	0	S	0	0	0	0	2	1	0	6	1					
10	0	0	0	0	0	0	0	0	3	6	5	4	1	0	1	1	2	S	3	1	1	4	4	5	5	0	6					
11	3	3	2	4	3	4	5	8	3	3	4	4	3	3	3	S	4	3	3	2	3	4	4	4	4	2	8					
12	4	5	6	6	6	7	10	10	4	C	C	C	C	C	C	C	C	C	C	2	2	3	2	1	1	10	5					
13	1	1	1	1	13	2	1	4	6	1	2	6	4	S	2	1	1	4	0	1	10	7	4	2	0	13						
14	1	2	4	3	4	7	18	9	6	5	3	2	S	2	1	1	1	2	3	3	4	7	9	7	1	18						
15	2	1	1	0	2	9	1	1	2	12	10	S	6	13	9	8	13	7	16	9	10	13	9	1	0	16						
16	4	9	3	5	5	8	6	9	10	5	S	6	6	3	3	3	4	4	3	2	2	5	4	4	2	10						
17	4	6	5	6	9	14	18	22	17	S	10	8	5	1	1	5	4	7	12	14	9	9	8	1	22	8						
18	6	5	4	11	11	6	5	6	S	13	5	2	2	1	1	3	7	17	6	1	8	14	30	21	1	30						
19	13	5	2	2	2	3	12	S	12	4	2	2	2	1	6	4	4	1	0	0	0	7	2	0	0	13						
20	0	0	0	1	1	1	S	5	3	4	4	3	3	2	1	1	0	1	0	0	0	0	1	0	0	5						
21	0	2	1	1	2	S	11	20	2	2	1	1	1	0	2	11	9	12	12	10	7	9	9	7	0	20						
22	3	3	3	2	S	4	5	3	3	3	2	3	2	2	2	2	2	3	6	7	3	4	5	5	2	7						
23	4	4	2	S	5	3	3	4	4	4	3	2	2	2	2	2	2	2	1	4	5	2	2	1	1	5						
24	1	1	S	2	2	1	2	5	6	6	6	6	5	6	4	9	2	2	2	2	4	2	3	3	1	9						
25	1	S	10	6	11	1	3	23	11	5	5	12	4	5	3	5	5	4	8	3	8	6	3	13	1	23						
26	S	4	5	8	7	11	17	17	7	2	1	2	1	1	1	1	3	1	2	4	3	3	S	1	17	5						
27	7	8	8	6	5	4	10	13	7	4	3	1	1	1	11	7	9	1	4	3	3	3	S	4	1	13						
28	2	5	7	5	1	2	3	3	7	4	2	2	3	7	5	1	5	5	9	1	1	S	2	2	1	9						
29	2	2	2	2	5	3	3	4	6	5	3	7	10	8	10	6	7	8	13	11	S	4	5	5	2	13						
30	5	5	5	6	5	8	8	10	3	1	4	4	4	4	4	3	4	3	1	S	4	3	5	7	1	10						
31	6	6	6	5	4	4	5	6	7	8	5	4	4	7	5	7	5	5	S	10	9	7	6	4	4	10						
HOURLY MAX	13	9	10	11	13	25	18	23	17	13	10	12	10	13	11	11	13	17	16	12	14	14	30	21								
HOURLY AVG	3	3	3	3	4	5	6	8	6	4	3	3	3	3	3	3	3	4	4	4	4	4	5	4								

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 159 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0			
NUMBER OF NON-ZERO READINGS:	633			
MINIMUM 1-HR AVERAGE:	0	ppb	@ HOUR	1 ON DAY
MAXIMUM 1-HR AVERAGE:	30	ppb	@ HOUR	22 ON DAY
MAXIMUM 24-HR AVERAGE:	8	ppb		ON DAY
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	744 hrs
MONTHLY CALIBRATION TIME:	9	hrs	AMD OPERATION UPTIME:	100.0 %
STANDARD DEVIATION:	4		MONTHLY AVERAGE:	4 ppb

24 HR AVERAGES October 2018



NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)



Wind: LICA MASKWA
 Poll.: LICA MASKWA-NO2[ppb]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 20.20% Calm Avg: 3.86 [ppb]

Direction	0.0-10.3	10.3-20.7	20.7-31.0	>31.0	Total
N	7.1	0.0	0.0	0.0	7.1
NE	11.7	0.1	0.0	0.0	11.8
E	3.8	0.0	0.0	0.0	3.8
SE	7.7	0.0	0.0	0.0	7.7
S	9.1	0.0	0.0	0.0	9.1
SW	13.4	1.6	0.1	0.0	15.1
W	13.5	1.7	0.1	0.0	15.4
NW	7.8	1.7	0.3	0.0	9.8
Summary	74.1	5.1	0.6	0.0	79.8

% Icon Classes (ppb)


74 0.0-10.3

5 10.3-20.7

1 20.7-31.0

0 >31.0

LICA MASKWA Poll.: LICA MASKWA-NO2[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 20.20% Calm Poll Avg: 3.86[ppb]



NO2[ppb] Calibration: LICA MASKWA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



WIND SPEED



WIND SPEED Hourly Averages (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	4.0	3.5	1.5	0.8	0.7	1.3	0.4	0.4	0.0	1.2	1.8	1.5	0.8	3.7	3.6	3.1	2.3	2.7	2.9	2.4	2.0	0.3	0.6	2.3	0.0	4.0	0.8	24				
2	0.9	0.8	1.3	3.9	5.0	0.9	1.9	4.0	4.0	4.4	4.9	5.6	5.7	7.2	7.3	8.5	7.3	5.7	3.0	3.1	3.1	2.9	2.1	3.9	0.8	8.5	3.0	24				
3	4.4	2.8	3.4	4.2	4.5	5.6	4.5	6.1	5.8	7.1	6.8	6.7	6.8	6.8	5.9	6.0	6.2	4.0	2.5	2.4	1.2	1.7	0.6	3.8	0.6	7.1	3.0	24				
4	5.2	4.3	5.6	3.5	1.5	1.2	2.0	2.2	1.5	3.1	3.3	4.5	5.8	5.2	6.0	5.4	4.5	2.8	5.0	6.3	6.2	7.2	6.6	5.7	1.2	7.2	3.8	24				
5	4.6	5.0	5.9	7.5	6.7	3.4	2.5	3.2	3.0	4.3	4.7	5.1	4.8	4.6	4.5	4.2	4.9	5.3	3.6	4.0	1.7	0.9	0.8	2.6	0.8	7.5	1.5	24				
6	1.9	1.6	1.8	1.3	0.9	3.0	3.3	5.9	5.3	7.1	7.3	6.2	5.0	5.6	1.9	4.0	3.6	1.9	2.0	2.7	1.8	1.5	4.9	4.5	0.9	7.3	2.9	24				
7	6.2	7.0	5.8	7.3	7.5	8.3	8.9	8.6	7.4	6.4	6.2	7.4	8.2	8.0	9.2	7.4	7.4	8.0	10.9	11.5	9.3	8.9	8.7	10.8	5.8	11.5	8.1	24				
8	10.1	10.2	10.9	10.1	9.7	7.7	7.8	7.3	8.8	7.8	6.4	6.8	7.1	6.1	3.9	3.6	4.6	5.7	0.5	0.5	0.6	0.3	0.3	0.1	0.1	10.9	5.5	24				
9	0.6	0.5	0.3	0.3	0.9	1.9	1.3	1.6	3.0	4.0	4.5	4.1	2.6	4.1	4.0	2.2	3.2	3.0	3.0	1.5	1.0	1.3	0.6	0.4	0.3	4.5	1.8	24				
10	0.7	0.7	1.0	0.9	0.4	0.0	0.8	0.7	2.2	0.8	3.4	4.4	3.8	3.6	2.1	0.3	1.7	0.5	0.7	0.6	0.6	2.6	1.3	1.4	0.0	4.4	0.9	24				
11	1.6	1.1	0.4	1.5	2.8	2.9	1.9	2.9	3.6	5.0	5.2	4.9	6.6	8.2	7.6	8.0	8.8	5.7	6.6	7.6	7.2	6.3	5.9	6.1	0.4	8.8	4.9	24				
12	7.2	5.2	3.0	3.0	2.0	2.3	2.1	2.2	2.8	4.3	3.1	5.2	9.0	10.7	11.7	11.5	10.3	9.7	10.4	6.6	4.1	5.8	8.3	5.0	2.0	11.7	4.1	24				
13	5.7	6.8	6.1	5.8	7.4	6.3	6.5	6.9	7.2	7.1	7.3	6.8	7.5	7.4	6.6	6.7	6.1	5.6	4.8	2.6	2.2	1.6	2.8	1.9	1.6	7.5	5.2	24				
14	3.4	3.9	3.2	3.7	3.2	5.5	4.1	6.3	5.1	6.3	7.3	6.8	7.9	9.9	10.4	10.2	10.4	6.1	6.5	7.4	5.7	5.3	4.9	3.8	3.2	10.4	6.1	24				
15	7.5	7.8	6.7	7.0	7.2	6.0	6.5	6.9	8.4	8.5	9.4	8.4	9.0	8.9	9.0	7.5	6.6	4.6	4.2	4.5	4.5	5.4	4.4	2.3	2.3	9.4	6.5	24				
16	1.4	1.2	1.7	1.7	0.8	0.7	0.7	0.2	1.2	3.4	2.8	4.6	3.6	3.3	4.8	3.3	3.8	3.1	0.9	2.0	2.5	0.6	5.5	6.3	0.2	6.3	2.1	24				
17	3.3	1.9	4.2	4.4	4.5	3.8	4.5	5.3	7.0	8.3	10.2	6.2	6.5	5.4	5.4	6.0	4.9	6.5	7.0	6.1	4.1	4.1	3.6	5.0	1.9	10.2	5.2	24				
18	3.9	2.4	4.2	5.9	4.2	5.9	7.5	5.8	5.8	6.0	4.4	5.6	7.4	10.6	9.6	7.8	7.3	8.6	6.2	8.0	7.8	6.1	5.9	5.7	2.4	10.6	5.7	24				
19	3.9	3.4	3.0	2.2	3.1	4.0	1.5	4.1	6.4	6.4	7.6	9.4	7.8	7.4	6.2	5.2	5.7	3.2	3.1	3.1	2.9	1.4	3.1	5.8	1.4	9.4	2.2	24				
20	7.7	5.5	3.7	4.5	4.4	4.4	5.1	5.7	6.6	6.3	8.2	6.5	6.6	8.2	11.6	10.4	9.3	7.4	5.5	7.1	7.0	6.8	6.5	3.7	11.6	3.5	24					
21	5.0	4.3	4.4	3.9	2.7	3.3	4.0	3.5	7.6	10.0	10.3	9.8	10.3	10.1	9.2	9.0	7.1	4.1	2.8	2.3	1.9	1.8	0.7	0.5	0.5	10.3	5.1	24				
22	0.2	0.8	0.4	0.7	1.2	0.2	0.2	0.6	0.9	2.7	4.5	5.9	5.8	6.3	6.3	7.5	5.2	3.2	4.2	5.8	6.7	5.5	4.5	2.8	0.2	7.5	3.1	24				
23	3.5	2.3	2.0	2.3	1.7	2.6	2.2	3.1	3.8	3.6	5.7	5.2	6.1	5.4	4.5	3.9	3.3	3.0	3.8	0.9	0.4	1.2	0.5	0.4	0.4	6.1	2.3	24				
24	2.0	1.0	0.6	1.2	1.9	1.0	1.3	2.1	3.1	4.4	3.9	5.0	5.3	5.7	3.9	3.5	1.4	1.4	0.8	1.8	0.6	2.2	2.4	2.2	0.6	5.7	2.3	24				
25	0.7	2.6	3.0	2.5	1.5	1.7	2.2	0.2	1.3	2.6	2.9	1.1	1.8	3.3	3.4	1.8	1.7	0.7	0.8	0.7	0.5	0.4	0.4	0.2	0.2	3.4	1.0	24				
26	1.5	1.2	1.4	0.7	1.1	2.4	2.0	2.6	2.4	6.0	7.5	4.9	5.9	5.4	4.8	3.4	2.1	1.5	2.5	0.9	1.6	1.9	0.8	0.7	0.7	7.5	2.1	24				
27	1.3	2.0	1.9	4.3	2.9	2.5	5.6	4.7	4.4	8.7	8.2	7.9	9.8	10.5	9.5	7.0	4.8	2.6	1.9	6.1	5.8	4.3	3.7	2.2	1.3	10.5	4.6	24				
28	2.8	2.1	1.1	2.8	1.8	1.9	1.1	0.3	3.5	2.5	3.5	4.0	4.6	5.9	4.8	4.0	3.6	3.4	3.1	5.8	4.9	5.4	6.2	4.3	0.3	6.2	2.5	24				
29	3.5	4.6	3.4	2.5	5.8	4.5	4.7	5.4	4.3	4.0	5.6	5.6	6.2	7.6	7.4	7.7	6.7	4.8	5.2	5.1	4.5	3.9	2.9	2.2	2.2	7.7	3.8	24				
30	2.2	2.4	2.1	2.0	2.0	1.1	1.6	0.6	1.8	2.3	3.6	2.8	3.2	3.3	3.1	3.0	2.0	1.7	1.9	1.8	0.4	1.0	1.8	3.5	0.4	3.6	1.3	24				
31	2.5	3.2	3.7	4.7	4.2	2.5	4.2	3.5	2.8	2.1	2.5	4.2	4.5	5.0	5.3	3.6	3.4	1.9	1.0	0.5	0.5	1.0	0.7	1.1	0.5	5.3	2.4	24				
HOURLY MAX	10.1	10.2	10.9	10.1	9.7	8.3	8.9	8.6	8.8	10.0	10.3	9.8	10.3	10.7	11.7	11.6	10.4	9.7	10.9	11.5	9.3	8.9	8.7	10.8								

STATUS FLAG CODES


C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION:	September 17, 2018
DECLINATION :	MAGNETIC DECLINATION 19 DEGREE EAST


MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	742
MINIMUM 1-HR AVERAGE	0.0 kph @ HOUR 8 ON DAY 1
MAXIMUM 1-HR AVERAGE:	11.7 kph @ HOUR 14 ON DAY 12
MAXIMUM 24-HR AVERAGE:	8.1 kph ON DAY 7
MONTHLY CALIBRATION TIME:	0 hrs
OPERATIONAL TIME:	744 hrs
AMSD OPERATION UPTIME:	100.0 %
STANDARD DEVIATION:	2.6
MONTHLY AVERAGE:	0.9 kph

24 HR AVERAGES October 2018



WIND SPEED Hourly Averages (WS kph)



Wind: LICA MASKWA
 Monitor: WSP [kph]
 Monthly: 18/10
 Type: WindRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 20.56%

Direction	1.8-2.4	2.4-4.7	4.7-7.1	7.1-9.4	9.4-11.8	>11.8	Total
N	1.6	2.3	2.4	0.7	0.9	0.0	7.9
NE	1.9	3.5	2.0	3.0	1.1	0.0	11.4
E	0.4	2.6	0.8	0.1	0.0	0.0	3.9
SE	0.4	4.2	2.8	0.1	0.0	0.0	7.5
S	0.7	3.5	3.4	1.2	0.3	0.0	9.0
SW	1.6	7.5	4.4	0.8	0.4	0.0	14.8
W	1.8	4.3	4.6	3.0	1.5	0.0	15.1
NW	0.5	2.2	4.4	2.6	0.1	0.0	9.8
Summary	8.9	30.0	24.9	11.4	4.3	0.0	79.4

% Icon Classes (kph)

9 1.8-2.4

30 2.4-4.7


25 4.7-7.1

11 7.1-9.4

4 9.4-11.8

0 >11.8

LICA MASKWA 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 20.56% Calm Wind Avg Speed: 0.94(kph)



WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - October 2018

WIND DIRECTION Hourly Averages (WD)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	24-HOUR AVG	24-HR	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	QUADRANT	RDGS.	
DAY 1	SE	SE	SE	ESE	ESE	SE	SSE	NE	SSE	SW	NNW	NNE	W	N	N	NNW	NNW	N	NNE	N	N	N	WSW	NNE	NNE	24	
2	NNW	NNE	N	NNE	NNE	NE	NNE	N	NNW	NNW	WNW	W	W	WNW	WNW	WNW	W	W	WSW	WSW	WSW	W	WSW	SSW	WNW	24	
3	SW	SW	W	W	W	WNW	W	WNW	NNW	NNW	NNW	NNW	NW	NW	NW	NNW	NNE	NNE	NE	NE	E	E	ENE	SSE	NW	24	
4	SSE	SE	SE	SE	SE	NE	NE	NE	E	SE	SE	ESE	ESE	SE	SE	SE	ESE	ESE	SSE	SSE	SSE	S	S	S	SE	24	
5	SSW	SSW	SSW	SSW	SSW	SW	SW	SSW	WNW	NW	NW	NW	NW	NW	NNW	NNW	N	N	NNE	NNE	NE	ENE	ENE	NE	WNW	24	
6	ENE	NE	NE	NE	ENE	ESE	ESE	ESE	ESE	ESE	ESE	E	SE	E	ENE	ENE	ENE	ENE	NNE	NNE	NNE	NNE	NE	E	24		
7	NNE	NNE	NE	NE	NE	NNE	NNE	NNE	NNE	NE	NE	NE	NE	NE	NE	NE	NE	NE	NE	NE	ENE	ENE	NE	NNE	NE	24	
8	NNE	NNE	NNE	NNE	NNE	NNE	NNE	NE	NNE	NNE	NE	NE	NE	NE	ENE	ENE	NE	NNE	NNE	SSE	SSE	ESE	E	ENE	NE	24	
9	ESE	ENE	ENE	ENE	ENE	E	E	ENE	NNE	NE	ENE	ENE	E	ENE	SE	SE	ENE	ENE	NE	ENE	E	NNE	NE	E	ENE	24	
10	ESE	E	E	ENE	E	NE	ESE	NE	NNE	SE	SSE	SSE	SSE	SSE	SSW	SSE	ENE	E	SSE	SE	SSW	SW	SW	SSW	SSE	24	
11	S	SE	SSE	S	S	SSW	SSW	SSW	SSW	S	S	S	SSW	SSW	S	SSW	S	S	S	SSW	SSW	SSW	SSW	SSW	S	24	
12	SSW	SSW	SW	SW	SW	WSW	W	NW	NNW	N	N	N	N	N	N	N	N	N	N	N	N	NNW	NNW	N	NNW	NNW	24
13	NNW	NNW	NW	NW	NW	NW	NW	NW	NW	NW	NW	NW	NW	NNW	NNW	NNW	NNW	NW	NW	NW	WNW	WSW	SSW	SSW	NW	24	
14	SSW	SW	SW	SSW	SW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SW	SW	SW	SW	SSW	24
15	W	WNW	W	W	W	W	W	W	WNW	WNW	WNW	NW	NW	NW	NW	NW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	24
16	WNW	WNW	WNW	WNW	W	S	S	S	S	S	SW	SSW	SW	SW	SSW	SW	S	S	SSE	S	S	E	S	SSW	SSW	24	
17	SSW	SW	SSW	SSW	SW	SW	SW	SW	SW	SSW	SSW	SW	SW	WSW	WSW	WSW	SW	SW	SSW	SW	WSW	SW	SSW	SSW	SW	24	
18	SW	WSW	SW	SSW	SW	W	W	W	WNW	WNW	W	W	WNW	WNW	W	WNW	WNW	W	NW	WNW	W	WNW	WNW	WNW	WNW	W	24
19	WNW	WSW	WSW	WSW	WSW	W	NNW	W	WNW	NW	N	N	N	N	NW	N	NNE	ENE	ENE	E	E	ESE	SE	SE	NNW	24	
20	SE	SE	SE	SE	SE	ESE	SE	SSE	S	SSW	SSW	SSW	SW	W	WNW	W	WNW	WNW	W	W	W	W	W	W	W	WSW	24
21	W	WSW	W	W	WSW	WSW	W	WSW	W	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	W	W	W	SSE	SSW	24
22	S	SSE	SE	E	ENE	SSW	SE	E	NNE	S	S	S	S	SSE	SSE	S	S	SSE	SE	SSE	SSE	SSE	SE	SE	SSE	24	
23	SSE	SE	NE	ENE	E	ESE	SE	SSE	SSE	SSE	S	S	SSE	SSE	ESE	E	ENE	ESE	SE	SE	NE	N	E	ENE	SE	24	
24	ENE	ENE	NE	ENE	ENE	E	NE	NE	NE	NNE	NE	NNE	NE	NE	ENE	NE	E	SE	ENE	E	NE	NE	NE	NE	NE	24	
25	ENE	E	NE	NE	NE	ENE	ENE	NNE	NNE	NE	NNE	NW	NW	NE	SE	SSE	SE	SSE	ESE	ESE	ENE	NE	W	NNE	ENE	24	
26	ENE	NE	SSW	NNW	WSW	WNW	WSW	SW	WSW	WNW	W	W	WNW	WSW	WSW	WSW	SW	SSW	SSW	SSW	SSW	SSW	E	SSE	WSW	24	
27	SSW	SSW	SW	SW	WSW	SW	SW	SW	W	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	W	W	W	W	W	W	24
28	WSW	W	W	SW	SSE	SSW	SE	ESE	SSW	S	SSE	SSE	ESE	ESE	E	ENE	ESE	E	E	ESE	SE	ESE	SE	ESE	SE	24	
29	ENE	ENE	NE	NNE	NNE	N	N	N	NNW	NNW	NW	NW	WNW	WNW	WNW	WNW	WNW	WNW	NW	NW	NW	NNW	NNW	NNW	NNW	24	
30	NNW	NNW	NNW	N	NNW	NW	W	W	WSW	WSW	SW	SW	SSW	SSW	SSW	SSW	W	WSW	WSW	SW	NW	SSE	S	S	WSW	24	
31	S	SSW	S	SSW	SSW	SSW	SSW	SSW	SSW	SW	SW	SW	SSW	SW	SSW	W	WNW	ENE	ESE	NNE	WSW	SW	SSW	SSW	SSW	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION:	September 17, 2018
DECLINATION :	MAGNETIC DECLINATION 19 DEGREE EAST

MONTHLY CALIBRATION TIME:	0	hrs	OPERATIONAL TIME:	744	hrs
STANDARD DEVIATION:	102		AMD OPERATION UPTIME:	100.0	%
			MONTHLY AVERAGE:	284 (WNW)	

WIND DIRECTION Hourly Averages (WD)



WDR[degwdr] Station: LICA MASKWA Monthly: 18/10 Type: AVG 1 Hr. [1 Hr.]

— WDR[degwdr]



STANDARD DEVIATION WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - October 2018

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59		
DAY																										
1	8	12	15	34	18	23	40	29	73	33	44	60	72	34	19	29	30	10	10	13	14	31	45	12	24	
2	26	16	15	10	10	40	17	21	20	27	26	29	23	22	17	9	10	11	11	13	19	15	18	9	24	
3	6	15	9	7	11	8	11	19	22	20	18	23	22	25	31	23	19	10	11	13	26	21	36	11	24	
4	6	9	8	12	20	28	17	7	28	32	30	23	21	23	20	19	10	12	4	6	11	7	7	7	24	
5	8	8	5	6	7	11	11	9	28	16	19	18	22	28	30	26	15	14	8	7	12	14	18	12	24	
6	13	19	29	18	43	16	20	9	12	11	18	18	20	30	54	41	22	30	17	12	15	21	7	11	24	
7	7	7	15	9	10	6	5	5	7	15	17	12	15	15	13	14	13	12	6	7	14	14	9	4	24	
8	4	4	5	5	5	5	4	6	7	7	9	15	19	19	33	36	23	9	33	44	22	22	39	48	24	
9	28	41	21	25	13	19	35	20	9	19	27	41	52	34	29	52	17	12	7	15	14	19	30	25	24	
10	20	22	10	29	18	59	30	40	9	67	26	20	31	28	29	68	31	35	25	33	27	9	17	13	24	
11	21	21	66	34	14	12	20	13	13	14	13	17	14	10	13	10	8	9	10	8	7	8	7	6	24	
12	6	8	12	12	13	12	24	21	17	15	18	15	9	9	8	9	11	11	12	14	19	12	10	14	24	
13	14	13	12	13	11	11	11	12	11	12	12	15	17	21	21	17	16	11	11	19	19	32	13	16	24	
14	7	6	8	9	11	6	9	5	8	9	11	13	13	8	8	7	5	7	8	8	9	7	7	18	24	
15	5	5	6	6	6	6	4	5	7	11	12	13	16	12	12	14	17	13	9	11	10	15	17	24		
16	26	37	25	23	24	36	47	68	27	17	23	16	23	32	20	21	14	17	23	12	33	69	23	6	24	
17	14	22	14	9	5	5	5	5	5	5	4	14	11	18	15	14	10	5	5	7	13	17	8	6	24	
18	8	14	18	6	16	13	8	12	12	14	13	13	8	8	7	20	19	10	9	6	5	6	9	9	24	
19	17	10	9	11	13	21	50	10	12	15	16	15	19	22	14	27	11	14	18	14	19	34	17	11	24	
20	9	14	13	9	12	11	13	17	15	17	11	10	17	31	11	7	6	7	6	8	6	7	6	6	24	
21	7	20	14	10	11	12	7	14	9	8	7	9	8	9	13	11	13	15	9	19	33	30	40	37	24	
22	64	49	38	34	37	58	53	40	10	19	18	21	22	21	19	15	14	7	6	5	6	6	9	11	24	
23	11	31	14	18	20	24	28	13	15	22	18	18	21	22	21	35	11	15	9	32	50	6	53	44	24	
24	24	23	31	18	17	21	15	13	9	14	11	16	12	25	19	22	21	57	15	22	15	18	21	24	24	
25	32	14	19	14	26	18	18	37	8	14	25	56	42	40	20	20	31	57	49	57	62	48	56	39	24	
26	13	50	38	59	41	31	28	10	24	10	8	22	20	25	22	21	14	23	11	51	15	17	55	34	24	
27	18	27	27	9	15	10	6	7	19	9	10	9	9	10	12	13	33	17	8	9	8	17	12	24	24	
28	19	20	46	11	31	25	41	53	22	19	20	28	17	13	19	18	15	23	15	12	12	11	10	19	24	
29	11	16	12	17	5	9	8	9	15	11	13	14	12	9	9	8	8	10	12	15	19	14	20	21	24	
30	23	19	27	29	34	45	19	58	27	32	24	34	24	21	17	19	27	22	16	23	56	38	32	11	24	
31	17	14	15	8	17	16	8	16	14	15	23	21	22	17	8	35	21	27	28	65	34	34	20	30	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION: September 17, 2018

CALIBRATION TIME: 0 hrs OPERATIONAL TIME: 744 hrs

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)



RELATIVE HUMIDITY




RELATIVE HUMIDITY Hourly Averages (RH %)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.					
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.						
DAY																																	
1	59	60	61	65	67	70	72	74	69	61	56	48	44	42	43	44	46	51	59	65	71	78	83	82	42	83	61	24					
2	83	85	90	92	91	89	88	82	70	63	56	47	44	38	39	43	45	49	55	57	59	62	66	70	38	92	65	24					
3	74	76	77	78	74	69	68	68	69	62	55	49	41	37	34	34	36	42	49	53	61	67	74	75	34	78	59	24					
4	66	64	62	64	68	76	79	78	69	62	57	48	45	45	44	46	49	53	60	59	57	58	59	62	44	79	60	24					
5	66	71	74	75	78	83	88	88	79	68	60	51	51	50	46	48	54	62	67	70	75	79	83	87	46	88	69	24					
6	92	92	93	92	93	92	90	88	81	75	70	67	65	61	58	65	76	81	80	83	85	86	92	95	58	95	81	24					
7	96	96	98	97	96	98	97	96	96	91	85	83	81	80	80	78	78	80	81	84	81	85	96	96	78	98	89	24					
8	96	95	95	93	90	89	89	87	85	84	82	73	71	72	72	71	73	74	83	93	95	91	91	91	71	96	85	24					
9	91	92	89	90	90	92	92	90	89	82	66	57	44	35	29	32	42	48	53	60	67	76	81	85	29	92	70	24					
10	87	88	90	92	91	91	91	90	85	68	56	47	40	37	37	39	44	52	63	70	71	70	75	81	37	92	69	24					
11	85	83	86	86	84	85	85	86	86	85	83	80	77	75	74	74	75	77	77	81	83	89	99	74	99	83	24						
12	100	100	100	100	100	100	100	100	100	99	99	97	88	85	81	79	70	64	60	60	68	68	67	69	60	100	86	24					
13	70	68	71	72	74	75	75	74	74	71	66	59	55	50	45	43	47	54	59	64	69	74	82	87	43	87	66	24					
14	89	85	91	95	94	94	95	92	86	84	77	68	63	58	50	48	51	57	63	68	71	76	78	79	48	95	76	24					
15	70	67	66	65	66	69	71	70	63	56	47	40	38	41	41	42	47	52	60	70	78	83	87	92	38	92	62	24					
16	93	88	85	85	96	100	100	98	89	84	86	84	83	78	78	81	83	89	100	100	99	100	98	94	78	100	90	24					
17	95	98	98	94	94	85	79	71	59	50	47	39	24	16	14	15	18	22	25	26	28	29	32	36	14	98	50	24					
18	36	38	36	42	41	37	33	36	39	38	39	36	37	35	35	38	41	40	42	44	43	47	50	52	33	52	40	24					
19	58	66	71	79	81	81	85	83	70	60	60	64	68	62	58	54	55	61	59	54	54	62	59	49	49	85	65	24					
20	51	54	60	63	65	69	72	73	71	70	67	61	55	44	36	30	32	34	36	39	39	41	43	44	30	73	52	24					
21	47	52	55	57	61	65	66	69	57	47	41	37	35	34	34	35	38	42	51	58	68	71	82	88	34	88	54	24					
22	89	91	91	92	93	95	96	94	85	66	51	47	43	40	36	33	39	50	55	54	55	59	62	72	33	96	66	24					
23	73	73	83	88	87	84	77	72	64	55	49	44	39	34	34	35	41	51	49	59	72	76	83	86	34	88	63	24					
24	86	90	88	93	94	96	95	95	88	78	66	54	47	43	39	41	48	64	74	84	88	87	89	91	39	96	76	24					
25	93	96	93	93	95	98	99	99	97	83	77	65	56	54	60	62	70	81	89	90	87	91	95	94	54	99	84	24					
26	93	95	96	98	96	91	92	97	94	82	68	58	51	47	45	47	55	69	80	86	94	92	93	94	45	98	80	24					
27	89	89	90	79	80	84	83	88	83	72	68	67	67	66	63	63	66	69	72	71	76	82	87	90	63	90	77	24					
28	90	90	96	100	100	100	99	99	100	92	80	71	73	75	71	74	72	71	74	69	66	68	68	73	66	100	82	24					
29	86	94	98	99	100	100	100	100	100	100	100	100	100	100	98	96	96	97	98	97	98	98	98	99	86	100	98	24					
30	99	99	97	98	100	100	100	99	99	98	97	94	91	92	93	93	93	94	96	97	97	98	100	100	91	100	97	24					
31	100	100	100	100	100	100	100	100	99	97	82	70	66	63	69	68	71	81	86	89	91	97	100	100	63	100	89	24					
HOURLY MAX	100	100	100	100	100	100	100	100	100	100	100	100	100	98	96	96	97	98	100	100	99	100	100	100	100	100	100	100	24				
HOURLY AVG	81	82	83	84	85	86	86	85	80	74	68	62	58	54	53	53	56	62	66	69	72	75	79	81	63	100	89	24					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	14	%	@ HOUR	14	ON DAY	17
MAXIMUM 1-HR AVERAGE:	100	%	@ HOUR	0	ON DAY	12
MAXIMUM 24-HR AVERAGE:	98	%			ON DAY	29
OPERATIONAL TIME:						744 hrs
AMD OPERATION UPTIME:						100.0 %
STANDARD DEVIATION:	20					MONTHLY AVERAGE: 72 %

RELATIVE HUMIDITY Hourly Averages (RH %)



BAROMETRIC PRESSURE




BAROMETRIC PRESSURE Hourly Averages (BP mbar)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	940	939	938	938	937	937	937	937	937	937	938	938	938	938	938	938	938	938	939	939	939	940	940	940	937	940	938	24	
2	940	940	940	940	940	940	940	940	940	940	939	939	938	937	936	935	934	934	933	933	933	932	932	932	932	932	940	937	24
3	932	932	932	932	933	933	934	935	936	937	937	938	938	939	939	939	940	940	940	940	940	940	940	941	932	941	937	24	
4	941	941	940	940	940	940	939	939	940	940	940	939	939	939	938	938	937	937	936	936	936	936	936	935	935	941	938	24	
5	935	934	934	934	934	934	934	934	935	936	936	937	937	937	937	938	938	938	938	938	938	938	938	938	938	934	938	936	24
6	938	938	938	938	938	938	939	939	939	939	939	939	939	939	939	939	939	939	940	940	941	941	941	941	938	941	939	24	
7	942	943	943	943	944	944	944	945	945	946	946	946	946	946	946	947	947	947	947	947	947	947	948	947	942	948	946	24	
8	948	948	948	948	949	949	949	950	950	950	950	950	950	950	950	950	949	949	949	949	949	949	949	950	948	950	949	24	
9	950	950	950	950	950	951	951	952	952	953	954	954	953	953	953	952	952	952	951	951	951	951	951	950	950	954	952	24	
10	950	950	949	949	949	949	948	948	948	948	948	948	947	947	946	946	945	945	944	944	944	944	943	943	943	950	947	24	
11	942	942	941	941	940	940	939	939	939	939	938	937	936	935	934	934	933	932	931	930	930	929	928	928	942	936	24		
12	927	926	926	925	924	924	923	923	924	924	924	925	926	927	929	930	932	933	934	935	936	937	937	923	937	928	24		
13	938	939	939	939	940	940	941	942	942	943	944	944	944	944	945	945	945	945	946	946	946	946	946	946	938	946	943	24	
14	945	945	945	944	944	943	943	943	942	942	942	941	940	939	938	937	936	935	934	933	932	932	931	931	931	945	940	24	
15	932	932	932	933	933	934	935	935	936	937	937	938	938	938	938	939	939	939	940	940	940	941	942	942	932	942	937	24	
16	943	943	943	944	943	943	944	944	944	945	945	945	945	944	944	944	944	943	942	942	941	941	941	941	941	941	945	943	24
17	940	939	938	938	938	937	937	937	938	938	938	938	937	937	936	935	934	934	934	933	933	933	932	932	932	940	936	24	
18	932	932	932	932	932	933	933	934	934	935	935	936	936	936	936	936	937	938	938	938	939	940	940	941	932	941	936	24	
19	942	942	943	943	943	944	944	945	946	948	948	949	949	949	949	948	947	947	947	947	946	946	945	942	949	946	24		
20	945	944	943	942	940	939	938	937	937	937	936	936	935	935	935	934	934	935	935	935	936	936	936	936	934	945	937	24	
21	936	936	936	937	937	937	937	937	938	939	939	940	940	940	941	941	941	941	942	942	942	942	943	943	936	943	939	24	
22	944	944	944	944	944	944	944	945	946	946	946	946	946	945	945	944	943	943	943	943	942	942	942	942	942	946	944	24	
23	941	941	941	940	940	939	939	939	939	939	939	939	939	938	938	938	937	936	935	935	934	934	933	933	933	941	938	24	
24	933	933	932	932	932	931	932	931	932	932	933	933	932	932	932	931	930	930	930	930	930	929	929	929	929	933	931	24	
25	929	929	929	930	930	929	930	930	930	931	932	932	932	932	932	931	930	930	930	930	930	930	930	929	932	930	24		
26	930	930	930	930	930	930	930	930	930	931	931	931	930	930	930	929	929	928	927	926	926	926	926	926	926	931	929	24	
27	926	926	927	927	927	927	928	928	929	930	931	931	931	931	932	933	933	933	934	934	935	935	935	935	926	935	931	24	
28	936	936	936	936	935	935	935	934	935	935	935	935	934	933	933	932	931	931	931	930	930	929	928	928	928	936	933	24	
29	927	926	926	925	925	925	926	926	927	928	929	929	930	930	930	931	931	932	932	932	933	933	933	934	925	934	929	24	
30	934	934	934	934	934	934	935	935	935	935	935	935	935	935	935	935	935	935	935	935	935	935	934	934	934	935	935	24	
31	934	934	933	933	932	931	931	930	930	930	931	930	930	929	929	929	928	928	928	929	929	929	929	928	928	934	930	24	
HOURLY MAX	950	950	950	950	950	951	951	952	952	953	954	954	954	953	953	952	952	952	951	951	951	951	950						
HOURLY AVG	938	938	937	937	937	937	937	938	938	939	939	939	938	938	938	938	938	938	938	938	938	938	937						

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	923 mbar	@ HOUR	6	ON DAY	12
MAXIMUM 1-HR AVERAGE:	954 mbar	@ HOUR	10	ON DAY	9
MAXIMUM 24-HR AVERAGE:	952 mbar			ON DAY	9
OPERATIONAL TIME:					744 hrs
AMD OPERATION UPTIME:					100.0 %
STANDARD DEVIATION:	7			MONTHLY AVERAGE:	938 mbar

BAROMETRIC PRESSURE Hourly Averages (BP mbar)



AMBIENT TEMPERATURE




AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.					
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.						
DAY																																	
1	-1.2	-1.1	-1.1	-1.6	-1.7	-1.7	-1.9	-1.8	-0.8	0.4	1.6	3.1	3.8	4.4	3.8	3.8	3.4	2.6	1.4	0.9	0.2	-1.1	-2.2	-1.7	-2.2	-2.2	4.4	0.5	24				
2	-2.6	-3.7	-4.6	-3.7	-2.9	-2.5	-2.6	-2.5	-1.7	-1.1	0.1	1.6	1.9	3.1	3.2	2.1	1.7	0.5	-1.1	-1.6	-2.0	-2.5	-3.1	-3.6	-4.6	3.2	-1.1	24					
3	-4.3	-4.7	-4.6	-4.3	-3.1	-2.3	-2.4	-2.1	-1.4	-0.9	-0.6	-0.2	0.3	1.4	1.8	1.6	0.5	-0.9	-2.8	-3.6	-5.3	-6.6	-7.4	-6.8	-7.4	1.8	-2.5	24					
4	-5.5	-5.5	-5.2	-5.4	-6.2	-7.2	-7.5	-6.9	-3.7	-1.2	0.5	2.9	4.2	4.7	5.0	5.0	4.8	3.5	1.7	1.8	2.0	1.6	0.9	0.2	-7.5	5.0	-0.7	24					
5	-0.6	-1.3	-1.1	-0.7	-0.9	-1.7	-2.8	-2.4	-0.3	1.7	3.0	4.6	4.4	4.9	5.7	5.4	3.7	1.6	0.4	-0.3	-1.6	-2.7	-3.7	-4.5	-4.5	5.7	0.5	24					
6	-4.6	-4.8	-5.1	-5.1	-4.9	-3.2	-2.5	-1.8	-1.1	-0.1	1.2	2.1	2.6	3.7	4.7	3.9	2.7	2.3	1.9	1.4	1.2	0.9	0.1	-0.6	-5.1	4.7	-0.2	24					
7	-0.6	-1.1	-1.7	-1.7	-1.7	-2.1	-2.6	-2.7	-2.2	-1.7	-1.3	-1.1	-0.9	-0.7	-0.8	-0.7	-0.9	-1.3	-1.8	-2.4	-2.7	-2.9	-3.5	-4.2	-4.2	-0.6	-1.8	24					
8	-4.6	-5.0	-5.3	-5.5	-5.8	-6.5	-6.9	-6.1	-5.3	-4.6	-4.2	-3.6	-3.6	-3.8	-3.9	-3.9	-4.0	-4.5	-6.3	-8.6	-9.5	-9.2	-9.1	-9.2	-9.2	-4.5	-5.8	24					
9	-10.4	-11.3	-12.0	-12.4	-12.7	-12.6	-12.4	-10.8	-7.3	-3.7	-1.5	-0.4	0.8	1.4	2.4	2.6	1.5	-0.4	-2.1	-3.6	-4.8	-6.1	-7.1	-7.9	-12.7	2.6	-5.5	24					
10	-8.7	-9.3	-9.8	-10.3	-10.6	-11.0	-11.4	-10.4	-6.3	-1.6	0.2	0.7	1.2	1.3	1.5	1.7	1.6	-0.1	-2.4	-3.6	-3.2	-2.4	-2.7	-3.5	-11.4	1.7	-4.1	24					
11	-3.8	-3.0	-3.1	-2.5	-1.9	-2.0	-1.8	-1.7	-1.4	-1.0	-0.7	-0.1	0.7	1.4	1.6	1.8	1.7	1.2	1.0	0.8	0.6	0.5	0.4	0.0	-3.8	1.8	-0.5	24					
12	0.0	0.1	0.1	0.2	0.2	0.3	0.4	0.9	1.4	1.6	1.8	2.0	2.0	1.1	0.5	-0.2	-0.7	-1.4	-1.7	-2.4	-2.8	-3.2	-4.2	-5.0	-5.0	2.0	-0.4	24					
13	-5.3	-5.0	-5.3	-5.6	-5.8	-5.9	-5.6	-5.3	-5.0	-4.4	-3.3	-1.8	-1.0	0.1	0.9	1.1	0.4	-0.9	-1.9	-2.8	-3.6	-4.9	-6.5	-7.3	-7.3	1.1	-3.5	24					
14	-7.4	-6.6	-7.3	-7.5	-7.2	-7.1	-7.6	-6.4	-4.0	-2.7	-0.2	2.7	4.4	5.6	6.4	6.9	6.3	5.4	5.1	4.9	4.3	3.3	3.0	2.8	-7.6	6.9	-0.1	24					
15	4.5	4.9	4.7	4.8	4.6	4.1	4.1	4.6	6.6	8.7	9.8	10.9	11.4	10.9	10.8	10.6	9.4	8.1	7.2	6.1	5.3	5.2	4.7	3.6	3.6	11.4	6.9	24					
16	2.8	2.8	2.7	2.4	-0.1	-1.9	-2.1	-0.9	1.8	4.1	5.2	6.1	6.7	7.7	8.0	7.9	7.7	6.4	3.6	2.6	2.4	0.1	2.6	4.6	-2.1	8.0	3.5	24					
17	3.7	1.8	1.3	1.9	1.1	1.5	2.1	3.2	5.7	7.8	9.2	13.1	17.4	20.6	21.8	21.5	19.4	17.0	15.7	15.3	14.2	13.3	11.1	9.4	1.1	21.8	10.4	24					
18	9.1	8.0	8.6	7.4	7.6	9.8	11.2	10.7	10.2	12.2	12.3	13.5	13.3	14.3	14.1	13.3	12.2	10.9	9.4	8.8	8.7	7.6	7.6	7.1	7.1	14.3	10.3	24					
19	5.8	3.6	2.5	1.1	1.0	1.1	0.1	0.5	3.4	5.5	6.8	6.4	5.5	5.9	6.2	6.4	5.0	2.4	0.2	-0.9	-1.3	-2.5	-2.1	-1.2	-2.5	6.8	2.6	24					
20	-1.2	-1.4	-2.1	-2.4	-1.9	-2.0	-1.5	-0.1	0.6	1.8	3.2	5.2	7.3	11.2	13.3	13.7	12.6	11.5	10.2	9.2	9.1	8.6	7.9	7.4	-2.4	13.7	5.0	24					
21	6.4	5.2	4.3	4.0	3.0	2.1	1.9	1.4	4.5	6.8	8.3	9.5	10.5	11.2	11.7	11.3	10.2	8.7	5.7	3.9	1.8	1.2	-1.1	-2.3	-2.3	11.7	5.4	24					
22	-2.7	-3.5	-4.0	-4.6	-5.2	-5.4	-5.8	-5.6	-2.9	2.8	6.3	7.6	9.2	10.4	11.3	11.0	9.4	6.4	5.4	5.6	5.4	4.3	3.4	1.1	-5.8	11.3	2.5	24					
23	0.7	0.6	-1.3	-2.1	-2.4	-2.3	-1.3	-0.3	1.5	3.8	5.8	7.8	9.9	11.7	12.3	11.3	8.9	5.8	6.4	4.1	1.0	-0.1	-1.4	-2.4	-2.4	12.3	3.2	24					
24	-2.9	-2.8	-3.3	-4.0	-4.5	-4.6	-4.9	-4.9	-2.0	1.5	5.4	8.3	9.9	11.0	12.7	12.2	9.7	5.8	3.0	0.7	-0.4	-0.6	-1.2	-1.1	-4.9	12.7	1.8	24					
25	-1.6	-2.1	-1.3	-1.3	-1.9	-2.7	-2.7	-2.9	-2.0	1.5	3.4	6.2	8.4	9.0	7.4	7.2	5.3	2.8	0.8	0.2	0.7	-0.3	-0.5	0.0	-2.9	9.0	1.4	24					
26	0.1	-0.4	-0.2	-0.7	0.3	1.5	1.8	1.3	3.1	5.7	7.3	8.6	9.5	10.4	10.5	9.8	7.4	4.0	1.6	0.0	-1.6	-1.4	-2.2	-1.8	-2.2	10.5	3.1	24					
27	-1.0	-0.5	-0.9	0.3	0.2	-0.1	0.4	-0.1	1.6	5.0	5.7	5.4	5.6	6.2	6.9	6.4	6.0	5.4	4.8	5.1	4.7	3.6	3.0	2.9	-1.0	6.9	3.2	24					
28	3.1	2.8	0.8	-0.4	-2.0	-2.4	-3.2	-3.8	-2.2	1.5	4.2	5.7	5.3	5.3	5.9	5.7	5.2	4.6	4.3	4.8	4.8	4.9	4.7	4.3	-3.8	5.9	2.7	24					
29	3.4	2.9	2.6	2.2	1.3	0.7	0.2	0.3	0.5	0.9	1.5	2.1	2.5	2.8	3.1	3.0	2.9	2.7	2.6	2.3	2.2	2.0	1.7	1.6	0.2	3.4	2.0	24					
30	1.6	1.5	1.3	0.9	0.8	0.5	0.5	0.7	0.8	1.2	1.4	1.8	2.5	2.8	2.8	3.0	2.9	2.7	2.7	2.7	2.8	2.5	1.9	1.9	0.5	3.0	1.8	24					
31	1.8	1.6	1.6	1.6	1.5	1.1	1.0	1.0	1.1	1.7	4.3	6.4	7.3	7.2	6.1	6.0	5.4	3.8	2.8	2.4	2.4	2.5	2.4	2.4	2.4	1.0	7.3	3.1	24				
HOURLY MAX	9.1	8.0	8.6	7.4	7.6	9.8	11.2	10.7	10.2	12.2	12.3	13.5	17.4	20.6	21.8	21.5	19.4	17.0	15.7	15.3	14.2	13.3	11.1	9.4									
HOURLY AVG	-0.8	-1.2	-1.6	-1.8	-2.0	-2.1	-2.1	-1.8	-0.2	1.7	3.1	4.4	5.3	6.0	6.4	6.2	5.2	3.8	2.5	1.7	1.1	0.5	-0.1	-0.4									

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

MINIMUM 1-HR AVERAGE:	-12.7 °C	@ HOUR	4	ON DAY	9
MAXIMUM 1-HR AVERAGE:	21.8 °C	@ HOUR	14	ON DAY	17
MAXIMUM 24-HR AVERAGE:	10.4 °C			ON DAY	17
OPERATIONAL TIME:				744	hrs
AMD OPERATION UPTIME:				100.0	%
STANDARD DEVIATION:	5.3	MONTHLY AVERAGE:		1.4	°C

AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)



PRECIPITATION


PRECIPITATION Hourly Totals (mm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.										
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	SUM											
DAY																																						
1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
6	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
7	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.2	0.2	0.0	0.0	0.2	0.4	0.0	0.0	0.0	0.0	0.0	0.0	24		
8	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
9	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
10	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
11	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.1	0.1	24	
12	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	C	C	C	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24	
14	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
15	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
17	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
19	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
20	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
22	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
24	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
25	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
26	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
27	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
28	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
29	0.0	0.0	0.2	0.3	0.8	0.8	0.5	0.4	0.6	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
30	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
31	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
HOURLY MAX	0.2	0.1	0.2	0.3	0.8	0.8	0.5	0.4	0.6	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
HOURLY SUM	0.4	0.2	0.2	0.3	0.8	0.9	0.5	0.4	0.7	0.3	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

24 HR TOTALS October 2018



PRECIPITATION Hourly Totals (mm)



***APPENDIX II
EQUIPMENT CALIBRATION RESULTS***

SULPHUR DIOXIDE



Thermo 431-TLE Sulphur Dioxide Analyzer Calibration

Date:	October 12, 2018	Barometer/B.P./units:	F.S. 05544 expires January 15, 2019	923	millibars
Company/Airshed:	LICA	Thermometer/Station Temp:	F.S. 170286131 expires April 19, 2019	22	°C
Location/Station Name:	Maskwa	Weather Conditions:	Cloudy/Overcast		
Parameter:	Sulphur Dioxide	Calibration Purpose:	routine monthly		
Start Time 24 hr. (mst):	9:23	Performed By/Reviewer:	Alex Yakupov	Rob Fisher	
End Time 24 hr. (mst):	15:11	Cal Gas Expiry Date:	October 24, 2020		
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	n/a		
Analyzer:					
Serial Number/Owner:	1180930031 LICA	Range ppb:	1000		
Last Calibration Date:	September 2, 2018	As Found C.F.:	1.020		
Previous C.F.:	1.000	New C.F.:	1.000		

Calibration Standards: Low Flow Meter ID/Expiry Date: Defender Low 152019 expires December 13, 2018 High Flow Meter ID/Expiry Date: Defender High 148944 expires December 13, 2018 Calibrator ID/Expiry Date: API id# 690 expires March 15, 2019 Cal Gas Cylinder I.D. #: LL 104225 Cal Gas Conc. (ppm): 49.2	Standard Calibration Points for Ranges <table border="1"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>780</td></tr> <tr><td>Mid</td><td>380</td></tr> <tr><td>Low</td><td>190</td></tr> </table>	Point	ppb	High	780	Mid	380	Low	190
Point	ppb								
High	780								
Mid	380								
Low	190								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration (ppb):	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total			
as found zero	4775	0.00	4775	0.0	0.1	n/a
as found high	4883	75.96	4959	753.6	739	1.020
adjusted zero	4775	0.00	4775	0.0	0	n/a
adjusted high	4883	75.96	4959	753.6	754	1.000
mid	4925	36.58	4962	362.7	365	0.994
low	4933	18.38	4951	182.6	182	1.004
calibrator zero	4775	0.00	4775	0.0	0	n/a
Average C.F. =						0.999

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	0.999		0.95-1.05
b (Intercept as % of full scale) =	-0.01%		± 3% F.S.
% change in C.F. from last cal =	-1.99%		± 10%

Thermo 431-TLE Sulphur Dioxide Analyzer Calibration




As found:		As left:	
Bkg:	1.88	Bkg:	2.16
Coef:	0.963	Coef:	0.977
Pmt:	-701.5	Pmt:	-700.8
Flash:	980	Flash:	980
Internal:	30.7	Internal:	30.6
Chamber:	45.2	Chamber:	44.9
Perm Oven Gas:	35.00	Perm Oven Gas:	35.00
Perm Oven Heater:	34.26	Perm Oven Heater:	34.26
Pressure:	661.0	Pressure:	663.1
Sample Flow:	0.453	Sample Flow:	0.455
Lamp Intensity:	92	Lamp Intensity:	92
Converter:	n/a	Converter:	n/a
Converter Set:	n/a	Converter Set:	n/a
Averaging Time:	120	Averaging Time:	120
Expected Value:	431.0	Expected Value:	439.0

Comments:

The analyzer sample inlet filter was changed. The analyzer cooling fan filter(s) were cleaned. The manifold blower was found to be working normally.

— SO2[ppb]



HYDROGEN SULPHIDE



Thermo 450i Hydrogen Sulphide Analyzer Calibration

Date:	October 12, 2018	Barometer/B.P./units:	F.S. 05544 expires January 15, 2019	923	millibars
Company/Airshed:	LICA	Thermometer/Station Temp:	F.S. 170286131 expires April 19, 2019	22	°C
Location/Station Name:	Maskwa	Weather Conditions:	Cloudy/Overcast		
Parameter:	Hydrogen Sulphide	Calibration Purpose:	routine monthly		
Start Time 24 hr. (mst):	14:19	Performed By/Reviewer:	Alex Yakupov	Rob Fisher	
End Time 24 hr. (mst):	21:09	Cal Gas Expiry Date:	October 20, 2020		
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	n/a		
Analyzer:					
Serial Number/Owner:	CM 17360005 LICA	Range ppb:	100		
Last Calibration Date:	September 3, 2018	As Found C.F.:	1.016		
Previous C.F.:	1.000	New C.F.:	1.000		

Calibration Standards: Low Flow Meter ID/Expiry Date: Defender Low 152019 expires December 13, 2018 High Flow Meter ID/Expiry Date: Defender High 148944 expires December 13, 2018 Calibrator ID/Expiry Date: Sabio id# 11900613 expires August 22, 2019 Cal Gas Cylinder I.D. #: EY 0001003 Cal Gas Conc. (ppm): 9.55	Standard Calibration Points for Ranges <table border="1"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19	SO2 Scrubber Check (10 minutes): Start/End Time 24 hr.: 14:22 / 14:38 SO2 Analyzer Range: 1000 Target Concentration (ppb): 780 As Found Zero: 1.4 Analyzer Response: (ppb): 1.4 Zero Corrected Result (ppb): 0.0
Point	ppb									
High	78									
Mid	38									
Low	19									


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration (ppb):	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total			
as found zero	7543	0.00	7543	0.0	1.4	n/a
as found high	7483	61.21	7544	77.5	77.7	1.016
adjusted zero	7543	0.00	7543	0.0	0.0	n/a
adjusted high	7483	61.21	7544	77.5	77.5	1.000
mid	7549	29.90	7579	37.7	37.6	1.002
low	7557	14.97	7572	18.9	18.9	0.999
calibrator zero	7543	0.00	7543	0.0	0.0	n/a
Average C.F. =						1.000

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	1.000		0.95-1.05
b (Intercept as % of full scale) =	0.03%		± 3% F.S.
% change in C.F. from last cal =	-1.55%		± 10%

Thermo 450i Hydrogen Sulphide Analyzer Calibration




As found:		As left:	
Bkg:	16.5	Bkg:	18.4
Coef:	0.851	Coef:	0.871
Pmt:	-602.0	Pmt:	-602.4
Flash:	811	Flash:	810
Internal:	34.4	Internal:	32.8
Chamber:	45.0	Chamber:	45.1
Converter Temp:	323.3	Converter Temp:	323.1
Converter Set:	325.0	Converter Set:	325.0
Perm Oven Gas:	35.00	Perm Oven Gas:	35.00
Perm Oven Htr:	34.31	Perm Oven Htr:	34.29
Pressure:	564.9	Pressure:	570.1
Sample Flow:	0.944	Sample Flow:	0.948
Lamp Intensity:	91	Lamp Intensity:	91
Averaging Time:	120	Averaging Time:	120
Expected Value:	54.0	Expected Value:	53.6

Comments:

The analyzer sample inlet filter was changed. The analyzer cooling fan filter(s) were cleaned. The manifold blower was found to be working normally.

H2S[ppb] Station: LICA MASKWA Daily: 18/10/12 Type: AVG 1 Min. [1 Min.]

— H2S[ppb]



TOTAL HYDROCARBON



Thermo 55i Methane/Non-Methane Analyzer Calibration

Date: October 3, 2018	Barometer/B.P./units: F.S. 05544 expires January 15, 2019	936	millibars
Company/Airshed: LICA	Thermometer/Station Temp: F.S. 170286131 expires April 19, 2019	22	°C
Location/Station Name: Maskwa	Weather Conditions: Mix of sun and clouds		
Parameter: CH4 / NMHC / THC	Calibration Purpose: shut down		
Start/End Time 24 hr. (mst): 9:08 / 11:02	Performed By/Reviewer: Alex Yakupov	Rob Fisher	
Calibration Method: Gas Dilution	Cal Gas Expiry Date: October 18, 2025		

Analyzer: Serial Number/Owner: 1236656107 LICA Measured Flow: 1108 Last Calibration Date: September 3, 2018 Range ppm: 20 CH4/20 NMHC/40 THC	Correction Factors: <table border="1" style="margin: auto;"> <tr> <th></th> <th>Previous C.F.:</th> <th>As Found C.F.:</th> <th>New C.F.:</th> </tr> <tr> <td>CH₄ =</td> <td>1.000</td> <td>1.029</td> <td>n/a</td> </tr> <tr> <td>NMHC =</td> <td>1.000</td> <td>0.996</td> <td>n/a</td> </tr> <tr> <td>THC =</td> <td>1.000</td> <td>1.013</td> <td>n/a</td> </tr> </table>		Previous C.F.:	As Found C.F.:	New C.F.:	CH ₄ =	1.000	1.029	n/a	NMHC =	1.000	0.996	n/a	THC =	1.000	1.013	n/a
	Previous C.F.:	As Found C.F.:	New C.F.:														
CH ₄ =	1.000	1.029	n/a														
NMHC =	1.000	0.996	n/a														
THC =	1.000	1.013	n/a														

Calibration Standards: Low Flow Meter ID/Expiry Date: Defender Low 152019 expires December 13, 2018 High Flow Meter ID/Expiry Date: Defender High 148944 expires December 13, 2018 Calibrator ID/Expiry Date: API id# 690 expires March 15, 2019 Cal Gas Cylinder I.D. #: LL 119471 CH ₄ Cylinder Conc.: 599.0 207.0 =C ₂ H ₆ Cylinder Conc. CH ₄ expressed as C ₂ H ₆ : 569.3 1168.3 =total CH ₄ equivalent	Standard Calibration Points for Analyzer Range of 20/20/40 ppm <table border="1" style="margin: auto;"> <tr> <th>Point</th> <th>CH₄</th> <th>NMHC</th> <th>THC</th> </tr> <tr> <td>High</td> <td>13.00</td> <td>13.00</td> <td>26.00</td> </tr> <tr> <td>Mid</td> <td>7.00</td> <td>7.00</td> <td>14.00</td> </tr> <tr> <td>Low</td> <td>3.00</td> <td>3.00</td> <td>6.00</td> </tr> </table>	Point	CH ₄	NMHC	THC	High	13.00	13.00	26.00	Mid	7.00	7.00	14.00	Low	3.00	3.00	6.00
Point	CH ₄	NMHC	THC														
High	13.00	13.00	26.00														
Mid	7.00	7.00	14.00														
Low	3.00	3.00	6.00														

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	Correction Factors:		
	Diluent	Cal Gas	Total Flow							CH ₄	NMHC	THC
as found zero	3478	0.00	3478	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
as found high	3475	78.13	3553	13.17	12.52	25.69	12.80	12.57	25.37	1.029	0.996	1.013
mid	3486	42.97	3529	7.29	6.93	14.22	7.11	7.04	14.15	1.026	0.985	1.005
low	3491	19.03	3510	3.25	3.09	6.33	3.21	3.10	6.31	1.012	0.996	1.004
Average C.F. =										1.022	0.992	1.007

Linear Regression/Calibration Results:

Correlation Coefficient = 1.000 Slope = 0.971 b (Intercept as % of full scale) = 0.13% % change in C.F. from last cal = -2.91%	<table border="1" style="margin: auto;"> <tr> <th>CH₄</th> <th>NMHC</th> <th>THC</th> </tr> <tr> <td>1.000</td> <td>1.005</td> <td>0.988</td> </tr> <tr> <td>0.971</td> <td>1.005</td> <td>0.988</td> </tr> <tr> <td>0.13%</td> <td>0.07%</td> <td>0.10%</td> </tr> <tr> <td>-2.91%</td> <td>0.42%</td> <td>-1.26%</td> </tr> </table>	CH ₄	NMHC	THC	1.000	1.005	0.988	0.971	1.005	0.988	0.13%	0.07%	0.10%	-2.91%	0.42%	-1.26%	LIMITS > or = 0.995 0.90-1.10 ± 3% F.S. ± 10%
CH ₄	NMHC	THC															
1.000	1.005	0.988															
0.971	1.005	0.988															
0.13%	0.07%	0.10%															
-2.91%	0.42%	-1.26%															

As Left Instrument Diagnostics:

Interface Board Voltages: Bias Supply: -292.8 Temperatures: Detector Oven: 175.1 Filter: 175.1 Column Oven: 75.0 Internal: 32.1 Cylinder Pressures/reg.: Carrier: 1500 50 Fuel: 1400 55 Span Gas: 1500 13 Zero Air Generator: 50 Internal Pressures: Carrier: 31.1 Fuel: 40.3 Air: 30.8 FID Status: Status: UT Counts: 26316 Flame: 360.3 Det Base: 175.1 Flame and Power Stats: Last Power On: Sep 2, 2018 Flameouts: 1 Det Oven at Start: 35.7 Col Oven at Start: 32.0 Calibration History: Time: Sep 3, 2018 / 11:15 Type: SPAN Status: GOOD Check/Adjust: ADJUST CH ₄ Span Conc: 14.45 CH ₄ SP Ratio: 0.000766 CH ₄ RT: 13.4 CH ₄ PK IDX: 27 CH ₄ PK HT: 18874 NM Span Conc: 13.72 NM SP Ratio: 0.000151	Calibration History cnt'd: NM Peak Area: 90183 Crucial Settings: Methane Start: n/a Methane End: n/a Backflush: n/a NMHV Start: n/a NMHC End: n/a Run History>1: Date: Oct 3, 2018 Time: 09:18 CH ₄ PK HT: 0 CH ₄ RT: 8.0 CH ₄ Baseline: 2266 CH ₄ LOD: 58 CH ₄ SD: 19 CH ₄ CONC: 0.00 NM PK HT: 0 NM Peak Area: 0 NM CONC: 0.00 NM Base Start: 2196 NM Base End: 2206 NM LOD: 7 NM Start IDX: 41 NM End IDX: 88 NM Max Slope: 5.7e-01 NM Min Slope: -3.4e-01 NM PT Count: 0 Previous CH ₄ : 10.13 Previous NMHC: 10.96 Previous THC: 21.08 New CH ₄ : n/a New NMHC: n/a New THC: n/a	Expected Values: CH ₄ RT: 13.4 CH ₄ PK IDX: 27 CH ₄ PK HT: 18874 NM Span Conc: 13.72 NM SP Ratio: 0.000151
---	---	--

Comments:




No zero adjustment was required/made. As found zero values were copied to adjusted zero values for linearity calculation purposes.

The manifold blower was found to be working normally.

Shutdown calibration was completed to re-install the analyzer at the Bonnyville monitoring site.


Date: October 3, 2018
Company/Airshed: LICA
Location/Station Name: Maskwa

Start/End Time 24 hr. (mst): 9:08 / 11:02
Calibration Purpose: shut down
Calibration Method: Gas Dilution



Station: LICA MASKWA Daily: 18/10/03 Type: AVG 1 Min. [1 Min.]

CH4[ppm] NMHC[ppm]





Thermo 51C Total Hydrocarbon Analyzer Calibration

Date: October 03, 2018 Company/Airshed: LICA Location/Station Name: Maskwa Parameter: Total Hydrocarbon Start/End Time 24 hr. (mst): 14:01 / 17:35 Calibration Method: Gas Dilution	Barometer/B.P./units: F.S. 05544 expires January 15, 2019 936 millibars Thermometer/Station Temp: F.S. 170286131 expires April 19, 2019 22 °C Weather Conditions: Mix of sun and clouds Calibration Purpose: installation Performed By/Reviewer: Alex Yakupov Rob Fisher Cal Gas Expiry Date: October 18, 2025
Analyzer: Serial Number/Owner: 46609738 LICA Last Calibration Date: n/a Previous Cal High Point C.F.: n/a	Range ppm: 50 As Found C.F.: n/a New C.F.: 1.000

Calibration Standards:

Low Flow Meter ID/Expiry Date: Defender Low 152019 expires December 13, 2018
 High Flow Meter ID/Expiry Date: Defender High 148944 expires December 13, 2018
 Calibrator ID/Expiry Date: API id# 690 expires March 15, 2019
 Cal Gas Cylinder I.D. #: LL 119471


CH₄/C₂H₆ Cylinder Conc. (ppm): CH ₄ as propane/total CH ₄ equivalents (ppm):	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th>Point</th> <th>Target ppm</th> </tr> <tr> <td>High</td> <td>38</td> </tr> <tr> <td>Mid</td> <td>18</td> </tr> <tr> <td>Low</td> <td>9</td> </tr> </table>	Point	Target ppm	High	38	Mid	18	Low	9
Point	Target ppm								
High	38								
Mid	18								
Low	9								

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration:	Indicated Concentration:	Correction Factors:
Point	Diluent	Cal Gas	Total	(ppm)	(ppm)	
adjusted zero	2478	0.00	2478	0.0	0.00	n/a
adjusted high	2480	82.15	2562	37.46	37.46	1.000
mid	2479	38.94	2518	18.07	18.02	1.003
low	2484	19.46	2503	9.08	9.02	1.007
calibrator zero	2478	0.00	2478	0.00	0.00	n/a
Average C.F.=						1.003

Linear Regression/Calibration Results:

Correlation Coefficient = 1.000 Slope = 1.000 b (Intercept as % of full scale) = 0.07% % change in C.F. from last cal = n/a	LIMITS > or = 0.995 0.95-1.05 ± 3% F.S. n/a
--	--



Cylinder/Regulator Pressures:	As found:	As left:
H2 Cylinder (psi):	n/a	1400
H2 cylinder reg set (psi):	n/a	30
Zero Air Gen Pressure:	n/a	39
Span Cylinder (psi):	n/a	2000
Span Cylinder reg set (psi):	n/a	24
Measured Flow:	n/a	0.884
Expected Value:	n/a	26.77


Comments:
 The analyzer sample inlet filter was changed.

A new span gas cylinder was installed.
 The analyzer cooling fan filter(s) were cleaned. The manifold blower was found to be working normally.

The THC analyzer was installed to replace a 55i NMHC analyzer.

THC[ppm] Station: LICA MASKWA Daily: 18/10/03 Type: AVG 1 Min. [1 Min.]

THC[ppm]





Thermo 51C Total Hydrocarbon Analyzer Calibration

Date:	October 12, 2018	Barometer/B.P./units:	F.S. 05544 expires January 15, 2019	923	millibars
Company/Airshed:	LICA	Thermometer/Station Temp:	F.S. 170286131 expires April 19, 2019	22	°C
Location/Station Name:	Maskwa	Weather Conditions:	Cloudy/Overcast		
Parameter:	Total Hydrocarbon	Calibration Purpose:	repeat		
Start/End Time 24 hr. (mst):	9:23 / 14:59	Performed By/Reviewer:	Alex Yakupov	Rob Fisher	
Calibration Method:	Gas Dilution	Cal Gas Expiry Date:	October 18, 2025		
Analyzer:		Range ppm:	50		
Serial Number/Owner:	436609738 LICA	As Found C.F.:	0.992		
Last Calibration Date:	October 3, 2018	New C.F.:	1.000		
Previous Cal High Point C.F.:	1.000				

Calibration Standards:

Low Flow Meter ID/Expiry Date: Defender Low 152019 expires December 13, 2018
 High Flow Meter ID/Expiry Date: Defender High 148944 expires December 13, 2018
 Calibrator ID/Expiry Date: Sabio id# 11900613 expires August 22, 2019
 Cal Gas Cylinder I.D. #: LL 119471

CH ₄ /C ₂ H ₆ Cylinder Conc. (ppm):	599.0	207.0
CH ₄ as propane/total CH ₄ equivalents (ppm):	569.3	1168.3

Standard Calibration Points for a Range of: 50 ppm	
Point	Target ppm
High	38
Mid	18
Low	9


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated Concentration: (ppm)	Indicated Concentration: (ppm)	Correction Factors:
	Diluent	Cal Gas	Total			
as found zero	2492	0.00	2492	0.0	-0.42	n/a
as found high	2415	81.40	2496	38.10	38.00	0.992
adjusted zero	2492	0.00	2492	0.00	0.00	n/a
adjusted high	2415	81.40	2496	38.10	38.10	1.000
mid	2457	38.73	2496	18.13	18.10	1.002
low	2450	19.44	2469	9.20	9.03	1.019
calibrator zero	2492	0.00	2492	0.0	0.00	n/a
Average C.F.=						1.007

Linear Regression/Calibration Results:

Correlation Coefficient = 1.000
 Slope = 0.999
 b (Intercept as % of full scale) = 0.15%
 % change in C.F. from last cal = 0.84%

LIMITS
 > or = 0.995
 0.95-1.05
 ± 3% F.S.
 ± 10%



As found:	As left:
measurement alarms: None	measurement alarms: None
service alarms: None	service alarms: None
cnt: 2520	cnt: 2552
rng: 1	rng: 1
try: 2	try: 2
flm: 212.2	flm: 212.6
det: 125.5	det: 125.5
Flame: 212	Flame: 212
Filter: 125	Filter: 125
Base: 125	Base: 125
Sample psi: 07.51	Sample psi: 07.51
Internal Air Pressure: 20	Internal Air Pressure: 20
Internal Fuel Pressure: 13	Internal Fuel Pressure: 13


Cylinder/Regulator Pressures:	H2 Cylinder (psi): 1100	H2 Cylinder (psi): 1100
H2 cylinder reg.set (psi): 30	H2 cylinder reg.set (psi): 30	H2 cylinder reg.set (psi): 30
Zero Air Gen Pressure: 42	Zero Air Gen Pressure: 42	Zero Air Gen Pressure: 42
Span Cylinder (psi): 1900	Span Cylinder (psi): 1900	Span Cylinder (psi): 1900
Span Cylinder reg.set (psi): 24	Span Cylinder reg.set (psi): 24	Span Cylinder reg.set (psi): 24
Measured Flow: 0.881	Measured Flow: n/a	Measured Flow: n/a
Expected Value: 26.77	Expected Value: 26.77	Expected Value: 26.77

Comments:
 The analyzer sample inlet filter was changed.

The analyzer cooling fan filter(s) were cleaned. The manifold blower was found to be working normally.

Repeat calibration was completed to correct for a zero drift. The SPAN value did not change after calibration.

THC[ppm]



NITROGEN DIOXIDE



Thermo 42i NO-NO2-NOx Analyzer Calibration

Date: <u>October 12, 2018</u>	Barometer/B.P./units: <u>F.S. 05544 expires January 15, 2019</u>	<u>923</u>	millibars
Company/Airshed: <u>LICA</u>	Thermometer/Station Temp: <u>F.S. 170286131 expires April 19, 2019</u>	<u>22</u>	°C
Location/Station Name: <u>Maskwa</u>	Weather Conditions: <u>Cloudy/Overcast</u>		
Start/End Time 24 hr. (mst): <u>9:23 / 17:45</u>	Calibration Purpose: <u>routine monthly</u>		
G.P.T. to be used for Ozone?: <u>No</u>	Performed By/Reviewer: <u>Alex Yakupov</u>	<u>Rob Fisher</u>	
Calibration Method: <u>Gas Dilution & Gas Phase Titration</u>	Cal Gas Expiry Date: <u>October 24, 2020</u>		

Analyzer: Serial Number/Owner: <u>1180930028</u> <u>LICA</u> Last Calibration Date: <u>September 2, 2018</u> Range ppb: <u>1000</u>	Correction Factors: <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th>Previous C.F.:</th> <th>As Found C.F.:</th> <th>New C.F.:</th> </tr> </thead> <tbody> <tr> <td>NO =</td> <td>1.000</td> <td>0.970</td> <td>1.000</td> </tr> <tr> <td>NO₂ =</td> <td>1.000</td> <td>1.000</td> <td>1.000</td> </tr> <tr> <td>NOx =</td> <td>1.000</td> <td>0.966</td> <td>1.000</td> </tr> </tbody> </table>		Previous C.F.:	As Found C.F.:	New C.F.:	NO =	1.000	0.970	1.000	NO ₂ =	1.000	1.000	1.000	NOx =	1.000	0.966	1.000
	Previous C.F.:	As Found C.F.:	New C.F.:														
NO =	1.000	0.970	1.000														
NO ₂ =	1.000	1.000	1.000														
NOx =	1.000	0.966	1.000														

Calibration Standards: Low Flow Meter ID/Expiry Date: <u>Defender Low 152019 expires December 13, 2018</u> High Flow Meter ID/Expiry Date: <u>Defender High 148944 expires December 13, 2018</u> Calibrator ID/Expiry Date: <u>API id# 690 expires March 15, 2019</u> Cal Gas Cylinder I.D. #: <u>LL 104225</u> Cal Gas Conc. (ppm): <u>51.5</u> <u>51.6</u>	Standard Calibration Points for a Range of: <u>1000 ppb</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Point</th> <th>Target NO (ppb)</th> <th>Target NO₂ (ppb)</th> <th>Cc Ozone ?</th> </tr> </thead> <tbody> <tr> <td>High</td> <td>780</td> <td>500</td> <td>n/a</td> </tr> <tr> <td>Mid</td> <td>380</td> <td>275</td> <td>n/a</td> </tr> <tr> <td>Low</td> <td>190</td> <td>100</td> <td>n/a</td> </tr> <tr> <td>Extra Point #1</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> <tr> <td>Extra Point #2</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> </tbody> </table>	Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?	High	780	500	n/a	Mid	380	275	n/a	Low	190	100	n/a	Extra Point #1	n/a	n/a	n/a	Extra Point #2	n/a	n/a	n/a
Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?																						
High	780	500	n/a																						
Mid	380	275	n/a																						
Low	190	100	n/a																						
Extra Point #1	n/a	n/a	n/a																						
Extra Point #2	n/a	n/a	n/a																						

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated NO	Calculated NOx	Indicated NO	Indicated NOx	NO C.F.	NOx C.F.
Point	Diluent	Cal Gas	Total Flow	(ppb)	(ppb)	(ppb)	(ppb)		
as found zero	4775	0.0	4775	0	0	0.0	0.0	n/a	n/a
as found high	4883	76.0	4959	788.9	790.4	813.0	818.0	0.970	0.966
adjusted zero	4775	0.00	4775	0.0	0.0	0.0	0.0	n/a	n/a
adjusted high	4883	75.96	4959	788.9	790.4	789.0	790.0	1.000	1.000
mid	4925	36.58	4962	379.7	380.4	384.0	385.0	0.989	0.988
low	4933	18.38	4951	191.2	191.6	193.0	193.0	0.991	0.993
calibrator zero	4775	0.00	4775	0	0	0.0	0.0	n/a	n/a
Average C.F.=								0.993	0.994

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calibrator Setting	Indicated NO	Indicated NOx	Indicated NO ₂	NO drop	NO ₂ gain	NO ₂ C.F.
Point	Diluent	Cal Gas	Total Flow	volts or ppb	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)
NOx reference	4883	75.96	4959	0.0	793.0	794.0	1.0	0.0	1.0	
as found high NO2	4883	75.96	4959	510.0	282.0	794.0	512.0	511.0	511.0	1.000
adjusted high NO2	4883	75.96	4959	510.0	282.0	794.0	512.0	511.0	511.0	1.000
gpt mid	4883	75.96	4959	275.0	521.0	796.0	275.0	272.0	274.0	0.993
gpt low	4883	75.96	4959	100.0	696.0	796.0	100.0	97.0	99.0	0.980
Average NO ₂ C.F.=										0.991

Linear Regression/Calibration Results:

	NO	NOx	NO ₂	LIMITS
Correlation Coefficient =	1.000	1.000	1.000	> or = 0.995
Slope =	1.000	1.001	1.002	0.95-1.05
b (Intercept as % of full scale)=	0.16%	0.16%	0.18%	± 3% F.S.
% change in C.F. from last cal=	2.97%	3.38%	0.00%	± 10%
NO2 converter efficiency			1.00	0.96 to 1.04

As found:	As left:
NO Bkg: <u>2.7</u>	NO Bkg: <u>2.7</u>
NOx Bkg: <u>2.8</u>	NOx Bkg: <u>2.8</u>
NO Coef: <u>1.060</u>	NO Coef: <u>1.027</u>
NO2 Coef: <u>1.000</u>	NO2 Coef: <u>1.000</u>
NOx Coef: <u>1.003</u>	NOx Coef: <u>0.999</u>
PMT: <u>-866.5</u>	PMT: <u>-866.5</u>
Internal: <u>30.0</u>	Internal: <u>29.9</u>
Chamber: <u>50.1</u>	Chamber: <u>50.5</u>
Cooler: <u>-3.0</u>	Cooler: <u>-3.0</u>
NO2 Converter: <u>326.3</u>	NO2 Converter: <u>326.3</u>
NO2 Converter Set: <u>325.0</u>	NO2 Converter Set: <u>325.0</u>
Perm Oven Gas: <u>44.98</u>	Perm Oven Gas: <u>45.01</u>
Perm Oven Heater: <u>44.16</u>	Perm Oven Heater: <u>44.19</u>
Pressure: <u>252.5</u>	Pressure: <u>254.0</u>
Flow: <u>0.545</u>	Flow: <u>0.548</u>
Ozonator Flow: <u>OK</u>	Ozonator Flow: <u>OK</u>
Expected Value NO: <u>5</u>	Expected Value NO: <u>6</u>
Expected Value NO2: <u>400</u>	Expected Value NO2: <u>395</u>
Expected Value NOx: <u>405</u>	Expected Value NOx: <u>400</u>

Comments:

The analyzer sample inlet filter was changed.

The manifold blower was found to be working normally.

The converter cooling fan filter was cleaned.




No high point NO2 adjustment was required/made. As found values were copied to adjusted high values for linearity calculation purposes.

The analyzer cooling fan filter(s) were cleaned.

Date: October 12, 2018
Company/Airshed: LICA
Location/Station Name: Maskwa


Start/End Time 24 hr. (mst): 9:23 / 17:45
Calibration Purpose: routine monthly
Calibration Method: Gas Dilution & Gas Phase Titration

Thermo 42i NO-NO2-NOx Analyzer Calibration



Station: LICA MASKWA Daily: 18/10/12 Type: AVG 1 Min. [1 Min.]

— NOX[ppb] — NO[ppb] — NO2[ppb]



WIND SYSTEM



Meteorological Sensor Audit/Calibration

Location Information

Company:	LICA	Performed By:	Alex Yakupov
Audit Location:	Maskwa	Reviewed By:	Rob Fisher
Audit Date:	September 17, 2018	Start/End Time (mst):	9:36 / 12:48
Calibration Purpose:	installation	Weather Conditions:	Cloudy/Overcast

Wind Sensor Information

Sensor ID Data:		Sensor Outputs:	
Sensor Make:	RM Young	Velocity Voltage Output Range:	0-1 V
Sensor Model:	05305VK	Velocity Unit Output Range:	0-200 km/h
Serial #:	161465	Direction Voltage Output Range:	0-1 V
Previous Cal/Audit Date:	May 17, 2018	Direction Unit Output Range:	0-360 degrees

Wind Calibrator Information

Calibrator I.D. and Expiry Date: Model 18860-90/18802 SN: CA 4744; expiration May 18, 2019

Wind Speed Audit Data ****+/- 2% of the average correction factor is the limit****

RPM	Wind Speed Generated kph	Clockwise Wind Speed kph	Counter Clockwise Wind Speed kph	Correction Factor
0	0	0.1	0.1	-
1000	18.4	18.5	18.5	0.995
2000	36.9	36.9	36.9	1.000
3000	55.3	55.4	55.4	0.998
4000	73.7	73.8	73.8	0.999
5000	92.2	92.3	92.3	0.999
6000	110.6	110.8	110.8	0.998
7000	129.0	129.3	129.3	0.998
8000	147.4	147.7	147.7	0.998
9000	165.9	166.1	166.1	0.999
10000	184.3	184.9	184.9	0.997
The audit meets AMD requirements.			Average Correction Factor=	0.998

Wind Direction Audit Data ****+/- 3° of the absolute average degrees difference for all points is the limit****

Generated Wind Direction 0-360 (Up)	Generated Wind Direction 360-0 (Down)	Indicated Wind Direction 0-360 (Up)	Indicated Wind Direction 360-0 (Down)	Degrees Difference 0-360 (Up)	Degrees Difference 360-0 (Down)	Average Absolute Degrees Difference
0	355	0	355	0.3	0.1	0.2
30	330	30	329	-0.4	0.7	0.5
60	300	62	300	-1.9	-0.3	1.1
90	270	91	270	-1.3	-0.3	0.8
120	240	121	241	-1.0	-0.8	0.9
150	210	152	212	-1.7	-1.7	1.7
180	180	181	182	-1.1	-2.0	1.6
210	150	211	152	-1.1	-1.8	1.5
240	120	241	122	-0.5	-1.8	1.2
270	90	270	91	-0.1	-0.8	0.5
300	60	300	61	0.4	-0.6	0.5
330	30	330	31	-0.1	-0.7	0.4
355	0	354	0	0.6	0.3	0.5
The audit meets AMD requirements.			Average Absolute Degrees Difference=		0.9	

Comments:

METEOROLOGICAL SYSTEM CHECK



Meteorological System Checklist

Date:	October 12, 2018		
Technician:	Alex Yakupov		
Reviewer:	Rob Fisher		
Station:	Maskwa		
Unit:	Make:	Model:	Serial #:
Precipitation Sampler:	Met One - Heated Rain Gauge	Part 387	F4418
PRECIPITATION SENSOR CHECK			
Checklist:	Reply:	Comments:	
Previous check date:	June 21, 2018	n/a	
Is the sensor Level?	yes	n/a	
Is the heater operating properly?	yes	n/a	
Are the bucket drain holes clean?	yes	n/a	
Is the screen on the housing? (screen should be on between July and September)	no	Screens removed for Winter season	
Is the housing clean?	yes	n/a	
Is the area around the housing clean and free from obstacles?	yes	n/a	
TIP TEST - Slowly pour water until 10 tip are heard. (10 tips = 1 ml)			
# of Tips	Data Logger Response (ml):	Manual Specification = +/- 0.1 ml	
10	1.00	0.00	

CALIBRATORS

Company: <u>Maxxam</u>		Operator: <u>Chris W</u>	
Calibrator:		Flow Measurement Device:	
Make/Model	<u>API 700</u>	Make/Model	<u>Mesa Defender 530</u>
Serial Number	<u>690</u>	Serial Number	<u>L-153351 H-152571</u>
Last Verification Date	<u>March 2016</u>	Temperature (°C)	<u>23.5 C</u>
NO Cylinder S/N	<u>LL108015</u>	Barometric Pressure	<u>695 mmHg</u>
NO [PPM]	<u>52.2</u>	NOx [PPM]	<u>52.3</u>
Expiry Date	<u>Oct 2020</u>		

Dilution Flow (sccm)			
Pt. #1	<u>5000</u>	Pt. #2	<u>5000</u>
Pt. #3	<u>5000</u>		
Gas Flow (sccm)			
Pt. #1	<u>80</u>	Pt. #2	<u>40</u>
Pt. #3	<u>20</u>		

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
5000	0.0	0.000	0.000	0.000	0.000	0.000	Limit ± 10%	
4959	75.0	0.789	0.791	0.793	0.000	0.793	1%	0%
4971	36.5	0.383	0.384	0.384	0.000	0.384	0%	0%
4967	18.2	0.191	0.192	0.191	0.000	0.191	0%	-1%
Absolute Average Percent Difference							0%	0%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO		LIMITS		NOx	
Correlation=	1.0000	≥ 0.990		Correlation=	1.0000
m (Slope)=	1.0054	0.90-1.10		m (Slope)=	1.0031
b (Intercept % of FS)=	-0.0583	± 3% F.S.		b (Intercept % of FS)=	-0.0795

Flow	O ₂ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4959	0.000	0.000	0.790	-0.001	0.789	NO ₂	% Diff. Limit
4959	0.500	0.497	0.293	0.493	0.786	-1%	± 10%
4959	0.275	0.273	0.517	0.269	0.787	-1%	± 10%
4959	0.100	0.102	0.688	0.099	0.787	-2%	± 10%
Absolute Average Percent Difference						1%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO₂		LIMITS
Correlation=	1.0000	≥ 0.995
m (Slope)=	0.9946	0.90-1.10
b (Intercept % of FS)=	-0.1817	± 3% F.S.

<p align="center">AENV Standards</p> <p align="center">Audit Calibrator</p> <p>Make/Model <u>Teco 146i</u></p> <p>Serial/AMU Number <u>AMU 1809</u></p> <p>SRM Gas Cylinder No. <u>APEX1170572</u></p> <p>Cylinder Conc. (ppm) <u>49.99</u></p>	<p align="center">NO_x Analyzer</p> <p>Make/Model <u>Teco 42i</u></p> <p>Serial/AMU Number <u>AMU 1868</u></p> <p>Last Calibration Date <u>March 14, 2018</u></p> <p>Full Scale (ppm) <u>1.0</u></p> <p>Cylinder Gas Expiry Date <u>November 2020</u></p>
---	--

COMMENTS: Cylinder contains 47.9 ppm SO₂.

Auditor: Al Clark

Operator Signature: *Chris W*

Date: March 15, 2018

Location: McIntyre Center Edmonton

Company Maxxam Operator: Mike

Calibrator:		Flow Measurement Device:	
Make/Model	<u>Sabio</u>	Make/Model	<u>Bios Definer 220</u>
Serial Number	<u>11900613</u>	Serial Number	<u>H=128686; L=129069</u>
Last Verification Date	<u>March 16, 2018</u>	Temperature (°C)	<u>22.9 C</u>
NO Cylinder S/N	<u>LL104183</u>	Barometric Pressure	<u>698 mmHg</u>
NO [PPM]	<u>50.8</u>	NOx [PPM]	<u>50.9</u>
Expiry Date	<u>October 24, 2020</u>		

Dilution Flow (sccm)			
Pt. #1	<u>5059</u>	Pt. #2	<u>5073</u>
		Pt. #3	<u>5073</u>
Gas Flow (sccm)			
Pt. #1	<u>77.5</u>	Pt. #2	<u>38.2</u>
		Pt. #3	<u>19.1</u>

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
5124	0.0	0.0000	0.0000	0.0000	-0.0001	0.0000	Limit ± 10%	
5059	77.5	0.7782	0.7797	0.7763	0.0005	0.7767	0%	0%
5073	38.2	0.3825	0.3833	0.3794	0.0000	0.3795	-1%	-1%
5073	19.1	0.1913	0.1916	0.1904	0.0000	0.1904	0%	-1%
Absolute Average Percent Difference							1%	1%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

<u>NO</u>		<u>LIMITS</u>		<u>NOx</u>	
Correlation=	1.0000	≥ 0.990		Correlation=	1.0000
m (Slope)=	0.9975	0.90-1.10		m (Slope)=	0.9960
b (Intercept % of FS)=	-0.0616	± 3% F.S.		b (Intercept % of FS)=	-0.0661

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
5059	0.0	0.0000	0.7741	0.0000	0.7741	NO ₂	% Diff. Limit
5059	500.0	0.4918	0.2823	0.4916	0.7739	0%	± 10%
5059	275.0	0.2774	0.4967	0.2780	0.7747	0%	± 10%
5059	100.0	0.1031	0.6710	0.1032	0.7743	0%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

<u>NO₂</u>		<u>LIMITS</u>	
Correlation=	1.0000	≥ 0.995	
m (Slope)=	0.9998	0.90-1.10	
b (Intercept % of FS)=	0.0173	± 3% F.S.	

<u>AENV Standards</u>		<u>NO_x Analyzer</u>	
<u>Audit Calibrator</u>			
Make/Model	<u>Thermo 146i</u>	Make/Model	<u>Thermo 42i</u>
Serial/AMU Number	<u>1809</u>	Serial/AMU Number	<u>1868</u>
SRM Gas Cylinder No.	<u>APEX1170572</u>	Last Calibration Date	<u>August 16, 2018</u>
Cylinder Conc. (ppm)	<u>49.99</u>	Full Scale (ppm)	<u>1.0</u>
		Cylinder Gas Expiry Date	<u>November 15, 2020</u>

COMMENTS: _____

Auditor: Shea Beaton
Operator Signature: _____

Date: August 22, 2018
Location: McIntyre Center Edmonton

CALIBRATION GASES



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2017-482CGA

Company: Maxxam **Operator's Name:** Mike
Cylinder #: LL104225 **Concentration PPM:** 49.2 **Tolerance(%)** 2 **Certified By:** Praxair
Expiry Date: October 2020

Reference Calibrator and Gas:	Flow Measurement Device:
Make/Model: <u>R&R MFC 201</u>	Make/Model: <u>Mesa Definer 220</u>
Serial Number: <u>AMU 1690</u>	Serial Number: <u>H-133034 / L-132702</u>
Last Verification Date: <u>December 13, 2017</u>	Temp. °C: <u>23.4 C</u>
Gas Type: <u>SO2</u> Conc. <u>98.07</u>	B.P. <u>707 mmHg</u>
Cylinder Number: <u>CAL016625</u>	
Expiry Date: <u>January 2019</u>	


Reference Analyzer:
 Make/Model: Teco 43C Serial/AMU Number: 1623
 Instrument Settings: Zero: 10.0 Span: 1.006 Range: 1.0
 Last Calibration: Date: Dec12/17 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.000			
4989	79.5	0.764	0.01594	62.755	47.9
4995	39.6	0.380	0.00793	126.136	47.9
4992	19.6	0.188	0.00393	254.694	47.9
Average Cylinder Concentration:					47.9

Previous Stated Concentration PPM: 49.2
 Percent variance from Stated: 3

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 < =5% Outside Manufacturer Tolerance. Use manufacturers concentration _____
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark Date: December 13, 2017
 Operator Signature: *Al Clark* Location: McIntyre Center Edmonton



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2017-493CGA

Company: Maxxam Operator's Name: Mike
 Cylinder #: EY0001003 Concentration PPM: 9.55 Tolerance(%): 2 Certified By: Praxair
 Expiry Date: October 2020

Reference Calibrator and Gas:
 Make/Model: Sabio 2010
 Serial Number: AMU 2092
 Last Verification Date: January 17, 2018
 Gas Type: H2S Conc. 20.43
 Cylinder Number: CAL015272
 Expiry Date: January 2019

Flow Measurement Device:
 Make/Model: Mesa Defender 530
 Serial Number: H-153961 / L-153874
 Temp. °C: 23.0 C
 B.P.: 697 mmHg

Reference Analyzer:
 Make/Model: Teco 450i Serial/AMU Number: 1980
 Instrument Settings: Zero: 12.9 Span: 0.955 Range: 0.1
 Last Calibration: Date: Jan 17/18 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.0000	0.0000	0.0000	0.0000
5051	39.6	0.0753	0.00784	127.551	9.60
5028	20.2	0.0387	0.00402	248.911	9.63
5033	10.5	0.0198	0.00209	479.333	9.49
Average Cylinder Concentration:					9.58

Previous Stated Concentration PPM: 9.55

Percent variance from Stated: 0

Meets Manufacturer Tolerance. Use manufacturers stated concentration COMMENTS: Used AEP regulator
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark

Date: January 18, 2018

Operator Signature: [Signature]

Location: McIntyre Center Edmonton



Calibration Gas Audit

CH4 / C3H8 Cylinder Gas

File No. 2017-481CGA

Company: Maxxam **Operators name:** Mike

Cylinder #: LL119471 Conc CH4 (PPM) 599/207 Tolerance (%) 2 Certified By: Praxair

Expiry Date: October 2025

Reference Calibrator and Gas:				Flow Measurement Device:	
Make/Model	<u>R&R MFC 201</u>			Make/Model	<u>Mesa Definer 220</u>
Serial Number	<u>AMU 1690</u>			Serial Number	<u>H-133034 / L-132702</u>
Last Verification Date	<u>December 13, 2017</u>			Temp. °C	<u>23.1 C</u>
Gas Type	<u>CH4</u>	Conc.	<u>990.4</u>	B.P.	<u>707 mmHg</u>
Cylinder Number	<u>5604875</u>	Expiry Date	<u>July 2021</u>		
Gas Type	<u>C3H8</u>	Conc.	<u>246.5</u>		
Cylinder Number	<u>XF003845B</u>	Expiry Date	<u>July 2022</u>		

Reference Analyzer:

Make/Model Teco 55i Serial/AMU Number: 2108

Instrument Settings Zero: N/A Span: N/A Range: 20.0

Last Calibration: Date: Dec 12/17 C.F. 1.000 Done By: Al Clark

Calibrator Flows (scem)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	CH4	C3H8			CH4	C3H8
3500	0.0	0.00	0.00	0.02	45.00	603	209
3618	80.4	13.41	12.75	0.02	45.00	603	209
3547	39.8	6.73	6.47	0.01	89.12	600	210
3560	19.8	3.34	3.21	0.01	179.80	601	210
Average Cylinder Concentration:						601	209

	<u>CH4</u>		<u>C3H8</u>
Previous Stated Concentration PPM:	<u>599</u>		<u>207</u>
Percent variance from Stated:	<u>0</u>		<u>1</u>

Cylinder gas tolerances based on CH4 only

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:**

< =5% Outside Manufacturer Tolerance. Use manufacturers concentration

> 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark Date: December 13, 2017

Operator Signature: *Al Clark* Location: McIntyre Center Edmonton



Calibration Gas Audit

NO Cylinder Gas

File No. 2017-483CGA

Company: Maxxam **Operators name:** Mike

Cylinder #: LL104225 Conc (PPM) 51.5/51.6 Tolerance (%) 2 Certified By: Praxair

Expiry Date: October 2020

Reference Calibrator and Gas:				Flow Measurement Device:	
Make/Model	<u>Teco 146i</u>			Make/Model	<u>Mesa Definer 220</u>
Serial Number	<u>AMU 1809</u>			Serial Number	<u>H-133034 / L-132702</u>
Last Verification Date	<u>December 13, 2017</u>			Temp. °C	<u>23.4 C</u>
Gas Type	<u>NO</u>	Conc.	<u>50.03</u>	B.P.	<u>707 mmHg</u>
Cylinder Number	<u>APEX 1223938</u>				
Expiry Date	<u>June 2020</u>				

Reference Analyzer:

Make/Model Teco 42i Serial/AMU Number: 1868

Instrument Settings Zero: 4.7 Span: 1.004 Range: 1.0

Last Calibration: Date: Dec12/17 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	NO	NOX			NO	NOX
5000	0.0	0.000	0.000				
4989	79.5	0.813	0.812	0.016	62.755	51.0	51.0
4995	39.6	0.407	0.406	0.008	126.136	51.3	51.2
4992	19.6	0.202	0.201	0.004	254.694	51.4	51.2
Average Cylinder Concentration:						51.3	51.1

<u>NO</u>	<u>NOx</u>
Previous Stated Concentration PPM: <u>51.5</u>	<u>51.6</u>
Percent variance from Stated: <u>0</u>	<u>1</u>

Cylinder gas tolerances based on NO only

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:**

<=5% Outside Manufacturer Tolerance. Use manufacturers concentration

> 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark Date: December 13, 2017

Operator Signature: *Al Clark* Location: McIntyre Center Edmonton

APPENDIX ~~@@~~
U ° ~~Q~~yU ° ~~V~~ai° Vu° V-lyo) ° u°



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - October 2018

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	5	0	S	0	0	1	2	2	1	1	4	6	3	1	1	3	1	1	1	1	1	0	0	0	0	6	2	24
2	0	S	0	0	0	0	0	1	1	2	8	6	3	1	0	0	0	2	0	0	0	0	0	0	0	8	1	24
3	S	0	0	0	2	6	5	16	8	1	2	13	9	6	8	1	0	0	0	0	0	0	0	S	0	16	4	24
4	0	0	0	0	0	0	0	0	1	1	1	2	1	1	1	1	2	0	1	3	3	0	S	1	0	3	1	24
5	1	0	0	1	1	0	0	0	12	26	21	13	7	11	6	8	0	1	0	0	0	S	1	0	0	26	5	24
6	0	0	0	0	0	0	1	3	4	4	4	3	5	1	3	6	6	0	0	0	S	0	0	0	0	6	2	24
7	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	24
8	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	24
9	0	0	0	0	0	0	0	0	0	0	2	2	2	3	4	2	0	S	0	0	0	0	0	0	0	4	1	24
10	0	0	0	0	0	0	0	0	1	2	2	0	1	2	1	2	S	2	0	0	0	0	0	0	0	2	1	24
11	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	1	0	0	0	1	0	24
12	0	0	0	1	0	0	0	0	1	C	C	C	C	C	C	C	0	0	0	0	1	0	0	0	0	1	0	24
13	0	0	1	1	19	10	1	11	12	5	6	13	9	S	3	0	7	11	2	2	12	5	1	0	0	19	6	24
14	0	0	0	0	0	0	0	0	1	1	1	S	0	0	0	0	0	0	0	0	0	1	1	1	0	1	0	24
15	0	0	0	0	0	1	0	0	4	17	16	S	11	22	19	16	15	14	13	8	5	4	20	0	0	22	8	24
16	13	12	2	7	1	1	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	13	2	24
17	0	0	0	0	1	1	2	3	2	S	2	2	2	0	0	0	3	2	2	5	5	3	3	3	0	5	2	24
18	3	1	2	2	1	1	2	2	S	20	6	1	1	0	0	7	17	21	7	0	1	4	35	22	0	35	7	24
19	19	2	0	0	0	0	3	S	28	21	1	0	0	0	13	11	1	0	0	0	1	3	2	0	0	28	5	24
20	0	0	0	0	0	0	S	2	1	0	0	1	1	1	0	0	0	0	0	0	0	0	0	0	0	2	0	24
21	0	2	0	0	0	S	0	0	0	1	2	0	0	0	2	15	11	10	8	3	3	2	2	1	0	15	3	24
22	0	0	0	0	S	0	0	0	0	1	0	0	0	0	0	0	0	1	3	4	1	2	2	2	0	4	1	24
23	1	1	0	S	0	0	1	1	1	1	1	1	2	2	1	1	0	0	0	1	2	0	0	0	0	2	1	24
24	0	0	S	0	0	0	0	0	0	0	2	3	1	2	2	0	0	0	0	0	0	0	0	0	0	3	1	24
25	0	S	1	0	0	0	0	0	0	0	2	20	3	1	1	1	1	0	0	0	0	0	0	0	0	20	1	24
26	S	0	0	0	0	5	2	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	5	0	24
27	0	0	0	1	0	0	0	0	0	2	1	0	0	0	0	16	19	7	1	1	0	1	0	S	1	19	2	24
28	0	0	1	0	0	0	0	0	0	0	0	1	1	7	5	1	3	5	4	1	1	S	1	1	0	7	1	24
29	1	1	0	0	0	0	0	0	0	0	0	12	13	9	22	5	3	11	17	15	S	1	1	1	0	22	5	24
30	1	1	1	2	1	2	2	2	1	1	2	2	2	2	1	1	0	0	0	S	0	0	0	0	0	2	1	24
31	0	0	0	0	0	0	1	1	1	1	1	0	1	1	3	4	2	0	S	0	0	1	0	0	0	4	1	24
HOURLY MAX	19	12	2	7	19	10	5	16	28	26	21	20	13	22	22	19	17	21	17	15	12	5	35	22				
HOURLY AVG	2	1	0	1	1	1	1	2	3	4	3	4	3	3	4	4	3	3	2	1	1	1	2	1				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	307
MAXIMUM INSTANTANEOUS VALUE:	35 ppb @ HOUR 22 ON DAY 18
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	7 hrs
STANDARD DEVIATION:	4
OPERATIONAL TIME:	744 hrs

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - October 2018

HYDROGEN SULPHIDE Instantaneous Maximum (H₂S ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.		
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.			
DAY 1	2	2	S	1	2	2	2	2	2	2	2	2	2	2	2	1	2	1	1	2	1	2	2	2	1	2	2	24		
2	2	S	2	2	1	2	1	2	1	2	2	2	1	2	2	2	1	2	1	2	2	1	2	2	1	2	2	24		
3	S	1	1	2	2	2	2	2	1	1	2	2	1	2	1	2	1	1	1	1	1	1	1	S	1	2	2	24		
4	2	1	1	1	1	1	3	3	2	2	2	3	2	1	2	2	3	2	2	1	2	2	S	2	1	3	2	24		
5	2	2	1	2	2	2	2	2	2	2	2	1	1	2	1	2	1	2	1	2	2	S	2	2	1	2	2	24		
6	1	1	2	2	2	4	2	1	3	2	2	2	2	2	2	3	3	2	2	1	S	2	2	2	1	4	2	24		
7	2	2	1	2	1	2	2	2	1	1	1	2	1	1	2	1	2	2	2	S	2	2	2	1	2	1	2	24		
8	2	1	2	2	2	2	2	2	1	1	1	1	2	2	2	2	2	2	S	1	2	2	1	2	1	2	2	24		
9	2	2	1	2	2	2	2	1	2	1	2	2	2	2	2	2	S	2	2	2	2	2	2	3	1	3	2	24		
10	3	2	3	4	4	2	2	2	4	4	1	2	2	2	2	2	S	2	2	2	2	2	2	2	2	1	4	2	24	
11	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	24	
12	2	2	2	2	2	2	2	2	2	2	2	2	2	2	C	C	C	C	C	C	C	C	C	0	0	0	2	2	24	
13	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	1	0	0	0	0	0	0	0	0	1	0	24	
14	0	0	0	0	0	0	0	1	0	0	0	0	S	0	0	0	0	1	0	0	0	0	1	0	0	1	0	1	0	24
15	0	0	0	0	0	0	0	0	1	1	1	S	1	2	0	1	1	1	1	1	1	1	1	1	1	0	2	1	24	
16	1	1	0	0	1	1	1	1	1	0	S	0	1	0	0	1	0	0	1	1	0	1	1	1	1	0	1	1	24	
17	1	1	1	1	0	0	1	1	1	S	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
18	0	0	0	0	0	1	0	0	S	1	1	0	0	0	0	1	1	2	1	1	0	1	2	2	2	0	2	1	24	
19	1	1	1	0	1	1	1	S	2	1	0	1	1	1	1	1	1	1	0	0	0	1	1	0	0	2	1	24		
20	1	1	1	0	0	1	S	1	1	1	1	1	1	1	1	1	1	1	1	0	0	1	1	1	0	1	1	24		
21	1	0	1	1	1	S	1	1	1	1	1	0	0	1	1	1	1	1	1	1	1	1	1	1	0	1	1	24		
22	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0	1	1	1	1	0	1	1	24		
23	1	1	1	S	2	1	1	1	1	1	1	1	1	1	1	2	1	1	1	1	1	1	2	2	1	1	2	1	24	
24	1	1	S	3	2	1	2	1	1	1	1	1	1	1	1	1	1	5	5	2	2	2	1	1	1	5	2	24		
25	1	S	2	1	1	1	1	1	1	1	1	1	1	1	2	1	1	1	1	1	1	1	2	1	1	2	1	24		
26	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	S	1	2	24	
27	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	2	1	1	1	1	1	1	S	1	1	2	1	24	
28	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	1	2	1	2	2	2	S	3	3	1	3	1	24	
29	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	3	2	2	1	2	1	S	1	1	1	1	3	1	24	
30	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	S	1	1	1	1	1	1	1	24	
31	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	S	1	1	1	1	1	1	1	1	1	24	
HOURLY MAX	3	2	3	4	4	4	3	3	4	4	2	3	2	2	3	3	3	5	5	2	2	2	3	3						
HOURLY AVG	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1		


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	602
MAXIMUM INSTANTANEOUS VALUE:	5 ppb @ HOUR 17 ON DAY 24
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	8 hrs
OPERATIONAL TIME:	744 hrs
STANDARD DEVIATION:	1

HYDROGEN SULPHIDE Instantaneous Maximum (H₂S ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - October 2018

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	2.04	2.03	S	2.02	2.03	2.06	2.08	2.09	2.18	2.19	2.14	2.07	2.01	2.00	2.00	1.99	1.98	1.99	2.01	1.99	2.01	2.01	2.02	2.04	1.98	2.19	2.04	24
2	2.03	S	2.03	2.02	1.97	1.97	1.97	1.97	1.96	1.96	1.96	1.96	1.95	1.97	1.99	1.98	1.95	1.97	1.96	1.97	1.97	1.98	1.99	1.99	1.95	2.03	1.98	24
3	S	2.00	2.00	2.00	2.00	1.99	1.98	2.28	1.96	C	C	Y	Y	Y	C	C	C	C	2.08	2.03	2.06	2.00	2.06	S	1.96	2.28	2.03	21
4	1.98	1.97	1.94	1.92	1.91	1.94	2.00	1.97	1.99	1.85	1.86	1.83	1.78	1.79	1.79	1.77	1.75	1.77	1.90	1.93	1.92	1.89	S	1.94	1.75	2.00	1.89	24
5	1.98	2.03	2.07	2.04	2.01	1.99	1.99	1.97	1.99	1.92	1.87	1.77	1.81	1.75	1.71	1.75	2.00	2.07	2.06	1.80	1.79	S	1.80	1.83	1.71	2.07	1.91	24
6	1.81	1.89	1.99	1.83	1.87	1.83	1.83	1.86	1.81	1.81	1.83	1.78	1.78	1.77	1.89	1.85	1.79	1.81	1.83	1.84	S	1.91	1.92	1.85	1.77	1.99	1.84	24
7	1.84	1.85	1.86	1.87	1.88	1.87	1.88	1.90	2.06	1.88	1.89	1.99	1.90	1.91	1.90	1.91	1.91	1.91	1.93	S	1.95	1.95	1.94	1.95	1.84	2.06	1.91	24
8	1.94	1.96	1.98	1.98	1.99	1.99	2.00	2.02	1.99	1.99	1.97	1.98	1.96	1.96	1.95	1.94	1.98	1.96	S	2.04	2.05	6.78	6.56	5.89	1.94	6.78	2.56	24
9	2.98	2.18	2.14	2.19	2.14	2.16	2.14	2.18	2.20	2.05	2.06	2.09	2.01	2.02	2.02	2.05	2.01	S	2.04	2.17	2.07	2.05	2.06	2.31	2.01	2.98	2.14	24
10	2.23	2.16	2.12	2.11	2.10	2.15	2.17	2.22	2.24	2.18	2.05	2.03	1.95	1.93	1.93	1.94	S	1.92	1.97	1.99	2.05	2.11	2.07	2.06	1.92	2.24	2.07	24
11	2.09	2.12	2.12	2.12	2.14	2.14	2.06	1.95	1.98	2.02	2.19	2.22	2.11	2.02	2.13	S	1.87	1.82	1.83	1.83	1.87	1.86	1.85	1.85	1.82	2.22	2.01	24
12	1.83	1.81	1.79	1.75	1.79	1.81	1.82	1.78	1.81	C1	C1	C1	C1	C1	C1	2.21	2.17	2.19	2.25	2.28	2.30	2.30	2.31	2.31	1.75	2.31	2.03	18
13	2.34	2.34	2.36	2.38	2.44	2.39	2.38	2.43	2.42	2.59	2.43	2.42	2.42	S	2.41	2.43	2.55	2.43	2.43	2.52	2.51	2.49	2.75	2.54	2.34	2.75	2.45	24
14	2.52	2.51	2.54	2.56	2.54	2.84	3.04	2.94	2.69	2.58	2.53	2.39	S	2.32	2.28	2.28	2.26	2.28	2.29	2.29	2.29	2.31	2.35	2.36	2.26	3.04	2.48	24
15	2.29	2.20	2.19	2.18	2.19	2.21	2.19	2.18	2.35	2.37	2.31	S	2.21	2.27	2.23	2.24	2.29	2.30	2.41	2.41	2.47	2.32	2.41	2.31	2.18	2.47	2.28	24
16	2.41	2.36	2.43	2.40	2.42	2.54	2.54	2.96	2.65	2.52	S	2.53	2.61	2.45	2.44	2.41	2.43	2.38	2.38	2.37	2.79	2.57	2.58	2.36	2.96	2.50	2.4	24
17	2.59	3.02	2.65	2.57	2.63	2.62	2.55	2.53	2.56	S	2.67	2.67	2.35	2.19	2.25	2.25	2.20	2.13	2.18	2.22	2.32	2.33	2.31	2.35	2.13	3.02	2.44	24
18	2.38	2.33	2.38	2.56	2.55	2.52	2.23	2.20	S	2.27	2.24	2.12	2.11	2.17	2.16	2.23	2.39	2.60	2.42	2.25	2.30	2.42	2.62	2.61	2.11	2.62	2.35	24
19	2.53	2.33	2.34	2.35	2.37	2.41	2.48	S	2.56	2.41	2.41	2.41	2.44	2.41	2.92	2.92	2.43	2.45	2.45	2.48	2.58	2.58	2.51	2.51	2.33	2.92	2.49	24
20	2.44	2.40	2.41	2.41	2.39	2.40	S	2.35	2.40	2.41	2.45	2.46	2.44	2.35	2.15	2.08	2.11	2.13	2.16	2.20	2.22	2.25	2.26	2.23	2.08	2.46	2.31	24
21	2.23	2.26	2.27	2.27	2.28	S	2.30	2.32	2.31	2.38	2.34	2.27	2.23	2.21	2.31	2.33	2.43	2.45	2.48	2.36	2.49	2.52	2.42	2.44	2.21	2.52	2.34	24
22	2.46	2.57	2.69	2.91	S	2.61	2.59	2.58	2.71	2.66	2.53	2.69	2.67	2.53	2.44	2.42	2.58	2.63	2.69	2.66	2.55	2.53	2.57	2.57	2.42	2.91	2.60	24
23	2.57	2.56	2.60	S	2.62	2.62	2.45	2.47	2.47	2.47	2.65	2.43	2.38	2.30	2.26	2.24	2.29	2.32	2.35	2.32	2.34	2.52	2.39	2.42	2.24	2.65	2.44	24
24	2.41	2.35	S	2.38	2.37	2.36	2.44	2.44	2.37	2.41	2.32	2.29	2.21	2.25	2.20	2.15	2.29	2.32	2.30	2.36	2.35	2.36	2.34	2.38	2.15	2.44	2.33	24
25	2.34	S	2.48	2.41	2.39	2.37	2.38	2.46	2.47	2.48	2.30	2.39	2.26	2.21	2.23	2.25	2.27	2.29	2.31	2.71	2.46	2.50	2.40	2.39	2.21	2.71	2.38	24
26	S	2.37	2.41	2.46	2.48	2.50	2.48	2.36	2.27	2.24	2.18	2.17	2.10	2.06	2.08	2.08	2.12	2.23	2.25	2.27	2.28	2.74	2.56	S	2.06	2.74	2.30	24
27	2.57	2.74	2.68	2.45	2.37	2.33	2.51	2.50	2.42	2.25	2.22	2.20	2.20	2.18	2.32	2.35	2.31	2.26	2.27	2.24	2.24	2.26	S	2.24	2.18	2.74	2.35	24
28	2.26	2.28	2.30	2.31	2.44	2.61	2.66	2.48	2.51	2.46	2.30	2.36	2.38	2.29	2.33	2.24	2.34	2.32	2.30	2.24	2.20	S	2.18	2.21	2.18	2.66	2.35	24
29	2.19	2.16	2.12	2.13	2.11	2.12	2.13	2.14	2.16	2.18	2.18	2.26	2.28	2.37	2.49	2.38	2.41	2.41	2.40	2.36	S	2.30	2.31	2.35	2.11	2.49	2.26	24
30	2.33	2.33	2.33	2.30	2.31	2.32	2.32	2.30	2.28	2.29	2.31	4.71	2.34	2.37	2.38	2.38	2.41	2.39	2.37	S	2.35	2.37	2.55	2.60	2.28	4.71	2.46	24
31	2.58	2.56	2.60	2.54	2.48	2.45	2.43	2.44	2.44	2.49	2.50	2.37	2.38	2.34	2.34	2.38	2.36	2.34	S	2.34	2.41	2.50	2.46	2.37	2.34	2.60	2.44	24
HOURLY MAX	2.98	3.02	2.69	2.91	2.63	2.84	3.04	2.96	2.71	2.66	2.67	4.71	2.67	2.53	2.92	2.92	2.58	2.63	2.69	2.71	2.58	6.78	6.56	5.89				
HOURLY AVG	2.28	2.26	2.27	2.25	2.24	2.27	2.27	2.28	2.27	2.26	2.24	2.32	2.19	2.16	2.19	2.19	2.20	2.21	2.22	2.22	2.23	2.45	2.43	2.40				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	697
MAXIMUM INSTANTANEOUS VALUE:	6.78 ppm @ HOUR 21 ON DAY 8
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	6 hrs
OPERATIONAL TIME:	735 hrs
STANDARD DEVIATION:	0.38

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - October 2018

METHANE MAX Instantaneous Maximum (CH₄ ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	2.04	2.03	S	2.02	2.03	2.06	2.08	2.09	2.18	2.19	2.14	2.07	2.01	2.00	2.00	1.99	1.98	1.99	2.01	1.99	2.01	2.01	2.02	2.04	1.98	2.19	2.04	24
2	2.03	S	2.03	2.02	1.97	1.97	1.97	1.97	1.96	1.96	1.96	1.96	1.95	1.97	1.99	1.98	1.95	1.97	1.96	1.97	1.97	1.98	1.99	1.99	1.95	2.03	1.98	24
3	S	2.00	2.00	2.00	2.00	1.99	1.98	2.28	1.96	C	C	-	-	-	-	-	-	-	-	-	-	-	-	-	1.96	2.28	2.03	11
4	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
5	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
6	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
7	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
8	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
9	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
10	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
11	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
12	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
13	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
14	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
15	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
16	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
17	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
18	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
19	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
20	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
21	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
22	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
23	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
24	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
25	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
26	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
27	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
28	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
29	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
30	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
31	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
HOURLY MAX	2.04	2.03	2.03	2.02	2.03	2.06	2.08	2.28	2.18	2.19	2.14	2.07	2.01	2.00	2.00	1.99	1.98	1.99	2.01	1.99	2.01	2.01	2.02	2.04				
HOURLY AVG	2.04	2.02	2.02	2.01	2.00	2.01	2.01	2.11	2.03	2.08	2.05	2.02	1.98	1.99	2.00	1.99	1.97	1.98	1.99	1.98	1.99	2.00	2.01	2.02				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	54
MAXIMUM INSTANTANEOUS VALUE:	2.28 ppm @ HOUR 7 ON DAY 3
IZS CALIBRATION TIME:	3 hrs
MONTHLY CALIBRATION TIME:	2 hrs
STANDARD DEVIATION:	0.06
OPERATIONAL TIME:	59 hrs

METHANE MAX Instantaneous Maximum (CH₄ ppm)





NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
2	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
3	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	C	C	-	-	-	-	-	-	-	-	-	-	-	-	-	0.00	0.00	0.00	11
4	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
5	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
6	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
7	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
8	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
9	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
10	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
11	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
12	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
13	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
14	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
15	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
16	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
17	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
18	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
19	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
20	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
21	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
22	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
23	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
24	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
25	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
26	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
27	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
28	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
29	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
30	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
31	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	0
HOURLY MAX	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00				
HOURLY AVG	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00				

STATUS FLAG CODES


C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	0
MAXIMUM INSTANTANEOUS VALUE:	0.00 ppm @ HOUR 0 ON DAY 1
IZS CALIBRATION TIME:	0 hrs
MONTHLY CALIBRATION TIME:	2 hrs
OPERATIONAL TIME:	59 hrs
STANDARD DEVIATION:	0.00



NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - October 2018

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	10	1	S	3	2	4	6	9	8	4	10	14	9	7	7	11	3	1	2	2	2	6	2	8	1	14	6	24	
2	3	S	7	6	6	5	9	11	5	5	18	13	5	3	2	24	8	13	1	6	28	2	1	3	1	28	8	24	
3	S	5	3	15	26	38	26	36	8	3	2	20	12	13	15	12	4	5	4	1	16	1	1	S	1	38	12	24	
4	3	3	2	2	1	38	64	22	16	4	7	8	6	3	3	2	8	2	3	11	10	3	S	8	1	64	10	24	
5	7	14	5	4	5	14	19	18	34	46	26	23	16	19	12	14	2	3	5	6	2	S	6	3	2	46	13	24	
6	2	18	19	1	3	2	8	17	23	20	17	10	16	4	11	18	20	4	5	6	S	7	8	1	1	23	11	24	
7	5	3	2	2	1	20	13	14	13	1	7	11	3	1	2	1	1	1	1	S	2	1	1	2	1	20	5	24	
8	2	2	2	3	3	11	6	3	11	8	9	6	4	5	1	4	6	S	13	2	16	4	3	1	1	16	6	24	
9	1	1	1	1	1	2	2	4	42	3	12	8	8	12	13	7	1	S	2	1	1	3	4	3	1	42	6	24	
10	2	1	1	1	1	1	2	59	85	20	9	3	3	4	3	5	S	7	3	2	10	8	7	10	1	85	11	24	
11	4	4	4	10	4	6	8	28	5	4	11	8	9	6	7	S	6	4	5	4	5	5	12	5	4	28	7	24	
12	5	7	8	7	8	18	28	15	17	C	C	C	C	C	C	C	C	C	C	6	6	12	5	6	3	3	28	10	24
13	3	3	3	5	39	15	6	27	27	21	24	29	23	S	10	4	26	32	5	8	20	11	9	8	3	39	16	24	
14	3	5	8	7	13	15	44	17	23	11	8	13	S	4	3	2	2	4	4	4	5	10	11	10	2	44	10	24	
15	5	1	1	1	5	20	6	7	22	35	35	S	24	44	42	28	33	22	48	18	20	21	36	2	1	48	21	24	
16	27	22	9	13	7	11	28	50	70	13	S	12	12	21	7	7	5	4	4	3	3	42	8	4	3	70	16	24	
17	5	101	8	7	15	39	37	49	49	S	18	32	9	2	2	6	9	18	9	16	18	12	10	10	2	101	21	24	
18	10	8	8	12	21	20	19	31	S	39	16	12	11	1	27	20	40	48	21	2	19	31	66	45	1	66	23	24	
19	32	10	3	3	3	14	41	S	58	40	10	13	5	7	30	22	16	1	1	2	6	13	8	1	1	58	15	24	
20	1	1	1	1	1	1	S	7	5	27	18	7	18	6	33	16	8	4	1	1	8	2	7	1	1	33	8	24	
21	1	5	3	1	4	S	21	53	42	5	5	3	2	1	7	37	24	33	27	14	13	18	10	10	1	53	15	24	
22	6	5	5	3	S	9	14	7	7	7	3	8	7	4	2	5	2	6	11	11	4	5	6	6	2	14	6	24	
23	5	4	3	S	6	5	5	5	6	6	5	5	4	4	3	2	4	2	7	8	4	2	2	2	2	8	4	24	
24	2	2	S	3	4	2	73	14	107	44	61	33	26	30	19	49	9	2	28	2	29	6	46	64	2	107	29	24	
25	8	S	192	33	42	13	54	98	79	15	22	50	9	12	6	33	7	7	30	15	20	16	7	43	6	192	35	24	
26	S	14	14	25	24	42	43	44	64	11	2	5	27	7	2	2	3	6	2	4	5	4	4	S	2	64	16	24	
27	18	25	19	12	9	8	12	60	36	15	12	8	1	3	39	37	19	4	7	7	14	12	S	12	1	60	17	24	
28	3	11	11	10	2	7	13	9	46	14	3	4	6	20	13	5	11	15	13	3	3	S	5	5	2	46	10	24	
29	5	5	5	6	13	9	13	13	12	24	7	33	35	30	52	19	17	24	39	34	S	7	10	10	5	52	18	24	
30	14	11	12	14	13	17	14	25	7	3	7	14	18	10	6	5	6	4	2	S	14	5	8	8	2	25	10	24	
31	7	7	7	7	6	6	22	150	11	16	15	62	27	8	23	13	15	14	S	45	16	11	11	7	6	150	22	24	
HOURLY MAX	32	101	192	33	42	42	73	150	107	46	61	62	35	44	52	49	40	48	48	45	29	42	66	64					
HOURLY AVG	7	10	13	7	10	14	22	30	31	16	14	16	12	10	14	14	11	10	10	9	11	10	11	10					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	703
MAXIMUM INSTANTANEOUS VALUE:	192 ppb @ HOUR 2 ON DAY 25
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	9 hrs
OPERATIONAL TIME:	744 hrs
STANDARD DEVIATION:	17

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - October 2018

NITRIC OXIDE Instantaneous Maximum (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	0	0	S	0	0	0	0	1	2	1	4	6	3	2	4	5	0	0	0	0	0	0	0	1	0	6	1	24				
2	0	S	0	1	1	0	2	3	2	2	8	6	2	1	1	14	4	2	0	2	14	0	0	0	0	14	3	24				
3	S	0	0	7	3	8	3	27	2	1	1	8	8	5	6	7	2	1	0	0	1	0	0	S	0	27	4	24				
4	0	0	0	0	0	14	34	12	5	1	2	3	2	1	1	0	1	0	0	0	0	0	S	0	34	3	24					
5	0	5	0	0	0	5	3	4	17	26	12	11	7	7	4	4	0	0	0	1	0	S	0	0	26	5	24					
6	0	5	4	0	0	0	1	4	7	6	7	4	7	1	3	4	5	1	0	1	S	1	1	0	7	3	24					
7	1	1	0	0	0	9	6	6	6	0	3	4	1	0	2	0	0	0	0	S	0	0	0	0	9	2	24					
8	1	0	0	1	1	4	2	1	5	3	3	4	2	2	2	0	1	2	S	0	0	1	0	0	5	2	24					
9	0	0	0	0	0	0	0	1	24	1	5	3	3	5	5	2	0	S	0	0	0	0	0	0	24	2	24					
10	0	0	0	0	0	0	0	43	66	11	2	1	1	1	1	0	S	0	0	0	1	0	0	0	66	6	24					
11	0	0	0	0	0	1	1	10	1	1	3	4	3	1	2	S	0	0	0	0	0	0	4	0	10	1	24					
12	0	1	0	1	0	6	15	5	6	C	C	C	C	C	C	C	C	C	1	2	2	1	1	0	15	3	24					
13	0	0	0	1	17	5	3	10	11	10	12	16	12	S	3	1	8	11	1	1	2	0	0	0	17	5	24					
14	0	0	0	0	1	1	19	4	10	4	4	7	S	0	1	0	0	0	0	0	0	0	0	0	19	2	24					
15	0	0	0	0	0	0	1	4	8	17	16	S	9	18	21	10	15	3	29	2	2	2	12	0	29	7	24					
16	6	3	1	2	0	0	16	36	52	4	S	4	4	8	2	1	0	0	0	0	0	25	2	0	52	7	24					
17	0	82	3	1	1	27	12	25	25	S	7	22	2	0	0	1	0	4	0	0	0	0	0	0	82	9	24					
18	0	0	0	0	3	13	1	5	S	13	7	3	3	0	29	8	16	16	3	0	0	2	30	15	30	7	24					
19	8	0	0	0	0	6	15	S	31	19	4	5	1	2	14	8	3	0	0	0	0	0	0	0	31	5	24					
20	0	0	0	0	0	0	S	0	0	12	8	3	9	2	32	11	3	0	0	0	3	0	3	0	32	4	24					
21	0	0	0	0	0	S	4	22	38	1	1	1	0	0	2	14	8	11	1	0	0	1	0	0	38	5	24					
22	0	0	0	0	S	0	5	1	3	3	1	5	3	1	0	1	0	0	0	0	0	0	0	0	5	1	24					
23	0	0	0	S	0	0	0	0	1	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0	1	0	24					
24	0	0	S	0	0	0	67	5	81	30	49	16	16	14	8	20	1	0	13	0	13	1	21	31	81	17	24					
25	2	S	149	10	17	4	30	70	48	6	10	29	3	5	1	16	0	0	4	4	13	6	1	24	149	20	24					
26	S	2	2	12	12	22	21	21	38	3	0	1	19	5	0	0	0	0	0	0	0	0	0	S	38	7	24					
27	7	5	3	2	1	0	1	39	15	5	3	2	0	0	19	14	5	0	0	0	1	0	S	0	39	5	24					
28	1	0	0	0	0	1	4	4	35	6	1	1	1	6	3	0	1	1	0	0	0	S	0	0	35	3	24					
29	1	1	1	2	5	4	5	7	5	13	3	15	17	13	31	7	5	10	21	16	S	2	3	3	31	8	24					
30	5	4	3	3	3	3	0	7	2	0	2	11	7	4	1	1	0	0	0	S	4	0	1	0	11	3	24					
31	0	0	0	1	0	0	13	134	3	7	9	34	16	2	13	2	1	4	S	25	2	0	1	0	134	12	24					
HOURLY MAX	8	82	149	12	17	27	67	134	81	30	49	34	19	18	32	20	16	16	29	25	14	25	30	31								
HOURLY AVG	1	4	6	1	2	4	9	17	18	7	6	8	6	4	7	5	3	2	3	2	2	1	3	3								


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	417
MAXIMUM INSTANTANEOUS VALUE:	149 ppb @ HOUR 2 ON DAY 25
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	9 hrs
STANDARD DEVIATION:	12
OPERATIONAL TIME:	744 hrs

NITRIC OXIDE Instantaneous Maximum (NO ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - October 2018

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	10	1	S	3	2	4	6	8	6	3	6	8	6	5	4	6	2	1	2	2	2	6	2	8	1	10	5	24	
2	3	S	7	5	5	5	8	8	4	3	9	7	3	2	2	12	3	11	1	4	14	2	2	3	1	14	5	24	
3	S	5	3	10	23	31	24	18	6	2	1	11	9	8	9	6	2	4	4	1	14	1	1	S	1	31	9	24	
4	3	3	2	2	1	25	31	13	11	3	5	6	4	2	2	2	7	2	3	11	10	3	S	8	1	31	7	24	
5	7	9	5	4	5	11	16	14	17	20	14	12	9	12	8	10	2	3	5	6	2	S	6	3	2	20	9	24	
6	2	14	15	1	2	2	7	13	16	14	11	7	10	3	9	14	16	3	4	6	S	6	8	1	1	16	8	24	
7	4	2	2	1	1	11	7	8	7	1	4	7	2	1	1	1	1	1	0	S	2	1	1	2	0	11	3	24	
8	2	1	2	2	2	8	4	2	7	5	5	5	4	3	4	1	4	4	S	13	2	15	4	3	1	15	4	24	
9	1	1	1	1	1	2	2	3	18	2	7	5	5	7	8	5	1	S	2	1	1	3	4	3	1	18	4	24	
10	2	1	1	1	1	1	2	17	23	10	7	3	3	3	3	5	S	7	3	2	9	8	7	9	1	23	5	24	
11	4	4	4	10	4	5	7	18	5	4	8	5	6	4	5	S	6	4	5	4	5	5	9	5	4	18	6	24	
12	5	6	7	7	7	12	15	10	12	C	C	C	C	C	C	C	C	C	C	4	5	10	4	5	3	3	15	7	24
13	3	3	3	4	23	11	4	17	16	11	12	15	12	S	7	3	17	21	4	8	18	11	8	8	3	23	10	24	
14	3	5	7	7	12	14	25	15	13	7	5	6	S	4	3	2	2	4	4	4	5	9	10	9	2	25	8	24	
15	5	1	1	1	5	20	4	3	14	18	20	S	14	26	22	20	21	19	30	16	19	20	24	2	1	30	14	24	
16	20	19	8	10	7	11	12	14	22	8	S	8	9	13	5	5	5	4	4	3	3	17	6	4	3	22	9	24	
17	5	21	6	7	14	17	26	28	26	S	11	11	7	2	2	5	9	13	9	16	18	12	10	10	2	28	12	24	
18	10	8	8	12	18	13	18	26	S	26	12	8	8	1	4	13	27	32	21	2	19	29	38	31	1	38	17	24	
19	25	10	3	3	3	9	27	S	27	22	6	8	4	5	16	14	13	1	1	2	6	13	8	1	1	27	10	24	
20	1	1	1	1	1	2	S	7	5	14	11	5	9	4	12	7	6	4	1	1	5	2	5	1	1	14	5	24	
21	1	5	3	1	4	S	18	32	8	4	4	2	1	1	5	23	17	22	26	14	13	17	10	10	1	32	11	24	
22	6	5	5	3	S	9	9	7	4	5	3	5	5	3	2	4	2	6	11	11	4	5	6	6	2	11	5	24	
23	5	4	3	S	6	5	5	5	5	4	4	3	4	3	3	3	2	4	2	7	8	3	3	2	2	8	4	24	
24	2	2	S	3	3	2	11	10	31	16	20	18	14	16	12	29	8	2	15	2	16	5	24	33	2	33	13	24	
25	6	S	43	23	26	9	26	31	31	9	12	21	7	9	5	17	6	7	26	11	16	12	6	20	5	43	16	24	
26	S	12	12	16	13	21	24	25	26	7	2	4	10	4	2	2	3	6	2	4	5	4	4	S	2	26	9	24	
27	12	20	16	10	8	8	12	21	21	10	8	6	1	2	21	23	14	4	7	7	14	12	S	12	1	23	12	24	
28	3	11	11	10	2	5	10	5	14	8	2	3	5	14	11	4	10	14	13	3	3	S	5	5	2	14	8	24	
29	5	5	4	6	8	5	8	6	8	11	4	17	18	17	22	12	12	14	19	18	S	5	8	7	4	22	10	24	
30	9	8	9	11	10	16	14	18	6	2	5	7	11	7	5	5	5	4	2	S	10	5	8	8	2	18	8	24	
31	7	7	7	7	5	5	10	21	8	9	8	31	11	7	12	11	14	12	S	20	15	11	10	7	5	31	11	24	
HOURLY MAX	25	21	43	23	26	31	31	32	31	26	20	31	18	26	22	29	27	32	30	20	19	29	38	33					
HOURLY AVG	6	7	7	6	7	10	13	14	14	9	8	9	7	6	8	9	8	8	8	7	9	8	8	8					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	702
MAXIMUM INSTANTANEOUS VALUE:	43 ppb @ HOUR 2 ON DAY 25
	VAR-VARIOUS
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	9 hrs
OPERATIONAL TIME:	744 hrs
STANDARD DEVIATION:	7

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
Maskwa Continuous Monitoring Station - October 2018

WIND SPEED Instantaneous Maximum (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	6.0	7.3	4.2	3.5	2.4	2.7	1.9	1.4	1.9	2.7	3.6	5.0	6.8	7.4	5.9	6.4	5.1	4.7	4.4	4.5	3.3	1.9	2.1	4.9	1.4	7.4	4.2	24
2	2.6	2.5	3.3	6.0	6.8	2.9	3.9	5.7	8.5	8.2	8.8	11.0	10.6	13.6	15.2	14.6	11.7	11.5	6.1	6.6	6.0	7.4	4.5	4.7	2.5	15.2	7.6	24
3	4.9	4.8	5.6	8.6	7.8	10.4	7.2	10.6	10.8	12.7	11.7	15.9	14.1	15.9	13.4	11.5	8.6	6.6	3.4	3.6	2.6	2.7	1.8	5.8	1.8	15.9	8.4	24
4	7.4	6.3	8.5	7.1	5.0	2.0	2.1	2.7	4.8	6.8	5.5	9.3	10.1	9.3	9.7	8.3	7.4	5.6	6.5	8.5	9.1	11.6	8.6	7.7	2.0	11.6	7.1	24
5	7.2	6.8	8.3	8.0	7.9	5.4	4.3	3.9	5.9	7.1	8.0	9.7	10.9	10.4	9.9	7.7	9.8	9.5	4.7	5.2	3.8	2.2	2.4	4.1	2.2	10.9	6.8	24
6	3.9	2.9	2.5	2.7	3.0	7.5	6.4	10.6	12.6	16.3	11.8	12.4	9.2	9.0	7.9	7.2	5.2	3.5	3.9	3.7	2.7	3.8	5.2	7.7	2.5	16.3	6.7	24
7	6.6	7.6	7.7	11.0	9.0	8.4	10.4	8.9	7.8	9.4	8.0	10.1	11.2	11.7	11.3	9.9	10.5	10.2	14.5	14.4	15.5	16.2	12.3	11.0	6.6	16.2	10.6	24
8	11.1	10.7	10.8	10.9	9.7	8.7	8.8	8.7	9.4	9.1	8.8	9.7	9.9	8.5	8.2	7.5	7.4	2.5	2.4	1.5	1.3	1.0	1.4	1.0	1.0	11.1	7.4	24
9	1.5	1.5	1.4	0.9	2.1	2.5	2.1	2.8	3.3	5.8	7.8	8.0	8.2	7.4	8.7	6.5	4.8	4.8	3.4	2.0	1.9	2.4	2.3	1.4	0.9	8.7	3.9	24
10	1.5	2.1	1.6	1.9	1.6	0.5	1.7	2.0	2.9	5.6	5.5	7.1	5.5	5.9	4.0	2.8	4.0	1.5	1.6	1.4	2.0	4.2	2.9	2.1	0.5	7.1	3.0	24
11	1.5	1.8	1.8	2.9	4.2	4.3	3.6	4.9	5.1	6.8	7.7	8.2	9.1	9.8	10.0	11.9	12.6	8.8	9.1	13.3	10.2	8.0	6.8	8.2	1.5	13.3	7.1	24
12	8.0	6.9	5.3	4.3	3.6	3.3	4.2	2.4	6.0	6.6	5.8	11.0	13.3	17.3	16.7	16.8	16.9	19.7	17.3	12.5	9.2	10.6	13.9	13.5	2.4	19.7	10.2	24
13	13.2	16.3	12.0	13.0	12.4	12.2	13.1	12.7	13.3	14.1	12.2	12.4	13.3	14.9	13.0	19.0	11.9	11.6	9.6	6.7	5.7	3.3	3.0	2.7	2.7	19.0	11.3	24
14	4.0	4.1	3.7	3.6	4.0	6.1	4.8	7.7	7.1	8.4	10.5	10.0	10.7	13.3	13.6	12.9	12.7	9.1	9.5	11.1	7.7	7.8	6.4	7.0	3.6	13.6	8.2	24
15	12.2	11.3	10.4	12.1	9.9	8.7	10.6	9.7	11.8	14.0	14.0	16.7	16.8	18.4	15.7	15.0	13.5	10.4	7.8	7.6	8.1	7.7	7.5	5.9	5.9	18.4	11.5	24
16	4.8	4.7	3.8	4.3	2.4	1.6	1.8	1.8	4.0	4.4	4.5	6.7	6.0	5.5	5.5	5.0	4.6	3.4	1.6	3.7	4.3	4.2	9.3	8.7	1.6	9.3	4.4	24
17	7.9	4.4	7.3	6.8	5.7	5.2	5.8	6.9	8.2	9.9	11.4	8.7	10.1	8.9	10.9	12.0	10.3	8.8	9.1	8.6	7.9	6.6	4.8	5.0	4.4	12.0	8.0	24
18	5.5	4.7	5.8	7.4	5.7	9.5	13.5	10.7	7.9	9.5	8.4	9.9	12.6	17.3	18.0	19.3	15.1	18.3	10.5	12.5	12.5	9.8	8.5	4.7	4.7	19.3	10.9	24
19	7.8	6.1	5.9	4.6	7.7	6.1	6.4	6.4	10.7	13.3	12.6	16.5	11.9	11.9	12.4	10.0	7.5	4.6	3.7	4.6	5.0	3.6	9.4	8.8	3.6	16.5	8.2	24
20	13.5	11.0	7.5	7.7	10.0	7.9	9.2	11.4	9.4	8.9	10.7	10.7	9.4	16.9	14.4	18.2	18.9	16.6	11.5	12.0	12.6	11.0	11.3	9.9	7.5	18.9	11.7	24
21	8.5	8.1	8.8	8.1	4.3	6.4	6.5	5.5	14.6	16.4	18.8	16.3	16.1	14.8	17.1	16.0	18.1	7.5	4.6	4.3	3.6	4.1	1.7	1.7	1.7	18.8	9.7	24
22	1.8	1.7	1.5	1.6	3.1	1.7	1.3	1.6	1.6	5.1	7.0	8.6	8.4	10.5	9.0	11.7	9.1	4.0	5.8	8.6	8.7	6.5	7.1	3.6	1.3	11.7	5.4	24
23	6.9	7.2	2.4	3.1	2.9	4.7	4.8	8.4	7.4	6.3	8.2	8.3	10.7	9.5	6.4	5.6	4.4	4.1	5.8	3.9	2.3	1.8	1.7	1.6	1.6	10.7	5.3	24
24	4.1	2.7	1.6	3.4	2.9	2.3	2.2	3.2	3.7	4.3	4.7	6.1	6.4	7.2	5.9	4.3	3.8	1.7	2.1	2.5	2.9	3.1	3.6	3.3	1.6	7.2	3.7	24
25	2.7	3.6	5.0	3.5	2.1	3.3	3.8	1.2	3.3	3.9	5.1	3.7	4.9	6.1	5.8	4.1	2.2	2.8	2.3	2.2	1.9	1.7	1.9	1.4	1.2	6.1	3.3	24
26	2.8	4.0	2.7	3.4	5.4	5.2	5.9	4.2	4.7	11.1	12.0	9.8	12.3	10.0	9.5	8.3	4.2	2.6	4.1	2.8	2.3	3.2	2.1	1.4	1.4	12.3	5.6	24
27	2.2	2.8	3.2	5.0	5.0	4.9	6.0	5.5	11.5	13.0	13.6	13.1	17.9	18.0	17.8	12.1	8.7	5.4	5.2	11.1	9.7	8.0	6.2	4.4	2.2	18.0	8.8	24
28	4.7	5.7	3.8	3.4	4.2	4.1	2.7	1.4	6.0	3.5	5.6	6.4	6.5	8.6	7.4	6.6	6.0	5.8	7.1	9.8	8.3	9.7	11.7	8.5	1.4	11.7	6.1	24
29	5.4	7.4	5.5	4.5	6.2	6.7	7.5	7.3	8.4	7.1	10.4	10.5	10.2	12.2	13.8	13.1	10.1	8.3	13.0	9.1	8.7	9.0	6.6	5.4	4.5	13.8	8.6	24
30	4.2	5.2	4.3	4.5	5.4	2.8	4.9	3.0	4.4	4.0	6.6	6.0	4.8	4.5	3.9	5.1	5.1	4.5	4.7	3.7	2.7	2.3	2.8	4.5	2.3	6.6	4.3	24
31	3.9	5.8	5.6	6.3	6.2	4.4	5.5	4.9	5.2	3.3	3.9	6.0	5.6	6.6	6.3	11.1	9.0	2.8	1.8	2.0	1.5	2.3	1.6	2.2	1.5	11.1	4.7	24
HOURLY MAX	13.5	16.3	12.0	13.0	12.4	12.2	13.5	12.7	14.6	16.4	18.8	16.7	17.9	18.4	18.0	19.3	18.9	19.7	17.3	14.4	15.5	16.2	13.9	13.5				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

MAXIMUM INSTANTANEOUS VALUE:	19.7	kph	@ HOUR	17	ON DAY	12	
OPERATIONAL TIME:						744	hrs

WIND SPEED Instantaneous Maximum (WS kph)



***APPENDIX IV
REPORT CERTIFICATION FORM***

Report Certification Form

Alberta Airshed (if applicable)	EPA Approval or Code of Practice Registration # (if applicable)
YES	NA
Company Name (if applicable)	Industrial Operation Name (if applicable)
LAKELAND INDUSTRY & COMMUNITY ASSOCIATION	Maskwa Continuous Monitoring Station
Name of the Representative of the Person Responsible	Position / Title of the Representative of the Person Responsible
Mike Bisaga	Environment Monitoring Program Manager
Is an External Party Certifying the Report?	
<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	
Name of External Person Certifying the Report	Position / Title of External Person Certifying the Report
Wunmi Adekanmbi	Project Team Lead, Customer Service, Air Services
Company Name for External Person Certifying the Report	Identification of Qualifications / Professional Designations of the External Person Certifying the Report
Maxxam Analytics, A Bureau Veritas Group Company	M.Sc., EPT., PMP

Maxxam Analytics is the designated contractor conducting monitoring and reporting activities. I certify that the submitted data has been (a) reviewed and validated as per the AMD Chapter 6: Ambient Data Quality. I certify that the submitted report (b) accurately reflects the monitoring results and reporting timeframe and (c) meets the specified analysis, summarization and reporting requirements as per the AMD Chapter 9: Reporting.



Signature of the External Person Certifying the Report

29 - Nov -2018

Report Issued Date (dd-mon-yyyy)

***APPENDIX V
DATA VALIDATION CERTIFICATION FORM***



Validation Certificate Form

Client: <u>Lakeland Industry & Community Association</u>	Project #: <u>2833-2018-10-30-C</u>
Site: <u>Maskwa Continuous Monitoring Station</u>	Contact: <u>Mike Bisaga</u>

Level 0 Preliminary Verification	<u>Maram Ghaleb</u>	Date <u>26 - Nov - 2018</u>
Level 1 Primary Validation	<u>Maram Ghaleb</u>	Date <u>26 - Nov - 2018</u>
Level 2 Final Validation	<u>Maram Ghaleb</u>	Date <u>27 - Nov - 2018</u>
Level 3 Independent Data Review	<u>cradamba</u>	Date <u>29 - Nov - 2018</u>
Post-Final Validation	<u>NA</u>	Date <u>NA</u>

Notes
The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. This validation is performed on an annual basis.

Alberta Environment and Parks (AEP)
Air.Reporting@gov.ab.ca

December 10, 2018

Subject: Monthly Report Submission for the LICA St. Lina station

Lakeland Industry & Community Association (LICA) is pleased to submit the ambient air monitoring monthly report for the LICA St. Lina AQM Station in the month of October 2018.

The air monitoring program consists of continuous air monitoring results for Sulphur Dioxide (SO₂), Hydrogen Sulphide (H₂S), Total Hydrocarbon (THC), Oxides of Nitrogen (NO_x), Nitric Oxides (NO), Nitrogen Dioxide (NO₂), Ozone (O₃), Particulate Matter 2.5 (PM_{2.5}), Relative Humidity (RH), Barometric Pressure (BP), Precipitation, Ambient Temperature (AmbTPX), Wind Speed (WS), Wind Direction (WD) and Standard Deviation Wind Direction (STDWD).

Sampling Program	Monitoring Activities Conducted By	Sample Analysis Conducted By	Data/Report Review and Prepared By	Electronic Submission Conducted By
Continuous ambient air	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics	Maxxam Analytics

All data collected in October 2018 was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement systems.

Precipitation: configuration error occurred during the datalogger upgrade on May 30 and was corrected on November 1. Consequently, data collected during this time frame, at rates greater than or equal to 0.3 millimeter /minute, might be compromised and should therefore be used with caution.

As the LICA Environmental Program Manager and Data & Reporting Specialist, we have reviewed and verified this report and that the information is complete, accurate and representative of the monitoring results, reporting timeframe and the specified analysis, summarization and reporting requirements. We also verify all air data that are required by the AMD to be electronically submitted to AEP and Alberta's Ambient Air Quality Data Warehouse have been submitted by the time of this report submission.

Should you have any questions, please don't hesitate to contact us.

Respectfully,



Lakeland Industry & Community Association
5107 50 St
Bonnyville, AB T9N 2J7

A handwritten signature in blue ink that reads "Michael Bisaga".

Michael Bisaga
Technical Program Manager
Lakeland Industry & Community Association
780-266-7068
monitoring@lica.ca

A handwritten signature in blue ink that reads "Lily Lin".

Lily Lin
Data & Reporting Specialist
587-225-2248
monitoring@lica.ca



MAXXAM ANALYTICS
#1 2080 39 Ave. NE, Calgary, AB
T2E 6P7

maxxam.ca
Toll Free 800-386-7247
Fax 403-219-3673

AMBIENT AIR MONITORING MONTHLY DATA REPORT
LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
ST. LINA CONTINUOUS MONITORING STATION

JOB #: 2833-2018-10-31-C

October 2018

Prepared for:

LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
5107 50 ST.
BONNYVILLE, ALBERTA
T9N 2J7

Attention: MIKE BISAGA

DATE: **November 28, 2018**

Prepared by: *Maram Ghaleb*

Maram Ghaleb, B.Sc.
Project Manager, Customer Service, Air Services

Reviewed by: *Wunmi Adekanmbi*

Wunmi Adekanmbi, M.Sc., EPT, PMP
Project Team Lead, Customer Service, Air Services

SUMMARY

In October 2018, Maxxam Analytics was contracted to manage the ambient air quality monitoring and maintenance activities at the St. Lina Continuous Monitoring Station, near Bonnyville, Alberta. The monitoring station provides continuous meteorological measurements and air quality data for non-compliance parameters, as requested by the Lakeland Industry and Community Association.

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

All data collected this month were within the Alberta Ambient Air Quality Objectives and Guidelines (November, 2018).

Power failure: Two power failure events occurred on October 12 and October 17, respectively. Downtime ranging from eight to ten hours were recorded across parameters due the power failures and the subsequent analyzer recovery periods.

H₂S: A repeat calibration was performed on October 28 to correct biased high zero and span drifts, incurring six hours of downtime.

NO_x/NO/NO₂: One hour of downtime was recorded on October 12 at hour 08:00, due to an additional zero-span check performed to assess a biased high drift in span response.

WS/WD/STDWD: Four hours of downtime were recorded on October 29 due to anomalous WS spikes that were consequently invalidated, along with the corresponding WD and STDWD data.

Precipitation: A configuration error occurred during the datalogger upgrade on May 30, and was corrected on November 1. Consequently, data collected during this time frame, at rates greater than or equal to 0.3 mm/min, might be compromised and should therefore be used with caution.

The summary of results is presented on the following pages.

Any deviations or modifications made to the sampling or analytical methods are outlined in Section 1.0, Discussion. On this basis, Maxxam Analytics is issuing this completed report to Lakeland Industry & Community Association, St. Lina Continuous Monitoring Station.

Should you have any questions concerning the results or if we can be of further assistance, please contact us at 403-219-3677 or toll-free at 1-800-386-7247.

Monthly Continuous Data Summary

Lakeland Industry & Community Association						MAXIMUM VALUES							OPERATIONAL TIME (%)
St. Lina Continuous Monitoring Station						1-HOUR					24-HOUR		
PARAMETER	OBJECTIVES		EXCEEDANCES		MONTHLY AVERAGE	READING	DAY	HOUR	WIND SPEED (kph)	WIND DIRECTION (sector)	READING	DAY	
	1-hr	24-hr	1-hr	24-hr									
SO ₂ (ppb)	172	48	0	0	0	5	17	18	16.3	WSW	1	17	98.9
H ₂ S (ppb)	10	3	0	0	0	1	10	2	12.1	SE	0	10	98.0
THC (ppm)	-	-	-	-	2.07	2.85	24	13	7.3	E	2.31	24	98.7
CH ₄ (ppm)	-	-	-	-	2.07	2.79	24	13	7.3	E	2.31	24	98.7
NMHC (ppm)	-	-	-	-	0.00	0.06	1	0	7.3	SSE	0.00	1	98.7
NO ₂ (ppb)	159	-	0	-	2	10	17	16	20.9	WSW	5	24	98.7
NO (ppb)	-	-	-	-	0	2	9	8	3.5	ENE	1	24	98.7
NO _x (ppb)	-	-	-	-	2	10	17	16	20.9	WSW	5	24	98.7
O ₃ (ppb)	82	-	0	-	28.0	56.4	17	14	23.7	WSW	40.9	17	98.9
PM _{2.5} (µg/m ³)	80	30	0	0	4	17	3	19	7.2	NE	11	11	98.9
RELATIVE HUMIDITY (%)	-	-	-	-	68	100	11	23	7.7	SW	96	30	98.9
BAROMETRIC PRESSURE (millibar)	-	-	-	-	930	944	9	12	4.5	ESE	942	9	98.9
AMBIENT TEMPERATURE (°C)	-	-	-	-	2.6	21.8	17	14	23.7	WSW	12.4	17	98.9
PRECIPITATION (mm)	-	-	-	-	17.0	1.5	29	5	X	X	3.9	29	98.9
VECTOR WS (kph)	-	-	-	-	3.0	25.8	12	16	-	N	16.6	17	98.4
VECTOR WD (sec)	-	-	-	-	264 (W)	-	-	-	-	-	-	-	98.4

* Precipitation: data represents the total (sum) for the indicated time frame.

X = Invalid WS/WD value.

Exceedance Summary Report

SO₂ 1-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 1-hour AAAQO of 172 ppb.

SO₂ 24-Hour Exceedances

Measured concentrations of sulphur dioxide were below the 24-hour AAAQO of 48.0 ppb.

H₂S 1-Hour Exceedances

Measured concentrations of hydrogen sulphide were below the 1-hour AAAQO of 10 ppb.

H₂S 24-Hour Exceedances

Measured concentrations of hydrogen sulphide were below the 24-hour AAAQO of 3 ppb.

NO₂ 1-Hour Exceedances

Measured concentrations of nitrogen dioxide were below the 1-hour AAAQO of 159 ppb.

PM_{2.5} 1-Hour Exceedances

Measured concentrations of fine particulate matter were below the 1-hour AAAQO of 80 µg/m³.

PM_{2.5} 24-Hour Exceedances

Measured concentrations of fine particulate matter were below the 24-hour AAAQO of 30 µg/m³.

O₃ 1-Hour Exceedances

Measured concentrations of ozone were below the 1-hour AAAQO of 82 ppb.

In accordance with EPEA and the Substance Release Regulation.

In accordance with A Guide to Release Reporting and the Alberta Ambient Air Quality Objectives and Guidelines Summary.

TABLE OF CONTENTS

<u>Title</u>	<u>Page</u>
SUMMARY	2
MONTHLY CONTINUOUS DATA SUMMARY REPORT	3
EXCEEDANCE SUMMARY REPORT	4
TABLE OF CONTENTS	5
1.0 Discussion	7
2.0 Project Personnel	11
3.0 Plant Monthly Required AMD Summary	11
4.0 Calculations and Results	11
5.0 Methods and Procedures	12
Appendix I	15
Continuous Monitoring Data Results	15
Sulphur Dioxide	16
Hydrogen Sulphide	22
Total Hydrocarbon	28
Methane	34
Non-Methane Hydrocarbon	40
Oxides of Nitrogen	46
Nitric Oxides	52
Nitrogen Dioxide	57
Ozone	63
Particulate Matter 2.5	69
Wind Speed	74
Wind Direction	79
Standard Deviation Wind Direction	83
Relative Humidity	86
Barometric Pressure	89
Ambient Temperature	92
Precipitation	95

Appendix II	Equipment Calibration Results	98
	Sulphur Dioxide	99
	Hydrogen Sulphide	102
	Total Hydrocarbon	107
	Nitrogen Dioxide	111
	Ozone	115
	Particulate Matter	118
	Wind System	120
	Meteorological System Check	122
	Calibrators	124
	Calibration Gases	126
Appendix III	Maximum Instantaneous Data	132
Appendix IV	Report Certification Form	153
Appendix V	Data Validation Certification Form	155

1.0 Discussion

This monthly report consists of continuous monitoring results for the following parameters: Sulphur Dioxide (SO₂), Hydrogen Sulphide (H₂S), Total Hydrocarbon (THC), Methane (CH₄), Non-Methane Hydrocarbon (NMHC), Oxides of Nitrogen (NO_x), Nitric Oxides (NO), Nitrogen Dioxide (NO₂), Ozone (O₃), Particulate Matter 2.5 (PM_{2.5}), Relative Humidity (RH), Barometric Pressure (BP), Precipitation, Ambient Temperature (AmbTPX), Wind Speed (WS), Wind Direction (WD) and Standard Deviation Wind Direction (STDWD).

The sample inlet filter for all continuous air analyzers are replaced before the calibration begins. The sample manifold is cleaned during the site visit each month.

Control checks, consisting of a zero and span, are conducted daily on all continuous air monitors. In place of the air sample, zero air (from scrubbed air or gas cylinders) is used for zero checks, and a known concentration of the pollutant being analyzed is used for span checks. These checks are controlled by automatic timers and valves. The total zero span cycle is completed within an hour, the commencement of the zero span cycle is at the beginning of the hour.

Multipoint calibrations are done a minimum of once a month for each continuous air monitor. An additional calibration is required under the following conditions: 1) within three days after the initial start-up and stabilization of a newly installed instrument, 2) prior to shut-down or moving of an instrument which has been working to specification, and 3) when major repair has been done on the instrument.

Time during the first multi-point calibration is not considered downtime (Data is flagged as C). If more than one calibration is performed during the month, the time during the additional calibration is considered as downtime (Data is flagged as C1).

Only one zero/span check is run per day. Time during the zero/span check is not considered as downtime (Data is flagged as S). If an extra zero/span check is performed, the time during the additional check is considered as downtime (Data is flagged as S1).

The AMD requires each instrument and accompanying data recording system to be operational 90% of the time, at a minimum, for each monthly monitoring period.

All sampling, analysis, and QA/QC for this project was performed by Maxxam Analytics and complies with the Alberta Air Monitoring Directive.

Data contained in this monthly report has undergone the verification and validation based on the requirements of the AMD Chapter 6: Ambient Data Quality (December, 2016). The descriptions of the data verification and validation process can be found in Section 5 of this report. Instantaneous data, where applicable, is provided for reference purposes and has not undergone zero correction. The minimum and maximum statistics are highlighted in the data table and are for reference only. The highlighted cells are based on the software's interpretation of the exact position of the minimum or maximum value. The visual presentation of these statistics may not be the obvious choice in a data range due to rounding, truncating or analyzer specifications.

Hourly/minute data have been reviewed based on daily zero/span results and multi-point calibration results. Data may be considered invalid if a zero-corrected span check in excess of +/- 10% of the span concentration (established by the previous multi-point calibration) is encountered and/or significant differences in the calibration factor occurs (greater than 10%).

SULPHUR DIOXIDE (SO₂)

- Operational time for the monitoring period was 98.9%, equivalent to 8 hours of downtime.
- The routine monthly calibration was performed on October 10.
- Two power failure events occurred on October 12, between hours 05:00 and 11:00, and October 17 at hour 09:00, incurring 7 hours and 1 hour of downtime, respectively. Minute data recorded at 12:02 – 12:05 on October 12 and 10:05 – 10:14 on October 17, were flagged as recovery period and the corresponding hour was re-averaged.
- One instance of maximum instantaneous data was invalidated on October 12 at hour 13:00 due to missing minute data that could not be recovered.

HYDROGEN SULPHIDE (H₂S)

- Operational time for the monitoring period was 98.0%, equivalent to 15 hours of downtime.
- The routine monthly calibration was performed on October 10.
- Two power failure events occurred on October 12, between hours 05:00 and 11:00, and October 17 at hour 09:00, incurring 7 hours and 1 hour of downtime, respectively. Hour 12:00 of October 12 was flagged as recovery period, incurring an additional hour of downtime. Minute data recorded at 10:05 – 10:14 on October 17 were flagged as recovery period and the corresponding hour was re-averaged.
- The analyzer exhibited biased high zero and span drifts beginning on October 25. This prompted a site visit on October 28, where a successful repeat calibration was completed. No further issues were identified. Six hours of downtime were, however, incurred due to the additional quality check.
- One instance of maximum instantaneous data was invalidated on October 12 at hour 13:00 due to missing minute data that could not be recovered.

TOTAL HYDROCARBONS (THC), METHANE (CH₄) and NON-METHANE HYDROCARBONS (NMHC)

- Operational time for the monitoring period was 98.7%, equivalent to 10 hours of downtime.
- The routine monthly calibration was performed on October 11.
- Two power failure events occurred on October 12, between hours 05:00 and 11:00, and October 17 at hour 09:00, incurring 7 hours and 1 hour of downtime, respectively. Hour 12:00 of October 12 and hour 10:00 of October 17 were flagged as recovery period, incurring two additional hours of downtime.
- Elevated minute concentrations were observed on October 25 from 08:50 to 08:52. These minutes were excluded and the corresponding hourly average was re-calculated. The corresponding maximum instantaneous data at hour 08:00 was also invalidated.
- One instance of maximum instantaneous data was invalidated on October 2 at hour 11:00, due to an anomalous spike. The anomalous minute data was removed and the hourly concentration was re-averaged.
- One instance of maximum instantaneous data was invalidated on October 12 at hour 13:00 due to missing minute data that could not be recovered.

OXIDES OF NITROGEN (NO_x), NITRIC OXIDE (NO) and NITROGEN DIOXIDE (NO₂)

- Operational time for the monitoring period was 98.7%, equivalent to 10 hours of downtime.
- The routine monthly calibration was performed on October 10.
- Two power failure events occurred on October 12, between hours 05:00 and 11:00, and October 17 at hour 09:00, incurring 7 hours and 1 hour of downtime, respectively. Hour 12:00 of October 12 was flagged as recovery period, incurring an additional hour of downtime. Minute data recorded at 10:05 – 10:14 on October 17 were flagged as recovery period and the corresponding hour was re-averaged.
- One hour of downtime was recorded on October 15 at hour 08:00, due to an additional zero-span check performed to assess a biased high drift in span response.
- One instance of maximum instantaneous data was invalidated on October 12 at hour 13:00 due to missing minute data that could not be recovered.

OZONE (O₃)

- Operational time for the monitoring period was 98.9%, equivalent to 8 hours of downtime.
- The routine monthly calibration was performed on October 11.
- Two power failure events occurred on October 12, between hours 05:00 and 11:00, and October 17 at hour 09:00, incurring 7 hours and 1 hour of downtime, respectively. Minute data recorded at 10:05 – 10:14 on October 17 were flagged as recovery period and the corresponding hour was re-averaged.
- One instance of maximum instantaneous data was invalidated on October 12 at hour 13:00 due to missing minute data that could not be recovered.

PARTICULATE MATTER < 2.5 MICRONS (PM_{2.5})

- Operational time for the monitoring period was 98.9%, equivalent to 8 hours of downtime.
- The routine monthly check was performed on October 11.
- Two power failure events occurred on October 12, between hours 05:00 and 11:00, and October 17 at hour 09:00, incurring 7 hours and 1 hour of downtime, respectively. Minute data recorded at 12:02 – 12:05 on October were flagged as recovery period and the corresponding hour was re-averaged.

WIND SPEED (WS), WIND DIRECTION (WD) and STANDARD DEVIATION WIND DIRECTION (STDWD)

- Operational time for the monitoring period was 98.4%, equivalent to 12 hours of downtime.
- Two power failure events occurred on October 12, between hours 05:00 and 11:00, and October 17 at hour 09:00, incurring 7 hours and 1 hour of downtime, respectively.
- For undetermined reasons, anomalous WS spikes were recorded on October 29 at hours 04:00 – 07:00, and were therefore invalidated, along with the corresponding WD and STDWD data. Four hours of downtime were incurred as a result.
- One instance of maximum instantaneous data was invalidated on October 12 at hour 13:00 due to missing minute data that could not be recovered.
- Wind data is reported as vector wind speed and vector wind direction. Wind direction is defined as the direction from which the wind is blowing and is measured in degrees from true north.

RELATIVE HUMIDITY (RH)

- Operational time for the monitoring period was 98.9%, equivalent to 8 hours of downtime.
- Two power failure events occurred on October 12, between hours 05:00 and 11:00, and October 17 at hour 09:00, incurring 7 hours and 1 hour of downtime, respectively.

BAROMETRIC PRESSURE (BP)

- Operational time for the monitoring period was 98.9%, equivalent to 8 hours of downtime.
- Two power failure events occurred on October 12, between hours 05:00 and 11:00, and October 17 at hour 09:00, incurring 7 hours and 1 hour of downtime, respectively.

PRECIPITATION (PRECIP)

- Operational time for the monitoring period was 98.9%, equivalent to 8 hours of downtime.
- A precipitation sensor check was conducted on October 11. The result was satisfactory.
- A configuration error occurred during the datalogger upgrade on May 30, and was corrected on November 1. Consequently, data collected during this time frame, at rates greater than or equal to 0.3 mm/min, might be compromised and should therefore be used with caution.
- Two power failure events occurred on October 12, between hours 05:00 and 11:00, and October 17 at hour 09:00, incurring 7 hours and 1 hour of downtime, respectively.

AMBIENT TEMPERATURE (AmbTPX)

- Operational time for the monitoring period was 98.9%, equivalent to 8 hours of downtime.
- Two power failure events occurred on October 12, between hours 05:00 and 11:00, and October 17 at hour 09:00, incurring 7 hours and 1 hour of downtime, respectively.

2.0 Project Personnel

Mike Bisaga and Lily Lin were the contacts for Lakeland Industry & Community Association and the Maxxam field technician was Alexander Yakupov.

3.0 Plant Monthly Required AMD Summary

All data collected this month was compliant with the requirements outlined in the Air Monitoring Directive (Alberta Environment and Parks, 2016).

The operational time for all continuous ambient air analyzers, meteorological systems and data acquisition systems were above the 90% requirement.

All data collected this month were within the Alberta Ambient Air Quality Objectives and Guidelines (November, 2018).

4.0 Calculations and Results

All calculations and reporting of results follow the methods described in the AMD, 2016.

A THC/CH₄/NMHC spike was recorded between 08:50 - 08:52 on October 25 and deemed to be valid data by Maxxam Analytics. At LICA's request, minute data recorded during this period of time was discarded as other monitoring parameters did not support this spike event. The corresponding hourly average was re-calculated.

5.0 Methods and Procedures

The following methods and procedures were used to complete the monitoring program:

Maxxam AIR SOP-00001: Methane, Non-Methane Hydrocarbon Analyzer Monitoring
Maxxam AIR SOP-00014: Measurement of Particulate Concentration Using the THERMO SHARP
Maxxam AIR SOP-00209: Ambient Sulphur Monitoring
Maxxam AIR SOP-00212: Ambient O₃ Monitoring
Maxxam AIR SOP-00213: Ambient NO/NO₂/NO_x Monitoring
Maxxam AIR SOP-00242: Precipitation Collector Installation/Maintenance
MET One Instruments: Operation Manual Document No. 50.5-9800

There were no deviations from the prescribed methods.

The following instruments were used to perform the test program:

Sulphur Dioxide - Thermo 43i-TLE UV Fluorescent Analyzer
Hydrogen Sulphide - Thermo 450i UV Fluorescent Analyzer
Methane, Non-Methane Hydrocarbon - Thermo 55i FID Analyzer
Oxides of Nitrogen - Thermo 42i Chemiluminescent Analyzer
Ozone - Thermo 49i Photometric Analyzer
Particulate Matter (PM_{2.5}) - Thermo SHARP 5030i Unit
Wind System - Met One Unit
Relative Humidity - Campbell Scientific Unit
Barometric Pressure - Met One Unit
Ambient Temperature - Campbell Scientific Unit
Precipitation - Met One Unit
Datalogger - Envista Ultimate

The following steps were used to complete the data verification and validation process:

Level 0 Preliminary Verification

Level 0 data are raw data obtained directly from the data acquisition system (DAS). Under the step of Level 0, these data undergo a certain amount of manual or automated screening and flagging. It included a) identification of periods of missing data; b) verification of time stamps against reference time; c) verification that instrument diagnostics/datalogger flags indicate normal operation; d) comparison of data to upper and lower limits; e) rate of change flagging indicating that data changed too rapidly or not at all; and f) verification that zero, span and multipoint performance checks are within specifications. This level of verification is performed on a daily basis.

Level 1 Primary Validation

Validation actions under the step of Level 1 include a) review of all screening flags assigned during preliminary verification; b) review of all supporting site information and documentation; c) review of operational acceptance limits for each parameter/analyzer; d) review of daily zero/span and monthly calibration results for all gaseous parameters; and e) application of any necessary adjustments to data (e.g. baseline adjustments, below zero adjustments). This level of validation is performed on a monthly basis.

Level 2 Final Validation


The purpose of Level 2 validation is to verify that there are no inconsistencies among related data, or among regional data measured at nearby sites.

Level 3 Independent Data Review

Level 3 validation is the last step of data review, and it is completed by an individual that is independent of both field operations and primary data validation. A final independent QA review and endorsement is performed during this step before data is submitted to Alberta Environment.

Post-Final Validation

The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. Any data issues or patterns which were not clear on a monthly basis are highlighted during this step. This validation is performed on an annual basis.



Source: Air Monitoring Directive (December 2016), Chapter 6, Ambient Data Quality; Figure 1 Data Collection and Management Process Flow Chart

APPENDIX I
CONTINUOUS MONITORING DATA RESULTS

SULPHUR DIOXIDE

SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
2	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
3	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
4	0	0	0	0	0	0	0	0	S	0	0	0	0	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
5	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
6	0	0	1	0	0	S	0	0	0	0	0	0	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
7	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
8	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
9	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
10	0	S	0	0	0	0	0	0	0	0	0	0	C	C	C	C	C	C	0	0	0	0	0	0	0	0	0	0	0	0	24	
11	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	24	
12	0	0	0	0	0	P	P	P	P	P	P	P	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	17	
13	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	24	
14	0	0	0	0	0	0	0	0	1	0	0	0	2	1	0	0	0	0	1	2	S	1	0	0	0	0	0	2	0	0	24	
15	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	24	
16	0	0	0	0	0	0	0	0	0	0	0	0	1	1	1	1	1	1	1	S	0	0	0	0	1	0	1	0	1	0	24	
17	1	1	1	1	1	1	0	0	1	P	1	1	1	0	1	1	3	S	5	2	1	1	1	1	1	0	0	5	1	0	23	
18	2	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	2	0	0	24	
19	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
20	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
21	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
22	0	0	0	0	0	0	0	0	0	0	0	0	S	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
23	0	0	0	0	0	0	0	0	0	0	0	1	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
24	0	0	0	0	0	0	1	1	1	1	0	S	0	0	1	1	1	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
25	0	0	0	0	0	0	1	0	1	S	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24	
26	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	1	0	1	0	0	0	0	0	1	0	24	
27	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
28	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
29	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
30	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24	
31	0	0	0	0	S	1	0	0	0	0	0	0	0	1	2	1	1	2	3	1	1	0	0	0	0	0	0	0	3	1	24	
HOURLY MAX	2	1	1	1	1	1	1	1	1	0	1	1	2	1	2	1	3	2	5	2	1	1	1	1	1	1	1					
HOURLY AVG	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT:	1-HR	172	ppb	24-HR	48	ppb
----------------------	------	-----	-----	-------	----	-----


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0		
NUMBER OF 24-HR EXCEEDANCES:	0		
NUMBER OF NON-ZERO READINGS:	68		
MINIMUM 1-HR AVERAGE:	0 ppb @ HOUR	0 ON DAY	1
MAXIMUM 1-HR AVERAGE:	5 ppb @ HOUR	18 ON DAY	17
MAXIMUM 24-HR AVERAGE:	1 ppb	ON DAY	17
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	736 hrs
MONTHLY CALIBRATION TIME:	6 hrs	AMD OPERATION UPTIME:	98.9 %
STANDARD DEVIATION:	0	MONTHLY AVERAGE:	0 ppb

24 HR AVERAGES October 2018



SULPHUR DIOXIDE Hourly Averages (SO₂ ppb)



Wind: LICA ST. LINA
 Poll.: LICA ST. LINA-SO2[ppb]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 0.00%

Calm Avg: 0.00 [ppb]

Direction	0.0-1.2	1.2-2.4	2.4-3.6	3.6-4.8	4.8-6.0	>6.0	Total
N	6.9	0.0	0.0	0.0	0.0	0.0	6.9
NE	6.0	0.1	0.0	0.0	0.0	0.0	6.2
E	12.7	0.1	0.1	0.0	0.0	0.0	12.9
SE	7.5	0.0	0.0	0.0	0.0	0.0	7.5
S	12.7	0.0	0.0	0.0	0.0	0.0	12.7
SW	14.1	1.0	0.1	0.1	0.0	0.0	15.4
W	16.6	0.3	0.0	0.0	0.0	0.0	16.8
NW	21.6	0.0	0.0	0.0	0.0	0.0	21.6
Summary	98.0	1.6	0.3	0.1	0.0	0.0	100.0

% Icon Classes (ppb)

98 0.0-1.2

2 1.2-2.4


0 2.4-3.6

0 3.6-4.8

0 4.8-6.0

0 >6.0

LICA ST. LINA Poll.: LICA ST. LINA-SO2[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 0.00%



SO2[ppb] Calibration: LICA ST. LINA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



HYDROGEN SULPHIDE



HYDROGEN SULPHIDE Hourly Averages (H₂S ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
2	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
3	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
4	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
5	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
6	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
7	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
8	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
9	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
10	0	S	1	1	1	1	1	1	1	1	1	1	C	C	C	C	C	C	C	0	0	0	0	0	0	0	0	0	1	1	1	24
11	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	24
12	0	0	0	0	0	P	P	P	P	P	P	P	R	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	16
13	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	24
14	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	24
15	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	24
16	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	24
17	0	0	0	0	0	0	0	0	0	P	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	23
18	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	24
19	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
20	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
21	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
22	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
23	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
24	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
25	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
26	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
27	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
28	0	0	0	0	0	S	0	0	0	0	1	1	1	C1	C1	C1	C1	C1	C1	0	0	0	0	0	0	0	0	0	0	1	0	18
29	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
30	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
31	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
HOURLY MAX	0	0	1	1	1	1	1	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0				
HOURLY AVG	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT:	1-HR	10	ppb	24-HR	3	ppb
----------------------	------	----	-----	-------	---	-----


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0				
NUMBER OF 24-HR EXCEEDANCES:	0				
NUMBER OF NON-ZERO READINGS:	12				
MINIMUM 1-HR AVERAGE:	0 ppb @ HOUR	0	ON DAY	1	
MAXIMUM 1-HR AVERAGE:	1 ppb @ HOUR	2	ON DAY	10	
MAXIMUM 24-HR AVERAGE:	0 ppb		ON DAY	10	
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	729	hrs
MONTHLY CALIBRATION TIME:	7	hrs	AMD OPERATION UPTIME:	98.0	%
STANDARD DEVIATION:	0		MONTHLY AVERAGE:	0	ppb

24 HR AVERAGES October 2018



HYDROGEN SULPHIDE Hourly Averages (H₂S ppb)



% Icon Classes (ppb)

99

0.0-0.7

1

0.7-1.3


0

1.3-2.0

0

>2.0

LICA ST. LINA Poll.: LICA ST. LINA-H2S[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 0.00%



H2S[ppb] Calibration: LICA ST. LINA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



TOTAL HYDROCARBON


TOTAL HYDROCARBONS Hourly Averages (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2.14	2.28	2.29	2.42	2.45	2.42	2.44	2.42	2.41	2.23	S	2.02	2.02	2.01	2.01	2.02	2.10	2.01	2.02	2.01	2.02	2.02	2.03	2.02	2.01	2.45	2.17	2.4	
2	2.02	2.01	2.01	2.01	2.00	2.00	2.00	1.99	1.98	S	1.97	1.99	1.98	1.99	1.97	1.98	2.00	1.99	2.01	1.98	1.98	1.98	1.98	1.97	1.97	2.02	1.99	2.4	
3	2.02	1.98	1.97	1.97	1.97	1.97	1.97	1.98	S	1.98	1.98	1.98	1.96	1.94	1.96	1.97	1.96	1.94	1.96	1.99	2.00	2.00	2.01	2.01	1.94	2.02	1.98	2.4	
4	2.00	2.04	2.09	2.11	2.16	2.16	2.20	S	2.29	2.31	2.35	2.35	2.29	2.16	2.10	1.99	1.96	1.96	1.96	1.98	1.98	2.00	2.03	2.04	1.96	2.35	2.11	2.4	
5	2.03	2.02	2.03	2.05	2.05	2.06	S	2.08	2.08	2.29	2.21	2.06	2.00	2.01	1.99	2.01	1.98	1.96	1.97	1.99	2.00	2.06	2.04	2.06	1.96	2.29	2.05	2.4	
6	2.11	2.09	2.06	2.07	2.08	S	2.12	2.15	2.12	2.13	2.17	2.26	2.23	2.16	2.06	2.02	2.00	2.05	2.08	2.07	2.11	2.12	2.11	2.19	2.00	2.26	2.11	2.4	
7	2.19	2.04	2.06	2.12	S	2.17	2.18	2.09	2.04	2.01	2.01	2.00	2.00	2.01	1.99	1.98	1.99	1.98	1.98	1.98	1.98	1.98	1.98	1.98	1.98	1.98	2.19	2.03	2.4
8	1.98	1.98	1.99	S	1.98	1.99	1.98	1.98	1.99	1.99	1.99	1.99	1.99	2.00	1.98	2.00	1.99	1.99	1.99	1.99	1.99	1.99	1.98	1.96	1.96	2.00	1.99	2.4	
9	1.95	2.07	S	2.05	2.09	2.06	2.20	2.26	2.22	2.08	2.02	1.98	1.99	2.02	1.99	1.98	2.01	2.02	2.00	1.99	2.00	2.02	2.07	2.03	1.95	2.26	2.05	2.4	
10	2.02	S	2.01	2.02	2.04	2.15	2.31	2.40	2.40	2.34	2.32	2.24	2.15	2.10	2.03	2.04	2.01	2.03	2.01	2.01	2.01	2.02	2.01	1.98	1.98	2.40	2.11	2.4	
11	S	1.97	1.97	1.97	1.99	1.99	1.99	1.98	1.99	C	C	C	C	C	2.03	2.04	2.01	2.02	2.04	2.04	2.03	2.04	2.05	2.05	S	1.97	2.05	2.01	2.4
12	2.07	2.05	2.07	2.08	2.08	P	P	P	P	P	P	P	R	2.05	2.04	2.00	2.04	2.03	2.06	2.01	2.00	2.00	S	2.00	2.00	2.08	2.04	16	
13	2.01	1.99	1.96	1.99	1.98	2.01	2.00	2.01	2.00	2.02	2.03	1.98	2.00	2.02	2.02	2.01	2.01	2.01	2.02	2.07	2.09	S	2.02	2.02	1.96	2.09	2.01	2.4	
14	2.02	2.03	2.05	2.06	2.05	2.09	2.11	2.11	2.08	2.05	2.04	2.02	2.01	2.01	2.00	2.01	2.02	2.03	2.04	2.04	S	2.03	2.02	2.01	2.00	2.11	2.04	2.4	
15	2.00	1.98	1.97	1.97	2.02	1.97	1.96	1.97	1.97	1.97	1.99	2.02	2.03	2.08	2.01	2.01	2.06	2.02	2.04	S	2.07	2.08	2.08	2.08	1.96	2.08	2.01	2.4	
16	2.11	2.12	2.01	2.09	2.01	2.01	2.02	2.04	2.05	2.08	2.04	2.01	2.01	2.00	2.00	1.99	2.05	2.03	S	2.07	2.05	2.05	2.07	2.08	1.99	2.12	2.04	2.4	
17	2.08	2.08	2.07	2.08	2.08	2.08	2.07	2.06	2.06	P	R	2.04	2.03	2.01	1.99	1.99	2.03	S	2.03	2.03	2.03	2.03	2.04	1.99	2.08	2.08	2.04	22	
18	2.06	2.06	2.05	2.04	2.03	1.99	1.97	1.99	2.02	2.01	1.99	1.98	1.97	1.99	2.01	1.99	S	2.03	2.00	1.98	1.98	1.97	1.98	1.98	1.97	2.06	2.00	2.4	
19	1.98	1.99	1.99	2.00	1.99	2.00	2.00	2.03	2.02	2.01	2.00	2.01	2.00	2.00	2.00	S	2.00	2.01	2.04	2.04	2.03	2.05	2.06	2.16	1.98	2.16	2.02	2.4	
20	2.23	2.17	2.16	2.15	2.07	2.06	2.08	2.09	2.10	2.10	2.09	2.08	2.05	2.04	S	1.98	1.98	2.00	2.05	1.99	1.99	1.99	1.98	1.98	1.98	2.23	2.06	2.4	
21	1.99	2.00	2.01	2.00	2.00	2.00	1.99	1.99	1.98	1.97	1.97	1.98	1.99	S	1.98	2.02	2.06	2.02	1.99	1.99	2.00	2.05	2.00	2.04	1.97	2.06	2.00	2.4	
22	2.06	1.99	2.06	2.08	2.06	2.08	2.10	2.06	2.07	2.08	2.04	2.03	S	2.03	2.04	2.03	2.04	2.05	2.06	2.08	2.11	2.11	2.12	2.15	1.99	2.15	2.07	2.4	
23	2.18	2.20	2.22	2.28	2.35	2.36	2.33	2.34	2.34	2.30	2.28	S	2.27	2.20	2.14	2.13	2.24	2.27	2.22	2.17	2.17	2.25	2.26	2.25	2.13	2.36	2.25	2.4	
24	2.20	2.26	2.32	2.35	2.37	2.44	2.37	2.35	2.38	2.43	S	2.37	2.34	2.46	2.31	2.29	2.37	2.32	2.28	2.21	2.18	2.14	2.17	2.28	2.14	2.46	2.31	2.4	
25	2.23	2.26	2.27	2.32	2.32	2.45	2.27	2.32	2.28	S	2.22	2.17	2.14	2.11	2.10	2.04	2.26	2.07	2.10	2.12	2.11	2.09	2.02	2.01	2.01	2.45	2.19	2.4	
26	2.04	2.05	2.09	2.05	2.12	2.01	2.01	2.01	S	2.00	1.99	2.00	2.00	1.99	2.00	2.03	2.00	2.00	2.00	2.00	2.01	2.00	2.02	2.04	1.99	2.12	2.02	2.4	
27	2.05	2.04	2.02	2.02	2.01	2.00	2.01	S	2.06	2.05	2.02	2.00	1.99	2.01	2.02	2.02	2.00	1.99	1.98	1.98	2.01	2.03	1.99	1.98	1.98	2.06	2.01	2.4	
28	1.98	1.99	2.01	2.01	2.02	2.03	S	2.06	2.06	2.05	2.03	2.02	2.09	2.31	2.15	2.19	2.24	2.27	2.25	2.17	2.16	2.18	2.16	1.98	2.31	2.11	2.4		
29	2.20	2.18	2.12	2.15	2.12	S	2.03	2.02	2.04	2.07	2.09	2.09	2.05	2.09	2.13	2.05	2.08	2.09	2.12	2.07	2.05	2.11	2.08	2.06	2.02	2.20	2.09	2.4	
30	2.13	2.07	2.04	2.10	S	2.09	2.09	2.12	2.13	2.11	2.12	2.10	2.06	2.02	1.99	2.00	2.02	2.03	2.02	2.03	2.05	2.06	2.06	2.06	1.99	2.13	2.07	2.4	
31	2.06	2.06	2.05	S	2.07	2.08	2.09	2.10	2.11	2.11	2.10	2.09	2.07	2.09	2.06	2.06	2.10	2.14	2.12	2.12	2.13	2.13	2.09	2.05	2.14	2.07	2.09	2.4	
HOURLY MAX	2.23	2.28	2.32	2.42	2.45	2.45	2.44	2.42	2.41	2.43	2.35	2.37	2.34	2.46	2.31	2.29	2.37	2.32	2.28	2.21	2.18	2.25	2.26	2.28					
HOURLY AVG	2.07	2.07	2.07	2.09	2.09	2.10	2.10	2.11	2.12	2.11	2.08	2.07	2.06	2.07	2.04	2.03	2.05	2.04	2.05	2.04	2.04	2.05	2.05	2.06					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	697			
MINIMUM 1-HR AVERAGE:	1.94 ppm	@ HOUR	13	ON DAY 3
MAXIMUM 1-HR AVERAGE:	2.85 ppm	@ HOUR	13	ON DAY 24
MAXIMUM 24-HR AVERAGE:	2.31 ppm			ON DAY 24
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	734 hrs	
MONTHLY CALIBRATION TIME:	5 hrs	AMD OPERATION UPTIME:	98.7 %	
STANDARD DEVIATION:	0.10	MONTHLY AVERAGE:	2.07 ppm	

TOTAL HYDROCARBONS Hourly Averages (THC ppm)




% Icon Classes (ppm)


0  0.0-0.8

0  0.8-1.6

100  1.6-2.5

0  >2.5

LICA ST. LINA Poll.: LICA ST. LINA-THC[ppm] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 0.00%



THC[ppm] Calibration: LICA ST. LINA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



METHANE




METHANE Hourly Averages (CH₄ ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2.14	2.28	2.29	2.42	2.45	2.42	2.44	2.42	2.41	2.23	S	2.02	2.02	2.01	2.01	2.02	2.10	2.01	2.02	2.01	2.02	2.02	2.03	2.02	2.01	2.45	2.17	2.17	24
2	2.02	2.01	2.01	2.01	2.00	2.00	2.00	1.99	1.98	S	1.97	1.99	1.98	1.99	1.97	1.98	2.00	1.99	2.01	1.98	1.98	1.98	1.98	1.97	1.97	2.02	1.99	2.02	24
3	2.02	1.98	1.97	1.97	1.97	1.97	1.97	1.98	S	1.98	1.98	1.98	1.96	1.94	1.96	1.97	1.96	1.94	1.96	1.99	2.00	2.00	2.01	2.01	1.94	2.02	1.98	2.02	24
4	2.00	2.04	2.09	2.11	2.16	2.16	2.20	S	2.29	2.31	2.35	2.35	2.29	2.16	2.10	1.99	1.96	1.96	1.96	1.98	1.98	2.00	2.03	2.04	1.96	2.35	2.11	2.11	24
5	2.03	2.02	2.03	2.05	2.05	2.06	S	2.08	2.08	2.29	2.21	2.06	2.00	2.01	1.99	2.01	1.98	1.96	1.97	1.99	2.00	2.06	2.04	2.06	1.96	2.29	2.05	2.05	24
6	2.11	2.09	2.06	2.07	2.08	S	2.12	2.15	2.12	2.13	2.17	2.26	2.23	2.16	2.06	2.02	2.00	2.05	2.08	2.07	2.11	2.12	2.11	2.19	2.00	2.26	2.11	2.11	24
7	2.19	2.04	2.06	2.12	S	2.17	2.18	2.09	2.04	2.01	2.01	2.00	2.00	2.01	1.99	1.98	1.99	1.98	1.98	1.98	1.98	1.99	1.98	1.97	1.97	2.19	2.03	2.03	24
8	1.98	1.98	1.99	S	1.98	1.99	1.98	1.98	1.99	1.99	1.99	1.99	1.99	2.00	1.98	2.00	1.99	1.99	1.99	1.99	1.99	1.99	1.98	1.96	1.96	2.00	1.99	2.00	24
9	1.95	2.07	S	2.05	2.09	2.06	2.20	2.26	2.22	2.08	2.02	1.98	1.99	2.02	1.99	1.98	2.01	2.02	2.00	1.99	2.00	2.02	2.07	2.03	1.95	2.26	2.05	2.05	24
10	2.02	S	2.01	2.02	2.04	2.15	2.31	2.40	2.40	2.34	2.32	2.24	2.15	2.10	2.03	2.04	2.01	2.03	2.01	2.01	2.01	2.02	2.01	1.98	1.98	2.40	2.11	2.11	24
11	S	1.97	1.97	1.97	1.99	1.99	1.98	1.98	1.99	C	C	C	C	C	C	2.01	2.00	2.02	2.04	2.04	2.03	2.04	2.05	2.05	S	1.97	2.05	2.01	24
12	2.07	2.05	2.07	2.07	2.08	P	P	P	P	P	P	P	R	2.05	2.04	2.00	2.04	2.03	2.06	2.01	2.00	2.00	S	2.00	2.00	2.08	2.04	2.04	16
13	2.01	1.99	1.96	1.99	1.98	2.01	2.00	2.01	2.00	2.02	2.03	1.98	2.00	2.02	2.02	2.01	2.01	2.01	2.02	2.07	2.09	S	2.02	2.02	1.96	2.09	2.01	2.01	24
14	2.02	2.03	2.05	2.06	2.05	2.09	2.11	2.11	2.08	2.05	2.04	2.02	2.01	2.01	2.00	2.01	2.02	2.03	2.04	2.04	S	2.03	2.02	2.01	2.00	2.11	2.04	2.04	24
15	2.00	1.98	1.97	1.97	2.02	1.97	1.96	1.97	1.97	1.97	1.99	2.02	2.03	2.08	2.01	2.01	2.06	2.02	2.04	S	2.07	2.08	2.08	2.08	1.96	2.08	2.01	2.01	24
16	2.11	2.12	2.01	2.09	2.01	2.01	2.02	2.04	2.05	2.08	2.04	2.01	2.01	2.00	2.00	1.99	2.05	2.03	S	2.07	2.05	2.05	2.07	2.08	1.99	2.12	2.04	2.04	24
17	2.08	2.08	2.07	2.08	2.08	2.08	2.07	2.06	2.06	P	R	2.04	2.03	2.01	1.99	1.99	2.03	S	2.03	2.03	2.03	2.03	2.04	1.99	2.08	2.08	2.04	2.04	22
18	2.06	2.06	2.05	2.04	2.03	1.99	1.97	1.99	2.02	2.01	1.99	1.98	1.97	1.98	2.01	1.99	S	2.03	2.00	1.98	1.98	1.97	1.98	1.98	1.97	2.06	2.00	2.00	24
19	1.98	1.99	1.99	2.00	1.99	2.00	2.00	2.03	2.02	2.01	2.00	2.01	2.00	2.00	2.00	S	2.00	2.01	2.04	2.04	2.04	2.03	2.05	2.06	2.16	1.98	2.16	2.02	24
20	2.23	2.17	2.16	2.15	2.07	2.06	2.08	2.09	2.10	2.10	2.09	2.08	2.05	2.04	S	1.98	1.98	2.00	2.05	1.99	1.99	1.99	1.98	1.98	1.98	2.23	2.06	2.06	24
21	1.99	2.00	2.01	2.00	2.00	1.99	1.99	1.99	1.98	1.97	1.97	1.98	1.99	S	1.98	2.01	2.06	2.02	1.99	1.99	2.00	2.05	2.00	2.04	1.97	2.06	2.00	2.00	24
22	2.06	1.99	2.06	2.08	2.06	2.08	2.10	2.06	2.07	2.07	2.04	2.03	S	2.03	2.04	2.03	2.04	2.05	2.06	2.08	2.11	2.11	2.12	2.15	1.99	2.15	2.07	2.07	24
23	2.18	2.20	2.22	2.28	2.35	2.36	2.33	2.34	2.34	2.30	2.28	S	2.27	2.20	2.14	2.13	2.24	2.27	2.21	2.17	2.17	2.25	2.26	2.25	2.13	2.36	2.25	2.25	24
24	2.20	2.26	2.32	2.35	2.37	2.44	2.37	2.35	2.38	2.43	S	2.37	2.34	2.46	2.31	2.29	2.37	2.32	2.28	2.21	2.18	2.13	2.17	2.28	2.13	2.46	2.31	2.31	24
25	2.23	2.26	2.27	2.32	2.32	2.45	2.27	2.32	2.28	S	2.22	2.17	2.14	2.11	2.10	2.04	2.26	2.07	2.10	2.12	2.11	2.09	2.02	2.01	2.01	2.45	2.19	2.19	24
26	2.04	2.05	2.09	2.05	2.12	2.01	2.01	2.01	S	2.00	1.99	2.00	2.00	1.99	2.00	2.03	2.00	2.00	2.00	2.00	2.01	2.00	2.02	2.04	1.99	2.12	2.02	2.02	24
27	2.05	2.04	2.02	2.02	2.01	2.00	2.01	S	2.06	2.05	2.02	2.00	1.99	2.01	2.02	2.02	2.00	1.99	1.98	1.98	2.01	2.03	1.99	1.98	1.98	2.06	2.01	2.01	24
28	1.98	1.99	2.01	2.01	2.02	2.03	S	2.06	2.06	2.05	2.03	2.02	2.09	2.31	2.15	2.19	2.24	2.27	2.25	2.17	2.16	2.18	2.16	1.98	2.31	2.11	2.11	24	
29	2.20	2.18	2.12	2.15	2.12	S	2.03	2.02	2.04	2.07	2.09	2.09	2.05	2.09	2.13	2.05	2.07	2.09	2.12	2.07	2.05	2.11	2.08	2.06	2.02	2.20	2.09	2.09	24
30	2.13	2.07	2.04	2.10	S	2.09	2.09	2.12	2.13	2.11	2.12	2.10	2.06	2.02	1.99	2.00	2.02	2.03	2.02	2.03	2.05	2.06	2.06	2.06	1.99	2.13	2.07	2.07	24
31	2.06	2.06	2.05	S	2.07	2.08	2.09	2.10	2.11	2.11	2.10	2.09	2.07	2.09	2.06	2.06	2.10	2.14	2.12	2.12	2.13	2.13	2.09	2.05	2.14	2.07	2.07	2.07	24
HOURLY MAX	2.23	2.28	2.32	2.42	2.45	2.45	2.44	2.42	2.41	2.43	2.35	2.37	2.34	2.46	2.31	2.29	2.37	2.32	2.28	2.21	2.18	2.25	2.26	2.28					
HOURLY AVG	2.07	2.07	2.07	2.09	2.09	2.10	2.10	2.11	2.12	2.11	2.08	2.07	2.06	2.07	2.04	2.03	2.05	2.04	2.05	2.04	2.04	2.05	2.05	2.06					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	697			
MINIMUM 1-HR AVERAGE:	1.94 ppm	@ HOUR	13	ON DAY 3
MAXIMUM 1-HR AVERAGE:	2.79 ppm	@ HOUR	13	ON DAY 24
MAXIMUM 24-HR AVERAGE:	2.31 ppm			ON DAY 24
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	734 hrs	
MONTHLY CALIBRATION TIME:	5 hrs	AMD OPERATION UPTIME:	98.7 %	
STANDARD DEVIATION:	0.10	MONTHLY AVERAGE:	2.07 ppm	

METHANE Hourly Averages (CH₄ ppm)



% Icon Classes (ppm)


0 0.0-0.8

0 0.8-1.6

100 1.6-2.5


0 >2.5

LICA ST. LINA Poll.: LICA ST. LINA-CH4[ppm] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 0.00%



CH4[ppm] Calibration: LICA ST. LINA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



NON-METHANE HYDROCARBON




NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.																						
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.																							
DAY																																																		
1	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
2	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
3	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
4	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
5	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
6	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
7	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
8	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
9	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
10	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
11	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	C	C	C	C	C	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	24																						
12	0.00	0.00	0.00	0.00	0.00	P	P	P	P	P	P	P	R	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	16																						
13	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	24																						
14	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	24																						
15	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
16	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
17	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	P	R	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	22																						
18	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
19	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
20	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
21	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
22	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
23	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
24	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
25	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
26	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
27	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
28	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
29	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
30	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
31	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24																						
HOURLY MAX	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00																						
HOURLY AVG	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00																						

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	0			
MINIMUM 1-HR AVERAGE:	0.00	ppm @ HOUR	0	ON DAY 1
MAXIMUM 1-HR AVERAGE:	0.06	ppm @ HOUR	0	ON DAY 1
MAXIMUM 24-HR AVERAGE:	0.00	ppm		ON DAY 1
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	734 hrs
MONTHLY CALIBRATION TIME:	5	hrs	AMD OPERATION UPTIME:	98.7 %
STANDARD DEVIATION:	0.00		MONTHLY AVERAGE:	0.00 ppm

NON-METHANE HYDROCARBONS Hourly Averages (NMHC ppm)



Wind: LICA ST. LINA
 Poll.: LICA ST. LINA-NMHC[ppm]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 0.00%

Calm Avg: 0.00 [ppm]

Direction	0-0.074	0.074-0.148	0.148-0.222	0.222-0.296	0.296-0.37	>0.4	Total
N	6.8	0.0	0.0	0.0	0.0	0.0	6.8
NE	6.2	0.0	0.0	0.0	0.0	0.0	6.2
E	13.0	0.0	0.0	0.0	0.0	0.0	13.0
SE	8.1	0.0	0.0	0.0	0.0	0.0	8.1
S	12.7	0.0	0.0	0.0	0.0	0.0	12.7
SW	15.0	0.0	0.0	0.0	0.0	0.0	15.0
W	16.7	0.0	0.0	0.0	0.0	0.0	16.7
NW	21.6	0.0	0.0	0.0	0.0	0.0	21.6
Summary	100.0	0.0	0.0	0.0	0.0	0.0	100.0

% Icon Classes (ppm) 100 0-0.074 0 0.074-0.148 0 0.148-0.222 0 0.222-0.296 0 0.296-0.37 0 >0.4

LICA ST. LINA Poll.: LICA ST. LINA-NMHC[ppm] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 0.00%



NMHC[ppm] Calibration: LICA ST. LINA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



OXIDES OF NITROGEN




OXIDES OF NITROGEN Hourly Averages (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2	3	3	3	3	3	3	3	3	3	2	S	1	1	1	1	1	2	2	2	2	1	1	1	1	1	3	2	24
2	1	0	0	1	1	1	2	2	1	S	1	1	1	1	1	1	1	1	1	1	0	0	0	0	0	0	2	1	24
3	0	0	0	0	0	1	0	1	S	2	1	0	0	0	0	0	0	0	0	0	1	2	2	1	0	0	2	1	24
4	2	1	2	2	2	2	3	S	5	5	5	5	4	3	3	2	1	1	1	1	1	2	1	2	1	1	5	2	24
5	2	2	2	2	2	3	S	4	3	3	2	1	1	1	1	1	1	1	1	1	1	3	2	3	1	4	4	2	24
6	3	2	2	1	1	S	3	2	2	3	3	4	4	5	5	4	3	2	3	3	3	3	3	2	1	5	5	3	24
7	2	1	1	3	S	7	7	6	2	1	1	1	1	1	2	1	1	0	1	1	1	1	1	1	0	7	2	2	24
8	1	1	2	S	2	2	1	2	2	2	2	1	1	0	0	0	0	0	0	0	0	0	0	0	0	2	1	2	24
9	0	0	S	4	3	4	8	7	7	3	1	0	0	0	0	0	1	1	1	1	1	1	2	1	0	8	2	2	24
10	1	S	1	1	1	2	4	5	5	4	5	C	C	C	C	C	C	C	C	3	1	1	1	0	0	5	2	24	
11	S	2	1	1	1	1	1	1	1	2	2	2	2	2	2	2	2	3	3	2	2	2	2	S	1	3	2	24	
12	3	3	3	3	3	P	P	P	P	P	P	P	R	1	2	2	2	1	0	0	0	0	S	1	0	3	2	16	
13	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	S	1	1	0	1	0	24
14	1	0	1	1	1	1	5	5	4	2	2	2	6	5	3	2	3	3	6	7	S	6	3	2	0	7	3	24	
15	1	1	0	0	0	0	1	1	S1	2	2	1	1	1	1	1	2	1	1	S	2	1	1	1	0	2	1	23	
16	0	0	0	0	1	1	3	2	3	2	2	2	2	2	2	2	1	2	S	5	3	2	2	2	0	5	2	24	
17	3	3	3	6	5	4	3	3	P	5	4	3	2	2	4	10	S	10	6	5	4	4	4	4	2	10	4	23	
18	5	5	3	2	1	1	1	0	1	1	1	0	0	0	0	0	S	2	1	1	1	1	1	0	0	5	1	24	
19	0	0	0	0	0	0	0	1	1	1	1	1	0	1	0	S	2	1	2	2	2	2	2	2	0	2	1	24	
20	3	2	1	1	1	1	1	1	1	2	2	3	3	2	S	1	1	0	0	0	1	0	0	0	0	3	1	24	
21	0	1	1	1	1	0	0	0	0	0	0	0	0	0	S	1	1	0	0	0	1	0	0	0	0	1	0	24	
22	0	0	0	1	1	1	2	2	2	2	2	1	S	2	1	1	1	1	2	2	2	2	2	2	0	2	1	24	
23	2	2	2	2	3	3	3	3	3	4	4	S	4	3	2	2	3	5	5	4	4	4	4	4	2	5	3	24	
24	3	4	4	4	4	5	4	5	5	7	S	8	7	8	6	6	6	6	5	4	4	3	4	4	3	8	5	24	
25	4	4	4	4	4	3	5	4	S	5	4	4	3	3	2	4	3	4	4	4	4	4	2	2	2	5	4	24	
26	2	3	2	1	1	1	1	1	S	2	1	1	1	1	1	2	3	5	4	4	3	4	3	2	1	5	2	24	
27	3	2	2	1	1	1	1	S	3	3	2	1	1	1	0	1	0	0	0	1	1	1	0	1	0	3	1	24	
28	0	0	1	1	1	1	S	3	3	3	2	1	2	4	2	2	2	2	2	2	2	2	1	1	0	4	2	24	
29	1	1	1	2	2	S	3	2	2	1	1	1	1	1	1	2	1	1	1	1	1	1	1	1	1	3	1	24	
30	1	1	1	1	S	2	2	2	3	4	3	3	2	1	1	1	1	2	2	2	1	1	2	2	1	4	2	24	
31	2	2	2	S	6	6	5	6	6	6	6	6	6	8	9	6	6	6	7	7	5	3	2	2	2	9	5	24	
HOURLY MAX	5	5	4	6	6	7	8	7	7	7	6	8	7	8	9	6	10	6	10	7	5	6	4	4	2	9	5	24	
HOURLY AVG	2	2	2	2	2	2	3	3	3	3	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	3	2	24

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	571			
MINIMUM 1-HR AVERAGE:	0	ppb @ HOUR	1	ON DAY 2
MAXIMUM 1-HR AVERAGE:	10	ppb @ HOUR	16	ON DAY 17
MAXIMUM 24-HR AVERAGE:	5	ppb		ON DAY 24
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	734 hrs
MONTHLY CALIBRATION TIME:	8	hrs	AMD OPERATION UPTIME:	98.7 %
STANDARD DEVIATION:	2		MONTHLY AVERAGE:	2 ppb

OXIDES OF NITROGEN Hourly Averages (NO_x ppb)



% Icon Classes (ppb)

85



0.0-3.7

14



3.7-7.3

1




7.3-11.0

0



>11.0

LICA ST. LINA Poll.: LICA ST. LINA-NOX[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 0.00%



NOX[ppb] Calibration: LICA ST. LINA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



NITRIC OXIDE


NITRIC OXIDE Hourly Averages (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.																						
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.																							
DAY																																																		
1	0	0	0	0	0	0	0	0	0	0	1	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24																					
2	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	1	1	0	0	0	0	0	1	0	24																					
3	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24																					
4	0	0	0	0	0	0	0	0	S	1	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0	1	0	24																					
5	0	0	0	0	0	0	S	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24																					
6	0	0	0	0	0	S	0	0	0	1	1	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0	1	0	24																					
7	0	0	0	0	S	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24																					
8	0	0	0	S	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24																					
9	0	0	S	0	0	0	0	1	2	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	2	0	24																					
10	0	S	0	0	0	0	0	0	1	1	1	C	C	C	C	C	C	C	C	C	C	C	C	C	0	0	1	0	24																					
11	S	0	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	S	0	1	0	24																					
12	0	0	0	0	0	P	P	P	P	P	P	P	R	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	16																					
13	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	24																					
14	0	0	0	0	0	0	0	0	1	1	1	1	2	2	1	1	0	0	0	0	0	S	0	0	0	0	2	0	24																					
15	0	0	0	0	0	0	0	0	S1	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	23																					
16	0	0	0	0	0	0	1	0	1	0	0	1	0	1	0	0	0	0	S	1	0	0	0	0	0	0	1	0	24																					
17	0	0	0	0	0	0	0	0	0	P	2	1	0	0	0	0	1	S	0	0	0	0	0	0	0	0	2	0	23																					
18	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	24																					
19	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	24																					
20	0	0	0	0	0	0	0	0	0	0	1	1	1	1	0	S	0	0	0	0	0	0	0	0	0	0	1	0	24																					
21	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	24																					
22	0	0	0	0	0	0	0	0	0	0	1	1	0	S	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24																					
23	0	0	0	0	0	0	0	0	1	1	1	1	S	1	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24																					
24	0	0	0	0	0	0	0	0	1	2	S	2	2	2	2	1	1	0	0	0	0	0	0	0	0	0	2	1	24																					
25	0	0	0	0	0	0	0	1	1	S	1	1	1	0	0	0	1	0	0	0	0	0	0	0	0	0	1	0	24																					
26	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	1	0	0	0	1	0	0	0	0	1	0	24																					
27	0	0	0	0	0	0	0	S	1	1	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	1	0	24																					
28	0	0	0	0	0	S	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	1	0	24																					
29	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24																					
30	0	0	0	0	S	0	0	0	1	2	1	1	1	1	0	0	0	0	1	0	0	0	0	0	0	0	2	0	24																					
31	0	0	0	S	0	0	0	2	1	1	2	2	1	2	2	1	0	0	0	0	0	0	0	0	0	0	2	1	24																					
HOURLY MAX	0	0	0	0	0	0	1	2	2	2	2	2	2	2	2	1	1	1	1	1	1	0	1	0	0																									
HOURLY AVG	0	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0																									

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	90			
MINIMUM 1-HR AVERAGE:	0 ppb @ HOUR	0	ON DAY	1
MAXIMUM 1-HR AVERAGE:	2 ppb @ HOUR	8	ON DAY	9
MAXIMUM 24-HR AVERAGE:	1 ppb		ON DAY	24
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	734 hrs	
MONTHLY CALIBRATION TIME:	8 hrs	AMD OPERATION UPTIME:	98.7 %	
STANDARD DEVIATION:	0	MONTHLY AVERAGE:	0 ppb	

NITRIC OXIDE Hourly Averages (NO ppb)



% Icon Classes (ppb)


95 0.0-1.0

4 1.0-2.0

1 2.0-3.0

0 >3.0

LICA ST. LINA Poll.: LICA ST. LINA-NO[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 0.00%



NITROGEN DIOXIDE

NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2	3	3	3	3	3	3	3	3	3	2	S	1	1	1	1	1	2	2	2	2	1	1	1	1	1	3	2	24
2	1	0	0	1	1	1	2	1	1	S	1	1	1	0	0	0	0	0	1	0	0	0	0	0	0	0	2	1	24
3	0	0	0	0	0	0	0	0	0	S	2	1	0	0	0	0	0	0	0	0	1	2	2	1	0	0	2	0	24
4	2	1	2	2	2	2	3	S	4	4	4	4	3	2	2	1	1	1	1	1	2	1	1	2	1	1	4	2	24
5	2	2	2	2	2	2	S	4	2	2	1	1	1	1	1	1	1	1	1	1	1	3	2	3	1	4	2	2	24
6	3	2	2	1	1	S	3	2	2	2	2	3	3	4	4	4	3	2	3	3	3	3	3	2	1	4	3	24	
7	2	1	1	3	S	7	7	5	2	1	1	1	1	1	1	1	0	1	1	1	1	1	1	1	0	7	2	24	
8	1	1	2	S	2	2	1	1	2	2	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	2	1	24	
9	0	0	S	4	3	4	8	7	5	2	1	0	0	0	0	0	1	1	1	1	1	1	2	1	0	8	2	24	
10	1	S	1	1	1	2	4	4	4	4	3	C	C	C	C	C	C	C	C	2	1	1	1	0	0	4	2	24	
11	S	2	1	1	1	1	1	1	1	1	1	1	1	1	1	2	2	3	3	2	2	2	2	S	1	3	2	24	
12	3	3	3	3	3	P	P	P	P	P	P	P	R	1	2	2	2	1	0	0	0	0	S	1	0	3	2	16	
13	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	1	0	1	0	24	
14	1	0	1	1	1	1	5	4	3	2	1	1	3	3	1	2	2	3	6	7	S	6	3	2	0	7	3	24	
15	1	1	0	0	0	0	1	1	S1	1	1	1	1	1	0	0	1	1	1	S	2	1	1	0	0	2	1	23	
16	0	0	0	0	0	1	2	1	2	2	1	1	1	2	1	1	1	2	S	4	3	2	2	2	0	4	1	24	
17	3	3	3	6	5	4	3	2	2	P	3	3	2	2	2	3	10	S	10	5	5	3	3	4	2	10	4	23	
18	5	5	3	2	1	1	1	0	1	1	1	0	0	0	0	0	S	2	1	1	1	1	1	0	0	5	1	24	
19	0	0	0	0	0	0	0	1	1	1	0	1	0	0	0	S	2	1	2	2	2	2	2	2	0	2	1	24	
20	3	2	1	1	1	1	1	1	1	1	1	2	2	1	S	1	1	0	0	0	1	0	0	0	0	3	1	24	
21	0	1	1	1	1	0	0	0	0	0	0	0	0	0	S	1	1	0	0	0	0	0	0	0	0	1	0	24	
22	0	0	0	1	1	1	2	2	2	2	1	1	S	2	1	1	1	1	2	2	2	2	2	2	0	2	1	24	
23	2	2	2	2	3	3	3	3	3	3	2	1	S	3	2	2	3	5	5	4	4	4	4	4	2	5	3	24	
24	3	4	4	4	4	5	4	4	4	4	5	S	6	5	6	5	5	6	6	5	4	4	3	4	3	6	5	24	
25	4	4	4	4	4	3	3	4	4	S	4	3	3	3	3	2	3	3	4	4	4	4	2	2	2	4	3	24	
26	2	3	2	1	1	1	1	1	1	S	2	1	1	1	1	1	3	4	4	3	3	3	2	1	4	4	2	24	
27	3	2	2	1	1	1	1	S	3	2	1	1	1	1	0	1	0	0	0	1	1	1	0	1	0	3	1	24	
28	0	0	1	1	1	1	S	3	3	2	2	1	2	3	2	2	2	2	2	2	2	2	1	1	0	3	2	24	
29	1	1	1	2	2	S	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	24	
30	1	1	1	1	S	2	2	2	2	2	2	2	2	1	1	1	1	1	1	1	1	1	2	2	1	2	1	24	
31	2	2	2	S	6	5	5	4	4	4	4	4	4	7	8	5	6	6	7	7	5	3	2	2	2	2	8	5	24
HOURLY MAX	5	5	4	6	6	7	8	7	5	5	4	6	5	7	8	5	10	6	10	7	5	6	4	4					
HOURLY AVG	2	2	2	2	2	2	2	2	2	2	2	2	2	2	1	1	2	2	2	2	2	2	2	1					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 159 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0			
NUMBER OF NON-ZERO READINGS:	555			
MINIMUM 1-HR AVERAGE:	0	ppb	@ HOUR	1 ON DAY 2
MAXIMUM 1-HR AVERAGE:	10	ppb	@ HOUR	16 ON DAY 17
MAXIMUM 24-HR AVERAGE:	5	ppb		ON DAY 24
IZS CALIBRATION TIME:	32	hrs	OPERATIONAL TIME:	734 hrs
MONTHLY CALIBRATION TIME:	8	hrs	AMD OPERATION UPTIME:	98.7 %
STANDARD DEVIATION:	2		MONTHLY AVERAGE:	2 ppb

24 HR AVERAGES October 2018



NITROGEN DIOXIDE Hourly Averages (NO₂ ppb)



% Icon Classes (ppb)


88 0.0-3.7

12 3.7-7.3

1 7.3-11.0


0 >11.0

LICA ST. LINA Poll.: LICA ST. LINA-NO2[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 0.00%



NO2[ppb] Calibration: LICA ST. LINA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



OZONE

OZONE Hourly Averages (O₃ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	23.1	21.6	20.5	19.4	18.4	17.9	17.5	17.3	18.2	22.1	S	30.9	30.7	31.8	31.8	32.1	31.9	30.3	28.8	29.4	29.0	27.2	27.0	27.0	17.3	32.1	25.4	24
2	27.3	27.7	27.9	24.9	22.1	21.4	20.0	20.1	22.5	S	27.9	28.9	29.5	31.1	31.9	32.6	32.2	32.5	31.1	30.0	30.2	29.7	30.1	29.9	20.0	32.6	27.9	24
3	30.6	30.8	30.5	30.7	30.2	28.9	28.3	26.9	S	29.2	30.6	33.4	33.9	33.8	33.0	32.9	32.7	32.7	32.3	33.5	31.7	30.4	30.9	30.3	26.9	33.9	31.2	24
4	29.8	28.6	26.6	26.1	24.6	22.7	21.6	S	20.5	20.8	23.1	26.1	30.5	33.9	34.8	38.6	40.5	39.9	38.1	35.9	34.6	33.6	31.6	29.9	20.5	40.5	30.1	24
5	29.4	28.2	25.8	23.0	20.4	18.1	S	18.5	20.7	22.6	26.2	31.0	33.0	33.7	33.4	32.8	32.1	31.0	28.4	27.1	25.5	23.5	24.7	23.7	18.1	33.7	26.6	24
6	22.9	23.3	24.4	24.0	23.2	S	21.4	21.6	22.1	23.2	24.7	25.8	28.2	27.3	29.7	30.9	31.1	31.3	28.0	26.7	23.5	22.3	21.5	21.7	21.4	31.3	25.2	24
7	23.3	24.4	23.5	18.6	S	12.5	11.3	13.4	18.2	21.1	22.1	22.3	23.4	23.3	23.3	23.0	23.5	22.9	22.5	23.2	23.8	23.4	24.9	26.0	11.3	26.0	21.5	24
8	25.2	25.5	23.9	S	25.3	24.9	25.3	26.1	26.5	27.4	28.1	29.7	31.1	31.3	31.6	31.5	31.5	31.6	31.8	31.5	30.4	30.3	30.0	30.2	23.9	31.8	28.7	24
9	30.3	28.9	S	24.3	23.8	23.3	17.7	18.0	20.7	26.5	28.6	30.5	31.0	31.7	32.8	32.5	31.8	31.5	31.7	30.7	30.2	28.8	26.9	26.2	17.7	32.8	27.8	24
10	28.8	S	31.9	31.2	29.2	25.4	21.1	18.2	17.6	21.6	23.2	24.6	27.1	28.1	30.9	31.2	33.4	31.8	28.8	27.5	25.9	24.1	22.0	23.0	17.6	33.4	26.4	24
11	S	23.2	23.1	22.3	20.6	21.2	20.8	20.8	19.8	C	C	C	C	C	C	C	25.9	22.8	21.5	22.0	22.0	21.9	21.4	S	19.8	25.9	22.0	24
12	18.7	17.2	15.4	13.5	12.2	P	P	P	P	P	P	P	18.9	18.6	18.3	20.9	25.9	28.7	28.9	29.9	30.9	30.6	S	28.5	12.2	30.9	22.3	17
13	27.5	28.3	29.6	28.4	29.6	30.8	30.1	28.8	29.1	30.4	31.2	33.1	33.5	34.0	34.1	34.0	34.2	34.3	34.6	34.0	34.0	S	32.7	32.1	27.5	34.6	24	
14	32.2	31.0	28.2	26.6	25.8	22.2	14.9	15.7	17.9	20.7	21.2	20.8	20.0	23.5	27.4	27.0	26.3	24.9	21.2	19.5	S	22.9	28.0	31.8	14.9	32.2	24	
15	34.5	36.9	36.2	35.6	34.8	33.6	32.7	31.5	30.7	31.2	32.0	33.7	34.5	33.6	33.2	32.8	31.5	31.0	30.2	S	29.0	28.2	28.1	28.1	28.1	36.9	32.3	24
16	27.9	27.8	27.5	26.6	25.3	22.9	19.9	17.6	14.4	16.7	20.7	23.2	23.4	27.4	29.7	30.8	27.1	26.6	S	24.8	28.4	31.4	31.4	29.6	14.4	31.4	25.3	24
17	27.7	27.0	27.8	25.5	27.1	30.2	31.0	31.8	35.0	P	40.8	46.6	51.3	55.1	56.4	54.1	48.2	S	47.8	51.1	49.5	48.0	45.6	42.7	25.5	56.4	40.9	23
18	39.6	39.9	42.6	43.7	44.5	43.0	42.3	40.8	39.4	39.0	39.7	41.3	41.8	41.7	42.0	42.2	S	40.6	39.8	38.9	38.4	37.6	37.0	36.8	36.8	44.5	40.6	24
19	35.9	35.3	35.0	35.0	34.8	34.2	33.7	29.0	30.1	32.0	33.7	34.1	34.0	33.9	33.4	S	32.5	32.0	29.6	27.6	27.5	25.8	26.4	25.1	25.1	35.9	31.8	24
20	24.4	26.1	24.6	24.1	24.5	24.6	23.8	22.2	21.6	22.6	24.4	26.4	30.2	34.3	S	43.2	42.7	42.0	40.7	39.7	38.6	38.7	38.4	38.1	21.6	43.2	31.1	24
21	36.0	34.0	33.0	34.0	32.7	32.5	33.5	33.9	34.4	36.0	36.8	37.7	38.0	S	37.5	36.8	36.8	36.5	36.0	35.7	35.9	36.4	36.1	35.6	32.5	38.0	35.5	24
22	36.2	36.2	33.6	29.1	29.6	28.9	25.6	29.0	30.2	27.8	31.5	34.0	S	37.3	38.2	38.8	37.2	35.2	33.2	32.3	31.4	30.1	29.2	27.7	25.6	38.8	32.3	24
23	27.3	26.7	26.4	24.4	22.5	21.7	20.9	20.6	21.1	22.0	23.9	S	27.2	30.3	32.9	33.8	28.3	24.4	23.2	23.3	23.1	22.7	23.6	24.4	20.6	33.8	25.0	24
24	26.2	22.4	20.3	21.0	21.0	20.6	21.5	22.0	21.2	19.7	S	23.2	25.6	24.4	26.1	25.7	24.1	24.3	25.8	27.3	28.6	31.1	28.4	26.7	19.7	31.1	24.2	24
25	26.9	26.1	24.1	22.9	23.5	21.8	22.8	20.7	23.1	S	22.3	24.9	28.3	31.2	31.9	34.9	35.1	30.7	27.6	27.3	27.2	28.0	33.6	36.6	20.7	36.6	27.4	24
26	33.4	30.4	28.4	31.4	30.2	28.6	27.8	26.8	S	27.0	31.1	34.2	35.6	37.4	36.2	35.1	34.0	33.7	34.0	34.2	34.8	34.6	33.9	32.0	26.8	37.4	32.4	24
27	29.9	29.4	30.9	32.1	31.0	29.7	29.1	S	23.8	25.7	31.6	35.3	36.5	35.8	35.9	36.1	36.9	35.6	35.2	34.0	32.8	32.3	32.0	33.1	23.8	36.9	32.4	24
28	34.1	32.0	28.8	28.4	28.9	27.5	S	22.2	24.7	24.7	26.8	31.8	32.1	27.8	30.4	28.9	28.4	27.8	26.8	25.9	25.3	24.5	24.1	23.4	22.2	34.1	27.6	24
29	22.4	21.5	20.1	17.8	19.3	S	21.5	23.2	22.5	22.6	20.9	21.1	20.4	22.7	23.7	23.8	23.5	22.5	21.5	22.6	21.6	19.7	19.3	18.4	17.8	23.8	21.4	24
30	17.4	17.4	16.9	14.3	S	14.1	11.4	9.3	9.3	10.1	10.2	13.7	19.6	25.9	29.3	27.6	22.6	20.0	21.1	18.9	18.2	17.1	15.0	13.5	9.3	29.3	17.1	24
31	14.7	15.3	14.8	S	12.0	11.3	11.1	10.4	11.7	13.4	16.5	18.6	19.6	19.0	20.1	26.5	25.2	22.7	18.3	16.4	19.5	23.9	23.6	22.6	10.4	26.5	17.7	24
HOURLY MAX	39.6	39.9	42.6	43.7	44.5	43.0	42.3	40.8	39.4	39.0	40.8	46.6	51.3	55.1	56.4	54.1	48.2	42.0	47.8	51.1	49.5	48.0	45.6	42.7				
HOURLY AVG	28.1	27.4	26.7	26.2	25.8	24.8	23.5	22.7	23.1	24.5	27.0	29.2	30.0	31.0	32.1	32.8	31.6	30.4	30.0	29.4	29.4	28.6	28.6	28.5				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT: 1-HR 82 ppb


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0			
NUMBER OF NON-ZERO READINGS:	697			
MINIMUM 1-HR AVERAGE:	9.3 ppb	@ HOUR	7	ON DAY 30
MAXIMUM 1-HR AVERAGE:	56.4 ppb	@ HOUR	14	ON DAY 17
MAXIMUM 24-HR AVERAGE:	40.9 ppb			ON DAY 17
IZS CALIBRATION TIME:	32 hrs	OPERATIONAL TIME:	736 hrs	
MONTHLY CALIBRATION TIME:	7 hrs	AMD OPERATION UPTIME:	98.9 %	
STANDARD DEVIATION:	7.1	MONTHLY AVERAGE:	28.0 ppb	

24 HR AVERAGES October 2018




OZONE Hourly Averages (O₃ ppb)



% Icon Classes (ppb) 8 0.0-18.8 84 18.8-37.7 7 37.7-56.5 0 >56.5

LICA ST. LINA Poll.: LICA ST. LINA-O3[ppb] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 0.00%



O3[ppb] Calibration: LICA ST. LINA Monthly: 18/10 Type: Span

Span Meas Span Ref Span Low Span High



PARTICULATE MATTER 2.5



PARTICULATE MATTER < 2.5 MICRONS Hourly Averages (PM_{2.5} µg/m³)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY																													
1	2	2	3	3	3	3	3	3	3	3	3	2	2	2	2	2	2	3	3	4	4	4	3	3	2	4	3	24	
2	4	5	3	3	3	13	5	3	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	13	2	24
3	1	1	1	1	1	1	1	1	2	2	2	1	1	1	1	1	1	2	1	17	2	1	1	1	1	1	17	2	24
4	1	1	1	1	1	1	2	2	3	7	9	10	12	12	12	10	7	5	4	5	5	4	5	5	5	1	12	5	24
5	6	8	10	11	13	14	16	15	9	5	3	2	2	2	2	2	2	2	3	3	3	3	3	3	3	2	16	6	24
6	3	4	4	4	4	5	5	5	4	4	5	7	8	9	8	8	6	5	6	7	8	11	10	10	3	11	6	24	
7	11	11	10	8	7	8	7	5	3	1	1	2	3	3	4	3	4	3	2	2	1	2	2	1	1	11	4	24	
8	1	1	1	1	1	1	1	1	1	0	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0	1	1	24
9	1	1	1	1	1	2	2	2	2	1	2	1	1	1	1	2	1	2	10	2	2	2	2	2	2	1	10	2	24
10	2	2	2	1	1	2	3	5	5	3	3	3	3	3	3	3	3	3	5	7	6	6	6	8	1	8	4	24	
11	12	17	16	15	13	13	10	9	10	12	16	15	13	C	6	6	8	9	10	10	11	11	10	10	6	17	11	24	
12	9	4	2	2	5	P	P	P	P	P	P	P	3	3	3	3	3	2	1	1	0	1	1	0	0	9	3	17	
13	1	1	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0	1	1	1	1	1	1	1	0	1	0	24	
14	1	1	1	1	1	1	3	3	3	2	2	4	3	3	2	2	1	1	2	2	4	4	3	3	1	4	2	24	
15	3	3	4	3	3	3	3	3	3	4	2	1	1	1	2	2	2	2	1	1	1	1	1	1	1	1	4	2	24
16	1	1	2	3	4	3	2	2	2	2	2	3	3	3	2	2	2	3	3	3	5	6	8	8	1	8	3	24	
17	10	9	7	8	9	9	8	7	6	P	5	5	5	4	5	5	7	7	8	7	7	7	6	6	4	10	7	23	
18	7	8	7	7	9	12	13	11	11	9	7	8	9	8	6	4	2	3	3	3	3	3	3	3	2	13	7	24	
19	3	2	2	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	3	2	3	2	2	1	3	2	24	
20	2	2	3	3	3	3	3	3	3	2	3	4	4	4	2	1	1	1	1	2	2	2	2	2	1	4	2	24	
21	2	3	3	3	3	3	3	3	2	2	2	2	1	1	2	1	1	2	2	2	2	2	2	2	1	3	2	24	
22	2	2	2	3	3	5	7	5	4	4	3	3	2	2	3	3	3	4	4	4	4	3	3	4	2	7	3	24	
23	3	4	4	4	4	4	4	4	4	4	4	5	5	5	5	5	7	10	8	7	7	7	7	7	3	10	5	24	
24	7	9	10	9	9	9	12	13	13	10	10	10	10	10	10	10	10	10	9	9	9	9	10	10	7	13	10	24	
25	9	9	9	9	9	8	9	9	10	9	9	11	12	12	12	11	10	11	12	11	12	12	8	7	7	12	10	24	
26	7	8	7	5	5	5	3	2	2	1	1	1	1	1	1	2	2	2	3	3	3	3	3	1	8	3	24		
27	3	5	4	3	2	2	3	3	4	4	3	3	3	3	3	4	4	3	2	3	3	2	2	2	2	5	3	24	
28	2	2	2	2	3	4	5	6	6	5	6	4	3	4	4	4	3	3	4	4	4	4	5	5	2	6	4	24	
29	5	4	5	7	3	1	1	2	3	4	4	4	6	6	7	7	6	6	5	6	7	8	7	6	1	8	5	24	
30	7	9	9	10	8	4	2	3	4	7	10	9	6	4	3	3	4	6	6	6	6	6	4	6	5	2	10	6	24
31	4	4	4	5	7	6	6	6	7	7	7	7	6	6	5	4	3	3	4	5	5	4	5	5	3	7	5	24	
HOURLY MAX	12	17	16	15	13	14	16	15	13	12	16	15	13	12	12	11	10	11	12	17	12	12	10	10					
HOURLY AVG	4	5	5	4	5	5	5	5	4	4	4	4	4	4	4	4	3	4	4	5	4	4	4	4					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


OBJECTIVE LIMIT:

ALBERTA ENVIRONMENT:	1-HR	80	µg/m ³	24-HR	30	µg/m ³
----------------------	------	----	-------------------	-------	----	-------------------


MONTHLY SUMMARY

NUMBER OF 1-HR EXCEEDANCES:	0				
NUMBER OF 24-HR EXCEEDANCES:	0				
NUMBER OF NON-ZERO READINGS:	723				
MINIMUM 1-HR AVERAGE:	0 µg/m ³ @ HOUR	9	ON DAY	8	
MAXIMUM 1-HR AVERAGE:	17 µg/m ³ @ HOUR	19	ON DAY	3	
MAXIMUM 24-HR AVERAGE:	11 µg/m ³		ON DAY	11	
MONTHLY CALIBRATION TIME:	1	hrs	OPERATIONAL TIME:	736	hrs
STANDARD DEVIATION:	3		AMD OPERATION UPTIME:	98.9	%
			MONTHLY AVERAGE:	4	µg/m ³

24 HR AVERAGES October 2018



PARTICULATE MATTER < 2.5 MICRONS Hourly Averages (PM_{2.5} µg/m³)



Wind: LICA ST. LINA
 Poll.: LICA ST. LINA-PM25[ug/m3]
 Monthly: 18/10
 Type: PollutionRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.


Calm: 0.00%

Calm Avg: 0.00 [ug/m3]

Direction	0.0-3.6	3.6-7.2	7.2-10.8	10.8-14.4	14.4-18.0	>18.0	Total
N	5.3	0.6	0.7	0.3	0.0	0.0	6.8
NE	4.2	0.8	1.0	0.0	0.1	0.0	6.2
E	7.4	3.0	2.2	0.3	0.0	0.0	12.9
SE	4.9	1.5	1.0	0.4	0.0	0.0	7.8
S	4.9	5.1	1.6	1.0	0.3	0.0	12.9
SW	6.4	5.5	2.1	1.0	0.4	0.0	15.3
W	10.7	3.8	1.9	0.3	0.3	0.0	17.0
NW	13.4	3.4	3.3	1.0	0.1	0.0	21.2
Summary	57.3	23.7	13.7	4.1	1.2	0.0	100.0

% Icon	Classes (ug/m3(L))	57		0.0-3.6	24		3.6-7.2	14		7.2-10.8	4		10.8-14.4	1		14.4-18.0	0		>18.0
--------	--------------------	----	---	---------	----	---	---------	----	---	----------	---	---	-----------	---	---	-----------	---	---	-------

LICA ST. LINA Poll.: LICA ST. LINA-PM25[ug/m3(L)] 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 0.00%



WIND SPEED



WIND SPEED Hourly Averages (WS kph)


HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY																												
1	7.3	7.7	6.4	6.9	5.7	5.2	5.3	4.6	4.6	3.9	3.1	3.3	3.2	3.4	4.4	2.6	3.3	5.3	6.0	6.5	7.7	7.0	6.7	6.5	2.6	7.7	0.6	24
2	6.0	6.6	7.6	9.5	6.6	5.4	5.2	5.4	5.5	6.4	6.8	7.1	8.4	11.4	11.5	12.6	11.0	9.8	8.5	8.5	9.2	8.1	9.3	10.1	5.2	12.6	6.5	24
3	11.5	10.6	12.8	13.2	13.0	11.4	11.0	10.1	10.0	11.4	10.6	12.1	14.2	11.9	12.1	10.5	7.3	7.1	5.9	7.2	7.0	7.1	8.3	8.7	5.9	14.2	6.0	24
4	10.2	11.3	11.8	9.5	9.1	10.3	10.5	7.7	9.6	8.6	9.6	9.0	10.5	12.4	11.2	12.9	12.1	11.5	8.9	10.1	12.2	12.8	13.4	12.6	7.7	13.4	9.3	24
5	14.6	12.5	11.8	11.7	12.8	11.7	10.2	9.3	10.3	8.7	9.6	10.8	12.9	11.4	11.8	10.4	8.7	9.1	7.9	7.2	9.5	11.6	12.8	11.7	7.2	14.6	4.1	24
6	10.9	10.4	12.1	12.1	12.0	11.7	12.8	11.4	12.9	13.9	10.3	7.5	11.8	10.4	9.9	14.7	13.8	8.1	9.6	9.6	9.3	9.6	7.2	6.4	6.4	14.7	2.1	24
7	7.9	5.9	6.4	6.3	5.2	6.1	7.5	10.7	12.3	15.3	13.9	11.3	10.8	10.6	13.1	13.1	13.7	12.2	12.5	13.6	15.0	16.3	17.5	13.3	5.2	17.5	10.7	24
8	12.1	10.9	11.5	12.4	12.1	10.9	11.5	11.9	11.1	11.2	10.4	10.5	8.7	7.8	7.0	7.1	6.8	8.4	9.0	7.8	6.5	7.4	7.9	8.4	6.5	12.4	9.1	24
9	9.0	5.7	6.7	7.5	6.7	8.1	7.1	4.8	3.5	6.0	9.2	8.6	4.5	5.8	4.7	2.7	4.1	4.2	5.8	8.5	9.2	10.0	7.3	7.1	2.7	10.0	5.9	24
10	7.4	9.2	12.1	11.7	11.2	10.1	7.2	7.6	6.4	7.1	8.2	8.1	9.3	10.7	9.7	8.8	10.5	6.1	6.9	6.9	6.1	5.4	7.7	7.9	5.4	12.1	6.7	24
11	6.9	6.4	6.6	4.8	6.1	7.5	8.9	9.6	9.8	9.7	9.5	12.5	11.8	13.0	14.9	16.4	14.7	10.7	12.0	13.9	12.6	9.7	7.3	7.7	4.8	16.4	9.9	24
12	6.9	7.8	8.4	7.8	7.5	P	P	P	P	P	P	P	16.6	17.3	10.6	25.6	25.8	20.6	14.3	15.1	17.0	15.5	14.4	14.5	6.9	25.8	10.8	17
13	13.3	13.2	13.1	12.7	12.9	14.4	10.8	13.7	15.8	16.8	19.3	19.6	18.7	19.4	19.3	18.3	13.5	12.9	12.0	8.3	7.0	6.0	7.7	8.3	6.0	19.6	12.9	24
14	7.7	10.6	10.2	12.0	13.0	8.8	11.0	14.8	13.4	11.5	11.9	13.3	15.7	17.7	22.0	21.3	16.5	14.7	14.6	16.1	16.7	15.5	15.4	16.0	7.7	22.0	13.4	24
15	15.1	15.0	14.8	14.0	18.6	16.4	12.8	11.5	12.2	13.8	16.3	18.5	19.4	17.6	18.1	15.3	14.4	13.2	9.1	10.9	11.1	8.6	9.1	9.1	8.6	19.4	13.6	24
16	8.7	7.3	8.0	5.2	6.1	5.6	5.1	5.3	3.8	4.9	6.6	7.2	2.3	2.7	4.3	3.9	5.4	7.3	8.7	9.7	9.7	10.7	12.3	13.1	2.3	13.1	3.9	24
17	12.6	12.7	15.3	17.8	15.3	15.8	18.5	19.1	20.8	P	18.2	19.1	20.7	21.7	23.7	24.3	20.9	20.1	16.3	15.9	9.9	8.8	10.6	9.8	8.8	24.3	16.6	23
18	10.4	10.9	10.5	12.1	11.9	14.5	14.4	12.9	12.2	11.0	13.1	16.6	21.3	22.1	25.2	22.5	17.9	14.4	11.3	10.7	9.8	11.9	12.2	12.7	9.8	25.2	14.0	24
19	13.4	12.6	12.1	11.4	11.9	12.0	13.1	12.3	11.7	13.5	15.1	13.7	11.4	7.1	7.2	6.9	8.8	12.8	16.6	11.9	12.7	14.3	13.9	6.9	16.6	3.2	24	
20	16.5	17.5	16.5	18.2	21.7	21.1	22.8	19.0	14.2	15.0	18.0	13.3	12.0	14.6	18.8	20.0	19.5	18.7	12.6	10.3	16.3	14.9	11.8	10.0	10.0	22.8	10.4	24
21	12.4	15.9	16.2	15.2	12.8	12.6	14.3	15.3	14.9	14.6	16.8	17.8	18.3	21.8	19.7	16.4	13.4	9.8	9.7	9.8	9.5	11.0	9.1	7.5	7.5	21.8	13.1	24
22	6.8	7.7	4.0	7.7	8.6	9.7	10.4	13.6	11.6	13.2	13.5	11.0	13.1	15.5	14.7	16.0	12.3	13.0	12.8	13.1	12.5	13.0	12.9	13.5	4.0	16.0	10.1	24
23	14.6	14.8	13.1	14.1	13.8	12.7	12.7	14.0	14.7	12.4	10.0	10.4	11.5	10.2	9.7	11.1	10.9	14.0	13.9	12.8	11.6	12.3	10.3	9.7	9.7	14.8	11.8	24
24	10.4	9.4	9.7	10.2	10.0	10.0	10.1	8.3	5.8	5.7	6.6	6.6	6.0	7.3	6.5	4.8	5.2	4.7	4.9	7.9	9.0	6.0	2.8	8.7	2.8	10.4	6.1	24
25	9.0	8.1	7.1	9.6	6.7	8.6	11.0	8.4	7.5	5.0	7.6	9.0	7.2	7.0	7.1	5.8	4.3	6.5	6.2	7.0	5.9	6.3	8.4	5.6	4.3	11.0	4.3	24
26	10.3	11.6	15.6	14.9	11.0	9.7	12.1	10.7	11.8	12.3	12.0	13.7	17.4	16.5	9.6	11.9	5.2	6.3	4.9	6.2	8.5	10.0	10.3	10.8	4.9	17.4	10.0	24
27	13.9	11.5	11.4	12.2	11.6	12.2	12.3	13.0	13.0	15.3	18.2	20.0	18.4	16.2	15.0	16.7	14.9	17.7	13.1	14.4	10.0	8.9	7.9	7.4	7.4	20.0	12.9	24
28	8.6	6.4	8.1	7.5	7.5	8.0	10.2	10.5	12.3	11.1	11.0	12.7	14.8	12.5	13.5	13.7	14.3	11.9	10.2	13.7	13.2	12.2	13.1	14.3	6.4	14.8	8.1	24
29	11.3	9.8	5.7	3.6	X	X	X	X	19.1	12.0	12.3	13.6	13.5	13.6	14.9	15.8	13.8	10.7	9.2	8.8	10.2	9.2	9.1	6.9	3.6	19.1	5.4	20
30	3.7	6.3	5.7	1.1	0.8	5.9	5.4	4.6	6.8	8.5	8.8	8.6	9.1	9.9	11.4	8.8	11.3	13.0	10.6	10.0	9.8	10.0	8.8	9.2	0.8	13.0	7.3	24
31	8.2	9.0	8.2	8.8	11.1	10.1	10.1	10.6	10.7	8.7	7.5	6.5	8.0	7.3	6.2	5.4	1.5	6.1	7.9	7.4	8.1	6.0	5.6	5.3	1.5	11.1	4.0	24
HOURLY MAX	16.5	17.5	16.5	18.2	21.7	21.1	22.8	19.1	20.8	16.8	19.3	20.0	21.3	22.1	25.2	25.6	25.8	20.6	16.3	16.6	17.0	16.3	17.5	16.0				

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION:	May 25, 2017
DECLINATION :	MAGNETIC DECLINATION 19 DEGREE EAST


24 HR AVERAGES October 2018



MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	732
MINIMUM 1-HR AVERAGE	0.8 kph @ HOUR 4 ON DAY 30
MAXIMUM 1-HR AVERAGE:	25.8 kph @ HOUR 16 ON DAY 12
MAXIMUM 24-HR AVERAGE:	16.6 kph ON DAY 17
MONTHLY CALIBRATION TIME:	0 hrs
OPERATIONAL TIME:	732 hrs
AMSD OPERATION UPTIME:	98.4 %
STANDARD DEVIATION:	4.1
MONTHLY AVERAGE:	3.0 kph

WIND SPEED Hourly Averages (WS kph)




Wind: LICA ST. LINA
 Monitor: WSP [kph]
 Monthly: 18/10
 Type: WindRose
 Direction: Blowing From (Wind Frequency)
 Based On 1 Hr.

Calm: 0.00%

Direction	0.4-5.2	5.2-10.4	10.4-15.5	15.5-20.7	20.7-25.9	>25.9	Total
N	0.4	3.6	1.5	1.1	0.3	0.0	6.8
NE	0.4	3.3	2.2	0.3	0.0	0.0	6.2
E	1.2	7.0	4.5	0.1	0.0	0.0	12.9
SE	0.1	4.2	3.3	0.1	0.0	0.0	7.8
S	0.8	4.5	6.2	1.0	0.4	0.0	12.9
SW	1.4	7.5	4.5	1.4	0.7	0.0	15.4
W	0.4	4.6	9.2	2.3	0.4	0.0	16.9
NW	0.8	7.5	8.5	3.7	0.7	0.0	21.2
Summary	5.6	42.2	39.8	10.0	2.5	0.0	100.0

% Icon	Classes (kph)	6	 0.4-5.2	42	 5.2-10.4	40	 10.4-15.5	10	 15.5-20.7	2	 20.7-25.9	0	 >25.9
--------	---------------	---	---	----	--	----	---	----	---	---	---	---	---

LICA ST. LINA 2018/10/01 00:00 - 2018/10/31 23:00 Calm: 0.00%



WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - October 2018

WIND DIRECTION Hourly Averages (WD)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	24-HOUR AVG	24-HR	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	QUADRANT	RDGS.	
DAY 1	SSE	SSE	SSE	S	S	S	S	SSE	S	SW	SW	NE	NE	SSW	WSW	NNW	WNW	N	N	N	N	NNE	NNE	NNE	SE	24	
2	N	N	N	NNE	NNE	NNE	N	NNE	NNE	NNE	N	NW	WNW	WNW	WNW	NW	NW	WNW	WNW	WNW	WNW	WNW	W	W	NW	24	
3	WNW	WNW	W	W	W	W	W	W	WNW	NW	NW	NW	NW	NNW	NNW	NNW	N	NNE	NE	NE	E	ESE	ESE	ESE	NW	24	
4	ESE	SE	SE	SE	ESE	ESE	ESE	ESE	SE	SE	SE	SSE	SSE	S	S	S	S	S	S	S	SSW	SSW	SSW	SSW	SSE	24	
5	SSW	SW	SSW	SW	SW	WSW	W	WNW	NW	NW	NNW	NNW	NNW	NNW	NNW	N	N	N	NE	NE	ENE	ENE	E	NW	24		
6	E	E	E	E	E	E	E	E	ESE	SE	S	SW	WSW	WSW	W	W	NW	NNW	NNW	NNW	NNW	N	N	ENE	24		
7	NNW	NNE	NNE	NE	NE	NE	NE	ENE	ENE	E	E	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	ENE	NE	ENE	NE	ENE	24	
8	NE	ENE	NE	NE	NE	NE	NE	NE	ENE	ENE	ENE	ENE	E	ENE	ENE	ENE	ENE	ENE	ENE	ENE	E	ESE	ESE	ENE	ENE	24	
9	ESE	ESE	ENE	E	ENE	ENE	ENE	E	ENE	ESE	ESE	ESE	ESE	ESE	ENE	E	NE	NE	ENE	ENE	E	SE	SE	E	24		
10	SE	SE	SE	SE	SE	ESE	ESE	ESE	ESE	SE	SSE	SSE	SSE	SSE	S	SSE	SSE	S	SSW	SW	SW	SSW	SW	SW	SSE	24	
11	SW	SW	SW	SSW	SSW	S	S	S	SSW	SSW	SW	SW	SSW	SW	SW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	SSW	24	
12	SW	SW	SW	SW	WSW	P	P	P	P	P	P	P	N	N	N	N	N	N	NNW	NNW	NNW	NNW	NNW	NNW	NNW	17	
13	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NNW	NW	NNW	NNW	NNW	NNW	NNW	NW	W	WSW	W	NNW	24
14	W	WSW	WSW	WSW	W	SW	SW	SW	SW	SW	SSW	SSW	SSW	SSW	SSW	SW	SW	SW	SW	SW	WSW	WSW	W	W	SW	24	
15	W	WNW	WNW	WNW	WNW	NW	WNW	WNW	NW	NW	NW	NW	NW	WNW	WNW	NW	NW	WNW	WNW	WNW	WNW	NW	NW	NW	WNW	24	
16	NW	NW	NNW	NW	W	WSW	SW	WSW	SW	SSW	SW	SW	SW	NNW	NNW	WSW	E	ESE	SSE	S	SSW	SSW	SSW	SSW	SW	24	
17	SSW	SW	WSW	W	WSW	WSW	WSW	WSW	WSW	P	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	WSW	SW	SW	WSW	23	
18	W	W	W	WNW	WNW	W	W	WNW	NW	NW	WNW	WNW	WNW	WNW	NW	WNW	WNW	NW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	24	
19	WNW	WNW	WNW	WNW	WNW	W	WSW	WSW	W	WNW	NW	NNW	N	N	NNE	NNE	NE	ENE	E	E	ESE	ESE	SE	SSE	NW	24	
20	SSE	SSE	SSE	SSE	S	S	S	SSW	SSW	SW	WSW	WSW	WSW	W	W	WNW	WNW	WNW	NW	W	W	WNW	WNW	W	SW	24	
21	WSW	WSW	W	WSW	WSW	WSW	W	W	W	WNW	WNW	WNW	WNW	WNW	WNW	NW	NW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	24
22	NNW	NNE	SSW	SSW	SSW	SSW	S	S	S	S	S	S	S	S	SSE	SSE	SSE	SSE	SSE	SSE	SSE	S	S	S	S	24	
23	S	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SSE	SE	ESE	ESE	ESE	ESE	SE	SSE	SSE	SSE	SSE	SSE	24	
24	SSE	ESE	E	E	E	E	ESE	SE	E	ENE	E	ENE	E	E	E	E	E	ENE	ENE	E	ENE	ENE	NNW	NW	E	24	
25	NNW	N	WNW	NW	N	NW	NW	NNW	NW	WNW	NW	NW	NNW	NW	NW	NW	NW	ENE	ESE	SSE	SSW	SW	SW	SW	NW	24	
26	W	W	WNW	NW	WNW	W	W	W	W	W	W	WSW	WSW	WSW	WSW	SW	WSW	WSW	SW	WSW	WSW	WSW	WSW	W	W	24	
27	WNW	WNW	WNW	W	WSW	W	W	WSW	WSW	WSW	W	WNW	WNW	NW	NW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	WNW	24
28	WNW	WSW	SW	S	S	S	S	S	S	SSE	SSE	SSE	SSE	SE	ESE	ESE	E	ESE	ESE	ESE	E	E	E	E	E	24	
29	E	E	ESE	NNE	X	X	X	X	SE	SE	WNW	NW	WNW	WNW	WNW	WNW	NW	NW	WNW	WNW	WNW	WNW	NW	NW	NW	20	
30	WNW	WNW	W	SSW	NW	SW	SSW	SW	SW	WSW	W	W	W	W	W	WSW	W	W	W	WSW	WSW	WSW	WSW	WSW	WSW	24	
31	SW	SSW	SW	SW	SW	SW	SW	SW	SW	WSW	SW	SW	SW	SW	WSW	WSW	SE	ESE	E	ENE	NE	E	E	NE	SW	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION:	May 25, 2017
DECLINATION :	MAGNETIC DECLINATION 19 DEGREE EAST

MONTHLY CALIBRATION TIME:	0	hrs	OPERATIONAL TIME:	732	hrs
STANDARD DEVIATION:	95		AMD OPERATION UPTIME:	98.4	%
			MONTHLY AVERAGE:	264	(W)

WIND DIRECTION Hourly Averages (WD)



WDR[degwdr] Station: LICA ST. LINA Monthly: 18/10 Type: AVG 1 Hr. [1 Hr.]

— WDR[degwdr]



STANDARD DEVIATION WIND DIRECTION



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - October 2018

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59		
DAY																										
1	9	4	5	4	4	11	3	7	11	17	47	34	44	48	21	28	36	9	5	4	9	4	4	3	24	
2	7	7	3	10	13	14	8	7	12	14	22	22	29	20	24	26	14	8	8	5	8	4	6	10	24	
3	7	9	5	4	4	4	6	7	10	10	13	11	11	19	12	16	20	16	7	6	18	10	12	4	24	
4	9	4	4	5	5	3	4	9	9	9	15	14	17	15	21	13	9	5	8	5	3	3	3	7	24	
5	3	3	7	6	5	6	12	8	6	11	11	14	10	18	10	11	13	7	14	9	9	6	4	6	24	
6	5	7	6	6	5	7	7	6	7	10	16	29	21	13	14	8	11	17	6	5	3	8	9	21	24	
7	11	17	9	7	7	5	6	6	7	6	8	10	11	11	9	6	7	6	5	5	6	4	4	5	24	
8	5	9	6	5	5	5	5	5	6	10	9	10	15	19	17	16	16	9	16	9	10	11	16	10	24	
9	10	6	13	6	12	7	8	24	18	23	18	22	44	19	28	46	18	26	12	3	10	11	17	16	24	
10	16	10	6	2	4	7	12	6	10	14	10	16	18	16	21	19	14	36	8	11	11	5	10	6	24	
11	8	12	6	11	8	7	5	6	9	9	14	10	11	15	10	7	9	5	6	5	9	5	7	5	24	
12	4	7	10	7	13	P	P	P	P	P	P	P	5	16	48	6	5	6	9	4	4	4	4	5	17	
13	6	5	6	6	4	3	5	4	7	8	8	9	8	10	11	7	9	5	8	12	7	19	5	9	24	
14	12	5	8	7	5	13	3	6	4	5	11	10	8	8	8	5	4	6	3	5	7	4	6	3	24	
15	4	4	5	4	5	5	5	9	8	9	7	9	9	10	9	8	6	7	5	5	6	6	6	9	24	
16	6	10	4	16	9	5	8	8	12	15	14	12	37	42	40	49	14	12	11	13	4	3	4	4	24	
17	7	11	5	3	4	4	4	2	3	P	7	5	5	7	6	7	6	3	3	2	4	8	6	12	23	
18	18	13	12	11	6	5	10	10	13	13	11	10	11	11	6	7	6	6	6	6	6	3	6	3	24	
19	3	4	2	2	4	7	11	6	9	10	10	10	12	11	17	13	11	11	10	3	11	6	7	5	24	
20	6	5	4	7	5	3	4	10	5	6	11	11	14	19	22	8	7	8	7	6	7	6	6	13	24	
21	8	5	5	5	4	6	6	3	6	9	9	11	13	10	11	11	8	4	5	5	10	6	6	15	24	
22	27	6	46	6	8	6	3	2	8	3	8	13	9	11	12	7	7	3	4	3	3	5	3	4	24	
23	3	3	3	4	2	7	3	3	4	9	11	10	13	15	17	9	10	5	4	4	6	6	5	5	24	
24	5	10	9	2	3	6	8	6	20	14	13	13	16	16	13	16	8	5	5	4	12	43	8	24		
25	16	9	16	15	30	10	6	12	9	29	18	23	15	23	11	17	31	7	20	10	6	8	3	5	24	
26	9	8	9	17	8	5	7	9	7	8	10	11	10	11	12	4	15	13	13	8	5	9	5	18	24	
27	4	5	7	7	4	4	4	5	8	6	11	10	14	8	11	9	6	8	6	8	9	8	8	8	24	
28	14	35	14	7	8	8	8	6	5	5	10	14	10	14	12	7	5	7	8	6	9	10	4	4	24	
29	4	4	16	35	X	X	X	X	4	13	17	7	7	7	8	5	9	6	6	11	11	8	7	12	20	
30	20	12	16	67	72	4	13	19	11	8	11	9	8	10	9	15	12	3	3	5	7	8	6	24		
31	11	6	6	8	4	10	5	5	4	10	11	7	6	6	25	27	56	12	17	10	17	22	6	12	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

LAST CALIBRATION: May 25, 2017


CALIBRATION TIME: 0 hrs OPERATIONAL TIME: 732 hrs

STANDARD DEVIATION WIND DIRECTION Hourly Averages (STDWD deg)



RELATIVE HUMIDITY

RELATIVE HUMIDITY Hourly Averages (RH %)



BAROMETRIC PRESSURE




BAROMETRIC PRESSURE Hourly Averages (BP mbar)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.		
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.			
DAY																														
1	929	929	928	928	927	927	927	927	928	928	929	929	929	930	930	930	930	930	930	931	931	931	931	931	931	927	931	929	24	
2	931	931	931	931	930	931	930	930	931	931	930	930	929	929	928	927	926	925	925	924	924	924	924	924	924	924	924	931	928	24
3	924	924	924	924	924	925	925	926	927	928	929	929	930	930	930	931	931	931	931	930	930	930	931	931	931	924	931	928	24	
4	930	930	930	930	929	929	929	929	929	929	929	929	929	929	929	928	928	927	927	927	926	926	926	926	926	926	930	928	24	
5	925	925	925	925	925	925	925	925	926	927	928	929	929	929	929	929	929	929	929	929	929	929	928	928	928	925	929	927	24	
6	928	928	928	928	928	928	928	928	928	928	929	930	930	930	930	931	931	932	932	932	932	932	932	932	932	928	932	930	24	
7	933	933	934	934	934	934	935	935	935	936	936	936	937	937	937	937	936	936	936	936	936	937	937	937	933	937	936	24		
8	938	938	938	938	938	939	939	940	940	941	941	940	940	940	940	940	939	939	939	939	939	939	940	940	940	938	941	939	24	
9	940	940	940	940	940	941	941	942	943	943	943	944	944	944	943	943	942	942	942	941	941	941	940	940	940	940	944	942	24	
10	940	940	939	939	939	939	938	938	938	938	938	938	938	937	937	937	936	936	935	935	935	934	934	933	933	940	937	24		
11	933	932	932	932	931	931	930	930	930	929	929	928	927	926	925	924	923	922	921	921	920	919	919	919	919	933	927	24		
12	918	918	917	917	916	P	P	P	P	P	P	P	916	917	919	921	922	924	925	926	927	928	928	929	916	929	922	17		
13	930	930	930	931	931	932	932	933	934	935	935	936	936	936	937	937	937	937	937	937	937	938	937	937	937	930	938	935	24	
14	937	936	936	935	935	934	934	933	933	933	933	933	932	932	931	930	929	928	927	927	926	925	925	925	925	925	937	931	24	
15	925	926	926	927	927	928	929	929	929	930	931	931	931	931	932	932	932	932	933	933	934	934	935	935	925	935	931	24		
16	936	936	936	936	936	936	936	936	936	936	937	937	937	937	937	937	936	935	935	934	934	934	933	933	933	937	936	24		
17	932	932	931	931	931	931	931	931	931	P	932	933	933	932	931	930	929	928	928	927	927	926	926	926	926	926	933	930	23	
18	926	926	926	926	927	927	927	928	928	929	929	930	930	930	931	931	932	932	932	932	933	933	934	934	926	934	930	24		
19	935	935	936	936	936	936	937	937	937	938	940	941	941	941	941	941	940	939	938	938	937	936	936	935	935	941	938	24		
20	935	934	933	932	930	929	928	928	928	928	928	929	928	929	929	929	929	929	929	929	929	930	930	930	928	935	930	24		
21	930	930	930	930	930	930	931	931	932	933	933	933	933	934	934	934	935	935	935	935	935	935	935	935	930	935	933	24		
22	936	936	936	936	935	935	935	935	936	936	937	937	937	937	937	936	936	935	935	934	934	933	933	933	933	937	935	24		
23	932	932	931	931	930	930	929	929	929	929	930	930	930	929	929	928	928	927	926	926	925	925	925	925	925	925	932	929	24	
24	925	924	923	923	923	922	922	922	923	923	923	924	924	924	924	924	923	923	923	922	922	922	922	922	922	922	925	923	24	
25	922	921	922	921	921	921	922	922	923	923	924	924	924	924	924	924	924	923	923	923	922	922	922	922	921	924	923	24		
26	922	922	923	923	923	923	923	923	923	923	923	923	923	923	922	922	921	920	919	919	919	919	919	919	919	919	923	921	24	
27	919	919	920	920	920	920	921	921	922	923	924	924	924	925	925	925	926	926	927	927	927	927	927	927	919	928	923	24		
28	927	927	927	927	926	926	926	925	925	925	925	925	925	924	923	923	922	921	921	920	919	919	918	918	918	927	924	24		
29	917	916	916	916	916	917	918	919	920	921	921	922	922	923	923	924	924	924	925	925	925	925	925	916	925	921	24			
30	925	925	925	925	925	925	925	925	925	926	926	926	926	927	927	927	927	927	927	927	927	926	926	926	925	927	926	24		
31	925	925	925	924	923	923	922	921	921	922	922	922	921	921	921	920	920	920	920	920	920	920	921	920	920	920	925	922	24	
HOURLY MAX	940	940	940	940	940	940	941	941	942	943	943	943	944	944	944	944	943	943	942	942	941	941	941	940						
HOURLY AVG	929	929	929	929	929	929	929	929	930	930	930	931	930	930	930	930	930	930	930	929	929	929	929	929						

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR AVERAGES October 2018



MONTHLY SUMMARY


MINIMUM 1-HR AVERAGE:	916 mbar	@ HOUR	4	ON DAY	12
MAXIMUM 1-HR AVERAGE:	944 mbar	@ HOUR	12	ON DAY	9
MAXIMUM 24-HR AVERAGE:	942 mbar			ON DAY	9
OPERATIONAL TIME:					736 hrs
AMD OPERATION UPTIME:					98.9 %
STANDARD DEVIATION:	6	MONTHLY AVERAGE:			930 mbar

BAROMETRIC PRESSURE Hourly Averages (BP mbar)



AMBIENT TEMPERATURE

AMBIENT TEMPERATURE Hourly Averages (AmbTPX °C)



PRECIPITATION




PRECIPITATION Hourly Totals (mm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	SUM		
DAY																													
1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
6	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
7	0.0	0.0	0.0	0.0	0.5	0.6	0.2	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.3	0.1	0.1	0.0	0.1	0.5	0.3	0.1	0.0	0.0	0.6	3.0	24
8	0.0	0.2	0.6	0.4	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.4	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.6	2.7	24
9	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
10	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
11	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	1.2	0.5	0.0	1.2	2.2	24	
12	0.1	0.0	0.0	0.0	0.0	P	P	P	P	P	P	P	0.0	0.0	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.3	17
13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
14	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
15	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
17	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	P	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	23
18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
19	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
20	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
22	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
24	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
25	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
26	0.0	0.0	0.0	0.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	0.5	24
27	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
28	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.0	0.4	0.4	0.1	0.1	0.0	1.0	2.0	24
29	0.0	0.0	0.1	0.5	1.0	1.5	0.6	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.5	3.9	24	
30	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	24
31	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	1.1	0.5	0.6	0.0	1.1	2.4	24	
HOURLY MAX	0.1	0.2	0.6	0.5	1.0	1.5	0.6	0.2	0.1	0.1	0.1	0.4	0.2	0.0	0.2	0.0	0.3	0.1	0.1	1.0	0.4	1.1	1.2	0.6					
HOURLY SUM	0.1	0.2	0.7	1.4	1.7	2.3	1.0	0.3	0.2	0.1	0.1	0.4	0.2	0.0	0.2	0.0	0.3	0.1	0.1	1.0	0.7	2.5	2.1	1.3					

STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE


24 HR TOTALS October 2018



MONTHLY SUMMARY

MINIMUM 1-HR TOTAL:	0.0	mm	@ HOUR	0	ON DAY	1
MAXIMUM 1-HR TOTAL:	1.5	mm	@ HOUR	5	ON DAY	29
MAXIMUM 24-HR TOTAL:	3.9	mm			ON DAY	29
MONTHLY TOTAL	17.0	mm				
OPERATIONAL TIME:						736 hrs
AMD OPERATION UPTIME:						98.9 %
STANDARD DEVIATION:	0.1					mm
MONTHLY TOTAL:						

PRECIPITATION Hourly Totals (mm)



APPENDIX II
EQUIPMENT CALIBRATION RESULTS

SULPHUR DIOXIDE



Thermo 43I-TLE Sulphur Dioxide Analyzer Calibration

Date:	October 10, 2018	Barometer/B.P./units:	F.S. 05544 expires January 15, 2019	938	millibars
Company/Airshed:	LICA	Thermometer/Station Temp:	F.S. 170286131 expires April 19, 2019	22	°C
Location/Station Name:	St. Lina	Weather Conditions:	A few clouds		
Parameter:	Sulphur Dioxide	Calibration Purpose:	routine monthly		
Start Time 24 hr. (mst):	10:51	Performed By/Reviewer:	Alex Yakupov	Rob Fisher	
End Time 24 hr. (mst):	16:23	Cal Gas Expiry Date:	October 24, 2020		
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	n/a		
Analyzer:					
Serial Number/Owner:	1180930030 LICA	Range ppb:	1000		
Last Calibration Date:	September 12, 2018	As Found C.F.:	1.055		
Previous C.F.:	1.000	New C.F.:	1.001		

Calibration Standards: Low Flow Meter ID/Expiry Date: Defender Low 152019 expires December 13, 2018 High Flow Meter ID/Expiry Date: Defender High 148944 expires December 13, 2018 Calibrator ID/Expiry Date: API id# 690 expires March 15, 2019 Cal Gas Cylinder I.D. #: LL 104225 Cal Gas Conc. (ppm): 49.2	Standard Calibration Points for Ranges <table border="1"> <tr><td>Point</td><td>ppb</td></tr> <tr><td>High</td><td>780</td></tr> <tr><td>Mid</td><td>380</td></tr> <tr><td>Low</td><td>190</td></tr> </table>	Point	ppb	High	780	Mid	380	Low	190
Point	ppb								
High	780								
Mid	380								
Low	190								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration (ppb):	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total			
as found zero	4944	0.00	4944	0.0	0.1	n/a
as found high	4870	76.04	4946	756.4	717	1.055
adjusted zero	4944	0.00	4944	0.0	0.0	n/a
adjusted high	4870	76.04	4946	756.4	756	1.001
mid	4915	36.93	4952	366.9	366	1.002
low	4932	18.48	4950	183.7	182	1.009
calibrator zero	4946	0.00	4946	0.0	0	n/a
Average C.F. =						1.004

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	0.998		0.95-1.05
b (Intercept as % of full scale) =	0.19%		± 3% F.S.
% change in C.F. from last cal =	-5.51%		± 10%

Thermo 43I-TLE Sulphur Dioxide Analyzer Calibration




As found:		As left:	
Bkg:	3.32	Bkg:	3.80
Coef:	0.991	Coef:	1.042
Pmt:	-694.9	Pmt:	-696.3
Flash:	984	Flash:	984
Internal:	29.6	Internal:	28.9
Chamber:	45.3	Chamber:	44.9
Perm Oven Gas:	45.00	Perm Oven Gas:	45.00
Perm Oven Heater:	44.15	Perm Oven Heater:	44.14
Pressure:	675.3	Pressure:	674.7
Sample Flow:	0.443	Sample Flow:	0.443
Lamp Intensity:	0.90	Lamp Intensity:	0.90
Converter:	n/a	Converter:	n/a
Converter Set:	n/a	Converter Set:	n/a
Averaging Time:	120	Averaging Time:	120
Expected Value:	642.0	Expected Value:	635.0

Comments:

The analyzer sample inlet filter was changed. The analyzer cooling fan filter(s) were cleaned. The manifold blower was found to be working normally.

SO2[ppb] Station: LICA ST. LINA Daily: 18/10/10 Type: AVG 1 Min. [1 Min.]

— SO2[ppb]



HYDROGEN SULPHIDE



Thermo 450i Hydrogen Sulphide Analyzer Calibration

Date:	October 10, 2018	Barometer/B.P./units:	F.S. 05544 expires January 15, 2019	938	millibars
Company/Airshed:	LICA	Thermometer/Station Temp:	F.S. 170286131 expires April 19, 2019	22	°C
Location/Station Name:	St. Lina	Weather Conditions:	A few clouds		
Parameter:	Hydrogen Sulphide	Calibration Purpose:	routine monthly		
Start Time 24 hr. (mst):	10:51	Performed By/Reviewer:	Alex Yakupov	Rob Fisher	
End Time 24 hr. (mst):	17:20	Cal Gas Expiry Date:	October 20, 2020		
Calibration Method:	Gas Dilution	Converter Model & s/n (if applicable):	n/a		
Analyzer:					
Serial Number/Owner:	CM 18010058 LICA	Range ppb:	100		
Last Calibration Date:	September 11, 2018	As Found C.F.:	1.005		
Previous C.F.:	0.995	New C.F.:	0.999		

Calibration Standards: Low Flow Meter ID/Expiry Date: Defender Low 152019 expires December 13, 2018 High Flow Meter ID/Expiry Date: Defender High 148944 expires December 13, 2018 Calibrator ID/Expiry Date: Sabio id# 11900613 expires August 22, 2019 Cal Gas Cylinder I.D. #: EY 0001003 Cal Gas Conc. (ppm): 9.55	Standard Calibration Points for Ranges <table border="1"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19	SO2 Scrubber Check (10 minutes): Start/End Time 24 hr.: 11:06 / 11:22 SO2 Analyzer Range: 1000 Target Concentration (ppb): 780 As Found Zero: 1.5 Analyzer Response: (ppb): 1.5 Zero Corrected Result (ppb): 0.0
Point	ppb									
High	78									
Mid	38									
Low	19									


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration (ppb):	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total			
as found zero	7556	0.00	7556	0.0	1.5	n/a
as found high	7496	61.54	7558	77.8	78.9	1.005
adjusted zero	7556	0.00	7556	0.0	0.0	n/a
adjusted high	7496	61.54	7558	77.8	77.8	0.999
mid	7546	29.89	7576	37.7	37.7	0.999
low	7554	15.00	7569	18.9	18.9	1.001
calibrator zero	7556	0.00	7556	0.0	0.0	n/a
Average C.F. =						1.000

Linear Regression/Calibration Results:

Correlation Coefficient =	1.000	LIMITS	> or = 0.995
Slope =	0.999		0.95-1.05
b (Intercept as % of full scale) =	0.03%		± 3% F.S.
% change in C.F. from last cal =	-0.97%		± 10%

Thermo 450i Hydrogen Sulphide Analyzer Calibration




As found: Bkg: 27.2 Coef: 0.887 Pmt: -634.2 Flash: 904 Internal: 32.4 Chamber: 45.3 Converter Temp: 324.4 Converter Set: 325.0 Perm Oven Gas: 45.00 Perm Oven Htr: 44.08 Pressure: 589.9 Sample Flow: 0.834 Lamp Intensity: 90 Averaging Time: 120 Expected Value: 31.5	As left: Bkg: 29.4 Coef: 0.898 Pmt: -634.2 Flash: 904 Internal: 31.2 Chamber: 45.2 Converter Temp: 323.3 Converter Set: 325.0 Perm Oven Gas: 45.00 Perm Oven Htr: 44.06 Pressure: 586.9 Sample Flow: 0.834 Lamp Intensity: 90 Averaging Time: 120 Expected Value: 30.4
---	--

Comments: The analyzer sample inlet filter was changed. The analyzer cooling fan filter(s) were cleaned. The manifold blower was found to be working normally.

Station: LICA ST. LINA Daily: 18/10/11 Type: AVG 1 Min. [1 Min.]

CH4[ppm] NMHC[ppm]





Thermo 450i Hydrogen Sulphide Analyzer Calibration

Date: <u>October 28, 2018</u>	Barometer/B.P./units: <u>F.S. 05544 expires January 15, 2019</u> <u>924</u> <u>millibars</u>	Thermometer/Station Temp: <u>F.S. 170286131 expires April 19, 2019</u> <u>22</u> <u>°C</u>
Company/Airshed: <u>LICA</u>	Weather Conditions: <u>A few clouds</u>	Calibration Purpose: <u>repeat</u>
Location/Station Name: <u>St. Lina</u>	Parameter: <u>Hydrogen Sulphide</u>	Performed By/Reviewer: <u>Alex Yakupov</u> <u>Rob Fisher</u>
Start Time 24 hr. (mst): <u>13:37</u>	End Time 24 hr. (mst): <u>18:44</u>	Cal Gas Expiry Date: <u>October 20, 2020</u>
Calibration Method: <u>Gas Dilution</u>	Converter Model & s/n (if applicable): <u>n/a</u>	
Analyzer: Serial Number/Owner: <u>CM 18010058</u> <u>LICA</u>	Range ppb: <u>100</u>	As Found C.F.: <u>0.936</u>
Last Calibration Date: <u>October 10, 2018</u>	New C.F.: <u>1.000</u>	Previous C.F.: <u>0.999</u>

Calibration Standards: Low Flow Meter ID/Expiry Date: <u>Defender Low 152019 expires December 13, 2018</u> High Flow Meter ID/Expiry Date: <u>Defender High 148944 expires December 13, 2018</u> Calibrator ID/Expiry Date: <u>Sabio id# 11900613 expires August 22, 2019</u> Cal Gas Cylinder I.D. #: <u>EY 0001003</u> Cal Gas Conc. (ppm): <u>9.55</u>	Standard Calibration Points for Ranges <table border="1" style="margin: auto;"> <tr><th>Point</th><th>ppb</th></tr> <tr><td>High</td><td>78</td></tr> <tr><td>Mid</td><td>38</td></tr> <tr><td>Low</td><td>19</td></tr> </table>	Point	ppb	High	78	Mid	38	Low	19
Point	ppb								
High	78								
Mid	38								
Low	19								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated Concentration (ppb):	Indicated Concentration (ppb):	Correction Factors (C.F.):
Point	Diluent	Cal Gas	Total			
as found zero	7520	0.00	7520	0.0	1.5	n/a
as found high	7450	60.88	7511	77.4	84.2	0.936
adjusted zero	7520	0.00	7520	0.0	0	n/a
adjusted high	7460	61.14	7521	77.6	77.6	1.000
mid	7448	29.80	7478	38.1	37.9	1.004
low	7471	14.90	7486	19.0	18.9	1.006
calibrator zero	7520	0.00	7520	0.0	0	n/a
Average C.F. =						1.003

Linear Regression/Calibration Results:

Correlation Coefficient = <u>1.000</u>	LIMITS <u>> or = 0.995</u>
Slope = <u>1.000</u>	<u>0.95-1.05</u>
b (Intercept as % of full scale) = <u>0.07%</u>	<u>± 3% F.S.</u>
% change in C.F. from last cal = <u>6.31%</u>	<u>± 10%</u>

Thermo 450i Hydrogen Sulphide Analyzer Calibration




As found: Bkg: <u>29.6</u> Coef: <u>0.898</u> Pmt: <u>-633.8</u> Flash: <u>905</u> Internal: <u>32.7</u> Chamber: <u>45.0</u> Converter Temp: <u>329.3</u> Converter Set: <u>325.0</u> Perm Oven Gas: <u>45.00</u> Perm Oven Htr: <u>44.09</u> Pressure: <u>580.3</u> Sample Flow: <u>0.823</u> Lamp Intensity: <u>90</u> Averaging Time: <u>120</u> Expected Value: <u>30.4</u>	As left: Bkg: <u>30.3</u> Coef: <u>0.867</u> Pmt: <u>-634.2</u> Flash: <u>903</u> Internal: <u>32.5</u> Chamber: <u>44.9</u> Converter Temp: <u>326.0</u> Converter Set: <u>325.0</u> Perm Oven Gas: <u>45.00</u> Perm Oven Htr: <u>44.07</u> Pressure: <u>577.0</u> Sample Flow: <u>0.821</u> Lamp Intensity: <u>90</u> Averaging Time: <u>120</u> Expected Value: <u>30.4</u>
--	---

Comments: The SO2 scrubber check was not performed, see comments below.
The manifold blower was found to be working normally.

A repeat calibration was completed to correct for a Zero and a Span drift.

H2S[ppb]



TOTAL HYDROCARBON



Thermo 55i Methane/Non-Methane Analyzer Calibration

Date:	October 11, 2018	Barometer/B.P./units:	F.S. 05544 expires January 15, 2019	929	millibars
Company/Airshed:	LICA	Thermometer/Station Temp:	F.S. 170286131 expires April 19, 2019	22	°C
Location/Station Name:	St. Lina	Weather Conditions:	A few clouds		
Parameter:	CH4 / NMHC / THC	Calibration Purpose:	routine monthly		
Start/End Time 24 hr. (mst):	9:42 / 13:22	Performed By/Reviewer:	Alex Yakupov	Rob Fisher	
Calibration Method:	Gas Dilution	Cal Gas Expiry Date:	October 18, 2025		

Analyzer:		Correction Factors:			
Serial Number/Owner:	1180930025 LICA	Previous C.F.:	As Found C.F.:	New C.F.:	
Measured Flow:	1245	CH ₄ =	1.005	0.999	1.000
Last Calibration Date:	September 11, 2018	NMHC =	0.999	0.990	1.000
Range ppm:	20 CH4/20 NMHC/40 THC	THC =	1.002	0.994	1.000

Calibration Standards:

Low Flow Meter ID/Expiry Date:	Defender Low 152019 expires December 13, 2018	Standard Calibration Points for Analyzer Range of 20/20/40 ppm			
High Flow Meter ID/Expiry Date:	Defender High 148944 expires December 13, 2018	Point	CH4	NMHC	THC
Calibrator ID/Expiry Date:	API id# 690 expires March 15, 2019	High	13.00	13.00	26.00
Cal Gas Cylinder I.D. #:	LL 119471	Mid	7.00	7.00	14.00
CH4 Cylinder Conc. =	599.0 207.0 = C ₂ H ₆ Cylinder Conc.	Low	3.00	3.00	6.00
CH ₄ expressed as C ₂ H ₆ =	569.3 1168.3 = total CH4 equivalent				

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rates (cc/min)			Calculated CH ₄ (ppm)	Calculated NMHC (ppm)	Calculated THC (ppm)	Indicated CH ₄ (ppm)	Indicated NMHC (ppm)	Indicated THC (ppm)	Correction Factors:		
	Diluent	Cal Gas	Total Flow							CH ₄	NMHC	THC
as found zero	2962	0.00	2962	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
as found high	2960	67.90	3028	13.43	12.76	26.20	13.45	12.90	26.35	0.999	0.990	0.994
adjusted zero	2962	0.00	2962	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
adjusted high	2960	67.90	3028	13.43	12.76	26.20	13.43	12.76	26.19	1.000	1.000	1.000
mid	2960	36.72	2997	7.34	6.97	14.31	7.42	7.01	14.44	0.989	0.995	0.991
low	2962	15.95	2978	3.21	3.05	6.26	3.24	3.08	6.32	0.990	0.990	0.990
calibrator zero	2962	0.00	2962	0.00	0.00	0.00	0.00	0.00	0.00	n/a	n/a	n/a
Average C.F. =										0.993	0.995	0.994

Linear Regression/Calibration Results:

Correlation Coefficient =	CH ₄	NMHC	THC	LIMITS
Slope =	1.000	1.000	1.000	> or = 0.995
b (Intercept as % of full scale) =	0.14%	0.10%	0.12%	0.95-1.05
% change in C.F. from last cal =	0.63%	0.95%	0.78%	± 3% F.S.
				± 10%




As Left Instrument Diagnostics:

Interface Board Voltages:	Bias Supply: -295.9	Calibration History cnt'd:	NM Peak Area: 82132
Temperatures:	Detector Oven: 175.0	Crucial Settings:	Methane Start: n/a
	Filter: 175.0		Methane End: n/a
	Column Oven: 75.0		Backflush: n/a
Cylinder Pressures/reg.:	Internal: 27.1	Run History>1:	NMHV Start: n/a
	Carrier: 1200 55		NMHC End: n/a
	Fuel: 1700 56		Date: Oct 11, 2018
Internal Pressures:	Span Gas: 1000 10	Time: 10:16	CH ₄ PK HT: 0
	Zero Air Generator: 44	CH ₄ PK RT: 8.0	CH ₄ Baseline: 3773
	Carrier: 32.0	CH ₄ LOD: 37	CH ₄ SD: 12
FID Status:	Fuel: 48.1	CH ₄ CONC: 0.00	NM PK HT: 0
	Air: 36.2	NM Peak Area: 0	NM CONC: 0.00
	Status: LIT	NM Base Start: 3720	NM Base End: 3774
Flame and Power Stats:	Counts: 41600	NM LOD: 51	NM Start IDX: 4
	Flame: 405.0	NM End IDX: 98	NM Max Slope: 2.2e+00
	Det Base: 175.1	NM Min Slope: -7.5e-01	NM PT Count: 0
Calibration History:	Det Base: 175.1	Previous CH ₄ : 10.28	Previous NMHC: 11.15
	Last Power On: Aug 23, 2018	Previous THC: 21.35	New CH ₄ : 10.30
	Flameouts: 2	New NMHC: 11.01	New THC: 21.31
Calibration History:	Det Oven at Start: 43.5	Expected Values:	
	Col Oven at Start: 24.8	CH ₄ SP Ratio: 0.000715	CH ₄ RT: 13.4
	Time: Sep 11, 2018 / 13:58	CH ₄ PK IDX: 27	CH ₄ PK HT: 18608
Calibration History:	Type: SPAN	NM Span Conc: 12.65	NM SP Ratio: 0.000154
	Status: GOOD		
	Check/Adjust: ADJUST		


Comments:
 The analyzer sample inlet filter was changed.
 No zero adjustment was required/made. As found zero values were copied to adjusted zero values for linearity calculation purposes.
 The analyzer cooling fan filter(s) were cleaned.
 The manifold blower was found to be working normally.

Date: October 11, 2018
Company/Airshed: LICA
Location/Station Name: St. Lina

Start/End Time 24 hr. (mst): 9:42 / 13:22
Calibration Purpose: routine monthly
Calibration Method: Gas Dilution



CH4[ppm] NMHC[ppm]



NITROGEN DIOXIDE



Thermo 42i NO-NO2-NOx Analyzer Calibration

Date: <u>October 10, 2018</u>	Barometer/B.P./units: <u>F.S. 05544 expires January 15, 2019</u>	<u>938</u>	millibars
Company/Airshed: <u>LICA</u>	Thermometer/Station Temp: <u>F.S. 170286131 expires April 19, 2019</u>	<u>22</u>	°C
Location/Station Name: <u>St. Lina</u>	Weather Conditions: <u>A few clouds</u>		
Start/End Time 24 hr. (mst): <u>10:51 / 18:37</u>	Calibration Purpose: <u>routine monthly</u>		
G.P.T. to be used for Ozone?: <u>No</u>	Performed By/Reviewer: <u>Alex Yakupov</u>	<u>Rob Fisher</u>	
Calibration Method: <u>Gas Dilution & Gas Phase Titration</u>	Cal Gas Expiry Date: <u>October 24, 2020</u>		

Analyzer: Serial Number/Owner: <u>1180930029</u> <u>LICA</u> Last Calibration Date: <u>September 12, 2018</u> Range ppb: <u>1000</u>	Correction Factors: <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th>Previous C.F.:</th> <th>As Found C.F.:</th> <th>New C.F.:</th> </tr> </thead> <tbody> <tr> <td>NO =</td> <td>1.000</td> <td>0.994</td> <td>1.000</td> </tr> <tr> <td>NO₂ =</td> <td>1.000</td> <td>1.000</td> <td>1.000</td> </tr> <tr> <td>NOx =</td> <td>1.000</td> <td>0.994</td> <td>1.000</td> </tr> </tbody> </table>		Previous C.F.:	As Found C.F.:	New C.F.:	NO =	1.000	0.994	1.000	NO ₂ =	1.000	1.000	1.000	NOx =	1.000	0.994	1.000
	Previous C.F.:	As Found C.F.:	New C.F.:														
NO =	1.000	0.994	1.000														
NO ₂ =	1.000	1.000	1.000														
NOx =	1.000	0.994	1.000														

Calibration Standards: Low Flow Meter ID/Expiry Date: <u>Defender Low 152019 expires December 13, 2018</u> High Flow Meter ID/Expiry Date: <u>Defender High 148944 expires December 13, 2018</u> Calibrator ID/Expiry Date: <u>API id# 690 expires March 15, 2019</u> Cal Gas Cylinder I.D. #: <u>LL 104225</u> Cal Gas Conc. (ppm): <u>51.5</u> <u>51.6</u>	Standard Calibration Points for a Range of: <u>1000 ppb</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Point</th> <th>Target NO (ppb)</th> <th>Target NO₂ (ppb)</th> <th>Cc Ozone ?</th> </tr> </thead> <tbody> <tr> <td>High</td> <td>780</td> <td>500</td> <td>n/a</td> </tr> <tr> <td>Mid</td> <td>380</td> <td>275</td> <td>n/a</td> </tr> <tr> <td>Low</td> <td>190</td> <td>100</td> <td>n/a</td> </tr> <tr> <td>Extra Point #1</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> <tr> <td>Extra Point #2</td> <td>n/a</td> <td>n/a</td> <td>n/a</td> </tr> </tbody> </table>	Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?	High	780	500	n/a	Mid	380	275	n/a	Low	190	100	n/a	Extra Point #1	n/a	n/a	n/a	Extra Point #2	n/a	n/a	n/a
Point	Target NO (ppb)	Target NO ₂ (ppb)	Cc Ozone ?																						
High	780	500	n/a																						
Mid	380	275	n/a																						
Low	190	100	n/a																						
Extra Point #1	n/a	n/a	n/a																						
Extra Point #2	n/a	n/a	n/a																						

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calculated NO	Calculated NOx	Indicated NO	Indicated NOx	NO C.F.	NOx C.F.
Point	Diluent	Cal Gas	Total Flow	(ppb)	(ppb)	(ppb)	(ppb)		
as found zero	4944	0.0	4944	0	0	0.1	0.1	n/a	n/a
as found high	4870	76.0	4946	791.8	793.3	797.0	798.0	0.994	0.994
adjusted zero	4944	0.00	4944	0.0	0.0	0.0	0.0	n/a	n/a
adjusted high	4870	76.04	4946	791.8	793.3	792.0	793.0	1.000	1.000
mid	4915	36.93	4952	384.1	384.8	385.0	386.0	0.998	0.997
low	4932	18.48	4950	192.3	192.6	193.0	194.0	0.996	0.993
calibrator zero	4944	0.00	4944	0	0	0.0	0.0	n/a	n/a
Average C.F.=								0.998	0.997

ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Calibrator Flow Rates (cc/min)				Calibrator Setting	Indicated NO	Indicated NOx	Indicated NO ₂	NO drop	NO ₂ gain	NO ₂ C.F.
Point	Diluent	Cal Gas	Total Flow	volts or ppb	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)	(ppb)
NOx reference	4870	76.04	4946	0.0	793.0	795.0	1.0	0.0	1.0	
as found high NO2	4870	76.04	4946	500.0	290.0	794.0	504.0	503.0	503.0	1.000
adjusted high NO2	4870	76.04	4946	500.0	290.0	795.0	504.0	503.0	503.0	1.000
gpt mid	4870	76.04	4946	275.0	520.0	794.0	274.0	273.0	273.0	1.000
gpt low	4870	76.04	4946	105.0	687.0	794.0	107.0	106.0	106.0	1.000
Average NO ₂ C.F.=										1.000

Linear Regression/Calibration Results:

	NO	NOx	NO ₂	LIMITS
Correlation Coefficient =	1.000	1.000	1.000	> or = 0.995
Slope =	1.000	1.001	1.002	0.95-1.05
b (Intercept as % of full scale)=	0.04%	0.08%	0.06%	± 3% F.S.
% change in C.F. from last cal=	0.64%	0.58%	0.00%	± 10%
NO2 converter efficiency			1.00	0.96 to 1.04

As found:		As left:	
NO Bkg:	5.2	NO Bkg:	5.2
NOx Bkg:	5.3	NOx Bkg:	5.3
NO Coef:	1.155	NO Coef:	1.048
NO2 Coef:	1.000	NO2 Coef:	0.995
NOx Coef:	1.001	NOx Coef:	1.000
PMT:	-824.0	PMT:	-824.0
Internal:	27.4	Internal:	26.6
Chamber:	50.1	Chamber:	50.6
Cooler:	-3.0	Cooler:	-2.7
NO2 Converter:	325.3	NO2 Converter:	326.3
NO2 Converter Set:	325.0	NO2 Converter Set:	325.0
Perm Oven Gas:	45.01	Perm Oven Gas:	45.01
Perm Oven Heater:	44.16	Perm Oven Heater:	44.15
Pressure:	256.3	Pressure:	256.6
Flow:	0.538	Flow:	0.536
Ozonator Flow:	OK	Ozonator Flow:	OK
Expected Value NO:	5	Expected Value NO:	5
Expected Value NO2:	458	Expected Value NO2:	455
Expected Value NOx:	464	Expected Value NOx:	459

Comments:

The converter cooling fan filter was cleaned.

The analyzer sample inlet filter was changed.




The manifold blower was found to be working normally.

The analyzer cooling fan filter(s) were cleaned.

Date: October 10, 2018
Company/Airshed: LICA
Location/Station Name: St. Lina


Start/End Time 24 hr. (mst): 10:51 / 18:37
Calibration Purpose: routine monthly
Calibration Method: Gas Dilution & Gas Phase Titration

Thermo 42i NO-NO2-NOx Analyzer Calibration



Station: LICA ST. LINA Daily: 18/10/10 Type: AVG 1 Min. [1 Min.]

— NOX[ppb] — NO[ppb] — NO2[ppb]



OZONE



Thermo 49i Ozone Analyzer Calibration

Date: October 11, 2018 Company/Airshed: LICA Location/Station Name: St. Lina Start/End Time 24 hr. (mst): 9:43 / 15:33 Ozone Calibration Method: Varying UV Lamp Power G.P.T. Date: n/a-done by Varying UV Lamp Power Analyzer: Serial Number/Owner: 1002240371 LICA Last Calibration Date: September 12, 2018 Previous Cal High Point C.F.: 1.000	Barometer/B.P./units: F.S. 05544 expires January 15, 2019 929 millibars Thermometer/Station Temp: F.S. 170286131 expires April 19, 2019 22 °C Weather Conditions: A few clouds Calibration Purpose: routine monthly Performed By/Reviewer: Alex Yakupov Rob Fisher Cal Gas Expiry Date: n/a-done by Varying UV Lamp Power Ozone Range ppb: 500 As Found C.F.: 0.989 New C.F.: 1.000
---	--

Calibration Standards:									
Low Flow Meter ID/Expiry Date: Defender Low 152019 expires December 13, 2018 High Flow Meter ID/Expiry Date: Defender High 148944 expires December 13, 2018 Calibrator ID/Expiry Date: Sabio id# 11900613 expires August 22, 2019 Cal Gas Cylinder I.D. #: n/a	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th>Point</th> <th>AMD Required Range of Ozone Calibration Points</th> </tr> <tr> <td>High</td> <td>300-400 ppb</td> </tr> <tr> <td>Mid</td> <td>150-200 ppb</td> </tr> <tr> <td>Low</td> <td>50-75 ppb</td> </tr> </table>	Point	AMD Required Range of Ozone Calibration Points	High	300-400 ppb	Mid	150-200 ppb	Low	50-75 ppb
Point	AMD Required Range of Ozone Calibration Points								
High	300-400 ppb								
Mid	150-200 ppb								
Low	50-75 ppb								


ALL POINTS ARE 15 MINUTES OF STABILITY AS OF SEPTEMBER 23, 2015

Point	Calibrator Flow Rate (cc/min)		Calculated Concentration:	Corrected Calculated Concentration:	Indicated Concentration:	Correction Factors:
	Total Flow @ Point Start	Total Flow @ Point Finish	(ppb)	(ppb)	(ppb)	
as found zero	5000	5000	0.0	n/a	-0.3	n/a
as found high	5000	5000	380.0	380.0	384.0	0.989
adjusted zero	5000	5000	0.0	0.0	0.0	n/a
adjusted high	5000	5000	380.0	380.0	380.0	1.000
mid	5000	5000	181.0	181.0	181.0	1.000
low	5000	5000	60.0	60.0	60.0	1.000
calibrator zero	5000	5000	0.0	n/a	0.0	n/a
Average C.F.=						1.000

Linear Regression/Calibration Results:

Correlation Coefficient = 1.000 Slope = 1.000 b (Intercept as % of full scale)= 0.00% % change in C.F. from last cal= 1.12%	LIMITS > or = 0.995 0.95-1.05 ± 3% F.S. ± 10%
--	--

Thermo 49i Ozone Analyzer Calibration



As found:

O3 Bkg:	-0.3
O3 Coef:	1.019
Photo Lamp:	10.7
O3 Lamp:	8.2
Bench:	27.4
Bench Lamp:	53.6
O3 Lamp:	67.7
Pressure:	680.6
Cell A lpm:	0.731
Cell B lpm:	0.776
O3 ppb:	-2.8
Cell A ppb:	2.1
Cell B ppb:	-7.6
Cell A int (Hz):	74393
Cell B int (Hz):	93638
Expected Value:	369.0


As left:

O3 Bkg:	-0.4
O3 Coef:	1.010
Photo Lamp:	10.7
O3 Lamp:	8.2
Bench:	26.7
Bench Lamp:	53.6
O3 Lamp:	67.7
Pressure:	678.5
Cell A lpm:	0.730
Cell B lpm:	0.780
O3 ppb:	0.4
Cell A ppb:	0.0
Cell B ppb:	0.4
Cell A int (Hz):	74415
Cell B int (Hz):	93667
Expected Value:	366.0

Comments:

The analyzer sample inlet filter was changed.
 The analyzer cooling fan filter(s) were cleaned.
 The manifold blower was found to be working normally.

O3[ppb]



PARTICULATE MATTER



Thermo 5030i SHARP Monitor Monthly Audit

Date: October 11, 2018
Company: LICA
Station Name/Location: St Lina
Previous Audit Date: September 19, 2018
Parameter: PM 2.5

Performed By/Reviewer: Alex Yakupov | Robert Fisher
Start Time (mst): 12:56
End Time (mst): 13:52
Calibration Purpose: Monthly
Weather Conditions: A few clouds

SHARP 5030i Information and Status:

Serial Number: CM17091001 **Filter Tape Counter** 461 / 0

Reference Standards:

Air Flow

	Manometer	Orifice	Pressure:	Temp / RH:
Make:	Dwyer	Chinook	Fisher Scientific	Fisher Scientific
Model:	475 Mk.III	CHN0901	FB 1291	11-661-7A, 11745843
Serial Number:	#3	#2	130168457 / 05544	170286131
Calibration Date:	January 9, 2019	April 24, 2019	January 15, 2019	April 19, 2019

Ambient Temperature (°C)

Reference	SHARP	Difference	Range	Action
#1	2.62	2.6	0.0	< ± 2°C OK
				2-3 °C Recalibrate
				> 3°C Fail

Ambient Relative Humidity (%RH)

As Found:	Reference	SHARP	Difference	Range	Action
#1	69.90	70.4	-0.5	< ± 2 %RH OK	
				2-5 %RH Recalibrate	
				> 5 %RH Fail	

Barometric Pressure (mmHg)

As Found:	Reference	SHARP	Difference	Range	Action
#1	695.3	697.5	-2.2	< ± 10 mmHg OK	
				10-12 mmHg Recalibrate	
				> 12 %RH Fail	


Flow Audit (L/min)

As Found:	Reference	SHARP	% Difference	Range	Action
#1	16.74	16.79	0.199840128	< ± 4% OK	
#2	16.62	16.64		4-5% Recalibrate	
#3	16.68	16.71		>5% Fail	
Average	16.68	16.71			

Leak Check (L/min)

#1	Without Leak Check Adapter			With leak Check Adapter			Leak Limit: 0.08 L/min
	Reference	SHARP	Difference	Reference	SHARP	Difference	
#1	16.68	16.71	-0.03	16.62	16.64	-0.02	
							LEAK RATE: 0.01

WIND SYSTEM



Sonic Wind Sensor Certificate of Calibration

Sensor Model No.: 50.5H
 Sensor Output Swing: 0V - 1.0V
 Customer: MAXXAM Analytics
 Tested per PO: 35-67600
 Calibrated by: David Frith *DF*

Sensor Serial No.: H12635
 Sensor Output Range: 0 - 50.0 MPS
 Sales Order No.: 122618
 Calibration Date: 05/25/2017

QC Inspection *Chris Paul*

Instrument Condition Within Tolerance: As Found As Left
 Corrective Action: No Adjustment Adjust Repair
 Preventative Maintenance

As Found Test Date: N/A As Left Test Date: 05/25/2017

Quality Control Manual Revision: September 16, 2013 MP42201 Rev. G.
 All Work Performed per Customer Purchase Order Requirements.
 Calibration Document No. 50.5-6100

Test Equipment Used for Calibration of Instruments

Description	Manufacturer	Model No.	Serial No.	Cal Date	Cal Due	Voltage Accuracy	Time Base Accuracy
Data Acquisition	Campbell Scientific	CR1000	6569	4/06/2015	4/06/2018	+/- 3mV	< 6 ppm
NIST Cupset	Met One Instruments	170-41	3309	1/26/2017	1/26/2022	Accuracy < 0.15 mph or 1% WS	

Environmental Data: Temperature 65 to 80 Deg F Vibration none
 Humidity 20 to 70% Radiation none

Firmware Version: 3194-01 R2.62

The standards used for calibration have accuracies equal to or greater than the instruments tested. These standards are on record and are traceable to NIST to the extent allowed by the institute's calibration facility. Unless otherwise stated heron, all instruments are calibrated to meet the manufacturer's published specifications. The calibration system complies with MIL-STD-45662A (8/1/88). Instrument's accuracy meets the requirements of Regulatory Guide 1.23 (2/72). Compliant with IS) 9001:2008 requirements

METEOROLOGICAL SYSTEM CHECK



Meteorological System Checklist

Date:	October 11, 2018		
Technician:	Alex Yakupov		
Reviewer:	Rob Fisher		
Station:	St. Lina		
Unit:	Make:	Model:	Serial #:
Precipitation Sampler:	Met One - Heated Rain Gauge	Part 387	n/a
PRECIPITATION SENSOR CHECK			
Checklist:	Reply:	Comments:	
Previous check date:	June 26, 2018	n/a	
Is the sensor Level?	yes	n/a	
Is the heater operating properly?	yes	n/a	
Are the bucket drain holes clean?	yes	cleaned	
Is the screen on the housing? (screen should be on between July and September)	no	The screen has been removed for Winter season	
Is the housing clean?	yes	cleaned	
Is the area around the housing clean and free from obstacles?	yes	n/a	
TIP TEST - Slowly pour water until 10 tip are heard. (10 tips = 1 ml)			
# of Tips	Data Logger Response (ml):	Manual Specification = +/- 0.1 ml	
10	1.00	0.00	

CALIBRATORS

Company: Maxxam Operator: Chris W

Calibrator:		Flow Measurement Device:	
Make/Model	<u>API 700</u>	Make/Model	<u>Mesa Defender 530</u>
Serial Number	<u>690</u>	Serial Number	<u>L-153351 H-152571</u>
Last Verification Date	<u>March 2016</u>	Temperature (°C)	<u>23.5 C</u>
NO Cylinder S/N	<u>LL108015</u>	Barometric Pressure	<u>695 mmHg</u>
NO [PPM]	<u>52.2</u>	NOx [PPM]	<u>52.3</u>
Expiry Date	<u>Oct 2020</u>		

Dilution Flow (sccm)			
Pt. #1	<u>5000</u>	Pt. #2	<u>5000</u>
Pt. #3	<u>5000</u>		
Gas Flow (sccm)			
Pt. #1	<u>80</u>	Pt. #2	<u>40</u>
Pt. #3	<u>20</u>		

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
5000	0.0	0.000	0.000	0.000	0.000	0.000	Limit ± 10%	
4959	75.0	0.789	0.791	0.793	0.000	0.793	1%	0%
4971	36.5	0.383	0.384	0.384	0.000	0.384	0%	0%
4967	18.2	0.191	0.192	0.191	0.000	0.191	0%	-1%
Absolute Average Percent Difference							0%	0%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO	LIMITS	NOx
Correlation= 1.0000	≥ 0.990	Correlation= 1.0000
m (Slope)= 1.0054	0.90-1.10	m (Slope)= 1.0031
b (Intercept % of FS)= -0.0583	± 3% F.S.	b (Intercept % of FS)= -0.0795

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
4959	0.000	0.000	0.790	-0.001	0.789	NO ₂	% Diff. Limit
4959	0.500	0.497	0.293	0.493	0.786	-1%	± 10%
4959	0.275	0.273	0.517	0.269	0.787	-1%	± 10%
4959	0.100	0.102	0.688	0.099	0.787	-2%	± 10%
Absolute Average Percent Difference						1%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

NO₂	LIMITS
Correlation= 1.0000	≥ 0.995
m (Slope)= 0.9946	0.90-1.10
b (Intercept % of FS)= -0.1817	± 3% F.S.

AENV Standards	NO_x Analyzer
Audit Calibrator	Make/Model <u>Teco 42i</u>
Make/Model <u>Teco 146i</u>	Serial/AMU Number <u>AMU 1868</u>
Serial/AMU Number <u>AMU 1809</u>	Last Calibration Date <u>March 14, 2018</u>
SRM Gas Cylinder No. <u>APEX1170572</u>	Full Scale (ppm) <u>1.0</u>
Cylinder Conc. (ppm) <u>49.99</u>	Cylinder Gas Expiry Date <u>November 2020</u>

COMMENTS: Cylinder contains 47.9 ppm SO2.

Auditor: Al Clark
Operator Signature: *Al Clark*

Date: March 15, 2018
Location: McIntyre Center Edmonton

Company Maxxam Operator: Mike

Calibrator:		Flow Measurement Device:	
Make/Model	<u>Sabio</u>	Make/Model	<u>Bios Definer 220</u>
Serial Number	<u>11900613</u>	Serial Number	<u>H=128686; L=129069</u>
Last Verification Date	<u>March 16, 2018</u>	Temperature (°C)	<u>22.9 C</u>
NO Cylinder S/N	<u>LL104183</u>	Barometric Pressure	<u>698 mmHg</u>
NO [PPM]	<u>50.8</u>	NOx [PPM]	<u>50.9</u>
Expiry Date	<u>October 24, 2020</u>		

Dilution Flow (sccm)
Pt. #1 5059 Pt. #2 5073 Pt. #3 5073

Gas Flow (sccm)
Pt. #1 77.5 Pt. #2 38.2 Pt. #3 19.1

Calibrator Flow (sccm)		Calculated Conc.(ppm)		Indicated Conc.(ppm)			% Difference vs Audit Gas	
Dilution	Gas	NO	NOx	NO	NO ₂	NOx	NO	NOx
5124	0.0	0.0000	0.0000	0.0000	-0.0001	0.0000	Limit ± 10%	
5059	77.5	0.7782	0.7797	0.7763	0.0005	0.7767	0%	0%
5073	38.2	0.3825	0.3833	0.3794	0.0000	0.3795	-1%	-1%
5073	19.1	0.1913	0.1916	0.1904	0.0000	0.1904	0%	-1%
Absolute Average Percent Difference							1%	1%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

<u>NO</u>		<u>LIMITS</u>		<u>NOx</u>	
Correlation=	1.0000	≥ 0.990		Correlation=	1.0000
m (Slope)=	0.9975	0.90-1.10		m (Slope)=	0.9960
b (Intercept % of FS)=	-0.0616	± 3% F.S.		b (Intercept % of FS)=	-0.0661

Flow	O ₃ Conc	NO Decrease	NO	NO ₂	NOX	% Diff. Vs Audit gas	
5059	0.0	0.0000	0.7741	0.0000	0.7741	NO ₂	% Diff. Limit
5059	500.0	0.4918	0.2823	0.4916	0.7739	0%	± 10%
5059	275.0	0.2774	0.4967	0.2780	0.7747	0%	± 10%
5059	100.0	0.1031	0.6710	0.1032	0.7743	0%	± 10%
Absolute Average Percent Difference						0%	± 10%

LINEAR REGRESSION ANALYSIS *y=mx+b (where x=calculated concentration, y=indicated concentration)*

<u>NO₂</u>		<u>LIMITS</u>	
Correlation=	1.0000	≥ 0.995	
m (Slope)=	0.9998	0.90-1.10	
b (Intercept % of FS)=	0.0173	± 3% F.S.	

AENV Standards		NO_x Analyzer	
Audit Calibrator		Make/Model	<u>Thermo 42i</u>
Make/Model	<u>Thermo 146i</u>	Serial/AMU Number	<u>1868</u>
Serial/AMU Number	<u>1809</u>	Last Calibration Date	<u>August 16, 2018</u>
SRM Gas Cylinder No.	<u>APEX1170572</u>	Full Scale (ppm)	<u>1.0</u>
Cylinder Conc. (ppm)	<u>49.99</u>	Cylinder Gas Expiry Date	<u>November 15, 2020</u>

COMMENTS:

Auditor: Shea Beaton
Operator Signature: [Signature]

Date: August 22, 2018
Location: McIntyre Center Edmonton

CALIBRATION GASES



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2017-482CGA

Company: Maxxam **Operator's Name:** Mike
Cylinder #: LL104225 **Concentration PPM:** 49.2 **Tolerance(%)** 2 **Certified By:** Praxair
Expiry Date: October 2020

Reference Calibrator and Gas:	Flow Measurement Device:
Make/Model: <u>R&R MFC 201</u>	Make/Model: <u>Mesa Definer 220</u>
Serial Number: <u>AMU 1690</u>	Serial Number: <u>H-133034 / L-132702</u>
Last Verification Date: <u>December 13, 2017</u>	Temp. °C: <u>23.4 C</u>
Gas Type: <u>SO2</u> Conc. <u>98.07</u>	B.P.: <u>707 mmHg</u>
Cylinder Number: <u>CAL016625</u>	
Expiry Date: <u>January 2019</u>	

Reference Analyzer:
 Make/Model: Teco 43C Serial/AMU Number: 1623
 Instrument Settings: Zero: 10.0 Span: 1.006 Range: 1.0
 Last Calibration: Date: Dec12/17 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.000	0.000	0.000	0.000
4989	79.5	0.764	0.01594	62.755	47.9
4995	39.6	0.380	0.00793	126.136	47.9
4992	19.6	0.188	0.00393	254.694	47.9
Average Cylinder Concentration:					47.9


Previous Stated Concentration PPM: 49.2

Percent variance from Stated: 3

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:** _____
 < =5% Outside Manufacturer Tolerance. Use manufacturers concentration _____
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder _____

Auditor: Al Clark
 Operator Signature: *Al Clark*

Date: December 13, 2017
 Location: McIntyre Center Edmonton



Calibration Gas Audit

Single Component Cylinder Gas

File No. 2017-493CGA

Company: Maxxam Operator's Name: Mike
 Cylinder #: EY0001003 Concentration PPM: 9.55 Tolerance(%): 2 Certified By: Praxair
 Expiry Date: October 2020

Reference Calibrator and Gas:
 Make/Model: Sabio 2010
 Serial Number: AMU 2092
 Last Verification Date: January 17, 2018
 Gas Type: H2S Conc. 20.43
 Cylinder Number: CAL015272
 Expiry Date: January 2019

Flow Measurement Device:
 Make/Model: Mesa Defender 530
 Serial Number: H-153961 / L-153874
 Temp. °C: 23.0 C
 B.P.: 697 mmHg

Reference Analyzer:
 Make/Model: Teco 450i Serial/AMU Number: 1980
 Instrument Settings: Zero: 12.9 Span: 0.955 Range: 0.1
 Last Calibration: Date: Jan 17/18 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Concentration (PPM)	Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration
Dilution	Gas				
5000	0.0	0.0000	0.0000	0.0000	0.0000
5051	39.6	0.0753	0.00784	127.551	9.60
5028	20.2	0.0387	0.00402	248.911	9.63
5033	10.5	0.0198	0.00209	479.333	9.49
Average Cylinder Concentration:					9.58

Previous Stated Concentration PPM: 9.55

Percent variance from Stated: 0

Meets Manufacturer Tolerance. Use manufacturers stated concentration COMMENTS: Used AEP regulator
 <=5% Outside Manufacturer Tolerance. Use manufacturers concentration
 > 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark

Date: January 18, 2018

Operator Signature: *Al Clark*

Location: McIntyre Center Edmonton



Calibration Gas Audit

NO Cylinder Gas

File No. 2017-483CGA

Company: Maxxam **Operators name:** Mike

Cylinder #: LL104225 Conc (PPM) 51.5/51.6 Tolerance (%) 2 Certified By: Praxair

Expiry Date: October 2020

Reference Calibrator and Gas:				Flow Measurement Device:	
Make/Model	<u>Teco 146i</u>			Make/Model	<u>Mesa Definer 220</u>
Serial Number	<u>AMU 1809</u>			Serial Number	<u>H-133034 / L-132702</u>
Last Verification Date	<u>December 13, 2017</u>			Temp. °C	<u>23.4 C</u>
Gas Type	<u>NO</u>	Conc.	<u>50.03</u>	B.P.	<u>707 mmHg</u>
Cylinder Number	<u>APEX 1223938</u>				
Expiry Date	<u>June 2020</u>				

Reference Analyzer:

Make/Model Teco 42i Serial/AMU Number: 1868

Instrument Settings Zero: 4.7 Span: 1.004 Range: 1.0

Last Calibration: Date: Dec12/17 C.F. 1.000 Done By: Al Clark

Calibrator Flows (sccm)		Indicated Conc. (ppm)		Gas Flow/ Dilution Flow	Concentration Factor	Cylinder Concentration	
Dilution	Gas	NO	NOX			NO	NOX
5000	0.0	0.000	0.000				
4989	79.5	0.813	0.812	0.016	62.755	51.0	51.0
4995	39.6	0.407	0.406	0.008	126.136	51.3	51.2
4992	19.6	0.202	0.201	0.004	254.694	51.4	51.2
Average Cylinder Concentration:						51.3	51.1

<u>NO</u>	<u>NOx</u>
Previous Stated Concentration PPM: <u>51.5</u>	<u>51.6</u>
Percent variance from Stated: <u>0</u>	<u>1</u>

Cylinder gas tolerances based on NO only

Meets Manufacturer Tolerance. Use manufacturers stated concentration **COMMENTS:**

<=5% Outside Manufacturer Tolerance. Use manufacturers concentration

> 5% Outside Manufacturer Tolerance. **DO NOT USE** this cylinder

Auditor: Al Clark Date: December 13, 2017

Operator Signature: *Al Clark* Location: McIntyre Center Edmonton

APPENDIX III
MAXIMUM INSTANTANEOUS DATA



LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - October 2018

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY 1	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	1	1	1	0	0	0	1	0	24	
2	0	0	0	0	0	0	0	0	0	1	S	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
3	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
4	0	0	0	0	0	0	0	0	S	0	0	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0	1	0	24
5	0	1	1	1	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	1	0	24
6	1	1	1	1	1	S	0	0	0	1	1	1	2	1	1	1	1	0	0	1	0	0	0	0	0	0	2	1	24
7	0	0	0	0	S	1	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
8	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
9	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
10	0	S	0	0	0	0	0	0	0	1	0	C	C	C	C	C	C	C	0	0	0	0	0	0	0	0	1	0	24
11	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	24
12	0	0	0	0	0	P	P	P	P	P	P	P	P	X	0	0	0	0	0	0	0	0	0	S	0	0	0	0	15
13	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	24
14	0	0	0	0	0	0	0	0	1	0	0	1	2	2	0	0	0	0	0	2	2	S	1	0	0	0	2	1	24
15	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	24
16	0	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	S	0	0	0	0	1	0	1	0	24
17	0	0	1	1	1	0	0	0	1	P	P	1	0	0	1	2	4	S	6	3	1	1	1	1	1	0	6	1	22
18	2	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	2	0	24
19	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	24
20	0	0	0	0	0	0	0	0	0	0	0	1	0	0	S	0	0	0	0	0	0	0	0	0	0	0	1	0	24
21	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	24
22	0	0	0	0	0	0	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	24
23	0	0	0	0	0	0	0	0	0	0	2	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	2	0	24
24	0	0	0	0	0	0	1	1	1	0	S	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	1	0	24
25	0	0	0	0	0	0	1	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24
26	0	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	1	0	24
27	0	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
28	0	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
29	0	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
30	0	0	0	0	S	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24
31	0	0	0	S	1	0	0	0	0	0	0	0	0	1	2	1	1	3	3	2	1	0	0	0	0	0	3	1	24
HOURLY MAX	2	1	1	1	1	1	1	1	1	1	2	1	2	2	2	2	4	3	6	3	1	1	1	1	1	0	3	1	24
HOURLY AVG	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	24


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	78
MAXIMUM INSTANTANEOUS VALUE:	6 ppb @ HOUR 18 ON DAY 17
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	7 hrs
OPERATIONAL TIME:	733 hrs
STANDARD DEVIATION:	0

SULPHUR DIOXIDE Instantaneous Maximum (SO₂ ppb)





HYDROGEN SULPHIDE Instantaneous Maximum (H₂S ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY 1	2	1	1	2	2	1	1	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	1	2	2	24	
2	2	2	2	2	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	24	
3	2	2	2	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	24	
4	2	2	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	24	
5	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	24	
6	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	24	
7	2	3	2	2	S	3	2	2	2	2	2	2	2	3	3	3	3	2	2	2	2	2	2	2	2	3	2	24	
8	2	2	2	S	2	3	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	3	2	24	
9	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	3	2	2	2	2	2	2	2	3	2	24	
10	2	S	2	2	2	2	3	2	2	2	C	C	C	C	C	C	C	C	1	1	0	1	1	1	0	3	2	24	
11	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	S	1	1	1	24	
12	1	1	1	1	1	P	P	P	P	P	P	P	X	1	1	1	1	1	1	1	1	1	S	1	1	1	1	15	
13	1	1	1	1	1	1	1	1	1	0	0	1	1	1	0	0	0	0	0	1	1	S	1	1	0	1	1	24	
14	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	S	1	1	1	1	1	1	24	
15	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	S	1	1	2	2	1	2	1	24	
16	1	1	2	1	2	2	2	1	2	2	2	2	2	2	2	2	2	2	S	2	2	2	2	2	2	1	2	2	24
17	2	2	1	2	1	1	1	1	1	P	P	1	2	2	1	1	1	S	1	1	1	1	1	1	1	2	1	22	
18	2	1	1	1	1	1	1	2	1	1	1	1	2	1	1	1	S	2	2	2	2	1	2	2	1	1	2	24	
19	2	1	2	1	1	2	2	2	1	2	1	2	2	2	2	S	2	2	2	2	2	2	2	2	2	1	2	2	24
20	2	2	2	2	2	2	2	2	2	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	24	
21	1	1	1	1	2	2	2	2	2	1	2	1	2	S	2	1	2	1	2	2	2	2	1	2	2	1	2	24	
22	2	2	2	2	2	2	2	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	24	
23	2	2	2	2	2	2	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	24	
24	2	2	2	2	2	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	24	
25	2	2	2	2	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	24	
26	2	2	2	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	24	
27	2	2	2	2	2	2	S	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	24	
28	2	2	2	2	2	S	2	2	2	2	2	2	C1	C1	C1	C1	C1	C1	1	1	1	1	1	1	1	1	2	18	
29	1	1	1	1	1	S	1	0	0	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0	1	1	24	
30	1	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24	
31	1	1	1	S	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	24	
HOURLY MAX	2	3	2	2	2	3	3	2	2	2	2	2	2	3	3	3	3	3	2	2	2	2	2	2	2	2	2	24	
HOURLY AVG	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	24	


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	677
MAXIMUM INSTANTANEOUS VALUE:	3 ppb @ HOUR 1 ON DAY 7
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	8 hrs
OPERATIONAL TIME:	727 hrs
STANDARD DEVIATION:	1

HYDROGEN SULPHIDE Instantaneous Maximum (H₂S ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - October 2018

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59				
DAY 1	2.20	2.33	2.33	2.50	2.50	2.45	2.47	2.47	2.47	2.36	S	2.05	2.04	2.04	2.03	2.09	3.19	2.06	2.03	2.02	2.03	2.04	2.04	2.04	2.02	3.19	2.25	24
2	2.04	2.05	2.03	2.03	2.01	2.02	2.02	2.00	1.99	S	2.00	X	2.01	2.04	2.02	2.05	2.04	2.16	2.17	2.04	2.09	2.00	1.99	1.98	1.98	2.17	2.03	23
3	2.45	2.10	1.98	2.09	2.00	1.99	1.98	1.99	S	2.02	2.07	2.27	2.19	1.97	1.98	1.99	2.00	1.96	1.98	2.01	2.16	2.13	2.21	2.13	1.96	2.45	2.07	24
4	2.12	2.11	2.27	2.15	2.49	2.17	2.30	S	2.32	2.34	2.40	2.37	2.36	2.20	2.15	2.06	1.99	1.99	2.00	2.02	2.03	2.02	2.05	2.06	1.99	2.49	2.17	24
5	2.05	2.04	2.05	2.06	2.07	2.08	S	2.16	2.34	3.35	3.22	3.32	2.13	2.15	2.17	2.31	2.53	2.04	1.98	2.01	2.03	2.09	2.09	2.26	1.98	3.35	2.28	24
6	2.31	2.25	2.28	2.36	2.26	S	2.37	2.30	2.22	2.17	2.21	2.33	2.30	2.19	2.10	2.04	2.03	2.27	2.11	2.11	2.32	2.14	2.20	3.08	2.03	3.08	2.26	24
7	2.76	2.16	2.13	2.18	S	2.20	2.20	2.18	2.31	2.09	2.09	2.08	2.09	2.09	2.05	2.11	2.04	2.04	2.03	2.02	2.00	2.00	1.99	2.01	1.99	2.76	2.12	24
8	2.00	2.02	2.01	S	2.01	2.00	1.99	2.02	2.00	2.14	2.03	2.12	2.11	2.34	2.15	2.15	2.13	2.13	2.15	2.27	2.13	2.21	2.21	2.05	1.99	2.34	2.10	24
9	2.07	2.72	S	2.18	2.16	2.09	2.26	2.49	2.45	2.34	2.16	2.17	2.09	2.24	2.12	2.01	2.21	2.25	2.03	2.02	2.02	2.11	2.12	2.10	2.01	2.72	2.19	24
10	2.06	S	2.04	2.05	2.07	2.37	3.44	2.65	2.38	2.36	2.31	2.20	2.13	2.05	2.08	2.04	2.06	2.04	2.03	2.03	2.04	2.04	2.04	2.00	2.00	3.44	2.22	24
11	S	1.99	1.99	2.00	2.01	2.02	2.01	2.00	2.02	C	C	C	C	C	C	2.03	2.03	2.12	2.06	2.07	2.04	2.06	2.06	S	1.99	2.12	2.03	24
12	2.09	2.06	2.12	2.31	2.11	P	P	P	P	P	P	P	P	X	2.10	2.10	2.10	2.11	2.24	2.11	2.03	2.06	S	2.14	2.03	2.31	2.12	15
13	2.08	2.06	2.09	2.19	2.07	2.06	2.22	2.11	2.06	2.62	2.06	2.05	2.02	2.03	2.03	2.04	2.07	2.04	2.16	2.70	2.56	S	2.04	2.09	2.02	2.70	2.15	24
14	2.03	2.04	2.14	2.07	2.07	2.11	2.13	2.17	2.10	2.06	2.05	2.04	2.03	2.04	2.03	2.04	2.04	2.11	2.05	S	2.04	2.03	2.02	2.02	2.17	2.06	24	
15	2.11	1.99	2.01	2.13	2.64	2.02	1.97	1.98	1.99	2.00	2.08	2.24	2.33	2.53	2.14	2.16	2.37	2.13	2.17	S	2.17	2.42	2.33	2.33	1.97	2.64	2.18	24
16	2.51	2.73	2.26	2.89	2.14	2.03	2.04	2.05	2.08	2.14	2.06	2.06	2.03	2.15	2.15	2.02	2.28	2.25	S	2.10	2.08	2.07	2.10	2.10	2.02	2.69	2.19	24
17	2.09	2.18	2.08	2.09	2.10	2.09	2.10	2.07	2.07	P	P	2.06	2.05	2.03	2.00	2.03	2.05	S	2.04	2.06	2.05	2.05	2.04	2.05	2.00	2.18	2.07	22
18	2.17	2.24	2.10	2.09	2.16	2.01	1.99	2.18	2.29	2.24	2.21	2.07	2.07	2.27	2.21	2.06	S	2.18	2.21	1.99	2.03	2.00	2.09	2.00	1.99	2.29	2.12	24
19	1.99	1.99	2.00	2.02	2.00	2.06	2.01	2.05	2.05	2.02	2.01	2.02	2.09	2.01	2.01	S	2.01	2.02	2.14	2.09	2.14	2.06	2.23	2.19	1.99	2.23	2.05	24
20	2.27	2.21	2.26	2.19	2.10	2.09	2.09	2.11	2.13	2.12	2.10	2.10	2.06	2.06	S	1.99	1.99	2.20	2.28	2.01	2.01	2.06	2.03	2.07	1.99	2.28	2.11	24
21	2.10	2.03	2.05	2.05	2.02	2.01	2.01	2.01	2.00	1.99	2.00	2.01	2.01	S	2.01	2.19	2.30	2.19	2.09	2.03	2.09	2.29	2.19	2.67	1.99	2.67	2.10	24
22	2.94	2.01	2.24	2.13	2.12	2.11	2.13	2.10	2.09	2.17	2.06	2.05	S	2.04	2.05	2.04	2.07	2.08	2.14	2.10	2.12	2.14	2.16	2.21	2.01	2.94	2.14	24
23	2.21	2.26	2.25	2.36	2.37	2.39	2.37	2.38	2.37	2.33	2.31	S	2.30	2.27	2.16	2.17	2.34	2.29	2.28	2.19	2.20	2.29	2.27	2.27	2.16	2.39	2.29	24
24	2.22	2.38	2.34	2.43	2.40	2.69	2.60	2.61	2.56	2.51	S	2.45	2.40	2.53	2.40	2.38	2.61	2.48	2.31	2.27	2.22	2.16	2.53	2.81	2.16	2.81	2.45	24
25	3.34	4.66	4.66	2.64	3.62	3.82	3.01	3.05	X	S	2.41	2.21	2.18	2.18	2.55	2.22	3.37	2.15	2.13	2.14	2.14	2.13	2.09	2.09	2.09	4.66	2.76	23
26	2.11	2.09	2.25	2.32	2.67	2.02	2.02	2.02	S	2.02	2.00	2.02	2.01	2.01	2.01	2.01	2.52	2.03	2.02	2.02	2.02	2.02	2.05	2.07	2.00	2.67	2.10	24
27	2.08	2.06	2.07	2.03	2.02	2.03	2.06	S	2.07	2.10	2.05	2.02	2.01	2.12	2.16	2.17	2.21	2.11	2.02	2.05	2.27	2.85	2.19	2.07	2.01	2.85	2.12	24
28	2.21	2.17	2.03	2.02	2.03	2.08	S	2.07	2.08	2.08	2.04	2.04	2.19	2.39	2.39	2.34	2.48	2.50	2.46	2.32	2.45	2.37	2.30	2.31	2.02	2.50	2.23	24
29	2.68	2.37	2.31	2.25	2.15	S	2.10	2.13	2.14	2.23	2.24	2.33	2.31	2.64	2.83	2.23	2.31	2.45	2.36	2.35	2.48	2.43	2.29	2.24	2.10	2.83	2.34	24
30	3.25	2.60	2.18	3.09	S	2.12	2.12	2.14	2.19	2.13	2.14	2.14	2.09	2.06	2.01	2.10	2.03	2.04	2.03	2.04	2.07	2.08	2.07	2.01	2.01	3.25	2.21	24
31	2.09	2.23	2.07	S	2.15	2.09	2.10	2.11	2.12	2.12	2.13	2.10	2.18	2.10	2.12	2.17	2.11	2.43	2.48	2.21	2.48	2.54	2.70	2.10	2.07	2.70	2.21	24
HOURLY MAX	3.34	4.66	4.66	3.09	3.62	3.82	3.44	3.05	2.65	3.35	3.22	3.32	2.40	2.64	2.83	2.38	3.37	2.50	2.48	2.70	2.56	2.85	2.70	3.08				
HOURLY AVG	2.29	2.27	2.22	2.24	2.22	2.19	2.22	2.20	2.19	2.23	2.17	2.19	2.14	2.17	2.14	2.11	2.25	2.16	2.14	2.11	2.15	2.16	2.16	2.19				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	694
MAXIMUM INSTANTANEOUS VALUE:	4.66 ppm @ HOUR 1 ON DAY 25
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	5 hrs
OPERATIONAL TIME:	731 hrs
STANDARD DEVIATION:	0.27

TOTAL HYDROCARBONS Instantaneous Maximum (THC ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - October 2018

METHANE MAX Instantaneous Maximum (CH₄ ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	2.20	2.33	2.33	2.50	2.50	2.45	2.47	2.47	2.44	2.36	S	2.05	2.04	2.04	2.03	2.09	3.22	2.06	2.03	2.02	2.03	2.04	2.04	2.04	2.02	3.22	2.25	24
2	2.04	2.05	2.03	2.03	2.01	2.02	2.02	2.00	1.99	S	2.00	X	2.01	2.04	2.02	2.05	2.04	2.16	2.17	2.04	2.09	2.00	1.99	1.98	1.98	2.17	2.03	23
3	2.45	2.10	1.98	2.09	2.00	1.99	1.98	1.99	S	2.02	2.07	2.27	2.19	1.97	1.98	1.99	2.00	1.96	1.98	2.01	2.16	2.13	2.21	2.13	1.96	2.45	2.07	24
4	2.12	2.08	2.11	2.15	2.49	2.17	2.30	S	2.32	2.34	2.40	2.37	2.36	2.20	2.15	2.06	1.99	1.99	2.00	2.02	1.99	2.02	2.05	2.06	1.99	2.49	2.16	24
5	2.05	2.04	2.05	2.06	2.07	2.08	S	2.16	2.34	3.35	3.22	3.32	2.13	2.15	2.17	2.31	2.53	1.99	1.98	2.01	2.03	2.09	2.09	2.26	1.98	3.35	2.28	24
6	2.31	2.25	2.28	2.36	2.26	S	2.37	2.30	2.22	2.17	2.21	2.33	2.30	2.19	2.10	2.04	2.03	2.27	2.11	2.11	2.32	2.14	2.20	3.08	2.03	3.08	2.26	24
7	2.76	2.16	2.13	2.18	S	2.20	2.20	2.18	2.31	2.09	2.09	2.08	2.09	2.09	2.05	2.11	2.04	2.04	2.03	2.02	2.00	2.00	1.99	1.99	1.99	2.76	2.12	24
8	2.00	2.02	2.01	S	2.01	2.00	1.99	1.99	2.00	2.14	2.03	2.12	2.11	2.35	2.15	2.15	2.13	2.13	2.15	2.27	2.13	2.21	2.21	2.05	1.99	2.35	2.10	24
9	2.07	2.72	S	2.18	2.16	2.09	2.26	2.49	2.45	2.34	2.16	2.17	2.09	2.24	2.12	2.01	2.21	2.25	2.03	2.02	2.02	2.11	2.12	2.10	2.01	2.72	2.19	24
10	2.06	S	2.04	2.05	2.07	2.37	3.44	2.65	2.65	2.38	2.36	2.31	2.20	2.13	2.05	2.08	2.04	2.06	2.04	2.03	2.03	2.04	2.04	2.00	2.00	3.44	2.22	24
11	S	1.99	1.99	2.00	2.01	2.02	2.01	2.00	2.02	C	C	C	C	C	2.03	2.03	2.12	2.06	2.07	2.04	2.06	2.06	S	1.99	2.12	2.03	24	
12	2.09	2.06	2.08	2.16	2.11	P	P	P	P	P	P	P	P	X	2.10	2.10	2.10	2.11	2.24	2.11	2.03	2.06	S	2.14	2.03	2.24	2.11	15
13	2.08	2.06	2.09	2.19	2.07	2.06	2.22	2.11	2.06	2.62	2.06	2.05	2.02	2.03	2.03	2.04	2.07	2.04	2.16	2.70	2.56	S	2.03	2.09	2.02	2.70	2.15	24
14	2.03	2.04	2.14	2.07	2.07	2.11	2.13	2.13	2.10	2.06	2.05	2.04	2.02	2.04	2.03	2.03	2.04	2.04	2.11	2.05	S	2.04	2.03	2.02	2.02	2.14	2.06	24
15	2.11	1.99	1.99	2.14	2.64	1.97	1.97	1.98	1.98	2.00	2.08	2.24	2.33	2.53	2.14	2.16	2.37	2.13	2.17	S	2.17	2.42	2.33	2.33	1.97	2.64	2.18	24
16	2.51	2.73	2.26	2.89	2.14	2.03	2.04	2.05	2.08	2.14	2.06	2.06	2.03	2.02	2.02	2.02	2.28	2.25	S	2.10	2.08	2.07	2.10	2.10	2.02	2.89	2.18	24
17	2.09	2.09	2.08	2.09	2.09	2.09	2.10	2.07	2.07	P	P	2.06	2.05	2.03	2.00	2.03	2.05	S	2.04	2.04	2.05	2.05	2.04	2.05	2.00	2.10	2.06	22
18	2.17	2.08	2.10	2.09	2.16	2.01	1.99	2.18	2.29	2.24	2.21	2.07	2.07	2.27	2.21	2.06	S	2.18	2.21	1.99	2.03	1.98	2.09	2.00	1.98	2.29	2.12	24
19	1.99	1.99	2.00	2.02	2.00	2.06	2.01	2.05	2.03	2.02	2.01	2.02	2.01	2.01	2.01	S	2.01	2.02	2.14	2.09	2.14	2.06	2.23	2.19	1.99	2.23	2.05	24
20	2.27	2.21	2.18	2.19	2.10	2.09	2.09	2.11	2.11	2.12	2.10	2.10	2.06	2.06	S	1.99	1.99	2.20	2.28	2.01	2.01	2.06	2.03	2.07	1.99	2.28	2.11	24
21	2.10	2.03	2.02	2.02	2.02	2.01	2.01	2.01	2.00	1.99	2.00	2.01	2.01	S	2.01	2.15	2.30	2.19	2.09	2.03	2.09	2.29	2.19	2.67	1.99	2.67	2.10	24
22	2.94	2.01	2.24	2.13	2.12	2.11	2.13	2.10	2.08	2.09	2.06	2.05	S	2.04	2.05	2.04	2.07	2.08	2.14	2.10	2.12	2.14	2.16	2.17	2.01	2.94	2.14	24
23	2.21	2.26	2.25	2.36	2.37	2.39	2.37	2.38	2.37	2.33	2.31	S	2.30	2.27	2.16	2.17	2.34	2.29	2.27	2.19	2.20	2.29	2.27	2.27	2.16	2.39	2.29	24
24	2.22	2.38	2.34	2.39	2.40	2.69	2.60	2.61	2.56	2.51	S	2.45	2.40	2.53	2.40	2.38	2.61	2.48	2.31	2.27	2.22	2.16	2.53	2.81	2.16	2.81	2.45	24
25	3.34	4.66	4.66	2.64	3.62	3.82	3.01	3.05	X	S	2.41	2.21	2.18	2.18	2.55	2.22	3.37	2.15	2.13	2.14	2.14	2.13	2.09	2.09	2.09	4.66	2.76	23
26	2.11	2.09	2.25	2.32	2.67	2.02	2.02	2.02	S	2.02	2.00	2.02	2.01	2.01	2.00	2.01	2.52	2.03	2.01	2.02	2.02	2.05	2.07	2.00	2.00	2.67	2.10	24
27	2.08	2.06	2.07	2.03	2.02	2.02	2.06	S	2.07	2.10	2.05	2.02	2.01	2.12	2.16	2.17	2.21	2.11	2.02	2.05	2.27	2.85	2.19	2.07	2.01	2.85	2.12	24
28	2.21	2.17	2.02	2.02	2.03	2.08	S	2.07	2.08	2.08	2.04	2.04	2.19	2.39	2.39	2.34	2.48	2.50	2.46	2.32	2.45	2.37	2.30	2.31	2.02	2.50	2.23	24
29	2.68	2.37	2.31	2.25	2.15	S	2.10	2.13	2.14	2.23	2.24	2.33	2.31	2.64	2.83	2.23	2.31	2.45	2.36	2.35	2.48	2.43	2.29	2.24	2.10	2.83	2.34	24
30	3.25	2.60	2.18	3.09	S	2.12	2.12	2.14	2.19	2.13	2.14	2.14	2.09	2.06	2.01	2.10	2.03	2.04	2.03	2.04	2.07	2.07	2.08	2.07	2.01	3.25	2.21	24
31	2.09	2.23	2.07	S	2.08	2.09	2.10	2.11	2.12	2.12	2.13	2.10	2.18	2.10	2.10	2.17	2.11	2.43	2.48	2.21	2.48	2.54	2.70	2.10	2.07	2.70	2.21	24
HOURLY MAX	3.34	4.66	4.66	3.09	3.62	3.82	3.44	3.05	2.65	3.35	3.22	3.32	2.40	2.64	2.83	2.38	3.37	2.50	2.48	2.70	2.56	2.85	2.70	3.08				
HOURLY AVG	2.29	2.26	2.21	2.23	2.22	2.18	2.22	2.20	2.19	2.23	2.17	2.19	2.14	2.17	2.14	2.11	2.25	2.16	2.14	2.11	2.15	2.16	2.16	2.19				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	694
MAXIMUM INSTANTANEOUS VALUE:	4.66 ppm @ HOUR 1 ON DAY 25
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	5 hrs
OPERATIONAL TIME:	731 hrs
STANDARD DEVIATION:	0.27

METHANE MAX Instantaneous Maximum (CH₄ ppm)





NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY MIN.	DAILY MAX.	24-HR AVG.	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59					
DAY 1	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.05	0.00	S	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.05	0.00	24
2	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	X	0.00	0.00	0.00	0.00	0.00	0.03	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.03	0.00	23
3	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.03	0.00	0.00	0.00	0.00	0.00	0.04	0.00	0.00	0.00	0.00	0.00	0.04	0.00	24
4	0.00	0.03	0.17	0.00	0.01	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.04	0.00	0.00	0.00	0.00	0.00	0.17	0.01	24
5	0.00	0.01	0.00	0.01	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.02	0.07	0.00	0.00	0.00	0.01	0.01	0.00	0.00	0.07	0.01	24	
6	0.00	0.02	0.00	0.00	0.00	S	0.00	0.00	0.06	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.01	0.00	0.02	0.00	0.00	0.01	0.00	0.00	0.00	0.06	0.01	24	
7	0.00	0.01	0.03	0.03	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.06	0.00	0.06	0.01	24	
8	0.00	0.00	0.00	S	0.00	0.00	0.00	0.05	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.04	0.00	0.00	0.05	0.00	24	
9	0.00	0.00	S	0.00	0.00	0.01	0.00	0.00	0.00	0.02	0.00	0.00	0.00	0.02	0.00	0.00	0.03	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.03	0.00	24
10	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	24
11	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	C	C	C	C	C	0.00	0.00	0.00	0.00	0.01	0.00	0.01	0.00	0.00	S	0.00	0.01	0.00	24	
12	0.00	0.01	0.07	0.24	0.00	P	P	P	P	P	P	P	X	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.03	0.00	S	0.00	0.00	0.24	0.03	15	
13	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.07	S	0.03	0.00	0.00	0.07	0.01	24	
14	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.05	0.00	0.00	0.00	0.00	0.02	0.01	0.01	0.00	0.00	0.01	0.00	0.00	S	0.00	0.00	0.01	0.00	0.05	0.01	24	
15	0.00	0.00	0.02	0.00	0.00	0.06	0.00	0.00	0.00	0.00	0.00	0.14	0.01	0.00	0.00	0.00	0.02	0.00	0.00	S	0.00	0.04	0.00	0.07	0.00	0.14	0.02	24	
16	0.00	0.00	0.00	0.00	0.00	0.02	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.13	0.13	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.13	0.01	24	
17	0.00	0.12	0.00	0.00	0.01	0.01	0.01	0.02	0.00	P	P	0.01	0.00	0.00	0.01	0.00	0.01	S	0.00	0.04	0.00	0.01	0.00	0.00	0.00	0.12	0.01	22	
18	0.00	0.18	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.04	0.00	0.05	0.04	0.00	S	0.00	0.01	0.00	0.00	0.03	0.00	0.00	0.00	0.18	0.02	24	
19	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.00	0.04	0.00	0.00	0.00	0.09	0.02	0.00	S	0.00	0.00	0.00	0.03	0.00	0.00	0.01	0.00	0.00	0.09	0.01	24	
20	0.00	0.00	0.12	0.00	0.00	0.01	0.00	0.00	0.03	0.00	0.00	0.00	0.00	0.00	S	0.00	0.01	0.06	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.12	0.01	24	
21	0.00	0.00	0.04	0.04	0.00	0.00	0.01	0.00	0.00	0.00	0.00	0.00	0.00	S	0.00	0.17	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.00	0.17	0.01	24	
22	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.03	0.09	0.00	0.00	S	0.00	0.00	0.00	0.01	0.00	0.01	0.00	0.02	0.00	0.00	0.06	0.00	0.09	0.01	24	
23	0.02	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.00	0.01	S	0.00	0.00	0.00	0.00	0.00	0.01	0.10	0.00	0.01	0.00	0.00	0.00	0.00	0.10	0.01	24	
24	0.00	0.00	0.00	0.10	0.00	0.00	0.02	0.00	0.00	0.00	S	0.00	0.05	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.10	0.01	24	
25	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	X	S	0.00	0.01	0.03	0.00	0.01	0.00	0.00	0.00	0.02	0.00	0.01	0.00	0.00	0.01	0.00	0.03	0.01	23	
26	0.00	0.00	0.00	0.00	0.02	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.01	0.02	0.00	0.00	0.04	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.04	0.00	24	
27	0.00	0.00	0.00	0.00	0.01	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.06	0.01	0.00	0.06	0.00	24	
28	0.00	0.01	0.03	0.00	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00	0.00	0.00	0.00	0.03	0.00	24	
29	0.00	0.00	0.00	0.05	0.02	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.04	0.00	0.00	0.00	0.00	0.02	0.00	0.00	0.00	0.00	0.05	0.01	24	
30	0.00	0.02	0.00	0.00	S	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.01	0.01	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.02	0.00	24	
31	0.00	0.00	0.00	S	0.08	0.00	0.00	0.00	0.01	0.00	0.00	0.02	0.00	0.00	0.03	0.00	0.03	0.00	0.01	0.00	0.00	0.00	0.01	0.01	0.00	0.08	0.01	24	
HOURLY MAX	0.02	0.18	0.17	0.24	0.08	0.06	0.02	0.05	0.06	0.09	0.01	0.14	0.09	0.13	0.13	0.17	0.04	0.07	0.10	0.04	0.07	0.04	0.06	0.07					
HOURLY AVG	0.00	0.01	0.02	0.02	0.01	0.00	0.00	0.01	0.01	0.00	0.00	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.00	0.01	0.01	0.01					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	147
MAXIMUM INSTANTANEOUS VALUE:	0.24 ppm @ HOUR 3 ON DAY 12
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	5 hrs
OPERATIONAL TIME:	731 hrs
STANDARD DEVIATION:	0.02

NON-METHANE HYDROCARBONS Instantaneous Maximum (NMHC ppm)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - October 2018

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	3	3	3	4	4	4	4	4	4	4	S	4	2	1	2	3	6	6	7	6	3	2	1	2	1	7	4	24
2	2	1	1	2	2	2	4	3	2	S	3	2	9	6	9	5	8	7	21	13	1	2	0	1	0	21	5	24
3	1	1	1	0	6	14	2	3	S	7	2	3	2	1	2	2	1	1	1	2	2	3	3	2	0	14	3	24
4	3	2	3	2	3	3	4	S	6	5	6	6	5	4	9	7	2	2	3	2	3	2	2	10	2	10	4	24
5	2	2	3	3	5	8	S	7	23	23	8	3	2	2	1	13	2	2	1	1	3	4	4	4	1	23	6	24
6	4	3	3	2	2	S	5	3	3	4	4	5	6	6	6	5	5	5	4	3	3	3	3	3	2	6	4	24
7	2	2	2	5	S	8	7	8	4	2	1	5	5	3	5	4	3	1	1	1	1	2	2	2	1	8	3	24
8	2	2	2	S	4	2	2	2	2	3	3	2	1	1	1	1	1	1	1	1	1	1	1	0	0	4	1	24
9	0	0	S	5	5	6	9	9	8	6	2	2	1	1	1	1	1	2	5	1	1	2	2	2	0	9	3	24
10	2	S	3	2	2	4	5	6	6	6	C	C	C	C	C	C	C	C	C	12	2	2	11	1	1	12	5	24
11	S	4	2	2	8	2	2	2	7	9	4	3	6	8	4	9	5	9	5	4	3	3	3	S	2	9	5	24
12	6	3	3	4	6	P	P	P	P	P	P	P	X	3	4	4	3	2	1	2	2	S	4	1	6	3	15	24
13	1	1	1	1	3	1	1	1	2	5	4	1	2	1	2	1	2	4	2	2	6	S	3	4	1	6	2	24
14	7	1	1	1	3	3	7	8	9	3	6	4	11	12	9	13	11	9	8	14	S	8	6	2	1	14	7	24
15	2	1	1	1	1	2	4	3	S1	5	4	4	2	4	5	7	4	5	S	5	3	1	1	1	1	7	3	23
16	1	1	1	1	10	5	36	7	15	5	7	21	2	4	6	10	2	3	S	18	10	3	2	3	1	36	7	24
17	3	3	4	6	9	4	4	5	6	P	P	4	3	6	9	11	14	S	15	8	25	5	11	5	3	25	8	22
18	6	6	5	3	3	1	4	2	4	4	4	1	1	1	2	2	S	5	4	1	2	3	7	1	1	7	3	24
19	1	1	1	1	1	1	6	16	3	1	1	25	1	3	2	S	3	2	3	3	3	3	3	3	1	25	4	24
20	4	2	2	2	1	1	1	4	2	8	3	4	9	3	S	2	4	3	2	3	9	1	4	2	1	9	3	24
21	1	2	1	1	3	1	1	2	1	2	1	0	1	S	4	5	2	1	2	3	4	1	1	1	0	5	2	24
22	1	1	1	2	2	2	3	3	3	4	22	2	S	3	3	2	1	2	2	2	2	2	2	2	1	22	3	24
23	2	2	2	3	3	3	3	4	4	4	S	5	3	2	3	4	6	6	4	4	5	4	5	4	2	6	4	24
24	4	6	5	5	5	5	5	6	6	8	S	10	8	9	8	7	7	7	6	5	4	4	27	4	4	27	7	24
25	4	4	4	4	4	5	5	38	6	S	6	7	5	6	6	6	32	3	6	6	5	5	4	3	3	38	8	24
26	4	3	3	2	2	2	6	11	S	3	2	8	7	5	10	3	19	29	29	5	6	22	3	5	2	29	8	24
27	3	3	2	2	1	1	2	S	8	8	4	2	3	1	1	3	1	3	1	7	5	32	1	6	1	32	4	24
28	1	1	1	1	1	2	S	5	3	3	3	2	3	5	3	2	3	3	2	2	2	2	2	2	1	5	2	24
29	2	2	2	2	3	S	16	3	5	4	3	2	3	2	5	2	4	3	1	1	3	2	3	1	16	3	24	
30	2	2	2	2	S	3	8	5	13	40	6	4	13	4	2	15	4	12	17	7	3	2	2	7	2	40	8	24
31	10	2	3	S	8	7	10	29	14	7	15	20	14	12	14	9	7	7	9	9	7	5	3	3	2	29	10	24
HOURLY MAX	10	6	5	6	10	14	36	38	23	40	22	25	14	12	14	15	32	29	29	18	25	32	27	10				
HOURLY AVG	3	2	2	2	4	4	6	7	6	7	5	6	5	4	5	5	6	5	6	5	4	5	4	3				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	685
MAXIMUM INSTANTANEOUS VALUE:	40 ppb @ HOUR 9 ON DAY 30
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	9 hrs
OPERATIONAL TIME:	732 hrs
STANDARD DEVIATION:	5

OXIDES OF NITROGEN Instantaneous Maximum (NO_x ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - October 2018

NITRIC OXIDE Instantaneous Maximum (NO ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.				
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.					
DAY																																
1	0	0	0	0	0	0	0	0	1	1	S	1	0	0	0	2	2	1	1	1	0	0	0	0	0	0	2	0	24			
2	0	0	0	0	0	0	0	0	0	S	1	0	2	6	4	2	3	4	13	7	0	1	0	0	0	0	13	2	24			
3	0	0	0	0	4	8	1	2	S	2	0	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0	8	1	24			
4	0	0	0	0	0	0	0	S	1	1	2	1	1	1	3	3	0	0	0	0	0	0	0	6	0	6	1	24				
5	0	0	0	0	1	5	S	1	14	14	4	1	0	0	0	7	0	0	0	0	0	0	0	0	0	0	14	2	24			
6	0	0	0	0	0	S	0	0	0	1	1	1	2	1	1	1	0	1	0	0	0	0	0	0	0	0	2	0	24			
7	0	0	0	0	S	0	0	1	1	0	0	1	1	0	1	1	0	0	0	0	0	0	0	0	0	0	1	0	24			
8	0	0	0	S	0	0	0	0	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	24			
9	0	0	S	0	0	0	0	2	2	2	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	2	0	24			
10	0	S	0	0	0	0	0	0	1	1	C	C	C	C	C	C	C	C	C	C	4	0	0	4	0	0	4	1	24			
11	S	0	0	0	4	0	0	0	3	3	1	1	4	4	1	7	1	2	0	0	0	0	0	0	S	0	7	1	24			
12	0	0	0	0	2	P	P	P	P	P	P	P	P	X	0	1	0	0	0	0	1	1	S	0	0	2	0	15				
13	0	0	0	0	1	0	0	0	1	3	2	0	1	0	1	0	0	2	0	0	2	S	0	1	0	3	1	24				
14	3	0	0	0	1	0	0	2	3	1	7	1	8	5	23	7	3	4	3	2	S	0	1	0	0	23	3	24				
15	0	0	0	0	0	1	2	1	S1	5	1	1	0	2	7	2	5	1	1	S	2	1	0	0	0	7	2	23				
16	0	0	0	0	7	2	23	3	8	2	3	7	1	1	2	4	0	0	S	8	2	0	0	0	0	23	3	24				
17	0	0	0	0	2	0	0	3	2	P	P	1	1	2	2	2	2	S	2	1	10	0	3	0	0	10	2	22				
18	0	0	0	0	0	0	1	0	1	3	1	0	0	0	0	1	S	1	1	0	0	1	6	0	0	6	1	24				
19	0	0	0	0	0	0	2	7	1	0	0	12	0	1	0	S	0	0	0	0	0	0	0	0	0	12	1	24				
20	0	0	0	0	0	0	0	1	0	3	1	1	5	1	S	0	1	1	0	2	3	0	1	0	0	5	1	24				
21	0	0	0	0	1	0	0	0	0	2	0	0	0	S	1	1	0	0	0	1	1	0	0	0	0	2	0	24				
22	0	0	0	0	0	0	0	0	0	1	21	0	S	0	0	0	0	0	0	0	0	0	0	0	0	21	1	24				
23	0	0	0	0	0	0	0	0	1	1	1	S	1	1	0	0	0	0	0	0	0	0	0	0	0	1	0	24				
24	0	0	0	0	0	0	0	0	1	2	S	3	2	3	2	1	1	0	0	0	0	0	14	0	0	14	1	24				
25	0	0	0	0	0	0	1	31	1	S	2	2	1	2	2	2	17	0	0	0	0	0	0	0	0	31	3	24				
26	0	0	0	0	0	0	3	7	S	0	0	5	3	1	6	0	8	16	16	0	1	12	0	3	0	16	4	24				
27	0	0	0	0	0	0	0	S	3	5	1	0	2	0	0	1	0	1	0	3	2	16	0	3	0	16	2	24				
28	0	0	0	0	0	0	S	0	0	0	1	0	0	1	0	0	0	0	0	0	0	0	0	0	0	1	0	24				
29	0	0	0	0	0	S	9	1	3	3	2	0	1	1	1	3	0	1	1	0	0	1	0	1	0	9	1	24				
30	0	0	0	0	S	0	5	2	7	34	3	1	9	2	1	8	3	6	15	7	0	0	0	4	0	34	5	24				
31	8	0	0	0	S	1	0	4	28	7	3	9	9	7	4	5	2	0	0	0	0	0	0	0	0	0	28	4	24			
HOURLY MAX	8	0	0	0	7	8	23	31	14	34	21	12	9	6	23	8	17	16	16	8	10	16	14	6								
HOURLY AVG	0	0	0	0	1	1	2	3	2	3	3	2	2	1	2	2	2	1	2	1	1	1	1	1								


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	271
MAXIMUM INSTANTANEOUS VALUE:	34 ppb @ HOUR 9 ON DAY 30
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	9 hrs
STANDARD DEVIATION:	3
OPERATIONAL TIME:	732 hrs

NITRIC OXIDE Instantaneous Maximum (NO ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - October 2018

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.	
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.		
DAY 1	3	3	3	4	4	4	4	4	3	3	S	3	2	1	2	2	4	6	7	5	3	2	2	2	1	7	3	24	
2	2	1	1	2	2	2	4	3	2	S	3	2	7	2	5	3	6	4	9	6	1	1	1	1	1	1	9	3	24
3	1	1	1	1	2	6	1	2	S	6	2	2	1	1	2	2	1	1	1	2	2	3	3	2	1	6	2	24	
4	3	2	3	2	3	3	4	S	6	4	4	4	4	3	7	4	2	2	3	2	4	2	2	5	2	7	3	24	
5	3	3	3	3	5	4	S	6	9	9	4	2	2	2	7	2	2	2	2	3	4	4	4	2	9	4	24		
6	4	3	3	2	2	S	5	3	3	3	3	4	5	5	5	5	4	4	3	3	3	3	3	2	5	4	24		
7	3	2	2	5	S	8	7	7	4	2	1	4	4	2	4	3	2	1	1	1	2	2	2	2	1	8	3	24	
8	2	2	2	S	4	2	2	2	2	2	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	4	1	24	
9	1	1	S	5	5	6	9	9	7	4	2	1	1	1	1	1	1	2	5	1	2	2	2	2	1	9	3	24	
10	2	S	3	2	2	4	5	6	6	5	C	C	C	C	C	C	C	C	C	8	2	2	7	2	2	8	4	24	
11	S	4	2	2	4	2	2	2	4	6	4	2	2	4	3	5	4	7	4	4	3	3	3	S	2	7	4	24	
12	6	4	4	4	4	P	P	P	P	P	P	P	X	3	3	4	2	2	1	2	1	S	4	1	6	3	15		
13	1	1	1	1	2	1	1	1	1	2	2	1	1	1	1	1	2	3	2	2	5	S	3	3	1	5	2	24	
14	5	1	1	1	2	3	7	6	6	2	3	3	5	7	4	6	7	5	9	12	S	8	5	2	1	12	5	24	
15	2	1	1	1	1	2	2	2	S1	3	3	3	2	2	2	3	3	4	S	3	2	1	1	1	1	4	2	23	
16	1	1	1	1	3	3	14	3	10	3	4	15	2	3	4	6	2	3	S	10	7	3	2	3	1	15	4	24	
17	3	3	4	7	8	4	4	3	4	P	P	4	3	4	7	10	12	S	13	8	15	5	8	5	3	15	6	22	
18	6	6	5	3	3	1	3	2	3	2	3	1	1	1	1	1	S	4	3	1	1	2	2	1	1	6	3	24	
19	1	1	1	1	1	1	4	9	2	1	1	13	1	2	1	S	3	2	3	3	3	3	3	3	1	13	3	24	
20	4	3	2	2	2	2	2	3	2	5	2	3	4	2	S	2	3	2	2	1	6	1	2	1	1	6	3	24	
21	1	2	2	1	3	1	1	1	1	1	1	1	1	S	3	3	1	1	2	2	3	1	1	1	1	3	1	24	
22	1	1	1	2	2	2	3	3	2	3	4	2	S	3	2	2	2	3	3	2	2	2	2	3	1	4	2	24	
23	2	2	2	3	3	3	3	3	3	3	3	S	5	3	2	2	4	6	6	5	5	5	5	5	2	6	4	24	
24	4	6	5	5	5	5	5	5	5	6	S	7	6	7	6	6	7	7	6	5	5	4	13	4	4	13	6	24	
25	4	5	4	5	5	5	5	9	5	S	6	5	4	5	4	4	16	4	6	6	5	5	4	4	4	16	5	24	
26	4	4	4	2	2	2	3	4	S	3	2	3	4	4	4	3	11	15	14	5	5	10	4	3	2	15	5	24	
27	3	3	2	2	2	1	2	S	5	5	3	1	2	1	1	2	1	2	1	4	3	16	1	4	1	16	3	24	
28	1	1	1	1	1	2	S	5	3	3	3	2	2	4	3	3	3	3	3	2	2	2	2	2	1	5	2	24	
29	2	2	2	2	3	S	8	2	4	2	2	2	2	2	2	2	2	3	2	1	2	2	2	2	1	8	2	24	
30	2	2	2	2	S	3	4	3	7	7	4	3	5	3	2	8	2	6	3	4	2	2	2	3	2	8	3	24	
31	5	2	3	S	7	7	6	6	7	5	7	11	7	8	10	7	7	7	9	9	7	5	3	3	2	11	7	24	
HOURLY MAX	6	6	5	7	8	8	14	9	10	9	7	15	7	8	10	10	16	15	14	12	15	16	13	5					
HOURLY AVG	3	2	2	3	3	3	4	4	4	4	3	4	3	3	3	4	4	4	4	4	4	4	3	3					


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	691
MAXIMUM INSTANTANEOUS VALUE:	16 ppb @ HOUR 16 ON DAY 25
	VAR-VARIOUS
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	9 hrs
OPERATIONAL TIME:	732 hrs
STANDARD DEVIATION:	2

NITROGEN DIOXIDE Instantaneous Maximum (NO₂ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - October 2018

OZONE Instantaneous Maximum (O₃ ppb)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	24.4	22.6	21.3	20.3	19.2	18.6	18.3	18.1	19.5	26.0	S	32.2	31.7	32.9	32.8	33.3	33.0	31.8	30.3	30.7	30.5	28.5	28.1	28.2	18.1	33.3	26.6	24
2	28.2	28.7	29.0	27.5	23.2	23.1	22.1	23.0	23.9	S	29.4	30.0	30.4	33.1	32.7	34.0	33.5	33.7	32.0	31.0	30.9	30.4	30.6	31.0	22.1	34.0	29.2	24
3	31.2	31.2	31.3	31.5	31.1	30.1	29.0	27.5	S	30.1	32.8	34.5	35.0	34.7	33.9	34.0	34.0	33.8	33.1	35.8	33.8	31.6	31.7	31.7	27.5	35.8	32.3	24
4	30.6	30.4	27.9	27.7	26.0	23.5	22.8	S	21.9	22.4	26.0	27.5	36.2	35.2	37.0	40.9	41.7	40.6	39.8	37.8	35.4	34.9	33.1	31.1	21.9	41.7	31.8	24
5	30.1	29.3	27.9	24.6	21.8	20.0	S	19.6	22.3	25.1	28.9	32.9	33.9	34.4	34.4	34.1	33.6	32.6	30.7	28.2	26.5	24.5	26.4	25.3	19.6	34.4	28.1	24
6	23.5	24.4	25.1	25.1	24.0	S	22.5	22.4	22.7	25.2	26.0	27.5	29.8	29.4	32.2	32.4	32.6	32.5	31.4	28.3	25.2	23.5	22.9	23.2	22.4	32.6	26.6	24
7	25.1	25.5	24.6	23.1	S	13.5	12.7	17.5	20.4	22.2	23.1	23.7	24.7	25.2	24.8	24.1	24.7	24.1	23.5	24.1	24.9	25.0	26.5	26.8	12.7	26.8	23.0	24
8	26.2	28.1	24.8	S	26.1	25.5	26.1	26.8	27.5	28.3	28.9	31.5	31.9	32.2	32.3	32.5	32.7	32.7	32.7	33.0	31.5	31.0	31.0	31.0	24.8	33.0	29.7	24
9	31.2	29.8	S	25.9	24.8	24.8	22.1	22.2	22.6	31.0	30.1	31.5	33.2	33.0	33.6	33.3	32.8	32.5	35.5	31.6	31.2	30.5	28.8	28.4	22.1	35.5	29.6	24
10	31.7	S	32.6	32.0	30.7	28.1	23.1	19.9	19.1	23.6	24.3	26.2	28.6	30.7	32.2	32.5	35.6	33.4	30.5	28.8	27.7	26.0	23.3	24.5	19.1	35.6	28.0	24
11	S	24.2	24.2	23.9	21.8	22.2	21.9	21.8	20.7	C	C	C	C	C	C	C	28.5	23.9	22.6	23.4	23.3	22.8	22.8	S	20.7	28.5	23.2	24
12	20.5	18.8	17.5	15.1	13.5	P	P	P	P	P	P	P	X	20.2	22.3	30.0	30.4	30.1	32.1	31.9	31.7	S	29.8	13.5	32.1	24.6	15	
13	28.6	30.0	31.2	29.8	30.9	32.4	32.3	29.9	30.1	31.8	33.2	34.1	34.4	34.8	35.1	35.3	35.9	36.0	35.9	35.5	35.0	S	34.4	33.4	28.6	36.0	33.0	24
14	33.4	32.2	30.6	27.9	26.7	26.3	18.8	17.0	20.3	22.2	22.4	21.6	21.3	28.0	28.7	28.2	27.7	26.2	24.1	20.5	S	25.4	31.5	33.1	17.0	33.4	25.8	24
15	36.5	37.8	37.5	36.4	35.6	34.6	33.8	32.8	31.9	31.9	33.5	34.8	35.4	35.0	34.1	33.4	32.5	32.8	31.0	S	29.7	28.9	28.6	28.7	28.6	37.8	33.4	24
16	28.7	28.4	28.4	27.4	26.8	24.5	23.7	19.1	18.2	20.1	23.1	24.9	26.5	30.3	31.1	32.8	28.2	27.7	S	26.6	31.1	33.2	32.5	31.3	18.2	33.2	27.2	24
17	28.8	28.2	28.9	27.4	30.3	31.5	32.1	33.9	36.9	P	P	50.6	52.9	57.1	58.0	57.6	49.4	S	50.3	52.4	51.2	49.1	47.8	44.7	27.4	58.0	42.8	22
18	41.2	42.8	45.5	45.1	45.4	44.7	43.7	42.2	40.5	39.9	41.5	42.6	42.9	43.1	43.0	43.5	S	42.2	40.8	39.9	39.3	38.8	37.8	37.8	37.8	45.5	41.9	24
19	37.0	36.2	35.8	35.8	35.6	35.2	35.0	31.3	32.2	33.7	34.6	35.3	34.9	34.9	34.7	S	33.5	33.1	32.5	29.2	29.5	27.1	27.8	26.4	26.4	37.0	33.1	24
20	26.5	27.1	25.8	25.0	25.3	25.4	25.1	23.2	22.5	24.4	25.4	29.0	32.5	41.0	S	44.2	44.0	43.0	41.9	40.8	39.5	39.5	39.4	39.0	22.5	44.2	32.6	24
21	38.6	35.2	34.4	35.3	33.5	33.7	34.7	34.9	35.8	37.4	37.9	38.5	39.0	S	38.3	37.8	37.7	37.3	37.1	36.7	37.0	37.1	36.8	36.8	33.5	39.0	36.6	24
22	37.0	37.0	36.8	33.0	31.3	30.6	27.2	31.4	31.6	29.8	33.6	35.8	S	38.3	39.4	40.0	39.0	36.7	34.5	33.1	32.6	31.4	30.3	28.9	27.2	40.0	33.9	24
23	28.1	27.4	27.3	26.7	23.5	22.8	21.7	21.4	22.0	23.8	24.9	S	29.1	32.8	34.2	35.0	33.0	26.6	24.3	24.6	24.3	23.8	24.6	26.2	21.4	35.0	26.4	24
24	27.3	27.3	21.1	21.9	22.4	21.7	22.6	22.9	22.7	22.5	S	25.6	27.4	26.4	27.4	26.9	26.0	25.1	27.3	28.8	30.8	31.9	31.7	27.8	21.1	31.9	25.9	24
25	28.5	26.9	26.6	24.6	27.1	23.4	24.2	23.5	24.1	S	24.9	26.8	31.1	33.0	34.9	37.1	38.3	34.0	30.2	28.5	29.0	31.1	37.8	41.8	23.4	41.8	29.9	24
26	42.0	33.4	30.2	35.3	33.8	30.0	29.8	29.1	S	29.5	33.7	36.8	37.3	39.1	37.7	36.8	35.3	35.5	35.1	35.6	35.7	36.1	35.6	34.4	29.1	42.0	34.7	24
27	31.0	30.7	32.6	33.8	33.4	31.1	30.1	S	25.6	28.9	34.6	37.2	37.2	36.7	36.6	37.7	38.0	36.5	36.1	35.3	33.9	33.4	33.0	34.5	25.6	38.0	33.8	24
28	34.8	34.7	30.2	29.9	29.7	28.9	S	23.6	25.8	25.7	29.3	33.9	33.4	31.4	32.4	29.9	29.4	28.8	27.8	27.5	26.4	25.8	25.7	24.5	23.6	34.8	29.1	24
29	23.3	22.6	22.3	19.6	21.8	S	22.7	24.7	24.1	24.4	22.0	22.2	24.6	25.2	24.9	24.9	23.8	22.9	24.1	23.3	21.0	20.3	19.3	19.3	19.3	25.2	22.9	24
30	18.8	18.7	19.3	18.1	S	16.1	13.2	13.5	10.8	11.9	13.4	16.9	23.6	29.4	30.6	30.1	24.0	22.2	22.5	20.6	19.9	18.5	16.8	15.1	10.8	30.6	19.3	24
31	17.6	16.9	16.7	S	14.4	13.3	12.3	12.1	13.0	15.0	21.1	20.5	21.5	20.3	24.5	29.6	27.2	24.1	21.5	18.7	24.1	31.9	25.4	23.8	12.1	31.9	20.2	24
HOURLY MAX	42.0	42.8	45.5	45.1	45.4	44.7	43.7	42.2	40.5	39.9	41.5	50.6	52.9	57.1	58.0	57.6	49.4	43.0	50.3	52.4	51.2	49.1	47.8	44.7				
HOURLY AVG	29.7	28.9	28.2	27.9	27.2	26.3	25.1	24.5	24.6	26.4	28.4	30.9	32.1	33.5	33.5	34.3	33.2	31.8	31.6	30.8	30.8	30.1	30.1	30.0				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

NUMBER OF NON-ZERO READINGS:	694
MAXIMUM INSTANTANEOUS VALUE:	58.0 ppb @ HOUR 14 ON DAY 17
IZS CALIBRATION TIME:	32 hrs
MONTHLY CALIBRATION TIME:	7 hrs
OPERATIONAL TIME:	733 hrs
STANDARD DEVIATION:	7.0

OZONE Instantaneous Maximum (O₃ ppb)





LAKELAND INDUSTRY & COMMUNITY ASSOCIATION
St. Lina Continuous Monitoring Station - October 2018

WIND SPEED Instantaneous Maximum (WS kph)

HR START (MST)	0:00	1:00	2:00	3:00	4:00	5:00	6:00	7:00	8:00	9:00	10:00	11:00	12:00	13:00	14:00	15:00	16:00	17:00	18:00	19:00	20:00	21:00	22:00	23:00	DAILY	DAILY	24-HR	RDGS.
HR END (MST)	0:59	1:59	2:59	3:59	4:59	5:59	6:59	7:59	8:59	9:59	10:59	11:59	12:59	13:59	14:59	15:59	16:59	17:59	18:59	19:59	20:59	21:59	22:59	23:59	MIN.	MAX.	AVG.	
DAY 1	15.9	14.3	13.2	11.6	9.0	8.6	8.6	7.9	9.7	9.9	13.4	13.2	15.6	13.6	12.5	9.7	11.9	12.3	12.3	16.0	21.9	16.2	13.8	13.6	7.9	21.9	12.7	24
2	11.6	13.6	20.7	27.0	18.2	10.9	10.1	15.8	14.2	17.6	17.8	18.9	33.1	28.7	34.0	32.7	29.2	26.1	19.5	19.3	21.1	15.1	16.0	17.1	10.1	34.0	20.3	24
3	23.0	21.9	22.4	23.2	23.9	22.8	21.0	17.6	23.5	27.4	24.1	28.9	38.5	31.6	34.8	30.0	23.7	19.1	13.0	12.1	10.5	10.3	12.5	12.1	10.3	38.5	22.0	24
4	16.2	20.6	21.1	16.7	15.6	16.5	18.2	16.2	22.2	17.8	21.0	19.1	30.5	31.8	28.9	30.0	32.0	26.1	19.3	18.1	27.2	30.9	27.8	24.8	15.6	32.0	22.9	24
5	27.2	21.7	24.3	20.0	19.5	15.6	19.7	20.6	23.1	17.8	20.4	27.2	29.4	30.0	29.8	24.3	22.8	23.7	26.5	14.1	19.6	27.6	26.5	30.7	14.1	30.7	23.4	24
6	30.9	23.0	23.7	21.7	23.2	25.0	27.0	21.1	27.2	30.9	22.1	19.1	32.9	27.2	25.9	30.2	34.8	21.3	27.6	23.2	19.1	20.4	15.8	14.3	14.3	34.8	24.5	24
7	17.1	13.8	11.9	16.0	12.5	12.7	16.7	24.3	26.3	43.8	41.0	29.8	28.7	39.2	30.0	33.8	38.8	31.1	32.0	32.7	42.5	38.1	39.7	29.4	11.9	43.8	28.4	24
8	28.3	29.4	26.7	28.5	29.4	28.9	25.4	28.9	26.3	26.3	30.3	27.8	25.0	19.3	19.1	19.3	21.1	19.1	20.8	22.4	14.9	15.8	27.0	21.5	14.9	30.3	24.2	24
9	22.4	8.4	11.0	11.9	10.3	10.8	11.0	11.4	8.6	15.8	22.8	21.5	17.6	15.8	13.4	13.6	10.8	10.6	9.7	10.5	12.5	14.3	10.5	10.3	8.4	22.8	13.1	24
10	9.0	16.2	19.3	20.0	22.6	18.6	13.8	14.3	13.8	17.8	17.4	16.9	25.2	28.9	26.9	22.8	24.3	15.1	11.9	11.9	10.3	9.9	15.6	17.8	9.0	28.9	17.5	24
11	18.0	12.9	12.1	9.9	11.2	16.5	15.8	18.9	25.7	20.4	21.5	24.6	26.3	28.5	30.5	36.8	28.5	21.9	29.6	33.1	34.0	17.4	15.4	15.1	9.9	36.8	21.8	24
12	10.5	11.9	12.7	12.1	12.9	P	P	P	P	P	P	P	P	X	50.4	61.1	61.8	62.2	40.1	46.2	39.9	37.1	28.8	32.0	10.5	62.2	34.6	15
13	30.3	35.3	32.4	25.0	33.8	31.8	28.5	29.2	39.9	44.5	53.0	50.6	51.3	44.3	44.9	40.5	36.6	32.9	30.5	25.0	16.5	11.0	12.6	14.7	11.0	53.0	33.1	24
14	13.0	16.9	13.8	15.8	17.8	14.0	19.5	20.7	20.2	22.4	23.7	26.3	30.5	34.4	36.6	37.0	27.8	25.0	25.0	28.0	27.6	24.6	28.7	33.1	13.0	37.0	24.3	24
15	27.5	32.9	31.8	31.1	41.4	36.1	26.5	23.9	30.2	30.0	35.7	39.9	44.0	44.7	43.8	32.2	33.8	29.8	19.9	26.3	23.7	17.4	19.5	22.1	17.4	44.7	31.0	24
16	19.1	14.0	15.1	13.0	9.7	8.6	10.4	10.1	8.8	10.3	12.1	12.9	10.6	12.7	16.0	15.1	10.6	13.2	14.3	16.0	18.9	16.2	19.9	20.4	8.6	20.4	13.7	24
17	20.6	19.7	21.3	30.2	21.5	25.4	32.9	28.9	33.1	P	P	30.5	33.6	39.0	47.3	42.5	34.4	29.4	27.4	23.2	19.0	13.6	16.7	14.5	13.6	47.3	27.5	22
18	18.0	23.0	22.6	24.3	27.8	37.7	35.3	34.2	31.8	26.3	32.0	54.7	56.0	57.1	54.3	51.9	41.4	39.0	24.5	21.5	19.1	21.5	22.2	20.9	18.0	57.1	33.2	24
19	22.6	21.5	22.0	22.1	20.2	19.1	17.4	16.5	21.0	25.8	28.0	34.0	34.6	30.7	21.5	15.8	13.6	16.1	38.6	39.2	29.8	22.6	29.4	35.1	13.6	39.2	24.9	24
20	39.9	40.1	36.1	43.8	50.1	52.8	54.7	47.5	25.0	28.9	27.6	23.7	23.4	45.8	50.6	51.3	46.0	56.1	34.6	19.7	36.4	32.9	24.3	19.7	19.7	56.1	38.0	24
21	17.6	36.8	31.6	28.9	23.5	18.4	39.4	25.8	27.6	34.2	36.6	43.6	54.3	48.6	49.7	42.3	30.7	24.1	17.1	18.0	18.0	22.6	17.6	15.1	15.1	54.3	30.1	24
22	11.4	11.0	9.7	10.1	12.3	12.3	14.0	24.8	25.2	28.8	26.1	27.6	28.9	34.4	37.0	40.3	34.6	25.2	26.1	28.3	27.2	29.7	25.5	26.9	9.7	40.3	24.1	24
23	31.1	29.4	28.5	28.9	28.9	27.6	22.8	29.4	32.0	30.1	19.5	20.8	28.0	21.7	20.2	22.8	20.6	26.9	24.3	21.9	21.5	19.9	17.8	14.9	14.9	32.0	24.6	24
24	16.5	13.6	12.9	18.6	14.0	14.5	17.3	14.0	10.5	11.0	16.0	15.0	12.5	15.4	12.8	9.7	11.0	7.1	7.3	12.1	12.3	8.4	9.9	13.4	7.1	18.6	12.7	24
25	15.6	12.5	13.6	16.2	13.2	14.5	17.8	16.1	17.1	14.5	16.2	23.3	17.7	15.8	18.8	20.6	10.3	10.2	9.9	8.6	7.7	9.0	10.1	9.0	7.7	23.3	14.1	24
26	25.0	20.8	34.6	35.3	25.6	15.4	16.3	15.8	20.2	31.3	21.7	24.5	34.8	46.4	17.8	20.2	11.6	9.0	9.4	8.4	11.2	14.5	16.9	18.7	8.4	46.4	21.1	24
27	26.3	21.7	22.8	19.1	17.8	17.8	21.3	19.3	23.7	31.8	50.1	50.4	45.3	35.5	34.8	44.0	41.9	47.1	30.7	47.1	25.0	20.8	15.6	13.9	13.9	50.4	30.1	24
28	16.9	12.5	12.1	11.2	9.5	14.7	17.3	16.0	25.4	29.2	25.9	32.7	31.1	26.3	27.8	38.3	37.6	39.0	23.7	30.3	43.4	37.7	37.7	37.5	9.5	43.4	26.4	24
29	36.2	18.2	14.3	16.2	X	X	X	X	33.8	30.5	29.0	34.8	31.8	30.3	31.1	34.2	31.3	24.6	19.5	22.8	19.3	22.8	21.0	13.8	13.8	36.2	25.8	20
30	8.8	11.0	10.5	9.1	8.8	8.4	8.1	7.7	14.5	16.0	16.0	16.9	18.4	20.6	22.1	17.3	21.0	20.6	16.7	14.7	17.1	14.5	12.5	13.9	7.7	22.1	14.4	24
31	15.1	14.0	12.3	12.3	14.9	14.1	14.0	13.8	15.1	12.9	12.6	11.0	13.4	12.7	13.2	12.8	5.3	9.0	11.6	14.5	17.8	13.8	14.0	9.0	5.3	17.8	12.9	24
HOURLY MAX	39.9	40.1	36.1	43.8	50.1	52.8	54.7	47.5	39.9	44.5	53.0	54.7	56.0	57.1	54.3	61.1	61.8	62.2	40.1	47.1	43.4	38.1	39.7	37.5				


STATUS FLAG CODES

C	- MONTHLY CALIBRATION	Q	- QUALITY ASSURANCE
C1	- REPEAT CALIBRATION	R	- RECOVERY
Y	- MAINTENANCE	X	- MACHINE MALFUNCTION
S	- DAILY ZERO/SPAN CHECK	G	- OUT FOR REPAIR
S1	- REPEAT ZERO/SPAN CHECK	P	- POWER FAILURE

MONTHLY SUMMARY

MAXIMUM INSTANTANEOUS VALUE:	62.2 kph	@ HOUR	17	ON DAY	12
OPERATIONAL TIME:			729	hrs	

WIND SPEED Instantaneous Maximum (WS kph)



***APPENDIX IV
REPORT CERTIFICATION FORM***

Report Certification Form

Alberta Airshed (if applicable)	EPA Approval or Code of Practice Registration # (if applicable)
YES	NA
Company Name (if applicable)	Industrial Operation Name (if applicable)
LAKELAND INDUSTRY & COMMUNITY ASSOCIATION	ST. LINA CONTINUOUS MONITORING STATION
Name of the Representative of the Person Responsible	Position / Title of the Representative of the Person Responsible
Mike Bisaga	Environment Monitoring Program Manager
Is an External Party Certifying the Report?	
<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	
Name of External Person Certifying the Report	Position / Title of External Person Certifying the Report
Wunmi Adekanmbi	Project Team Lead, Customer Service - Air Services
Company Name for External Person Certifying the Report	Identification of Qualifications / Professional Designations of the External Person Certifying the Report
Maxxam Analytics, A Bureau Veritas Group Company	M.Sc., EPt., PMP

Maxxam Analytics is the designated contractor conducting monitoring and reporting activities. I certify that the submitted data has been (a) reviewed and validated as per the AMD Chapter 6: Ambient Data Quality. I certify that the submitted report (b) accurately reflects the monitoring results and reporting timeframe and (c) meets the specified analysis, summarization and reporting requirements as per the AMD Chapter 9: Reporting.



 Signature of the External Person Certifying the Report

28 - Nov - 2018

 Report Issued Date (dd-mon-yyyy)

APPENDIX V
DATA VALIDATION CERTIFICATION FORM



Validation Certificate Form

Client: <u>Lakeland Industry & Community Association</u>	Project #: <u>2833-2018-10-31-C</u>
Site: <u>St. Lina Continuous Monitoring Station</u>	Contact: <u>Mike Bisaga</u>

Level 0 Preliminary Verification	<u>Maram Ghaleb</u>	Date <u>20 - Nov -2018</u>
Level 1 Primary Validation	<u>Maram Ghaleb</u>	Date <u>20 - Nov -2018</u>
Level 2 Final Validation	<u>Maram Ghaleb</u>	Date <u>23 - Nov -2018</u>
Level 3 Independent Data Review	<u>CS-LMBG</u>	Date <u>28 - Nov -2018</u>
Post-Final Validation	<u>NA</u>	Date <u>NA</u>

Notes
The Post-Final Validation step serves to re-evaluate the data that errors or omissions are discovered and/or suspected after the initial submittal of data. This validation is performed on an annual basis.