



Box 8237  
5107W-50th Street  
Bonnyville, AB T9N 2J5  
Phone: (780) 812-2182  
Fax: (780) 812-2186  
Toll Free: 1-877-737-2182  
E-Mail: [lica2@lica.ca](mailto:lica2@lica.ca)  
Website: <http://www.lica.ca>

Alberta Environment  
Environmental Response Centre  
1st fl Twin Atria Building  
4999 - 98 Avenue  
Edmonton, AB T6B 2X3

Fax: 780 427-3178

January 26, 2012

**RE: Particulate Matter less than 90% uptime at St. Lina Monitoring Station - Reference# 255191**

---

The particulate matter analyzer at the St. Lina monitoring station failed during routine calibration & maintenance work. A replacement analyzer has been installed at the station while the main unit is being repaired. A review of the data revealed that some of the concentrations recorded prior to the equipment failure will have to be invalidated, therefore it will likely result in less than 90% coverage for particulate matter during the month of January 2012.

Should you have any questions, please don't hesitate to contact me directly at (780) 266-7068.

Respectfully,

A handwritten signature in blue ink that reads "Michael Bisaga".

Michael Bisaga

Airshed Program Manager  
Lakeland Industry and Community Association

cc (email): Lili Goncalves (Imperial Oil), Monty Moore (Husky), Kathleen Zellweger (Shell), Roxane Bretzlaff (CNRL), Shannon Makinson (Cenovus), Charmaine Code (LICA)